

NISTIR 8261-05

NIST Time and Frequency Bulletin

Kathryn Stephenson, Editor

This publication is available free of charge from:
<https://doi.org/10.6028/NIST.IR.8261-05>

NIST
National Institute of
Standards and Technology
U.S. Department of Commerce

NISTIR 8261-05

NIST Time and Frequency Bulletin

Kathryn Stephenson, Editor
Time and Frequency Division
Physical Measurement Laboratory

This publication is available free of charge from:
<https://doi.org/10.6028/NIST.IR.8261-05>

May 2019



U.S. Department of Commerce
Wilbur L. Ross, Jr., Secretary

National Institute of Standards and Technology
Walter Copan, NIST Director and Under Secretary of Commerce for Standards and Technology

NIST TIME AND FREQUENCY BULLETIN
NIST IR 8261-05

No. 738 May 2019

1. GENERAL BACKGROUND INFORMATION.....	2
2. TIME SCALE INFORMATION	2
3. BROADCAST OUTAGES OVER FIVE MINUTES AND WWVB PHASE PERTURBATIONS	4
4. NOTES ON NIST TIME SCALES AND PRIMARY STANDARDS	4
5. UTC (NIST) – AT1 PARAMETERS	5

This bulletin is published monthly. Address correspondence to:

Kathryn Stephenson, Editor
Time and Frequency Division
National Institute of Standards and Technology
325 Broadway
Boulder, CO 80305
(303) 497-3295
Email: kathryn.stephenson@nist.gov



U.S. DEPARTMENT OF COMMERCE, Wilbur L. Ross, Jr., Secretary
NATIONAL INSTITUTE OF STANDARDS AND TECHNOLOGY, Walter Copan, NIST Director
and Under Secretary of Commerce for Standards and Technology

1. GENERAL BACKGROUND INFORMATION

ACRONYMS AND ABBREVIATIONS USED IN THIS BULLETIN

ACTS	- Automated Computer Time Service		
BIPM	- Bureau International des Poids et Mesures		
GPS	- Global Positioning System		
IERS	- International Earth Rotation Service		
MC	- Master Clock		
MJD	- Modified Julian Date		
NIST	- National Institute of Standards and Technology	ns	- nanosecond
SI	- International System of Units	µs	- microsecond
TA	- Atomic Time	ms	- millisecond
TAI	- International Atomic Time	s	- second
USNO	- United States Naval Observatory	min	- minute
UT1	- Universal Time (Astronomical)		
UTC	- Coordinated Universal Time		

2. TIME SCALE INFORMATION

The values listed below are based on data from the IERS, the USNO, and NIST. The UTC(USNO,MC) - UTC(NIST) values are averaged measurements from all available common-view GPS satellites (see bibliography on page 5). UTC - UTC(NIST) data are on page 3.

0000 HOURS COORDINATED UNIVERSAL TIME			
Apr 2019	MJD	UT1-UTC(NIST) (±5 ms)	UTC(USNO,MC) - UTC(NIST) (±20 ns)
4	58577	-122.47 ms	-3 ns
11	58584	-130.73 ms	-3.1 ns
18	58591	-138.07 ms	-2.2 ns
25	58598	-144.49 ms	-2.7 ns

The master clock pulses used by the WWV, WWVH, and WWVB time-code transmissions are referenced to the UTC (NIST) time scale. Occasionally, 1 s is added to the UTC time scale. This second is called a leap second. Its purpose is to keep the UTC time scale within ±0.9 s of the UT1 astronomical time scale, which changes slightly due to variations in the Earth's period of rotation.

NOTE: No leap second will be added at the end of June 2019.

Positive leap seconds, beginning at 23 h 59 min 60 s UTC and ending at 0 h 0 min 0 s UTC, were inserted in the UTC time scale on 30 June 1972, 1981-1983, 1985, 1992-1994, 1997, 2012, 2015 and on 31 December 1972-1979, 1987, 1989, 1990,1995, 1998, 2005, 2008, 2016.

The use of leap seconds ensures that UT1 - UTC will always be held within ±0.9 s. The current value of UT1 - UTC is called the DUT1 correction. DUT1 corrections are broadcast by WWV, WWVH, WWVB, and ACTS and are printed below. These corrections may be added to the received UTC time signals in order to obtain UT1.

DUT1 = UT1 - UTC =	-0.1 s beginning 0000 UTC 17 January 2019 +0.0 s beginning 0000 UTC 21 September 2018 +0.1 s beginning 0000 UTC 15 March 2018 +0.2 s beginning 0000 UTC 30 November 2017 +0.3 s beginning 0000 UTC 29 June 2017 +0.4 s beginning 0000 UTC 30 March 2017 +0.5 s beginning 0000 UTC 26 January 2017 +0.6 s beginning 0000 UTC 01 January 2017 -0.4 s beginning 0000 UTC 17 November 2016 -0.3 s beginning 0000 UTC 01 September 2016
--------------------	---

The difference between UTC(NIST) and UTC has been within ± 100 ns since July 6, 1994. The table below shows values of UTC - UTC(NIST) as supplied by the BIPM in their *Circular T* publication for the most recent 310-day period in which data are available. Data are given at ten-day intervals. Five-day interval data are available in *Circular T*.

0000 Hours Coordinated Universal Time		
DATE	MJD	UTC-UTC(NIST), ns
Apr. 26, 2019	58599	0.4
Apr. 16, 2019	58589	0.6
Apr. 6, 2019	58579	-0.3
Mar. 27, 2019	58569	-1.9
Mar. 17, 2019	58559	-1.0
Mar. 7, 2019	58549	0.7
Feb. 25, 2019	58539	-2.4
Feb. 15, 2019	58529	-5.3
Feb. 5, 2019	58519	-3.3
Jan. 26, 2019	58509	-2.1
Jan. 16, 2019	58499	-2.4
Jan. 6, 2019	58489	-2.1
Dec. 27, 2018	58479	0.3
Dec. 17, 2018	58469	3.0
Dec. 7, 2018	58459	3.7
Nov. 27, 2018	58449	2.0
Nov. 17, 2018	58439	1.1
Nov. 7, 2018	58429	0.7
Oct 28, 2018	58419	0.1
Oct. 18, 2018	58409	-0.6
Oct. 8, 2018	58399	-1.3
Sep. 28, 2018	58389	0.5
Sep. 18, 2018	58379	2.5
Sep. 8, 2018	57369	2.2
Aug. 29, 2018	58359	0.1
Aug. 19, 2018	58349	0.2
Aug. 9, 2018	58339	-0.5
Jul. 30, 2018	58329	-1.0
Jul. 20, 2018	58319	-0.4
Jul. 10, 2018	58309	0.3
Jun. 30, 2018	58299	1.5
Jun. 20, 2018	58289	1.7
Jun. 10, 2018	58379	2.4

3. BROADCAST OUTAGES OVER FIVE MINUTES AND WWVB PHASE PERTURBATIONS

OUTAGES OF 5 MINUTES OR MORE						PHASE PERTURBATIONS 2 ms			
Station	Apr 2019	MJD	Began UTC	Ended UTC	Freq.	Apr 2019	MJD	Began UTC	End UTC
WWVB	04-04-19	58577	0647	0702	60 kHz	None			
WWV	None					None			
WWVH	None					None			

4. NOTES ON NIST TIME SCALES AND PRIMARY STANDARDS

Primary frequency standards developed and operated by NIST are used to provide accuracy (rate) input to the BIPM and to provide the best possible realization of the SI second. NIST-F1 and NIST F-2, cold-atom cesium fountain frequency standards, have served as the U.S. primary standards of time and frequency since 1999. The uncertainty of NIST-F2 is currently about 1 part in 10^{16} .

The AT1 scale is run in real-time by use of data from an ensemble of cesium standards and hydrogen masers. It is a free-running scale whose frequency is maintained as nearly constant as possible by choosing the optimum weight for each clock that contributes to the computation.

UTC(NIST) is generated as an offset from our real-time scale AT1. It is steered in frequency towards UTC by use of data published by the BIPM in its *Circular T*. Changes in the steering frequency will be made, if necessary, at 0000 UTC on the first day of the month, and occasionally at mid-month. A change in frequency is limited to no more than ± 2 ns/day. The frequency of UTC(NIST) is kept as stable as possible at other times.

UTC is generated at the BIPM by use of a post-processed time-scale algorithm and is not available in real-time. The parameters that we use to generate UTC(NIST) in real-time are therefore based on an extrapolation of UTC from the most recent available data.

References:

Allan, D.W.; Hellwig, H.; and Glaze, D.J., "An accuracy algorithm for an atomic time scale," *Metrologia*, Vol.11, No.3, pp. 133-138 (1975).

Allan, D.W.; Davis, D.D.; Weiss, M.A.; Clements, A.; Guinot, B.; Granveaud, M.; Dorenwendt, K.; Fischer, B.; Hetzel, P.; Aoki, S.; Fujimoto, M.; Charron, L.; and Ashby, N., "Accuracy of international time and frequency comparisons via global positioning system satellites in common-view," *IEEE Transactions on Instrumentation and Measurement*, Vol. IM-34, pp.118-125 (1985).

Heavner, T.P.; Jefferts, S.R.; Donley, E.A.; Shirley, J.H. and Parker, T.E., "NIST F1; recent improvements and accuracy evaluations," *Metrologia*, Vol. 42, pp. 411-422 (2005).

Jefferts, S.R.; Shirley, J.; Parker, T.E.; Heavner, T.P.; Meekhof, D.M.; Nelson, C., Levi, F.; Costanza, G.; De Marchi, A.; Drullinger, R.; Hollberg, L.; Lee, W.D.; and Walls, F.L., "Accuracy evaluation of NIST-F1," *Metrologia*, Vol. 39, pp. 321-336 (2002).

Lewandowski, W. and Thomas, C., "GPS Time transfer," *Proceedings of the IEEE*, Vol. 79, pp. 991-1000 (1991).

Parker, T.E.; Jefferts, S.R.; Heavner, T.P.; and Donley, E.A., "Operation of the NIST-F1 caesium fountain primary frequency standard with a maser ensemble, including the impact of frequency transfer noise," *Metrologia*, Vol. 42, pp. 423-430 (2005).

Weiss, M.A.; Allan, D.W., "An NBS Calibration Procedure for Providing Time and Frequency at a Remote Site by Weighting and Smoothing of GPS Common View Data," *IEEE Transactions on Instrumentation and Measurement*, Vol. IM-36, pp. 572-578 (1987).

5. UTC(NIST) – AT1 PARAMETERS

The table below lists parameters that are used to define UTC(NIST) with respect to our real-time scale AT1. To find the value of UTC(NIST) - AT1 at any time T (expressed as a Modified Julian Date, including a fraction if needed), the appropriate equation to use is the one for which the desired T is greater than or equal to the entry in the T_0 column and less than the entry in the last column. The values of x_{ls} , x , and y for that month are then used in the equation below to find the desired value. The parameters x and y represent the offsets in time and frequency, respectively, between UTC(NIST) and AT1; the parameter x_{ls} is the number of leap seconds applied to both UTC(NIST) and UTC, as specified by the IERS. Leap seconds are not applied to AT1.

$UTC(NIST) - AT1 = x_{ls} + x + y(T - T_0)$					
Month	x_{ls} (s)	x (ns)	y (ns/d)	T_0 (MJD)	Valid until 0000 on: (MJD)
May 19	-37	-471321.7	-36.70*	58604	58619
Apr 19	-37	-471107.5	-36.70	58598	58604*
Apr 19	-37	-470592.3	-36.80	58584	58598†
Apr 19	-37	-470222.3	-37.00	58574	58584†
Mar 19	-37	-470074.3	-37.00	58570	58574
Mar 19	-37	-469817.05	-36.75	58563	58570†
Mar 19	-37	-469560.5	-36.65	58556	58563†
Mar 19	-37	-469304.65	-36.55	58549	58556†
Mar 19	-37	-469083.25	-36.90	58543	58549†
Feb 19	-37	-468788.05	-36.90	58535	58543
Feb 19	-37	-468528.35	-37.10	58528	58535†
Feb 19	-37	-468053.85	-36.50*	58515	58528†
Jan 19	-37	-468017.35	-36.50	58514	58515
Jan 19	-37	-467244.55	-36.80	58493	58514†
Jan 19	-37	-466915.15	-36.60	58484	58493†
Dec 18	-37	-466219.75	-36.60	58465	58484
Dec 18	-37	-465962.15	-36.80	58458	58465†
Dec 18	-37	-465962.15	37.00	58453	58458†
Nov 18	-37	-464667.15	-37.00	58423	58453
Oct 18	-37	-463888.05	-37.10	58402	58423
Oct 18	-37	-463520.05	-36.80	58392	58402†
Sep 18	-37	-462857.65	-36.80	58374	58392
Sep 18	-37	-462411.25	-37.20	58362	58374†
Aug 18	-37	-462336.85	-37.20	58360	58362
Aug 18	-37	-461260.95	-37.10	58331	58360†
Jul 18	-37	-461038.35	-37.10	58325	58331
Jul 18	-37	-460113.35	-37.00	58300	58325†
Jun 18	-37	-459003.35	-37.00	58270	58300
May 18	-37	-458966.35	-37.00	58269	58270
May 18	-37	-457921.95	-37.30	58241	58269†
May 18	-37	-457848.35	-36.80	58239	58241†

† Rate change in mid-month

*Provisional value