

NATL INST OF STAND & TECH R.I.C.



A11104 023092

NBS  
PUBLICATIONS



TECHNICAL ASSOCIATION OF THE  
PULP AND PAPER INDUSTRY

COLLABORATIVE REFERENCE PROGRAM  
FOR PAPER

REPORT NO. 58G



U.S. DEPARTMENT OF COMMERCE  
National Bureau of Standards

QC  
100  
.U56  
79-1374  
1979  
C.2

NBS COLLABORATIVE REFERENCE PROGRAMS

TAPPI Paper and Board (6 times per year)

Bursting strength	Smoothness
Tearing strength	Surface pick strength
Tensile breaking strength	K & N ink absorption
Elongation to break	pH
Tensile energy absorption	Opacity
Folding endurance	Blue reflectance (brightness)
Stiffness	Specular gloss, 75°
Air resistance	Thickness
Grammage	Concora (flat crush)
	Ring crush

FKBG-API Containerboard (48 times per year)

Mullen burst of linerboard  
Concora test of medium

MCCA Color and Appearance (4 times per year)

Gloss at 60°  
Color and color difference

CTS Rubber (4 times per year)

Tensile strength, ultimate elongation and tensile stress  
Hardness  
Mooney viscosity  
Vulcanization properties

CTS Thermal Insulation Materials (2 times per year)

19 test methods for thermal insulation materials covering:  
thermal properties; strength properties; dimensions, stability,  
and density properties; fire properties; and properties of  
vapor barriers

ASTM Cement (2 times per year)

Chemical (11 chemical components)  
Physical (8 characteristics)

AASHTO Bituminous

Asphalt cement (2 times per year)  
Cutbacks (once a year)

NBS Collaborative Reference Programs  
A05 Technology Building  
National Bureau of Standards  
Washington, DC 20234

FEB 27 1980

not acc - lens  
QC 1000  
USA  
79-1179  
1079  
c. 2

TECHNICAL ASSOCIATION OF THE  
PULP AND PAPER INDUSTRY

COLLABORATIVE REFERENCE PROGRAM  
FOR PAPER

Report No. 58G

R. G. Powell  
TAPPI-NBS Research Associate  
Collaborative Testing Services, Inc.

J. Horlick  
Office of Testing Laboratory Evaluation Technology  
Office of Engineering Standards  
National Engineering Laboratory

NBSIR 79-1374

U. S. DEPARTMENT OF COMMERCE  
National Bureau of Standards



## INTRODUCTION

Reports 58S and 58G comprise the fourth set of reports for the 78-79 program year. Participants in tests which involve strength properties of paper will receive only the S report; those in tests which measure other properties will receive only the G report.

Notes and comments to individual laboratories and "Best Values" applicable to a particular method are given following Table 1 for each method. See page 4 of this report for an explanation of "Best Values". Please do not confuse these Best Values with provisional values included with the samples to detect serious discrepancies at the time of test. NBS results, identified as L502 in the optical tests, are included on some of the tables.

If there are any questions on the notes, the analyses, or the reports in general, contact Robert G. Powell or Jeffrey Horlick on 301/921-2946.



Jeffrey Horlick, Administrator  
NBS-TAPPI Collaborative Reference Program  
Office of Testing Laboratory Evaluation Technology

June 8, 1979

## TAPPI-NBS COLLABORATIVE REFERENCE PROGRAM

### BACKGROUND AND PURPOSE

In 1969, the National Bureau of Standards and the Technical Association of the Pulp and Paper Industry established a collaborative reference program to provide a participating laboratory with a means to check periodically the level and uniformity of its testing in comparison with that of other laboratories.

The interchange of paper and board products and of the raw materials for these products requires agreement among raw material suppliers, paper and board producers, converters, distributors, retailers, commercial testing laboratories, user organizations and the ultimate consumer as to the meaning of test results, an agreement that cannot be achieved without accurate and precise testing. This program is designed to help assure agreement.

### HOW THE PROGRAM WORKS

Participants Select the Tests in which they wish to participate. This choice is made on joining the program, but additional tests may be added at any time. Also new participants may enter the program at any time.

Test Samples are Distributed Bimonthly; i.e. every 2 months.

Provisional Values are Provided with the Samples for one or both of the test levels, depending on method. The provisional values permit serious discrepancies to be detected without delay. (It is left to the discretion of the laboratory supervisor as to whether these values should be known to the operator.)

Each Participant Tests the Samples, following instructions provided for each test method. The full check on a single instrument should normally take no more than 30 minutes. The test results are then sent to NBS for analysis. The participant is also asked to report other information relevant to an accurate analysis, such as test conditions and the instruments used.

Industry Means, Best Values and Other Statistics are developed from the data by NBS. The best values are estimates based on a careful examination of all data, both current and past, with special attention to results obtained by the National Bureau of Standards and other recognized reference laboratories in this and other countries.

A Quick Report is Prepared for each participating laboratory reporting data on time. This report shows the industry mean values, and the deviations of the laboratory's results from these values for each test method.

A Longer Summary Report, Showing the Data from all Participants, is also prepared. In the summary report, of which this report is an example, each laboratory is identified by a code number so that the information is maintained on a confidential basis. However, instruments are identified by type so participants can compare their results with those obtained on similar instruments of different manufacture. This report includes test averages, best values and standard deviations for individual participants and for the group as a whole. A participant should be able to readily determine the level and variability of his results in comparison with those of the other laboratories.

Repeatability and Reproducibility Statements such as Contained in ASTM, TAPPI and ISO Standards are included at the end of the report. Participants can check their performance level against the precision statement given in the test method or specification.



TABLE OF CONTENTS

Analyses In This Report

PAGE  
-----

i	Introduction
ii	Description of Program
iv	Metric Conversion Table
1	Key to Tables and Graphs
5	40-1 Air Resistance, Gurley Oil type
8	40-2 Air Resistance, Sheffield type
11	41-1 Air Resistance, Gurley Mercury type
13	44-1 Smoothness, Parker Printsurf
14	45-1 Smoothness, Sheffield type
19	45-2 Smoothness, Bekk type
20	47-1 Smoothness, Bendtsen type
21	56-1 K & N Ink Absorption
22	57-1 pH, Cold Extraction
23	57-2 pH, Hot Extraction
24	60-1 Opacity, White (89%) Backing
29	60-2 Opacity, Paper Backing, B & L type
30	60-3 Opacity, Paper Backing, Elrepho type
32	65-1 Blue Reflectance (Brightness), Directional
35	65-2 Blue Reflectance, Diffuse, Elrepho (Gloss Trap)
37	65-3 Blue Reflectance, Diffuse, Elrepho (No Gloss Trap)
39	75-1 Specular Gloss, 75 degree
42	90-1 Thickness (Caliper)
47	95-1 Grammage (Basis Weight)
49	Summary

Analyses In The S Report

10-1	Bursting Strength - Up to 45 psi
10-2	Bursting Strength - Up to 45 psi, Air Clamps
11-1	Bursting Strength - Up to 100 psi
15-1	Tearing Strength - Deep Cutout
17-1	Tearing Strength - No Cutout
19-1	Tensile Breaking Strength - Packaging Papers
20-1	Tensile Breaking Strength - Printing Papers, CRE
20-2	Tensile Breaking Strength - Printing Papers, Pendulum
25-1	Tensile Energy Absorption - Packaging Papers
26-1	Tensile Energy Absorption - Printing Papers
28-1	Elongation to Break - Packaging Papers
29-1	Elongation to Break - Printing Papers
30-1	Folding Endurance, MIT type
30-2	Folding Endurance, MIT type, log (base 10)
35-1	Stiffness, Gurley
36-1	Stiffness, Taber
49-1	Surface Pick Strength, IGT
50-1	Surface Pick Strength, Wax
91-1	Concora (Flat Crush)
96-1	Ring Crush

TABLE OF CONVERSION FACTORS TO METRIC (SI) UNITS

<u>Physical Quantity</u>	<u>To Convert From</u>	<u>To</u>	<u>Multiply by</u>
Bursting strength	psi	kPa	6.895
	kg/cm <sup>2</sup>	kPa	98.07
	bar	kPa	100.00
Tearing strength	g	mN	9.807
Tensile strength	lb/in.	kN/m	.1751
	lb/0.5 in.	kN/m	.3502
	lb/15 mm	kN/m	.2965
	kg/15 mm	kN/m	.6538
	kg/25 mm	kN/m	.3923
	kg/mm	kN/m	9.807
Tensile energy absorption	ft-lb/ft <sup>2</sup>	J/m <sup>2</sup>	14.59
	in.-lb/in. <sup>2</sup>	J/m <sup>2</sup>	175.1
	kg-m/m <sup>2</sup>	J/m <sup>2</sup>	9.807
Bending stiffness	g·cm	μN·m	98.07
Flat-crush strength (Concora)	lb	N	4.448
Ring-crush (TAPPI)	lb	N	4.448
	(ISO)	lb/6.00 in.	kN/m
Thickness	mil	μm	25.40



## KEY TO TABLES AND GRAPHS

- MEAN - The average of individual TEST DETERMINATIONS. The number of TEST DETERMINATIONS in the mean is given in the upper right corner of the first table (TEST D.) and again at the bottom of this table.
- GRAND MEAN - (GR. MEAN) The average of the individual laboratory MEANS, excluding laboratories flagged (see column F) with an X, #, or +. The GRAND MEAN is given in US customary units and, where applicable, in SI metric units.
- SD OF MEANS - (SD MEANS) The standard deviation of the laboratory MEANS about the GRAND MEAN; an index of the among-laboratory precision.
- DEV - The deviation or difference of the laboratory MEAN from the GRAND MEAN.
- N. DEV - The normal deviate or ratio of the DEV to the SD OF MEANS; an indication of the degree of divergence of the laboratory MEAN from the GRAND MEAN. A N. DEV of more than 2 or less than -2 may indicate that the participant is not following the procedure considered standard for this analysis.
- SDR - The standard deviation of repeated measurements; that is, of individual test determinations about their MEAN.
- AVERAGE SDR - The average of the individual laboratory SDR's; an index of the within-laboratory precision of repeated measurements.
- R. SDR - The relative standard deviation of repeated measurements; that is, the ratio of the SDR to the AVERAGE SDR; an indication of the ability of a participant to repeat his measurements relative to the average ability. The greater the number of TEST DETERMINATIONS the closer the R. SDR should be to unity. If R. SDR is outside the limits given below, the participant may not be following the procedure considered standard for this analysis:

<u>No. of test Determinations</u>	<u>Lower limit for R. SDR</u>	<u>Upper limit for R. SDR</u>
3	0.09	2.58
5	0.27	2.06
8	0.40	1.77
10	0.46	1.67
15	0.56	1.53
20	0.61	1.45
25	0.65	1.39

- VAR - Code for instrument type or variation in condition, see second table.
- F - Flag, with following meaning:
- + - Excluded from grand means because VAR non-standard for this analysis.
  - # - Excluded because data were not understood or because of a non-coded variation reported by the laboratory. (See NOTES following Table 1 for each method).
  - M - Excluded because data for one sample are missing.
  - X - Excluded because plotted point would fall outside of the 99% error ellipse, (see below for explanation of Graph).
  - \* - Included in grand means but plotted point falls outside of the 95% error ellipse. The participants should take this as a warning to reexamine his testing procedure.
  - S - Included in grand mean but only after omission of one or more 'wild' values; that is, test determinations more than 3 times AVERAGE SDR from the laboratory's MEAN. Not more than 20% of the test determination may be excluded in this manner without rejecting the laboratory.
  - O - Included in grand mean and inside 95% error ellipse.
- COORDINATES - Distances along major and minor axes of error ellipse. If special additive or concurrent model of the measuring process applies to this method, the distance along the minor axis represents the random error within a laboratory while that along the major axis also includes a systematic laboratory component of error.

95% ELLIPSE - Lengths of the major and minor axes of the ellipse and the angle that the major axis makes with the horizontal axis.

AVG R. SDR - Average of the R. SDR for the two samples; an indication of the laboratory's precision of repeated measurements.

Graph - For each laboratory the MEAN for the second sample is plotted against the MEAN for the first sample, with each point representing a laboratory. The horizontal and vertical lines are the GRAND MEANS. The dashed line is drawn at 45°. The solid sloping line, which may or may not lie close to the 45° line, is along the major axis of the error ellipse. The ellipse is drawn so that, on the average, it will include 95% of the points representing the laboratories.

Plotted symbols are as explained above (under F), except that an 'S' is plotted as an 'O'. A participant whose plotted point falls outside of the ellipse should carefully reexamine the testing procedure he is following.

The graph is plotted with an ellipse when there are 20 or more laboratories in the analysis. When there are 10 through 19 laboratories in the analysis the graph is plotted but the ellipse is omitted. When there are fewer than 10 laboratories retained in the analysis the graph is not plotted.

The International System of Units (SI) is used on the plots wherever possible to aid participants in familiarizing themselves with SI. Grand means in SI units are given at the top of the plot, and supplementary scales in SI units are drawn along the axes allowing the reader to compare means and variability in common units and SI units for the same data.

Summary - In addition to several quantities already defined  
(At end of above, the summary shows the following values for  
report) each test method:

REPL CRP - The number of replicate test determinations used  
in this Collaborative Reference Program.

REPL TAPPI - The number of replicate test determinations in a  
test result required by the applicable TAPPI  
Standard or assumed here if there is no TAPPI  
Standard. This quantity is needed in the compu-  
tation of TAPPI repeatability and reproducibility  
from the SD OF MEANS and the AVER SDR. See TAPPI  
Standard T1206 for definitions and computations.

REPEAT - TAPPI repeatability, a measure of the within-  
laboratory precision of a test result.

REPROD - TAPPI reproducibility, a measure of the between-  
laboratory precision of a test result.

Best values - Given at the end of Table 1 for each method  
for which sufficient information is available.  
These best values are estimates based on a  
careful examination of all data, both current  
and past, with special attention to results  
obtained by the National Bureau of Standards  
and other recognized reference laboratories  
in this and other countries. All participants  
using equipment that is standard for the  
analysis should be able to achieve results  
within the plus-minus (+) limits, when these  
are shown along with the best values.



TAPPI COLLABORATIVE REFERENCE PROGRAM  
ANALYSIS T40-1 TABLE 1  
AIR RESISTANCE, GURLEY UNITS (SECONDS/100 CC)  
TAPPI STANDARD T460 GS-75, AIR RESISTANCE OF PAPER

LAB CODE	SAMPLE K21 MEAN	PRINTING 103 GRAMS PER SQUARE METER				SAMPLE K43 MEAN	PRINTING 60 GRAMS PER SQUARE METER				TEST D <sub>0</sub> = 10		
		DEV	N <sub>0</sub> DEV	SDR	R <sub>0</sub> SDR		DEV	N <sub>0</sub> DEV	SDR	R <sub>0</sub> SDR	VAR	F	LAB
L100	56.6	5.4	1.78	7.3	1.70	17.0	.6	.73	1.3	1.11	40D	Ø	L100
L106	40.5	-10.7	-3.56	1.2	.28	12.6	-3.8	-4.52	.7	.59	40D	X	L106
L107	52.6	1.3	.44	5.8	1.35	17.1	.7	.89	.6	.49	40D	Ø	L107
L121	50.9	-0.4	-.12	4.2	.97	15.7	-0.7	-.80	.7	.58	40D	Ø	L121
L122	50.0	-1.2	-.40	3.2	.74	16.0	-.4	-.43	1.7	1.45	40D	Ø	L122
L123	52.1	.8	.27	4.1	.97	15.7	-0.7	-.80	1.4	1.22	40D	Ø	L123
L124G	48.8	-2.4	-.81	2.9	.67	16.0	-.4	-.44	.8	.66	40D	Ø	L124G
L125	50.0	-1.3	-.42	3.3	.77	16.6	.2	.27	1.3	1.07	40D	Ø	L125
L128	47.6	-3.6	-1.21	4.4	1.03	16.4	.0	.03	1.5	1.27	40D	Ø	L128
L141	52.0	.8	.25	3.7	.87	17.1	.8	.90	1.0	.83	40D	Ø	L141
L148	53.6	2.4	.78	4.3	1.01	16.6	.2	.24	1.4	1.19	40D	Ø	L148
L153	48.4	-2.8	-.94	5.9	1.38	16.9	.5	.62	1.4	1.17	40D	Ø	L153
L158	46.9	-4.3	-1.44	4.9	1.14	15.8	-0.6	-.65	1.3	1.11	40D	Ø	L158
L159	50.9	-0.3	-.10	1.7	.39	15.9	-0.5	-.62	.8	.71	40D	Ø	L159
L163	52.3	1.0	.34	5.8	1.36	16.6	.2	.29	.8	.69	40D	Ø	L163
L166	51.6	.3	.11	4.1	.96	16.5	.2	.18	1.3	1.07	40D	Ø	L166
L174	51.4	.2	.05	4.8	1.12	16.6	.2	.22	1.2	1.02	40D	Ø	L174
L182G	50.9	-0.3	-.11	7.1	1.66	16.4	-0.0	-.02	1.6	1.38	40D	Ø	L182G
L183	56.7	5.5	1.81	4.9	1.14	17.3	1.0	1.15	1.2	1.04	40D	Ø	L183
L190C	52.7	1.5	.49	5.6	1.32	16.4	.0	.01	1.7	1.43	40D	Ø	L190C
L190R	54.4	3.2	1.05	6.7	1.57	16.1	-0.3	-.31	1.2	1.05	40D	Ø	L190R
L212	47.4	-3.8	-1.27	5.8	1.36	14.5	-1.9	-2.23	.8	.66	40D	Ø	L212
L219	47.4	-3.8	-1.26	2.9	.68	15.9	-0.5	-.60	1.0	.85	40D	Ø	L219
L223	53.6	2.4	.78	4.1	.96	17.6	1.2	1.43	1.3	1.10	40D	Ø	L223
L230G	59.7	8.5	2.81	5.5	1.28	18.3	1.9	2.30	1.3	1.06	40D	*	L230G
L232	43.0	-8.3	-2.75	7.8	1.83	14.6	-1.8	-2.13	1.1	.93	40D	*	L232
L236	52.7	1.5	.48	5.4	1.26	17.0	.6	.76	1.0	.83	40D	Ø	L236
L238A	51.7	.5	.15	2.5	.58	17.0	.6	.76	1.0	.86	40D	Ø	L238A
L241	47.8	-3.4	-1.14	3.8	.88	15.7	-0.7	-.81	.9	.80	40D	Ø	L241
L242	50.0	-1.2	-.41	1.5	.35	15.3	-1.1	-1.32	1.2	1.05	40D	Ø	L242
L243G	49.6	-1.6	-.54	4.8	1.13	15.7	-0.7	-.85	.8	.70	40D	Ø	L243G
L254	52.2	1.0	.32	3.8	.89	17.2	.8	.88	.9	.78	40D	Ø	L254
L259	47.9	-3.3	-1.11	3.1	.73	14.8	-1.5	-1.83	1.1	.92	40D	Ø	L259
L261	51.8	.6	.19	5.6	1.30	16.1	-0.3	-.31	1.4	1.19	40D	Ø	L261
L262G	50.4	-.8	-.27	1.8	.43	16.6	.2	.22	1.2	1.05	40D	Ø	L262G
L265	48.5	-2.8	-.91	4.7	1.11	15.9	-0.5	-.62	1.1	.92	40D	Ø	L265
L278	51.0	-0.3	-.09	3.8	.88	16.2	-0.2	-.19	1.4	1.16	40D	Ø	L278
L285	52.4	1.2	.40	5.4	1.26	16.7	.3	.36	1.2	1.02	40D	Ø	L285
L301	50.1	-1.2	-.38	4.3	1.00	16.3	-0.0	-.06	1.3	1.08	40D	Ø	L301
L308	50.9	-0.3	-.11	3.7	.86	16.6	.2	.29	1.3	1.14	40D	Ø	L308
L320	46.8	-4.4	-1.47	2.7	.63	15.6	-0.8	-.93	1.8	1.55	40D	Ø	L320
L321	46.0	-5.2	-1.74	3.6	.85	12.8	-3.6	-4.28	.6	.53	40D	X	L321
L324	50.4	-0.9	-.28	4.2	.98	16.4	.1	.06	1.0	.81	40D	Ø	L324
L326	54.3	3.1	1.03	7.1	1.65	18.4	2.0	2.40	1.0	.88	40D	*	L326
L328	50.8	-0.4	-.14	4.8	1.12	16.7	.3	.35	1.4	1.20	40D	Ø	L328
L339	40.4	-10.8	-3.60	5.0	1.18	12.1	-4.3	-5.10	1.0	.86	40D	X	L339
L344	47.4	-3.8	-1.27	4.8	1.12	15.6	-0.8	-.98	.9	.76	40D	Ø	L344
L376	53.6	2.4	.78	5.0	1.17	17.0	.6	.70	1.4	1.17	40D	Ø	L376
L380	50.6	-0.6	-.21	2.1	.48	15.6	-0.8	-.93	1.3	1.07	40D	Ø	L380
L388	59.5	8.2	2.73	5.0	1.16	17.2	.9	1.03	1.6	1.39	40D	*	L388
L394	54.4	3.2	1.05	5.1	1.19	17.6	1.2	1.43	1.3	1.13	40D	Ø	L394
L396M	48.5	-2.7	-.89	5.5	1.29	46.5	30.1	35.99	2.3	1.94	40D	#	L396M
L567	53.1	1.9	.62	3.6	.85	16.7	.3	.37	1.6	1.37	40D	Ø	L567
L576	51.5	.2	.07	4.0	.94	15.1	-1.3	-1.52	1.1	.92	40D	Ø	L576
L585	53.7	2.5	.82	5.3	1.23	17.2	.8	.96	1.4	1.18	40D	Ø	L585
L597	50.5	-0.7	-.24	3.1	.72	16.3	-0.1	-.09	1.8	1.55	40D	Ø	L597
L604	51.3	.1	.02	5.1	1.19	16.1	-0.3	-.33	.6	.48	40D	Ø	L604
L616	52.3	1.1	.37	3.3	.77	17.7	1.3	1.54	.3	.24	40D	Ø	L616
L651	49.4	-1.8	-.61	3.8	.88	15.0	-1.4	-1.65	1.1	.89	40D	Ø	L651

GR. MEAN = 51.2 GURLEY UNITS      GR. MEAN = 16.4 GURLEY UNITS      TEST DETERMINATIONS = 10  
SD MEANS = 3.0 GURLEY UNITS      SD OF MEANS = .8 GURLEY UNITS      55 LABS IN GRAND MEAN  
AVERAGE SDR = 4.3 GURLEY UNITS      AVERAGE SDR = 1.2 GURLEY UNITS

L291 50.9 -0.3 -0.11 5.0 1.16 17.6 1.2 1.46 1.1 .91 40U \* L291  
TOTAL NUMBER OF LABORATORIES REPORTING = 60

Best values: K21 51.0 ± 4.1 Gurley units  
K43 16.3 ± 1.2 Gurley units

The following laboratories were omitted from the grand means because of extreme test results: 396M.


TAPPI COLLABORATIVE REFERENCE PROGRAM  
 ANALYSIS T40-1 TABLE 2  
 AIR RESISTANCE, GURLEY UNITS (SECONDS/100 CC)  
 TAPPI STANDARD T460 OS-75, AIR RESISTANCE OF PAPER

LAB CODE	F	MEANS		COORDINATES		AVG R <sub>s</sub> SDR	VAR	PROPERTY---	TEST INSTRUMENT---	CONDITIONS
		K21	K43	MAJOR	MINOR					
L339	X	40.4	12.1	-11.5	-1.9	1.02	40D	AIR RESISTANCE,	GURLEY DENSOMETER	- OIL FLOTATION
L106	X	40.5	12.6	-11.3	-1.5	.43	40D	AIR RESISTANCE,	GURLEY DENSOMETER	- OIL FLOTATION
L232	*	43.0	14.6	-8.5	-.0	1.38	40D	AIR RESISTANCE,	GURLEY DENSOMETER	- OIL FLOTATION
L321	X	46.0	12.8	-5.9	-2.4	.69	40D	AIR RESISTANCE,	GURLEY DENSOMETER	- OIL FLOTATION
L320	Ø	46.8	15.6	-4.5	.2	1.09	40D	AIR RESISTANCE,	GURLEY DENSOMETER	- OIL FLOTATION
L158	Ø	46.9	15.8	-4.4	.3	1.12	40D	AIR RESISTANCE,	GURLEY DENSOMETER	- OIL FLOTATION
L344	Ø	47.4	15.6	-3.9	-.0	.94	40D	AIR RESISTANCE,	GURLEY DENSOMETER	- OIL FLOTATION
L212	Ø	47.4	14.5	-4.1	-1.0	1.01	40D	AIR RESISTANCE,	GURLEY DENSOMETER	- OIL FLOTATION
L219	Ø	47.4	15.9	-3.8	.3	.76	40D	AIR RESISTANCE,	GURLEY DENSOMETER	- OIL FLOTATION
L128	Ø	47.6	16.4	-3.6	.8	1.15	40D	AIR RESISTANCE,	GURLEY DENSOMETER	- OIL FLOTATION
L241	Ø	47.8	15.7	-3.5	.0	.84	40D	AIR RESISTANCE,	GURLEY DENSOMETER	- OIL FLOTATION
L259	Ø	47.9	14.8	-3.6	-.8	.82	40D	AIR RESISTANCE,	GURLEY DENSOMETER	- OIL FLOTATION
L153	Ø	48.4	16.9	-2.7	1.1	1.27	40D	AIR RESISTANCE,	GURLEY DENSOMETER	- OIL FLOTATION
L265	Ø	48.5	15.9	-2.8	.1	1.01	40D	AIR RESISTANCE,	GURLEY DENSOMETER	- OIL FLOTATION
L396H	#	48.5	46.5	3.6	30.0	1.62	40D	AIR RESISTANCE,	GURLEY DENSOMETER	- OIL FLOTATION
L124G	Ø	48.8	16.0	-2.5	.1	.66	40D	AIR RESISTANCE,	GURLEY DENSOMETER	- OIL FLOTATION
L651	Ø	49.4	15.0	-2.1	-1.0	.89	40D	AIR RESISTANCE,	GURLEY DENSOMETER	- OIL FLOTATION
L243G	Ø	49.6	15.7	-1.7	-.4	.91	40D	AIR RESISTANCE,	GURLEY DENSOMETER	- OIL FLOTATION
L125	Ø	50.0	16.6	-1.2	.5	.92	40D	AIR RESISTANCE,	GURLEY DENSOMETER	- OIL FLOTATION
L242	Ø	50.0	15.3	-1.4	-.8	.70	40D	AIR RESISTANCE,	GURLEY DENSOMETER	- OIL FLOTATION
L122	Ø	50.0	16.0	-1.3	-.1	1.09	40D	AIR RESISTANCE,	GURLEY DENSOMETER	- OIL FLOTATION
L301	Ø	50.1	16.3	-1.1	.2	1.04	40D	AIR RESISTANCE,	GURLEY DENSOMETER	- OIL FLOTATION
L324	Ø	50.4	16.4	-.8	.2	.90	40D	AIR RESISTANCE,	GURLEY DENSOMETER	- OIL FLOTATION
L262G	Ø	50.4	16.6	-.8	.3	.74	40D	AIR RESISTANCE,	GURLEY DENSOMETER	- OIL FLOTATION
L597	Ø	50.5	16.3	-.7	.1	1.13	40D	AIR RESISTANCE,	GURLEY DENSOMETER	- OIL FLOTATION
L380	Ø	50.6	15.6	-.8	-.6	.78	40D	AIR RESISTANCE,	GURLEY DENSOMETER	- OIL FLOTATION
L328	Ø	50.8	16.7	-.4	.4	1.16	40D	AIR RESISTANCE,	GURLEY DENSOMETER	- OIL FLOTATION
L121	Ø	50.9	15.7	-.5	-.6	.77	40D	AIR RESISTANCE,	GURLEY DENSOMETER	- OIL FLOTATION
L291	*	50.9	17.6	-.1	1.3	1.04	40U	AIR RESISTANCE,	SHEFFIELD IN GURLEY UNITS	
L308	Ø	50.9	16.6	-.3	.3	1.00	40D	AIR RESISTANCE,	GURLEY DENSOMETER	- OIL FLOTATION
L182G	Ø	50.9	16.4	-.3	.1	1.52	40D	AIR RESISTANCE,	GURLEY DENSOMETER	- OIL FLOTATION
L159	Ø	50.9	15.9	-.4	-.4	.55	40D	AIR RESISTANCE,	GURLEY DENSOMETER	- OIL FLOTATION
L278	Ø	51.0	16.2	-.3	-.1	1.02	40D	AIR RESISTANCE,	GURLEY DENSOMETER	- OIL FLOTATION
L604	Ø	51.3	16.1	.0	-.3	.84	40D	AIR RESISTANCE,	GURLEY DENSOMETER	- OIL FLOTATION
L174	Ø	51.4	16.6	.2	.1	1.07	40D	AIR RESISTANCE,	GURLEY DENSOMETER	- OIL FLOTATION
L576	Ø	51.5	15.1	-.0	-1.3	.93	40D	AIR RESISTANCE,	GURLEY DENSOMETER	- OIL FLOTATION
L166	Ø	51.6	16.5	.3	.1	1.02	40D	AIR RESISTANCE,	GURLEY DENSOMETER	- OIL FLOTATION
L238A	Ø	51.7	17.0	.6	.5	.72	40D	AIR RESISTANCE,	GURLEY DENSOMETER	- OIL FLOTATION
L261	Ø	51.8	16.1	.5	-.4	1.24	40D	AIR RESISTANCE,	GURLEY DENSOMETER	- OIL FLOTATION
L141	Ø	52.0	17.1	.9	.6	.85	40D	AIR RESISTANCE,	GURLEY DENSOMETER	- OIL FLOTATION
L123	Ø	52.1	15.7	.7	-.8	1.09	40D	AIR RESISTANCE,	GURLEY DENSOMETER	- OIL FLOTATION
L254	Ø	52.2	17.2	1.1	.6	.83	40D	AIR RESISTANCE,	GURLEY DENSOMETER	- OIL FLOTATION
L163	Ø	52.3	16.6	1.1	.0	1.02	40D	AIR RESISTANCE,	GURLEY DENSOMETER	- OIL FLOTATION
L616	Ø	52.3	17.7	1.4	1.0	.50	40D	AIR RESISTANCE,	GURLEY DENSOMETER	- OIL FLOTATION
L285	Ø	52.4	16.7	1.2	.0	1.14	40D	AIR RESISTANCE,	GURLEY DENSOMETER	- OIL FLOTATION
I107	Ø	52.6	17.1	1.4	.5	.92	40D	AIR RESISTANCE,	GURLEY DENSOMETER	- OIL FLOTATION
L236	Ø	52.7	17.0	1.6	.3	1.05	40D	AIR RESISTANCE,	GURLEY DENSOMETER	- OIL FLOTATION
L190C	Ø	52.7	16.4	1.4	-.3	1.37	40D	AIR RESISTANCE,	GURLEY DENSOMETER	- OIL FLOTATION
L567	Ø	53.1	16.7	1.9	-.1	1.11	40D	AIR RESISTANCE,	GURLEY DENSOMETER	- OIL FLOTATION
L376	Ø	53.6	17.0	2.4	.1	1.17	40D	AIR RESISTANCE,	GURLEY DENSOMETER	- OIL FLOTATION
L148	Ø	53.6	16.6	2.4	-.3	1.10	40D	AIR RESISTANCE,	GURLEY DENSOMETER	- OIL FLOTATION
L223	Ø	53.6	17.6	2.6	.7	1.03	40D	AIR RESISTANCE,	GURLEY DENSOMETER	- OIL FLOTATION
L585	Ø	53.7	17.2	2.6	.3	1.20	40D	AIR RESISTANCE,	GURLEY DENSOMETER	- OIL FLOTATION
L326	*	54.3	18.4	3.4	1.3	1.27	40D	AIR RESISTANCE,	GURLEY DENSOMETER	- OIL FLOTATION
L394	Ø	54.4	17.6	3.3	.5	1.16	40D	AIR RESISTANCE,	GURLEY DENSOMETER	- OIL FLOTATION
L190R	Ø	54.4	16.1	3.1	-.9	1.31	40D	AIR RESISTANCE,	GURLEY DENSOMETER	- OIL FLOTATION
L100	Ø	56.6	17.0	5.4	-.5	1.40	40D	AIR RESISTANCE,	GURLEY DENSOMETER	- OIL FLOTATION
L183	Ø	56.7	17.3	5.5	-.2	1.09	40D	AIR RESISTANCE,	GURLEY DENSOMETER	- OIL FLOTATION
L388	*	59.5	17.2	8.2	-.9	1.27	40D	AIR RESISTANCE,	GURLEY DENSOMETER	- OIL FLOTATION
L230G	*	59.7	18.3	8.7	.1	1.17	40D	AIR RESISTANCE,	GURLEY DENSOMETER	- OIL FLOTATION
OMEANS:		51.2	16.4			1.00				
		95% ELLIPSE:		7.8	1.4					WITH GAMMA = 11 DEGREES



# AIR RESISTANCE, GURLEY

SAMPLE K21 = 51.2 GURLEY UNITS    SAMPLE K43 = 16.4 GURLEY UNITS



AIR RESISTANCE, SHEFFIELD UNITS (CC/MIN) FOR 0.442 SQ. IN (3/4 IN. DIA) ORIFICE  
SHEFFIELD TESTER IS STANDARD FOR THIS ANALYSIS


LAB CODE	SAMPLE K21 103 GRAMS PER SQUARE METER					SAMPLE K43 60 GRAMS PER SQUARE METER					TEST D. = 10		
	MEAN	DEV	N.DEV	SDR	R. SDR	MEAN	DEV	N.DEV	SDR	R. SDR	VAR	F	LAB
L114	79.4	8.3	1.40	4.9	.82	177.7	10.1	1.10	13.0	1.48	40S	θ	L114
L121	89.0	17.9	3.03	5.8	.97	172.9	5.3	.58	9.8	1.11	40S	*	L121
L124S	70.2	-0.9	-0.15	6.8	1.15	168.0	.4	.05	7.4	.84	40S	θ	L124S
L132	69.0	-2.1	-0.35	8.2	1.38	168.4	.8	.09	11.8	1.35	40S	θ	L132
L148	79.9	8.8	1.49	4.5	.75	174.6	7.0	.77	8.9	1.02	40S	θ	L148
L150	67.2	-3.6	-0.66	4.4	.74	194.5	26.9	2.93	23.9	2.72	40S	*	L150
L157	67.8	-3.3	-0.56	5.8	.97	174.6	7.0	.77	10.9	1.25	40S	θ	L157
L158	73.0	1.6	.32	5.9	.99	162.0	-5.6	-0.60	11.4	1.29	40S	θ	L158
L173B	67.5	-3.6	-0.61	4.9	.82	177.0	9.4	1.03	4.8	.55	40S	θ	L173B
L190C	78.8	7.7	1.30	6.4	1.08	162.6	-5.0	-0.54	8.8	1.00	40S	θ	L190C
L213	75.7	4.6	.78	5.7	.96	165.2	-2.4	-0.26	8.1	.92	40S	θ	L213
L223	62.1	-9.0	-1.52	3.7	.62	163.0	-4.6	-0.50	10.2	1.17	40S	θ	L223
L228	66.4	-4.7	-0.79	6.4	1.08	165.9	-1.7	-0.18	5.8	.66	40S	θ	L228
L230S	69.6	-1.5	-0.25	7.0	1.17	159.9	-7.7	-0.83	4.0	.46	40S	θ	L230S
L233	55.4	-15.7	-2.66	7.2	1.21	158.0	-9.6	-1.04	8.2	.93	40S	*	L233
L241	79.7	8.6	1.46	5.7	.95	176.4	8.8	.96	7.6	.87	40S	θ	L241
L249	65.9	-5.2	-0.88	6.4	1.08	164.4	-3.2	-0.34	10.0	1.14	40S	θ	L249
L255	92.1	21.0	3.55	6.2	1.03	188.7	21.1	2.30	5.8	.66	40S	X	L255
L257A	77.6	6.5	1.10	5.0	.85	157.5	-10.1	-1.09	8.7	1.00	40S	θ	L257A
L257B	68.6	-2.5	-0.42	5.6	.94	169.0	1.4	.16	13.2	1.51	40S	θ	L257B
L257C	78.5	7.4	1.25	8.0	1.35	161.7	-5.9	-0.64	8.0	.92	40S	θ	L257C
L260	71.2	.1	.02	6.7	1.12	173.4	5.8	.64	13.9	1.59	40S	θ	L260
L262S	71.4	.3	.05	4.1	.68	157.2	-10.4	-1.13	8.4	.96	40S	θ	L262S
L288	73.3	2.2	.37	4.1	.69	187.3	19.7	2.15	13.6	1.55	40S	θ	L288
L301	68.2	-2.6	-0.49	8.3	1.40	168.6	1.0	.11	9.4	1.07	40S	θ	L301
L305	70.0	-1.1	-0.19	5.8	.97	171.0	3.4	.37	6.5	.74	40S	θ	L305
L318	70.2	-0.9	-0.15	6.4	1.07	158.5	-9.1	-0.98	5.6	.64	40S	θ	L318
L352	78.5	7.4	1.25	2.3	.39	168.6	1.0	.11	8.7	.99	40S	θ	L352
L354	73.6	2.5	.42	6.8	1.15	168.8	1.2	.14	9.3	1.06	40S	θ	L354
L360	69.1	-2.0	-0.34	2.8	.48	157.2	-10.4	-1.13	6.5	.74	40S	θ	L360
L370	67.3	-3.8	-0.64	5.5	.92	156.3	-11.3	-1.22	8.2	.93	40S	θ	L370
L372	69.5	-1.6	-0.27	9.9	1.66	170.4	2.8	.31	6.5	.74	40S	θ	L372
L390	66.5	-4.6	-0.78	5.8	.97	161.0	-6.6	-0.71	8.4	.96	40S	θ	L390
L562	71.5	.4	.07	6.3	1.06	181.8	14.2	1.55	9.1	1.03	40S	θ	L562
L575	71.1	.0	.00	4.4	.74	168.2	.6	.07	12.3	1.40	40S	θ	L575
L585	71.5	.4	.07	8.2	1.38	167.5	-0.1	-0.01	11.8	1.35	40S	θ	L585
L597	70.6	-0.5	-0.08	6.9	1.16	158.9	-8.7	-0.94	6.0	.68	40S	θ	L597
L600	71.0	-0.1	-0.02	7.1	1.20	160.2	-7.4	-0.80	5.5	.63	40S	θ	L600
L626	64.4	-6.7	-1.13	8.8	1.47	146.3	-21.3	-2.31	6.7	.77	40S	θ	L626
L684	64.5	-6.6	-1.12	3.7	.62	176.7	9.1	.99	8.9	1.02	40S	θ	L684
L687	69.1	-2.0	-0.34	5.5	.92	171.1	3.5	.39	6.5	.74	40S	θ	L687
GR. MEAN =	71.1	SHEFF. UNITS				GRAND MEAN =	167.6	SHEFF. UNITS			TEST DETERMINATIONS = 10		
SD MEANS =	5.9	SHEFF. UNITS				SD OF MEANS =	9.2	SHEFF. UNITS			40 LABS IN GRAND MEANS		
		AVERAGE SDR =				5.9	SHEFF. UNITS				8.8 SHEFF. UNITS		
L182B	246.5	175.4	29.67	18.3	3.07	762.5	594.9	64.67	35.8	4.09	40B	*	L182B
L243B	280.2	209.1	35.37	24.5	4.12	821.4	653.6	71.07	52.2	5.95	40B	*	L243B
L312	80.0	8.6	1.51	3.8	.63	167.5	-0.1	-0.01	6.5	.74	40T	*	L312
L587	74.0	2.9	.49	5.7	.95	169.5	1.9	.21	8.3	.95	40T	*	L587
TOTAL NUMBER OF LABORATORIES REPORTING =	45												
Best values:	K21	71	± 8 Sheffield units										
	K43	167	± 14 Sheffield units										

AIR RESISTANCE, SHEFFIELD UNITS (CC/MIN) FOR 0.442 SQ. IN (3/4 IN. DIA) ORIFICE  
SHEFFIELD TESTER IS STANDARD FOR THIS ANALYSIS

LAB CODE	F	MEANS		COORDINATES		AVG		PROPERTY---	TEST INSTRUMENT---	CONDITIONS
		K21	K43	MAJOR	MINOR	R.SDR	VAR			
L233	*	55.4	158.0	-12.6	13.3	1.07	40S	AIR RESISTANCE,	SHEFFIELD (3/4 INCH DIAMETER ORIFICE)	
L223	Ø	62.1	163.0	-6.3	7.8	.89	40S	AIR RESISTANCE,	SHEFFIELD (3/4 INCH DIAMETER ORIFICE)	
L626	Ø	64.4	146.3	-22.2	2.1	1.12	40S	AIR RESISTANCE,	SHEFFIELD (3/4 INCH DIAMETER ORIFICE)	
L684	Ø	64.5	176.7	7.6	8.4	.82	40S	AIR RESISTANCE,	SHEFFIELD (3/4 INCH DIAMETER ORIFICE)	
L249	Ø	65.9	164.4	-4.2	4.4	1.11	40S	AIR RESISTANCE,	SHEFFIELD (3/4 INCH DIAMETER ORIFICE)	
L228	Ø	66.4	165.9	-2.6	4.2	.87	40S	AIR RESISTANCE,	SHEFFIELD (3/4 INCH DIAMETER ORIFICE)	
L390	Ø	66.5	161.0	-7.4	3.1	.97	40S	AIR RESISTANCE,	SHEFFIELD (3/4 INCH DIAMETER ORIFICE)	
L150	*	67.2	154.5	25.5	5.5	1.73	40S	AIR RESISTANCE,	SHEFFIELD (3/4 INCH DIAMETER ORIFICE)	
L370	Ø	67.3	156.3	-11.8	1.3	.92	40S	AIR RESISTANCE,	SHEFFIELD (3/4 INCH DIAMETER ORIFICE)	
L173B	Ø	67.5	177.0	8.5	5.5	.68	40S	AIR RESISTANCE,	SHEFFIELD (3/4 INCH DIAMETER ORIFICE)	
L157	Ø	67.8	174.6	6.2	4.7	1.11	40S	AIR RESISTANCE,	SHEFFIELD (3/4 INCH DIAMETER ORIFICE)	
L301	Ø	68.2	168.6	.4	3.0	1.23	40S	AIR RESISTANCE,	SHEFFIELD (3/4 INCH DIAMETER ORIFICE)	
L257B	Ø	68.6	169.0	.9	2.7	1.23	40S	AIR RESISTANCE,	SHEFFIELD (3/4 INCH DIAMETER ORIFICE)	
L132	Ø	69.0	168.4	.4	2.2	1.37	40S	AIR RESISTANCE,	SHEFFIELD (3/4 INCH DIAMETER ORIFICE)	
L687	Ø	69.1	171.1	3.0	2.7	.83	40S	AIR RESISTANCE,	SHEFFIELD (3/4 INCH DIAMETER ORIFICE)	
L360	Ø	69.1	157.2	-10.5	-.2	.61	40S	AIR RESISTANCE,	SHEFFIELD (3/4 INCH DIAMETER ORIFICE)	
L372	Ø	69.5	170.4	2.4	2.2	1.20	40S	AIR RESISTANCE,	SHEFFIELD (3/4 INCH DIAMETER ORIFICE)	
L230S	Ø	69.6	159.9	-7.8	-.1	.82	40S	AIR RESISTANCE,	SHEFFIELD (3/4 INCH DIAMETER ORIFICE)	
L305	Ø	70.0	171.0	3.1	1.8	.85	40S	AIR RESISTANCE,	SHEFFIELD (3/4 INCH DIAMETER ORIFICE)	
L124S	Ø	70.2	168.0	.2	1.0	1.00	40S	AIR RESISTANCE,	SHEFFIELD (3/4 INCH DIAMETER ORIFICE)	
L318	Ø	70.2	158.5	-9.0	-1.0	.86	40S	AIR RESISTANCE,	SHEFFIELD (3/4 INCH DIAMETER ORIFICE)	
L597	Ø	70.6	158.9	-8.6	-1.3	.92	40S	AIR RESISTANCE,	SHEFFIELD (3/4 INCH DIAMETER ORIFICE)	
L600	Ø	71.0	160.2	-7.2	-1.5	.91	40S	AIR RESISTANCE,	SHEFFIELD (3/4 INCH DIAMETER ORIFICE)	
L575	Ø	71.1	168.2	.6	.1	1.07	40S	AIR RESISTANCE,	SHEFFIELD (3/4 INCH DIAMETER ORIFICE)	
L260	Ø	71.2	173.4	5.7	1.1	1.35	40S	AIR RESISTANCE,	SHEFFIELD (3/4 INCH DIAMETER ORIFICE)	
L262S	Ø	71.4	157.2	-10.1	-2.5	.82	40S	AIR RESISTANCE,	SHEFFIELD (3/4 INCH DIAMETER ORIFICE)	
L585	Ø	71.5	167.5	.0	-.4	1.36	40S	AIR RESISTANCE,	SHEFFIELD (3/4 INCH DIAMETER ORIFICE)	
L562	Ø	71.5	181.8	14.0	2.6	1.04	40S	AIR RESISTANCE,	SHEFFIELD (3/4 INCH DIAMETER ORIFICE)	
L158	Ø	73.0	162.0	-5.0	-3.0	1.14	40S	AIR RESISTANCE,	SHEFFIELD (3/4 INCH DIAMETER ORIFICE)	
L288	Ø	73.3	187.3	19.8	2.0	1.12	40S	AIR RESISTANCE,	SHEFFIELD (3/4 INCH DIAMETER ORIFICE)	
L354	Ø	73.6	168.8	1.7	-2.2	1.10	40S	AIR RESISTANCE,	SHEFFIELD (3/4 INCH DIAMETER ORIFICE)	
L587	*	74.0	169.5	2.5	-2.4	.95	40T	AIR RESISTANCE,	SHEFFIELD (3 INCH DIAMETER ORIFICE)	
L213	Ø	75.7	165.2	-1.3	-5.0	.94	40S	AIR RESISTANCE,	SHEFFIELD (3/4 INCH DIAMETER ORIFICE)	
L257A	Ø	77.6	157.5	-8.5	-8.5	.92	40S	AIR RESISTANCE,	SHEFFIELD (3/4 INCH DIAMETER ORIFICE)	
L352	Ø	78.5	168.6	2.6	-7.0	.69	40S	AIR RESISTANCE,	SHEFFIELD (3/4 INCH DIAMETER ORIFICE)	
L257C	Ø	78.5	161.7	-4.2	-8.5	1.13	40S	AIR RESISTANCE,	SHEFFIELD (3/4 INCH DIAMETER ORIFICE)	
L190C	Ø	78.8	162.6	-3.2	-8.6	1.04	40S	AIR RESISTANCE,	SHEFFIELD (3/4 INCH DIAMETER ORIFICE)	
L114	Ø	79.4	177.7	11.7	-6.0	1.15	40S	AIR RESISTANCE,	SHEFFIELD (3/4 INCH DIAMETER ORIFICE)	
L241	Ø	79.7	176.4	10.5	-6.6	.91	40S	AIR RESISTANCE,	SHEFFIELD (3/4 INCH DIAMETER ORIFICE)	
L148	Ø	79.9	174.6	8.7	-7.1	.88	40S	AIR RESISTANCE,	SHEFFIELD (3/4 INCH DIAMETER ORIFICE)	
L312	*	80.0	167.5	1.8	-8.7	.68	40T	AIR RESISTANCE,	SHEFFIELD (3 INCH DIAMETER ORIFICE)	
L121	*	89.0	172.9	9.0	-16.4	1.04	40S	AIR RESISTANCE,	SHEFFIELD (3/4 INCH DIAMETER ORIFICE)	
L255	X	92.1	188.7	25.1	-16.1	.85	40S	AIR RESISTANCE,	SHEFFIELD (3/4 INCH DIAMETER ORIFICE)	
L182B	*	246.5	762.5	618.5	-46.6	3.58	40B	AIR RESISTANCE,	BENDTSEN, WG 150	
L243B	*	280.2	821.4	683.2	-67.2	5.04	40B	AIR RESISTANCE,	BENDTSEN, WG 150	
GMEANS:		71.1	167.6			1.00				
		95% ELLIPSE:		24.1	14.7			WITH GAMMA = 77 DEGREES		

# AIR RESISTANCE, SHEFFIELD

SAMPLE K21 = 71.    SHEFF. UNITS    SAMPLE K43 = 168.    SHEFF. UNITS



AIR RESISTANCE, HIGH RANGE, GURLEY MERCURY FLOTATION  
DIRECT READING, SEC/10 CC, MERCURY DENSITY

LAB CODE	SAMPLE E68		RELEASE 95 GRAMS PER SQUARE METER			SAMPLE E65		BACKING 98 GRAMS PER SQUARE METER			TEST D. = 10		
	MEAN	DEV	N. DEV	SDR	R. SDR	MEAN	DEV	N. DEV	SDR	R. SDR	VAR	F	LAB
L122	622.	-29.	-.47	144.	1.69	657.	-73.	-1.17	202.	1.37	41G	Ø	L122
L128	620.	-31.	-.50	56.	.66	780.	50.	.80	86.	.58	41G	Ø	L128
L134	691.	40.	.65	81.	.95	716.	-14.	-.23	177.	1.20	41G	Ø	L134
L166M	603.	-48.	-.77	100.	1.18	781.	51.	.81	154.	1.04	41G	Ø	L166M
L230	713.	62.	1.01	57.	.67	697.	-34.	-.54	124.	.84	41G	Ø	L230
L259	625.	-26.	-.42	66.	.77	739.	8.	.13	218.	1.48	41G	Ø	L259
L312	619.	-32.	-.52	70.	.82	662.	-69.	-1.10	157.	1.06	41G	Ø	L312
L358	536.	-115.	-1.86	86.	1.01	629.	-102.	-1.63	90.	.61	41G	Ø	L358
L557	770.	118.	1.92	142.	1.67	754.	23.	.37	116.	.78	41G	Ø	L557
L558	697.	46.	.74	78.	.92	838.	107.	1.71	177.	1.20	41G	Ø	L558
L574	684.	33.	.53	51.	.60	722.	-9.	-.14	153.	1.03	41G	Ø	L574
L576	632.	-19.	-.31	91.	1.07	793.	63.	1.00	119.	.81	41G	Ø	L576

GR. MEAN = 651. SEC/10 CC GRAND MEAN = 731. SEC/10 CC TEST DETERMINATIONS = 10  
 SD MEANS = 62. SEC/10 CC SD OF MEANS = 63. SEC/10 CC 12 LABS IN GRAND MEANS  
 AVERAGE SDR = 85. SEC/10 CC AVERAGE SDR = 148. SEC/10 CC  
 TOTAL NUMBER OF LABORATORIES REPORTING = 12

Best values: E68 650 seconds per 10cc,  
 E65 730 mercury density  
 (direct reading)

The values reported here are the time in seconds required for the displacement of 10 ml of air through an area of 1.0 sq. in. of the specimen. The values are not converted to 100ml of air nor to oil density.


AIR RESISTANCE, HIGH RANGE, GURLEY MERCURY FLOTATION  
DIRECT READING, SEC/10 CC, MERCURY DENSITY

LAB CODE	F	MEANS		COORDINATES		AVG R. SDR VAR	PROPERTY---TEST INSTRUMENT---CONDITIONS
		E68	E65	MAJOR	MINOR		
L358	Ø	536.	629.	-153.	12.	.81 41G	AIR RESISTANCE, HIGH RANGE, GURLEY MERCURY FLOTATION
L166M	Ø	603.	781.	3.	70.	1.11 41G	AIR RESISTANCE, HIGH RANGE, GURLEY MERCURY FLOTATION
L312	Ø	619.	662.	-72.	-25.	.94 41G	AIR RESISTANCE, HIGH RANGE, GURLEY MERCURY FLOTATION
L128	Ø	620.	780.	14.	57.	.62 41G	AIR RESISTANCE, HIGH RANGE, GURLEY MERCURY FLOTATION
L122	Ø	622.	657.	-73.	-30.	1.53 41G	AIR RESISTANCE, HIGH RANGE, GURLEY MERCURY FLOTATION
L259	Ø	625.	739.	-13.	24.	1.13 41G	AIR RESISTANCE, HIGH RANGE, GURLEY MERCURY FLOTATION
L576	Ø	632.	793.	32.	57.	.94 41G	AIR RESISTANCE, HIGH RANGE, GURLEY MERCURY FLOTATION
L574	Ø	684.	722.	17.	-30.	.82 41G	AIR RESISTANCE, HIGH RANGE, GURLEY MERCURY FLOTATION
L134	Ø	691.	716.	18.	-39.	1.08 41G	AIR RESISTANCE, HIGH RANGE, GURLEY MERCURY FLOTATION
L558	Ø	697.	838.	109.	42.	1.06 41G	AIR RESISTANCE, HIGH RANGE, GURLEY MERCURY FLOTATION
L230	Ø	713.	697.	19.	-68.	.75 41G	AIR RESISTANCE, HIGH RANGE, GURLEY MERCURY FLOTATION
L557	Ø	770.	754.	99.	-69.	1.22 41G	AIR RESISTANCE, HIGH RANGE, GURLEY MERCURY FLOTATION
GMEANS:		651.	731.			1.00	
		95% ELLIPSE:	218.	149.		WITH GAMMA = 45 DEGREES	

# AIR RESISTANCE, GURLEY HG FLOTATION

SAMPLE E68 = 651. SEC/10 CC

SAMPLE E65 = 731. SEC/10 CC





LAB CODE	SAMPLE K45 MEAN	PRINTING 60 GEAMS PER SQUARE METER				SAMPLE A83 MEAN	WAVE ENVELOPE PAPER 75 GEAMS PER SQUARE METER				TEST D.° 10		
		DEV	N. DEV	SDR	R. SDR		DEV	N. DEV	SDR	R. SDR	VAR	F	LAB
L122	6.19	.72	1.66	.11	1.11	5.43	.60	1.50	.39	1.61	44P	Ø	L122
L136	5.29	-.18	-.42	.17	1.71	4.36	-.48	-1.20	.22	.91	44P	Ø	L136
L182	5.49	.02	.04	.03	.32	4.76	-.07	-.19	.16	.65	44P	Ø	L182
L183	5.04	-.43	-1.00	.05	.53	4.54	-.30	-.75	.16	.65	44P	Ø	L183
L223	5.46	-.01	-.03	.10	1.06	4.90	.06	.15	.34	1.38	44P	Ø	L223
L288	6.00	.53	1.21	.09	.97	5.45	.61	1.54	.35	1.44	44P	Ø	L288
L317	5.57	.10	.22	.15	1.53	5.00	.16	.41	.24	.97	44P	Ø	L317
L588	4.78	-.69	-1.59	.12	1.26	4.46	-.38	-.95	.18	.75	44P	Ø	L588
L669	5.43	-.04	-.10	.05	.50	4.64	-.20	-.50	.16	.65	44P	Ø	L669

GR. MEAN = 5.47 MICRONS    GRAND MEAN = 4.84 MICRONS    TEST DETERMINATIONS = 10  
SD MEANS = .43 MICRONS    SD OF MEANS = .40 MICRONS    9 LABS IN GRAND MEANS  
AVERAGE SDR = .10 MICRONS    AVERAGE SDR = .24 MICRONS

TOTAL NUMBER OF LABORATORIES REPORTING = 9  
Best values: K45 5.5 microns  
A83 4.7 microns

LAB CODE	F	MEANS		COORDINATES		AVG		PROPERTY---TEST INSTRUMENT---	CONDITIONS
		K45	A83	MAJOR	MINOR	R. SDR	VAR		
L588	Ø	4.78	4.46	-.77	.18	1.01	44P	SMOOTHNESS, PARKER PRINTSURF	
L183	Ø	5.04	4.54	-.52	.07	.59	44P	SMOOTHNESS, PARKER PRINTSURF	
L136	Ø	5.29	4.36	-.46	-.23	1.31	44P	SMOOTHNESS, PARKER PRINTSURF	
L669	Ø	5.43	4.64	-.16	-.12	.57	44P	SMOOTHNESS, PARKER PRINTSURF	
L223	Ø	5.46	4.90	.03	.05	1.22	44P	SMOOTHNESS, PARKER PRINTSURF	
L182	Ø	5.49	4.76	-.04	-.07	.49	44P	SMOOTHNESS, PARKER PRINTSURF	
L317	Ø	5.57	5.00	.18	.05	1.25	44P	SMOOTHNESS, PARKER PRINTSURF	
L288	Ø	6.00	5.45	.80	.10	1.20	44P	SMOOTHNESS, PARKER PRINTSURF	
L122	Ø	6.19	5.43	.93	-.04	1.36	44P	SMOOTHNESS, PARKER PRINTSURF	
GMEANS:		5.47	4.84			1.00			
		95% ELLIPSE:	1.89	.42		WITH GAMMA = 42 DEGREES			

TAPPI COLLABORATIVE REFERENCE PROGRAM  
ANALYSIS T45-1 TABLE 1  
SMOOTHNESS, SHEFFIELD UNITS  
SHEFFIELD TESTER IS STANDARD FOR THIS ANALYSIS

FEBRUARY 1979

LAB CODE	SAMPLE K45 MEAN	PRINTING 60 GRAMS PER SQUARE METER				SAMPLE A83 MEAN	WAVE ENVELOPE PAPER 75 GRAMS PER SQUARE METER				TEST D. # 15		
		DEV	N.DEV	SDR	R.SDR		DEV	N.DEV	SDR	R.SDR	VAR	F	LAB
L100	162.5	2.3	.32	5.0	.55	96.2	-9.2	-1.36	6.9	.65	45S	Ø	L100
L108	155.0	-5.3	-.75	6.2	.68	106.7	1.3	.19	8.7	.81	45S	Ø	L108
L114	164.1	3.8	.54	9.9	1.08	98.6	-6.8	-1.00	11.4	1.06	45S	Ø	L114
L121	156.3	-4.0	-.57	7.4	.80	106.1	.7	.10	11.5	1.08	45S	Ø	L121
L122	161.9	1.6	.23	8.4	.92	112.1	6.7	.98	13.3	1.24	45S	Ø	L122
L123	163.4	3.1	.45	11.2	1.23	101.9	-3.6	-.52	11.4	1.06	45S	Ø	L123
L124	153.6	-6.7	-.95	11.0	1.20	103.1	-2.4	-.35	11.0	1.03	45S	Ø	L124
L125	167.0	6.7	.96	11.5	1.25	103.0	-2.4	-.36	16.8	1.56	45S	Ø	L125
L126	164.5	4.3	.61	15.3	1.68	106.2	.8	.11	14.8	1.38	45S	Ø	L126
L128	160.5	.2	.03	11.7	1.28	102.9	-2.6	-.38	8.7	.82	45S	Ø	L128
L132	169.5	9.3	1.32	13.6	1.48	101.3	-4.2	-.61	14.3	1.33	45S	Ø	L132
L134	150.7	-9.6	-1.37	7.5	.82	97.7	-7.8	-1.14	6.2	.58	45S	Ø	L134
L139S	163.9	3.6	.51	6.7	.74	118.5	13.1	1.92	9.0	.84	45S	Ø	L139S
L148	169.0	8.7	1.24	9.5	1.03	116.3	10.8	1.59	13.2	1.23	45S	Ø	L148
L150	170.9	10.6	1.51	18.6	2.03	105.5	.0	.00	14.9	1.39	45S	Ø	L150
L152	163.2	2.9	.42	5.4	.59	111.7	6.3	.92	11.3	1.06	45S	Ø	L152
L153	173.7	13.4	1.91	7.3	.79	114.7	9.2	1.35	7.6	.71	45S	Ø	L153
L157	168.5	8.2	1.17	11.5	1.26	116.6	11.2	1.64	14.5	1.36	45S	Ø	L157
L158	158.0	-2.3	-.32	7.3	.79	110.7	5.2	.77	10.2	.95	45S	Ø	L158
L159	170.1	9.8	1.39	11.8	1.28	114.9	9.5	1.39	10.7	1.00	45S	Ø	L159
L162	156.7	-3.6	-.51	6.2	.67	107.0	1.6	.23	7.0	.65	45S	Ø	L162
L166	153.9	-6.3	-.90	8.0	.87	109.3	3.9	.57	9.2	.86	45S	Ø	L166
L167	162.3	2.1	.29	4.2	.46	112.7	7.2	1.06	5.0	.46	45S	Ø	L167
L173B	161.7	1.4	.20	7.5	.82	88.7	-16.8	-2.46	6.9	.65	45S	*	L173B
L183S	162.9	2.7	.38	8.8	.96	111.9	6.4	.94	16.7	1.56	45S	Ø	L183S
L190C	159.4	-.9	-.12	7.4	.81	108.3	2.9	.42	9.2	.86	45S	Ø	L190C
L190R	157.1	-3.1	-.45	9.3	1.01	93.1	-12.3	-1.81	12.0	1.11	45S	Ø	L190R
L206	163.6	3.3	.47	6.3	.68	114.4	9.0	1.31	11.6	1.08	45S	Ø	L206
L211	145.4	-10.9	-1.55	7.4	.81	101.1	-4.4	-.64	10.5	.98	45S	Ø	L211
L213	146.3	-13.9	-1.99	10.4	1.14	103.2	-2.2	-.33	10.2	.95	45S	Ø	L213
L219	170.2	9.5	1.41	16.9	1.85	108.3	2.9	.42	10.8	1.01	45S	Ø	L219
L223	157.7	-2.6	-.37	6.3	.69	103.9	-1.5	-.22	14.0	1.31	45S	Ø	L223
L226B	149.9	-10.3	-1.47	9.1	1.00	96.1	-9.3	-1.37	11.2	1.05	45S	Ø	L226B
L228	163.3	3.1	.44	10.5	1.15	108.1	2.7	.39	13.4	1.25	45S	Ø	L228
L230S	162.1	1.5	.26	8.6	.94	101.5	-3.9	-.57	10.6	.99	45S	Ø	L230S
L231	167.1	6.8	.97	9.5	1.04	103.8	-1.6	-.24	13.0	1.21	45S	Ø	L231
L232S	160.3	.1	.01	5.8	.64	92.7	-12.8	-1.87	8.8	.82	45S	Ø	L232S
L233	155.2	-5.1	-.72	11.6	1.26	105.6	.2	.02	10.6	.99	45S	Ø	L233
L237	160.8	.5	.07	5.2	.56	106.0	.6	.08	4.3	.40	45S	Ø	L237
L241	150.1	-10.1	-1.44	8.4	.92	98.9	-6.6	-.96	7.7	.71	45S	Ø	L241
L249	156.1	-4.1	-.59	8.3	.90	106.1	.6	.09	10.3	.96	45S	Ø	L249
L254	166.7	6.4	.91	10.3	1.12	106.5	1.1	.16	8.4	.78	45S	Ø	L254
L255	161.3	1.0	.14	4.3	.47	120.6	15.2	2.22	2.6	.24	45S	Ø	L255
L257A	157.1	-3.1	-.45	10.2	1.11	102.8	-2.6	-.39	7.7	.72	45S	Ø	L257A
L257B	165.3	5.1	.72	9.9	1.08	103.2	-2.2	-.33	9.7	.90	45S	Ø	L257B
L257C	160.0	-.3	-.04	7.1	.78	106.5	1.1	.16	7.7	.72	45S	Ø	L257C
L259	174.2	13.5	1.98	14.8	1.62	116.4	11.0	1.61	10.4	.97	45S	Ø	L259
L260	160.6	.3	.05	6.2	.67	111.4	6.0	.87	16.2	1.51	45S	Ø	L260
L261	156.4	-3.9	-.55	7.3	.80	103.3	-2.2	-.32	8.6	.80	45S	Ø	L261
L262	157.6	-2.7	-.38	8.3	.91	102.8	-2.6	-.39	8.3	.78	45S	Ø	L262
L275	156.7	-3.5	-.50	12.2	1.33	102.9	-2.6	-.38	13.7	1.28	45S	Ø	L275
L278	167.0	6.7	.96	8.9	.97	113.7	8.2	1.21	13.3	1.24	45S	Ø	L278
L281	162.7	2.4	.34	10.2	1.11	109.9	4.5	.66	14.5	1.35	45S	Ø	L281
L285	148.7	-11.6	-1.65	8.3	.91	101.7	-3.8	-.55	12.5	1.16	45S	Ø	L285
L288	169.2	8.9	1.27	7.7	.84	106.9	1.4	.21	11.2	1.04	45S	Ø	L288
L290	152.3	-7.5	-1.13	7.3	.80	98.3	-7.1	-1.04	7.5	.70	45S	Ø	L290
L291S	162.4	2.1	.30	5.8	.63	103.9	-1.6	-.23	11.2	1.04	45S	Ø	L291S
L301	163.0	2.7	.39	7.4	.81	116.3	10.8	1.59	12.4	1.15	45S	Ø	L301
L305	165.2	4.9	.70	8.4	.92	109.1	3.6	.53	7.4	.69	45S	Ø	L305
L308	156.9	-3.4	-.49	7.0	.77	99.5	-6.0	-.88	11.2	1.05	45S	Ø	L308
L312	161.4	1.1	.16	13.4	1.46	115.0	9.6	1.40	9.4	.88	45S	Ø	L312
L317	157.3	-3.0	-.43	13.4	1.46	94.1	-11.4	-1.67	7.5	.70	45S	Ø	L317
L318	160.9	.6	.08	11.0	1.20	117.5	12.1	1.77	16.7	1.55	45S	Ø	L318
L321	140.8	-19.5	-2.77	5.6	.61	98.7	-6.7	-.98	12.4	1.15	45S	*	L321
L323	164.0	3.7	.53	10.4	1.13	103.7	-1.8	-.26	9.5	.89	45S	Ø	L323

TAPPI COLLABORATIVE REFERENCE PROGRAM  
 ANALYSIS T45-1 TABLE 1  
 SMOOTHNESS, SHEFFIELD UNITS  
 SHEFFIELD TESTER IS STANDARD FOR THIS ANALYSIS

LAB CODE	SAMPLE K45 MEAN	PRINTING 60 GRAMS PER SQUARE METER				SAMPLE A83 MEAN	WAVE ENVELOPE PAPER 75 GRAMS PER SQUARE METER				TEST D <sub>0</sub> = 15		
		DEV	N.DEV	SDR	R.SDR		DEV	N.DEV	SDR	R.SDR	VAR	F	LAB
L326	159.7	-0.6	-0.09	5.1	.56	110.8	5.4	.79	9.2	.86	45S	Ø	L326
L328	168.9	8.6	1.22	12.8	1.40	100.7	-4.8	-0.70	10.1	.95	45S	Ø	L328
L349	150.5	-9.8	-1.40	7.6	.83	102.5	-2.9	-0.43	8.7	.81	45S	Ø	L349
L352	159.7	-0.5	-0.08	7.5	.82	110.7	5.3	.78	12.1	1.13	45S	Ø	L352
L360	162.1	1.9	.26	5.6	.62	102.0	-3.4	-0.51	12.8	1.20	45S	Ø	L360
L370	154.3	-5.5	-0.85	4.2	.46	105.4	-0	-0.01	10.5	.98	45S	Ø	L370
L372	159.7	-0.6	-0.09	6.1	.67	99.0	-6.4	-0.95	5.5	.88	45S	Ø	L372
L376	178.5	18.2	2.59	19.1	2.09	107.7	2.2	.33	4.9	.46	45S	*	L376
L380	156.1	-4.1	-0.59	10.8	1.18	96.9	-8.5	-1.25	9.2	.85	45S	Ø	L380
L382	159.9	-0.4	-0.06	10.2	1.11	110.5	5.1	.75	12.6	1.19	45S	Ø	L382
L390	155.7	-4.6	-0.66	9.0	.99	107.3	1.9	.28	19.5	1.82	45S	Ø	L390
L396M	154.7	-5.5	-0.79	9.9	1.09	96.3	-9.2	-1.35	8.0	.74	45S	Ø	L396M
L554	152.3	-8.0	-1.14	9.7	1.06	96.0	-9.4	-1.39	12.0	1.12	45S	*	L554
L567	159.1	-1.1	-0.16	6.8	.74	109.7	4.3	.63	7.1	.66	45S	Ø	L567
L575	169.0	8.7	1.24	11.0	1.20	109.9	4.4	.65	10.3	.96	45S	Ø	L575
L585	149.3	-10.9	-1.56	16.8	1.83	93.3	-12.1	-1.78	11.0	1.02	45S	Ø	L585
L587	169.0	8.7	1.24	9.1	.99	104.3	-1.1	-0.16	5.9	.55	45S	Ø	L587
L597	155.2	-5.1	-0.72	8.2	.90	103.4	-2.0	-0.30	13.6	1.27	45S	Ø	L597
L600	154.5	-5.7	-0.82	12.5	1.36	90.1	-15.4	-2.26	12.9	1.21	45S	Ø	L600
L626	155.5	-4.7	-0.68	7.7	.84	113.6	8.2	1.20	10.9	1.02	45S	Ø	L626
L648	154.7	-5.5	-0.79	9.4	1.03	100.2	-5.2	-0.77	10.8	1.00	45S	Ø	L648
L651	145.9	-14.3	-2.04	7.8	.86	107.6	2.2	.32	13.0	1.21	45S	Ø	L651
L670	175.6	15.3	2.18	13.5	1.48	112.8	7.4	1.08	12.7	1.18	45S	Ø	L670
L679	134.5	-25.7	-3.67	7.7	.84	83.0	-22.4	-3.29	7.9	.74	45S	#	L679
L688	164.3	4.1	.58	11.3	1.24	103.9	-1.6	-0.23	9.8	.91	45S	Ø	L688

GR. MEAN = 160.3 SHEFF. UNITS      GRAND MEAN = 105.4 SHEFF. UNITS      TEST DETERMINATIONS = 15  
 SD MEANS = 7.0 SHEFF. UNITS      SD OF MEANS = 6.8 SHEFF. UNITS      89 LABS IN GRAND MEANS  
 AVERAGE SDR = 9.2 SHEFF. UNITS      AVERAGE SDR = 10.7 SHEFF. UNITS

L174      245.5      85.2      12.13      5.4      .59      216.3      110.9      16.27      6.5      .61      45R      \*      L174  
 TOTAL NUMBER OF LABORATORIES REPORTING = 91

Best values: K45 160 ± 11 Sheffield units  
 A83 105 ± 11 Sheffield units

The following laboratories were omitted from the grand means because of extreme test results: 679.

TAPPI COLLABORATIVE REFERENCE PROGRAM  
ANALYSIS T45-1 TABLE 2  
SMOOTHNESS, SHEFFIELD UNITS  
SHEFFIELD TESTER IS STANDARD FOR THIS ANALYSIS

FEBRUARY 1979

LAB CODE	F	MEANS		COORDINATES		AVG		PROPERTY---	TEST INSTRUMENT---	CONDITIONS
		K45	A83	MAJOR	MINOR	R <sub>s</sub>	SDR VAR			
L679	#	134.5	83.0	-34.1	1.2	.79	45S	SMOOTHNESS,	SHEFFIELD	
L321	*	140.8	98.7	-18.8	8.4	.88	45S	SMOOTHNESS,	SHEFFIELD	
L651	Ø	145.9	107.6	-9.0	11.4	1.03	45S	SMOOTHNESS,	SHEFFIELD	
L213	Ø	146.3	103.2	-11.7	7.9	1.04	45S	SMOOTHNESS,	SHEFFIELD	
L285	Ø	148.7	101.7	-11.1	5.2	1.04	45S	SMOOTHNESS,	SHEFFIELD	
L585	Ø	149.3	93.3	-16.3	-1.4	1.43	45S	SMOOTHNESS,	SHEFFIELD	
L211	Ø	149.4	101.1	-10.9	4.2	.90	45S	SMOOTHNESS,	SHEFFIELD	
L226B	Ø	149.5	96.1	-13.9	.3	1.02	45S	SMOOTHNESS,	SHEFFIELD	
L241	Ø	150.1	98.9	-11.9	2.1	.82	45S	SMOOTHNESS,	SHEFFIELD	
L349	Ø	150.5	102.5	-9.2	4.6	.82	45S	SMOOTHNESS,	SHEFFIELD	
L134	Ø	150.7	97.7	-12.3	.9	.70	45S	SMOOTHNESS,	SHEFFIELD	
L554	Ø	152.3	96.0	-12.3	-1.4	1.09	45S	SMOOTHNESS,	SHEFFIELD	
L290	Ø	152.3	98.3	-10.7	.2	.75	45S	SMOOTHNESS,	SHEFFIELD	
L124	Ø	153.6	103.1	-6.5	2.8	1.11	45S	SMOOTHNESS,	SHEFFIELD	
L166	Ø	153.9	109.3	-2.0	7.2	.87	45S	SMOOTHNESS,	SHEFFIELD	
L370	Ø	154.3	105.4	-4.4	4.0	.72	45S	SMOOTHNESS,	SHEFFIELD	
L600	Ø	154.5	50.1	-14.7	-7.3	1.28	45S	SMOOTHNESS,	SHEFFIELD	
L648	Ø	154.7	100.2	-7.6	-.0	1.02	45S	SMOOTHNESS,	SHEFFIELD	
L396M	Ø	154.7	96.3	-10.3	-2.9	.91	45S	SMOOTHNESS,	SHEFFIELD	
L108	Ø	155.0	106.7	-3.0	4.5	.75	45S	SMOOTHNESS,	SHEFFIELD	
L597	Ø	155.2	103.4	-5.1	2.0	1.08	45S	SMOOTHNESS,	SHEFFIELD	
L233	Ø	155.2	105.6	-3.6	3.6	1.13	45S	SMOOTHNESS,	SHEFFIELD	
L626	Ø	155.5	113.6	2.1	9.2	.93	45S	SMOOTHNESS,	SHEFFIELD	
L390	Ø	155.7	107.3	-2.1	4.5	1.40	45S	SMOOTHNESS,	SHEFFIELD	
L380	Ø	156.1	96.9	-8.8	-3.4	1.02	45S	SMOOTHNESS,	SHEFFIELD	
L249	Ø	156.1	106.1	-2.6	3.3	.93	45S	SMOOTHNESS,	SHEFFIELD	
L121	Ø	156.3	106.1	-2.5	3.2	.94	45S	SMOOTHNESS,	SHEFFIELD	
L261	Ø	156.4	103.3	-4.3	1.1	.80	45S	SMOOTHNESS,	SHEFFIELD	
L162	Ø	156.7	107.0	-1.6	3.6	.66	45S	SMOOTHNESS,	SHEFFIELD	
L275	Ø	156.7	102.9	-4.3	.5	1.30	45S	SMOOTHNESS,	SHEFFIELD	
L308	Ø	156.9	99.5	-6.6	-2.0	.91	45S	SMOOTHNESS,	SHEFFIELD	
L190R	Ø	157.1	93.1	-10.7	-6.8	1.06	45S	SMOOTHNESS,	SHEFFIELD	
L257A	Ø	157.1	102.8	-4.1	.2	.91	45S	SMOOTHNESS,	SHEFFIELD	
L317	Ø	157.3	94.1	-10.0	-6.3	1.08	45S	SMOOTHNESS,	SHEFFIELD	
L262	Ø	157.6	102.8	-3.8	-.1	.84	45S	SMOOTHNESS,	SHEFFIELD	
L223	Ø	157.7	103.5	-2.9	.7	1.00	45S	SMOOTHNESS,	SHEFFIELD	
L158	Ø	158.0	110.7	1.9	5.4	.87	45S	SMOOTHNESS,	SHEFFIELD	
L567	Ø	159.1	109.7	2.1	3.9	.70	45S	SMOOTHNESS,	SHEFFIELD	
L190C	Ø	159.4	108.3	1.3	2.7	.83	45S	SMOOTHNESS,	SHEFFIELD	
L372	Ø	159.7	99.0	-4.8	-4.3	.78	45S	SMOOTHNESS,	SHEFFIELD	
L326	Ø	159.7	110.8	3.2	4.3	.71	45S	SMOOTHNESS,	SHEFFIELD	
L352	Ø	159.7	110.7	3.2	4.2	.97	45S	SMOOTHNESS,	SHEFFIELD	
L382	Ø	159.9	110.5	3.2	4.0	1.15	45S	SMOOTHNESS,	SHEFFIELD	
L257C	Ø	160.0	106.5	.5	1.0	.75	45S	SMOOTHNESS,	SHEFFIELD	
L232S	Ø	160.3	92.7	-8.7	-9.4	.73	45S	SMOOTHNESS,	SHEFFIELD	
L128	Ø	160.5	102.9	-1.6	-2.0	1.05	45S	SMOOTHNESS,	SHEFFIELD	
L260	Ø	160.6	111.4	4.3	4.1	1.09	45S	SMOOTHNESS,	SHEFFIELD	
L237	Ø	160.8	106.0	.8	.0	.48	45S	SMOOTHNESS,	SHEFFIELD	
L318	Ø	160.9	117.5	8.7	8.4	1.38	45S	SMOOTHNESS,	SHEFFIELD	
L255	Ø	161.3	120.6	11.1	10.4	.35	45S	SMOOTHNESS,	SHEFFIELD	
L312	Ø	161.4	115.0	7.3	6.2	1.17	45S	SMOOTHNESS,	SHEFFIELD	
L173B	*	161.7	88.7	-10.4	-13.2	.73	45S	SMOOTHNESS,	SHEFFIELD	
L122	Ø	161.9	112.1	5.7	3.8	1.08	45S	SMOOTHNESS,	SHEFFIELD	
L230S	Ø	162.1	101.5	-1.3	-4.1	.97	45S	SMOOTHNESS,	SHEFFIELD	
L360	Ø	162.1	102.0	-1.0	-3.8	.91	45S	SMOOTHNESS,	SHEFFIELD	
L167	Ø	162.3	112.7	6.4	3.9	.46	45S	SMOOTHNESS,	SHEFFIELD	
L291S	Ø	162.4	103.9	.5	-2.6	.84	45S	SMOOTHNESS,	SHEFFIELD	
L100	Ø	162.5	96.2	-4.7	-8.3	.60	45S	SMOOTHNESS,	SHEFFIELD	
L281	Ø	162.7	109.9	4.8	1.6	1.23	45S	SMOOTHNESS,	SHEFFIELD	
L183S	Ø	162.9	111.9	6.3	2.9	1.26	45S	SMOOTHNESS,	SHEFFIELD	
L301	Ø	163.0	116.3	9.4	6.0	.98	45S	SMOOTHNESS,	SHEFFIELD	
L152	Ø	163.2	111.7	6.4	2.6	.82	45S	SMOOTHNESS,	SHEFFIELD	
L228	Ø	163.3	108.1	4.1	-.1	1.20	45S	SMOOTHNESS,	SHEFFIELD	
L123	Ø	163.4	101.9	-.2	-4.7	1.14	45S	SMOOTHNESS,	SHEFFIELD	
L206	Ø	163.6	114.4	8.5	4.3	.88	45S	SMOOTHNESS,	SHEFFIELD	


TAPPI COLLABORATIVE REFERENCE PROGRAM  
ANALYSIS T45-1 TABLE 2  
SMOOTHNESS, SHEFFIELD UNITS  
SHEFFIELD TESTER IS STANDARD FOR THIS ANALYSIS

LAB CODE	F	MEANS		COORDINATES		AVG		PROPERTY---	TEST INSTRUMENT---	CONDITIONS
		R45	A83	MAJOR	MINOR	R,SDR	VAR			
L139S	Ø	163.9	118.5	11.6	7.1	.79	45S	SMOOTHNESS,	SHEFFIELD	
L323	Ø	164.0	103.7	1.5	-3.8	1.01	45S	SMOOTHNESS,	SHEFFIELD	
L114	Ø	164.1	98.6	-1.9	-7.6	1.07	45S	SMOOTHNESS,	SHEFFIELD	
L688	Ø	164.3	103.9	1.9	-3.9	1.07	45S	SMOOTHNESS,	SHEFFIELD	
L126	Ø	164.5	106.2	3.6	-2.4	1.53	45S	SMOOTHNESS,	SHEFFIELD	
L305	Ø	165.2	109.1	6.1	-.7	.80	45S	SMOOTHNESS,	SHEFFIELD	
L257B	Ø	165.3	103.2	2.2	-5.1	.99	45S	SMOOTHNESS,	SHEFFIELD	
L254	Ø	166.7	106.5	5.4	-3.6	.95	45S	SMOOTHNESS,	SHEFFIELD	
L125	Ø	167.0	103.0	3.2	-6.4	1.41	45S	SMOOTHNESS,	SHEFFIELD	
L278	Ø	167.0	113.7	10.5	1.4	1.11	45S	SMOOTHNESS,	SHEFFIELD	
L231	Ø	167.1	103.8	3.8	-5.8	1.12	45S	SMOOTHNESS,	SHEFFIELD	
L157	Ø	168.5	116.6	13.6	2.6	1.31	45S	SMOOTHNESS,	SHEFFIELD	
L328	Ø	168.9	100.7	3.0	-9.4	1.17	45S	SMOOTHNESS,	SHEFFIELD	
L148	Ø	169.0	116.3	13.8	1.9	1.13	45S	SMOOTHNESS,	SHEFFIELD	
L587	Ø	169.0	104.3	5.6	-6.8	.77	45S	SMOOTHNESS,	SHEFFIELD	
L575	Ø	169.0	109.9	9.4	-2.7	1.08	45S	SMOOTHNESS,	SHEFFIELD	
L288	Ø	169.2	106.9	7.5	-5.1	.94	45S	SMOOTHNESS,	SHEFFIELD	
L132	Ø	169.5	101.3	3.9	-9.4	1.41	45S	SMOOTHNESS,	SHEFFIELD	
L159	Ø	170.1	114.9	13.6	.2	1.14	45S	SMOOTHNESS,	SHEFFIELD	
L219	Ø	170.2	108.3	9.2	-4.7	1.43	45S	SMOOTHNESS,	SHEFFIELD	
L150	Ø	170.9	105.5	7.8	-7.2	1.71	45S	SMOOTHNESS,	SHEFFIELD	
L153	Ø	173.7	114.7	16.1	-2.4	.75	45S	SMOOTHNESS,	SHEFFIELD	
L259	Ø	174.2	116.4	17.7	-1.5	1.29	45S	SMOOTHNESS,	SHEFFIELD	
L670	Ø	175.6	112.8	16.2	-5.1	1.33	45S	SMOOTHNESS,	SHEFFIELD	
L376	*	178.5	107.7	14.8	-10.8	1.27	45S	SMOOTHNESS,	SHEFFIELD	
L174	*	245.5	216.3	138.0	22.8	.60	45R	SMOOTHNESS,	SHEFFIELD,	NON-STANDARD INSTRUMENT
GMEANS:		160.3	105.4			1.00				
95% ELLIPSE:				20.8	13.0	WITH GAMMA = 43 DEGREES				



# SMOOTHNESS, SHEFFIELD

SAMPLE K45 = 160. SHEFF. UNITS    SAMPLE A83 = 105. SHEFF. UNITS





TAPPI SUGGESTED METHOD T479 SU-71, SMOOTHNESS OF PAPER (BEKK METHOD)

LAB CODE	SAMPLE K45 MEAN	PRINTING 60 GRAMS PER SQUARE METER				SAMPLE A83 MEAN	WOVE ENVELOPE PAPER 75 GRAMS PER SQUARE METER				TEST D. = 15		
		DEV	N. DEV	SDR	R. SDR		DEV	N. DEV	SDR	R. SDR	VAR	F	LAB
L139B	31.9	1.8	.45	2.6	1.00	47.8	-7.1	-.94	7.3	.79	45K	Ø	L139B
L162	24.1	-6.1	-1.56	2.3	.90	48.2	-6.7	-.88	8.8	.95	45K	Ø	L162
L182K	26.5	-3.7	-.95	2.3	.90	48.5	-6.4	-.84	8.4	.91	45K	Ø	L182K
L190C	33.2	3.0	.77	2.9	1.10	61.7	6.8	.89	14.0	1.51	45K	Ø	L190C
L230B	27.7	-2.4	-.63	3.6	1.38	52.9	-2.0	-.27	6.4	.69	45K	Ø	L230B
L232B	35.1	4.9	1.26	1.7	.67	67.3	12.4	1.64	8.1	.88	45K	Ø	L232B
L243K	30.9	.8	.19	3.0	1.16	50.4	-4.5	-.59	9.8	1.07	45K	Ø	L243K
L291K	34.7	4.5	1.16	2.9	1.13	64.6	9.7	1.28	9.2	1.00	45K	Ø	L291K
L581	27.5	-2.7	-.70	2.0	.77	52.6	-2.3	-.30	11.0	1.19	45K	Ø	L581

GR. MEAN = 30.2 BEKK SECONDS      GRAND MEAN = 54.9 BEKK SECONDS      TEST DETERMINATIONS = 15  
SD MEANS = 3.9 BEKK SECONDS      SD OF MEANS = 7.6 BEKK SECONDS      9 LABS IN GRAND MEANS  
AVERAGE SDR = 2.6 BEKK SECONDS      AVERAGE SDR = 9.2 BEKK SECONDS

L250M	28.3	-1.8	-.47	2.4	.91	40.5	-14.4	-1.89	6.0	.65	45L	*	L250M
L251	30.2	.0	.00	2.4	.92	52.4	-2.5	-.33	11.2	1.21	45L	*	L251

TOTAL NUMBER OF LABORATORIES REPORTING = 11

Best values: K45 30 Bekk seconds  
A83 52 Bekk seconds

TAPPI SUGGESTED METHOD T479 SU-71, SMOOTHNESS OF PAPER (BEKK METHOD)

LAB CODE	F	MEANS		COORDINATES		AVG		PROPERTY---	TEST INSTRUMENT---	CONDITIONS
		K45	A83	MAJOR	MINOR	R. SDR	VAR			
L162	Ø	24.1	48.2	-8.0	2.9	.92	45K	SMOOTHNESS,	BEKK	
L182K	Ø	26.5	48.5	-7.3	.8	.91	45K	SMOOTHNESS,	BEKK	
L581	Ø	27.5	52.6	-3.2	1.6	.98	45K	SMOOTHNESS,	BEKK	
L230B	Ø	27.7	52.9	-2.8	1.4	1.04	45K	SMOOTHNESS,	BEKK	
L250M	*	28.3	40.5	-13.9	-4.1	.78	45L	SMOOTHNESS,	BEKK,	20 C, 65% RH
L251	*	30.2	52.4	-2.3	-1.0	1.06	45L	SMOOTHNESS,	BEKK,	20 C, 65% RH
L243K	Ø	30.9	50.4	-3.8	-2.5	1.11	45K	SMOOTHNESS,	BEKK	
L139B	Ø	31.9	47.8	-5.8	-4.5	.90	45K	SMOOTHNESS,	BEKK	
L190C	Ø	33.2	61.7	7.4	-.0	1.31	45K	SMOOTHNESS,	BEKK	
L291K	Ø	34.7	64.6	10.7	-.2	1.06	45K	SMOOTHNESS,	BEKK	
L232B	Ø	35.1	67.3	13.4	.5	.77	45K	SMOOTHNESS,	BEKK	
GMEANS:		30.2	54.9			1.00				
		95% ELLIPSE:	27.1	7.4						WITH GAMMA = 66 DEGREES

LAB CODE	SAMPLE K45		PRINTING 60 GRAMS PER SQUARE METER			SAMPLE A83		WOVE ENVELOPE PAPER 75 GRAMS PER SQUARE METER			TEST D <sub>50</sub> = 10		
	MEAN	DEV	N. DEV	SDR	R <sub>50</sub> SDR	MEAN	DEV	N. DEV	SDR	R <sub>50</sub> SDR	VAR	F	LAB
L1C0	193.	-10.	-.67	13.	.70	91.	-22.	-1.69	12.	.65	47B	Ø	L100
L182B	230.	28.	1.92	27.	1.46	122.	9.	.71	26.	1.47	47B	Ø	L182B
L236	200.	-2.	-.16	13.	.71	117.	4.	.32	12.	.66	47B	Ø	L236
L242	189.	-14.	-.94	18.	.97	110.	-3.	-.21	9.	.53	47B	Ø	L242
L243B	212.	10.	.67	17.	.90	114.	1.	.07	14.	.77	47B	Ø	L243B
L244	189.	-13.	-.91	20.	1.10	107.	-6.	-.45	23.	1.32	47B	Ø	L244
L248	195.	-8.	-.53	22.	1.18	136.	23.	1.75	24.	1.34	47B	Ø	L248
L333	211.	9.	.61	18.	.98	106.	-6.	-.49	22.	1.25	47B	Ø	L333

GR. MEAN = 203. ML/MIN                      GRAND MEAN = 113. ML/MIN                      TEST DETERMINATIONS = 10  
 SD MEANS = 14. ML/MIN                      SD OF MEANS = 13. ML/MIN                      8 LABS IN GRAND MEANS  
 AVERAGE SDR = 19. ML/MIN                      AVERAGE SDR = 18. ML/MIN  
 TOTAL NUMBER OF LABORATORIES REPORTING = 8

Best values: K45 200 milliliter per minute  
 A83 120 milliliter per minute

LAB CODE	F	MEANS		COORDINATES		AVG		PROPERTY---TEST INSTRUMENT---CONDITIONS
		K45	A83	MAJOR	MINOR	R <sub>50</sub> SDR	VAR	
L242	Ø	189.	110.	-13.	5.	.75	47B	SMOOTHNESS, BENDTSEN, WG 150
L244	Ø	189.	107.	-14.	3.	1.21	47B	SMOOTHNESS, BENDTSEN, WG 150
L100	Ø	193.	91.	-20.	-13.	.68	47B	SMOOTHNESS, BENDTSEN, WG 150
L248	Ø	195.	136.	7.	23.	1.26	47B	SMOOTHNESS, BENDTSEN, WG 150
L236	Ø	200.	117.	1.	5.	.69	47B	SMOOTHNESS, BENDTSEN, WG 150
L333	Ø	211.	106.	4.	-10.	1.11	47B	SMOOTHNESS, BENDTSEN, WG 150
L243B	Ø	212.	114.	9.	-5.	.83	47B	SMOOTHNESS, BENDTSEN, WG 150
L182B	Ø	230.	122.	28.	-8.	1.47	47B	SMOOTHNESS, BENDTSEN, WG 150
GMEANS:		203.	113.			1.00		
		95% ELLIPSE:		54.	40.	WITH GAMMA = 34 DEGREES		

LAB CODE	SAMPLE E48 MEAN	COATED GLOSS 91 GFAMS PER SQUARE METER				SAMPLE B58 MEAN	HEAT SET OFFSET BOOK 76 GRAMS PER SQUARE METER				TEST D. * 4		
		DEV	N.DEV	SDR	R.SDR		DEV	N.DEV	SDR	R.SDR	VAR	F	LAB
L126	74.9	50.2	15.40	.6	.66	42.2	-19.2	-2.77	.6	.82	56K	# L126	
L149	23.7	-1.0	-0.29	.5	.60	54.7	-6.6	-0.96	1.0	1.28	56K	Ø L149	
L182	28.7	4.0	1.23	.6	.70	61.9	.5	.08	.2	.23	56K	Ø L182	
L213	28.4	3.7	1.13	1.0	1.14	64.7	3.4	.49	.6	.86	56K	Ø L213	
L291	22.3	-2.4	-0.73	1.3	1.53	64.9	3.5	.51	1.2	1.56	56K	Ø L291	
L333	20.5	-4.2	-1.28	.7	.89	53.1	-8.3	-1.19	1.1	1.48	56K	Ø L333	
L339	28.2	3.5	1.09	.5	.60	71.6	10.2	1.48	.5	.64	56K	Ø L339	
L616	23.7	-1.0	-0.29	.5	.60	66.7	5.4	.77	.5	.67	56K	Ø L616	
L643	22.0	-2.7	-0.83	1.6	1.95	53.2	-8.1	-1.17	1.0	1.28	56K	Ø L643	
GR. MEAN = 24.7 K & N UNITS		GRAND MEAN = 61.4 K & N UNITS				TEST DETERMINATIONS = 4							
SD MEANS = 3.3 K & N UNITS		SD OF MEANS = 6.9 K & N UNITS				8 LABS IN GRAND MEANS							
AVERAGE SDR = .8 K & N UNITS		AVERAGE SDR = .7 K & N UNITS											
L651	61.5	36.8	11.30	.8	1.00	31.7	-29.6	-4.28	.9	1.16	56K	* L651	
L688	60.2	35.5	10.90	.9	1.10	30.1	-31.3	-4.52	.3	.46	56K	* L688	
TOTAL NUMBER OF LABORATORIES REPORTING = 11													
Best values: E48 25 K & N units													
B58 61 K & N units													

The following laboratories were omitted from the grand means because of extreme test results; 126.

LAB CODE	F	MEANS		COORDINATES		AVG		PROPERTY---TEST INSTRUMENT---CONDITIONS	
		E48	B58	MAJOR	MINOR	R.SDR	VAR		
L333	Ø	20.5	53.1	-9.2	1.3	1.18	56K	INK ABSORPTION, K&N INK TEST	
L643	Ø	22.0	53.2	-8.6	-0	1.62	56K	INK ABSORPTION, K&N INK TEST	
L291	Ø	22.3	64.9	2.6	3.4	1.55	56K	INK ABSORPTION, K&N INK TEST	
L149	Ø	23.7	54.7	-6.6	-1.2	.54	56K	INK ABSORPTION, K&N INK TEST	
L616	Ø	23.7	66.7	4.8	2.6	.63	56K	INK ABSORPTION, K&N INK TEST	
L339	Ø	28.2	71.6	10.8	-0.1	.62	56K	INK ABSORPTION, K&N INK TEST	
L213	Ø	28.4	64.7	4.4	-2.4	1.00	56K	INK ABSORPTION, K&N INK TEST	
L182	Ø	28.7	61.9	1.8	-3.6	.46	56K	INK ABSORPTION, K&N INK TEST	
L688	*	60.2	30.1	-18.3	-43.7	.78	56Ø	INK ABSORPTION: ØWN METHOD	
L651	*	61.5	31.7	-16.3	-44.4	1.08	56Ø	INK ABSORPTION: ØWN METHOD	
L126	#	74.9	42.2	-2.1	-53.7	.74	56K	INK ABSORPTION, K&N INK TEST	
GMEANS:		24.7	61.4			1.00			
95% ELLIPSE:		25.2	8.3	WITH GAMMA = 71 DEGREES					

TAPPI COLLABORATIVE REFERENCE PROGRAM  
 ANALYSIS 157-1 TABLE 1  
 HYDROGEN ION CONCENTRATION (PH), COLD  
 TAPPI STANDARD T509 6S-77

FEBRUARY 1979

LAB CODE	SAMPLE J78 MEAN	PRINTING 89 GRAMS PER SQUARE METER				R.SDR	SAMPLE J75 MEAN	PRINTING 76 GRAMS PER SQUARE METER				VAR	F	LAB
		DEV	N.DEV	SDR	R.SDR			DEV	N.DEV	SDR	R.SDR			
L174C	8.180	.841	5.86	.045	1.05	7.200	1.011	4.46	.000	.00	57F	#	L174C	
L182C	7.300	-.039	-.27	.014	.33	6.240	.051	.22	.055	1.92	57D	Ø	L182C	
L251C	7.472	.133	.93	.011	.26	6.288	.059	.44	.011	.38	57P	Ø	L251C	
L328	7.120	-.219	-1.53	.027	.64	5.800	-.389	-1.72	.035	1.24	57M	Ø	L328	
L356	7.342	.003	.02	.030	.71	6.388	.199	.88	.011	.38	57V	Ø	L356	
L442	7.462	.123	.86	.130	3.05	6.231	.041	.18	.031	1.08	57Ø	Ø	L442	
GR. MEAN	7.339	PH UNITS				GRAND MEAN	6.189	PH UNITS				TEST DETERMINATIONS	5	
SD MEANS	.143	PH UNITS				SD OF MEANS	.226	PH UNITS				5 LABS IN GRAND MEANS		
AVERAGE SDR = .043 PH UNITS						AVERAGE SDR = .029 PH UNITS								
TOTAL NUMBER OF LABORATORIES REPORTING = 6														
Best values: J78 7.3 pH units J75 6.2 pH units														

The following laboratories were omitted from the grand means because of extreme test results: 174C.

TAPPI COLLABORATIVE REFERENCE PROGRAM  
 ANALYSIS 157-1 TABLE 2  
 HYDROGEN ION CONCENTRATION (PH), COLD  
 TAPPI STANDARD T509 6S-77

FEBRUARY 1979

LAB CODE	F	MEANS		COORDINATES		AVG		PROPERTY---	TEST INSTRUMENT---	CONDITIONS
		J78	J75	MAJOR	MINOR	R.SDR	VAR			
L328	Ø	7.120	5.800	-.447	-.001	.94	57M	PH, COLD,	BECKMAN ZEROMATIC	
L182C	Ø	7.300	6.240	.025	.059	1.13	57D	PH, COLD,	RADIOMETER TYPE PH M 28	
L356	Ø	7.342	6.388	.174	.095	.55	57V	PH, COLD,	BECKMAN EXPANOMATIC	
L442	Ø	7.462	6.231	.096	-.086	2.06	57Ø	PH, COLD,	ORION DIGITAL IONALYZER	
L251C	Ø	7.472	6.288	.151	-.067	.32	57P	PH, COLD,	RADIOMETER TYPE PH M64	
L174C	#	8.180	7.200	1.294	-.234	.53	57F	PH, COLD,	FISHER ACCUMET MODEL 220	
GMEANS:		7.339	6.189			1.00				
		95% ELLIPSE:		1.294	.355	WITH GAMMA = 60 DEGREES				

TAPPI COLLABORATIVE REFERENCE PROGRAM  
 ANALYSIS T57-2 TABLE 1  
 HYDROGEN ION CONCENTRATION (PH), HOT  
 TAPPI STANDARD T435 6S-77

LAB CODE	SAMPLE J78		PRINTING 89 GRAMS PER SQUARE METER				SAMPLE J75		PRINTING 76 GRAMS PER SQUARE METER				TEST D. = 5		
	MEAN	DEV	N.DEV	SDR	R.SDR	MEAN	DEV	N.DEV	SDR	R.SDR	VAR	F	LAB		
L128	7.812	-.103	-.49	.046	1.32	5.756	-.085	-.22	.034	.98	57L	Ø	L128		
L162	7.926	.012	.06	.067	1.90	5.470	-.371	-.94	.070	1.99	57C	Ø	L162		
L174H	8.200	.286	1.37	.000	.00	6.400	.559	1.42	.000	.00	57G	Ø	L174H		
L182H	7.720	-.194	-.93	.027	.78	5.738	-.103	-.26	.036	1.04	57E	Ø	L182H		

GR. MEAN = 7.914 PH UNITS                      GRAND MEAN = 5.841 PH UNITS                      TEST DETERMINATIONS = 5  
 SD MEANS = .208 PH UNITS                      SD OF MEANS = .395 PH UNITS                      4 LABS IN GRAND MEANS  
 AVERAGE SDR = .035 PH UNITS                      AVERAGE SDR = .035 PH UNITS  
 TOTAL NUMBER OF LABORATORIES REPORTING = 4

TAPPI COLLABORATIVE REFERENCE PROGRAM  
 ANALYSIS T57-2 TABLE 2  
 HYDROGEN ION CONCENTRATION (PH), HOT  
 TAPPI STANDARD T435 6S-77

LAB CODE	F	MEANS		COORDINATES		AVG		PROPERTY---TEST INSTRUMENT---CONDITIONS
		J78	J75	MAJOR	MINOR	R.SDR	VAR	
L182H	Ø	7.720	5.738	-.173	.137	.91	57E PH, HOT, RADIOMETER TYPE PH M 28	
L128	Ø	7.812	5.756	-.119	.060	1.15	57L PH, HOT, L*N	
L162	Ø	7.926	5.470	-.335	-.160	1.94	57C PH, HOT, CORNING MODEL 12 RESEARCH METER	
L174H	Ø	8.200	6.400	.627	-.036	.00	57G PH, HOT, FISHER ACCUMET MODEL 220	

GMEANS: 7.914 5.841                      1.00  
 95% ELLIPSE: 3.250 .972                      WITH GAMMA = 66 DEGREES



OPACITY (89% REFLECTANCE BACKING) IN PERCENT  
TAPPI STANDARD 1425 GS-75, OPACITY OF PAPER (15 DEG./DIFFUSE, ILLUMINANT A) - B&L TYPE

LAB CODE	SAMPLE E86 MEAN	BOND PER SQUARE METER				SAMPLE B21 MEAN	SEMI BLEACHED 65 GRAMS PER SQUARE METER				TEST D. <sup>o</sup> = 10		
		79 GRAMS PER SQUARE METER DEV	N <sub>o</sub> DEV	SDR	R <sub>o</sub> SDR		65 GRAMS PER SQUARE METER DEV	N <sub>o</sub> DEV	SDR	R <sub>o</sub> SDR	VAR	F	LAB
L105	90.40	.65	1.21	.59	1.66	74.00	1.64	1.54	.85	.84	60H	Ø	L105
L109	89.90	.15	.28	.41	1.16	72.25	-.11	-.11	.44	.44	60B	Ø	L108
L118	89.73	-.02	-.03	.35	1.00	72.49	.13	.13	1.19	1.17	60B	Ø	L118
L121	90.09	.34	.63	.51	1.44	72.48	.12	.12	1.56	1.53	60B	Ø	L121
L122	89.41	-.34	-.63	.30	.86	71.38	-.98	-.98	.74	.73	60D	Ø	L122
L123	89.25	-.50	-.92	.32	.91	72.79	.43	.43	1.09	1.07	60W	Ø	L123
L124	89.39	-.36	-.66	.49	1.40	72.63	.27	.27	.95	.94	60B	Ø	L124
L125	89.27	-.48	-.89	.30	.85	71.02	-1.34	-1.34	1.13	1.11	60H	Ø	L125
L131	88.43	-1.32	-2.44	.21	.58	70.67	-1.65	-1.69	.91	.89	60R	Ø	L131
L132	89.48	-.27	-.50	.25	.72	71.02	-1.34	-1.34	1.21	1.19	60B	Ø	L132
L134	89.98	.23	.43	.28	.79	72.84	.48	.48	1.05	1.03	60R	Ø	L134
L136	89.40	-.35	-.65	.33	.94	72.41	.05	.05	1.00	.98	60H	Ø	L136
L139	89.50	-.25	-.46	.35	.99	71.68	-.68	-.68	.85	.84	60B	Ø	L139
L148H	89.29	-.46	-.85	.30	.84	71.20	-1.16	-1.16	1.22	1.20	60H	Ø	L148H
L150	90.20	.45	.84	.42	1.20	74.10	1.74	1.74	.81	.80	60B	Ø	L150
L152	90.05	.30	.56	.23	.64	73.08	.72	.72	.88	.86	60B	Ø	L152
L153	89.95	.20	.37	.64	1.83	72.95	.59	.59	.69	.67	60B	Ø	L153
L157	90.35	.60	1.12	.41	1.17	72.95	.59	.59	1.82	1.79	60B	Ø	L157
L158	90.38	.63	1.17	.50	1.41	73.80	1.44	1.44	1.19	1.17	60D	Ø	L158
L159	89.61	-.14	-.26	.20	.57	71.04	-1.32	-1.32	1.03	1.01	60R	Ø	L159
L162	90.01	.26	.49	.33	.93	72.22	-.14	-.14	.89	.87	60W	Ø	L162
L166	88.76	-.99	-1.83	.58	1.63	71.37	-.95	-.99	1.57	1.54	60B	Ø	L166
L173A	90.30	.55	1.02	.48	1.37	72.20	-.16	-.16	.92	.90	60B	Ø	L173A
L190C	89.08	-.67	-1.24	.34	.97	70.16	-2.20	-2.21	1.41	1.39	60B	Ø	L190C
L190R	89.77	.02	.04	.37	1.04	72.64	.28	.28	1.21	1.19	60B	Ø	L190R
L206	90.00	.25	.47	.34	.95	73.38	1.02	1.02	1.29	1.26	60B	Ø	L206
L210B	89.93	.18	.34	.29	.82	73.21	.85	.85	.93	.92	60B	Ø	L210B
L210D	90.14	.35	.73	.35	1.00	73.07	.71	.71	.79	.78	60D	Ø	L210D
L211S	88.42	-1.33	-2.46	.38	1.08	69.68	-2.68	-2.69	1.21	1.19	60R	*	L211S
L212	90.50	.75	1.40	.71	2.01	72.80	.44	.44	.79	.78	60B	Ø	L212
L213	90.11	.36	.67	.45	1.26	72.84	.48	.48	.91	.89	60B	Ø	L213
L223B	90.11	.36	.67	.39	1.10	72.41	.05	.05	1.40	1.37	60B	Ø	L223B
L225	90.91	1.16	2.16	.64	1.82	74.62	2.26	2.26	1.11	1.09	60B	Ø	L225
L226B	88.86	-.85	-1.65	.30	.86	70.84	-1.52	-1.52	.74	.73	60B	Ø	L226B
L228	89.44	-.31	-.57	.27	.76	72.35	-.01	-.01	.91	.90	60H	Ø	L228
L230	89.92	.17	.32	.18	.51	72.02	-.34	-.34	.50	.49	60B	Ø	L230
L236B	89.02	-.73	-1.35	.64	1.81	71.22	-1.14	-1.14	.45	.49	60B	Ø	L236B
L238A	88.38	-1.37	-2.54	.24	.69	70.79	-1.57	-1.57	.54	.53	60R	*	L238A
L241	89.55	-.20	-.37	.18	.52	72.77	.41	.41	1.06	1.04	60B	Ø	L241
L243	89.56	-.15	-.35	.44	1.24	72.01	-.35	-.35	.74	.73	60B	Ø	L243
L254	90.20	.45	.84	.41	1.17	72.06	-.30	-.30	1.09	1.07	60H	Ø	L254
L255	89.87	.12	.23	.21	.60	72.00	-.36	-.36	.65	.64	60B	Ø	L255
L259	89.99	.24	.45	.45	1.28	73.10	.74	.74	.84	.83	60B	Ø	L259
L261	90.47	.72	1.34	.35	1.00	73.20	.84	.84	1.06	1.04	60B	Ø	L261
L262	90.36	.61	1.14	.27	.77	72.80	.44	.44	.88	.87	60R	Ø	L262
L275	89.34	-.41	-.76	.37	1.05	71.53	-.83	-.83	.85	.83	60R	Ø	L275
L278	90.00	.25	.47	.59	1.67	72.93	.57	.57	.90	.89	60B	Ø	L278
L281	89.77	.02	.04	.24	.68	72.92	.56	.56	.91	.89	60D	Ø	L281
L285D	89.52	-.23	-.42	.28	.80	72.08	-.28	-.28	1.38	1.35	60D	Ø	L285D
L285R	89.14	-.61	-1.13	.21	.60	71.26	-1.10	-1.10	.88	.86	60R	Ø	L285R
L288	90.03	.28	.52	.46	1.30	72.13	-.23	-.23	1.12	1.10	60D	Ø	L288
L301	89.61	-.14	-.26	.23	.66	72.22	-.14	-.14	.71	.70	60B	Ø	L301
L305	89.67	-.08	-.14	.18	.52	71.52	-.84	-.84	.87	.85	60R	Ø	L305
L308	90.45	.70	1.30	.22	.62	73.13	.77	.77	1.35	1.33	60H	Ø	L308
L315	89.21	-.54	-1.00	.35	1.00	71.26	-1.10	-1.10	1.42	1.39	60D	Ø	L315
L317	90.04	.29	.54	.30	.86	72.55	.19	.19	.63	.62	60B	Ø	L317
L318	90.05	.30	.56	.44	1.24	73.35	.99	.99	.75	.73	60B	Ø	L318
L323	89.92	.17	.32	.36	1.01	72.36	-.00	-.00	1.14	1.12	60W	Ø	L323
L326	89.60	-.15	-.27	.44	1.25	72.84	.48	.48	1.71	1.68	60B	Ø	L326
L328	90.60	.85	1.58	.70	1.98	72.60	.24	.24	1.43	1.41	60B	Ø	L328
L339	89.00	-.75	-1.39	.47	1.34	90.00	17.64	17.67	.00	.00	60B	#	L339
L349	89.75	.00	.00	.19	.54	72.80	.44	.44	.75	.74	60D	Ø	L349
L352	89.18	-.57	-1.05	.22	.62	72.10	-.26	-.26	.69	.68	60R	Ø	L352
L354	89.00	-.75	-1.39	.00	.00	71.20	-1.16	-1.16	.79	.78	60B	Ø	L354
L390	89.92	.17	.32	.27	.75	73.92	1.56	1.56	1.92	1.89	60B	Ø	L390



OPACITY (89% REFLECTANCE BACKING) IN PERCENT  
TAPPI STANDARD T425 GS-75, OPACITY OF PAPER (15 DEG./DIFFUSE, ILLUMINANT A) - B&L TYPE

LAB CODE	SAMPLE E86 MEAN	BOND 79 GRAMS PER SQUARE METER				SAMPLE B21 MEAN	SEMI BLEACHED 65 GRAMS PER SQUARE METER				TEST D. # 10		
		DEV	N. DEV	SDR	R. SDR		DEV	N. DEV	SDR	R. SDR	VAR	F	LAB
L396	89.60	-.15	-.27	.44	1.24	71.38	-.98	-.98	.80	.79	60B	Ø	L396
L523	89.98	.23	.43	.18	.50	72.60	.24	.24	.86	.84	60R	Ø	L523
L543	89.40	-.35	-.65	.49	1.38	72.67	.31	.31	.53	.52	60D	Ø	L543
L567	91.02	1.27	2.36	.38	1.08	75.25	2.89	2.89	1.50	1.47	60D	*	L567
L573	89.81	.06	.12	.22	.63	73.24	.88	.88	.50	.49	60H	Ø	L573
L581	89.92	.17	.32	.42	1.20	72.13	-.23	-.23	1.89	1.86	60B	Ø	L581
L587	89.68	-.07	-.13	.43	1.21	71.79	-.57	-.57	1.22	1.20	60B	Ø	L587
L592	88.64	-1.11	-2.06	.37	1.06	71.11	-1.25	-1.25	.86	.84	60W	Ø	L592
L594	89.73	-.02	-.03	.35	.99	73.68	1.32	1.32	1.72	1.69	60D	Ø	L594
L597	89.98	.23	.43	.06	.18	71.35	-1.01	-1.01	.88	.87	60B	Ø	L597
L599	89.60	-.15	-.27	.32	.90	74.35	1.99	1.99	1.31	1.29	60B	X	L599
L608	91.28	1.53	2.84	.32	.91	76.65	4.29	4.30	.72	.71	60D	#	L608
L673R	89.93	.18	.34	.21	.60	72.71	.35	.35	.81	.80	60B	Ø	L673R
L673T	89.75	.00	.00	.20	.56	73.08	.72	.72	.91	.89	60B	Ø	L673T
L688	90.51	.76	1.41	.22	.63	73.33	.97	.97	.79	.77	60B	Ø	L688
L692	89.87	.12	.23	.44	1.26	72.61	.25	.25	1.57	1.55	60D	Ø	L692
GR. MEAN = 89.75 PERCENT		GRAND MEAN = 72.36 PERCENT				TEST DETERMINATIONS = 10							
SD MEANS = .54 PERCENT		SD OF MEANS = 1.00 PERCENT				78 LABS IN GRAND MEANS							
AVERAGE SDR = .35 PERCENT		AVERAGE SDR = 1.02 PERCENT											
L100	89.70	-.05	-.09	.21	.58	72.01	-.35	-.35	.43	.43	60E	*	L100
L232	89.05	-.70	-1.29	.44	1.24	71.30	-1.06	-1.06	.26	.25	60P	*	L232
L249	88.90	-.85	-1.57	.32	.90	71.28	-1.08	-1.08	1.17	1.15	60P	*	L249
L256	88.70	-1.05	-1.94	.19	.53	71.44	-.92	-.92	1.53	1.51	60N	*	L256
L260	89.75	.00	.00	.42	1.21	72.00	-.36	-.36	.67	.66	60P	*	L260
L312	88.30	-1.45	-2.69	.26	.73	70.10	-2.26	-2.27	.88	.86	60P	*	L312
L380	89.00	-.75	-1.39	.47	1.34	70.40	-1.96	-1.96	.84	.83	60P	*	L380
L687	89.25	-.50	-.92	.35	1.00	71.60	-.76	-.76	.84	.83	60P	*	L687

TOTAL NUMBER OF LABORATORIES REPORTING = 89

Best values: E86 89.8 ± 0.9 percent  
B21 72.4 ± 1.5 percent

The following laboratories were omitted from the grand means because of extreme test results: 339, 608.

OPACITY (89% REFLECTANCE BACKING) IN PERCENT  
TAPPI STANDARD T425 8S-75, OPACITY OF PAPER (15 DEG./DIFFUSE, ILLUMINANT A) - B&L TYPE

LAB CODE	F	MEANS		COORDINATES		AVG		PROPERTY---TEST INSTRUMENT---CONDITIONS
		E86	B21	MAJOR	MINOR	E,SDR	VAR	
L312	*	88.30	70.10	-2.66	.35	.80	60P	OPACITY (WHITE BACKING), PHOTOVOLT
L238A	*	88.38	70.79	-2.00	.57	.61	60R	OPACITY (WHITE BACKING), THWING-ALBERT (FORMERLY SRL)
L211S	*	88.42	69.68	-2.99	.07	1.14	60R	OPACITY (WHITE BACKING), THWING-ALBERT (FORMERLY SRL)
L131	Ø	88.43	70.67	-2.09	.48	.74	60R	OPACITY (WHITE BACKING), THWING-ALBERT (FORMERLY SRL)
L592	Ø	88.64	71.11	-1.60	.47	.95	60W	OPACITY (WHITE BACKING), HUYGEN, DIGITAL
L256	*	88.70	71.44	-1.28	.56	1.02	60N	OPACITY (WHITE BACKING), HUNTER
L166	Ø	88.76	71.37	-1.32	.48	1.59	60B	OPACITY (WHITE BACKING), BAUSCH * LOMB
L226B	Ø	88.86	70.84	-1.75	.16	.79	60B	OPACITY (WHITE BACKING), BAUSCH * LOMB
L249	*	88.90	71.28	-1.34	.31	1.03	60P	OPACITY (WHITE BACKING), PHOTOVOLT
L380	*	89.00	70.40	-2.09	-.15	1.08	60P	OPACITY (WHITE BACKING), PHOTOVOLT
L354	Ø	89.00	71.20	-1.37	.19	.39	60B	OPACITY (WHITE BACKING), BAUSCH * LOMB
L339	Ø	89.00	90.00	15.66	8.15	.67	60B	OPACITY (WHITE BACKING), BAUSCH * LOMB
L236B	Ø	89.02	71.22	-1.34	.18	1.15	60B	OPACITY (WHITE BACKING), BAUSCH * LOMB
L232	*	89.05	71.30	-1.26	.18	.75	60P	OPACITY (WHITE BACKING), PHOTOVOLT
L190C	Ø	89.08	70.16	-2.28	-.33	1.18	60B	OPACITY (WHITE BACKING), BAUSCH * LOMB
L285R	Ø	89.14	71.26	-1.25	.08	.73	60R	OPACITY (WHITE BACKING), THWING-ALBERT (FORMERLY SRL)
L352	Ø	89.18	72.10	-.48	.40	.65	60R	OPACITY (WHITE BACKING), THWING-ALBERT (FORMERLY SRL)
L315	Ø	89.21	71.26	-1.23	.02	1.19	60P	OPACITY (WHITE BACKING), BNL-2
L687	*	89.25	71.60	-.90	.13	.92	60D	OPACITY (WHITE BACKING), PHOTOVOLT
L123	Ø	89.25	72.79	.18	.63	.99	60W	OPACITY (WHITE BACKING), HUYGEN, DIGITAL
L125	Ø	89.27	71.02	-1.42	-.13	.98	60H	OPACITY (WHITE BACKING), HUYGEN
L148H	Ø	89.29	71.20	-1.25	-.08	1.02	60H	OPACITY (WHITE BACKING), HUYGEN
L275	Ø	89.34	71.53	-.93	.02	.94	60R	OPACITY (WHITE BACKING), THWING-ALBERT (FORMERLY SRL)
L124	Ø	89.39	72.63	.09	.44	1.17	60B	OPACITY (WHITE BACKING), BAUSCH * LOMB
L543	Ø	89.40	72.67	.13	.45	.95	60D	OPACITY (WHITE BACKING), BNL-2
L136	Ø	89.40	72.41	-.10	.34	.96	60H	OPACITY (WHITE BACKING), HUYGEN
L122	Ø	89.41	71.38	-1.03	-.11	.79	60D	OPACITY (WHITE BACKING), BNL-2
L228	Ø	89.44	72.35	-.14	.27	.83	60H	OPACITY (WHITE BACKING), HUYGEN
L132	Ø	89.48	71.02	-1.33	-.33	.95	60B	OPACITY (WHITE BACKING), BAUSCH * LOMB
L139	Ø	89.50	71.68	-.72	-.06	.91	60B	OPACITY (WHITE BACKING), BAUSCH * LOMB
L285D	Ø	89.52	72.08	-.35	.09	1.08	60D	OPACITY (WHITE BACKING), BNL-2
L241	Ø	89.55	72.77	.29	.35	.78	60B	OPACITY (WHITE BACKING), BAUSCH * LOMB
L243	Ø	89.56	72.01	-.40	.02	.99	60B	OPACITY (WHITE BACKING), BAUSCH * LOMB
L599	X	89.60	74.35	1.74	.98	1.09	60B	OPACITY (WHITE BACKING), BAUSCH * LOMB
L396	Ø	89.60	71.38	-.95	-.28	1.01	60B	OPACITY (WHITE BACKING), BAUSCH * LOMB
L326	Ø	89.60	72.84	.37	.34	1.46	60B	OPACITY (WHITE BACKING), BAUSCH * LOMB
L159	Ø	89.61	71.04	-1.26	-.43	.79	60R	OPACITY (WHITE BACKING), THWING-ALBERT (FORMERLY SRL)
L301	Ø	89.61	72.22	-.19	.07	.68	60B	OPACITY (WHITE BACKING), BAUSCH * LOMB
L305	Ø	89.67	71.52	-.80	-.29	.69	60R	OPACITY (WHITE BACKING), THWING-ALBERT (FORMERLY SRL)
L587	Ø	89.68	71.79	-.55	-.18	1.20	60B	OPACITY (WHITE BACKING), BAUSCH * LOMB
L100	*	89.70	72.01	-.34	-.11	.50	60E	OPACITY (WHITE BACKING), ZEISS ELKEP80, FMY-C(10) FILTER
L118	Ø	89.73	72.49	.11	.07	1.08	60B	OPACITY (WHITE BACKING), BAUSCH * LOMB
L594	Ø	89.73	73.68	1.19	.57	1.34	60D	OPACITY (WHITE BACKING), BNL-2
L260	*	89.75	72.00	-.33	-.15	.93	60P	OPACITY (WHITE BACKING), PHOTOVOLT
L673T	Ø	89.75	73.08	.65	.30	.72	60B	OPACITY (WHITE BACKING), BAUSCH * LOMB
L349	Ø	89.75	72.80	.40	.18	.64	60D	OPACITY (WHITE BACKING), BNL-2
L190R	Ø	89.77	72.64	.26	.10	1.11	60B	OPACITY (WHITE BACKING), BAUSCH * LOMB
L281	Ø	89.77	72.92	.52	.22	.79	60D	OPACITY (WHITE BACKING), BNL-2
L573	Ø	89.81	73.24	.82	.32	.56	60H	OPACITY (WHITE BACKING), HUYGEN
L255	Ø	89.87	72.00	-.28	-.26	.62	60H	OPACITY (WHITE BACKING), BAUSCH * LOMB
L692	Ø	89.87	72.61	.28	-.01	1.40	60D	OPACITY (WHITE BACKING), BNL-2
L108	Ø	89.90	72.25	-.04	-.18	.80	60B	OPACITY (WHITE BACKING), BAUSCH * LOMB
L581	Ø	89.92	72.13	-.14	-.25	1.53	60B	OPACITY (WHITE BACKING), BAUSCH * LOMB
L390	Ø	89.92	73.92	1.49	.50	1.32	60B	OPACITY (WHITE BACKING), BAUSCH * LOMB
L230	Ø	89.92	72.02	-.24	-.30	.50	60B	OPACITY (WHITE BACKING), BAUSCH * LOMB
L323	Ø	89.92	72.36	.07	-.16	1.06	60W	OPACITY (WHITE BACKING), HUYGEN, DIGITAL
L673R	Ø	89.93	72.71	.39	-.02	.70	60B	OPACITY (WHITE BACKING), BAUSCH * LOMB
L210B	Ø	89.93	73.21	.85	.19	.87	60B	OPACITY (WHITE BACKING), BAUSCH * LOMB
L153	Ø	89.95	72.95	.62	.07	1.25	60B	OPACITY (WHITE BACKING), BAUSCH * LOMB
L134	Ø	89.98	72.84	.53	-.01	.91	60R	OPACITY (WHITE BACKING), THWING-ALBERT (FORMERLY SRL)
L597	Ø	89.98	71.35	-.82	-.64	.52	60B	OPACITY (WHITE BACKING), BAUSCH * LOMB
L523	Ø	89.98	72.60	.31	-.11	.67	60R	OPACITY (WHITE BACKING), THWING-ALBERT (FORMERLY SRL)
L259	Ø	89.99	73.10	.77	.09	1.05	60B	OPACITY (WHITE BACKING), BAUSCH * LOMB
L206	Ø	90.00	73.38	1.03	.20	1.11	60B	OPACITY (WHITE BACKING), BAUSCH * LOMB
L278	Ø	90.00	72.93	.62	.01	1.28	60B	OPACITY (WHITE BACKING), BAUSCH * LOMB

ANALYSIS T60-1 TABLE 2

OPACITY (89% REFLECTANCE BACKING) IN PERCENT


TAPPI STANDARD T425 GS-75, OPACITY OF PAPER (15 DEG./DIFFUSE, ILLUMINANT A) - B&L TYPE

LAB CODE	F	MEANS		COORDINATES		AVG		PROPERTY---TEST INSTRUMENT---CONDITIONS
		E86	B21	MAJOR	MINOR	R.SDR	VAR	
L162	Ø	90.01	72.22	-.02	-.30	.90	60W	OPACITY (WHITE BACKING), HUYGEN, DIGITAL
L288	Ø	90.03	72.13	-.09	-.35	1.20	60D	OPACITY (WHITE BACKING), BNL-2
L317	Ø	90.04	72.55	.29	-.18	.74	60B	OPACITY (WHITE BACKING), BAUSCH * LOMB
L152	Ø	90.05	73.08	.78	.03	.75	60B	OPACITY (WHITE BACKING), BAUSCH * LOMB
L318	Ø	90.05	73.35	1.02	.15	.99	60B	OPACITY (WHITE BACKING), BAUSCH * LOMB
L121	Ø	90.09	72.48	.25	-.26	1.48	60B	OPACITY (WHITE BACKING), BAUSCH * LOMB
L223B	Ø	90.11	72.41	.20	-.31	1.24	60B	OPACITY (WHITE BACKING), BAUSCH * LOMB
L213	Ø	90.11	72.84	.55	-.13	1.08	60B	OPACITY (WHITE BACKING), BAUSCH * LOMB
L210D	Ø	90.14	73.07	.81	-.05	.89	60D	OPACITY (WHITE BACKING), BNL-2
L150	Ø	90.20	74.10	1.77	.33	1.00	60B	OPACITY (WHITE BACKING), BAUSCH * LOMB
L254	Ø	90.20	72.06	-.08	-.54	1.12	60B	OPACITY (WHITE BACKING), HUYGEN
L173A	Ø	90.30	72.20	.05	-.57	1.14	60B	OPACITY (WHITE BACKING), BAUSCH * LOMB
L157	Ø	90.35	72.95	.79	-.30	1.48	60B	OPACITY (WHITE BACKING), BAUSCH * LOMB
L262	Ø	90.36	72.80	.66	-.37	.82	60R	OPACITY (WHITE BACKING), THWING-ALBERT (FORMERLY SRL)
L158	Ø	90.38	73.80	1.57	.04	1.29	60D	OPACITY (WHITE BACKING), BNL-2
L105	Ø	90.40	74.00	1.76	.10	1.25	60B	OPACITY (WHITE BACKING), HUYGEN
L308	Ø	90.45	73.13	.99	-.31	.97	60H	OPACITY (WHITE BACKING), HUYGEN
L261	Ø	90.47	73.20	1.07	-.30	1.02	60B	OPACITY (WHITE BACKING), BAUSCH * LOMB
L212	Ø	90.50	72.80	.72	-.50	1.39	60B	OPACITY (WHITE BACKING), BAUSCH * LOMB
L688	Ø	90.51	73.33	1.20	-.28	.70	60B	OPACITY (WHITE BACKING), BAUSCH * LOMB
L328	Ø	90.60	72.60	.58	-.67	1.69	60B	OPACITY (WHITE BACKING), BAUSCH * LOMB
L225	Ø	90.91	74.62	2.54	-.10	1.46	60B	OPACITY (WHITE BACKING), BAUSCH * LOMB
L567	*	91.02	75.25	3.16	.07	1.28	60D	OPACITY (WHITE BACKING), BNL-2
L608	#	91.28	76.65	4.53	.43	.81	60D	OPACITY (WHITE BACKING), BNL-2
GMEANS:		89.75	72.36			1.00		
		95% ELLIPSE:	2.75	.77				WITH GAMMA = 64 DEGREES

OPACITY, B&L TYPE, 89% BACKING

SAMPLE E86 = 89.7 PERCENT

SAMPLE B21 = 72.4 PERCENT



OPACITY (PAPER BACKING) IN PERCENT  
TAPPI STANDARD T425 GS-75, OPACITY OF PAPER (15 DEG./DIFFUSE, ILLUMINANT A) - B&L TYPE

LAB CODE	SAMPLE E86 MEAN	BOND 79 GRAMS PER SQUARE METER				SAMPLE B21 MEAN	SEMI BLEACHED 65 GRAMS PER SQUARE METER				TEST D <sub>0</sub> = 10		
		DEV	N.DEV	SDR	R.SDR		DEV	N.DEV	SDR	R.SDR	VAR	F	LAB
L118	91.40	-.09	-.32	.37	1.24	75.61	.00	.01	.91	.83	60C	Ø	L118
L190C	91.50	.01	.04	.30	1.01	75.59	-.02	-.06	1.03	.93	60C	Ø	L190C
L190R	91.83	.34	1.23	.24	.80	76.06	.45	1.66	1.71	1.55	60C	Ø	L190R
L236B	91.43	-.06	-.21	.43	1.44	75.67	.06	.23	.66	.60	60C	Ø	L236B
L243	91.73	.24	.87	.21	.71	75.49	-.12	-.43	.97	.89	60C	Ø	L243
L543	91.04	-.45	-1.61	.24	.80	75.22	-.39	-1.42	1.32	1.20	60V	Ø	L543
GR. MEAN = 91.49 PERCENT		AVERAGE SDR = .30 PERCENT				GRAND MEAN = 75.61 PERCENT				TEST DETERMINATIONS = 10			
SD MEANS = .28 PERCENT						SD OF MEANS = .27 PERCENT				6 LABS IN GRAND MEANS			
TOTAL NUMBER OF LABORATORIES REPORTING = 6												AVERAGE SDR = 1.10 PERCENT	
Best values: E86 91.5 percent B21 75.5 percent													

OPACITY (PAPER BACKING) IN PERCENT  
TAPPI STANDARD T425 GS-75, OPACITY OF PAPER (15 DEG./DIFFUSE, ILLUMINANT A) - B&L TYPE

LAB CODE	F	MEANS		COORDINATES		AVG		PROPERTY---TEST INSTRUMENT---CONDITIONS	
		E86	B21	MAJOR	MINOR	R.SDR	VAR		
L543	Ø	91.04	75.22	-.59	.04	1.00	60V	OPACITY (PAPER BACKING), DIANC/BNL	
L118	Ø	91.40	75.61	-.06	.06	1.03	60C	OPACITY (PAPER BACKING), BAUSCH * LØMB	
L236B	Ø	91.43	75.67	.00	.09	1.02	60C	OPACITY (PAPER BACKING), BAUSCH * LØMB	
L190C	Ø	91.50	75.59	-.00	-.02	.97	60C	OPACITY (PAPER BACKING), BAUSCH * LØMB	
L243	Ø	91.73	75.49	.09	-.25	.80	60C	OPACITY (PAPER BACKING), BAUSCH * LØMB	
L190R	Ø	91.83	76.06	.56	.09	1.17	60C	OPACITY (PAPER BACKING), HAUSCH * LØMB	
GMEANS:		91.49	75.61			1.00			
		95% ELLIPSE:		1.53	.54	WITH GAMMA = 44 DEGREES			



ANALYSIS T60-3 TABLE 1

OPACITY (PAPER BACKING) IN PERCENT

TAPPI SUGGESTED METHOD T519 6S-78, DIFFUSE OPACITY OF PAPER - ILLUMINANT C, ELREPHO TYPE

LAB CODE	E86		BOND			SAMPLE B21		SEMI BLEACHED				TEST D <sub>0</sub> = 10		
	MEAN	DEV	N <sub>0</sub> DEV	SDR	R <sub>0</sub> SDR	MEAN	DEV	N <sub>0</sub> DEV	SDR	R <sub>0</sub> SDR	VAR	F	LAB	
L100	92.07	.00	.02	.22	1.14	76.65	-.43	-.66	.46	.60	60J	Ø	L100	
L150	92.05	-.01	-.08	.21	1.08	77.05	-.03	-.04	.94	1.23	60J	Ø	L150	
L182E	92.04	-.03	-.13	.20	1.04	77.15	.07	.10	.69	.91	60J	Ø	L182E	
L233F	91.85	-.22	-1.13	.17	.88	76.22	-.86	-1.31	.80	1.05	60F	Ø	L233F	
L236	91.93	-.14	-.71	.19	.97	76.74	-.34	-.52	.86	1.14	60J	Ø	L236	
L242	92.17	.10	.55	.22	1.14	77.65	.57	.86	.93	1.22	60J	Ø	L242	
L244	91.95	-.12	-.60	.25	1.29	76.88	-.20	-.31	.73	.95	60F	Ø	L244	
L250T	91.98	-.09	-.45	.27	1.37	76.83	-.25	-.38	.78	1.02	60J	Ø	L250T	
L251	92.04	-.02	-.12	.22	1.13	76.46	-.62	-.95	.89	1.17	60F	Ø	L251	
L309	88.88	-3.19	-16.69	.30	1.53	70.36	-6.72	-10.20	.42	.55	60J	#	L309	
L360	91.77	-.30	-1.55	.12	.60	76.41	-.67	-1.02	.68	.89	60F	Ø	L360	
L446	92.01	-.05	-.28	.19	.96	77.14	.05	.08	.67	.88	60J	Ø	L446	
L575	92.21	.14	.76	.18	.92	77.42	.34	.51	.97	1.27	60J	Ø	L575	
L598	92.50	.43	2.28	.15	.77	78.69	1.61	2.44	.63	.83	60J	Ø	L598	
L678	92.34	.28	1.44	.13	.69	77.87	.75	1.19	.64	.84	60J	Ø	L678	

GR. MEAN = 92.07 PERCENT

GRAND MEAN = 77.08 PERCENT

TEST DETERMINATIONS = 10

SD MEANS = .19 PERCENT

SD OF MEANS = .66 PERCENT

14 LABS IN GRAND MEANS

AVERAGE SDR = .19 PERCENT

AVERAGE SDR = .76 PERCENT

L626	91.25	-.82	-4.27	.26	1.36	75.70	-1.38	-2.10	.59	.77	60Q	*	L626
TOTAL NUMBER OF LABORATORIES REPORTING = 16													

Best values: E86 92.1 ± 0.3 percent  
B21 77.0 ± 1.0 percent

The following laboratories were omitted from the grand means because of extreme test results: 309.

ANALYSIS T60-3 TABLE 2

OPACITY (PAPER BACKING) IN PERCENT


TAPPI SUGGESTED METHOD T519 6S-78, DIFFUSE OPACITY OF PAPER - ILLUMINANT C, ELREPHO TYPE

LAB CODE	F	MEANS		COORDINATES		AVG R <sub>0</sub> SDR	VAR	PROPERTY---	TEST INSTRUMENT---	CONDITIONS
		E86	B21	MAJØS	MINØR					
L309	#	88.88	70.36	-7.32	1.32	1.04	60J	OPACITY (PAPER BACKING),	ZEISS ELREPHØ,	FMY-C(10) FILTER
L626	*	91.25	75.70	-1.55	.43	1.06	60Q	OPACITY (PAPER BACKING),	PHOTØVOLT	
L360	Ø	91.77	76.41	-.73	.11	.74	60F	OPACITY (PAPER BACKING),	ZEISS ELREPHØ,	FMY-C(10) NØ TRAP
L233F	Ø	91.85	76.22	-.89	-.02	.97	60F	OPACITY (PAPER BACKING),	ZEISS ELREPHØ,	FMY-C(10) NØ TRAP
L236	Ø	91.93	76.74	-.37	.04	1.05	60J	OPACITY (PAPER BACKING),	ZEISS ELREPHØ,	FMY-C(10) FILTER
L244	Ø	91.95	76.88	-.23	.06	1.12	60F	OPACITY (PAPER BACKING),	ZEISS ELREPHØ,	FMY-C(10) NØ TRAP
L250T	Ø	91.98	76.83	-.27	.02	1.20	60J	OPACITY (PAPER BACKING),	ZEISS ELREPHØ,	FMY-C(10) FILTER
L446	Ø	92.01	77.14	.04	.07	.92	60J	OPACITY (PAPER BACKING),	ZEISS ELREPHØ,	FMY-C(10) FILTER
L182E	Ø	92.04	77.15	.06	.04	.97	60J	OPACITY (PAPER BACKING),	ZEISS ELREPHØ,	FMY-C(10) FILTER
L251	Ø	92.04	76.46	-.61	-.14	1.15	60F	OPACITY (PAPER BACKING),	ZEISS ELREPHØ,	FMY-C(10) NØ TRAP
L150	Ø	92.05	77.05	-.03	.01	1.16	60J	OPACITY (PAPER BACKING),	ZEISS ELREPHØ,	FMY-C(10) FILTER
L100	Ø	92.07	76.65	-.42	-.12	.87	60J	OPACITY (PAPER BACKING),	ZEISS ELREPHØ,	FMY-C(10) FILTER
L242	Ø	92.17	77.65	.58	.05	1.18	60J	OPACITY (PAPER BACKING),	ZEISS ELREPHØ,	FMY-C(10) FILTER
L575	Ø	92.21	77.42	.36	-.05	1.10	60J	OPACITY (PAPER BACKING),	ZEISS ELREPHØ,	FMY-C(10) FILTER
L678	Ø	92.34	77.87	.83	-.06	.77	60J	OPACITY (PAPER BACKING),	ZEISS ELREPHØ,	FMY-C(10) FILTER
L598	Ø	92.50	78.69	1.67	-.00	.80	60J	OPACITY (PAPER BACKING),	ZEISS ELREPHØ,	FMY-C(10) FILTER
GMEANS:		92.07	77.08			1.00				
		95% ELLIPSE:		1.96	.21	WITH GAMMA = 74 DEGREES				

OPACITY, ELREPHO TYPE, PAPER BACKING

SAMPLE E86 = 92.07 PERCENT

SAMPLE B21 = 77.08 PERCENT



DIRECTIONAL BLUE REFLECTANCE IN PERCENT  
TAPPI STANDARD T452 6S-77, 'BRIGHTNESS'; MARTIN SWEETS (ACBT & GE) IS STANDARD FOR THIS ANALYSIS

LAB CODE	SAMPLE J36 94 GEAMS PER SQUARE METER					SAMPLE J79 73 GRAMS PER SQUARE METER					TEST D. " 8		
	MEAN	DEV	N. DEV	SDR	R. SDR	MEAN	DEV	N. DEV	SDR	R. SDR	VAR	F	LAB
L108	83.96	.03	.05	.12	.96	68.59	.45	1.28	.24	1.09	65M	0	L108
L122	83.40	-.53	-.68	.09	.75	68.01	-.13	-.36	.16	.76	65N	0	L122
L132	83.45	-.48	-.80	.13	1.06	68.01	-.13	-.36	.26	1.20	65N	0	L132
L158	84.10	.17	.28	.09	.75	68.47	.34	.96	.16	.73	65N	0	L158
L190C	83.69	-.24	-.40	.08	.68	68.24	.10	.28	.18	.85	65A	0	L190C
L210M	84.30	.37	.62	.12	.97	68.21	.07	.21	.20	.91	65M	0	L210M
L210N	84.32	.40	.66	.18	1.48	67.76	-.38	-1.07	.28	1.28	65N	0	L210N
L211	81.64	-2.29	-3.82	.19	1.56	68.02	-.11	-.32	.35	1.63	65N	#	L211
L225	83.49	-.44	-.74	.11	.91	68.29	.15	.43	.40	1.85	65N	0	L225
L243	83.31	-.62	-1.03	.17	1.40	67.80	-.34	-.97	.21	.96	65A	0	L243
L259	83.84	-.05	-.15	.12	.96	67.61	-.53	-1.50	.11	.52	65M	0	L259
L275	83.70	-.23	-.38	.12	.97	68.57	.44	1.25	.10	.48	65M	0	L275
L285	85.56	1.63	2.72	.05	.42	68.17	.04	.11	.17	.77	65N	*	L285
L288	83.72	-.20	-.34	.09	.72	67.76	-.38	-1.07	.30	1.37	65N	0	L288
L308	85.11	1.18	1.97	.14	1.10	68.86	.72	2.07	.14	.65	65N	0	L308
L315	83.61	-.32	-.53	.06	.52	67.95	-.19	-.54	.37	1.71	65N	0	L315
L317	83.69	-.24	-.40	.19	1.53	68.00	-.14	-.39	.11	.49	65M	0	L317
L523	83.91	-.02	-.03	.11	.91	68.32	.19	.53	.23	1.04	65N	0	L523
L543	83.29	-.64	-1.07	.24	1.96	68.55	.41	1.18	.23	1.08	65M	0	L543
L565	83.56	-.37	-.61	.07	.60	68.27	.14	.39	.07	.33	65A	0	L565
L598	NO DATA REPORTED FOR SAMPLE J36					67.65	-.49	-1.39	.17	.78	65M	M	L598
L636	84.82	.90	1.49	.07	.57	67.57	-.56	-1.61	.14	.64	65M	0	L636
L673R	83.67	-.25	-.42	.15	1.21	67.85	-.29	-.82	.45	2.07	65N	0	L673R
L692	80.51	-3.42	-5.69	.38	3.04	66.36	-1.78	-5.07	.46	2.12	65N	#	L692

GR. MEAN = 83.93 PERCENT      GRAND MEAN = 68.14 PERCENT      TEST DETERMINATIONS = 8  
SD MEANS = .60 PERCENT      SD OF MEANS = .35 PERCENT      21 LABS IN GRAND MEANS  
AVERAGE SDR = .12 PERCENT      AVERAGE SDR = .22 PERCENT

L105	83.65	-.28	-.47	.08	.61	67.49	-.65	-1.86	.11	.52	65T	*	L105
L213	84.25	.32	.53	.09	.75	67.87	-.26	-.75	.22	1.01	65T	*	L213
L219	85.20	1.27	2.12	.13	1.06	68.81	.67	1.93	.37	1.72	65P	*	L219
L223	85.86	1.93	3.22	.07	.60	68.75	.61	1.75	.13	.60	65G	*	L223
L232	85.00	1.07	1.78	.00	.00	67.94	-.20	-.57	.18	.82	65P	*	L232
L241	84.16	.23	.39	.23	1.88	68.51	.37	1.07	.10	.46	65I	*	L241
L249	84.51	.58	.57	.11	.91	69.22	1.09	3.10	.16	.73	65P	*	L249
L256	83.41	-.52	-.86	.10	.80	67.34	-.80	-2.29	.12	.55	65H	*	L256
L260	84.26	.33	.55	.13	1.05	69.05	.91	2.60	.08	.35	65P	*	L260
L278	85.56	1.63	2.72	.32	2.59	71.15	3.01	8.60	.35	1.62	65P	*	L278
L301	84.05	.12	.20	.14	1.15	68.32	.19	.53	.13	.59	65G	*	L301
L312	84.87	.95	1.57	.23	1.87	70.94	2.80	7.99	.18	.82	65P	*	L312
L321	85.94	2.01	3.34	.18	1.43	70.00	1.86	5.32	.00	.00	65P	*	L321
L328	87.65	3.72	6.20	.18	1.44	71.64	3.50	9.99	.23	1.07	65P	*	L328
L339	86.62	2.70	4.49	.35	2.86	71.87	3.74	10.67	.23	1.07	65P	*	L339
L380	84.12	.20	.33	.23	1.87	72.00	3.86	11.03	.00	.00	65P	*	L380
L442	81.00	7.07	11.77	.17	1.37	72.97	4.84	13.81	.24	1.12	65T	*	L442
L562	87.00	3.07	5.11	.00	.00	74.50	6.36	18.17	.00	.00	65P	*	L562
L587	84.37	.45	.74	.07	.57	68.01	-.13	-.36	.30	1.41	65I	*	L587
L591	83.85	-.08	-.13	.08	.63	66.79	-1.35	-3.85	.09	.39	65H	*	L591
L626	86.34	2.41	4.01	.19	1.56	70.24	2.10	6.00	.23	1.07	65P	*	L626
L684	83.77	-.15	-.26	.16	1.28	66.79	-1.35	-3.86	.83	3.84	65H	*	L684

TOTAL NUMBER OF LABORATORIES REPORTING = 46

Best values: J36 83.7 ± 1.2 percent  
J79 68.0 ± 0.6 percent

The following laboratories were omitted from the grand means because of extreme test results: 211, 692.


DIRECTIONAL BLUE REFLECTANCE IN PERCENT  
TAPPI STANDARD T452 6S-77, 'BRIGHTNESS'; MARTIN SWEETS (ACBT & GE) IS STANDARD FOR THIS ANALYSIS

LAB CODE	F	MEANS		COORDINATES		AVG R <sub>0</sub> SDR	VAR	PROPERTY---	TEST INSTRUMENT---	CONDITIONS
		J36	J79	MAJ05	MIN06					
L598	M		67.65			.78	65M	BLUE REFLECTANCE (DIRECTIONAL),	MARTIN SWEETS (GE),	S-1
L692	#	80.51	66.36	-3.58	-1.43	2.58	65N	BLUE REFLECTANCE (DIRECTIONAL),	TECHNIDYNE/DIAN0/M.S.,	S-4
L211	#	81.64	68.02	-2.29	.12	1.60	65N	BLUE REFLECTANCE (DIRECTIONAL),	TECHNIDYNE/DIAN0/M.S.,	S-4
L543	0	83.29	68.55	-.60	.47	1.52	65M	BLUE REFLECTANCE (DIRECTIONAL),	MARTIN SWEETS (GE),	S-1
L243	0	83.31	67.80	-.65	-.27	1.18	65A	BLUE REFLECTANCE (DIRECTIONAL),	MARTIN SWEETS (ACBT),	S-2
L122	0	83.40	68.01	-.54	-.07	.75	65N	BLUE REFLECTANCE (DIRECTIONAL),	TECHNIDYNE/DIAN0/M.S.,	S-4
L256	*	83.41	67.34	-.59	-.74	.68	65H	BLUE REFLECTANCE (DIRECTIONAL),	HUNTER	
L132	0	83.45	68.01	-.49	-.08	1.13	65N	BLUE REFLECTANCE (DIRECTIONAL),	TECHNIDYNE/DIAN0/M.S.,	S-4
L225	0	83.49	68.29	-.43	.19	1.38	65N	BLUE REFLECTANCE (DIRECTIONAL),	TECHNIDYNE/DIAN0/M.S.,	S-4
L565	0	83.56	68.27	-.35	.17	.46	65A	BLUE REFLECTANCE (DIRECTIONAL),	MARTIN SWEETS (ACBT),	S-2
L315	0	83.61	67.95	-.33	-.16	1.11	65N	BLUE REFLECTANCE (DIRECTIONAL),	TECHNIDYNE/DIAN0/M.S.,	S-4
L105	*	83.65	67.49	-.34	-.62	.57	65T	BLUE REFLECTANCE (DIRECTIONAL),	HUNTER D25D2M	
L673R	0	83.67	67.85	-.28	-.26	1.64	65N	BLUE REFLECTANCE (DIRECTIONAL),	TECHNIDYNE/DIAN0/M.S.,	S-4
L150C	0	83.69	68.24	-.23	.12	.76	65A	BLUE REFLECTANCE (DIRECTIONAL),	MARTIN SWEETS (ACBT),	S-2
L317	0	83.69	68.00	-.25	-.11	1.01	65M	BLUE REFLECTANCE (DIRECTIONAL),	MARTIN SWEETS (GE),	S-1
L275	0	83.70	68.57	-.18	.46	.72	65M	BLUE REFLECTANCE (DIRECTIONAL),	MARTIN SWEETS (GE),	S-1
L288	0	83.72	67.76	-.24	-.35	1.05	65N	BLUE REFLECTANCE (DIRECTIONAL),	TECHNIDYNE/DIAN0/M.S.,	S-4
L684	*	83.77	68.79	-.29	-1.33	2.56	65H	BLUE REFLECTANCE (DIRECTIONAL),	HUNTER	
L259	0	83.84	67.61	-.14	-.51	.74	65M	BLUE REFLECTANCE (DIRECTIONAL),	MARTIN SWEETS (GE),	S-1
L591	*	83.85	66.79	-.22	-1.33	.51	65H	BLUE REFLECTANCE (DIRECTIONAL),	HUNTER	
L523	0	83.91	68.32	.00	.19	.98	65N	BLUE REFLECTANCE (DIRECTIONAL),	TECHNIDYNE/DIAN0/M.S.,	S-4
L108	0	83.96	68.59	.08	.44	1.03	65M	BLUE REFLECTANCE (DIRECTIONAL),	MARTIN SWEETS (GE),	S-1
L301	*	84.05	68.32	.14	.17	.87	65G	BLUE REFLECTANCE (DIRECTIONAL),	GARDNER	
L158	0	84.10	68.47	.20	.32	.74	65N	BLUE REFLECTANCE (DIRECTIONAL),	TECHNIDYNE/DIAN0/M.S.,	S-4
L380	*	84.12	72.00	.58	3.82	.94	65P	BLUE REFLECTANCE (DIRECTIONAL),	PH0T0V0LT	
L241	*	84.16	68.51	.27	.35	1.17	65I	BLUE REFLECTANCE (DIRECTIONAL),	HUNTER D25D2A	
L213	*	84.25	67.87	.29	-.25	.88	65T	BLUE REFLECTANCE (DIRECTIONAL),	HUNTER D25D2M	
L260	*	84.26	69.05	.42	.87	.70	65P	BLUE REFLECTANCE (DIRECTIONAL),	PH0T0V0LT	
L210M	0	84.30	68.21	.38	.04	.94	65M	BLUE REFLECTANCE (DIRECTIONAL),	MARTIN SWEETS (GE),	S-1
L210N	0	84.32	67.76	.36	-.41	1.38	65N	BLUE REFLECTANCE (DIRECTIONAL),	TECHNIDYNE/DIAN0/M.S.,	S-4
L587	*	84.37	68.01	.43	-.17	.99	65I	BLUE REFLECTANCE (DIRECTIONAL),	HUNTER D25D2A	
L249	*	84.51	69.22	.69	1.02	.82	65P	BLUE REFLECTANCE (DIRECTIONAL),	PH0T0V0LT	
L636	0	84.82	67.57	.83	-.65	.61	65M	BLUE REFLECTANCE (DIRECTIONAL),	MARTIN SWEETS (GE),	S-1
L312	*	84.87	70.94	1.22	2.69	1.35	65P	BLUE REFLECTANCE (DIRECTIONAL),	PH0T0V0LT	
L232	*	85.00	67.94	1.04	-.31	.41	65P	BLUE REFLECTANCE (DIRECTIONAL),	PH0T0V0LT	
L308	0	85.11	68.86	1.25	.60	.87	65N	BLUE REFLECTANCE (DIRECTIONAL),	TECHNIDYNE/DIAN0/M.S.,	S-4
L219	*	85.20	68.81	1.33	.54	1.39	65P	BLUE REFLECTANCE (DIRECTIONAL),	PH0T0V0LT	
L278	*	85.56	71.15	1.93	2.83	2.11	65P	BLUE REFLECTANCE (DIRECTIONAL),	PH0T0V0LT	
L285	*	85.56	68.17	1.63	-.13	.60	65N	BLUE REFLECTANCE (DIRECTIONAL),	TECHNIDYNE/DIAN0/M.S.,	S-4
L223	*	85.86	68.75	1.98	.42	.60	65G	BLUE REFLECTANCE (DIRECTIONAL),	GARDNER	
L321	*	85.94	70.60	2.18	1.65	.72	65P	BLUE REFLECTANCE (DIRECTIONAL),	PH0T0V0LT	
L626	*	86.34	70.24	2.61	1.85	1.32	65P	BLUE REFLECTANCE (DIRECTIONAL),	PH0T0V0LT	
L339	*	86.62	71.87	3.06	3.45	1.97	65P	BLUE REFLECTANCE (DIRECTIONAL),	PH0T0V0LT	
L562	*	87.00	74.50	3.69	6.02	.00	65P	BLUE REFLECTANCE (DIRECTIONAL),	PH0T0V0LT	
L328	*	87.65	71.64	4.05	3.11	1.26	65P	BLUE REFLECTANCE (DIRECTIONAL),	PH0T0V0LT	
L442	*	91.00	72.97	7.52	4.11	1.25	65T	BLUE REFLECTANCE (DIRECTIONAL),	HUNTER D25D2M	
GMEANS:		83.93	68.14			1.00				
		95% ELLIPSE:		1.64	.94	WITH GAMMA = 5 DEGREES				

# BLUE REFLECTANCE, DIRECTIONAL

SAMPLE J36 = 83.9 PERCENT

SAMPLE J79 = 68.1 PERCENT





DIFFUSE BLUE REFLECTANCE IN PERCENT (GLOSS TRAP)  
TAPPI SUGGESTED METHOD T525 SU-72, BRIGHTNESS OF PULP (DIFFUSE ILLUMINATION AND 0 DEG. OBSERVATION)

LAB CODE	SAMPLE J36		PRINTING 94 GFAMS PER SQUARE METER				SAMPLE J79		PRINTING 73 GRAMS PER SQUARE METER				TEST D. # 8		
	MEAN	DEV	N. DEV	SDR	R. SDR	MEAN	DEV	N. DEV	SDR	R. SDR	VAR	F	LAB		
L100	84.62	.17	.36	.09	1.10	67.47	-.40	-.55	.16	.92	65F	Ø	L100		
L121	84.58	.13	.28	.08	.89	68.74	.87	1.21	.13	.73	65K	Ø	L121		
L136	84.53	.08	.17	.16	1.85	67.65	-.22	-.31	.17	.95	65F	Ø	L136		
L150	84.15	-.30	-.64	.06	.77	67.01	-.87	-1.20	.14	.82	65Q	Ø	L150		
L170	84.00	-.45	-.96	.00	.00	67.60	-.27	-.38	.23	1.33	65B	Ø	L170		
L210K	85.05	.60	1.27	.05	.61	69.28	1.41	1.95	.20	1.12	65K	Ø	L210K		
L236	84.82	.37	.78	.08	.90	67.94	.07	.09	.09	.53	65F	Ø	L236		
L242	84.30	-.15	-.31	.08	1.00	66.75	-1.12	-1.55	.32	1.82	65F	Ø	L242		
L250T	84.36	-.10	-.21	.13	1.55	67.66	-.21	-.29	.16	.91	65F	Ø	L250T		
L280	84.29	-.16	-.35	.09	1.06	67.72	-.15	-.21	.23	1.33	65Q	Ø	L280		
L325	85.08	.62	1.32	.09	1.10	67.98	.10	.15	.25	1.41	65F	Ø	L325		
L349	84.00	-.46	-.97	.16	1.94	67.17	-.71	-.98	.20	1.14	65K	Ø	L349		
L446	84.09	-.36	-.77	.11	1.28	67.38	-.49	-.68	.15	.85	65F	Ø	L446		
L573	85.53	1.08	2.29	.08	.90	68.33	.46	.64	.11	.65	65F	Ø	L573		
L575	83.98	-.48	-1.01	.05	.62	67.56	-.31	-.43	.25	1.41	65F	Ø	L575		
L598	84.25	-.21	-.44	.05	.61	69.21	1.33	1.85	.09	.52	65K	Ø	L598		
L636	84.83	.38	.81	.05	.63	68.76	.88	1.22	.16	.89	65K	Ø	L636		
L680	83.70	-.76	-1.61	.10	1.17	67.49	-.38	-.53	.12	.67	65K	Ø	L680		

GR. MEAN = 84.45 PERCENT GRAND MEAN = 67.87 PERCENT TEST DETERMINATIONS = 8  
 SD MEANS = .47 PERCENT SD OF MEANS = .72 PERCENT 18 LABS IN GRAND MEANS  
 AVERAGE SDR = .08 PERCENT AVERAGE SDR = .18 PERCENT

L289 83.37 -1.08 -2.29 .05 .55 67.34 -.53 -.74 .07 .42 65Ø \* L289  
 TOTAL NUMBER OF LABORATORIES REPORTING = 19

Best values: J36 84.4 ± 0.7 percent  
 J79 67.8 ± 1.0 percent

DIFFUSE BLUE REFLECTANCE IN PERCENT (GLOSS TRAP)  
TAPPI SUGGESTED METHOD T525 SU-72, BRIGHTNESS OF PULP (DIFFUSE ILLUMINATION AND 0 DEG. OBSERVATION)


LAB CODE	F	MEANS		COORDINATES		AVG		PROPERTY---TEST INSTRUMENT---CONDITIONS
		J36	J79	MAJOR	MINOR	R. SDR	VAR	
L289	*	83.37	67.34	-.94	.75	.49	65Ø	DIFFUSE REFLECTANCE, ELREPHØ, GL. TRAP, SPECIFIC CALIBRATION
L680	Ø	83.70	67.49	-.67	.52	.92	65K	DIFFUSE REFLECTANCE, ELREPHØ, GL. TRAP, MØØ (ZEISS) BASE
L575	Ø	83.98	67.56	-.48	.30	1.01	65F	DIFFUSE REFLECTANCE, ELREPHØ, GL. TRAP, NRC-PTB ABSOLUTE BASE
L349	Ø	84.00	67.17	-.83	.11	1.54	65K	DIFFUSE REFLECTANCE, ELREPHØ, GL. TRAP, MØØ (ZEISS) BASE
L170	Ø	84.00	67.60	-.44	.29	.66	65B	DIFFUSE REFLECTANCE, ELREPHØ, GL. TRAP, NBS ABSOLUTE BASE
L446	Ø	84.09	67.38	-.60	.12	1.07	65F	DIFFUSE REFLECTANCE, ELREPHØ, GL. TRAP, NRC-PTB ABSOLUTE BASE
L150	Ø	84.15	67.01	-.91	-.09	.79	65Q	DIFFUSE REFLECTANCE, ELREPHØ, GL. TRAP, ZEISS ABSOLUTE BASE
L598	Ø	84.25	69.21	1.12	.75	.57	65K	DIFFUSE REFLECTANCE, ELREPHØ, GL. TRAP, MØØ (ZEISS) BASE
L280	Ø	84.29	67.72	-.21	.09	1.19	65Q	DIFFUSE REFLECTANCE, ELREPHØ, GL. TRAP, ZEISS ABSOLUTE BASE
L242	Ø	84.30	66.75	-1.08	-.34	1.41	65F	DIFFUSE REFLECTANCE, ELREPHØ, GL. TRAP, NRC-PTB ABSOLUTE BASE
L250T	Ø	84.36	67.66	-.23	-.00	1.23	65F	DIFFUSE REFLECTANCE, ELREPHØ, GL. TRAP, NRC-PTB ABSOLUTE BASE
L136	Ø	84.53	67.65	-.17	-.17	1.40	65F	DIFFUSE REFLECTANCE, ELREPHØ, GL. TRAP, NRC-PTB ABSOLUTE BASE
L121	Ø	84.58	68.74	.84	.25	.81	65K	DIFFUSE REFLECTANCE, ELREPHØ, GL. TRAP, MØØ (ZEISS) BASE
L100	Ø	84.62	67.47	-.29	-.32	1.01	65F	DIFFUSE REFLECTANCE, ELREPHØ, GL. TRAP, NRC-PTB ABSOLUTE BASE
L236	Ø	84.82	67.94	.22	-.31	.72	65F	DIFFUSE REFLECTANCE, ELREPHØ, GL. TRAP, NRC-PTB ABSOLUTE BASE
L636	Ø	84.83	68.76	.96	.03	.76	65K	DIFFUSE REFLECTANCE, ELREPHØ, GL. TRAP, MØØ (ZEISS) BASE
L210K	Ø	85.05	69.28	1.53	.06	.87	65K	DIFFUSE REFLECTANCE, ELREPHØ, GL. TRAP, MØØ (ZEISS) BASE
L325	Ø	85.08	67.98	.36	-.52	1.26	65F	DIFFUSE REFLECTANCE, ELREPHØ, GL. TRAP, NRC-PTB ABSOLUTE BASE
L573	Ø	85.53	68.33	.87	-.78	.78	65F	DIFFUSE REFLECTANCE, ELREPHØ, GL. TRAP, NRC-PTB ABSOLUTE BASE

GMEANS: 84.45 67.87 1.00  
 95% ELLIPSE: 2.16 1.03 WITH GAMMA = 64 DEGREES

BLUE REFLECTANCE, DIFFUSE, WITH TRAP

SAMPLE J36 = 84.5 PERCENT

SAMPLE J79 = 67.9 PERCENT



DIFFUSE BLUE REFLECTANCE IN PERCENT (NO GLOSS TRAP)  
TAPPI SUGGESTED METHOD T525 SU-72, BRIGHTNESS OF PULP (DIFFUSE ILLUMINATION AND 0 DEG. OBSERVATION)

LAB CODE	SAMPLE J36 54 GRAMS PER SQUARE METER					SAMPLE J79 73 GRAMS PER SQUARE METER					TEST D. = 8		
	MEAN	DEV	N. DEV	SDR	R. SDR	MEAN	DEV	N. DEV	SDR	R. SDR	VAR	F	LAB
L152	83.40	-.65	-1.06	.05	.65	68.23	-.24	-.37	.09	.66	65E	Ø	L152
L157	84.41	.36	.60	.05	.63	68.82	.35	.53	.13	.93	65E	Ø	L157
L161	83.92	-.13	-.21	.05	.64	68.97	.50	.76	.08	.61	65E	Ø	L161
L173A	83.76	-.28	-.47	.10	1.20	68.44	-.03	-.04	.11	.77	65E	Ø	L173A
L194	85.34	1.30	2.13	.08	.92	73.30	4.83	7.42	.10	.75	65E	#	L194
L219	84.20	.15	.25	.11	1.30	68.68	.21	.32	.22	1.61	65E	Ø	L219
L238A	84.25	.20	.33	.05	.65	69.09	.62	.95	.08	.55	65E	Ø	L238A
L241	83.12	-.92	-1.51	.14	1.68	66.75	-1.72	-2.64	.20	1.44	65E	Ø	L241
L244	84.65	.61	1.00	.09	1.10	68.16	-.31	-.48	.20	1.45	65D	Ø	L244
L251	83.79	-.26	-.43	.15	1.83	68.32	-.15	-.23	.05	.33	65E	Ø	L251
L255	85.25	1.21	1.98	.08	.92	69.44	.97	1.50	.13	.97	65D	Ø	L255
L305	80.70	-3.34	-5.49	.17	2.10	67.83	-.64	-.99	.09	.68	65D	#	L305
L309	84.71	.66	1.09	.09	1.12	68.09	-.38	-.58	.30	2.14	65J	Ø	L309
L360	83.79	-.26	-.43	.05	.63	68.42	-.05	-.08	.14	1.00	65E	Ø	L360
L384	80.10	-3.95	-6.48	.12	1.45	67.32	-1.15	-1.76	.09	.64	65S	#	L384
L565	83.35	-.70	-1.14	.05	.65	68.70	.23	.35	.08	.55	65W	Ø	L565

GR. MEAN = 84.05 PERCENT GRAND MEAN = 68.47 PERCENT TEST DETERMINATIONS = 8  
 SD MEANS = .61 PERCENT SD OF MEANS = .65 PERCENT 13 LABS IN GRAND MEANS  
 AVERAGE SDR = .08 PERCENT AVERAGE SDR = .14 PERCENT  
 TOTAL NUMBER OF LABORATORIES REPORTING = 16  
 Best values: J36 84.0 ± 1.2 percent  
 J79 68.5 ± 1.1 percent

The following laboratories were omitted from the grand means because of extreme test results: 194, 305, 384.


DIFFUSE BLUE REFLECTANCE IN PERCENT (NO GLOSS TRAP)  
TAPPI SUGGESTED METHOD T525 SU-72, BRIGHTNESS OF PULP (DIFFUSE ILLUMINATION AND 0 DEG. OBSERVATION)

LAB CODE	F	MEANS		COORDINATES		AVG		PROPERTY---	TEST INSTRUMENT---	CONDITIONS
		J36	J79	MAJOR	MINOR	R. SDR	VAR			
L384	#	80.10	67.32	-3.47	2.20	1.05	65S	DIFFUSE REFLECTANCE,	ELREPHØ,	NØ TRAP, ABSOLUTE-UNKNOWN BASE
L305	#	80.70	67.83	-2.69	2.08	1.39	65D	DIFFUSE REFLECTANCE,	ELREPHØ,	NØ TRAP, NRC-PTB ABSOLUTE
L241	Ø	83.12	66.75	-1.90	-.45	1.56	65E	DIFFUSE REFLECTANCE,	ELREPHØ,	NØ TRAP, MØØ (ZEISS) BASE
L565	Ø	83.35	68.70	-.29	.67	.60	65W	DIFFUSE REFLECTANCE,	ELREPHØ,	NØ TRAP, NBS MØØ BASE
L152	Ø	83.40	68.23	-.61	.33	.65	65E	DIFFUSE REFLECTANCE,	ELREPHØ,	NØ TRAP, MØØ (ZEISS) BASE
L173A	Ø	83.76	68.44	-.21	.20	.99	65E	DIFFUSE REFLECTANCE,	ELREPHØ,	NØ TRAP, MØØ (ZEISS) BASE
L360	Ø	83.79	68.42	-.21	.16	.82	65E	DIFFUSE REFLECTANCE,	ELREPHØ,	NØ TRAP, MØØ (ZEISS) BASE
L251	Ø	83.79	68.32	-.29	.09	1.08	65E	DIFFUSE REFLECTANCE,	ELREPHØ,	NØ TRAP, MØØ (ZEISS) BASE
L161	Ø	83.92	68.97	.29	.43	.63	65E	DIFFUSE REFLECTANCE,	ELREPHØ,	NØ TRAP, MØØ (ZEISS) BASE
L219	Ø	84.20	68.68	.26	.03	1.45	65E	DIFFUSE REFLECTANCE,	ELREPHØ,	NØ TRAP, MØØ (ZEISS) BASE
L238A	Ø	84.25	69.09	.60	.26	.60	65E	DIFFUSE REFLECTANCE,	ELREPHØ,	NØ TRAP, MØØ (ZEISS) BASE
L157	Ø	84.41	68.82	.50	-.04	.78	65E	DIFFUSE REFLECTANCE,	ELREPHØ,	NØ TRAP, MØØ (ZEISS) BASE
L244	Ø	84.65	68.16	.17	-.66	1.27	65D	DIFFUSE REFLECTANCE,	ELREPHØ,	NØ TRAP, NRC-PTB ABSOLUTE
L309	Ø	84.71	68.09	.16	-.75	1.63	65J	DIFFUSE REFLECTANCE,	ELREPHØ,	NØ TRAP, NBS ABSOLUTE
L255	Ø	85.25	69.44	1.53	-.26	.94	65D	DIFFUSE REFLECTANCE,	ELREPHØ,	NØ TRAP, NRC-PTB ABSOLUTE
L194	#	85.34	73.30	4.48	2.23	.83	65E	DIFFUSE REFLECTANCE,	ELREPHØ,	NØ TRAP, MØØ (ZEISS) BASE
GMEANS:		84.05	68.47			1.00				
		95% ELLIPSE:		2.31	1.25			WITH GAMMA = 48 DEGREES		

BLUE REFLECTANCE, DIFFUSE, NO TRAP

SAMPLE J36 = 84.0 PERCENT

SAMPLE J79 = 68.5 PERCENT



SPECULAR GLOSS AT 75 DEGREES, IN GLOSS UNITS  
TAPPI STANDARD T480 GS-78, SPECULAR GLOSS OF PAPER AND PAPERBOARD AT 75 DEGREES

LAB CODE	SAMPLE E87 COATED OFFSET 117 GRAMS PER SQUARE METER					SAMPLE E51 COATED B66K 118 GRAMS PER SQUARE METER					TEST D. = 10		
	MEAN	DEV	N. DEV	SDR	R. SDR	MEAN	DEV	N. DEV	SDR	R. SDR	VAR	F	LAB
L108	68.0	-.3	-.20	2.4	1.25	69.3	1.2	.75	.6	.56	75H	Ø	L108
L121	69.8	1.5	.86	2.9	1.50	70.0	1.9	1.20	1.3	1.17	75H	Ø	L121
L122	68.5	.1	.08	2.7	1.40	67.3	-.8	-.54	.8	.74	75H	Ø	L122
L128	66.8	-1.5	-.89	2.5	1.29	68.7	.6	.38	.9	.87	75G	Ø	L128
L134	69.2	.5	.51	2.0	1.06	69.5	1.4	.89	1.3	1.16	75H	Ø	L134
L136	70.5	2.2	1.27	1.7	.86	71.2	3.1	1.99	.6	.58	75G	Ø	L136
L149	61.9	-6.4	-3.76	2.2	1.13	62.9	-5.2	-3.33	1.2	1.09	75G	#	L149
L153	70.4	2.1	1.21	2.0	1.04	70.7	2.6	1.65	.7	.62	75G	Ø	L153
L162	73.0	4.7	2.74	2.1	1.09	72.4	4.3	2.74	1.0	.92	75G	#	L162
L173A	64.8	-3.5	-2.06	2.7	1.40	65.6	-2.5	-1.60	1.3	1.23	75G	Ø	L173A
L182	69.1	.8	.47	1.8	.95	67.7	-.4	-.26	1.0	.96	75H	Ø	L182
L189	68.2	-.1	-.05	.9	.46	65.1	-3.0	-1.92	1.7	1.55	75P	Ø	L189
L190C	65.4	-2.9	-1.71	1.9	.98	66.6	-1.5	-.97	1.7	1.56	75G	Ø	L190C
L190R	67.3	-1.1	-.62	1.8	.95	68.1	-.0	-.03	1.5	1.37	75G	Ø	L190R
L206	68.6	.3	.18	1.4	.73	67.7	-.5	-.29	1.3	1.14	75H	Ø	L206
L210	72.2	3.5	2.26	1.6	.84	69.4	1.3	.83	1.4	1.31	75H	Ø	L210
L211	69.5	1.2	.71	1.2	.63	68.8	.7	.43	.9	.79	75H	Ø	L211
L212	71.6	3.3	1.91	2.3	1.20	69.4	1.3	.82	1.0	.88	75P	Ø	L212
L213	68.5	.2	.11	1.6	.82	67.1	-1.0	-.61	.8	.70	75H	Ø	L213
L223	68.7	.4	.22	1.5	.79	68.0	-.1	-.08	.9	.84	75H	Ø	L223
L230	68.6	.3	.16	1.3	.66	66.7	-1.4	-.90	.9	.87	75H	Ø	L230
L243	69.8	1.5	.86	1.9	.97	69.3	1.2	.76	1.1	.97	75B	Ø	L243
L251	69.4	1.1	.63	1.8	.91	67.4	-.7	-.42	1.1	1.04	75G	Ø	L251
L255	67.8	-.5	-.31	1.2	.64	69.5	1.4	.89	.8	.78	75G	Ø	L255
L256	69.2	.5	.52	1.7	.90	68.0	-.1	-.05	1.0	.92	75H	Ø	L256
L259	67.7	-.6	-.37	2.8	1.47	68.8	.7	.44	1.6	1.48	75H	Ø	L259
L262	68.9	.6	.33	2.5	1.29	70.8	2.7	1.75	.9	.83	75K	Ø	L262
L278	62.9	-5.4	-3.18	2.8	1.45	61.8	-6.3	-4.00	1.7	1.55	75G	#	L278
L291	66.6	-1.7	-1.02	3.1	1.59	67.1	-1.0	-.62	2.0	1.85	75H	Ø	L291
L301	67.4	-1.0	-.57	2.4	1.27	68.5	.3	.22	.9	.79	75H	Ø	L301
L315	67.9	-.4	-.25	2.6	1.35	67.1	-1.0	-.65	1.0	.91	75G	Ø	L315
L317	68.2	-.1	-.09	2.2	1.14	67.9	-.2	-.14	1.1	1.02	75H	Ø	L317
L321	69.4	1.1	.63	2.0	1.06	70.1	2.0	1.30	1.1	1.01	75G	Ø	L321
L323	66.9	-1.4	-.82	2.7	1.38	68.5	.4	.27	.7	.67	75H	Ø	L323
L328	67.2	-1.1	-.65	1.4	.70	66.2	-1.9	-1.21	1.1	1.04	75H	Ø	L328
L339	65.4	-2.5	-1.71	3.2	1.64	66.6	-1.5	-.97	2.0	1.83	75P	Ø	L339
L349	65.8	-2.5	-1.46	1.3	.70	67.9	-.2	-.10	1.1	.96	75H	Ø	L349
L372	67.7	-.6	-.34	1.6	.84	67.2	-.9	-.58	1.1	.97	75B	Ø	L372
L388	60.3	-8.0	-4.70	1.1	.55	59.9	-8.2	-5.25	1.0	.91	75P	#	L388
L396	69.3	1.0	.57	1.0	.51	69.2	1.1	.70	1.0	.89	75G	Ø	L396
L456	68.7	.3	.19	1.6	.84	67.3	-.8	-.53	1.0	.95	75H	Ø	L456
L483	68.0	-.3	-.20	1.7	.86	67.9	-.2	-.15	1.3	1.19	75H	Ø	L483
L573	65.4	-2.9	-1.71	2.2	1.15	66.4	-1.7	-1.09	1.0	.88	75G	Ø	L573
L574	66.9	-1.5	-.85	1.7	.88	66.7	-1.4	-.88	1.0	.93	75G	Ø	L574
L583	69.5	1.1	.66	2.0	1.02	69.1	1.0	.66	.6	.53	75H	Ø	L583
L587	69.6	1.3	.74	.8	.44	69.4	1.3	.82	1.0	.88	75H	Ø	L587
L592	68.6	.3	.18	1.5	.76	67.1	-1.0	-.62	1.4	1.27	75H	Ø	L592
L598	67.8	-.5	-.29	1.2	.62	66.9	-1.2	-.78	1.1	1.05	75H	Ø	L598
L643	68.3	-.0	-.03	1.7	.87	65.7	-2.5	-1.57	.9	.78	75H	Ø	L643
L668	68.4	.0	.02	2.5	1.32	66.0	-2.2	-1.37	1.1	1.05	75G	Ø	L668
L670	65.2	-3.1	-1.83	2.8	1.43	66.6	-1.5	-.95	.8	.70	75H	Ø	L670
L688	68.4	.0	.02	1.2	.63	68.7	.6	.40	1.1	.97	75G	Ø	L688

GR. MEAN = 68.3 GLOSS UNITS      GRAND MEAN = 68.1 GLOSS UNITS      TEST DETERMINATIONS = 10  
SD MEANS = 1.7 GLOSS UNITS      SD OF MEANS = 1.6 GLOSS UNITS      49 LABS IN GRAND MEANS  
AVERAGE SDR = 1.9 GLOSS UNITS      AVERAGE SDR = 1.1 GLOSS UNITS

L250    63.4    -4.9    -2.88    1.3    .70    59.5    -8.6    -5.50    1.2    1.08    75Q    \*    L250  
L288    68.0    -.3    -.20    2.8    1.44    68.7    .6    .41    .8    .71    75I    \*    L288

TOTAL NUMBER OF LABORATORIES REPORTING = 54

Best values: E87 68 ± 3 gloss units  
E51 68 ± 2 gloss units

The following laboratories were omitted from the grand means because of extreme test results: 149, 278, 388.




ANALYSIS T75-1 TABLE 2  
 SPECULAR GLOSS AT 75 DEGREES, IN GLOSS UNITS  
 TAPPI STANDARD T480 6S-78, SPECULAR GLOSS OF PAPER AND PAPERBOARD AT 75 DEGREES

LAB CODE	F	MEANS		COORDINATES		AVG		PROPERTY---	TEST INSTRUMENT---	CONDITIONS
		E87	ES1	MAJOR	MINOR	R.SDR	VAR			
L388	#	60.3	59.9	-11.5	-0.9	.73	75P	SPECULAR GLOSS (75 DEGREE)	PHOTOVOLT	
L149	#	61.9	62.9	-8.3	.3	1.11	75G	SPECULAR GLOSS (75 DEGREE)	GARDNER	
L278	#	62.9	61.8	-8.2	-1.1	1.50	75G	SPECULAR GLOSS (75 DEGREE)	GARDNER	
L250	*	63.4	59.5	-9.4	-3.2	.89	75Q	SPECULAR GLOSS (75 DEGREE)	PHOTOVOLT	20 C, 65% RH
L173A	Ø	64.8	65.6	-4.3	.4	1.32	75G	SPECULAR GLOSS (75 DEGREE)	GARDNER	
L670	Ø	65.2	66.6	-3.3	.9	1.07	75H	SPECULAR GLOSS (75 DEGREE)	HUNTER	
L573	Ø	65.4	66.4	-3.3	.6	1.02	75G	SPECULAR GLOSS (75 DEGREE)	GARDNER	
L339	Ø	65.4	66.6	-3.2	.8	1.74	75P	SPECULAR GLOSS (75 DEGREE)	PHOTOVOLT	
L190C	Ø	65.4	66.6	-3.2	.8	1.27	75G	SPECULAR GLOSS (75 DEGREE)	GARDNER	
L349	Ø	65.8	67.9	-2.0	1.5	.83	75H	SPECULAR GLOSS (75 DEGREE)	HUNTER	
L291	Ø	66.6	67.1	-1.9	.4	1.72	75H	SPECULAR GLOSS (75 DEGREE)	HUNTER	
L128	Ø	66.8	68.7	-.8	1.5	1.08	75G	SPECULAR GLOSS (75 DEGREE)	GARDNER	
L574	Ø	66.9	66.7	-2.0	-1.1	.90	75G	SPECULAR GLOSS (75 DEGREE)	GARDNER	
L323	Ø	66.9	68.5	-.8	1.2	1.02	75H	SPECULAR GLOSS (75 DEGREE)	HUNTER	
L328	Ø	67.2	66.2	-2.1	-.7	.87	75H	SPECULAR GLOSS (75 DEGREE)	HUNTER	
L190R	Ø	67.3	68.1	-.8	.7	1.16	75G	SPECULAR GLOSS (75 DEGREE)	GARDNER	
L301	Ø	67.4	68.5	-.5	.9	1.03	75H	SPECULAR GLOSS (75 DEGREE)	HUNTER	
L259	Ø	67.7	68.8	-.0	.9	1.47	75H	SPECULAR GLOSS (75 DEGREE)	HUNTER	
L372	Ø	67.7	67.2	-1.0	-.3	.90	75B	SPECULAR GLOSS (75 DEGREE)	BAUSCH • LOMB	
L255	Ø	67.8	69.5	.5	1.4	.71	75G	SPECULAR GLOSS (75 DEGREE)	GARDNER	
L598	Ø	67.8	66.9	-1.2	-.6	.83	75H	SPECULAR GLOSS (75 DEGREE)	HUNTER	
L315	Ø	67.9	67.1	-1.0	-.5	1.13	75G	SPECULAR GLOSS (75 DEGREE)	GARDNER	
L483	Ø	68.0	67.9	-.4	.1	1.03	75H	SPECULAR GLOSS (75 DEGREE)	HUNTER	
L288	*	68.0	68.7	.2	.7	1.07	75I	SPECULAR GLOSS (75 DEGREE)	HUNTER	20 C, 65% RH
L108	Ø	68.0	69.3	.5	1.1	.90	75H	SPECULAR GLOSS (75 DEGREE)	HUNTER	
L317	Ø	68.2	67.9	-.3	-.1	1.08	75H	SPECULAR GLOSS (75 DEGREE)	HUNTER	
L189	Ø	68.2	65.1	-2.0	-2.2	1.00	75P	SPECULAR GLOSS (75 DEGREE)	PHOTOVOLT	
L643	Ø	68.3	65.7	-1.7	-1.8	.83	75H	SPECULAR GLOSS (75 DEGREE)	HUNTER	
L688	Ø	68.4	68.7	.4	.4	.80	75G	SPECULAR GLOSS (75 DEGREE)	GARDNER	
L668	Ø	68.4	66.0	-1.4	-1.6	1.18	75G	SPECULAR GLOSS (75 DEGREE)	GARDNER	
L122	Ø	68.5	67.3	-.4	-.7	1.07	75H	SPECULAR GLOSS (75 DEGREE)	HUNTER	
L213	Ø	68.5	67.1	-.5	-.8	.76	75H	SPECULAR GLOSS (75 DEGREE)	HUNTER	
L230	Ø	68.6	66.7	-.7	-1.2	.76	75H	SPECULAR GLOSS (75 DEGREE)	HUNTER	
L206	Ø	68.6	67.7	-.1	-.5	.94	75H	SPECULAR GLOSS (75 DEGREE)	HUNTER	
L592	Ø	68.6	67.1	-.4	-.9	1.01	75H	SPECULAR GLOSS (75 DEGREE)	HUNTER	
L456	Ø	68.7	67.3	-.3	-.8	.89	75H	SPECULAR GLOSS (75 DEGREE)	HUNTER	
L223	Ø	68.7	68.0	.2	-.3	.82	75H	SPECULAR GLOSS (75 DEGREE)	HUNTER	
L262	Ø	68.9	70.8	2.2	1.7	1.06	75K	SPECULAR GLOSS (75 DEGREE)	GAERTNER (K-C TYPE)	
L182	Ø	69.1	67.7	.3	-.8	.95	75H	SPECULAR GLOSS (75 DEGREE)	HUNTER	
L134	Ø	69.2	69.5	1.6	.5	1.11	75H	SPECULAR GLOSS (75 DEGREE)	HUNTER	
L256	Ø	69.2	68.0	.6	-.6	.91	75H	SPECULAR GLOSS (75 DEGREE)	HUNTER	
L396	Ø	69.3	69.2	1.4	.2	.70	75G	SPECULAR GLOSS (75 DEGREE)	GARDNER	
L251	Ø	69.4	67.4	.4	-1.2	.98	75G	SPECULAR GLOSS (75 DEGREE)	GARDNER	
L321	Ø	69.4	70.1	2.1	.8	1.03	75G	SPECULAR GLOSS (75 DEGREE)	GARDNER	
L583	Ø	69.5	69.1	1.5	.0	.77	75H	SPECULAR GLOSS (75 DEGREE)	HUNTER	
L211	Ø	69.5	68.8	1.4	-.3	.71	75H	SPECULAR GLOSS (75 DEGREE)	HUNTER	
L587	Ø	69.6	69.4	1.8	.1	.66	75H	SPECULAR GLOSS (75 DEGREE)	HUNTER	
L121	Ø	69.8	70.0	2.3	.4	1.34	75H	SPECULAR GLOSS (75 DEGREE)	HUNTER	
L243	Ø	69.8	69.3	1.9	-.1	.97	75B	SPECULAR GLOSS (75 DEGREE)	BAUSCH • LOMB	
L153	Ø	70.4	70.7	3.3	.6	.83	75G	SPECULAR GLOSS (75 DEGREE)	GARDNER	
L136	Ø	70.5	71.2	3.7	.9	.72	75G	SPECULAR GLOSS (75 DEGREE)	GARDNER	
L212	Ø	71.6	69.4	3.3	-1.2	1.04	75P	SPECULAR GLOSS (75 DEGREE)	PHOTOVOLT	
L210	Ø	72.2	69.4	3.8	-1.6	1.07	75H	SPECULAR GLOSS (75 DEGREE)	HUNTER	
L162	*	73.0	72.4	6.3	.1	1.01	75G	SPECULAR GLOSS (75 DEGREE)	GARDNER	
GMEANS:		68.3	68.1			1.00				
		95% ELLIPSE:		5.4	2.4	WITH GAMMA = 41 DEGREES				

SPECULAR GLOSS, 75 DEGREE

SAMPLE E87 = 68.3 GLOSS UNITS    SAMPLE E51 = 68.1 GLOSS UNITS



TAPPI COLLABORATIVE REFERENCE PROGRAM  
ANALYSIS T90-1 TABLE 1  
THICKNESS (CALIPER), THOUSANDTHS OF AN INCH  
TAPPI STANDARD T411 6S-76

LAB CODE	SAMPLE J64		PRINTING 85 GRAMS PER SQUARE METER				SAMPLE J83		PRINTING 73 GRAMS PER SQUARE METER				TEST D <sub>s</sub> = 10		
	MEAN	DEV	N <sub>o</sub> DEV	SDR	R <sub>o</sub> SDR	MEAN	DEV	N <sub>o</sub> DEV	SDR	R <sub>o</sub> SDR	VAR	F	LAB		
L100	5.409	.150	1.66	.044	.99	2.806	.063	.83	.028	.71	90V	Ø	L100		
L105	5.244	-.015	-.17	.027	.61	2.783	.045	.55	.050	1.27	90Q	Ø	L105		
L118	5.379	.120	1.33	.033	.74	2.759	.021	.26	.043	1.10	90Q	Ø	L118		
L122	5.375	.116	1.29	.066	1.48	2.830	.092	1.13	.034	.85	90V	Ø	L122		
L123F	5.356	.097	1.08	.028	.64	2.880	.142	1.73	.048	1.22	90F	Ø	L123F		
L125	5.266	.007	.08	.053	1.19	2.781	.043	.53	.035	.89	90T	Ø	L125		
L128	5.292	.033	.37	.050	1.13	2.751	.013	.16	.038	.95	90T	Ø	L128		
L141	5.199	-.060	-.67	.057	1.28	2.734	-.004	-.04	.038	.97	90T	Ø	L141		
L153	5.388	.129	1.43	.054	1.21	2.937	.199	2.43	.040	1.01	90T	*	L153		
L158	5.254	-.005	-.06	.046	1.04	5.711	2.973	36.22	.033	.84	90T	#	L158		
L159	5.353	.094	1.04	.049	1.11	2.827	.089	1.09	.041	1.03	90T	Ø	L159		
L162	5.190	-.069	-.77	.057	1.28	2.490	-.248	-3.02	.032	.80	90D	X	L162		
L166	5.261	.002	.02	.023	.51	2.817	.079	.97	.030	.76	90T	Ø	L166		
L173B	5.370	.111	1.23	.067	1.52	2.740	.002	.03	.052	1.31	90F	Ø	L173B		
L174	5.230	-.029	-.32	.082	1.85	2.710	-.028	-.34	.074	1.87	90T	Ø	L174		
L182	5.272	.013	.14	.058	1.31	2.723	-.015	-.18	.051	1.28	90L	Ø	L182		
L183	5.363	.104	1.15	.038	.85	2.760	.022	.27	.042	1.07	90T	Ø	L183		
L190C	5.150	-.109	-1.21	.053	1.19	2.680	-.058	-.70	.042	1.07	90T	Ø	L190C		
L212	5.345	.086	.95	.044	.98	2.780	.042	.52	.026	.65	90T	Ø	L212		
L213	5.110	-.149	-1.65	.032	.71	2.590	-.148	-1.80	.032	.80	90T	Ø	L213		
L223	5.260	.001	.01	.030	.67	2.699	-.039	-.47	.028	.72	90V	Ø	L223		
L228	5.245	-.014	-.16	.050	1.12	2.765	.027	.33	.063	1.58	90T	Ø	L228		
L233	5.579	.320	3.55	.047	1.05	2.753	.015	.19	.028	.71	90Q	X	L233		
L238A	5.320	.061	.68	.063	1.42	2.710	-.028	-.34	.032	.80	90T	Ø	L238A		
L241	5.315	.056	.62	.041	.93	2.845	.107	1.31	.064	1.63	90T	Ø	L241		
L249	5.220	-.039	-.43	.046	1.03	2.637	-.101	-1.23	.059	1.50	90T	Ø	L249		
L259	10.121	4.862	53.94	15.066	338.78	2.734	-.004	-.04	.022	.55	90T	X	L259		
L260	5.199	-.060	-.67	.026	.58	2.643	-.095	-1.15	.031	.77	90T	Ø	L260		
L261	5.293	.034	.38	.046	1.03	2.690	-.048	-.58	.017	.43	90T	Ø	L261		
L262	5.195	-.064	-.71	.037	.83	2.670	-.068	-.82	.054	1.36	90T	Ø	L262		
L285	5.193	-.066	-.73	.031	.69	2.798	.060	.74	.023	.57	90T	Ø	L285		
L291	5.340	.081	.90	.039	.89	2.815	.077	.94	.034	.85	90T	Ø	L291		
L305	5.255	-.004	-.04	.055	1.24	2.765	.027	.33	.041	1.04	90T	Ø	L305		
L309	5.240	-.019	-.21	.084	1.90	2.650	-.088	-1.07	.053	1.33	90T	Ø	L309		
L318	5.100	-.159	-1.76	.053	1.19	2.590	-.148	-1.80	.032	.80	90T	Ø	L318		
L320	5.110	-.149	-1.65	.032	.71	2.585	-.153	-1.86	.034	.85	90T	Ø	L320		
L323	5.037	-.222	-2.46	.043	.97	2.549	-.189	-2.30	.021	.53	90T	Ø	L323		
L324	5.206	-.053	-.59	.020	.44	2.712	-.026	-.31	.011	.29	90T	Ø	L324		
L326	5.315	.056	.62	.034	.76	2.745	.007	.09	.037	.93	90T	Ø	L326		
L328	5.210	-.049	-.54	.056	1.25	2.724	-.014	-.17	.027	.68	90T	Ø	L328		
L331	5.229	-.030	-.33	.079	1.78	2.813	.075	.92	.027	.68	90T	Ø	L331		
L339	5.195	-.064	-.71	.055	1.24	2.755	.017	.21	.037	.93	90T	Ø	L339		
L352	5.267	.008	.09	.032	.73	2.761	.023	.28	.019	.47	90Q	Ø	L352		
L356	5.202	-.057	-.63	.038	.85	2.700	-.038	-.46	.028	.71	90T	Ø	L356		
L358	5.160	-.099	-1.10	.066	1.48	2.661	-.077	-.93	.039	.98	90T	Ø	L358		
L376	5.340	.081	.90	.052	1.16	2.815	.077	.94	.041	1.04	90T	Ø	L376		
L380	5.200	-.059	-.65	.000	.00	2.800	.062	.76	.000	.00	90T	Ø	L380		
L382	5.310	.051	.57	.032	.71	2.710	-.028	-.34	.152	3.86	90T	Ø	L382		
L390	5.452	.193	2.14	.043	.98	2.880	.142	1.73	.039	.98	90T	Ø	L390		
L442	5.399	.140	1.55	.044	.99	2.879	.141	1.72	.029	.74	90V	Ø	L442		
L556	5.262	.003	.03	.034	.77	2.690	-.048	-.58	.047	1.18	90T	Ø	L556		
L557	5.170	-.089	-.99	.067	1.52	2.660	-.078	-.95	.084	2.13	90T	Ø	L557		
L574	5.095	-.164	-1.82	.045	1.02	2.663	-.075	-.91	.034	.86	90V	Ø	L574		
L575	5.216	-.043	-.48	.048	1.07	2.714	-.024	-.29	.025	.62	90T	Ø	L575		
L581	5.380	.121	1.34	.026	.58	2.830	.092	1.13	.026	.65	90T	Ø	L581		
L585	5.300	.041	.45	.000	.00	2.710	-.028	-.34	.032	.80	90T	Ø	L585		
L587	5.220	-.039	-.43	.042	.95	2.680	-.058	-.70	.063	1.60	90T	Ø	L587		
L626	4.940	-.319	-3.54	.076	1.72	2.377	-.361	-4.39	.027	.69	90T	#	L626		
L679	5.175	-.084	-.93	.054	1.21	2.630	-.108	-1.31	.026	.65	90T	Ø	L679		
L693	5.258	-.001	-.01	.056	1.26	2.738	.000	.00	.025	.63	90T	Ø	L693		

GR. MEAN = 5.259 MILS  
SD MEANS = .090 MILS

GRAND MEAN = 2.738 MILS  
SD OF MEANS = .082 MILS

TEST DETERMINATIONS = 10  
55 LABS IN GRAND MEANS

AVERAGE SDR = .044 MILS

AVERAGE SDR = .040 MILS

GR. MEAN = 133.58 MICROMETER

GRAND MEAN = 69.54 MICROMETER

TAPPI COLLABORATIVE REFERENCE PROGRAM  
ANALYSIS 190-1 TABLE 1  
THICKNESS (CALIPER), THOUSANDTHS OF AN INCH  
TAPPI STANDARD T411 GS-76

FEBRUARY 1979

LAB CODE	SAMPLE J64 MEAN	PRINTING 85 GRAMS PER SQUARE METER				SAMPLE J83 MEAN	PRINTING 73 GRAMS PER SQUARE METER				TEST D. = 10		
		DEV	N.DEV	SDR	R.SDR		DEV	N.DEV	SDR	R.SDR	VAR	F	LAB
L106	5.090	-.169	-1.87	.057	1.28	2.860	.122	1.49	.070	1.77	90C	*	L106
L185	5.190	-.069	-.77	.025	.57	2.698	-.040	-.48	.020	.50	90B	*	L185
L242B	5.303	.044	.49	.020	.46	2.742	.005	.06	.028	.70	90B	*	L242B
L242P	5.406	.147	1.63	.041	.92	2.825	.087	1.06	.027	.67	90P	*	L242P
L243	5.162	-.097	-1.08	.026	.58	2.672	-.066	-.80	.029	.72	90S	*	L243
L251	5.236	-.023	-.25	.034	.78	2.726	-.011	-.14	.031	.79	90W	*	L251
L344	5.280	.021	.23	.063	1.42	2.690	-.048	-.58	.032	.80	90U	*	L344
L396M	51.200	45.941	509.63	.789	17.74	26.400	23.662	288.24	.516	13.07	90S	*	L396M
L563	5.340	.081	.90	.052	1.16	2.820	.082	1.00	.063	1.60	90U	*	L563
L567	5.266	.007	.08	.047	1.06	2.674	-.064	-.77	.052	1.32	90B	*	L567
L576	5.135	-.124	-1.38	.048	1.08	2.662	-.076	-.92	.043	1.09	90C	*	L576
L616	5.336	.077	.85	.039	.87	2.832	.094	1.15	.041	1.05	90C	*	L616
L684	5.211	-.048	-.53	.020	.44	2.830	.092	1.13	.036	.90	90U	*	L684

TOTAL NUMBER OF LABORATORIES REPORTING = 73

Best values: J64 5.26 ± 0.14 mils  
J83 2.74 ± 0.13 mils

The following laboratories were omitted from the grand means because of extreme test results: 158, 626.



TAPPI COLLABORATIVE REFERENCE PROGRAM  
ANALYSIS T90-1 TABLE 2  
THICKNESS (CALIPER), THOUSANDTHS OF AN INCH  
TAPPI STANDARD T411 6S-76

LAB CODE	F	MEANS		COORDINATES		AVG		PROPERTY---TEST INSTRUMENT---	CONDITIONS
		J64	J83	MAJOR	MINOR	R.SDR	VAR		
L626	#	4.940	2.377	-.478	-.058	1.20	90T	THICKNESS (CALIPER), TMI,	MOTOR DRIVEN
L323	Ø	5.037	2.549	-.291	.006	.75	90T	THICKNESS (CALIPER), TMI,	MOTOR DRIVEN
L106	*	5.090	2.860	-.045	.204	1.52	90C	THICKNESS (CALIPER), CADY,	HAND DRIVEN
L574	Ø	5.095	2.663	-.172	.053	.94	90V	THICKNESS (CALIPER), TMI,	MOTOR DRIVEN, DIGITIZED
L318	Ø	5.100	2.590	-.217	-.005	.99	90T	THICKNESS (CALIPER), TMI,	MOTOR DRIVEN
L320	Ø	5.110	2.585	-.213	-.015	.78	90T	THICKNESS (CALIPER), TMI,	MOTOR DRIVEN
L213	Ø	5.110	2.590	-.205	-.011	.76	90T	THICKNESS (CALIPER), TMI,	MOTOR DRIVEN
L576	*	5.135	2.662	-.143	.026	1.08	90C	THICKNESS (CALIPER), CADY,	HAND DRIVEN
L190C	Ø	5.150	2.680	-.120	.029	1.13	90T	THICKNESS (CALIPER), TMI,	MOTOR DRIVEN
L358	Ø	5.160	2.661	-.125	.008	1.23	90T	THICKNESS (CALIPER), TMI,	MOTOR DRIVEN
L243	*	5.162	2.672	-.116	.015	.65	90S	THICKNESS (CALIPER), SCHÖPPER,	HAND DRIVEN
L557	Ø	5.170	2.660	-.118	.001	1.83	90T	THICKNESS (CALIPER), TMI,	MOTOR DRIVEN
L679	Ø	5.175	2.630	-.134	-.025	.93	90T	THICKNESS (CALIPER), TMI,	MOTOR DRIVEN
L185	*	5.190	2.698	-.078	.016	.54	90B	THICKNESS (CALIPER), AMTHOR,	HAND DRIVEN
L162	X	5.190	2.490	-.216	-.139	1.04	90D	THICKNESS (CALIPER), CADY,	MOTOR DRIVEN
L285	Ø	5.193	2.798	-.009	.089	.63	90T	THICKNESS (CALIPER), TMI,	MOTOR DRIVEN
L339	Ø	5.195	2.755	-.036	.055	1.09	90T	THICKNESS (CALIPER), TMI,	MOTOR DRIVEN
L262	Ø	5.195	2.670	-.093	-.008	1.09	90T	THICKNESS (CALIPER), TMI,	MOTOR DRIVEN
L141	Ø	5.199	2.734	-.047	.037	1.13	90T	THICKNESS (CALIPER), TMI,	MOTOR DRIVEN
L260	Ø	5.199	2.643	-.108	-.031	.68	90T	THICKNESS (CALIPER), TMI,	MOTOR DRIVEN
L380	Ø	5.200	2.800	-.003	.086	.00	90T	THICKNESS (CALIPER), TMI,	MOTOR DRIVEN
L356	Ø	5.202	2.700	-.067	.010	.78	90T	THICKNESS (CALIPER), TMI,	MOTOR DRIVEN
L324	Ø	5.206	2.712	-.057	.016	.36	90T	THICKNESS (CALIPER), TMI,	MOTOR DRIVEN
L328	Ø	5.210	2.724	-.046	.022	.96	90T	THICKNESS (CALIPER), TMI,	MOTOR DRIVEN
L684	*	5.211	2.830	.025	.101	.67	90U	THICKNESS (CALIPER), TMI,	HAND DRIVEN
L575	Ø	5.216	2.714	-.048	.011	.85	90T	THICKNESS (CALIPER), TMI,	MOTOR DRIVEN
L249	Ø	5.220	2.637	-.056	-.049	1.27	90T	THICKNESS (CALIPER), TMI,	MOTOR DRIVEN
L587	Ø	5.220	2.680	-.067	-.017	1.27	90T	THICKNESS (CALIPER), TMI,	MOTOR DRIVEN
L331	Ø	5.229	2.813	.028	.076	1.23	90T	THICKNESS (CALIPER), TMI,	MOTOR DRIVEN
L174	Ø	5.230	2.710	-.040	-.601	1.86	90T	THICKNESS (CALIPER), TMI,	MOTOR DRIVEN
L251	*	5.236	2.726	-.024	.007	.78	90W	THICKNESS (CALIPER), L * W,	MOTOR DRIVEN, 20 C, 65% RH
L309	Ø	5.240	2.650	-.072	-.053	1.62	90T	THICKNESS (CALIPER), TMI,	MOTOR DRIVEN
L105	Ø	5.244	2.783	.019	.044	.94	90Q	THICKNESS (CALIPER), EMVECC,	MOTOR DRIVEN
L228	Ø	5.245	2.765	.008	.030	1.35	90T	THICKNESS (CALIPER), TMI,	MOTOR DRIVEN
L158	#	5.254	5.711	1.970	2.227	.94	90T	THICKNESS (CALIPER), TMI,	MOTOR DRIVEN
L305	Ø	5.255	2.765	.015	.023	1.14	90T	THICKNESS (CALIPER), TMI,	MOTOR DRIVEN
L693	Ø	5.258	2.738	-.000	.001	.95	90T	THICKNESS (CALIPER), TMI,	MOTOR DRIVEN
L223	Ø	5.260	2.699	-.025	-.030	.70	90V	THICKNESS (CALIPER), TMI,	MOTOR DRIVEN, DIGITIZED
L166	Ø	5.261	2.817	.054	.058	.64	90I	THICKNESS (CALIPER), TMI,	MOTOR DRIVEN
L556	Ø	5.262	2.690	-.029	-.038	.98	90T	THICKNESS (CALIPER), TMI,	MOTOR DRIVEN
L567	*	5.266	2.674	-.037	-.052	1.19	90B	THICKNESS (CALIPER), AMTHOR,	HAND DRIVEN
L125	Ø	5.266	2.781	.034	.028	1.04	90T	THICKNESS (CALIPER), TMI,	MOTOR DRIVEN
L352	Ø	5.267	2.761	.022	.012	.60	90Q	THICKNESS (CALIPER), EMVECC,	MOTOR DRIVEN
L182	Ø	5.272	2.723	-.000	-.020	1.30	90L	THICKNESS (CALIPER), L * W,	MOTOR DRIVEN
L344	*	5.280	2.690	-.016	-.050	1.11	90U	THICKNESS (CALIPER), TMI,	HAND DRIVEN
L128	Ø	5.292	2.751	.034	-.012	1.04	90T	THICKNESS (CALIPER), TMI,	MOTOR DRIVEN
L261	Ø	5.293	2.690	-.006	-.058	.73	90T	THICKNESS (CALIPER), TMI,	MOTOR DRIVEN
L585	Ø	5.300	2.710	.012	-.048	.40	90T	THICKNESS (CALIPER), TMI,	MOTOR DRIVEN
L242Ø	*	5.303	2.742	.036	-.026	.58	90Ø	THICKNESS (CALIPER), MESSMER,	MOTOR DRIVEN, BS3983
L382	Ø	5.310	2.710	.020	-.055	2.28	90T	THICKNESS (CALIPER), TMI,	MOTOR DRIVEN
L326	Ø	5.315	2.745	.047	-.032	.85	90T	THICKNESS (CALIPER), TMI,	MOTOR DRIVEN
L241	Ø	5.315	2.845	.113	.043	1.28	90T	THICKNESS (CALIPER), TMI,	MOTOR DRIVEN
L238A	Ø	5.320	2.710	.027	-.061	1.11	90T	THICKNESS (CALIPER), TMI,	MOTOR DRIVEN
L616	*	5.336	2.832	.120	.019	.96	90C	THICKNESS (CALIPER), CADY,	HAND DRIVEN
L563	*	5.340	2.820	.115	.008	1.38	90U	THICKNESS (CALIPER), TMI,	HAND DRIVEN
L291	Ø	5.340	2.815	.112	.004	.87	90T	THICKNESS (CALIPER), TMI,	MOTOR DRIVEN
L376	Ø	5.340	2.815	.112	.004	1.10	90T	THICKNESS (CALIPER), TMI,	MOTOR DRIVEN
L212	Ø	5.345	2.780	.092	-.025	.82	90T	THICKNESS (CALIPER), TMI,	MOTOR DRIVEN
L159	Ø	5.353	2.827	.130	.004	1.07	90T	THICKNESS (CALIPER), TMI,	MOTOR DRIVEN
L123F	Ø	5.356	2.880	.167	.042	.93	90F	THICKNESS (CALIPER), FEDERAL,	MOTOR DRIVEN
L183	Ø	5.363	2.760	.093	-.052	.56	90T	THICKNESS (CALIPER), TMI,	MOTOR DRIVEN
L173B	Ø	5.370	2.740	.085	-.072	1.41	90F	THICKNESS (CALIPER), FEDERAL,	MOTOR DRIVEN
L122	Ø	5.375	2.830	.148	-.008	1.17	90V	THICKNESS (CALIPER), TMI,	MOTOR DRIVEN, DIGITIZED
L118	Ø	5.379	2.759	.104	-.064	.92	90Q	THICKNESS (CALIPER), EMVECC,	MOTOR DRIVEN
L581	Ø	5.380	2.830	.152	-.011	.62	90T	THICKNESS (CALIPER), TMI,	MOTOR DRIVEN



TAPPI COLLABORATIVE REFERENCE PROGRAM  
 ANALYSIS T90-1 TABLE 2  
 THICKNESS (CALIPER), THOUSANDTHS OF AN INCH  
 TAPPI STANDARD T411 6S-76

LAB CODE	F	MEANS		COORDINATES		AVG		PROPERTY---TEST INSTRUMENT---CONDITIONS
		J64	J83	MAJØR	MINØR	R.SDR	VAR	
L153	*	5.388	2.937	.229	.063	1.11	90T	THICKNESS (CALIPER), TMI, MØTØR DRIVEN
L442	Ø	5.399	2.879	.199	.013	.87	90V	THICKNESS (CALIPER), TMI, MØTØR DRIVEN, DIGITIZED
L242P	*	5.406	2.825	.168	-.032	.79	90P	THICKNESS (CALIPER), MESSMER, MØTØR DRIVEN, ISO R534
L100	Ø	5.409	2.806	.158	-.048	.85	90V	THICKNESS (CALIPER), TMI, MØTØR DRIVEN, DIGITIZED
L390	Ø	5.452	2.860	.239	-.022	.98	90T	THICKNESS (CALIPER), TMI, MØTØR DRIVEN
L233	X	5.579	2.753	.250	-.201	.88	90Q	THICKNESS (CALIPER), EMVECO, MØTØR DRIVEN
L259	X	10.121	2.734	3.633	-3.231	169.67	90T	THICKNESS (CALIPER), TMI, MØTØR DRIVEN
L396M	*	51.200	26.400	50.065	-12.806	15.40	90S	THICKNESS (CALIPER), SCHØPPER, HAND DRIVEN
GMEANS:		5.259	2.738			1.00		
		55% ELLIPSE:		.293	.101	WITH GAMMA = 41 DEGREES		


# THICKNESS (CALIPER)

SAMPLE J64 = 5.26 MILS

SAMPLE J83 = 2.74 MILS

SAMPLE J64 = 133.6 MICRØMETER

SAMPLE J83 = 69.5 MICRØMETER






GRAMMAGE (MASS PER UNIT AREA)

SAMPLE D32 = 79.3 G/SQ.METER

SAMPLE D33 = 117.8 G/SQ.METER



SUMMARY TABLE

TEST METHOD	SAMPLE CODE	GRAND MEAN	SD OF MEAN	AVER SDR	REPL CRP	LABS INCL	LABS PARTIC	REPL TAPPI	REPEAT	REPROD	
AIR RESISTANCE, GURLEY T40-1	GURLEY UNITS	K21	51.2	3.0	4.3	10	55	60	10	3.8	8.3
	GURLEY UNITS	K43	16.4	.8	1.2					1.0	2.3
AIR RESISTANCE, SHEFFIELD T40-2	SHEFF. UNITS	K21	71.1	5.9	5.9	10	40	45	10	5.2	16.4
	SHEFF. UNITS	K43	167.6	9.2	8.8					7.7	25.5
AIR RESISTANCE, GURLEY HG FLOTATION T41-1	SEC/10 CC	E68	651.	62.	85.	10	12	12	10	75.	171.
	SEC/10 CC	E65	731.	63.	148.					129.	173.
SMOOTHNESS, PARKER PRINTSURF T44-1	MICRONS	K45	5.47	.43	.10	10	9	9	10	.09	1.20
	MICRONS	A83	4.84	.40	.24					.21	1.10
SMOOTHNESS, SHEFFIELD T45-1	SHEFF. UNITS	K45	160.3	7.0	9.2	15	89	91	10	8.0	20.0
	SHEFF. UNITS	A83	105.4	6.8	10.7					9.4	19.6
SMOOTHNESS, HEKK T45-2	HEKK SECONDS	K45	30.2	3.9	2.6	15	9	11	10	2.3	10.9
	HEKK SECONDS	A83	54.9	7.6	9.2					8.1	21.5
SMOOTHNESS, HENDTSEN T47-1	ML/MIN	K45	203.	14.	19.	10	8	8	10	16.	40.
	ML/MIN	A83	113.	13.	18.					16.	36.
K & N INK ABSORPTION T56-1	K & N UNITS	E48	24.7	3.3	.8	4	8	11	4	1.2	9.0
	K & N UNITS	H58	61.4	6.9	.7					1.0	19.2
PH, COLD T57-1	PH UNITS	J78	7.339	.143	.043	5	5	6	2	.083	.402
	PH UNITS	J75	6.189	.226	.029					.056	.629
PH, HOT T57-2	PH UNITS	J78	7.914	.208	.035	5	4	4	2	.069	.579
	PH UNITS	J75	5.841	.395	.035					.069	1.095
OPACITY, H&L TYPE, 89% BACKING T60-1	PERCENT	E86	89.75	.54	.35	10	78	89	5	.44	1.52
	PERCENT	H21	72.36	1.00	1.02					1.26	2.90
OPACITY, H&L TYPE, PAPER BACKING T60-2	PERCENT	E86	91.49	.28	.30	10	6	6	5	.37	.81
	PERCENT	H21	75.61	.27	1.10					1.36	1.23
OPACITY, ELREPHO TYPE, PAPER BACKING T60-3	PERCENT	E86	92.07	.19	.19	10	14	16	5	.24	.56
	PERCENT	H21	77.08	.66	.76					.94	1.94
BLUE REFLECTANCE, DIRECTIONAL T65-1	PERCENT	J36	83.93	.60	.12	8	21	46	6	.14	1.66
	PERCENT	J79	68.14	.35	.22					.24	.98
BLUE REFLECTANCE, DIFFUSE, WITH TRAP T65-2	PERCENT	J36	84.45	.47	.08	8	18	19	6	.10	1.31
	PERCENT	J79	67.87	.72	.18					.20	2.00
BLUE REFLECTANCE, DIFFUSE, NO TRAP T65-3	PERCENT	J36	84.05	.61	.08	8	13	16	6	.09	1.69
	PERCENT	J79	68.47	.65	.14					.16	1.81
SPECULAR GLOSS, 75 DEGREE T75-1	GLOSS UNITS	E87	68.3	1.7	1.9	10	49	54	5	2.4	5.0
	GLOSS UNITS	E51	68.1	1.6	1.1					1.4	4.4
THICKNESS (CALIPER) T90-1	MILS	J64	5.259	.090	.044	10	55	73	10	.039	.250
	MILS	J83	2.738	.082	.040					.035	.227
GRAMMAGE (MASS PER UNIT AREA) T95-1	G/SG. METER	D32	79.28	.56	.34	10	18	20	3	.55	1.61
	G/SG. METER	D33	117.80	.94	1.28					2.05	3.13



U.S. DEPT. OF COMM. BIBLIOGRAPHIC DATA SHEET	1. PUBLICATION OR REPORT NO. <b>TAPPI CRP 58G</b>	2. Gov't. Accession No.	3. Recipient's Accession No.
4. TITLE AND SUBTITLE <b>Technical Association of the Pulp and Paper Industry COLLABORATIVE REFERENCE PROGRAM FOR PAPER Report #58G</b>		5. Publication Date <b>June 8, 1979</b>	6. Performing Organization Code
7. AUTHOR(S) <b>R. G. Powell, J. Horlick</b>		8. Performing Organ. Report No. <b>NBSIR 79-1374</b>	
9. PERFORMING ORGANIZATION NAME AND ADDRESS  <b>NATIONAL BUREAU OF STANDARDS DEPARTMENT OF COMMERCE WASHINGTON, DC 20234</b>		10. Project/Task/Work Unit No.	11. Contract/Grant No.
12. SPONSORING ORGANIZATION NAME AND COMPLETE ADDRESS (Street, City, State, ZIP) <b>Collaborative Testing Services, Inc. 9241 Wood Glade Drive, Great Falls, VA 22066 and Technical Association of the Pulp and Paper Industry</b>		13. Type of Report & Period Covered <b>FINAL</b>	
15. SUPPLEMENTARY NOTES  <input type="checkbox"/> Document describes a computer program; SF-185, FIPS Software Summary, is attached.		14. Sponsoring Agency Code	
16. ABSTRACT (A 200-word or less factual summary of most significant information. If document includes a significant bibliography or literature survey, mention it here.)  <b>Collaborative Reference Programs provide participating labora- tories with the means for checking periodically the level and uniformity of their testing in comparison with that of other participating laboratories. An important by-product of the pro- grams is the provision of realistic pictures of the state of the testing art. This is one of the periodic reports showing averages for each participant, within and between laboratory variability, and other information for participants and stan- dards committees.</b>			
17. KEY WORDS (six to twelve entries; alphabetical order; capitalize only the first letter of the first key word unless a proper name; separated by semicolons)  <b>Collaborative reference program; Laboratory evaluation; Paper; Precision; Reference samples, Testing calibration</b>			
18. AVAILABILITY <input type="checkbox"/> Unlimited  <input checked="" type="checkbox"/> For Official Distribution. Do Not Release to NTIS  <input type="checkbox"/> Order From Sup. of Doc., U.S. Government Printing Office, Washington, DC 20402, SD Stock No. SN003-003-  <input type="checkbox"/> Order From National Technical Information Service (NTIS), Springfield, VA, 22161		19. SECURITY CLASS (THIS REPORT)  UNCLASSIFIED	21. NO. OF PRINTED PAGES  56
		20. SECURITY CLASS (THIS PAGE)  UNCLASSIFIED	22. Price



