

A11100 986230

NBS
PUBLICATIONS

NAT'L INST. OF STAND & TECH R.I.C.



A11104 938205



TECHNICAL ASSOCIATION OF THE
PULP AND PAPER INDUSTRY

COLLABORATIVE REFERENCE PROGRAM
FOR PAPER

REPORT NO. 59G



U.S. DEPARTMENT OF COMMERCE
National Bureau of Standards

QC
100
.U56
79-1801
1979
C.2

NBS COLLABORATIVE REFERENCE PROGRAMS

TAPPI Paper and Board (6 times per year)

Bursting strength	Smoothness
Tearing strength	Surface pick strength
Tensile breaking strength	K & N ink absorption
Elongation to break	pH
Tensile energy absorption	Opacity
Folding endurance	Blue reflectance (brightness)
Stiffness	Specular gloss, 75°
Air resistance	Thickness
Grammage	Concora (flat crush)
	Ring crush

FKBG-API Containerboard (48 times per year)

Mullen burst of linerboard
Concora test of medium

MCCA Color and Appearance (4 times per year)

Gloss at 60°
Color and color difference

CTS Rubber (4 times per year)

Tensile strength, ultimate elongation and tensile stress
Hardness
Mooney viscosity
Vulcanization properties

CTS Thermal Insulation Materials (2 times per year)

19 test methods for thermal insulation materials covering:
thermal properties; strength properties; dimensions, stability,
and density properties; fire properties; and properties of
vapor barriers

ASTM Cement (2 times per year)

Chemical (11 chemical components)
Physical (8 characteristics)

AASHTO Bituminous

Asphalt cement (2 times per year)
Cutbacks (once a year)

NBS Collaborative Reference Programs
A05 Technology Building
National Bureau of Standards
Washington, DC 20234

National Bureau of Standards
FEB 27 1980
not Acc-Linc
QC100
456
79-1801
1979
C-A

TECHNICAL ASSOCIATION OF THE
PULP AND PAPER INDUSTRY

COLLABORATIVE REFERENCE PROGRAM
FOR PAPER

Report No. 59G

R. G. Powell
TAPPI-NBS Research Associate
Collaborative Testing Services, Inc.

J. Horlick
Office of Testing Laboratory Evaluation Technology
Office of Engineering Standards
National Engineering Laboratory

NBSIR 79-1801


U. S. DEPARTMENT OF COMMERCE
National Bureau of Standards

INTRODUCTION

Reports 59S and 59G comprise the fifth set of reports for the 78-79 program year. Participants in tests which involve strength properties of paper will receive only the S report; those in tests which measure other properties will receive only the G report.

Notes and comments to individual laboratories and "Best Values" applicable to a particular method are given following Table 1 for each method. See page 4 of this report for an explanation of "Best Values". Please do not confuse these Best Values with provisional values included with the samples to detect serious discrepancies at the time of test.

If there are any questions on the notes, the analyses, or the reports in general, contact Robert G. Powell or Jeffrey Horlick on 301/921-2946.



Jeffrey Horlick, Administrator
NBS-TAPPI Collaborative Reference Program
Office of Testing Laboratory Evaluation Technology

July 27, 1979

BACKGROUND AND PURPOSE

In 1969, the National Bureau of Standards and the Technical Association of the Pulp and Paper Industry established a collaborative reference program to provide a participating laboratory with a means to check periodically the level and uniformity of its testing in comparison with that of other laboratories.

The interchange of paper and board products and of the raw materials for these products requires agreement among raw material suppliers, paper and board producers, converters, distributors, retailers, commercial testing laboratories, user organizations and the ultimate consumer as to the meaning of test results, an agreement that cannot be achieved without accurate and precise testing. This program is designed to help assure agreement.

HOW THE PROGRAM WORKS

Participants Select the Tests in which they wish to participate. This choice is made on joining the program, but additional tests may be added at any time. Also new participants may enter the program at any time.

Test Samples are Distributed Bimonthly; i.e. every 2 months.

Provisional Values are Provided with the Samples for one or both of the test levels, depending on method. The provisional values permit serious discrepancies to be detected without delay. (It is left to the discretion of the laboratory supervisor as to whether these values should be known to the operator.)

Each Participant Tests the Samples, following instructions provided for each test method. The full check on a single instrument should normally take no more than 30 minutes. The test results are then sent to NBS for analysis. The participant is also asked to report other information relevant to an accurate analysis, such as test conditions and the instruments used.

Industry Means, Best Values and Other Statistics are developed from the data by NBS. The best values are estimates based on a careful examination of all data, both current and past, with special attention to results obtained by the National Bureau of Standards and other recognized reference laboratories in this and other countries.

A Quick Report is Prepared for each participating laboratory reporting data on time. This report shows the industry mean values, and the deviations of the laboratory's results from these values for each test method.

A Longer Summary Report, Showing the Data from all Participants, is also prepared. In the summary report, of which this report is an example, each laboratory is identified by a code number so that the information is maintained on a confidential basis. However, instruments are identified by type so participants can compare their results with those obtained on similar instruments of different manufacture. This report includes test averages, best values and standard deviations for individual participants and for the group as a whole. A participant should be able to readily determine the level and variability of his results in comparison with those of the other laboratories.

Repeatability and Reproducibility Statements such as Contained in ASTM, TAPPI and ISO Standards are included at the end of the report. Participants can check their performance level against the precision statement given in the test method or specification.

TABLE OF CONTENTS

Analyses In This Report

PAGE

1	Introduction
ii	Description of Program
iv	Metric Conversion Table
1	Key to Tables and Graphs
5	40-1 Air Resistance, Gurley Oil type
8	40-2 Air Resistance, Sheffield type
11	41-1 Air Resistance, Gurley Mercury type
13	44-1 Smoothness, Parker Printsurf
14	45-1 Smoothness, Sheffield type
19	45-2 Smoothness, Bekk type
21	47-1 Smoothness, Bendtsen type
22	56-1 K & N Ink Absorption
23	57-1 pH, Cold Extraction
24	57-2 pH, Hot Extraction
25	60-1 Opacity, White (89%) Backing
30	60-2 Opacity, Paper Backing, B & L type
31	60-3 Opacity, Paper Backing, Elrepho type
33	65-1 Blue Reflectance (Brightness), Directional
36	65-2 Blue Reflectance, Diffuse, Elrepho (Gloss Trap)
38	65-3 Blue Reflectance, Diffuse, Elrepho (No Gloss Trap)
40	75-1 Specular Gloss, 75 degree
43	90-1 Thickness (Caliper)
48	95-1 Grammage (Basis Weight)
50	Summary

Analyses In The S Report

10-1	Bursting Strength - Up to 45 psi
10-2	Bursting Strength - Up to 45 psi, Air Clamps
11-1	Bursting Strength - Up to 100 psi
15-1	Tearing Strength - Deep Cutout
17-1	Tearing Strength - No Cutout
19-1	Tensile Breaking Strength - Packaging Papers
20-1	Tensile Breaking Strength - Printing Papers, CRE
20-2	Tensile Breaking Strength - Printing Papers, Pendulum
25-1	Tensile Energy Absorption - Packaging Papers
26-1	Tensile Energy Absorption - Printing Papers
28-1	Elongation to Break - Packaging Papers
29-1	Elongation to Break - Printing Papers
30-1	Folding Endurance, MIT type
30-2	Folding Endurance, MIT type, log (base 10)
35-1	Stiffness, Gurley
36-1	Stiffness, Taber
49-1	Surface Pick Strength, IGT
50-1	Surface Pick Strength, Wax
91-1	Concora (Flat Crush)
96-1	Ring Crush

TABLE OF CONVERSION FACTORS TO METRIC (SI) UNITS

<u>Physical Quantity</u>	<u>To Convert From</u>	<u>To</u>	<u>Multiply by</u>
Bursting strength	psi	kPa	6.895
	kg/cm ²	kPa	98.07
	bar	kPa	100.00
Tearing strength	g	mN	9.807
Tensile strength	lb/in.	kN/m	.1751
	lb/0.5 in.	kN/m	.3502
	lb/15 mm	kN/m	.2965
	kg/15 mm	kN/m	.6538
	kg/25 mm	kN/m	.3923
	kg/mm	kN/m	9.807
Tensile energy absorption	ft-lb/ft ²	J/m ²	14.59
	in.-lb/in. ²	J/m ²	175.1
	kg-m/m ²	J/m ²	9.807
Bending stiffness	g·cm	μN·m	98.07
Flat-crush strength (Concora)	lb	N	4.448
Ring-crush (TAPPI)	lb	N	4.448
	lb/6.00 in.	kN/m	0.0292
Thickness	mil	μm	25.40

KEY TO TABLES AND GRAPHS

MEAN -	The average of individual TEST DETERMINATIONS. The number of TEST DETERMINATIONS in the mean is given in the upper right corner of the first table (TEST D.) and again at the bottom of this table.
GRAND MEAN - (GR. MEAN)	The average of the individual laboratory MEANS, excluding laboratories flagged (see column F) with an X, #, or +. The GRAND MEAN is given in US customary units and, where applicable, in SI metric units.
SD OF MEANS - (SD MEANS)	The standard deviation of the laboratory MEANS about the GRAND MEAN; an index of the among-laboratory precision.
DEV -	The deviation or difference of the laboratory MEAN from the GRAND MEAN.
N. DEV -	The normal deviate or ratio of the DEV to the SD OF MEANS; an indication of the degree of divergence of the laboratory MEAN from the GRAND MEAN. A N. DEV of more than 2 or less than -2 may indicate that the participant is not following the procedure considered standard for this analysis.
SDR -	The standard deviation of repeated measurements; that is, of individual test determinations about their MEAN.
AVERAGE SDR -	The average of the individual laboratory SDR's; an index of the within-laboratory precision of repeated measurements.
R. SDR -	The relative standard deviation of repeated measurements; that is, the ratio of the SDR to the AVERAGE SDR; an indication of the ability of a participant to repeat his measurements relative to the average ability. The greater the number of TEST DETERMINATIONS the closer the R. SDR should be to unity. If R. SDR is outside the limits given below, the participant may not be following the procedure considered standard for this analysis:

No. of test Determinations	Lower limit for R. SDR	Upper limit for R. SDR
3	0.09	2.58
5	0.27	2.06
8	0.40	1.77
10	0.46	1.67
15	0.56	1.53
20	0.61	1.45
25	0.65	1.39

VAR - Code for instrument type or variation in condition, see second table.

F - Flag, with following meaning:

+ - Excluded from grand means because VAR non-standard for this analysis.

- Excluded because data were not understood or because of a non-coded variation reported by the laboratory. (See NOTES following Table 1 for each method).

M - Excluded because data for one sample are missing.

X - Excluded because plotted point would fall outside of the 99% error ellipse, (see below for explanation of Graph).

* - Included in grand means but plotted point falls outside of the 95% error ellipse. The participants should take this as a warning to reexamine his testing procedure.

S - Included in grand mean but only after omission of one or more 'wild' values; that is, test determinations more than 3 times AVERAGE SDR from the laboratory's MEAN. Not more than 20% of the test determination may be excluded in this manner without rejecting the laboratory.

0 - Included in grand mean and inside 95% error ellipse.

COORDINATES - Distances along major and minor axes of error ellipse. If special additive or concurrent model of the measuring process applies to this method, the distance along the minor axis represents the random error within a laboratory while that along the major axis also includes a systematic laboratory component of error.

95% ELLIPSE -

Lengths of the major and minor axes of the ellipse and the angle that the major axis makes with the horizontal axis.

AVG R. SDR -

Average of the R. SDR for the two samples; an indication of the laboratory's precision of repeated measurements.

Graph -

For each laboratory the MEAN for the second sample is plotted against the MEAN for the first sample, with each point representing a laboratory. The horizontal and vertical lines are the GRAND MEANS. The dashed line is drawn at 45°. The solid sloping line, which may or may not lie close to the 45° line, is along the major axis of the error ellipse. The ellipse is drawn so that, on the average, it will include 95% of the points representing the laboratories.

Plotted symbols are as explained above (under F), except that an 'S' is plotted as an 'O'. A participant whose plotted point falls outside of the ellipse should carefully reexamine the testing procedure he is following.

The graph is plotted with an ellipse when there are 20 or more laboratories in the analysis. When there are 10 through 19 laboratories in the analysis the graph is plotted but the ellipse is omitted. When there are fewer than 10 laboratories retained in the analysis the graph is not plotted.

The International System of Units (SI) is used on the plots wherever possible to aid participants in familiarizing themselves with SI. Grand means in SI units are given at the top of the plot, and supplementary scales in SI units are drawn along the axes allowing the reader to compare means and variability in common units and SI units for the same data.

Summary -
(At end of
report)

In addition to several quantities already defined above, the summary shows the following values for each test method:

REPL CRP -

The number of replicate test determinations used in this Collaborative Reference Program.

REPL TAPPI -

The number of replicate test determinations in a test result required by the applicable TAPPI Standard or assumed here if there is no TAPPI Standard. This quantity is needed in the computation of TAPPI repeatability and reproducibility from the SD OF MEANS and the AVER SDR. See TAPPI Standard T1206 for definitions and computations.

REPEAT -

TAPPI repeatability, a measure of the within-laboratory precision of a test result.

REPROD -

TAPPI reproducibility, a measure of the between-laboratory precision of a test result.

Best values -

Given at the end of Table 1 for each method for which sufficient information is available. These best values are estimates based on a careful examination of all data, both current and past, with special attention to results obtained by the National Bureau of Standards and other recognized reference laboratories in this and other countries. All participants using equipment that is standard for the analysis should be able to achieve results within the plus-minus (+) limits, when these are shown along with the best values.

TAPPI COLLABORATIVE REFERENCE PROGRAM
ANALYSIS T40-1 TABLE I
AIR RESISTANCE, GURLEY UNITS (SECONDS/100 CC)
TAPPI STANDARD T460 GS-75, AIR RESISTANCE OF PAPER

LAB CODE	SAMPLE K43	PRINTING					SAMPLE B95	HEAT SET OFFSET BOOK					TEST D _e = 10		
		MEAN	DEV	N _e DEV	SDR	R _e SDR		MEAN	DEV	N _e DEV	SDR	R _e SDR	VAR	F	LAB
L100	14.0	-2.3	-2.69	1.2	1.05	38.9	-9.7	-3.03	2.6	.49	40D	* L100			
L106	20.0	3.7	4.24	6.9	6.23	44.8	-3.8	-1.18	3.9	.75	40D	X L106			
L107	15.5	-0.8	-0.97	0.8	.77	48.3	-3	-0.08	2.1	.39	40D	0 L107			
L121	16.4	.0	.04	.8	.72	48.2	-3	-0.10	4.8	.92	40D	0 L121			
L122	16.8	.5	.56	1.4	1.25	45.1	-3.4	-1.08	4.9	.93	40D	0 L122			
L123	14.8	-1.5	-1.73	1.0	.95	45.5	-3.0	-0.95	5.7	1.10	40D	0 L123			
L124G	15.9	-0.4	-0.48	.7	.61	44.9	-3.7	-1.15	3.5	.67	40D	0 L124G			
L125	16.7	.3	.38	.9	.82	49.5	1.0	.31	2.8	.53	40D	0 L125			
L128	17.4	1.1	1.23	1.3	1.23	49.5	.9	.29	3.2	.61	40D	0 L128			
L141	17.0	.6	.75	1.1	.96	47.9	-0.7	-0.21	5.6	1.06	40D	0 L141			
L148	17.5	1.1	1.30	.8	.74	45.6	-3.0	-0.94	4.9	.93	40D	* L148			
L153	16.3	-0.0	-0.04	1.7	1.54	45.5	-3.1	-0.96	4.3	.82	40D	0 L153			
L158	14.3	-2.0	-2.36	.8	.75	45.1	-3.5	-1.09	6.3	1.21	40D	0 L158			
L159	17.2	.5	1.00	1.1	1.03	51.0	2.4	.75	6.7	1.28	40D	0 L159			
L163	17.0	.7	.80	1.2	1.06	52.5	3.9	1.23	3.1	.60	40D	0 L163			
L166	17.5	1.2	1.35	1.3	1.19	50.5	1.9	.61	6.0	1.16	40D	0 L166			
L174	16.3	-0.1	-0.09	1.0	.87	47.2	-1.4	-0.43	8.6	1.64	40D	0 L174			
L182G	15.8	-0.6	-0.64	1.0	.91	47.3	-1.3	-0.40	5.9	1.12	40D	0 L182G			
L183	17.1	.7	.84	.8	.76	48.3	-0.3	-0.08	4.7	.90	40D	0 L183			
L190C	16.5	.2	.24	1.1	.96	50.6	2.0	.64	4.5	.85	40D	0 L190C			
L212	14.2	-2.1	-2.43	1.2	1.06	42.3	-6.3	-1.97	4.8	.91	40D	0 L212			
L223	16.9	.6	.70	1.1	.99	51.1	2.5	.80	4.5	.87	40D	0 L223			
L230G	16.4	.1	.07	.5	.47	49.2	.6	.20	6.1	1.16	40D	0 L230G			
L232	16.6	.3	.33	.6	.50	35.6	-13.0	-4.08	12.5	2.38	40D	* L232			
L236	16.7	.4	.48	.9	.82	45.3	-3.2	-1.01	4.9	.94	40D	0 L236			
L238A	17.0	.7	.83	.9	.83	50.8	2.3	.71	5.4	1.04	40D	0 L238A			
L241	15.9	-0.4	-0.50	1.2	1.09	45.5	-3.1	-0.96	6.1	1.16	40D	0 L241			
L242	16.0	-0.3	-0.35	.6	.59	49.5	.9	.29	5.7	1.09	40D	0 L242			
L243G	16.3	-0.0	-0.01	.4	.36	46.9	-1.7	-0.52	6.2	1.19	40D	0 L243G			
L254	16.6	.3	.31	.8	.77	50.1	1.5	.48	6.1	1.17	40D	0 L254			
L259	15.7	-0.6	-0.74	1.7	1.51	46.3	-2.3	-0.71	4.9	.95	40D	0 L259			
L261	16.2	-0.2	-0.18	1.0	.92	50.5	2.0	.62	5.0	.96	40D	0 L261			
L262G	16.6	.3	.31	1.5	1.33	47.1	-1.5	-0.46	3.3	.62	40D	0 L262G			
L265	15.7	-0.6	-0.72	1.2	1.08	47.0	-1.5	-0.48	2.4	.46	40D	0 L265			
L278	16.7	.4	.48	1.3	1.22	48.3	-0.3	-0.09	3.7	.71	40D	0 L278			
L285	16.7	.4	.43	.7	.60	50.0	1.4	.44	4.2	.80	40D	0 L285			
L301	16.6	.3	.33	1.4	1.26	48.8	.2	.07	5.6	1.08	40D	0 L301			
L308	17.0	.6	.73	1.2	1.07	56.4	7.8	2.46	6.1	1.17	40D	* L308			
L320	14.2	-2.1	-2.47	1.4	1.27	43.2	-5.4	-1.68	5.9	1.13	40D	0 L320			
L321	16.8	.4	.50	1.5	1.32	50.4	1.8	.58	4.8	.93	40D	0 L321			
L324	16.1	-0.2	-0.27	1.2	1.06	48.4	-0.2	-0.06	7.5	1.44	40D	0 L324			
L326	17.6	1.3	1.50	.9	.86	52.7	4.1	1.30	5.5	1.05	40D	0 L326			
L328	16.7	.4	.42	1.0	.92	54.0	5.5	1.72	9.7	1.85	40D	0 L328			
L339	11.9	-4.4	-5.09	.7	.62	35.5	-13.0	-4.09	4.2	.81	40D	* L339			
L344	15.9	-0.4	-0.50	1.9	1.68	52.6	4.0	1.27	12.3	2.34	40D	0 L344			
L376	16.9	.6	.64	1.1	1.01	52.7	4.1	1.29	3.6	.69	40D	0 L376			
L388	17.2	.9	1.00	1.4	1.28	54.7	6.1	1.92	5.0	.96	40D	0 L388			
L394	16.1	-0.2	-0.27	.9	.80	49.4	.8	.26	4.5	.86	40D	0 L394			
L396M	15.6	-0.7	-0.79	1.2	1.09	46.7	-1.9	-0.59	5.0	.96	40D	0 L396M			
L567	16.2	-0.2	-0.18	1.1	1.04	45.9	-2.7	-0.84	4.2	.81	40D	0 L567			
L576	15.3	-1.0	-1.18	1.0	.87	47.8	-.8	-0.24	4.3	.82	40D	0 L576			
L585	17.2	.8	.95	1.2	1.05	52.2	3.7	1.16	7.2	1.37	40D	0 L585			
L597	17.2	.9	1.00	.9	.83	50.3	1.7	.55	4.7	.89	40D	0 L597			
L604	16.0	-0.3	-0.39	1.5	1.35	47.8	-.8	-0.24	5.8	1.10	40D	0 L604			
L651	15.8	-0.5	-0.62	1.1	1.03	48.4	-.2	-0.05	4.2	.80	40D	0 L651			
L676	17.6	1.2	1.42	1.1	1.04	51.5	3.0	.94	5.4	1.02	40D	0 L676			
L697	16.4	.0	.03	1.4	1.23	47.0	-1.6	-0.49	5.6	1.07	40D	0 L697			

GR. MEAN = 16.3 GURLEY UNITS GRAND MEAN = 48.6 GURLEY UNITS TEST DETERMINATIONS = 10
 SD MEANS = .9 GURLEY UNITS SD OF MEANS = 3.2 GURLEY UNITS 54 LABS IN GRAND MEANS
 AVERAGE SDR = 1.1 GURLEY UNITS AVERAGE SDR = 5.2 GURLEY UNITS

L115 15.6 -0.7 -.85 1.1 .98 46.9 -1.7 -.52 4.3 .82 40U * L115
 L291 17.3 1.0 1.12 1.3 1.21 44.0 -4.6 -1.43 4.1 .79 40U * L291
 TOTAL NUMBER OF LABORATORIES REPORTING = 59

Best values: K43 16.3 ± 1.4 Gurley units
 B95 48.5 ± 5.5 Gurley units


TAPPI COLLABORATIVE REFERENCE PROGRAM
ANALYSIS T40-1 TABLE 2
AIR RESISTANCE, GURLEY UNITS (SECONDS/100 CC)
TAPPI STANDARD T460 GS-75, AIR RESISTANCE OF PAPER

APRIL 1979

LAB CODE	F	MEANS K43	MEANS B95	COORDINATES MAJOR	COORDINATES MINOR	Avg R. SDR VAR	PROPERTY---TEST INSTRUMENT---CONDITIONS
L339	#	11.9	35.5	-13.6	1.9	.71	40D AIR RESISTANCE, GURLEY DENSOMETER - GIL FLOATATION
L100	*	14.0	38.9	-9.9	.5	.77	40D AIR RESISTANCE, GURLEY DENSOMETER - GIL FLOATATION
L320	G	14.2	43.2	-5.7	1.1	1.20	40D AIR RESISTANCE, GURLEY DENSOMETER - GIL FLOATATION
L212	G	14.2	42.3	-6.5	.9	.99	40D AIR RESISTANCE, GURLEY DENSOMETER - GIL FLOATATION
L158	G	14.3	45.1	-3.8	1.3	.98	40D AIR RESISTANCE, GURLEY DENSOMETER - GIL FLOATATION
L123	G	14.8	45.5	-3.3	.9	1.02	40D AIR RESISTANCE, GURLEY DENSOMETER - GIL FLOATATION
L576	G	15.3	47.8	-9.9	.9	.84	40D AIR RESISTANCE, GURLEY DENSOMETER - GIL FLOATATION
L107	G	15.5	48.3	-4	.8	.58	40D AIR RESISTANCE, GURLEY DENSOMETER - GIL FLOATATION
L115	*	15.6	46.9	-1.8	.4	.90	40U AIR RESISTANCE, SHEFFIELD IN GURLEY UNITS
L396M	G	15.6	46.7	-2.0	.3	1.03	40D AIR RESISTANCE, GURLEY DENSOMETER - GIL FLOATATION
L259	G	15.7	46.3	-2.3	.2	1.23	40D AIR RESISTANCE, GURLEY DENSOMETER - GIL FLOATATION
L265	G	15.7	47.0	-1.6	.3	.77	40D AIR RESISTANCE, GURLEY DENSOMETER - GIL FLOATATION
L182G	G	15.8	47.3	-1.3	.3	1.01	40D AIR RESISTANCE, GURLEY DENSOMETER - GIL FLOATATION
L651	G	15.8	48.4	-3	.5	.92	40D AIR RESISTANCE, GURLEY DENSOMETER - GIL FLOATATION
L344	G	15.9	52.6	3.9	1.2	2.01	40D AIR RESISTANCE, GURLEY DENSOMETER - GIL FLOATATION
L241	G	15.9	45.5	-3.1	.2	1.12	40D AIR RESISTANCE, GURLEY DENSOMETER - GIL FLOATATION
L124G	G	15.9	44.9	-3.7	.3	.64	40D AIR RESISTANCE, GURLEY DENSOMETER - GIL FLOATATION
L604	G	16.0	47.8	-8	.2	1.23	40D AIR RESISTANCE, GURLEY DENSOMETER - GIL FLOATATION
L242	G	16.0	49.5	.9	.5	.84	40D AIR RESISTANCE, GURLEY DENSOMETER - GIL FLOATATION
L324	G	16.1	48.4	-2	.2	1.25	40D AIR RESISTANCE, GURLEY DENSOMETER - GIL FLOATATION
L394	G	16.1	49.4	.8	.4	.83	40D AIR RESISTANCE, GURLEY DENSOMETER - GIL FLOATATION
L567	G	16.2	45.9	-2.7	.4	.92	40D AIR RESISTANCE, GURLEY DENSOMETER - GIL FLOATATION
L261	G	16.2	50.5	1.9	.5	.94	40D AIR RESISTANCE, GURLEY DENSOMETER - GIL FLOATATION
L174	G	16.3	47.2	-1.4	.2	1.25	40D AIR RESISTANCE, GURLEY DENSOMETER - GIL FLOATATION
L153	G	16.3	45.5	-3.0	.5	1.18	40D AIR RESISTANCE, GURLEY DENSOMETER - GIL FLOATATION
L243G	G	16.3	46.9	-1.6	.3	.78	40D AIR RESISTANCE, GURLEY DENSOMETER - GIL FLOATATION
L697	G	16.4	47.0	-1.5	.3	1.15	40D AIR RESISTANCE, GURLEY DENSOMETER - GIL FLOATATION
L121	G	16.4	48.2	-3	.1	.82	40D AIR RESISTANCE, GURLEY DENSOMETER - GIL FLOATATION
L230G	G	16.4	49.2	.6	.1	.82	40D AIR RESISTANCE, GURLEY DENSOMETER - GIL FLOATATION
L190C	G	16.5	50.6	2.0	.2	.90	40D AIR RESISTANCE, GURLEY DENSOMETER - GIL FLOATATION
L262G	G	16.6	47.1	-1.4	.5	.98	40D AIR RESISTANCE, GURLEY DENSOMETER - GIL FLOATATION
L254	G	16.6	50.1	1.6	.0	.97	40D AIR RESISTANCE, GURLEY DENSOMETER - GIL FLOATATION
L232	#	16.6	35.6	-12.7	-2.7	1.44	40D AIR RESISTANCE, GURLEY DENSOMETER - GIL FLOATATION
L301	G	16.6	48.8	.3	.2	1.17	40D AIR RESISTANCE, GURLEY DENSOMETER - GIL FLOATATION
L125	G	16.7	49.5	1.0	.1	.67	40D AIR RESISTANCE, GURLEY DENSOMETER - GIL FLOATATION
L328	G	16.7	54.0	5.4	.7	1.39	40D AIR RESISTANCE, GURLEY DENSOMETER - GIL FLOATATION
L285	G	16.7	50.0	1.5	.1	.70	40D AIR RESISTANCE, GURLEY DENSOMETER - GIL FLOATATION
L236	G	16.7	45.3	-3.1	-1.0	.88	40D AIR RESISTANCE, GURLEY DENSOMETER - GIL FLOATATION
L278	G	16.7	48.3	-2	.5	.97	40D AIR RESISTANCE, GURLEY DENSOMETER - GIL FLOATATION
L321	G	16.8	50.4	1.9	.1	1.12	40D AIR RESISTANCE, GURLEY DENSOMETER - GIL FLOATATION
L122	G	16.8	45.1	-3.3	-1.1	1.09	40D AIR RESISTANCE, GURLEY DENSOMETER - GIL FLOATATION
L376	G	16.9	52.7	4.1	.2	.85	40D AIR RESISTANCE, GURLEY DENSOMETER - GIL FLOATATION
L223	G	16.9	51.1	2.6	.1	.93	40D AIR RESISTANCE, GURLEY DENSOMETER - GIL FLOATATION
L308	*	17.0	56.4	7.8	.9	1.12	40D AIR RESISTANCE, GURLEY DENSOMETER - GIL FLOATATION
L141	G	17.0	47.9	-5	.8	1.01	40D AIR RESISTANCE, GURLEY DENSOMETER - GIL FLOATATION
L163	G	17.0	52.5	4.0	.1	.83	40D AIR RESISTANCE, GURLEY DENSOMETER - GIL FLOATATION
L238A	G	17.0	50.8	2.4	.3	.93	40D AIR RESISTANCE, GURLEY DENSOMETER - GIL FLOATATION
L183	G	17.1	48.3	-1	.8	.83	40D AIR RESISTANCE, GURLEY DENSOMETER - GIL FLOATATION
L585	G	17.2	52.2	3.8	.1	1.21	40D AIR RESISTANCE, GURLEY DENSOMETER - GIL FLOATATION
L159	G	17.2	51.0	2.5	.4	1.15	40D AIR RESISTANCE, GURLEY DENSOMETER - GIL FLOATATION
L388	G	17.2	54.7	6.2	.3	1.12	40D AIR RESISTANCE, GURLEY DENSOMETER - GIL FLOATATION
L597	G	17.2	50.3	1.9	.5	.86	40D AIR RESISTANCE, GURLEY DENSOMETER - GIL FLOATATION
L291	*	17.3	44.0	-4.3	-1.8	1.00	40U AIR RESISTANCE, SHEFFIELD IN GURLEY UNITS
L128	G	17.4	49.5	1.1	.9	.92	40D AIR RESISTANCE, GURLEY DENSOMETER - GIL FLOATATION
L148	*	17.5	45.6	-2.7	-1.7	.84	40D AIR RESISTANCE, GURLEY DENSOMETER - GIL FLOATATION
L166	G	17.5	50.5	2.1	.8	1.17	40D AIR RESISTANCE, GURLEY DENSOMETER - GIL FLOATATION
L676	G	17.6	51.5	3.2	.6	1.03	40D AIR RESISTANCE, GURLEY DENSOMETER - GIL FLOATATION
L326	G	17.6	52.7	4.3	.5	.95	40D AIR RESISTANCE, GURLEY DENSOMETER - GIL FLOATATION
L106	X	20.0	44.8	-3.0	-4.3	3.49	40D AIR RESISTANCE, GURLEY DENSOMETER - GIL FLOATATION
GMEANS:		16.3	48.6		1.00		
95% ELLIPSE:		8.2	1.6		WITH GAMMA = 79 DEGREES		

AIR RESISTANCE, GURLEY

SAMPLE K43 = 16.3 GURLEY UNITS SAMPLE B95 = 48.6 GURLEY UNITS



ANALYSIS T40-2 TABLE I
AIR RESISTANCE, SHEFFIELD UNITS (CC/MIN) FOR 0.442 SQ. IN (3/4 IN. DIA) GRIFFICE
SHEFFIELD TESTER IS STANDARD FOR THIS ANALYSIS

LAB CODE	SAMPLE K43 MEAN	PRINTING				SAMPLE B95 MEAN	HEAT SET OFFSET BOOK				TEST D. = 10		
		60 GRAMS DEV	N. DEV	SDR	R. SDR		91 GRAMS PER SQUARE METER	DEV	N. DEV	SDR	R. SDR	VAR	F
L114	178.6	12.1	1.88	9.4	1.07	77.3	5.1	1.18	3.9	.65	40S	G	L114
L121	171.0	4.5	.70	9.2	1.05	94.0	21.8	5.07	4.7	.79	40S	X	L121
L122S	164.8	-1.7	-1.26	8.2	.94	79.7	7.5	1.74	5.7	.95	40S	G	L122S
L124S	158.8	-7.7	-1.20	9.7	1.11	70.9	-1.3	-1.31	7.7	1.29	40S	G	L124S
L132	165.4	-1.1	-1.17	5.2	.59	68.3	-3.9	-1.91	7.2	1.21	40S	G	L132
L148	171.9	5.4	.84	6.4	.73	77.3	5.1	1.18	6.6	1.11	40S	G	L148
L150	184.4	17.5	2.78	11.0	1.26	74.3	2.1	.48	9.3	1.56	40S	#	L150
L157	163.0	-3.5	-1.54	11.7	1.33	68.0	-4.2	-1.98	4.9	.83	40S	G	L157
L158	164.0	-2.5	-1.39	7.4	.84	76.0	3.8	.88	3.9	.66	40S	G	L158
L173B	166.5	.0	.00	9.1	1.04	70.0	-2.2	-1.52	8.8	1.48	40S	G	L173B
L190C	178.4	11.9	1.85	5.3	.61	81.9	9.7	2.25	6.4	1.07	40S	G	L190C
L213	167.2	.7	.11	9.3	1.06	79.0	6.8	1.58	5.7	.96	40S	G	L213
L223	166.0	-.5	-1.08	9.0	1.03	69.6	-2.6	-1.61	5.0	.84	40S	G	L223
L228	206.8	40.3	6.27	9.7	1.10	131.2	59.0	13.72	11.8	1.98	40S	#	L228
L230S	160.6	-5.9	-1.92	7.2	.82	69.0	-3.2	-1.75	6.9	1.16	40S	G	L230S
L233	166.2	-.3	-1.05	11.9	1.36	73.4	1.2	.27	3.2	.53	40S	G	L233
L241	197.3	30.8	4.79	20.3	2.32	87.5	15.3	3.55	4.3	.73	40S	#	L241
L249	166.6	.1	.02	12.8	1.46	70.2	-2.0	-1.47	7.2	1.21	40S	G	L249
L255	183.8	17.3	2.69	6.2	.71	96.5	24.3	5.65	5.3	.90	40S	#	L255
L257A	162.7	-3.8	-1.59	7.9	.90	73.8	1.6	.37	4.0	.68	40S	G	L257A
L257B	162.1	-4.4	-1.68	9.5	1.09	64.7	-7.5	-1.75	7.5	1.25	40S	G	L257B
L257C	168.9	2.4	.37	5.6	.64	77.7	5.5	1.27	3.9	.65	40S	G	L257C
L260	168.8	2.3	.36	3.9	.44	72.7	.5	.11	6.7	1.12	40S	G	L260
L262S	167.8	1.3	.20	5.5	.63	72.2	-0.0	-1.01	7.4	1.24	40S	G	L262S
L288	168.1	1.6	.25	16.5	1.89	73.7	1.5	.34	5.3	.90	40S	G	L288
L301	163.5	-3.0	-1.47	7.5	.86	71.9	-.3	-1.08	4.3	.72	40S	G	L301
L305	159.5	-7.0	-1.09	6.4	.73	73.2	1.0	.23	5.7	.95	40S	G	L305
L318	157.4	-9.1	-1.41	5.7	.66	67.4	-4.8	-1.12	5.7	.95	40S	G	L318
L352	166.2	-.3	-1.05	5.7	.65	78.4	6.2	1.44	5.2	.87	40S	G	L352
L354	162.5	-4.0	-1.62	12.3	1.40	67.1	-5.1	-1.19	5.7	.96	40S	G	L354
L360	162.0	-4.5	-1.70	5.4	.61	70.9	-1.3	-1.31	4.6	.77	40S	G	L360
L372	168.5	2.0	.31	8.5	.97	72.6	.4	.09	6.6	1.10	40S	G	L372
L390	201.0	34.5	5.36	13.3	1.52	92.0	19.8	4.60	5.9	.99	40S	#	L390
L562	177.6	11.1	1.73	11.2	1.28	69.6	-2.6	-1.61	6.3	1.07	40S	G	L562
L575	169.8	3.3	.51	9.2	1.05	71.6	-.6	-1.15	6.2	1.04	40S	G	L575
L585	166.5	.0	.00	9.7	1.11	69.0	-3.2	-1.75	6.6	1.11	40S	G	L585
L597	165.0	-1.5	-1.23	9.7	1.11	67.2	-5.0	-1.17	8.4	1.42	40S	G	L597
L600	165.9	-.6	-1.09	6.2	.71	78.1	5.9	1.37	10.7	1.80	40S	G	L600
L626	151.5	-15.0	-2.33	10.6	1.21	67.5	-4.7	-1.10	4.2	.71	40S	G	L626
L684	173.4	6.9	1.07	11.3	1.29	66.6	-5.6	-1.31	5.0	.83	40S	G	L684
L687	171.8	5.3	.83	10.5	1.20	73.9	1.7	.39	3.6	.60	40S	G	L687
L698	158.3	-8.2	-1.27	13.7	1.56	67.7	-4.5	-1.05	7.7	1.29	40S	G	L698

GR. MEAN = 166.5 SHEFF. UNITS

SD MEANS = 6.4 SHEFF. UNITS

AVERAGE SDR = 8.8 SHEFF. UNITS

GRAND MEAN = 72.2 SHEFF. UNITS

SD OF MEANS = 4.3 SHEFF. UNITS

AVERAGE SDR = 6.0 SHEFF. UNITS

TEST DETERMINATIONS = 10

37 LABS IN GRAND MEANS

TOTAL NUMBER OF LABORATORIES REPORTING = 45

Best values: K43 166 + 12 Sheffield units
B95 72 + 7 Sheffield units


The following laboratories were omitted from the grand means because of extreme test results: 228, 241, 255, 390.

AIR RESISTANCE, SHEFFIELD UNITS (CC/MIN) FOR 0.442 SQ. IN (3/4 IN. DIA) ORIFICE
SHEFFIELD TESTER IS STANDARD FOR THIS ANALYSIS

LAB CODE	F	MEANS E43	MEANS B95	COORDINATES MAJOR	COORDINATES MINOR	AVG R, SDR VAR	PROPERTY---TEST INSTRUMENT---CONDITIONS
L626	Ø	151.5	67.5	-15.6	1.5	.96 40S AIR RESISTANCE,	SHEFFIELD (3/4 INCH DIAMETER ORIFICE)
L318	Ø	157.4	67.4	-10.3	-0.9	.80 40S AIR RESISTANCE,	SHEFFIELD (3/4 INCH DIAMETER ORIFICE)
L698	Ø	158.3	67.7	-9.3	-0.9	1.43 40S AIR RESISTANCE,	SHEFFIELD (3/4 INCH DIAMETER ORIFICE)
L124S	Ø	158.8	70.9	-7.6	1.8	1.20 40S AIR RESISTANCE,	SHEFFIELD (3/4 INCH DIAMETER ORIFICE)
L305	Ø	159.5	73.2	-6.0	3.6	.84 40S AIR RESISTANCE,	SHEFFIELD (3/4 INCH DIAMETER ORIFICE)
L230S	Ø	160.6	69.0	-6.7	-0.7	.99 40S AIR RESISTANCE,	SHEFFIELD (3/4 INCH DIAMETER ORIFICE)
L360	Ø	162.0	70.9	-4.7	.5	.69 40S AIR RESISTANCE,	SHEFFIELD (3/4 INCH DIAMETER ORIFICE)
L257B	Ø	162.1	64.7	-7.0	-5.2	1.17 40S AIR RESISTANCE,	SHEFFIELD (3/4 INCH DIAMETER ORIFICE)
L354	Ø	162.5	67.1	-5.7	-3.1	1.18 40S AIR RESISTANCE,	SHEFFIELD (3/4 INCH DIAMETER ORIFICE)
L257A	Ø	162.7	73.8	-2.9	2.9	.79 40S AIR RESISTANCE,	SHEFFIELD (3/4 INCH DIAMETER ORIFICE)
L157	Ø	163.0	68.0	-4.9	-2.5	1.08 40S AIR RESISTANCE,	SHEFFIELD (3/4 INCH DIAMETER ORIFICE)
L301	Ø	163.5	71.9	-2.9	.9	.79 40S AIR RESISTANCE,	SHEFFIELD (3/4 INCH DIAMETER ORIFICE)
L158	Ø	164.0	76.0	-8	4.4	.75 40S AIR RESISTANCE,	SHEFFIELD (3/4 INCH DIAMETER ORIFICE)
L122S	Ø	164.8	79.7	1.4	7.5	.94 40S AIR RESISTANCE,	SHEFFIELD (3/4 INCH DIAMETER ORIFICE)
L597	Ø	165.0	67.2	-3.3	-4.0	1.26 40S AIR RESISTANCE,	SHEFFIELD (3/4 INCH DIAMETER ORIFICE)
L132	Ø	165.4	68.3	-2.5	-3.2	.90 40S AIR RESISTANCE,	SHEFFIELD (3/4 INCH DIAMETER ORIFICE)
L600	Ø	165.9	78.1	1.8	5.6	1.25 40S AIR RESISTANCE,	SHEFFIELD (3/4 INCH DIAMETER ORIFICE)
L223	Ø	166.0	69.6	-1.5	-2.2	.93 40S AIR RESISTANCE,	SHEFFIELD (3/4 INCH DIAMETER ORIFICE)
L312	+	166.2	79.6	2.6	6.9	.44 40T AIR RESISTANCE,	SHEFFIELD (3 INCH DIAMETER ORIFICE)
L352	Ø	166.2	78.4	2.2	5.8	.76 40S AIR RESISTANCE,	SHEFFIELD (3/4 INCH DIAMETER ORIFICE)
L233	Ø	166.2	73.4	.2	1.2	.95 40S AIR RESISTANCE,	SHEFFIELD (3/4 INCH DIAMETER ORIFICE)
L585	Ø	166.5	69.0	-1.3	-3.0	1.11 40S AIR RESISTANCE,	SHEFFIELD (3/4 INCH DIAMETER ORIFICE)
L173B	Ø	166.5	70.0	-0.9	-2.1	1.26 40S AIR RESISTANCE,	SHEFFIELD (3/4 INCH DIAMETER ORIFICE)
L249	Ø	166.6	70.2	-0.7	-1.9	1.34 40S AIR RESISTANCE,	SHEFFIELD (3/4 INCH DIAMETER ORIFICE)
L213	Ø	167.2	79.0	3.3	6.0	1.01 40S AIR RESISTANCE,	SHEFFIELD (3/4 INCH DIAMETER ORIFICE)
L262S	Ø	167.8	72.2	1.2	-0.5	.93 40S AIR RESISTANCE,	SHEFFIELD (3/4 INCH DIAMETER ORIFICE)
L288	Ø	168.1	73.7	2.1	.7	1.39 40S AIR RESISTANCE,	SHEFFIELD (3/4 INCH DIAMETER ORIFICE)
L372	Ø	168.5	72.6	2.0	-0.4	1.04 40S AIR RESISTANCE,	SHEFFIELD (3/4 INCH DIAMETER ORIFICE)
L260	Ø	168.8	72.7	2.3	-0.5	.78 40S AIR RESISTANCE,	SHEFFIELD (3/4 INCH DIAMETER ORIFICE)
L257C	Ø	168.9	77.7	4.4	4.1	.64 40S AIR RESISTANCE,	SHEFFIELD (3/4 INCH DIAMETER ORIFICE)
L575	Ø	169.8	71.6	2.8	-1.9	1.05 40S AIR RESISTANCE,	SHEFFIELD (3/4 INCH DIAMETER ORIFICE)
L121	X	171.0	94.0	12.7	18.3	.92 40S AIR RESISTANCE,	SHEFFIELD (3/4 INCH DIAMETER ORIFICE)
L687	Ø	171.8	73.9	5.5	-0.5	.90 40S AIR RESISTANCE,	SHEFFIELD (3/4 INCH DIAMETER ORIFICE)
L148	Ø	171.9	77.3	7.0	2.5	.92 40S AIR RESISTANCE,	SHEFFIELD (3/4 INCH DIAMETER ORIFICE)
L684	Ø	173.4	66.6	4.1	-7.9	1.06 40S AIR RESISTANCE,	SHEFFIELD (3/4 INCH DIAMETER ORIFICE)
L562	Ø	177.6	69.6	9.2	-6.8	1.17 40S AIR RESISTANCE,	SHEFFIELD (3/4 INCH DIAMETER ORIFICE)
L190C	Ø	178.4	81.9	14.7	4.2	.84 40S AIR RESISTANCE,	SHEFFIELD (3/4 INCH DIAMETER ORIFICE)
L114	Ø	178.6	77.3	13.1	-0.1	.86 40S AIR RESISTANCE,	SHEFFIELD (3/4 INCH DIAMETER ORIFICE)
L255	#	183.8	96.5	25.4	15.5	.80 40S AIR RESISTANCE,	SHEFFIELD (3/4 INCH DIAMETER ORIFICE)
L150	#	184.4	74.3	17.3	-5.1	1.41 40S AIR RESISTANCE,	SHEFFIELD (3/4 INCH DIAMETER ORIFICE)
L241	#	197.3	87.5	34.3	1.9	1.52 40S AIR RESISTANCE,	SHEFFIELD (3/4 INCH DIAMETER ORIFICE)
L390	#	201.0	92.0	39.5	4.6	1.25 40S AIR RESISTANCE,	SHEFFIELD (3/4 INCH DIAMETER ORIFICE)
L228	#	206.8	131.2	60.2	38.4	1.54 40S AIR RESISTANCE,	SHEFFIELD (3/4 INCH DIAMETER ORIFICE)
L243B	+	647.9	224.9	502.7	-48.6	4.24 40B AIR RESISTANCE,	BENDTSEN, WG 150
L182B	+	755.0	270.0	618.9	-49.2	4.52 40B AIR RESISTANCE,	BENDTSEN, WG 150
GMEANS:		166.5	72.2			1.00	
		95% ELLIPSE:	17.7	9.5		WITH GAMMA = 23 DEGREES	

AIR RESISTANCE. SHEFFIELD

SAMPLE K43 = 166. SHEFF. UNITS SAMPLE B95 = 72. SHEFF. UNITS



LAB CODE	SAMPLE E65 MEAN	BACKING				SAMPLE B73 MEAN	RELEASE BASE				TEST D _e = 10		
		98 GRAMS PER SQUARE METER	DEV	N _e DEV	SDR		116 GRAMS PER SQUARE METER	DEV	N _e DEV	SDR	R _e SDR	VAR	F
L122	861.	78.	1.04	148.	.86	1601.	287.	1.02	874.	1.69	41G	G	L122
L128	814.	32.	.42	139.	.81	1398.	84.	.30	418.	.81	41G	G	L128
L134	769.	-14.	-.18	161.	.93	1449.	136.	.48	637.	1.23	41G	G	L134
L166M	821.	38.	.51	137.	.79	1214.	-100.	-.36	293.	.57	41G	G	L166M
L195	703.	-79.	-1.05	152.	.88	886.	-428.	-1.53	274.	.53	41G	G	L195
L230	706.	-77.	-1.02	224.	1.29	942.	-372.	-1.33	563.	1.09	41G	G	L230
L259	78.	-705.	-9.35	14.	.08	111.	-1203.	-4.29	30.	.06	41G	#	L259
L312	750.	-32.	-.43	176.	1.02	1284.	-30.	-.11	552.	1.07	41G	G	L312
L358	877.	94.	1.25	144.	.83	1516.	202.	.72	693.	1.34	41G	G	L358
L557	773.	-10.	-.13	181.	1.05	1579.	265.	.94	745.	1.44	41G	G	L557
L558	941.	158.	2.10	361.	2.08	1704.	390.	1.39	673.	1.30	41G	G	L558
L574	722.	-60.	-.80	112.	.65	1203.	-111.	-.40	227.	.44	41G	G	L574
L576	726.	-56.	-.74	109.	.63	855.	-459.	-1.64	216.	.42	41G	G	L576
L618	7248.	6465.	85.78	1874.	10.83	13889.	12575.	44.80	5543.	10.73	41G	#	L618
L697	709.	-73.	-.97	207.	1.20	1452.	138.	.49	550.	1.06	41G	G	L697

GR. MEAN = 783. SEC/10 CC

SD MEANS = 75 SEC/10 CC

AVERAGES

TOTAL NUMBER OF LABORATORIES REPORTING - 15


Data from the following laboratories appear to be off by a multiplicative factor: 259, 618

LAB CODE	F	MEANS		COORDINATES		AVG E, SDR	VAR	PROPERTY---TEST INSTRUMENT---CONDITIONS				
		E65	B73	MAJOR	MINOR							
L259	#	78.	111.	-1318.	454.	.07	41G AIR RESISTANCE,	HIGH RANGE,	GURLEY	MERCURY	FLotation	
L195	Ø	703.	886.	-435.	-7.	.70	41G AIR RESISTANCE,	HIGH RANGE,	GURLEY	MERCURY	FLotation	
L230	Ø	706.	942.	-380.	2.	1.19	41G AIR RESISTANCE,	HIGH RANGE,	GURLEY	MERCURY	FLotation	
L697	Ø	709.	1452.	121.	99.	1.13	41G AIR RESISTANCE,	HIGH RANGE,	GURLEY	MERCURY	FLotation	
L574	Ø	722.	1203.	-121.	37.	.54	41G AIR RESISTANCE,	HIGH RANGE,	GURLEY	MERCURY	FLotation	
L576	Ø	726.	855.	-461.	-35.	.52	41G AIR RESISTANCE,	HIGH RANGE,	GURLEY	MERCURY	FLotation	
L312	Ø	750.	1284.	-35.	26.	1.04	41G AIR RESISTANCE,	HIGH RANGE,	GURLEY	MERCURY	FLotation	
L134	Ø	769.	1449.	130.	40.	1.08	41G AIR RESISTANCE,	HIGH RANGE,	GURLEY	MERCURY	FLotation	
L557	Ø	773.	1579.	258.	62.	1.24	41G AIR RESISTANCE,	HIGH RANGE,	GURLEY	MERCURY	FLotation	
L128	Ø	814.	1398.	88.	-15.	.81	41G AIR RESISTANCE,	HIGH RANGE,	GURLEY	MERCURY	FLotation	
L166N	Ø	821.	1214.	-91.	-57.	.68	41G AIR RESISTANCE,	HIGH RANGE,	GURLEY	MERCURY	FLotation	
L122	Ø	861.	1601.	297.	-20.	1.27	41G AIR RESISTANCE,	HIGH RANGE,	GURLEY	MERCURY	FLotation	
L358	Ø	877.	1516.	216.	-53.	1.09	41G AIR RESISTANCE,	HIGH RANGE,	GURLEY	MERCURY	FLotation	
L558	Ø	941.	1704.	413.	-78.	1.69	41G AIR RESISTANCE,	HIGH RANGE,	GURLEY	MERCURY	FLotation	
L618	#	7248.	13889.	13601.	-3865.	10.78	41G AIR RESISTANCE,	HIGH RANGE,	GURLEY	MERCURY	FLotation	
GMEANS:		783.	1314.			1.00						
95% ELLIPSE:		843.	151.			WITH GAMMA = 78 DEGREES						

AIR RESISTANCE, GURLEY HG FLOTATION

SAMPLE E65 = 783. SEC/10 CC

SAMPLE B73 = 1314. SEC/10 CC



TAPPI COLLABORATIVE REFERENCE PROGRAM
 ANALYSIS T44-1 TABLE 1
 SMOOTHNESS, PARKER PRINTSURF

APRIL 1979

LAB CODE	SAMPLE A83	WAVE ENVELOPE PAPER 75 GRAMS PER SQUARE METER				SAMPLE J50	PRINTING 94 GRAMS PER SQUARE METER				TEST D. = 10		
		MEAN	DEV	N. DEV	SDR		MEAN	DEV	N. DEV	SDR	R _e SDR	VAR	F
L122	4.653	.040	.24	.179	1.19	6.079	.278	1.25	.063	.72	44P	G	L122
L182	4.565	-.048	-.28	.170	1.13	5.755	-.046	-.21	.090	1.02	44P	G	L182
L183	4.420	-.193	-1.15	.092	.61	5.760	-.041	-.19	.070	.80	44P	G	L183
L223	4.591	-.022	-.13	.185	1.23	6.015	.214	.96	.101	1.16	44P	G	L223
L288	4.940	.327	1.95	.171	1.14	6.980	1.179	5.29	.155	1.77	44P	#	L288
L317	4.870	.257	1.53	.116	.77	5.730	-.071	-.32	.116	1.33	44P	G	L317
L588	4.420	-.193	-1.15	.215	1.43	5.400	-.401	-1.80	.105	1.21	44P	G	L588
L669	4.770	.157	.94	.095	.63	5.870	.069	.31	.067	.77	44P	G	L669

GR. MEAN = 4.613 MICRONS
 SD MEANS = .168 MICRONS

GRAND MEAN = 5.801 MICRONS
 SD OF MEANS = .223 MICRONS

TEST DETERMINATIONS = 10
 7 LABS IN GRAND MEANS

AVERAGE SDR = .150 MICRONS

AVERAGE SDR = .087 MICRONS

TOTAL NUMBER OF LABORATORIES REPORTING = 8

Best values: A83 4.6 microns
 J50 6.0 microns

The following laboratories were omitted from the
 grand means because of extreme test results: 288.

TAPPI COLLABORATIVE REFERENCE PROGRAM
 ANALYSIS T44-1 TABLE 2
 SMOOTHNESS, PARKER PRINTSURF

APRIL 1979

LAB CODE	F	MEANS		COORDINATES		AVG R _e SDR	VAR	PROPERTY---TEST INSTRUMENT---CONDITIONS
		A83	J50	MAJOR	MINOR			
L183	G	4.420	5.760	-.124	.153	.71	44P	SMOOTHNESS, PARKER PRINTSURF
L588	G	4.420	5.400	-.445	-.009	1.32	44P	SMOOTHNESS, PARKER PRINTSURF
L182	G	4.565	5.755	-.063	.022	1.08	44P	SMOOTHNESS, PARKER PRINTSURF
L223	G	4.591	6.015	.181	.116	1.19	44P	SMOOTHNESS, PARKER PRINTSURF
L122	G	4.653	6.079	.266	.089	.95	44P	SMOOTHNESS, PARKER PRINTSURF
L669	G	4.770	5.870	.132	-.109	.70	44P	SMOOTHNESS, PARKER PRINTSURF
L317	G	4.870	5.730	.052	-.262	1.05	44P	SMOOTHNESS, PARKER PRINTSURF
L288	#	4.940	6.980	1.200	.239	1.46	44P	SMOOTHNESS, PARKER PRINTSURF

GMEANS: 4.613 5.801
 95% ELLIPSE: .889 .540 WITH GAMMA = 63 DEGREES

TAPPI COLLABORATIVE REFERENCE PROGRAM
ANALYSIS T45-1 TABLE 1
SMOOTHNESS, SHEFFIELD UNITS
SHEFFIELD TESTER IS STANDARD FOR THIS ANALYSIS

APRIL 1979

LAB CGDE	SAMPLE A83	WAVE ENVELOPE PAPER				SAMPLE J50	PRINTING				TEST D _e = 15		
		75 GRAMS PER SQUARE METER	N _e DEV	SDR	R _e SDR		MEAN	94 GRAMS PER SQUARE METER	N _e DEV	SDR	R _e SDR	VAR	F
L100	112.7	7.1	.96	14.4	1.48	285.3	10.9	.83	13.7	1.84	45S	G	L100
L107	133.7	28.1	3.80	14.3	1.46	302.7	28.2	2.15	6.8	.91	45S	X	L107
L114	115.7	10.1	1.37	11.6	1.18	272.7	-1.8	-0.14	7.8	1.04	45S	G	L114
L115	107.0	1.4	.20	9.2	.94	259.7	-14.8	-1.13	14.5	1.94	45S	G	L115
L121	90.3	-15.2	-2.06	10.0	1.02	257.8	-16.7	-1.27	4.7	.63	45S	G	L121
L122	115.5	9.9	1.34	16.5	1.69	266.7	-7.8	-0.60	10.3	1.38	45S	G	L122
L123	102.6	-3.0	-0.40	12.6	1.29	278.0	3.5	.27	10.1	1.36	45S	G	L123
L124	91.4	-14.2	-1.91	8.6	.88	260.7	-13.7	-1.05	6.9	.92	45S	G	L124
L125	103.0	-2.6	-0.34	11.5	1.17	259.3	-15.1	-1.16	10.0	1.34	45S	G	L125
L126	107.1	1.5	.20	15.7	1.60	276.4	1.9	.15	7.8	1.04	45S	G	L126
L128	111.7	6.2	.83	11.5	1.18	276.7	2.2	.17	4.9	.66	45S	G	L128
L132	107.0	1.4	.20	19.0	1.94	283.1	8.6	.66	6.4	.85	45S	G	L132
L134	100.7	-6.9	-0.66	8.6	.88	284.0	9.5	.73	8.1	1.08	45S	G	L134
L139S	113.2	7.6	1.03	5.8	.60	267.0	-7.5	-0.57	3.7	.49	45S	G	L139S
L148	114.5	8.9	1.20	5.9	.61	280.7	6.2	.47	6.8	.91	45S	G	L148
L150	103.3	-2.3	-0.31	8.9	.91	292.7	18.3	1.39	6.5	.88	45S	G	L150
L152	120.1	14.6	1.97	8.1	.83	257.5	-17.0	-1.30	4.8	.65	45S	#	L152
L153	109.9	4.3	.58	5.3	.54	293.0	18.5	1.41	5.9	.79	45S	G	L153
L157	111.5	5.9	.80	10.0	1.02	285.7	11.3	.86	5.9	.80	45S	G	L157
L158	105.3	-2	-0.03	12.7	1.30	266.7	-7.8	-0.60	7.0	.94	45S	G	L158
L159	110.3	4.7	.64	10.1	1.03	282.3	7.8	.59	11.8	1.58	45S	G	L159
L162	108.0	2.4	.33	11.8	1.20	277.3	2.9	.22	8.0	1.07	45S	G	L162
L166	94.1	-11.4	-1.54	6.0	.61	261.5	-13.0	-0.99	8.5	1.14	45S	G	L166
L167	108.3	2.7	.37	3.1	.31	261.7	-12.8	-0.98	5.9	.79	45S	G	L167
L173B	98.0	-7.6	-1.02	6.8	.69	256.7	-17.8	-1.36	9.0	1.21	45S	G	L173B
L183S	102.5	-3.0	-0.41	7.7	.79	266.3	-8.2	-0.63	6.5	.87	45S	G	L183S
L190C	113.7	8.1	1.10	10.3	1.05	261.3	-13.1	-1.00	6.4	.86	45S	G	L190C
L195	116.8	11.2	1.52	11.7	1.19	271.8	-2.7	-0.21	8.9	1.20	45S	G	L195
L203	115.3	9.7	1.31	13.4	1.37	276.2	1.7	.13	7.4	1.00	45S	G	L203
L206	114.3	8.8	1.19	11.6	1.19	275.3	.8	.06	5.4	.73	45S	G	L206
L211	100.5	-5.0	-0.68	10.1	1.03	260.5	-14.0	-1.07	5.9	.79	45S	G	L211
L213	93.1	-12.5	-1.69	8.5	.87	255.1	-19.3	-1.48	6.7	.91	45S	G	L213
L223	104.5	-1.1	-0.15	9.1	.93	264.2	-10.3	-0.78	7.2	.97	45S	G	L223
L226B	100.1	-5.5	-0.74	8.5	.87	266.3	-8.2	-0.63	9.5	1.28	45S	G	L226B
L228	109.5	4.0	.54	7.1	.73	287.5	13.1	1.00	9.4	1.27	45S	G	L228
L230S	99.1	-6.5	-0.88	11.4	1.16	284.5	10.1	.77	7.3	.98	45S	G	L230S
L231	106.2	.6	.09	7.7	.78	272.6	-1.9	-0.14	8.1	1.09	45S	G	L231
L232S	104.7	-.9	-0.12	4.8	.49	285.0	10.5	.80	4.6	.62	45S	G	L232S
L233	98.9	-6.6	-0.89	11.0	1.13	280.7	6.2	.47	5.5	.74	45S	G	L233
L237	104.3	-1.2	-0.16	6.2	.64	272.3	-2.1	-0.16	5.3	.71	45S	G	L237
L241	110.3	4.8	.65	10.4	1.07	254.7	-19.8	-1.51	4.8	.65	45S	G	L241
L249	106.9	1.3	.18	12.8	1.31	283.3	8.8	.67	11.3	1.52	45S	G	L249
L254	109.2	3.6	.49	10.2	1.04	271.0	-3.5	-0.27	7.9	1.06	45S	G	L254
L255	114.4	8.8	1.20	6.7	.68	279.9	5.5	.42	5.2	.70	45S	G	L255
L257A	100.9	-4.7	-0.63	6.3	.65	273.5	-.9	-0.07	7.7	1.03	45S	G	L257A
L257B	108.7	3.2	.43	6.4	.66	277.3	2.9	.22	6.5	.87	45S	G	L257B
L257C	108.3	2.7	.37	8.1	.83	267.5	-7.0	-0.54	9.4	1.26	45S	G	L257C
L259	109.2	3.6	.49	9.3	.95	272.3	-2.2	-0.17	11.6	1.55	45S	G	L259
L260	108.6	3.0	.41	10.8	1.11	266.1	-8.3	-0.64	2.7	.36	45S	G	L260
L261	92.9	-12.6	-1.71	8.5	.87	278.3	3.9	.29	6.7	.90	45S	G	L261
L262	104.5	-1.1	-0.15	9.1	.93	270.1	-4.3	-0.33	10.9	1.47	45S	G	L262
L275	102.7	2.9	.39	11.5	1.17	280.7	6.2	.47	5.6	.76	45S	G	L275
L277	104.4	-1.2	-0.16	10.4	1.06	294.2	19.7	1.50	7.7	1.03	45S	G	L277
L278	113.0	7.4	1.01	10.6	1.09	294.0	19.5	1.49	9.7	1.30	45S	G	L278
L281	110.7	5.2	.70	9.0	.92	268.1	-6.3	-0.48	9.0	1.21	45S	G	L281
L285	93.9	-11.7	-1.58	10.6	1.09	277.6	3.1	.24	7.3	.99	45S	G	L285
L288	121.7	16.1	2.18	14.0	1.43	290.8	16.3	1.25	7.9	1.07	45S	G	L288
L290	107.7	2.2	.29	9.1	.93	247.8	-26.7	-2.04	5.5	.74	45S	G	L290
L291S	104.5	-1.1	-0.15	6.2	.64	285.6	11.1	.85	9.4	1.27	45S	G	L291S
L301	99.7	-5.8	-0.79	7.0	.71	276.2	1.7	.13	7.3	.98	45S	G	L301
L305	108.4	2.8	.38	8.7	.89	259.3	-15.2	-1.16	7.4	1.00	45S	G	L305
L308	98.0	-7.6	-1.02	9.7	.99	274.0	-.5	-.04	6.8	.92	45S	G	L308
L312	118.1	12.5	1.69	16.4	1.67	292.1	17.7	1.35	4.4	.60	45S	G	L312
L317	98.0	-7.6	-1.02	8.0	.81	286.3	11.9	.90	7.7	1.03	45S	G	L317
L318	103.3	-2.2	-0.30	7.2	.74	284.4	9.9	.76	10.1	1.36	45S	G	L318

TAPPI COLLABORATIVE REFERENCE PROGRAM
ANALYSIS T45-1 TABLE 2
SMOOTHNESS, SHEFFIELD UNITS
SHEFFIELD TESTER IS STANDARD FOR THIS ANALYSIS


LAB CODE	F	MEANS		COORDINATES		E, SDR VAR	PROPERTY---TEST INSTRUMENT---CONDITIONS
		A83	J50	MAJOR	MINOR		
L121	G	90.3	257.8	-18.3	13.2	.82 45S	SMOOTHNESS, SHEFFIELD
L567	G	91.0	268.2	-7.9	13.7	1.00 45S	SMOOTHNESS, SHEFFIELD
L585	G	91.2	268.0	-8.1	13.5	.99 45S	SMOOTHNESS, SHEFFIELD
L124	G	91.4	260.7	-15.3	12.5	.90 45S	SMOOTHNESS, SHEFFIELD
L698	G	92.0	287.1	11.0	14.9	.90 45S	SMOOTHNESS, SHEFFIELD
L261	G	92.9	278.3	2.4	13.0	.89 45S	SMOOTHNESS, SHEFFIELD
L213	G	93.1	255.1	-20.7	10.2	.89 45S	SMOOTHNESS, SHEFFIELD
L285	G	93.9	277.6	1.8	12.0	1.04 45S	SMOOTHNESS, SHEFFIELD
L166	G	94.1	261.5	-14.2	9.8	.88 45S	SMOOTHNESS, SHEFFIELD
L321	G	95.3	266.7	-8.9	9.3	.73 45S	SMOOTHNESS, SHEFFIELD
L597	G	96.0	272.7	-2.9	9.3	1.17 45S	SMOOTHNESS, SHEFFIELD
L679	G	96.5	246.9	-28.5	5.8	.96 45S	SMOOTHNESS, SHEFFIELD
L317	G	98.0	286.3	10.9	8.9	.92 45S	SMOOTHNESS, SHEFFIELD
L308	G	98.0	274.0	-1.3	7.4	.95 45S	SMOOTHNESS, SHEFFIELD
L173H	G	98.0	256.7	-18.6	5.5	.95 45S	SMOOTHNESS, SHEFFIELD
L233	G	98.9	280.7	5.4	7.3	.93 45S	SMOOTHNESS, SHEFFIELD
L230S	G	99.1	284.5	9.2	7.6	1.07 45S	SMOOTHNESS, SHEFFIELD
L301	G	99.7	276.2	1.0	6.0	.85 45S	SMOOTHNESS, SHEFFIELD
L349	*	99.8	304.2	28.9	9.1	.83 45S	SMOOTHNESS, SHEFFIELD
L226B	G	100.1	266.3	-8.8	4.5	1.07 45S	SMOOTHNESS, SHEFFIELD
L211	G	100.5	260.5	-14.5	3.4	.91 45S	SMOOTHNESS, SHEFFIELD
L134	G	100.7	284.0	8.9	6.0	.98 45S	SMOOTHNESS, SHEFFIELD
L257A	G	100.9	273.5	-1.5	4.5	.84 45S	SMOOTHNESS, SHEFFIELD
L670	G	101.4	288.1	13.0	5.7	.80 45S	SMOOTHNESS, SHEFFIELD
L323	G	101.7	281.7	6.7	4.7	.89 45S	SMOOTHNESS, SHEFFIELD
L360M	G	101.9	261.4	-13.4	2.2	.66 45S	SMOOTHNESS, SHEFFIELD
L382	G	102.0	283.3	8.4	4.6	.89 45S	SMOOTHNESS, SHEFFIELD
L360	G	102.0	276.9	2.0	3.8	.98 45S	SMOOTHNESS, SHEFFIELD
L575	G	102.1	303.2	28.1	6.7	.82 45S	SMOOTHNESS, SHEFFIELD
L183S	G	102.5	266.3	-8.5	2.1	.83 45S	SMOOTHNESS, SHEFFIELD
L123	G	102.6	278.0	3.2	3.3	1.33 45S	SMOOTHNESS, SHEFFIELD
L275	G	102.7	280.7	5.8	3.6	.96 45S	SMOOTHNESS, SHEFFIELD
L125	G	103.0	259.3	-15.3	.8	1.26 45S	SMOOTHNESS, SHEFFIELD
L150	G	103.3	292.7	17.9	4.4	.89 45S	SMOOTHNESS, SHEFFIELD
L318	G	103.3	284.4	9.6	3.3	1.05 45S	SMOOTHNESS, SHEFFIELD
L372	*	104.3	306.9	32.0	5.0	.60 45S	SMOOTHNESS, SHEFFIELD
L237	G	104.3	272.3	-2.3	1.0	.67 45S	SMOOTHNESS, SHEFFIELD
L277	G	104.4	294.2	19.5	3.4	1.04 45S	SMOOTHNESS, SHEFFIELD
L291S	G	104.5	285.6	10.9	2.4	.95 45S	SMOOTHNESS, SHEFFIELD
L262	G	104.5	270.1	-4.4	.6	1.20 45S	SMOOTHNESS, SHEFFIELD
L223	G	104.5	264.2	-10.3	-.1	.95 45S	SMOOTHNESS, SHEFFIELD
L626	G	104.7	277.3	2.7	1.2	1.00 45S	SMOOTHNESS, SHEFFIELD
L554	G	104.7	274.0	-.6	.8	1.59 45S	SMOOTHNESS, SHEFFIELD
L232S	G	104.7	285.0	10.3	2.1	.56 45S	SMOOTHNESS, SHEFFIELD
L158	G	105.3	266.7	-7.8	-.7	1.12 45S	SMOOTHNESS, SHEFFIELD
L688	G	105.5	271.9	-2.5	-.3	1.12 45S	SMOOTHNESS, SHEFFIELD
L231	G	106.2	272.6	-1.8	-.9	.94 45S	SMOOTHNESS, SHEFFIELD
L249	G	106.9	283.3	8.9	-.3	1.41 45S	SMOOTHNESS, SHEFFIELD
L132	G	107.0	283.1	8.7	-.4	1.40 45S	SMOOTHNESS, SHEFFIELD
L326	G	107.0	255.0	-19.2	-.3	.82 45S	SMOOTHNESS, SHEFFIELD
L115	G	107.0	259.7	-14.5	-.3	1.44 45S	SMOOTHNESS, SHEFFIELD
L126	G	107.1	276.4	2.1	-.3	1.32 45S	SMOOTHNESS, SHEFFIELD
L648	G	107.5	275.9	1.7	-.8	1.18 45S	SMOOTHNESS, SHEFFIELD
L290	G	107.7	247.8	-26.2	-.5	.84 45S	SMOOTHNESS, SHEFFIELD
L162	G	108.0	277.3	3.1	-.2	1.14 45S	SMOOTHNESS, SHEFFIELD
L167	G	108.3	261.7	-12.4	-.4	.55 45S	SMOOTHNESS, SHEFFIELD
L257C	G	108.3	267.5	-.6	-.5	1.05 45S	SMOOTHNESS, SHEFFIELD
L305	G	108.4	259.3	-14.8	-.4	.94 45S	SMOOTHNESS, SHEFFIELD
L260	G	108.6	266.1	-.7	-.4	.73 45S	SMOOTHNESS, SHEFFIELD
L257B	G	108.7	277.3	3.2	-.2	.76 45S	SMOOTHNESS, SHEFFIELD
L651	*	108.7	239.3	-34.6	-.7	.75 45S	SMOOTHNESS, SHEFFIELD
L259	G	109.2	272.3	-.1	-.3	1.25 45S	SMOOTHNESS, SHEFFIELD
L254	G	109.2	271.0	-.3	-.4	1.05 45S	SMOOTHNESS, SHEFFIELD
L228	G	109.5	287.5	13.4	-.2	1.00 45S	SMOOTHNESS, SHEFFIELD
L352I	G	109.7	253.1	-20.7	-.6	1.24 45S	SMOOTHNESS, SHEFFIELD

TAPPI COLLABORATIVE REFERENCE PROGRAM
ANALYSIS T45-1 TABLE 2
SMOOTHNESS, SHEFFIELD UNITS
SHEFFIELD TESTER IS STANDARD FOR THIS ANALYSIS

LAB CODE	MEANS F	A83	J50	COORDINATES MAJOR	MINOR	AVG R,SDR VAR	PROPERTY---TEST INSTRUMENT---CONDITIONS
L153	G	109.9	293.0	18.9	-2.1	.67 45S	SMOOTHNESS, SHEFFIELD
L159	G	110.3	282.3	8.3	-3.8	1.31 45S	SMOOTHNESS, SHEFFIELD
L241	G	110.3	254.7	-19.1	-7.0	.86 45S	SMOOTHNESS, SHEFFIELD
L281	G	110.7	268.1	-5.7	-5.9	1.07 45S	SMOOTHNESS, SHEFFIELD
L376	G	111.3	279.9	6.0	-5.1	1.11 45S	SMOOTHNESS, SHEFFIELD
L157	G	111.5	285.7	11.9	-4.6	.91 45S	SMOOTHNESS, SHEFFIELD
L128	G	111.7	276.7	2.9	-5.9	.92 45S	SMOOTHNESS, SHEFFIELD
L100	G	112.7	285.3	11.6	-5.8	1.66 45S	SMOOTHNESS, SHEFFIELD
L278	G	113.0	294.0	20.2	-5.1	1.19 45S	SMOOTHNESS, SHEFFIELD
L139S	G	113.2	267.0	-6.5	-8.5	.55 45S	SMOOTHNESS, SHEFFIELD
L190C	G	113.7	261.3	-12.1	-9.6	.95 45S	SMOOTHNESS, SHEFFIELD
L206	G	114.3	275.3	1.8	-8.6	.96 45S	SMOOTHNESS, SHEFFIELD
L255	G	114.4	279.9	6.4	-8.2	.69 45S	SMOOTHNESS, SHEFFIELD
L148	G	114.5	280.7	7.2	-8.1	.76 45S	SMOOTHNESS, SHEFFIELD
L203	G	115.3	276.2	2.8	-9.5	1.18 45S	SMOOTHNESS, SHEFFIELD
L122	G	115.5	266.7	-6.6	-10.7	1.53 45S	SMOOTHNESS, SHEFFIELD
L114	G	115.7	272.7	-6	-10.3	1.11 45S	SMOOTHNESS, SHEFFIELD
L328	G	115.7	282.5	9.2	-9.2	.78 45S	SMOOTHNESS, SHEFFIELD
L195	G	116.8	271.8	-1.4	-11.5	1.19 45S	SMOOTHNESS, SHEFFIELD
L312	G	118.1	292.1	19.0	-10.4	1.13 45S	SMOOTHNESS, SHEFFIELD
L600	G	119.2	280.3	7.4	-12.9	.98 45S	SMOOTHNESS, SHEFFIELD
L152	*	120.1	257.5	-15.2	-16.4	.74 45S	SMOOTHNESS, SHEFFIELD
L571	G	120.7	296.3	23.5	-12.5	.96 45S	SMOOTHNESS, SHEFFIELD
L288	G	121.7	290.8	18.1	-14.1	1.25 45S	SMOOTHNESS, SHEFFIELD
I390	X	128.7	303.0	31.0	-19.7	1.22 45S	SMOOTHNESS, SHEFFIELD
L107	X	133.7	302.7	31.2	-24.7	1.19 45S	SMOOTHNESS, SHEFFIELD
L174	*	215.9	315.2	53.2	-105.0	.52 45R	SMOOTHNESS, SHEFFIELD, NON-STANDARD INSTRUMENT
GMEANS: 105.6 274.5 1.00 95% ELLIPSE: 33.0 18.3 WITH GAMMA = 83 DEGREES							

SMOOTHNESS, SHEFFIELD

SAMPLE A83 = 106. SHEFF. UNITS SAMPLE J50 = 274. SHEFF. UNITS



TAPPI COLLABORATIVE REFERENCE PROGRAM
ANALYSIS T45-2 TABLE 1

APRIL 1979

TAPPI SUGGESTED METHOD T479 SU-71, SMOOTHNESS OF PAPER (BEKK METHOD)

LAB CODE	SAMPLE A83 MEAN	WAVE ENVELOPE PAPER 75 GRAMS PER SQUARE METER				SAMPLE J50 MEAN	PRINTING 94 GRAMS PER SQUARE METER				TEST D = 15		
		DEV	N. DBV	SDR	R. SDR		DEV	N. DEV	SDR	R. SDR	VAR	F	LAB
L139B	53.60	.75	.20	9.54	1.26	11.13	-.14	-.08	1.19	1.66	45K	G	L139B
L162	51.07	-1.78	-.48	4.71	.62	10.00	-1.27	-.71	.38	.53	45K	G	L162
L182K	54.40	1.55	.42	9.98	1.32	9.79	-1.48	-.83	.83	1.16	45K	G	L182K
L190C	58.47	5.62	1.51	9.82	1.30	11.07	-.20	-.11	.75	1.05	45K	G	L190C
L230B	56.87	4.02	1.08	6.75	.89	9.27	-2.00	-.11	.70	.98	45K	G	L230B
L232B	49.61	-3.24	-.87	5.55	.73	15.56	4.29	2.38	.87	1.22	45K	G	L232B
L243K	50.60	-2.25	-.60	5.22	.69	11.73	.46	.26	.60	.84	45K	G	L243K
L291K	56.85	4.00	1.08	11.67	1.54	12.67	1.40	.78	.99	1.38	45K	G	L291K
L581	49.47	-3.38	-.91	7.28	.96	11.00	-.27	-.15	.38	.53	45K	G	L581
L697	47.56	-5.29	-1.42	5.10	.67	10.49	-.78	-.43	.46	.65	45K	G	L697

GR. MEAN = 52.85 BEKK SECONDS

SD MEANS = 3.72 BEKK SECONDS

AVERAGE SDR = 7.56 BEKK SECONDS

GRAND MEAN = 11.27 BEKK SECONDS

SD OF MEANS = 1.80 BEKK SECONDS

AVERAGE SDR = .71 BEKK SECONDS

TEST DETERMINATIONS = 15

10 LABS IN GRAND MEANS

L250M 44.20 -6.65 -2.32 5.10 .67
TOTAL NUMBER OF LABORATORIES REPORTING = 11Best values: A83 52 Bekk seconds
J50 11 Bekk secondsTAPPI COLLABORATIVE REFERENCE PROGRAM
ANALYSIS T45-2 TABLE 2
SMOOTHNESS, BEKK SECONDS


APRIL 1979

TAPPI SUGGESTED METHOD T479 SU-71, SMOOTHNESS OF PAPER (BEKK METHOD)

LAB CODE	F	MEANS		COORDINATES		AVG R. SDR	VAR	PROPERTY---TEST INSTRUMENT---CONDITIONS		
		A83	J50	MAJOR	MINOR			PROPERTY---TEST INSTRUMENT---CONDITIONS		
L250M	+	44.20	9.60	-8.34	-2.85	.69	45L	SMOOTHNESS, BEKK, 20 C, 65% RH		
L697	G	47.56	10.49	-5.13	-1.50	.66	45K	SMOOTHNESS, BEKK		
L581	G	49.47	11.00	-3.31	-0.74	.75	45K	SMOOTHNESS, BEKK		
L232B	G	49.61	15.56	-3.80	3.80	.98	45K	SMOOTHNESS, BEKK		
L243K	G	50.60	11.73	-2.29	.15	.77	45K	SMOOTHNESS, BEKK		
L162	G	51.07	10.00	-1.59	-1.51	.58	45K	SMOOTHNESS, BEKK		
L139B	G	53.60	11.13	.76	-.03	1.46	45K	SMOOTHNESS, BEKK		
L182K	G	54.40	9.79	1.74	-1.26	1.24	45K	SMOOTHNESS, BEKK		
L291K	G	56.85	12.67	3.77	1.93	1.46	45K	SMOOTHNESS, BEKK		
L230B	G	56.87	9.27	4.26	-1.43	.94	45K	SMOOTHNESS, BEKK		
L190C	G	58.47	11.07	5.59	.58	1.17	45K	SMOOTHNESS, BEKK		
GMEANS:		52.85	11.27			1.00				
95% ELLIPSE:		51.89	11.89			5.51	WITH GAMMA = -7 DEGREES			

SMOOTHNESS, BEKK

SAMPLE A83 = 52.8 BEKK SECONDS SAMPLE J50 = 11.3 BEKK SECONDS



TAPPI COLLABORATIVE REFERENCE PROGRAM
ANALYSIS T47-1 TABLE 1
SMOOTHNESS, BENDTSEN

APRIL 1979

LAB CODE	SAMPLE A83	WAVE ENVELOPE PAPER					SAMPLE J50	PRINTING					TEST D = 10		
		MEAN	DEV	N. DEV	SDR	R. SDR		MEAN	DEV	N. DEV	SDR	R. SDR	VAR	F	LAB
L100	106.	-13.	-1.03	12.	.73	.540.	24.	.54	43.	.83	47B	6	L100		
L182B	144.	26.	2.08	27.	1.67	530.	14.	.31	45.	.88	47B	6	L182B		
L236	123.	4.	.35	10.	.62	514.	-2.	-0.04	78.	1.51	47B	6	L236		
L242	116.	-3.	-0.25	17.	1.02	495.	-20.	-0.45	78.	1.51	47B	6	L242		
L243B	114.	-5.	-0.40	13.	.77	596.	80.	1.77	56.	1.10	47B	6	L243B		
L244	114.	-5.	-0.43	17.	1.05	467.	-49.	-1.09	34.	.66	47B	6	L244		
L248	115.	-4.	-0.31	19.	1.14	469.	-47.	-1.04	26.	.51	47B	6	L248		
L333	100.	-19.	-1.57	52.	3.19	75.	-441.	-9.79	1.	.02	47B	#	L333		
L685	47.	-72.	-5.87	1.	.08	84.	-432.	-9.59	12.	.24	47B	#	L685		

GR. MEAN = 119. ML/MIN

SD MEANS = 12. ML/MIN

AVERAGE SDR = 16. ML/MIN

TOTAL NUMBER OF LABORATORIES REPORTING = 9

Best values: A83 115 milliliter per minute
J50 510 milliliter per minute

The following laboratories were omitted from the grand means because of extreme test results: 333, 685.

GRAND MEAN = 516. ML/MIN

SD OF MEANS = 45. ML/MIN

AVERAGE SDR = 51. ML/MIN

TEST DETERMINATIONS = 10

7 LABS IN GRAND MEANS

TAPPI COLLABORATIVE REFERENCE PROGRAM
ANALYSIS T47-1 TABLE 2
SMOOTHNESS, BENDTSEN

APRIL 1979

LAB CODE	F	MEANS		COORDINATES		AVG R. SDR VAR	PROPERTY---TEST INSTRUMENT---CONDITIONS			
		A83	J50	MAJOR	MINOR					
L685	#	47.	84.	-433.	66.	.16 47B	SMOOTHNESS, BENDTSEN, WG 150			
L333	#	100.	75.	-441.	13.	1.60 47B	SMOOTHNESS, BENDTSEN, WG 150			
L100	6	106.	540.	24.	13.	.78 47B	SMOOTHNESS, BENDTSEN, WG 150			
L244	6	114.	467.	-49.	5.	.86 47B	SMOOTHNESS, BENDTSEN, WG 150			
L243B	6	114.	556.	80.	6.	.93 47B	SMOOTHNESS, BENDTSEN, WG 150			
L248	6	115.	469.	-47.	3.	.82 47B	SMOOTHNESS, BENDTSEN, WG 150			
L242	6	116.	495.	-20.	3.	1.27 47B	SMOOTHNESS, BENDTSEN, WG 150			
L236	6	123.	514.	-2.	-4.	1.06 47B	SMOOTHNESS, BENDTSEN, WG 150			
L182B	6	144.	530.	14.	-25.	1.27 47B	SMOOTHNESS, BENDTSEN, WG 150			

GMEANS: 119. 516.

95% ELLIPSE: 168. 46.

1.00

WITH GAMMA = 89 DEGREES

LAB CODE	SAMPLE B58 MEAN	BEAT SET OFFSET BACK				SAMPLE E50 MEAN	PRINTING				TEST D.E. = 4 VAR	F	LAB	
		76 GRAMS PER SQUARE METER	N.DEV	SDR	R _e SDR		91 GRAMS PER SQUARE METER	N.DEV	SDR	R _e SDR				
L126	53.8	-6.9	-1.02	.7	1.03	57.5	-6.2	-1.04	.3	.66	56K	G	L126	
L149	56.7	-3.9	-0.58	.5	.71	59.2	-4.5	-0.75	.5	1.11	56K	G	L149	
L182	61.8	1.2	.18	.5	.78	66.1	2.3	.39	.1	.28	56K	G	L182	
L213	64.5	3.9	.58	.9	1.26	68.2	4.5	.75	.3	.59	56K	G	L213	
L277	67.2	6.6	.98	1.3	1.77	70.0	6.2	1.04	.0	.00	56K	G	L277	
L291	63.8	3.1	.47	.6	.78	65.2	1.5	.24	.4	.90	56K	G	L291	
L333	53.2	-7.4	-1.10	.5	.71	57.5	-6.3	-1.04	1.0	2.22	56K	G	L333	
L339	71.5	10.9	1.61	.6	.81	72.7	9.0	1.50	.5	1.11	56K	G	L339	
L643	53.0	-7.6	-1.13	.8	1.15	57.2	-6.5	-1.08	1.0	2.13	56K	G	L643	
GR. MEAN =	60.6	K & N UNITS				GRAND MEAN =	63.8	K & N UNITS			TEST DETERMINATIONS =	4		
SD MEANS =	6.7	K & N UNITS				SD OF MEANS =	6.0	K & N UNITS			9 LABS IN GRAND MEANS			
AVERAGE SDR =	.7	K & N UNITS				AVERAGE SDR =	.4	K & N UNITS						
L651	33.4	-27.3	-4.04	.2	.27	32.2	-31.5	-5.25	.1	.28	56G	*	L651	
L688	27.7	-32.9	-4.88	.4	.52	26.1	-37.7	-6.28	.4	1.00	56G	*	L688	
TOTAL NUMBER OF LAB CATEGORIES REPORTING =	11													
Best values:	B58	61	K & N units											
	E50	64	K & N units											

LAB CODE	F	MEANS		COORDINATES		AVG R _e SDR VAR	PROPERTY---TEST INSTRUMENT---CONDITIONS		
		B58	E50	MAJOR	MINOR				
L688	*	27.7	26.1	-49.6	-6.3	.76	56G INK ABSORPTION, GWN METHOD		
L651	*	33.4	32.2	-41.3	-5.4	.27	56G INK ABSORPTION, GWN METHOD		
L643	G	53.0	57.2	-10.0	.2	1.64	56K INK ABSORPTION, K&N INK TEST		
L333	G	53.2	57.5	-9.7	.2	1.46	56K INK ABSORPTION, K&N INK TEST		
L126	G	53.8	57.5	-9.3	.1	.84	56K INK ABSORPTION, K&N INK TEST		
L149	G	56.7	59.2	-5.9	-.8	.91	56K INK ABSORPTION, K&N INK TEST		
L182	G	61.8	66.1	2.4	.9	.53	56K INK ABSORPTION, K&N INK TEST		
L291	G	63.8	65.2	3.3	-1.0	.84	56K INK ABSORPTION, K&N INK TEST		
L213	G	64.5	68.2	5.9	.8	.93	56K INK ABSORPTION, K&N INK TEST		
L277	G	67.2	70.0	9.1	.3	.89	56K INK ABSORPTION, K&N INK TEST		
L339	G	71.5	72.7	14.1	-.5	.96	56K INK ABSORPTION, K&N INK TEST		
GMEANS:		60.6	63.8			1.00			
		95% ELLIPSE:	29.6	2.2			WITH GAMMA = 41 DEGREES		

TAPPI COLLABORATIVE REFERENCE PROGRAM
ANALYSIS T57-1 TABLE 1
HYDROGEN ION CONCENTRATION (PH), COLD
TAPPI STANDARD T509 GS-77

APRIL 1979

LAB CODE	SAMPLE J75	PRINTING 76 GRAMS PER SQUARE METER				SAMPLE J17	PRINTING 93 GRAMS PER SQUARE METER				TEST D. = 5		
		MEAN	DEV	N. DEV	SDR		MEAN	DEV	N. DEV	SDR	R. SDR	VAR	F
L174C	6.340	.167	1.92	.055	.81	6.300	1.647	18.05	.141	3.55	57F	#	L174C
L182C	6.206	.033	.38	.031	.46	4.676	.023	.25	.018	.46	57D	0	L182C
L328	6.180	.007	.08	.130	1.93	4.560	-.093	-1.02	.089	2.25	57M	0	L328
L356	6.050	-.123	-1.40	.071	1.05	4.770	.117	1.28	.027	.69	57V	0	L356
L442	6.254	.082	.94	.038	.56	4.607	-.046	-.51	.024	.61	57G	0	L442

GR. MEAN = 6.173 PH UNITS

GRAND MEAN = 4.653 PH UNITS

TEST DETERMINATIONS = 5

SD MEANS = .087 PH UNITS

SD OF MEANS = .091 PH UNITS

4 LABS IN GRAND MEANS

AVERAGE SDR = .067 PH UNITS

AVERAGE SDR = .040 PH UNITS

TOTAL NUMBER OF LABORATORIES REPORTING = 5

The following laboratories were omitted from the
grand means because of extreme test results: 174C.

TAPPI COLLABORATIVE REFERENCE PROGRAM
ANALYSIS T57-1 TABLE 2
HYDROGEN ION CONCENTRATION (PH), COLD
TAPPI STANDARD T509 GS-77

APRIL 1979

LAB CODE	F	MEANS J75	MEANS J17	COORDINATES	AVG	PROPERTY---TEST INSTRUMENT---CONDITIONS
MAJOR	MINOR	R. SDR	VAR			
L356	0	6.050	4.770	-.169	-.009	.87 57V PH, COLD, BECKMAN EXPANDOMATIC
L328	0	6.180	4.560	.073	-.059	2.09 57M PH, COLD, BECKMAN ZEROMATIC
L182C	0	6.206	4.676	.006	.040	.46 57D PH, COLD, RADIGMETER TYPE PH M 28
L442	0	6.254	4.607	.090	.028	.52 57G PH, COLD, ORION DIGITAL IONALYZER
L174C #		6.340	6.300	-.1.082	1.252	2.18 57F PH, COLD, FISHER ACCUMET MODEL 220

GMEANS: 6.173 4.653
95% ELLIPSE: .899 .337 WITH GAMMA = -46 DEGREES

TAPPI COLLABORATIVE REFERENCE PROGRAM
ANALYSIS T57-2 TABLE 1
HYDROGEN ION CONCENTRATION (PH), HOT
TAPPI STANDARD T435 GS-77

APRIL 1979

LAB CODE	SAMPLE J75	PRINTING				SAMPLE J17	PRINTING				TEST No. = 5	
		MEAN	DEV	N _e DEV	SDR		MEAN	DEV	N _e DEV	SDR	R _e SDR	
L128	5.660	.042	.47	.055	1.48	4.320	.001	.05	.045	1.39	57L	G L128
L162	5.574	-.044	-.49	.030	.80	4.324	.005	.25	.025	.78	57C	G L162
L174H	7.760	2.142	23.90	.055	1.48	8.120	3.801	19.037	.084	2.59	57G	# L174H
L182H	5.518	-.100	-1.12	.019	.52	4.292	-.027	-1.35	.004	.14	57E	G L182H
L334	5.720	.102	1.14	.045	1.21	4.340	.021	1.05	.055	1.70	57C	G L334

GR_e MEAN = 5.618 PH UNITS

SD MEANS = .090 PH UNITS

AVERAGE SDR = .037 PH UNITS

TOTAL NUMBER OF LABORATORIES REPORTING = 5

GRAND MEAN = 4.319 PH UNITS

SD OF MEANS = .020 PH UNITS

TEST DETERMINATIONS = 5

4 LABS IN GRAND MEANS

AVERAGE SDR = .032 PH UNITS

The following laboratories were omitted from the
grand means because of extreme test results: 174H.

TAPPI COLLABORATIVE REFERENCE PROGRAM
ANALYSIS T57-2 TABLE 2
HYDROGEN ION CONCENTRATION (PH), HOT
TAPPI STANDARD T435 GS-77

APRIL 1979

LAB CODE	F	MEANS J75	COORDINATES	Avg	PROPERTY---TEST INSTRUMENT---CONDITIONS
MAJOR	MINOR	R _e SDR	VAR		
L182H G	5.518	4.292	-.103	-.007	.33 57E PH, HOT, RADIOMETER TYPE PH M 28
L162 G	5.574	4.324	-.042	.013	.79 57C PH, HOT, CORNING MODEL 12 RESEARCH METER
L128 G	5.660	4.320	.041	-.007	1.43 57L PH, HOT, L+N
L334 G	5.720	4.340	.104	.001	1.45 57C PH, HOT, CORNING MODEL 12 RESEARCH METER
L174H #	7.760	8.120	2.832	3.319	2.03 57G PH, HOT, FISHER ACCUMET MODEL 220

GMSEANS: 5.618 4.319

95% ELLIPSE:

.694 .074

WITH GAMMA = 11 DEGREES

TAPPI COLLABORATIVE REFERENCE PROGRAM
ANALYSIS T60-1 TABLE 1

APRIL 1979

OPACITY (89% REFLECTANCE BACKING) IN PERCENT
TAPPI STANDARD T425 GS-75, OPACITY OF PAPER (15 DEG./DIFFUSE, ILLUMINANT A) - B&L TYPE

LAB CODE	SAMPLE B21	SEMI BLEACHED				SAMPLE K24	PRINTING				TEST D = 10				
		MEAN	65 GRAMS PER SQUARE METER	DEV	N. DEV		MEAN	103 GRAMS PER SQUARE METER	DEV	N. DEV	SDR	R. SDR	VAR	F	IAB
L105	72.64	.21	.20	.84	.79	95.48	.05	.12	.17	.54	60H	G L105			
L115	73.06	.63	.58	1.39	1.30	95.28	-.15	-.40	.27	.87	60B	G L115			
L118	72.78	.35	.32	1.36	1.28	95.41	-.02	-.06	.32	1.04	60B	G L118			
L121	72.48	.05	.05	1.46	1.37	94.83	-.60	-.158	.33	1.05	60B	G L121			
L122	71.25	-1.18	-1.08	1.04	.98	95.39	-.04	-.11	.25	.81	60D	G L122			
L123	71.12	-1.31	-1.20	1.08	1.02	95.27	-.16	-.43	.53	1.69	60W	G L123			
L124	73.45	1.02	.94	1.16	1.10	96.01	.58	1.51	.25	.80	60B	G L124			
L125	72.11	-.32	-.29	.99	.93	95.14	-.29	-.77	.43	1.39	60H	G L125			
L131	70.55	-1.88	-1.72	.78	.74	94.57	-.86	-2.25	.15	.48	60R	G L131			
L132	71.59	-.84	-.77	1.52	1.44	95.36	-.07	-.19	.32	1.04	60B	G L132			
L134	72.50	.07	.07	.71	.67	95.60	.17	.44	.52	1.67	60R	G L134			
L139	71.87	-.56	-.51	1.36	1.28	95.08	-.35	-.92	.45	1.46	60B	G L139			
L148H	72.10	-.33	-.30	.96	.90	95.21	-.22	-.58	.33	1.07	60H	G L148H			
L150	73.65	1.22	1.12	.85	.80	95.95	.52	1.35	.28	.92	60B	G L150			
L152	73.36	.93	.86	.81	.76	95.60	.17	.44	.16	.53	60B	G L152			
L153	71.35	-1.08	-.98	1.29	1.22	95.50	.07	.17	.41	1.32	60B	G L153			
L157	73.30	.87	.80	1.06	1.00	95.80	.37	.96	.26	.83	60B	G L157			
L158	73.50	1.07	.98	.69	.65	95.82	.39	1.01	.36	1.18	60D	G L158			
L159	71.67	-.76	-.69	.74	.70	95.67	.24	.62	.19	.63	60R	G L159			
L162	71.52	-.91	-.83	1.35	1.27	95.61	.18	.46	.32	1.05	60W	G L162			
L166	71.00	-1.43	-1.31	1.34	1.27	95.14	-.29	-.77	.22	.70	60B	G L166			
L173A	71.74	-.66	-.63	1.35	1.27	95.39	-.04	-.11	.38	1.24	60B	G L173A			
L206	72.49	.06	.06	.93	.88	95.46	.03	.07	.25	.82	60B	G L206			
L210B	72.66	.23	.21	1.00	.94	95.33	-.10	-.27	.23	.75	60B	G L210B			
L210D	73.38	.95	.87	1.40	1.32	95.72	.29	.75	.20	.64	60D	G L210D			
L211S	72.53	.10	.10	1.19	1.12	95.28	-.15	-.40	.29	.95	60R	G L211S			
L212	74.51	2.08	1.91	1.86	1.75	95.85	.42	1.09	.58	1.88	60B	G L212			
L213	72.65	.22	.21	.79	.75	95.68	.25	.64	.24	.79	60B	G L213			
L223B	72.81	.38	.35	.56	.53	95.49	.06	.15	.39	1.27	60B	G L223B			
L225	72.35	-.08	-.07	.66	.62	95.75	.32	.83	.53	1.70	60B	G L225			
L226B	71.31	-1.12	-1.02	1.03	.97	95.21	-.22	-.58	.45	1.45	60B	G L226B			
L228	71.00	-1.43	-1.31	1.00	.94	95.14	-.29	-.77	.31	1.01	60H	G L228			
L230	71.81	-.62	-.56	1.19	1.12	95.63	.20	.51	.38	1.23	60B	G L230			
L236B	70.74	-1.69	-1.54	1.50	1.41	95.08	-.35	-.92	.38	1.22	60B	G L236B			
L238A	70.46	-1.97	-1.80	.82	.77	94.66	-.77	-2.02	.29	.94	60R	G L238A			
L241	72.05	-.38	-.34	1.22	1.15	95.19	-.24	-.64	.30	.97	60B	G L241			
L243	71.35	-1.08	-.98	1.68	1.58	95.21	-.22	-.58	.24	.78	60B	G L243			
L254	72.87	.44	.41	1.15	1.09	95.53	.10	.25	.21	.66	60H	G L254			
L259	73.49	1.06	.97	.48	.45	95.89	.46	1.19	.19	.60	60B	G L259			
L261	74.46	2.03	1.86	1.03	.97	96.08	.65	1.69	.20	.64	60B	G L261			
L262	72.46	.03	.03	.64	.60	96.40	.67	2.53	.28	.91	60R	* L262			
L275	71.46	-.97	-.88	.61	.58	95.40	-.03	-.09	.18	.59	60R	G L275			
L278	73.09	.66	.61	.98	.92	95.47	.04	.10	.28	.90	60B	G L278			
L281	72.83	.40	.37	.86	.81	95.43	-.00	-.01	.34	1.10	60D	G L281			
L285D	72.17	-.26	-.23	.98	.93	94.92	-.51	-1.34	.39	1.26	60D	G L285D			
L288	72.25	-.18	-.16	1.01	.95	95.49	.06	.15	.30	.98	60D	G L288			
L301	72.71	.26	.26	1.06	.99	95.48	.05	.12	.29	.95	60B	G L301			
L305	71.62	-.81	-.74	.60	.57	95.67	.24	.62	.14	.46	60R	G L305			
L308	73.38	.95	.87	1.59	1.50	95.70	.27	.70	.36	1.17	60H	G L308			
L317	72.24	-.15	-.17	.71	.67	95.03	-.40	-1.05	.37	1.18	60B	G L317			
L318	73.20	.77	.71	1.25	1.18	95.25	-.18	-.48	.55	1.90	60B	G L318			
L323	72.33	-.10	-.09	1.02	.96	95.51	.08	.20	.31	.99	60W	G L323			
L326	73.98	1.55	1.42	1.10	1.04	95.42	-.01	-.03	.40	1.28	60B	G L326			
L328	71.99	-.44	-.40	1.27	1.20	94.40	-.103	-2.70	.84	2.72	60B	* L328			
L339	72.00	-.43	-.39	.47	.44	96.00	.57	1.48	.00	.00	60B	G L339			
L341	70.97	-1.46	-1.33	.78	.73	95.15	-.28	-.74	.25	.82	60R	G L341			
L349	73.60	1.17	1.08	.84	.79	95.00	-.43	-1.13	.58	1.86	60D	G L349			
L352	71.72	-.71	-.65	.76	.71	95.16	-.27	-.71	.07	.23	60R	G L352			
L354	70.30	-2.13	-1.95	1.16	1.09	94.70	-.73	-1.92	.48	1.56	60B	G L354			
L390	72.92	.45	.45	1.16	1.09	95.34	-.09	-.24	.27	.86	60B	G L390			
L396	74.50	2.07	1.90	1.08	1.02	95.80	.37	.96	.42	1.36	60B	G L396			
L523	72.61	.18	.17	.93	.88	95.64	.21	.54	.20	.65	60R	G L523			
L567	75.65	3.22	2.95	.85	.80	96.44	1.01	2.63	.30	.98	60D	* L567			
L571	73.12	.69	.64	1.56	1.47	95.63	.20	.51	.31	1.00	60D	G L571			
L573	72.75	.32	.30	1.09	1.03	95.60	.17	.44	.19	.61	60H	G L573			

OPACITY (89% REFLECTANCE BACKING) IN PERCENT
TAPPI STANDARD T425 GS-75, OPACITY OF PAPER (15 DEG./DIFFUSE, ILLUMINANT A) - B&L TYPE

LAB CODE	SAMPLE B21	SEMI BLEACHED				SAMPLE				PRINTING				TEST D = 10				
		MEAN	DEV	N _{DEV}	SDR	MEAN	DEV	N _{DEV}	SDR	R _e SDR	MEAN	DEV	N _{DEV}	SDR	R _e SDR	VAR	F	LAB
L581	72.65	.22	.21	1.47	1.38	95.35	-.08	.22	.14	.44	60B	6	L581					
L592	71.24	-1.19	-1.09	1.27	1.19	94.95	-.48	-1.26	.40	1.30	60W	6	L592					
L594	75.02	2.55	2.38	.62	.58	95.52	.09	.23	.26	.83	60D	*	L594					
L597	71.00	-1.43	-1.31	1.33	1.25	95.40	-.03	-.09	.70	2.26	60B	6	L597					
L599	74.40	1.97	1.81	.66	.62	95.70	.27	.70	.35	1.13	60B	6	L599					
L673R	72.89	.46	.43	1.41	1.33	95.45	.02	.04	.21	.67	60B	6	L673R					
L673T	71.90	-.53	-.48	.71	.67	95.67	.24	.62	.19	.61	60B	6	L673T					
L688	72.16	-.27	-.24	.66	.62	96.04	.61	1.58	.28	.89	60B	6	L688					
L692	71.89	-.54	-.49	1.18	1.11	95.25	-.18	-.48	.14	.46	60D	6	L692					
L698	71.40	-.03	-.94	1.47	1.38	94.84	-.59	-1.55	.30	.96	60D	6	L698					
GR. MEAN =	72.43	PERCENT				GRAND MEAN =	95.43	PERCENT						TEST DETERMINATIONS =	10			
SD MEANS =	1.09	PERCENT				SD OF MEANS =	.38	PERCENT						75 LABS IN GRAND MEANS				
	AVERAGE SDR =	1.06	PERCENT				AVERAGE SDR =		.31	PERCENT								
L100	72.91	.48	.44	.70	.66	95.59	.16	.41	.23	.74	60E	*	L100					
L232	71.50	-.93	-.85	.53	.50	95.50	.07	.17	.24	.76	60P	*	L232					
L249	73.66	1.23	1.13	.99	.93	95.43	-.00	-.01	.28	.89	60P	*	L249					
L256	70.54	-1.89	-1.73	1.14	1.07	94.72	-.71	-1.86	.40	1.30	60N	*	L256					
L260	72.00	-.43	-.39	.41	.38	95.65	.22	.57	.24	.78	60P	*	L260					
L277	28.90	-43.53	-39.86	.99	.94	4.90	-90.53	-236.46	.57	1.83	60P	*	L277					
L312	69.70	-2.73	-2.50	.86	.81	94.70	-.73	-1.92	.42	1.36	60P	*	L312					
L685B	71.17	-.126	-.115	.96	.90	95.48	.05	.12	.32	1.03	60P	*	L685B					
L687	70.82	-.161	-.147	.52	.49	95.94	.51	1.32	.37	1.21	60P	*	L687					

TOTAL NUMBER OF LABORATORIES REPORTING = 84

Best values: B21 72.4 + 1.9 percent
K24 95.4 + 0.6 percent

ANALYSIS T60-1 TABLE 2

OPACITY (89% REFLECTANCE BACKING) IN PERCENT

TAPPI STANDARD T425 GS-75, OPACITY OF PAPER (15 DEG./DIPFUSE, ILLUMINANT A) - B&L TYPE

LAB CODE	F	MEANS B21	COORDINATES K24	MAJOR MINOR	Avg E SDR VAR	PROPERTY--TEST INSTRUMENT--CONDITIONS
L277	+	28.90	4.90	-62.35	-78.76	1.38 60P OPACITY (WBITE BACKING), PHOTOVOLT
L312	+	69.70	94.70	-2.82	-0.12	1.08 60P OPACITY (WHITE BACKING), PHOTOVOLT
L354	0	70.30	94.70	-2.23	-0.25	1.32 60B OPACITY (WHITE BACKING), BAUSCH + LOMB
L238A	0	70.46	94.66	-2.09	-0.32	.85 60R OPACITY (WBITE BACKING), THWING-ALBERT (FORMERLY SRL)
L256	+	70.54	94.72	-2.00	-0.28	1.18 60N OPACITY (WHITE BACKING), HUNTER
L131	0	70.55	94.57	-2.02	-0.43	.61 60R OPACITY (WHITE BACKING), THWING-ALBERT (FORMERLY SRL)
L236B	0	70.74	95.08	-1.72	.03	1.32 60B OPACITY (WBITE BACKING), BAUSCH + LOMB
L687	+	70.82	95.94	-1.45	.05	.85 60P OPACITY (WHITE BACKING), PHOTOVOLT
L341	0	70.97	95.15	-1.48	.04	.78 60R OPACITY (WHITE BACKING), THWING-ALBERT (FORMERLY SRL)
L228	0	71.00	95.14	-1.46	.03	.98 60H OPACITY (WHITE BACKING), HUYGEN
L166	0	71.00	95.14	-1.46	.03	.98 60B OPACITY (WHITE BACKING), BAUSCH + LOMB
L597	0	71.00	95.40	-1.40	.28	1.76 60B OPACITY (WBITE BACKING), BAUSCB + LOMB
L123	0	71.12	95.27	-1.31	.13	1.36 60W OPACITY (WBITE BACKING), HUYGEN, DIGITAL
L685B	+	71.17	95.48	-1.21	.32	.97 60P OPACITY (WHITE BACKING), PHOTOVOLT
L592	0	71.24	94.95	-1.26	-0.21	1.25 60W OPACITY (WBITE BACKING), HUYGEN, DIGITAL
L122	0	71.25	95.39	-1.16	.22	.90 60D OPACITY (WHITE BACKING), BNL-2
L226B	0	71.31	95.21	-1.14	.03	1.21 60B OPACITY (WHITE BACKING), BAUSCH + LOMB
L153	0	71.35	95.50	-1.03	.30	1.27 60B OPACITY (WBITE BACKING), BAUSCH + LOMB
L243	0	71.35	95.21	-1.10	.02	1.18 60B OPACITY (WHITE BACKING), BAUSCH + LOMB
L698	0	71.40	94.84	-1.13	-0.35	1.17 60D OPACITY (WBITE BACKING), BNL-2
L275	0	71.46	95.40	-0.95	.18	.58 60R OPACITY (WHITE BACKING), TBWING-ALBERT (FORMERLY SRL)
L232	+	71.50	95.50	-0.89	.27	.63 60P OPACITY (WHITE BACKING), PHOTOVOLT
L162	0	71.52	95.61	-0.84	.37	1.16 60W OPACITY (WBITE BACKING), HUYGEN, DIGITAL
L132	0	71.59	95.36	-0.83	.11	1.24 60B OPACITY (WHITE BACKING), BAUSCB + LOMB
L305	0	71.62	95.67	-0.73	.41	.51 60R OPACITY (WHITE BACKING), THWING-ALBERT (FORMERLY SRL)
L159	0	71.67	95.67	-0.68	.40	.66 60R OPACITY (WBITE BACKING), THWING-ALBERT (FORMERLY SRL)
L352	0	71.72	95.16	-0.75	-0.11	.47 60R OPACITY (WHITE BACKING), TBWING-ALBERT (FORMERLY SRL)
L173A	0	71.74	95.39	-0.68	.11	1.25 60B OPACITY (WHITE BACKING), BAUSCB + LOMB
L230	0	71.81	95.63	-0.56	.33	1.18 60B OPACITY (WHITE BACKING), BAUSCH + LOMB
L139	0	71.87	95.08	-0.62	-0.22	1.37 60B OPACITY (WHITE BACKING), BAUSCH + LOMB
L692	0	71.89	95.25	-0.56	-0.06	.79 60D OPACITY (WBITE BACKING), BNL-2
L673T	0	71.90	95.67	-0.46	.35	.64 60B OPACITY (WHITE BACKING), BAUSCB + LOMB
L328	*	71.99	94.40	-0.65	-0.91	1.96 60B OPACITY (WBITE BACKING), BAUSCB + LOMB
L260	*	72.00	95.65	-0.37	.30	.58 60P OPACITY (WHITE BACKING), PHOTOVOLT
L339	0	72.00	96.00	-0.29	.65	.22 60B OPACITY (WBITE BACKING), BAUSCH + LOMB
L241	0	72.05	95.19	-0.42	-0.15	1.06 60B OPACITY (WHITE BACKING), BAUSCH + LOMB
L148H	0	72.10	95.21	-0.37	-0.15	.99 60H OPACITY (WHITE BACKING), HUYGEN
L125	0	72.11	95.14	-0.37	-0.22	1.16 60H OPACITY (WBITE BACKING), HUYGEN
L688	0	72.16	96.04	-0.13	.65	.75 60B OPACITY (WHITE BACKING), BAUSCB + LOMB
L285D	0	72.17	94.92	-0.36	-0.44	1.09 60D OPACITY (WHITE BACKING), BNL-2
L317	0	72.24	95.03	-0.27	-0.35	.93 60B OPACITY (WBITE BACKING), BAUSCH + LOMB
L288	0	72.25	95.49	-0.16	.09	.97 60D OPACITY (WHITE BACKING), BNL-2
L323	0	72.33	95.51	-0.08	.10	.97 60W OPACITY (WBITE BACKING), HUYGEN, DIGITAL
L225	0	72.35	95.75	-0.00	.33	1.16 60B OPACITY (WBITE BACKING), BAUSCH + LOMB
L262	*	72.46	96.40	.25	.94	.76 60R OPACITY (WHITE BACKING), THWING-ALBERT (FORMERLY SRL)
L121	0	72.48	94.83	-0.08	-0.60	1.21 60B OPACITY (WHITE BACKING), BAUSCH + LOMB
L206	0	72.49	95.46	.07	.01	.85 60B OPACITY (WHITE BACKING), BAUSCH + LOMB
L134	0	72.50	95.60	.11	.15	1.17 60R OPACITY (WHITE BACKING), THWING-ALBERT (FORMERLY SRL)
L211S	0	72.53	95.28	.07	-0.17	1.03 60R OPACITY (WBITE BACKING), TBWING-ALBERT (FORMERLY SRL)
L523	0	72.61	95.64	.23	.16	.76 60R OPACITY (WBITE BACKING), TBWING-ALBERT (FORMERLY SRL)
L105	0	72.64	95.48	.22	-0.00	.67 60H OPACITY (WBITE BACKING), HUYGEN
L213	0	72.65	95.68	.27	.19	.77 60B OPACITY (WHITE BACKING), BAUSCB + LOMB
L581	0	72.65	95.35	.20	-0.13	.91 60B OPACITY (WHITE BACKING), BAUSCB + LOMB
L210B	0	72.66	95.33	.21	-0.15	.84 60B OPACITY (WBITE BACKING), BAUSCH + LOMB
L301	0	72.71	95.48	.29	-0.02	.97 60B OPACITY (WBITE BACKING), BAUSCB + LOMB
L573	0	72.75	95.60	.35	.09	.82 60B OPACITY (WHITE BACKING), HUYGEN
L118	0	72.78	95.41	.34	-0.10	1.16 60B OPACITY (WHITE BACKING), BAUSCB + LOMB
L223B	0	72.81	95.49	.39	-0.03	.90 60B OPACITY (WBITE BACKING), BAUSCH + LOMB
L281	0	72.83	95.43	.39	-0.09	.95 60D OPACITY (WBITE BACKING), BNL-2
L254	0	72.87	95.53	.45	-0.00	.88 60H OPACITY (WHITE BACKING), HUYGEN
L673R	0	72.89	95.45	.46	-0.09	1.00 60B OPACITY (WBITE BACKING), BAUSCH + LOMB
L100	*	72.91	95.59	.51	.05	.70 60E OPACITY (WHITE BACKING), ZEISS ELREPBG, FMY-C(10) FILTER
L390	0	72.92	95.34	.46	-0.20	.98 60B OPACITY (WBITE BACKING), BAUSCB + LOMB
L115	0	73.06	95.28	.59	-0.29	1.09 60B OPACITY (WHITE BACKING), BAUSCB + LOMB
L278	0	73.09	95.47	.66	-0.11	.91 60B OPACITY (WBITE BACKING), BAUSCH + LOMB

ANALYSIS T60-1 TABLE 2

OPACITY (89% REFLECTANCE BACKING) IN PERCENT


TAPPI STANDARD T425 GS-75, OPACITY OF PAPER (15 DEG./DIFFUSE, ILLUMINANT A) - B&L TYPE

LAB CODE	F	MEANS		COORDINATES		AVG R_e	SDR VAR	PROPERTY---TEST INSTRUMENT---CONDITIONS		
		B21	K24	MAJOR	MINOR					
L571	G	73.12	95.63	.72	.04	1.23	60D	OPACITY (WHITE BACKING),	BNL-2	
L318	G	73.20	95.25	.72	-.035	1.54	60B	OPACITY (WHITE BACKING),	BAUSCH + LOMB	
L157	G	73.30	95.80	.93	.17	.92	60B	OPACITY (WHITE BACKING),	BAUSCH + LOMB	
L152	G	73.36	95.60	.95	-.04	.64	60B	OPACITY (WHITE BACKING),	BAUSCH + LOMB	
L308	G	73.38	95.70	.99	.05	1.33	60H	OPACITY (WHITE BACKING),	HUYGEN	
L210D	G	73.38	95.72	.99	.07	.98	60D	OPACITY (WHITE BACKING),	BNL-2	
L124	G	73.45	96.01	1.13	.34	.95	60B	OPACITY (WHITE BACKING),	BAUSCH + LOMB	
L259	G	73.49	95.89	1.14	.21	.52	60B	OPACITY (WHITE BACKING),	BAUSCH + LOMB	
L158	G	73.50	95.82	1.13	.14	.91	60D	OPACITY (WHITE BACKING),	BNL-2	
L349	G	73.60	95.00	1.05	-.68	1.33	60D	OPACITY (WHITE BACKING),	BNL-2	
L150	G	73.65	95.95	1.31	.24	.86	60B	OPACITY (WHITE BACKING),	BAUSCH + LOMB	
L249	*	73.66	95.43	1.20	-.27	.91	60P	OPACITY (WHITE BACKING),	PHOTOVOLT	
L326	G	73.98	95.42	1.51	-.35	1.16	60B	OPACITY (WHITE BACKING),	BAUSCH + LOMB	
L599	G	74.40	95.70	1.98	-.17	.87	60B	OPACITY (WHITE BACKING),	BAUSCH + LOMB	
L271	G	74.46	96.08	2.13	.18	.81	60B	OPACITY (WHITE BACKING),	BAUSCH + LOMB	
L396	G	74.50	95.80	2.10	-.10	1.19	60B	OPACITY (WHITE BACKING),	BAUSCH + LOMB	
L212	G	74.51	95.85	2.13	-.05	1.81	60B	OPACITY (WHITE BACKING),	BAUSCH + LOMB	
L594	*	75.02	95.52	2.55	-.49	.71	60D	OPACITY (WHITE BACKING),	BNL-2	
L567	*	75.65	96.44	3.37	.27	.89	60D	OPACITY (WHITE BACKING),	BNL-2	
GMEANS:		72.43	95.43			1.00				
95% ELLIPSE:		2.81		.76		WITH GAMMA = 12 DEGREES				

OPACITY, B&L TYPE, 89% BACKING

SAMPLE B21 = 72.4 PERCENT

SAMPLE K24 = 95.4 PERCENT



TAPPI STANDARD T425 GS-75, SPACIITY OF PAPER (15 DEG./DIFFUSE, ILLUMINANT A) - B&L TYPE

TAPPI STANDARD T425 GS-75. OPACITY OF PAPER (15 DEG.-DIFUSE, ILLUMINANT A) - BAL TYPE

LAB CODE	F	MEANS		COORDINATES		Ave R. SDR	Var	PROPERTY---TEST INSTRUMENT---CONDITIONS
		B21	K24	MAJOR	MINOR			
L236B	G	74.86	95.66	-.89	.05	1.41	60C	OPACITY (PAPER BACKING), BAUSCH + LOMB
L115	G	75.56	95.60	-.20	-.10	.65	60C	OPACITY (PAPER BACKING), BAUSCH + LOMB
L118	G	75.94	95.77	.20	.02	1.00	60C	OPACITY (PAPER BACKING), BAUSCH + LOMB
L243	G	76.62	95.86	.89	.02	.94	60C	OPACITY (PAPER BACKING), BAUSCH + LOMB
GMEANS:		75.74	95.72			1.00		
		95% ELLIPSE:		5.63	.51	WITH GAMMA = 7 DEGREES		

TAPPI COLLABORATIVE REFERENCE PROGRAM
ANALYSIS T60-3 TABLE 1
OPACITY (PAPER BACKING) IN PERCENT

APRIL 1979

TAPPI SUGGESTED METHOD T519 GS-78, DIFFUSE OPACITY OF PAPER - ILLUMINANT C, ELREPHO TYPE

LAB CODE	SAMPLE B21	SEMI BLEACHED				SAMPLE K24	PRINTING				TEST D. = 10		
		MEAN	DEV	N _o DEV	SDR		MEAN	DEV	N _o DEV	SDR	R _e SDR	VAR	F
L100	77.26	.29	.45	.65	.94	96.30	.07	.58	.20	1.15	60J	G	L100
L150	76.98	.02	.03	.45	.65	96.17	-.06	-.49	.13	.76	60J	G	L150
L182E	76.92	-.05	-.07	.94	1.36	96.14	-.09	-.71	.25	1.47	60J	G	L182E
L233F	76.82	-.15	-.22	.76	1.09	96.29	.06	.50	.19	1.10	60F	G	L233F
L236	76.49	-.46	-.72	.50	.72	96.05	-.18	-.144	.16	.91	60J	G	L236
L242	77.78	.81	1.23	.89	1.28	96.28	.06	.46	.24	1.36	60J	G	L242
L244	76.10	-.87	-1.31	.75	1.08	96.08	-.15	-.19	.14	.81	60F	G	L244
L250T	76.67	-.30	-.45	.58	.83	96.21	-.02	-.14	.29	1.69	60J	G	L250T
L309	76.01	-.96	-1.45	.60	.86	96.18	-.05	-.39	.13	.76	60J	G	L309
L360	76.55	-.42	-.63	.75	1.09	96.29	.06	.50	.19	1.07	60F	G	L360
L446	76.65	-.31	-.47	1.12	1.62	96.31	.08	.63	.07	.43	60J	G	L446
L575	78.04	1.07	1.63	.52	.75	96.40	.17	1.39	.08	.47	60J	G	L575
L598	77.24	.27	.42	.53	.76	96.09	-.14	-.11	.14	.79	60J	G	L598
L678	78.29	1.32	2.01	.61	.87	96.49	.26	2.12	.21	1.20	60J	G	L678
L685A	76.68	-.29	-.43	.77	1.10	96.14	-.09	-.71	.18	1.03	60F	G	L685A
GR _e MEAN = 76.97 PERCENT						GRAND MEAN = 96.23 PERCENT					TEST DETERMINATIONS = 10		
SD MEANS = .66 PERCENT						SD OF MEANS = .12 PERCENT					15 LABS IN GRAND MEANS		
AVERAGE SDR = .70 PERCENT						AVERAGE SDR = .17 PERCENT							
L253C	77.00	.03	.05	.97	1.40	95.83	-.40	-3.21	.13	.77	60G	G	L253C
L626	75.40	-1.57	-2.38	.46	.66	96.00	-.23	-1.84	.00	.00	60Q	G	L626
TOTAL NUMBER OF LABORATORIES REPORTING = 17													
Best values: B21 77.0 ± 1.0 percent						K24 96.3 ± 0.2 percent							

TAPPI COLLABORATIVE REFERENCE PROGRAM
ANALYSIS T60-3 TABLE 2
OPACITY (PAPER BACKING) IN PERCENT

APRIL 1979


TAPPI SUGGESTED METHOD T519 GS-78, DIFFUSE OPACITY OF PAPER - ILLUMINANT C, ELREPHO TYPE

LAB CODE	F	MEANS		COORDINATES		AVG R _e SDR VAR	PROPERTY---TEST INSTRUMENT---CONDITIONS		
		B21	K24	MAJOR	MINOR		ZEISS ELREPHO, FMY-C(10) FILTER	ZEISS ELREPHO, FMY-C(10) FILTER	ZEISS ELREPHO, FMY-C(10) NG TRAP
L626	+	75.40	96.00	-1.58	-.02	.33	60Q OPACITY (PAPER BACKING), PHOTOVOLT		
L309	G	76.01	96.18	-.95	.08	.81	60J OPACITY (PAPER BACKING), ZEISS ELREPHO, FMY-C(10) FILTER		
L244	G	76.10	96.08	-.88	-.03	.94	60F OPACITY (PAPER BACKING), ZEISS ELREPHO, FMY-C(10) NG TRAP		
L236	G	76.49	96.05	-.49	-.11	.82	60J OPACITY (PAPER BACKING), ZEISS ELREPHO, FMY-C(10) FILTER		
L360	G	76.55	96.29	-.40	.12	1.08	60F OPACITY (PAPER BACKING), ZEISS ELREPHO, FMY-C(10) NG TRAP		
L446	G	76.65	96.31	-.30	.12	1.02	60J OPACITY (PAPER BACKING), ZEISS ELREPHO, FMY-C(10) FILTER		
L250T	G	76.67	96.21	-.30	.02	1.26	60J OPACITY (PAPER BACKING), ZEISS ELREPHO, FMY-C(10) FILTER		
L665A	G	76.68	96.14	-.29	-.05	1.06	60F OPACITY (PAPER BACKING), ZEISS ELREPHO, FMY-C(10) NG TRAP		
L233F	G	76.82	96.29	-.14	.08	1.10	60F OPACITY (PAPER BACKING), ZEISS ELREPHO, FMY-C(10) NG TRAP		
L182E	G	76.92	96.14	-.06	-.08	1.41	60J OPACITY (PAPER BACKING), ZEISS ELREPHO, FMY-C(10) FILTER		
L150	G	76.98	96.17	.01	-.06	.70	60J OPACITY (PAPER BACKING), ZEISS ELREPHO, FMY-C(10) FILTER		
L253C	+	77.00	95.83	-.02	-.40	1.08	60G OPACITY (PAPER BACKING), GARDNER		
L598	G	77.24	96.09	.25	-.17	.77	60J OPACITY (PAPER BACKING), ZEISS ELREPHO, FMY-C(10) FILTER		
L100	G	77.26	96.30	.30	.03	1.05	60J OPACITY (PAPER BACKING), ZEISS ELREPHO, FMY-C(10) FILTER		
L242	G	77.78	96.28	.81	-.05	1.32	60J OPACITY (PAPER BACKING), ZEISS ELREPHO, FMY-C(10) FILTER		
L575	G	78.04	96.40	1.09	.03	.61	60J OPACITY (PAPER BACKING), ZEISS ELREPHO, FMY-C(10) FILTER		
L678	G	78.29	96.49	1.35	.09	1.04	60J OPACITY (PAPER BACKING), ZEISS ELREPHO, FMY-C(10) FILTER		
GMEANS: 76.97 96.23						1.00	WITH GAMMA = 7 DEGREES		
95% ELLIPSE: 1.90									

OPACITY, ELREPHO TYPE, PAPER BACKING

SAMPLE B21 = 76.97 PERCENT

SAMPLE K24 = 96.23 PERCENT



LAB CODE	SAMPLE					PRINTING					SAMPLE					PRINTING					TEST D. ^a		
	J79	73 GRAMS PER SQUARE METER				E79	116 GRAMS PER SQUARE METER				MEAN	DEV	N _e DEV	SDR	R _e SDR	MEAN	DEV	N _e DEV	SDR	R _e SDR	VAR	F	LAB
L122	67.90	.04	.10	.13	.59	95.61	-.17	-.18	.06	.56	65N	0	L122										
L132	68.09	.23	.57	.17	.78	96.01	.23	.23	.06	.56	65N	0	L132										
L158	68.50	.64	1.61	.29	1.33	97.21	1.43	1.45	.14	1.19	65N	0	L158										
L176A	67.12	-.73	-1.84	.28	1.25	94.69	-1.10	-1.11	.04	.31	65A	0	L176A										
L190C	67.91	.05	.14	.14	.62	95.80	.01	.01	.14	1.24	65A	0	L190C										
L210M	67.76	-.10	-.24	.29	1.33	95.64	-.15	-.15	.21	1.81	65M	0	L210M										
L210N	67.86	.00	.01	.15	.68	96.44	.65	.66	.13	1.14	65N	0	L210N										
L211	68.20	.34	.86	.13	.59	93.59	-2.20	-2.23	.50	4.38	65N	*	L211										
L225	67.94	.08	.20	.35	1.57	96.14	.35	.36	.43	3.74	65N	0	L225										
L243	67.92	.07	.17	.15	.67	95.92	.14	.14	.09	.78	65A	0	L243										
L259	67.92	.07	.17	.16	.72	95.76	-.02	-.02	.05	.45	65M	0	L259										
L275	68.56	.70	1.77	.16	.72	96.40	.61	.62	.11	.94	65M	0	L275										
L285	66.76	-1.10	-2.75	.38	1.71	93.09	-2.70	-2.74	.20	1.72	65N	*	L285										
L288	67.67	-.18	-.46	.10	.47	95.60	-.19	-.19	.05	.47	65N	0	L288										
L308	68.04	.18	.45	.13	.59	96.37	.59	.60	.09	.78	65N	0	L308										
L317	67.75	-.11	-.27	.14	.64	95.24	-.55	-.56	.19	1.68	65M	0	L317										
L523	68.01	.15	.39	.10	.45	95.31	-.47	-.48	.06	.56	65N	0	L523										
L565	68.10	.24	.61	.09	.42	95.96	.18	.18	.13	1.14	65A	0	L565										
L598	67.42	-.43	-1.09	.55	2.48	96.41	.63	.64	.08	.73	65M	0	L598										
L636	68.00	.14	.36	.61	2.75	97.00	1.21	1.23	.00	.00	65M	0	L636										
L673R	67.51	-.35	-.87	.24	1.07	96.42	.64	.65	.10	.91	65N	0	L673R										
L692	67.91	.05	.14	.19	.85	96.67	.89	.90	.12	1.02	65N	0	L692										
GR. MEAN = 67.86 PERCENT						GRAND MEAN = 95.79 PERCENT					TEST DETERMINATIONS = 8												
SD MEANS = .40 PERCENT						SD OF MEANS = .99 PERCENT					22 LABS IN GRAND MEANS												
AVERAGE SDR = .22 PERCENT						AVERAGE SDR = .11 PERCENT																	
L105	67.44	-.42	-1.06	.13	.59	97.77	1.99	2.02	.12	1.02	65T	0	L105										
L213	68.17	.32	.79	.14	.63	97.06	1.28	1.29	.20	1.75	65T	0	L213										
L223	68.54	.68	1.70	.11	.48	97.84	2.05	2.08	.11	.93	65G	0	L223										
L232	69.50	1.64	4.12	.27	1.21	98.00	2.21	2.25	.00	.00	65P	0	L232										
L241	68.14	.28	.70	.21	.97	96.44	.65	.66	.12	1.04	65I	0	L241										
L249	69.12	1.27	3.18	.10	.47	95.97	.19	.19	.07	.62	65P	0	L249										
L256	66.94	-.92	-2.31	.14	.64	96.05	2.26	2.30	.13	1.15	65H	0	L256										
L260	68.31	.45	1.14	.36	1.65	96.64	.85	.86	.27	2.38	65P	0	L260										
L277	75.00	7.14	17.92	.00	.00	93.00	-2.79	-2.83	.00	.00	65P	0	L277										
L278	68.00	.14	.36	.00	.00	96.00	.21	.22	.00	.00	65P	0	L278										
L301	68.19	.33	.83	.17	.78	97.36	1.58	1.60	.07	.65	65G	0	L301										
L312	70.12	2.27	5.69	.35	1.60	96.00	.21	.22	.00	.00	65P	0	L312										
L321	70.44	2.58	6.47	.18	.80	99.00	3.21	3.26	.27	2.34	65P	0	L321										
L328	70.00	2.14	5.37	.00	.00	95.22	-.56	-.57	.32	2.84	65P	0	L328										
L339	70.39	2.53	6.35	.26	1.17	98.51	2.73	2.77	.18	1.58	65P	0	L339										
L442	67.25	-.61	-1.53	.09	.42	97.45	1.66	1.69	.12	1.05	65T	0	L442										
L562	74.00	6.14	15.41	.00	.00	95.50	-.29	-.29	.00	.00	65P	0	L562										
L591	66.70	-1.16	-2.90	.09	.40	98.75	2.97	3.01	.12	1.08	65H	0	L591										
L617	70.00	2.14	5.37	.20	.91	97.00	1.21	1.23	.11	.94	65G	0	L617										
L626	70.06	2.20	5.53	.18	.80	96.25	.46	.47	.27	2.34	65P	0	L626										
L684	67.16	-.70	-1.75	.24	1.08	92.16	-3.62	-3.68	.21	1.81	65H	0	L684										
L695	70.06	2.20	5.53	.18	.80	99.69	3.90	3.96	.26	2.27	65P	0	L695										
L698	67.84	-.02	-.05	.18	.80	96.70	.91	.93	.08	.66	65I	0	L698										

TOTAL NUMBER OF LABORATORIES REPORTING = 45

Best values: J79 67.9 + 0.7 percent
E79 95.9 + 1.5 percent

DIRECTIONAL BLUE REFLECTANCE IN PERCENT


TAPPI STANDARD T452 GS-77, "BRIGHTNESS"; MARTIN SWEETS (ACBT & GE) IS STANDARD FOR THIS ANALYSIS

LAB CGDE	F	MEANS J79	MEANS E79	COORDINATES MAJOR	MINOR	R _s SDR VAR	PROPERTY---TEST INSTRUMENT---CONDITIONS
L591	+	66.70	98.75	2.61	1.82	.74 65B	BLUE REFLECTANCE (DIRECTIONAL), HUNTER
L285	*	66.76	93.09	-2.88	.44	1.71 65N	BLUE REFLECTANCE (DIRECTIONAL), TECBNIDYNE/DIANG/M. _s , S-4
L256	*	66.94	98.05	1.99	1.42	.89 65H	BLUE REFLECTANCE (DIRECTIONAL), HUNTER
L176A	G	67.12	94.69	-1.24	.46	.78 65A	BLUE REFLECTANCE (DIRECTIONAL), MARTIN SWEETS (ACBT), S-2
L684	+	67.16	92.16	-3.69	-1.17	1.45 65H	BLUE REFLECTANCE (DIRECTIONAL), HUNTER
L442	+	67.25	97.45	1.48	.98	.73 65T	BLUE RBFLCTANC(B (DIRECTIONAL), HUNTER D25D2M
L598	G	67.42	96.41	.51	.57	1.61 65M	BLUE REFLECTANCE (DIRECTIONAL), MARTIN SWEETS (GE), S-1
L105	*	67.44	97.77	1.84	.87	.81 65T	BLUE REFLECTANCE (DIRECTIONAL), HUNTER D25D2M
L673R	G	67.51	96.42	.54	.49	.99 65N	BLUE REFLECTANCE (DIRECTIONAL), TECBNIDYNE/DIANG/M. _s , S-4
L288	G	67.67	95.60	-2.22	.13	.47 65N	BLUE REFLECTANCE (DIRECTIONAL), TECBNIDYNE/DIANG/M. _s , S-4
L317	G	67.75	95.24	-56	-02	1.16 65M	BLUE REFLECTANCE (DIRECTIONAL), MARTIN SWEETS (GE), S-1
L210M	G	67.76	95.64	-17	.06	1.57 65N	BLUE REFLECTANCE (DIRECTIONAL), MARTIN SWEETS (GE), S-1
L698	*	67.84	96.70	.88	.23	.73 65I	BLUE REFLECTANCE (DIRECTIONAL), HUNTER D25D2A
L210N	G	67.86	96.44	.63	.15	.91 65N	BLUE REFLECTANCE (DIRECTIONAL), TECHNIDYNE/DIANG/M. _s , S-4
L122	G	67.90	95.61	-16	-08	.58 65N	BLUE REFLECTANCE (DIRECTIONAL), TECHNIDYNE/DIANG/M. _s , S-4
L190C	G	67.91	95.80	.03	.05	.93 65A	BLUE REFLECTANCE (DIRECTIONAL), MARTIN SWEETS (ACBT), S-2
L692	G	67.91	96.67	.88	.16	.94 65N	BLUE REFLECTANCE (DIRECTIONAL), TECBNIDYNE/DIANG/M. _s , S-4
L259	G	67.92	95.76	-01	.07	.59 65M	BLUE REFLECTANCE (DIRECTIONAL), MARTIN SWEETS (GE), S-1
L243	G	67.92	95.92	.15	.03	.73 65A	BLUE REFLECTANCE (DIRECTIONAL), MARTIN SWEETS (ACBT), S-2
L225	G	67.94	96.14	.36	.01	2.66 65N	BLUE REFLECTANCE (DIRECTIONAL), TECHNIDYNE/DIANG/M. _s , S-4
L636	G	68.00	97.00	1.21	.15	1.38 65M	BLUE REFLECTANCE (DIRECTIONAL), MARTIN SWEETS (GE), S-1
L278	+	68.00	96.00	.24	-09	.00 65P	BLUE REFLECTANCE (DIRECTIONAL), PHOTOVOLT
L523	G	68.01	95.31	-42	.26	.51 65N	BLUE REFLECTANCE (DIRECTIONAL), TECHNIDYNE/DIANG/M. _s , S-4
L308	G	68.04	96.37	.61	.04	.68 65N	BLUE REFLECTANCE (DIRECTIONAL), TECHNIDYNE/DIANG/M. _s , S-4
L132	G	68.09	96.01	.27	.17	.67 65N	BLUE REFLECTANCE (DIRECTIONAL), TECHNIDYNE/DIANG/M. _s , S-4
L565	G	68.10	95.96	.23	.19	.78 65A	BLUE REFLECTANCE (DIRECTIONAL), MARTIN SWEETS (ACBT), S-2
L241	+	68.14	96.44	.70	.12	1.00 65I	BLUE REFLECTANCE (DIRECTIONAL), HUNTER D25D2A
L213	+	68.17	97.06	1.31	.01	1.19 65T	BLUE REFLECTANCE (DIRECTIONAL), HUNTER D25D2M
L301	+	68.19	97.36	1.61	.05	.72 65G	BLUE REFLECTANCE (DIRECTIONAL), GARDNER
L211	*	68.20	93.59	-2.06	-85	2.49 65N	BLUE REFLECTANCE (DIRECTIONAL), TECHNIDYNE/DIANG/M. _s , S-4
L260	+	68.31	96.64	.93	.24	2.02 65P	BLUE REFLECTANCE (DIRECTIONAL), PHOTOVOLT
L158	G	68.50	97.21	1.54	.29	1.26 65N	BLUE REFLECTANCE (DIRECTIONAL), TECHNIDYNE/DIANG/M. _s , S-4
L223	+	68.54	97.84	2.15	.18	.70 65G	BLUE REFLECTANCE (DIRECTIONAL), GARDNER
L275	G	68.56	96.40	.76	.54	.83 65M	BLUE REFLECTANCE (DIRECTIONAL), MARTIN SWEETS (GE), S-1
L249	+	69.12	95.97	.48	-1.19	.54 65P	BLUE REFLECTANCE (DIRECTIONAL), PHOTOVOLT
L232	+	69.50	98.00	2.54	-1.08	.61 65P	BLUE REFLECTANCE (DIRECTIONAL), PHOTOVOLT
L617	*	70.00	97.00	1.68	-1.80	.92 65G	BLUE REFLECTANCE (DIRECTIONAL), GARDNER
L328	*	70.00	95.22	-05	-2.21	1.42 65F	BLUE REFLECTANCE (DIRECTIONAL), PHOTOVOLT
L695	*	70.06	99.69	4.31	-1.23	1.53 65P	BLUE REFLECTANCE (DIRECTIONAL), PHOTOVOLT
L626	*	70.06	96.25	.97	-2.03	1.57 65P	BLUE REFLECTANCE (DIRECTIONAL), PHOTOVOLT
L312	*	70.12	96.00	.74	-2.15	.80 65P	BLUE REFLECTANCE (DIRECTIONAL), PHOTOVOLT
L339	*	70.39	98.51	3.24	-1.82	1.38 65P	BLUE REFLECTANCE (DIRECTIONAL), PHOTOVOLT
L321	*	70.44	99.00	3.73	-1.76	1.57 65P	BLUE REFLECTANCE (DIRECTIONAL), PHOTOVOLT
L562	*	74.00	95.50	1.16	-6.04	.00 65F	BLUE REFLECTANCE (DIRECTIONAL), PHOTOVOLT
L277	*	75.00	93.00	-1.04	-7.59	.00 65F	BLUE REFLECTANCE (DIRECTIONAL), PHOTOVOLT
GMEANS:		67.86	95.79			1.00	
		95% ELLIPSE:		2.74	.89		WITH GAMMA = 76 DEGREES

BLUE REFLECTANCE, DIRECTIONAL

SAMPLE J79 = 67.9 PERCENT

SAMPLE E79 = 95.8 PERCENT



TAPPI COLLABORATIVE REFERENCE PROGRAM
ANALYSIS T65-2 TABLE 1

APRIL 1979

DIFFUSE BLUE REFLECTANCE IN PERCENT (GLOSS TRAP)
TAPPI SUGGESTED METHOD TS25 SU-72, BRIGHTNESS OF PULP (DIFFUSE ILLUMINATION AND 0 DEG. OBSERVATION)

LAB CODE	SAMPLE J79	PRINTING					SAMPLE E79	PRINTING					TEST D. #	S
		73 GRAMS MEAN	DEV	N _e DEV	SDR	R _e SDR		116 GRAMS MEAN	DEV	N _e DEV	SDR	R _e SDR	VAR	
L100	67.09	-0.77	-0.92	-0.14	1.02	97.03	0.68	-0.97	-0.10	1.34	65F	6	L100	
L121	69.52	1.65	1.96	0.12	.86	96.80	0.45	-0.64	-0.11	1.37	65K	6	L121	
L150	66.97	-0.89	-1.06	-0.12	.89	96.83	0.48	-0.69	-0.14	1.75	65Q	6	L150	
L170	67.62	-0.24	-0.29	-0.05	.33	95.11	-1.24	-1.77	-0.04	0.45	65B	6	L170	
L182	67.89	0.02	0.03	0.09	.67	96.02	-0.33	-0.47	0.06	1.15	65P	6	L182	
L210K	69.69	1.82	2.17	0.16	1.17	95.85	-0.50	-0.72	-0.10	1.30	65K	6	L210K	
L236	67.86	-0.00	-0.00	0.19	1.39	97.27	0.92	1.32	0.06	.83	65F	6	L236	
L242	67.44	-0.43	-0.51	0.17	1.19	96.50	0.15	-0.22	0.09	1.13	65F	6	L242	
L250T	67.96	0.10	0.12	0.07	.51	96.47	0.12	-0.18	0.07	.92	65F	6	L250T	
L280	67.49	-0.37	-0.44	0.09	.67	95.96	-0.39	-0.56	0.08	.97	65Q	6	L280	
L325	66.57	-1.29	-1.53	0.26	1.86	93.15	-3.20	-4.59	0.09	1.19	65F	#	L325	
L349	67.57	-0.30	-0.35	0.17	1.22	97.07	0.72	1.03	0.12	1.50	65K	6	L349	
L446	67.41	-0.45	-0.54	0.08	.55	96.06	-2.29	-0.42	0.02	.28	65F	6	L446	
L573	67.33	-0.53	-0.63	0.11	.78	97.50	1.15	1.64	0.08	.97	65F	6	L573	
L575	67.75	-0.12	-0.14	0.11	.82	95.08	-1.27	-1.82	0.08	.98	65F	6	L575	
L598	67.33	-0.54	-0.64	0.07	.53	96.32	-0.03	-0.04	0.07	.91	65K	6	L598	
L636	69.41	1.54	1.83	0.52	3.74	96.40	0.05	0.06	0.05	.69	65K	6	L636	
L680	67.37	-0.50	-0.59	0.09	.67	95.68	-0.67	-0.96	0.04	.45	65K	6	L680	

GR. MEAN = 67.87 PERCENT
SD MEANS = .84 PERCENTGRAND MEAN = 96.35 PERCENT
SD OF MEANS = .70 PERCENTTEST DETERMINATIONS = 8
17 LABS IN GRAND MEANS

AVERAGE SDR = .14 PERCENT

AVERAGE SDR = .08 PERCENT

L289 67.81 -0.05 -0.06 0.16 1.12 95.09 -1.26 -1.81 .17 2.22 65G L289

TOTAL NUMBER OF LABORATORIES REPORTING = 19

Best values: J79 67.8 + 1.7 percent
E79 96.4 + 1.2 percentThe following laboratories were omitted from the
grand means because of extreme test results: 325.TAPPI COLLABORATIVE REFERENCE PROGRAM
ANALYSIS T65-2 TABLE 2

APRIL 1979


TAPPI SUGGESTED METHOD TS25 SU-72, BRIGHTNESS OF PULP (DIFFUSE ILLUMINATION AND 0 DEG. OBSERVATION)

LAB CODE	F	MEANS J79	MEANS E79	COORDINATES MAJOR	COORDINATES MINOR	Avg R _e SDR	Var	PROPERTY---TEST INSTRUMENT---CONDITIONS
L325	#	66.57	93.15	-0.53	-3.41	1.53	65F	DIFFUSE REFLECTANCE, ELREPHG, GL _e TRAP, NRC-PTB ABSOLUTE BASE
L150	6	66.97	96.83	-0.98	0.26	1.32	65Q	DIFFUSE REFLECTANCE, ELREPHG, GL _e TRAP, ZEISS ABSOLUTE BASE
L100	6	67.09	97.03	-0.91	0.49	1.18	65F	DIFFUSE REFLECTANCE, ELREPHG, GL _e TRAP, NRC-PTB ABSOLUTE BASE
L598	6	67.33	96.32	-0.52	-0.15	0.72	65K	DIFFUSE REFLECTANCE, ELREPHG, GL _e TRAP, MGG (ZEISS) BASE
L573	6	67.33	97.50	-0.78	1.00	0.87	65F	DIFFUSE REFLECTANCE, ELREPHG, GL _e TRAP, NRC-PTB ABSOLUTE BASE
L680	6	67.37	95.68	-0.33	-0.76	0.56	65K	DIFFUSE REFLECTANCE, ELREPHG, GL _e TRAP, MGG (ZEISS) BASE
L446	6	67.41	96.06	-0.37	-0.38	0.41	65F	DIFFUSE REFLECTANCE, ELREPHG, GL _e TRAP, NRC-PTB ABSOLUTE BASE
L242	6	67.44	96.50	-0.45	0.05	1.16	65F	DIFFUSE REFLECTANCE, ELREPHG, GL _e TRAP, NRC-PTB ABSOLUTE BASE
L280	6	67.49	95.96	-0.27	-0.46	0.82	65Q	DIFFUSE REFLECTANCE, ELREPHG, GL _e TRAP, ZEISS ABSOLUTE BASE
L349	6	67.57	97.07	-0.45	0.63	1.36	65K	DIFFUSE REFLECTANCE, ELREPHG, GL _e TRAP, MGG (ZEISS) BASE
L170	6	67.62	95.11	0.05	-1.26	0.39	65B	DIFFUSE REFLECTANCE, ELREPHG, GL _e TRAP, NBS ABSOLUTE BASE
L575	6	67.75	95.08	0.17	-1.26	0.90	65F	DIFFUSE REFLECTANCE, ELREPHG, GL _e TRAP, NRC-PTB ABSOLUTE BASE
L289	+	67.81	95.09	0.24	-1.24	1.67	65G	DIFFUSE REFLECTANCE, ELREPHG, GL _e TRAP, SPECIFIC CALIBRATION
L236	6	67.86	97.27	-0.21	0.90	1.11	65F	DIFFUSE REFLECTANCE, ELREPHG, GL _e TRAP, NRC-PTB ABSOLUTE BASE
L182	6	67.89	96.02	0.10	-0.32	0.91	65F	DIFFUSE REFLECTANCE, ELREPHG, GL _e TRAP, NRC-PTB ABSOLUTE BASE
L250T	6	67.96	96.47	0.07	0.14	0.72	65F	DIFFUSE REFLECTANCE, ELREPHG, GL _e TRAP, NRC-PTB ABSOLUTE BASE
L636	6	69.41	96.40	1.49	0.39	2.22	65K	DIFFUSE REFLECTANCE, ELREPHG, GL _e TRAP, MGG (ZEISS) BASE
L121	6	69.52	96.80	1.51	0.81	1.12	65K	DIFFUSE REFLECTANCE, ELREPHG, GL _e TRAP, MGG (ZEISS) BASE
L210K	6	69.69	95.85	1.89	-0.07	1.24	65K	DIFFUSE REFLECTANCE, ELREPHG, GL _e TRAP, MGG (ZEISS) BASE
GMEANS:		67.87	96.35		1.00			
95% ELLIPSE:		2.38	1.93		WITH GAMMA =-13 DEGREES			

BLUE REFLECTANCE, DIFFUSE, WITH TRAP

SAMPLE J79 = 67.9 PERCENT

SAMPLE E79 = 96.4 PERCENT



TAPPI SUGGESTED METHOD T525 SU-72, BRIGHTNESS OF PULP (DIFFUSE ILLUMINATION AND 0 DEG. OBSERVATION)

LAB CODE	SAMPLE	PRINTING					SAMPLE	PRINTING					TEST D.	=	8	
	J79	73 GRAMS PER SQUARE METER	MEAN	DEV	N _e DEV	SDR	R _e SDR	E79	116 GRAMS PER SQUARE METER	MEAN	DEV	N _e DEV	SDR	R _e SDR	VAR	F
L115	67.05	-1.33	-1.76	.05	.39	92.95	-2.01	-1.59	.12	1.59	65E	6	L115			
L152	68.19	-0.18	-0.24	.11	.77	93.95	-1.00	-0.79	.05	.71	65E	6	L152			
L157	68.78	.40	.53	.10	.72	97.34	2.39	1.89	.05	.71	65E	6	L157			
L161	69.09	.71	.95	.10	.72	94.70	-2.26	-0.21	.06	.89	65E	6	L161			
L173A	68.53	.15	.20	.07	.54	93.82	-1.14	-0.90	.05	.73	65E	6	L173A			
L238A	68.77	.39	.52	.38	2.76	94.33	-0.63	-0.50	.04	.48	65E	6	L238A			
L241	67.06	-1.32	-1.75	.17	1.22	94.68	-2.28	-0.22	.07	1.02	65E	6	L241			
L244	69.28	.90	1.20	.09	.62	95.84	.88	.70	.09	1.29	65D	6	L244			
L255	69.61	1.23	1.63	.11	.83	96.14	1.19	.94	.05	.75	65D	6	L255			
L305	67.55	-0.83	-1.10	.16	1.15	93.09	-1.87	-1.48	.11	1.49	65D	6	L305			
L309	68.49	.12	.15	.26	1.92	95.71	.75	.59	.08	1.04	65J	6	L309			
L360	68.81	.43	.57	.14	.98	95.76	.81	.64	.08	1.03	65E	6	L360			
L384	68.31	-0.07	-0.09	.08	.61	95.97	1.02	.80	.07	.97	65S	6	L384			
L565	68.36	-0.02	-0.02	.13	.95	94.06	-0.89	-0.71	.09	1.26	65W	6	L565			
L685	67.78	-0.59	-0.79	.11	.82	96.01	1.05	.83	.08	1.03	65E	6	L685			
GR. MEAN = 68.38 PERCENT						GRAND MEAN = 94.96 PERCENT					TEST DETERMINATIONS = 8					
SD MEANS = .75 PERCENT						SD OF MEANS = 1.27 PERCENT					15 LABS IN GRAND MEANS					
AVERAGE SDR = .14 PERCENT											AVERAGE SDR = .07 PERCENT					
TOTAL NUMBER OF LABORATORIES REPORTING = 15																
Best values: J79 68.4 ± 1.3 percent																
E79 95.0 ± 1.9 percent																


TAPPI SUGGESTED METHOD T525 SU-72, BRIGHTNESS OF PULP (DIFFUSE ILLUMINATION AND 0 DEG. OBSERVATION)

LAB CODE	F	MEANS	COORDINATES	AVG	R _e SDR	VAR	PROPERTY---TEST INSTRUMENT---CONDITIONS
J79	E79	MAJOR	MINOR				
L115	6	67.05	92.95	-2.37	.47	.99	65E DIFFUSE REFLECTANCE, ELREPHG, NG TRAP, MGG (ZEISS) BASE
L241	6	67.06	94.68	-.76	1.11	1.12	65E DIFFUSE REFLECTANCE, ELREPHG, NG TRAP, MGG (ZEISS) BASE
L305	6	67.55	93.09	-2.05	.06	1.32	65D DIFFUSE REFLECTANCE, ELREPHG, NG TRAP, NRC-PTB ABSOLUTE
L665	6	67.78	96.01	.75	.95	.92	65E DIFFUSE REFLECTANCE, ELREPHG, NG TRAP, MGG (ZEISS) BASE
L152	6	68.19	93.95	-1.00	-.21	.74	65E DIFFUSE REFLECTANCE, ELREPHG, NG TRAP, MGG (ZEISS) BASE
L384	6	68.31	95.97	.92	.45	.79	65S DIFFUSE REFLECTANCE, ELREPHG, NG TRAP, ABSOLUTE-UNKNOWN BASE
L565	6	68.36	94.06	-.83	-.33	1.10	65W DIFFUSE REFLECTANCE, ELREPHG, NG TRAP, NBS MGG BASE
L309	6	68.49	95.71	.74	.18	1.48	65J DIFFUSE REFLECTANCE, ELREPHG, NG TRAP, NBS ABSOLUTE
L173A	6	68.53	93.82	-1.00	-.57	.63	65E DIFFUSE REFLECTANCE, ELREPHG, NG TRAP, MGG (ZEISS) BASE
L238A	6	68.77	94.33	-.43	-.60	1.62	65E DIFFUSE REFLECTANCE, ELREPHG, NG TRAP, MGG (ZEISS) BASE
L157	6	68.78	97.34	2.36	.53	.71	65E DIFFUSE REFLECTANCE, ELREPHG, NG TRAP, MGG (ZEISS) BASE
L360	6	68.81	95.76	.91	-.09	1.01	65E DIFFUSE REFLECTANCE, ELREPHG, NG TRAP, MGG (ZEISS) BASE
L161	6	69.09	94.70	.03	-.76	.81	65E DIFFUSE REFLECTANCE, ELREPHG, NG TRAP, MGG (ZEISS) BASE
L244	6	69.28	95.84	1.16	-.50	.96	65D DIFFUSE REFLECTANCE, ELREPHG, NG TRAP, NRC-PTB ABSOLUTE
L255	6	69.61	96.14	1.57	-.69	.79	65D DIFFUSE REFLECTANCE, ELREPHG, NG TRAP, NRC-PTB ABSOLUTE
GMEANS:	68.38	94.96		1.00			
95% ELLIPSE:	3.85	1.72		WITH GAMMA = 67 DEGREES			

BLUE REFLECTANCE, DIFFUSE, NO TRAP

SAMPLE J79 = 68.4 PERCENT

SAMPLE E79 = 95.0 PERCENT



SPECULAR GLOSS AT 75 DEGREES, IN GLOSS UNITS
TAPPI STANDARD T480 GS-78, SPECULAR GLOSS OF PAPER AND PAPERBOARD AT 75 DEGREES

LAB CODE	SAMPLE E51	COATED BACK				SAMPLE P91	COATED PRINTING				TEST D. = 10
		MEAN	DEV	N. DEV	SDR		MEAN	DEV	N. DEV	SDR	
L121	67.05	-0.99	-0.49	1.19	1.15	47.88	-1.92	-1.04	.54	.63	75H G L121
L122	67.49	-0.55	-0.27	.83	.80	49.39	-0.41	-0.22	.77	.89	75H G L122
L128	66.40	-1.64	-0.81	1.07	1.03	51.70	1.90	1.03	1.06	1.22	75G # L128
L134	70.00	1.96	.96	1.15	1.11	50.50	.70	.38	1.08	1.24	75H G L134
L149	77.60	9.56	4.69	1.26	1.22	62.70	12.90	7.02	1.16	1.34	75G # L149
L153	72.35	4.31	2.11	1.53	1.47	52.95	3.15	1.71	.60	.69	75G G L153
L162	72.48	4.44	2.18	1.74	1.67	53.48	3.68	2.00	.81	.93	75G G L162
L173A	63.90	-4.14	-2.03	.57	.55	46.00	-3.80	-2.07	.67	.77	75G G L173A
L182	67.63	-0.41	-0.20	.81	.78	48.86	-0.94	-0.51	.69	.80	75H G L182
L189	66.00	-2.04	-1.00	1.03	.99	50.95	1.15	.63	.55	.63	75P G L189
L190C	66.06	-1.98	-0.97	1.17	1.13	48.04	-1.76	-0.96	1.14	1.31	75G G L190C
L206	68.14	.10	.05	.80	.77	49.19	-0.61	-0.33	.98	1.13	75H G L206
L210	72.26	4.22	2.07	.96	.92	52.69	2.89	1.57	1.03	1.18	75H G L210
L211	66.96	-1.08	-0.53	.82	.78	49.66	-0.14	-0.08	1.37	1.58	75R G L211
L212	66.80	-1.24	-0.61	1.32	1.27	52.00	2.20	1.20	.94	1.09	75P G L212
L213	67.83	-0.21	-0.10	.73	.70	49.35	-0.45	-0.24	.89	1.02	75H G L213
L223	66.25	-1.79	-0.88	.95	.91	48.28	-1.52	-0.83	1.16	1.34	75H G L223
L230	66.80	-1.24	-0.61	1.14	1.09	50.20	.40	.22	.92	1.06	75H G L230
L243	69.50	1.46	.72	1.51	1.45	50.00	.20	.11	.82	.94	75B G L243
L253P	70.45	2.41	1.18	.98	.95	50.09	.29	.16	1.06	1.22	75G G L253P
L255	68.70	.66	.32	.82	.79	50.50	.70	.38	.53	.61	75G G L255
L256	68.05	.01	.00	.83	.80	49.67	-0.13	-0.07	.57	.66	75H G L256
L259	70.61	2.57	1.26	1.36	1.31	51.69	1.89	1.03	1.19	1.37	75H G L259
L262	68.20	.16	.08	.63	.61	50.10	.30	.16	1.10	1.27	75K G L262
L277A	71.76	3.72	1.82	1.60	1.54	52.38	2.58	1.40	.79	.91	75H G L277A
L277B	71.90	3.86	1.89	.83	.80	53.20	3.40	1.85	.78	.90	75H G L277B
L278	66.18	-1.86	-0.91	1.10	1.06	48.86	-0.94	-0.51	.55	.63	75G G L278
L279	67.40	-0.64	-0.31	1.78	1.71	49.20	-0.60	-0.33	1.11	1.28	75G G L279
L291	68.21	.17	.08	1.07	1.03	49.97	.17	.09	.99	1.14	75H G L291
L301	68.38	.34	.17	.61	.59	49.74	-0.06	-0.03	.97	1.12	75H G L301
L317	68.90	.86	.42	.88	.84	49.80	.00	.00	1.32	1.52	75H G L317
L321	71.35	3.31	1.62	.24	.23	50.85	1.05	.57	.71	.82	75G G L321
L323	67.77	-0.27	-0.13	1.19	1.15	48.80	-1.00	-0.54	.71	.82	75H G L323
L328	67.77	-0.27	-0.13	1.03	.99	49.91	.11	.06	.82	.94	75H G L328
L339	67.80	-0.24	-0.12	.42	.41	53.00	3.20	1.74	.00	.00	75P # L339
L349	66.88	-1.16	-0.57	.99	.95	45.65	-4.15	-2.26	.91	1.05	75H G L349
L372	67.80	-0.24	-0.12	.48	.46	50.08	.28	.15	.58	.66	75B G L372
L388	60.15	-7.89	-3.87	1.03	.99	48.25	-1.55	-0.84	1.57	1.81	75P # L388
L396	63.65	-4.39	-2.15	3.06	2.94	50.95	1.15	.63	1.67	1.93	75G X L396
L456	67.02	-1.02	-0.50	1.24	1.19	48.03	-1.77	-0.96	.79	.91	75H G L456
L483	67.36	-0.68	-0.33	1.21	1.16	48.69	-1.11	-0.60	.58	.67	75H G L483
L573	65.80	-2.24	-1.10	1.40	1.34	49.70	-1.10	-0.05	.82	.95	75G G L573
L574	65.00	-3.04	-1.49	1.15	1.11	46.90	-2.90	-1.58	.57	.65	75G G L574
L583	68.92	.88	.43	.57	.55	51.01	1.21	.66	1.38	1.59	75H G L583
L592	65.64	-2.40	-1.18	1.29	1.24	47.08	-2.72	-1.48	1.06	1.22	75H G L592
L598	65.78	-2.26	-1.11	1.56	1.50	48.08	-1.72	-0.94	.88	1.01	75H G L598
L643	67.25	-0.79	-0.39	.87	.83	46.75	-0.05	-0.03	.65	.75	75H G L643
L668	66.15	-1.89	-0.93	1.14	1.09	47.49	-2.31	-1.26	.79	.91	75G G L668
L670	69.28	1.24	.61	1.02	.98	51.12	1.32	.72	.96	1.11	75H G L670
L688	67.74	-0.30	-0.15	.66	.64	48.10	-1.70	-0.92	.98	1.13	75G G L688

GR. MEAN = 68.04 GLOSS UNITS GRAND MEAN = 49.80 GLOSS UNITS TEST DETERMINATIONS = 10
 SD MEANS = 2.04 GLOSS UNITS SD OF MEANS = 1.84 GLOSS UNITS 47 LABS IN GRAND MEANS
 AVERAGE SDR = 1.04 GLOSS UNITS AVERAGE SDR = .87 GLOSS UNITS

L250 58.30 -9.74 -4.78 2.54 2.44 45.40 -4.40 -2.39 .84 .97 75Q * L250
 L288 68.45 .41 .20 .91 .87 49.71 -.09 -.05 .83 .96 75I * L288
 L697 66.79 -1.25 -0.61 1.24 1.19 47.43 -2.37 -1.29 .56 .65 75X * L697

TOTAL NUMBER OF LABORATORIES REPORTING = 53

Best values: E51 68 + 4 gloss units
E91 50 ± 3 gloss units


The following laboratories were omitted from the grand means because of extreme test results: 149, 388.

ANALYSIS T75-1 TABLE 2
SPECULAR GLOSS AT 75 DEGREES, IN GLOSS UNITS
TAPPI STANDARD T480 GS-78, SPECULAR GLOSS OF PAPER AND PAPERBOARD AT 75 DEGREES

LAB CODE	F	MEANS E51	MEANS E91	COORDINATES MAJOR	MINOR	R. SDR	VAR	PROPERTY---TEST INSTRUMENT---CONDITIONS
L250	+	58.30	45.40	-10.24	3.08	1.71	75Q	SPECULAR GLOSS (75 DEGREE), PHOTOVOLTAIC, 20 C, 65% RH
L388	#	60.15	48.25	-6.57	4.01	1.40	75P	SPECULAR GLOSS (75 DEGREE), PHOTOVOLTAIC
L396	X	63.65	50.95	-2.56	3.75	2.43	75G	SPECULAR GLOSS (75 DEGREE), GARDNER
L173A	0	63.90	46.00	-5.62	-0.15	.66	75G	SPECULAR GLOSS (75 DEGREE), GARDNER
L574	0	65.00	46.90	-4.20	-0.19	.88	75G	SPECULAR GLOSS (75 DEGREE), GARDNER
L592	0	65.64	47.08	-3.60	-0.47	1.23	75H	SPECULAR GLOSS (75 DEGREE), HUNTER
L598	0	65.78	48.08	-2.83	.19	1.25	75H	SPECULAR GLOSS (75 DEGREE), HUNTER
L573	0	65.80	49.70	-1.76	1.40	1.15	75G	SPECULAR GLOSS (75 DEGREE), GARDNER
L189	0	66.00	50.95	-0.76	2.21	.81	75P	SPECULAR GLOSS (75 DEGREE), PHOTOVOLTAIC
L190C	0	66.06	48.04	+2.65	-0.03	1.22	75G	SPECULAR GLOSS (75 DEGREE), GARDNER
L668	0	66.15	47.49	-2.94	-0.50	1.00	75G	SPECULAR GLOSS (75 DEGREE), GARDNER
L278	0	66.18	48.86	-2.02	.51	.84	75G	SPECULAR GLOSS (75 DEGREE), GARDNER
L223	0	66.25	48.28	-2.35	.03	1.12	75H	SPECULAR GLOSS (75 DEGREE), HUNTER
L128	*	66.40	51.70	.01	2.51	1.13	75G	SPECULAR GLOSS (75 DEGREE), GARDNER
L697	+	66.79	47.43	-2.50	-0.97	.92	75X	SPECULAR GLOSS (75 DEGREE): GIVE INSTRUMENT MAKE + MODEL
L212	0	66.80	52.00	.51	2.48	1.18	75P	SPECULAR GLOSS (75 DEGREE), PHOTOVOLTAIC
L230	0	66.80	50.20	-0.67	1.12	1.07	75H	SPECULAR GLOSS (75 DEGREE), HUNTER
L349	*	66.88	45.65	-3.60	-2.37	1.00	75B	SPECULAR GLOSS (75 DEGREE), HUNTER
L211	0	66.96	49.66	-0.91	.61	1.18	75H	SPECULAR GLOSS (75 DEGREE), HUNTER
L456	0	67.02	48.03	-1.93	-0.66	1.05	75H	SPECULAR GLOSS (75 DEGREE), HUNTER
L121	0	67.05	47.88	-2.01	-0.80	.89	75H	SPECULAR GLOSS (75 DEGREE), HUNTER
L643	0	67.25	49.75	-0.63	.48	.79	75H	SPECULAR GLOSS (75 DEGREE), HUNTER
L483	0	67.36	48.69	-1.24	-0.39	.92	75H	SPECULAR GLOSS (75 DEGREE), HUNTER
L279	0	67.40	49.20	-0.88	-0.03	1.49	75G	SPECULAR GLOSS (75 DEGREE), GARDNER
L122	0	67.49	49.39	-0.68	.05	.84	75H	SPECULAR GLOSS (75 DEGREE), HUNTER
L182	0	67.63	48.86	-0.93	-0.44	.79	75B	SPECULAR GLOSS (75 DEGREE), HUNTER
L668	0	67.74	48.10	-1.34	-1.08	.88	75G	SPECULAR GLOSS (75 DEGREE), GARDNER
L328	0	67.77	49.91	-0.13	.26	.97	75B	SPECULAR GLOSS (75 DEGREE), HUNTER
L323	0	67.77	48.80	-0.86	-0.58	.99	75H	SPECULAR GLOSS (75 DEGREE), HUNTER
L372	0	67.80	50.02	.00	.37	.56	75B	SPECULAR GLOSS (75 DEGREE), BAUSCH + LOMB
L339	*	67.80	53.00	1.92	2.57	.20	75P	SPECULAR GLOSS (75 DEGREE), PHOTOVOLTAIC
L213	0	67.83	49.35	-0.45	-0.20	.86	75H	SPECULAR GLOSS (75 DEGREE), HUNTER
L256	0	68.05	45.67	-0.08	.10	.73	75H	SPECULAR GLOSS (75 DEGREE), HUNTER
L206	0	68.14	49.19	-0.33	-0.52	.95	75H	SPECULAR GLOSS (75 DEGREE), HUNTER
L262	0	68.20	50.10	.32	.12	.94	75K	SPECULAR GLOSS (75 DEGREE), GAERTNER (K-C TYPE)
L291	0	68.21	49.97	.24	.02	1.08	75H	SPECULAR GLOSS (75 DEGREE), HUNTER
L301	0	68.38	49.74	.22	-0.27	.85	75H	SPECULAR GLOSS (75 DEGREE), HUNTER
L288	+	68.45	49.71	.25	-0.34	.92	75I	SPECULAR GLOSS (75 DEGREE), HUNTER, 20 C, 65% RH
L255	0	68.70	50.50	.96	.10	.70	75G	SPECULAR GLOSS (75 DEGREE), GARDNER
L317	0	68.90	49.80	.65	-0.56	1.18	75H	SPECULAR GLOSS (75 DEGREE), HUNTER
L583	0	68.92	51.01	1.46	.34	1.07	75H	SPECULAR GLOSS (75 DEGREE), HUNTER
L670	0	69.28	51.12	1.80	.18	1.04	75H	SPECULAR GLOSS (75 DEGREE), HUNTER
L243	0	69.50	50.00	1.23	-0.81	1.20	75B	SPECULAR GLOSS (75 DEGREE), BAUSCH + LOMB
L134	0	70.00	50.50	1.94	-0.76	1.18	75K	SPECULAR GLOSS (75 DEGREE), HUNTER
L253P	0	70.45	50.09	2.01	-1.36	1.08	75G	SPECULAR GLOSS (75 DEGREE), GARDNER
L259	0	70.61	51.69	3.18	-0.26	1.34	75M	SPECULAR GLOSS (75 DEGREE), HUNTER
L321	0	71.35	50.85	3.19	-1.38	.52	75G	SPECULAR GLOSS (75 DEGREE), GARDNER
L277A	0	71.76	52.38	4.50	-0.49	1.23	75H	SPECULAR GLOSS (75 DEGREE), HUNTER
L277B	0	71.90	53.20	5.14	.03	.85	75H	SPECULAR GLOSS (75 DEGREE), HUNTER
L210	0	72.26	52.69	5.08	-0.59	1.05	75H	SPECULAR GLOSS (75 DEGREE), HUNTER
L153	0	72.35	52.95	5.32	-0.45	1.08	75G	SPECULAR GLOSS (75 DEGREE), GARDNER
L162	0	72.48	53.48	5.76	-0.14	1.30	75G	SPECULAR GLOSS (75 DEGREE), GARDNER
L149	#	77.60	62.70	15.68	3.46	1.28	75G	SPECULAR GLOSS (75 DEGREE), GARDNER
GMEANS:		68.04	49.80			1.00		
		95% ELLIPSE:	6.56	2.50			WITH GAMMA = 41 DEGREES	

SPECULAR GLOSS, 75 DEGREE

SAMPLE E51 = 68.0 GLOSS UNITS SAMPLE E91 = 49.8 GLOSS UNITS



LAB CGDE	SAMPLE J83 MEAN	PRINTING					SAMPLE J66 MEAN	PRINTING					TEST D. = 10 VAR F LAB		
		73 GRAMS PER SQUARE METER			SDR	R. SDR		93 GRAMS PER SQUARE METER			SDR	R. SDR			
		DEV	N. DEV					DEV	N. DEV						
L100	2.810	.098	1.42	.028	.70		6.451	.162	1.38	.040	.34	90V	G L100		
L105	2.771	.059	.85	.029	.73		6.438	.149	1.26	.100	.84	90Q	G L105		
L118	2.697	-.015	-.22	.037	.92		6.332	.043	.36	.127	1.06	90Q	G L118		
L122	2.760	.048	.69	.057	1.43		6.296	.007	.06	.126	1.06	90V	G L122		
L123F	2.865	.153	2.21	.034	.84		6.510	.221	1.88	.188	1.57	90F	G L123F		
L125	2.602	-.110	-1.60	.051	1.26		6.047	-.242	-2.06	.107	.89	90T	G L125		
L128	2.712	-.000	-.00	.029	.73		6.228	-.061	-.52	.095	.79	90T	G L128		
L141	2.686	-.026	-.38	.040	1.01		6.173	-.116	-.99	.124	1.04	90T	G L141		
L153	2.782	.070	1.01	.034	.84		6.144	-.145	-1.23	.099	.83	90T	* L153		
L158	2.710	-.002	-.03	.032	.79		6.390	.101	.86	.057	.47	90T	G L158		
L159	2.717	.005	.07	.023	.58		6.377	.088	.75	.093	.77	90T	G L159		
L162	2.680	-.032	-.47	.063	1.58		6.240	-.049	-.42	.143	1.19	90D	G L162		
L166	2.653	-.059	-.86	.032	.80		6.299	.010	.08	.078	.65	90T	G L166		
L173B	2.786	.074	1.07	.047	1.17		6.370	.081	.69	.086	.71	90F	G L173B		
L174	2.700	-.012	-.18	.047	1.18		6.280	-.005	-.08	.215	1.79	90T	G L174		
L182	2.723	.010	.15	.051	1.27		6.308	.019	.16	.109	.91	90L	G L182		
L183	2.744	.032	.46	.024	.60		6.425	.136	1.15	.071	.60	90T	G L183		
L190C	2.640	-.072	-1.05	.052	1.29		6.140	-.149	-1.27	.117	.98	90T	G L190C		
L203A	2.600	-.112	-1.62	.053	1.32		6.125	-.164	-1.39	.177	1.48	90T	G L203A		
L203C	2.760	.048	.69	.074	1.84		6.360	.071	.60	1.17	.98	90T	G L203C		
L212	2.785	.073	1.05	.034	.84		6.470	.181	1.54	.103	.86	90T	G L212		
L213	2.590	-.122	-1.77	.032	.79		6.360	.071	.60	1.17	.98	90T	* L213		
L223	2.690	-.022	-.32	.042	1.06		6.254	-.035	-.30	.116	.97	90V	G L223		
L228	2.740	.028	.40	.061	1.53		6.300	.011	.09	.131	1.10	90T	G L228		
L233	2.724	.012	.17	.028	.70		6.323	.034	.29	.155	1.29	90Q	G L233		
L238A	2.642	-.070	-1.02	.027	.68		6.178	-.111	-.94	.180	1.50	90T	G L238A		
L241	2.895	.183	2.65	.016	.39		6.275	-.014	-.12	.142	1.18	90T	X L241		
L242D	2.752	.040	.58	.039	.98		6.198	-.091	-.77	.135	1.12	90G	G L242D		
L242P	2.706	-.007	-.10	.031	.78		6.207	-.082	-.70	.088	.73	90P	G L242P		
L249	2.648	-.064	-.93	.034	.86		6.286	-.003	-.03	.070	.59	90T	G L249		
L259	2.645	-.067	-.97	.035	.87		6.038	-.251	-2.13	.206	1.72	90T	G L259		
L260	2.664	-.048	-.70	.020	.50		6.362	.073	.62	.033	.27	90T	G L260		
L261	2.760	.048	.69	.024	.61		6.458	.169	1.43	.049	.41	90T	G L261		
L262	2.695	-.017	-.25	.028	.71		6.120	-.169	-1.44	.132	1.10	90T	G L262		
L285	2.610	-.102	-1.48	.032	.79		6.300	.011	.09	.094	.79	90T	G L285		
L291	2.850	.138	2.00	.053	1.32		6.380	.091	.77	.193	1.61	90T	G L291		
L305	2.720	.002	.11	.042	1.05		6.330	.041	.35	.114	.95	90T	G L305		
L309	2.650	-.062	-.90	.053	1.32		6.140	-.149	-1.27	.184	1.53	90T	G L309		
L318	2.625	-.087	-1.26	.042	1.06		6.185	-.104	-.88	.153	1.28	90T	G L318		
L320	.003	-2.705	-39.24	.000	.00		.006	-6.283	-53.36	.000	.00	90T	* L320		
L323	2.686	-.026	-.38	.024	.59		6.296	.007	.06	.145	1.21	90T	G L323		
L324	2.740	.028	.40	.039	.58		6.265	-.024	-.20	.041	.34	90T	G L324		
L326	2.765	.053	.77	.047	1.18		6.515	.226	1.92	.078	.65	90T	G L326		
L328	2.706	-.006	-.09	.055	1.36		6.300	.011	.09	.081	.67	90T	G L328		
L331	2.789	.077	1.11	.042	1.04		6.216	-.073	-.62	.062	.52	90T	G L331		
L339	2.660	-.052	-.76	.032	.79		6.245	-.044	-.37	.171	1.42	90T	G L339		
L341	2.780	.062	.98	.039	.98		6.436	.147	1.25	.138	1.15	90T	G L341		
L352	2.720	.008	.11	.032	.80		6.343	.054	.46	.068	.57	90Q	G L352		
L356	2.684	-.028	-.41	.050	1.26		6.240	-.049	-.42	.188	1.57	90T	G L356		
L358	2.702	-.010	-.15	.038	.95		6.291	.002	.02	.119	.99	90T	G L358		
L376	2.800	.068	1.27	.053	1.32		6.370	.081	.69	.134	1.12	90T	G L376		
L382	2.790	.078	1.13	.057	1.42		6.320	.031	.26	.220	1.84	90T	G L382		
L390	2.846	.134	1.94	.028	.71		6.504	.215	1.83	.047	.39	90T	G L390		
L442	2.816	.104	1.50	.023	.57		6.548	.259	2.20	.091	.76	90V	G L442		
L556	2.770	.058	.84	.042	1.05		6.265	-.024	-.20	.180	1.50	90T	G L556		
L557	2.620	-.092	-1.33	.042	1.05		6.140	-.149	-1.27	.084	.70	90T	G L557		
L567	2.672	-.040	-.58	.060	1.50		6.269	-.020	-.17	.153	1.27	90V	G L567		
L571	2.710	-.002	-.03	.074	1.84		6.180	-.109	-.93	.193	1.61	90V	G L571		
L574	2.650	-.062	-.90	.034	.86		6.232	-.057	-.48	.125	1.04	90V	G L574		
L575	2.720	.008	.11	.023	.58		6.207	-.082	-.70	.046	.38	90T	G L575		
L581	2.760	.048	.69	.052	1.29		6.420	.131	1.11	.114	.95	90T	G L581		
L585	2.530	-.182	-2.64	.048	1.21		6.210	-.079	-.67	.179	1.50	90T	* L585		
L626	2.610	-.102	-1.48	.011	.26		6.128	-.161	-1.37	.156	1.30	90T	G L626		
L679	2.710	-.002	-.03	.046	1.15		6.130	-.159	-.35	.160	1.34	90T	G L679		
L693	2.737	.025	.36	.032	.79		6.319	.030	.25	.089	.75	90T	G L693		

GR. MEAN = 2.712 MILS
SD MEAN = .069 MILS

GRAND MEAN = 6.289 MILS
SD OF MEANS = .118 MILS

TEST DETERMINATIONS - 10
63 LABS IN GRAND MEANS

MILS

• MILS

GRÖMETER **GRAND**

OMETER

TAPPI COLLABORATIVE REFERENCE PROGRAM
ANALYSIS T90-1 TABLE 1
THICKNESS (CALIPER), THOUSANDS OF AN INCH
TAPPI STANDARD T411 GS-76

APRIL 1979

LAB CODE	SAMPLE J83	PRINTING					SAMPLE J66	PRINTING					TEST D _e = 10			
		MEAN	73 GRAMS PER SQUARE METER	DEV	N _o DEV	SDR		MEAN	93 GRAMS PER SQUARE METER	DEV	N _o DEV	SDR	R _e SDR	VAR	F	LAB
L106	27.100	24.388	353.23	.994	24.83	60.000	53.711	456.18	.000	.00	90C	. L106				
L185	2.660	-.052	-.76	.038	.94	6.136	-.153	-.1.30	.103	.86	90B	. L185				
L203B	2.540	-.172	-2.49	.052	1.29	6.110	-.179	-.1.52	.129	1.07	90C	. L203B				
L243	2.684	-.028	-.41	.032	.81	6.138	-.151	-.1.28	.059	.49	90S	. L243				
L344	2.587	-.125	-.1.81	.036	.90	6.164	-.125	-.1.06	.184	1.54	90U	. L344				
L396M	2.700	-.012	-.18	.033	.83	6.035	-.254	-.2.16	.145	1.21	90S	. L396M				
L563	2.730	-.018	.26	.048	1.21	6.320	-.031	.26	.123	1.03	90U	. L563				
L576	2.640	-.072	-1.05	.070	1.75	6.590	-.301	2.56	.137	1.14	90C	. L576				
L684	2.690	-.022	-.32	.032	.79	6.250	-.039	-.33	.097	.81	90U	. L684				
TOTAL NUMBER OF LABORATORIES REPORTING = 74																

Best values: J83 2.72 ± 0.11 mils
 J66 6.29 ± 0.18 mils

Data from the following laboratories appear to be off by a multiplicative factor: 320.

TAPPI COLLABORATIVE REFERENCE PROGRAM
ANALYSIS T90-1 TABLE 2
THICKNESS (CALIPER), THOUSANDS OF AN INCH
TAPPI STANDARD T411 GS-76

APRIL 1979

LAB CODE	MEANS		COORDINATES		AVG E.S.DR VAR	PROPERTY---TEST INSTRUMENT---CONDITIONS
	F	J83	J66	MAJOR	MINOR	
L320 #	.003	.006	-6.840	-.150	.00	90T THICKNESS (CALIPER), TMI, MOTOR DRIVEN
L585 *	2.530	6.210	-.148	.133	1.35	90T THICKNESS (CALIPER), TMI, MOTOR DRIVEN
L203B *	2.540	6.110	-.234	.082	1.18	90C THICKNESS (CALIPER), CADY, HAND DRIVEN
L344 *	2.587	6.164	-.166	.062	1.22	90U THICKNESS (CALIPER), TMI, HAND DRIVEN
L213 *	2.590	6.360	.014	.141	.88	90T THICKNESS (CALIPER), TMI, MOTOR DRIVEN
L203A @	2.600	6.125	-.196	.034	1.40	90T THICKNESS (CALIPER), TMI, MOTOR DRIVEN
L125 @	2.602	6.047	-.266	-.001	1.08	90T THICKNESS (CALIPER), TMI, MOTOR DRIVEN
L626 @	2.610	6.128	-.189	.026	.78	90T THICKNESS (CALIPER), TMI, MOTOR DRIVEN
L285 @	2.610	6.300	-.033	.097	.79	90T THICKNESS (CALIPER), TMI, MOTOR DRIVEN
L557 @	2.620	6.140	-.174	.022	.88	90T THICKNESS (CALIPER), TMI, MOTOR DRIVEN
L318 @	2.625	6.185	-.131	.036	1.17	90T THICKNESS (CALIPER), TMI, MOTOR DRIVEN
L576 *	2.640	6.590	.244	.191	1.44	90C THICKNESS (CALIPER), CADY, HAND DRIVEN
L190C @	2.640	6.140	-.166	.004	1.13	90T THICKNESS (CALIPER), TMI, MOTOR DRIVEN
L238A @	2.642	6.172	-.130	.018	1.09	90T THICKNESS (CALIPER), TMI, MOTOR DRIVEN
L259 @	2.645	6.038	-.256	-.043	1.30	90T THICKNESS (CALIPER), TMI, MOTOR DRIVEN
L249 @	2.648	6.286	-.029	.057	.72	90T THICKNESS (CALIPER), TMI, MOTOR DRIVEN
L574 @	2.650	6.232	-.078	.033	.95	90V THICKNESS (CALIPER), TMI, MOTOR DRIVEN, DIGITIZED
L309 @	2.650	6.140	-.161	-.005	1.43	90T THICKNESS (CALIPER), TMI, MOTOR DRIVEN
L166 @	2.653	6.299	-.016	.058	.72	90T THICKNESS (CALIPER), TMI, MOTOR DRIVEN
L185 *	2.660	6.136	-.161	-.016	.90	90B THICKNESS (CALIPER), AMTHOR, HAND DRIVEN
L339 @	2.660	6.245	-.062	.029	1.11	90T THICKNESS (CALIPER), TMI, MOTOR DRIVEN
L260 @	2.664	6.362	.046	.074	.39	90T THICKNESS (CALIPER), TMI, MOTOR DRIVEN
L567 @	2.672	6.269	-.035	.028	1.39	90V THICKNESS (CALIPER), TMI, MOTOR DRIVEN, DIGITIZED
L162 @	2.680	6.240	-.058	.009	1.39	90D THICKNESS (CALIPER), CADY, MOTOR DRIVEN
L243 *	2.684	6.138	-.149	-.037	.65	90S THICKNESS (CALIPER), SCHOPPER, HAND DRIVEN
L356 @	2.684	6.240	-.056	.005	1.41	90T THICKNESS (CALIPER), TMI, MOTOR DRIVEN
L141 @	2.686	6.173	-.116	-.024	1.02	90T THICKNESS (CALIPER), TMI, MOTOR DRIVEN
L323 @	2.686	6.296	-.005	.027	.90	90T THICKNESS (CALIPER), TMI, MOTOR DRIVEN
L684 *	2.690	6.250	-.045	.004	.80	90U THICKNESS (CALIPER), TMI, HAND DRIVEN
L223 @	2.690	6.254	-.041	.006	1.01	90V THICKNESS (CALIPER), TMI, MOTOR DRIVEN, DIGITIZED
L262 @	2.695	6.120	-.161	-.055	.90	90T THICKNESS (CALIPER), TMI, MOTOR DRIVEN
L118 @	2.697	6.332	.033	.032	.99	90Q THICKNESS (CALIPER), EMVECG, MOTOR DRIVEN
L396N *	2.700	6.035	-.236	-.095	1.02	90S THICKNESS (CALIPER), SCHOPPER, HAND DRIVEN
L174 @	2.700	6.280	-.013	.007	1.49	90T THICKNESS (CALIPER), TMI, MOTOR DRIVEN
L358 @	2.702	6.291	-.002	.010	.97	90T THICKNESS (CALIPER), TMI, MOTOR DRIVEN
L242P @	2.706	6.207	-.077	-.028	.76	90P THICKNESS (CALIPER), MESSMER, MOTOR DRIVEN, ISG R534
L328 @	2.706	6.300	.007	.010	1.02	90T THICKNESS (CALIPER), TMI, MOTOR DRIVEN
L679 @	2.710	6.130	-.146	-.064	1.24	90T THICKNESS (CALIPER), TMI, MOTOR DRIVEN
L571 @	2.710	6.180	-.100	-.043	1.73	90V THICKNESS (CALIPER), TMI, MOTOR DRIVEN, DIGITIZED
L158 @	2.710	6.390	.091	.044	.63	90T THICKNESS (CALIPER), TMI, MOTOR DRIVEN
L128 @	2.712	6.228	-.056	-.025	.76	90T THICKNESS (CALIPER), TMI, MOTOR DRIVEN
L159 @	2.717	6.377	.082	.032	.68	90T THICKNESS (CALIPER), TMI, MOTOR DRIVEN
L575 @	2.720	6.207	-.071	-.041	.48	90T THICKNESS (CALIPER), TMI, MOTOR DRIVEN
L352 @	2.720	6.343	.052	.015	.68	90Q THICKNESS (CALIPER), EMVECG, MOTOR DRIVEN
L305 @	2.720	6.330	.040	.010	1.00	90T THICKNESS (CALIPER), TMI, MOTOR DRIVEN
L182 @	2.723	6.308	.021	-.002	1.09	90L THICKNESS (CALIPER), L + W, MOTOR DRIVEN
L233 @	2.724	6.323	.036	.003	.59	90Q THICKNESS (CALIPER), EMVECG, MOTOR DRIVEN
L563 *	2.730	6.320	.036	-.003	1.12	90U THICKNESS (CALIPER), TMI, HAND DRIVEN
L693 @	2.737	6.319	.038	-.010	.77	90T THICKNESS (CALIPER), TMI, MOTOR DRIVEN
L228 @	2.740	6.300	.022	-.021	1.31	90T THICKNESS (CALIPER), TMI, MOTOR DRIVEN
L324 @	2.740	6.265	-.010	-.035	.66	90T THICKNESS (CALIPER), TMI, MOTOR DRIVEN
L183 @	2.744	6.425	.137	.028	.60	90T THICKNESS (CALIPER), TMI, MOTOR DRIVEN
L242@	2.752	6.198	-.066	-.074	1.05	90G THICKNESS (CALIPER), MESSMER, MOTOR DRIVEN, BS3983
L261 @	2.760	6.458	.174	.027	.51	90T THICKNESS (CALIPER), TMI, MOTOR DRIVEN
L122 @	2.760	6.296	.026	-.041	1.24	90V THICKNESS (CALIPER), TMI, MOTOR DRIVEN, DIGITIZED
L203C @	2.760	6.360	.084	-.014	1.41	90T THICKNESS (CALIPER), TMI, MOTOR DRIVEN
L581 @	2.760	6.420	.139	.011	1.12	90T THICKNESS (CALIPER), TMI, MOTOR DRIVEN
L326 @	2.765	6.515	.227	.046	.92	90T THICKNESS (CALIPER), TMI, MOTOR DRIVEN
L556 @	2.770	6.265	.002	-.063	1.28	90T THICKNESS (CALIPER), TMI, MOTOR DRIVEN
L105 @	2.771	6.438	.160	.008	.78	90Q THICKNESS (CALIPER), EMVECG, MOTOR DRIVEN
L341 *	2.780	6.436	.162	-.001	1.06	90T THICKNESS (CALIPER), TMI, MOTOR DRIVEN
L153 *	2.782	6.144	-.103	-.124	.83	90T THICKNESS (CALIPER), TMI, MOTOR DRIVEN
L212 @	2.785	6.470	.195	.009	.85	90T THICKNESS (CALIPER), TMI, MOTOR DRIVEN
L173B @	2.786	6.370	.104	-.033	.94	90F THICKNESS (CALIPER), FEDERAL, MOTOR DRIVEN
L331 @	2.789	6.216	-.034	-.100	.78	90T THICKNESS (CALIPER), TMI, MOTOR DRIVEN

TAPPI COLLABORATIVE REFERENCE PROGRAM
 ANALYSIS T90-1 TABLE 2
 THICKNESS (CALIPER), THOUSANDS OF AN INCH
 TAPPI STANDARD T411 GS-76

APRIL 1979

LAB CODE	F	MEANS J83	MEANS J66	COORDINATES MAJOR	COORDINATES MINOR	Avg R.E.S.D.R VAR	PROPERTY---TEST INSTRUMENT---CONDITIONS
L382	G	2.790	6.320	.061	-.058	1.63 90T THICKNESS (CALIPER), TMI,	MOTOR DRIVEN
L376	G	2.800	6.370	.110	-.046	1.22 90T THICKNESS (CALIPER), TMI,	MOTOR DRIVEN
L100	G	2.810	6.451	.188	-.022	.52 90V THICKNESS (CALIPER), TMI,	MOTOR DRIVEN, DIGITIZED
L442	G	2.816	6.548	.279	.013	.66 90V THICKNESS (CALIPER), TMI,	MOTOR DRIVEN, DIGITIZED
L390	G	2.846	6.504	.251	-.032	.55 90T THICKNESS (CALIPER), TMI,	MOTOR DRIVEN
L291	G	2.850	6.380	.140	-.087	1.46 90T THICKNESS (CALIPER), TMI,	MOTOR DRIVEN
L123F	G	2.865	6.510	.264	-.047	1.21 90F THICKNESS (CALIPER), FEDERAL,	MOTOR DRIVEN
L241	X	2.895	6.275	.063	-.172	.79 90T THICKNESS (CALIPER), TMI,	MOTOR DRIVEN
L106	*	27.100	60.000	58.988	.172	12.41 90C THICKNESS (CALIPER), CADY,	HAND DRIVEN
GMEANS:		2.712	6.289			1.00	
95% ELLIPSE:				.323	.123	WITH GAMMA = 65 DEGREES	


THICKNESS (CALIPER)

SAMPLE J83 = 2.71 MILS

SAMPLE J83 = 68.9 MICROMETER

SAMPLE J66 = 6.29 MILS

SAMPLE J66 = 159.7 MICRÖMETER



TAPPI COLLABORATIVE REFERENCE PROGRAM
 ANALYSIS T95-1 TABLE 1
 GRAMMAGE (MASS PER UNIT AREA)
 TAPPI STANDARD T410 GS-68

APRIL 1979

LAB CODE	SAMPLE D33 MEAN	MANILA ENVELOPE 118 GRAMS PER SQUARE METER				SAMPLE D34 MEAN	OFFSET ENAMEL COATED 117 GRAMS PER SQUARE METER				TEST D. = 10 VAR F LAB
		DEV	N. DEV	SDR	R. SDR		DEV	N. DEV	SDR	R. SDR	
L100	118.10	.31	.57	1.10	.82	119.10	.41	.49	.74	.86	95C # L100
L121	116.88	-.91	-1.66	1.03	.77	117.07	-1.62	-1.93	1.04	1.20	95B # L121
L162	118.08	.29	.54	1.91	1.43	119.93	1.24	1.48	1.17	1.36	95K # L162
L213	116.85	-.93	-1.71	1.73	1.29	118.04	-.65	-.78	.74	.85	95F # L213
L233	118.09	.30	.56	1.55	1.16	117.20	-1.49	-1.78	.58	.67	95T # L233
L249	123.86	6.07	11.14	1.01	.75	118.48	-.21	-.25	.64	.74	95I # L249
L280	117.65	-.14	-.25	1.72	1.29	119.22	.53	.63	1.15	1.34	95T # L280
L305	118.64	.85	1.57	1.51	1.13	118.38	-.31	-.37	.74	.86	95T # L305
L339	117.70	-.09	-.16	.48	.36	119.00	.31	.37	.00	.00	95T # L339
L342	117.83	.04	.07	1.60	1.19	119.58	.89	1.07	.44	.51	95C # L342
L344	118.33	.54	1.00	.33	.25	119.13	.44	.52	.13	.15	95T # L344
L442	117.78	-.01	-.01	.81	.61	118.46	-.23	-.28	.28	.33	95K # L442
L557	117.31	-.48	-.87	2.06	1.54	118.58	-.11	-.13	1.93	2.24	95C # L557
L559	118.13	.34	.63	1.56	1.17	118.99	.30	.36	.84	.97	95K # L559
L567	116.78	-1.01	-1.85	1.01	.75	117.45	-1.24	-1.48	1.62	1.88	95E # L567
L571	117.40	-.39	-.71	1.61	1.20	118.05	-.64	-.77	2.29	2.66	95P # L571
L574	117.73	-.06	-.10	1.23	.92	119.70	1.01	1.21	1.47	1.71	95D # L574
L597	118.16	.37	.68	1.24	.93	119.34	.65	.78	.55	.64	95C # L597
L688	117.91	.12	.23	2.23	1.67	118.46	-.23	-.28	.44	.51	95T # L688
L693	118.60	.81	1.49	.67	.50	119.44	.75	.90	.22	.25	95G # L693

GR. MEAN = 117.79 G/SQ.METER

SD MEANS = .55 G/SQ.METER

AVERAGE SDR = 1.34 G/SQ.METER

TOTAL NUMBER OF LABORATORIES REPORTING = 20

GRAND MEAN = 118.69 G/SQ.METER

SD OF MEANS = .84 G/SQ.METER

AVERAGE SDR =

TEST DETERMINATIONS = 10

19 LABS IN GRAND MEANS

.86 G/SQ.METER

Best values: D33 117.8 + 1.0 grams per square meter
 D34 118.7 + 1.3 grams per square meterThe following laboratories were omitted from the
 grand means because of extreme test results: 249.

TAPPI COLLABORATIVE REFERENCE PROGRAM
 ANALYSIS T95-1 TABLE 2
 GRAMMAGE (MASS PER UNIT AREA)
 TAPPI STANDARD T410 GS-68

APRIL 1979

LAB CODE	F	MEANS D33	MEANS D34	COORDINATES MAJOR	COORDINATES MINOR	Avg R. SDR VAR	PROPERTY---TEST INSTRUMENT---CONDITIONS
L567	#	116.78	117.45	-1.55	.38	1.32 95E	BASIS WEIGHT (GRAMMAGE), GUILLOTINE TYPE CUTTER
L213	#	116.85	118.04	-.99	.56	1.07 95F	BASIS WEIGHT (GRAMMAGE), FOUR-SQUARE CUTTER
L121	#	116.88	117.07	-1.85	.13	.99 95B	BASIS WEIGHT (GRAMMAGE), CONCRA CUTTER
L557	#	117.31	118.58	-.30	.38	1.89 95C	BASIS WEIGHT (GRAMMAGE), CUTTING BOARD
L571	#	117.40	118.05	-.75	.07	1.93 95P	BASIS WEIGHT (GRAMMAGE), PRODUCTION REAM CUTTER
L280	#	117.65	119.22	.42	.35	1.32 95T	BASIS WEIGHT (GRAMMAGE), TEMPLATE CUT
L339	#	117.70	119.00	.24	.21	1.18 95T	BASIS WEIGHT (GRAMMAGE), TEMPLATE CUT
L574	#	117.73	119.70	.89	.48	1.31 95D	BASIS WEIGHT (GRAMMAGE), DIE CUT
L442	#	117.78	118.46	-.21	-.09	1.47 95K	BASIS WEIGHT (GRAMMAGE), WEIGHED AS RECEIVED
L342	#	117.83	119.58	.82	.35	1.85 95C	BASIS WEIGHT (GRAMMAGE), CUTTING BOARD
L688	#	117.91	118.46	-.16	-.21	1.09 95T	BASIS WEIGHT (GRAMMAGE), TEMPLATE CUT
L162	#	118.08	119.93	1.24	.27	1.39 95K	BASIS WEIGHT (GRAMMAGE), WEIGHED AS RECEIVED
L223	#	118.09	117.20	-1.22	-.91	1.92 95T	BASIS WEIGHT (GRAMMAGE), TEMPLATE CUT
L100	#	118.10	119.10	.50	-.11	1.84 95C	BASIS WEIGHT (GRAMMAGE), CUTTING BOARD
L559	#	118.13	118.99	.42	-.18	1.07 95K	BASIS WEIGHT (GRAMMAGE), WEIGHED AS RECEIVED
L597	#	118.16	119.34	.75	-.06	.78 95C	BASIS WEIGHT (GRAMMAGE), CUTTING BOARD
L344	#	118.33	119.13	.63	-.30	1.20 95T	BASIS WEIGHT (GRAMMAGE), TEMPLATE CUT
L693	#	118.60	119.44	1.03	-.41	1.38 95G	BASIS WEIGHT (GRAMMAGE), PRECISION CUTTER
L305	#	118.64	118.38	.09	-.90	1.00 95T	BASIS WEIGHT (GRAMMAGE), TEMPLATE CUT
L249	#	123.86	118.48	2.42	-5.58	.75 95I	BASIS WEIGHT (GRAMMAGE), INGENTO PAPER CUTTER


GMSEANS: 117.79 118.69

95% ELLIPSE: 2.49 1.17 WITH GAMMA = 64 DEGREES

GRAMMAGE (MASS PER UNIT AREA)

SAMPLE D33 = 117.8 G/SQ.METER

SAMPLE D34 = 118.7 G/SQ.METER



SUMMARY TABLE

TEST METHOD		SAMPLE CODE	GRAND MEAN	SD OF MEAN	AVER SDR	REPL CRP	LABS INCL	LABS PARTIC	REPL TAPPI	REPEAT	REPROD
AIR RESISTANCE, GURLEY T40-1	GURLEY UNITS	K43 B95	16.3 48.6	.9 3.2	1.1 5.2	10	54	59	10	1.0 4.6	2.4 8.8
AIR RESISTANCE, SHEFFIELD T40-2	SHEFF. UNITS	K43 B95	166.5 72.2	6.4 4.3	8.8 6.0	10	37	45	10	7.7 5.2	17.8 11.9
AIR RESISTANCE, GURLEY BG FLOTATION T41-1	SEC/10 CC	E65 B73	783. 1314.	75. 281.	173. 517.	10	13	15	10	152. 453.	209. 777.
SMOOTHNESS, PARKER PRINISURF T44-1	MICRONS	A83 J50	4.613 5.801	.168 .223	.150 .087	10	7	8	10	.132 .077	.466 .617
SMOOTHNESS, SHEFFIELD T45-1	SHEFF. UNITS	A83 J50	105.6 274.5	7.4 13.1	9.8 7.4	15	89	92	10	8.6 6.5	21.1 36.5
SMOOTHNESS, BEKK T45-2	BEKK SECONDS	A83 J50	52.85 11.27	3.72 1.80	7.56 .71	15	10	11	10	6.62 .63	11.00 5.00
SMOOTHNESS, BENDTSEN T47-1	ML/MIN	A83 J50	119. 516.	12. 45.	16. 51.	10	7	9	10	14. 45.	34. 125.
K & N INK ABSORPTION T56-1	K & N UNITS	B58 E50	60.6 63.8	6.7 6.0	.7 .4	4	9	11	4	1.0 .6	18.7 16.6
PH, COLD T57-1	PH UNITS	J75 J17	6.173 4.653	.087 .091	.067 .040	5	4	5	2	.132 .078	.263 .260
PH, HOT T57-2	PH UNITS	J75 J17	5.618 4.319	.090 .020	.037 .032	5	4	5	2	.073 .063	.255 .074
OPACITY, B&L TYPE, 89% BACKING T60-1	PERCENT	B21 K24	72.43 95.43	1.09 .38	1.06 .31	10	75	84	5	1.32 .38	3.16 1.09
OPACITY, B&L TYPE, PAPER BACKING T60-2	PERCENT	B21 K24	75.74 95.72	.74 .12	1.02 .34	10	4	4	5	1.27 .42	2.23 .44
OPACITY, ELREPHC TYPE, PAPER BACKING T60-3	PERCENT	B21 K24	76.97 96.23	.66 .12	.70 .17	10	15	17	5	.86 .21	1.92 .38
BLUE REFLECTANCE, DIRECTIONAL T65-1	PERCENT	J79 E79	67.86 95.79	.40 .99	.22 .11	8	22	45	6	.25 .13	1.11 2.73
BLUE REFLECTANCE, DIFFUSE, WITH TRAP T65-2	PERCENT	J79 E79	67.87 96.35	.84 .70	.14 .08	8	17	19	6	.16 .09	2.33 1.93
BLUE REFLECTANCE, DIFFUSE, NO TRAP T65-3	PERCENT	J79 E79	68.38 94.96	.75 1.27	.14 .07	8	15	15	6	.16 .08	2.09 3.51
SPECULAR GLOSS, 75 DEGREE T75-1	GLOSS UNITS	E51 E91	68.04 49.80	2.04 1.84	1.04 .87	10	47	53	5	1.29 1.08	5.72 5.15
THICKNESS (CALIPER) T90-1	MILS	J83 J66	2.712 6.289	.069 .118	.040 .120	10	63	74	10	.035 .105	.191 .326
GRAMMAGE (MASS PER UNIT AREA) T95-1	G/SQ. METER	D33 D34	117.79 118.69	.55 .84	1.34 .86	10	19	20	3	2.14 1.38	2.34 2.59

U.S. DEPT. OF COMM. BIBLIOGRAPHIC DATA SHEET		1. PUBLICATION OR REPORT NO. TAPPI CRP 59G	2. Gov't. Accession No.	3. Recipient's Accession No.
4. TITLE AND SUBTITLE Technical Association of the Pulp and Paper Industry COLLABORATIVE REFERENCE PROGRAM FOR PAPER Report #59G		5. Publication Date July 27, 1979		
7. AUTHOR(S) R. G. Powell, J. Horlick		8. Performing Organ. Report No. NBSIR 79-1801		
9. PERFORMING ORGANIZATION NAME AND ADDRESS NATIONAL BUREAU OF STANDARDS DEPARTMENT OF COMMERCE WASHINGTON, DC 20234		10. Project/Task/Work Unit No.		
12. SPONSORING ORGANIZATION NAME AND COMPLETE ADDRESS (<i>Street, City, State, ZIP</i>) Collaborative Testing Services, Inc. 5241 Wood Glade Drive, Great Falls, VA 22066 and Technical Association of the Pulp and Paper Industry		13. Type of Report & Period Covered FINAL		
15. SUPPLEMENTARY NOTES <input type="checkbox"/> Document describes a computer program; SF-185, FIPS Software Summary, is attached.		16. ABSTRACT (<i>A 200-word or less factual summary of most significant information. If document includes a significant bibliography or literature survey, mention it here.</i>) Collaborative Reference Programs provide participating laboratories with the means for checking periodically the level and uniformity of their testing in comparison with that of other participating laboratories. An important by-product of the programs is the provision of realistic pictures of the state of the testing art. This is one of the periodic reports showing averages for each participant, within and between laboratory variability, and other information for participants and standards committees.		
17. KEY WORDS (<i>six to twelve entries; alphabetical order; capitalize only the first letter of the first key word unless a proper name; separated by semicolons</i>) Collaborative reference program; Laboratory evaluation; Paper; Precision; Reference samples; Testing calibration				
18. AVAILABILITY <input type="checkbox"/> Unlimited XX For Official Distribution. Do Not Release to NTIS <input type="checkbox"/> Order From Sup. of Doc., U.S. Government Printing Office, Washington, DC 20402, SD Stock No. SN003-003- <input type="checkbox"/> Order From National Technical Information Service (NTIS), Springfield, VA, 22161		19. SECURITY CLASS (THIS REPORT) UNCLASSIFIED		21. NO. OF PRINTED PAGES 57
		20. SECURITY CLASS (THIS PAGE) UNCLASSIFIED		22. Price

