

# Analysis of Results of Mini Round Robin Reflectance Test

U.S. DEPARTMENT OF COMMERCE National Bureau of Standards Thermal Processes Division National Engineering Laboratory Center for Chemical Engineering Washington, DC 20234

August 1981



U.S. DEPARTMENT OF COMMERCE NATIONAL BUREAU OF STANDARDS

NBSIR 81-2311

## ANALYSIS OF RESULTS OF MINI ROUND ROBIN REFLECTANCE TEST

MATIONAL BUREAU OF STANDARDS LIBBARY SEP 16 1981 Not 00 11 QCIDO UE 1001231 1981 C.D

J. C. Richmond

U.S. DEPARTMENT OF COMMERCE National Bureau of Standards Thermal Processes Division National Engineering Laboratory Center for Chemical Engineering Washington, DC 20234

August 1981

U.S. DEPARTMENT OF COMMERCE, Malcolm Baldrige, Secretary NATIONAL BUREAU OF STANDARDS, Ernest Ambler, Director



### CONTENTS

	Page
LIST OF TABLES	iv
LIST OF FIGURES	v
INTRODUCTION	1
PARTICIPATING LABORATORIES	1
PRELIMINARY DISUCSSION	1
SAMPLES FOR MEASUREMENT	3
DATA REQUESTED	3
RESULTS OF MEASUREMENTS	3
APPENDIX I	7
MEASUREMENT PROCEDURE	8

### LIST OF TABLES

Table 1 - Results of Round Robin Test on White Samples

~

Table 2 - Results of Round Robin Test on Black Samples

Table 3 - Reported Values for Solar Properties

.


Table 4 - Procedures Used for Computing Solar Properties

Table 5 - Differences as Fraction of Reported Solar Values

#### LIST OF FIGURES

- Figure 1 Difference between the data reported by Laboratory 1 and calibration data for the white sample, plotted as a function of wavelength.
- Figure 2 Difference between data reported by Laboratory 2 and calibration data for the white sample, plotted as a function of wavelength.
- Figure 3 Difference between data reported by Laboratory 3 and calibration data for the white sample, plotted as a function of wavelength.
- Figure 4 Difference between data reported by Laboratory 4 and calibration data for the black sample, plotted as a function of wavelength.
- Figure 5 Difference between the data reported by Laboratory 1 and the calibration data for the black sample, plotted as a function of wavelength.
- Figure 6 Difference between the data reported by Laboratory 2 and calibration data for the black sample, plotted as a function of wavelength.
- Figure 7 Difference between the data reported by Laboratory 3 and the calibration data for the black sample, plotted as a function of wavelength.
- Figure 8 Difference between data reported by Laboratory 4 and the calibration data for the balck sample, plotted as a function of wavelength.

v



#### Introduction

A round-robin test of directional-hemispherical reflectance measurements was conducted by the National Bureau of Standards at the request of the Solar Energy Research Institute. Four national laboratories participated. Measurements were made of the directional-hemispherical reflectance of two samples, having high and low reflectance, respectively, in the wavelength range of 250 to 2500 nm.

Measured directional-hemispherical reflectance is used to compute solar reflectance of materials, and solar absorptance of opaque materials. The overall objective of the round-robin test was to obtain data on the performance of the several laboratories in making such measurements. Since all participants were national laboratories, it is to be expected that the accuracy of such measurements will approach the state of the art for measurements made with commercially available instruments.

#### Participating Laboratories

The participating laboratories, in alphabetical order are:

- 1. Battelle Pacific Northwest Laboratories, Richland, Washington
- National Bureau of Standards, Center for Building Technology, National Engineering Laboratory, Gaithersburg, Maryland (Mail address: Washington, D.C.)
- 3. Sandia National Laboratory, Albuquerque, New Mexico
- 4. Solar Energy Research Institute, Golden, Colorado

In addition to the above, the reference measurements on the samples to which the reported values were compared, were made in the Radiometric Physics Division of the Center for Radiation Research, National Measurement Laboratory of NBS. The high-accuracy reflectometer designed and built at NBS was used for these measurements.

#### Preliminary Discussion

An integrating sphere reflectometer, the instrument used in evaluating directional-hemispherical reflectance, actually measures directionalnear-hemispherical reflectance factor. Reflectance is defined as the ratio of reflected flux to incident flux. Reflectance factor is defined as the ratio of the flux reflected by a sample under specified conditions of irradiation and viewing, to that reflected by the ideal completely reflecting, isotropically diffusing surface. Such an ideal surface does not exist, but it is closely approximated by several materials which are near-perfect isotropic diffusers, and have high reflectance. If the reflectance of such a material is known, a correction can be made for its reflectance. It is physically impossible to design an instrument that has true directional irradiance and hemispherical collection, because the solid angles of irradiance and collection are mutually exclusive. Some of the reflected flux is lost out the entrance aperture of the sphere, and some is lost out the detector aperture. The fraction of the flux lost in this manner is approximately equal to the ratio of the total area of all apertures to the total area of the sphere wall, apertures included. If the diameter of the apertures is small compared to the diameter of the sphere, the fractional losses will be small.

In a reflectance factor measurement in which a comparison standard of known reflectance is used, the flux losses through the ports will be exactly compensated for if the relative geometric (directional) distribution of the reflected flux is the same for sample and standard. If both sample and standard reflect diffusely, the losses will be nearly compensated.

For true geometrical conditions of hemispherical collection, reflectance factor is identically equal to reflectance. For diffusely reflecting samples, the measured directional-hemispherical reflectance factor will be identically equal to the directional-hemispherical reflectance if the ideal comparison standard is used, and the error will be small if a real diffusely reflecting standard is used and a correction is made for the absolute reflectance of the standard.

Measurement of the absolute reflectance of a diffusely reflecting sample is difficult and time-consuming. It is normally done only by a primary standards laboratory, and then only at rather long intervals.

An integrating sphere reflectometer measures only the ratio of the reflectance factor of a sample to that of the comparison standard used in the measurement. The measured value will include uncertainties due to (1) the error, if any, in the reflectance value assigned to the comparison standard, (2) differences in the fractional flux losses out the ports, due to differences in the geometrical distributions of flux by the sample and reference, (3) nonlinearity of the detector-amplifier system, and (4) scattering in double-beam instruments, so that some of the flux in the reference beam is scattered into the sample beam, or vice versa. To this must be added an uncertainty due to the statistical fluctuation in the output of the instrument, usually referred to as noise.

With double-beam instruments, it is customary to run a 100% curve and a zero curve. The 100% curve is obtained with identical samples in the two beams. The zero curve is usually obtained when the sample beam is blocked before entering the sphere. A correction for deviation of the two curves from 100% and 0% respectively is made by use of the following equation.

$$R_{\lambda} = \frac{S_{\lambda}^{-0}\lambda}{H_{\lambda}^{-0}\lambda}$$
(1)

-2-

where R<sub> $\lambda$ </sub> is the corrected directional-hemispherical reflectance factor, S<sub> $\lambda$ </sub> is the sample reading, 0<sub> $\lambda$ </sub> is the zero curve reading, and H<sub> $\lambda$ </sub> is the 100% curve reading, all at wavelength  $\lambda$ .

The results of a round-robin test can be analyzed to give an estimate of the fractions of the total error that may be due to bias and random error, respectively, and may give some indication of the source of some of the bias.

#### Samples for Measurement

The samples sent to the participants were black and white samples, either 50.8x50.8 mm (2x2 inch) or 25.4x25.4 mm (1x1 inch) for the black, and 50.8x50.8 mm (2x2 inch) or 25.4x76.2 mm (1-1/2x3 inch) for the white. The smaller samples were intended for measurement in instruments utilizing an Edwards-type integrating sphere in which the sample is held at the center of the sphere by a rod inserted through an aperture in the top of the sphere. The aperture in these spheres is too small to admit a 50.8x50.8 mm (2x2 inch) sample. The black samples were of black porcelain enamel, with spectral directional-hemispherical reflectance in the range of about 0.06 to 0.11 in the wavelength range of 250 to 2500 nm. The white samples were white ceramic tile, with reflectance generally above 0.8 at wavelengths from about 500 to 2500 nm, dropping from about 0.8 at 500 nm to about 0.15 at 250 nm.

These samples were among those calibrated for directional-hemispherical reflectance in the Radiometric Physics Division of NBS.

#### Data Requested

The data requested from all participants was the measured directionalhemispherical reflectance for near-normal incidence, including the specular component, and if possible, the same data for angles of incidence of 15, 30, 45 and 60 degrees from normal. Data were requested at 10 nm intervals over as much of the wavelength range of 250 to 2500 nm as the measurement instrument was capable of providing. The instructions for the test are given in Appendix 1.

#### Results of Measurements

The results reported by the different laboratories are shown in Table 1 for the white samples and Table 2 for the black samples, together with the calibration values for the sample measured in each case.

The values reported for the solar reflectance of the white sample and the solar absorptance of the black sample are given in Table 3. The procedures used to compute the solar properties are given in Table 4.

Because different samples were measured by each laboratory, direct comparison of the reported values is not justified. The calibration values for each sample at each wavelength are included in Table 1 in the column headed "Cal", and the algebraic difference between the calibration value and the reported value is included in the column headed "Diff". The difference values obtained by subtracting the calibration value from the reported value were plotted as a function of wavelength as shown in figures 1 to 3. The average differences for the white samples were about 0.029 for Lab 1, 0.012 for Lab 2, 0.019 for Lab 3 and 0.010 for Lab 4. The standard deviations associated with these values were about 0.02 for Lab 1, 0.01 for Lab 2, 0.012 for Lab 3 and 0.003 for Lab 4.

A review of figures 1 to 4, where the data for the white samples are plotted, gives some clues to the errors that may be present in the reported values. Figure 1 shows that the data are slightly wavelength dependent. The average value of the difference is about 0.003, and the standard deviation about 0.006. The wavelength dependence and average difference suggests that the values assigned to the halon reference are in error, particularly at wavelengths below 600 nm.

Figure 2, showing the differences for Lab 2, indicates that the values assigned to the Halon reference are high at wavelengths below about 360 nm, and uniformly low by about .0013 at wavelengths beyond about 400 nm.

Figure 3, showing the differences for Lab 3, indicates that the errors are unacceptably high at wavelengths of 800 to 850 nm and at 1380 nm, and at the very short wavelengths in the ultraviolet and at wavelengths beyond 1900 nm in the infrared. The peaks at about 820 nm and 1380 nm may be at the points where detectors were changed, and the signal-tp-noise ratio is small. The signal-to-noise ratio may also be small at the two ends of the spectral range. The comparison standard used for the measurements was not identified.

Figure 4, showing the differences for Lab 4, indicates that the value assigned to the pressed Halon comparison standard may be low by about 0.01. The random error is the lowest of any of the laboratories.

The average differences for the black samples are about 0.0056 for Lab 1, 0.002 for Lab 2, 0.004 for Lab 3 and -.005 for Lab 4, and the standard deviations are about 0.0013 for Lab 1, 0.0028 for Lab 2, 0.0043 for Lab 3 and 0.0037 for Lab 4.

Figure 5 is a plot of the differences for Lab 1 and shows an average difference of 0.0056, and the standard deviation is about 0.0013. The standard deviation is the lowest reported for the black samples, and indicates a very low random error. There is a slight tendency for the error to be greater at the short and long wavelengths. The bias in the data, of about 0.5%, is much too large to be accounted for by an incorrect value for comparison standard, since the value assigned to the comparison standard would have to be about 8% high to account for such a large bias, and the data on the white sample indicates that such a large error is highly unlikely. The bias must be due to a zero line error. The equation used for reducing the data as reported by the investigator, does not show that a zero line correction was made. Figure 6 shows that the bias in the results reported by Lab 2 is very low, which is also indicated by the average difference of 0.00178, the lowest reported by any of the laboratories for the black samples. The standard deviation is about double that reported by Lab 1. There is no marked variation of the differences with wavelength.

Figure 7 shows that there is an overall trend of increasing difference with wavelength. The peak at about 800 to 850 nm, which is very prominent in figure 3, is still present, but the peak at about 1380 nm, if present, is masked by the random error, which is the largest reported for a black sample, as indicated by the standard deviation of about 0.004.

Figure 8 shows a slight trend toward increasing difference with wavelength. There is a very pronounced peak at about 900 nm and valley at about 1000 nm. This is near the wavelength where detectors are likely to be changed, and may be due to nonlinearity of the detectors in spectral regions where the response is low. The random error appears to be low, and the high standard deviation, 0.00433, appears to be due largely to fluctuations of the difference with wavelength. Again, it appears likely that the observed differences are due to zero line errors.

The coefficient of variation, the ratio of the standard deviation to the mean, may be a more useful measure of the significance of the errors in the measured reflectances. These are shown in Table 5, in which the mean reflectance is taken as the solar reflectance reported by each laboratory. The values for the white samples range from  $0.0038 \pm$ 0.0069 for Lab 1 to  $0.0225 \pm 0.0140$  for Lab 3, and for the black samples from  $0.0250 \pm 0.0399$  for Lab 2 to  $-.0811 \pm .0524$  for Lab 1. The average of the absolute values is  $0.0128 \pm .00915$  for the white samples, and  $0.0456 \pm 0.0432$ . In each case the  $\pm$  uncertainty is taken as one standard deviation. This is probably much too large for the uncertainty of the average values, for which the individual values vary from 47 to 220, but may be more reasonable for the uncertainty in a single value at one wavelength. The reported data provided essentially no information on which a valid evaluation of the true random error could be based, which would involve multiple measurements at each wavelength with each reflectometer.

Within the limitations of the available data, the figures in Table 5 give an estimate of the state of the art in reflectance measurements in the participating laboratories.

Some general conclusions can be drawn from the results of this round-robin test. The bias in the reported values can be largely eliminated by proper use of the reflectance standards recently made available by the National Bureau of Standards as SRM's 2019, 2020, 2021, and 2022 under the sponsorship of SERI. The errors are due partly to the use of incorrect reflectance values for the reference standards, and partly to 100% line and zero-line errors. The white standards can be used to check the values assigned to the reference standards, and also to evaluate the 100% errors. The black standards can be used to detect, and to help correct the zero line errors. The errors in the values reported by the participating laboratories are larger than were expected, and especially the errors in the measurement of solar absorptance. Such errors introduce rather large errors into the computations of the thermal efficiencies of solar collectors from the measured absorptance of the absorber.

#### Appendix I

#### Instructions for Round Robin Test

#### 1. Preparation of Samples

The reflectance of the samples may be significantly affected by dirt or surface films that are not easily seen. They should be cleaned before measurement, using the following procedure.

The white samples are glazed ceramic tiles that have been cut to size. The ceramic body is porous, and would absorb water if not sealed. The back and edges have been sealed with an epoxy base masonry sealant. If this seal is breached, the sample may absorb water, particularly if immersed. Absorbed water will change the reflectance, primarily in the infrared. The samples should be handled gently to avoid chipping or a breach of the seal, and should never be immersed in water. Momentary exposure to water as described in the cleaning procedure will not affect them if the seal is not broken.

The cleaning procedure is as follows for both black and white samples.

1. Rinse with tepid tap water.

2. Wash with a dilute solution of a mild liquid soap, such as Ivory Liquid, using gentle rubbing with a soft cloth or rubber sponge. Never use on abrasive cleaner.

3. Rinse with tepid tap water, to remove soap.

4. Rinse with distilled water. -

5. Blot dry with a fresh facial tissue. Do not rub.

6. Allow to air dry for at least 30 minutes before measuring.

Any other noncorrosive liquid soap is also suitable for this use.

#### Measurement Procedure

The data desired are the spectral directional-hemispherical reflectance for near-normal incidence, including the specular component. In addition, if you have the facilities, also the spectral directional-hemishperical reflectance for angles of incidence of 15, 30, 45 and 60 degrees from normal, including the specular component.

- 1. Calibrate your reflectometer, using your normal procedure.
- 2. Measure the samples, using your normal procedure, making corrections as required.
- 3. Report data at 10 nm intervals over as much of the wavelength range of 250 to 2500 nm as your reflectometer can cover.
- Compute solar reflectance of the white sample and solar absorptance of the black sample, using your normal procedures, and report the results.
- 5. Send the data to

J. C. Richmond Room B126 Metrology National Bureau of Standards Washington, DC 20234

6. Include with the data a detailed description of all procedures used, including the manufacturer and model number of all instruments used, the method of reducing data, and the solar spectral distribution and method of computation used for computing solar properties.

#### TABLE 1 - RESULTS OF ROUND ROBIN TEST ON WHITE SAMPLES

-

.

					-		·					
λ		Lab 1			Lab 2			Lab 3			Lab 4	
വമ	p	Cal	Diff	ρ	Cal	Diff	ρ	Cal	Diff	p	Cal	Diff
300	.134	. 147	013	.1470	.147	.0000	,2126	.140	.0726	.1446	.147	0024
310	.1.34	. 147	015	.1940	.190	.0040	.2212	.190	.0312	.1829	.190	0071
320				.2025	.238	0355	.2630	.238	.025 <b>0</b>	.2365	.238	0015
330				.3251	.303	0221	.3241	.303	.0211	.3063	. 303	.0033
340				.3304	.377	0466	.3995	.377	.0225	. 3082	.377	.0032
350	.437	.451	014	.4556	.460	0044	.4783	.460	.0183	.4620	.451	.0110
360				.5377	.541	0033	.5614	.541	.0204	. 54 38	.533	.0108
370				.6118	.608	.0038	.6266	.608	.0186	.6119	.601	.0109
380				.6616	.654	.0076	. 67 39	.654	.0199	.6593	.648	.0113
390				.7025	.690	.0125	.7096	.690	.0196	.6937	. 684	.0097
400	.703	.714	011	.734T	.720	.0147	.7382	.720	.0182	.7245	.714	.0105
410				.7536	.737	.0166	.7537	.737	.0167	.7427	.731	.0117
420				.7663	.749	.0173	.7743	.749	.0253	.7520	.742	.0100
430				.7755	.757	.0185	.7736	.767	.0166	.7631	.751	.0121
440				.7807	.762	.0187	.7782	.762	.0162	.7665	.756	.0105
450	.755	.761	006	.7824	.767	.0154	.7815	.767	.0145	.7716	.761	.0106
460				.7890	.772	.0170	.7858	.772	.0138	.7770	.765	.0120
470				.7934	.776	.0174	.7891	.776	.0131	.7819	.770	.0119
480				.7994	.780	.0194	.7938	.780	.0138	.7867	.774	.0127
490				.8046	.787	.0176	.8010	.787	.0140	.7930	.781	.0120
500	.783	.787	004	.8114	.793	.0184	.8074	.793	.0144	. 7993	.787	.0123
510				.8177	.799	.0187	.8125	.799	.0135	.8056	.793	.0126
520				.8229	.805	.0179	.8163	.805	.0113	.8111	.799	.0121
530				.8288	.811	.0178	.8217	.811	.0107	.8166	.805	.0116
540				.8339	.816	.0179	.8271	.816	.0111	.8231	.811	.0121
550	.813	.815	002	.8379	.820	.0179	.8295	.820	.0095	.8279	.815	.0129
560				•.8416	.825	.0166	.8346	.825	.0096	.8322	.819	.0132
570				.8451	.829	.0161	.8383	.829	.0093	.8351	.824	.0111
580				.8464	.831	.0154	.8367	.831	.0057	.8359	.826	.0099
590				.8484	.834	.0144	.8427	.834	.0087	. 8402	. 829	.0112
600	.832	.833	-,001	.8546	.837	.0176	.8468	.837	.0098	.8451	.833	.0121
610				.8568	.839	.0178	.8484	.839	.0094	.8478	.835	.0128
620				.8591	.842	.0171	.8523	.842	.0103	.8497	.836	.0137
630				.8610	.843	.0180	.8546	.843	.0116	.8503	.838	.0123
640				.8638	.846	.0178	8561	.846	.0101	.8513	.340	.0113
650	.841	.838	.003	.8629	. 844	.0189	.8567	.344	.0127	. 8519	.838	.0139
660				.8628	.848	.0148	.8614	. 848	. 01 34	.8561	.842	.0141
670				.8698	.851	.0188	.3646	.851	.0136	.8589	.844	.0149
680				.8722	.851	.0212	.8662	.851	.0152	.8609	.345	.0159
690				.8724	.853	.0194	.8660	.853	.0130	.8618	.847	.0148
700	.853	.851	.002									
700	.854	.851	.003	.8745	.857	.0175	.8717	.857	.0147	.8642	.851	.0132
710							.8768	.858	.0188	.8667	.853	.0137
720							.8801	.859	.0211	.8690	.354	.0150
730							.8842	.860	.0242	.8699	.854	.0159
740				.8718	.861	.0108	.8877	.861	.0267	.8686	.855	.0136
750	. 857	.856	.001									
750	.858	. 856	.002				. 8903	.862 .862	.0283	.8652	.856	.0092
760							.8986		.0366	.3648		
770							.9008	.863	.0378	.8642	.857	.0072
780				.8768	.863	.0138	.9051	.863	.0421	.8628	.856	.0068
790							.9100	.861	.0490	.8636	.855	.0086
800	.856	.854	.002									
800	.858	.854	.004.				.9129	.860	.0529	.8637	.854	.0097
810				•			.9128	.862	.0508	.3604	.856	.0044
820				.8718	.860	.0118	.9136	.860	.0536	.8610	.854	.0070
830							.9129	.859	.0539	.8606	.853	.0076
840							.9100	.857	.0530	.8591	.852	.0071
850	.851	.850	.001				.9043	.856	.0483	.3588	.850	.0088
860				.8640	.855	.0090	.8977	.855	.0427	.8597	.849	.0107
870							.8860	.853	.0330	.3581	.847	.0111
880							.8780	.852	.0260	.8567	.846	.0107
890							.8692	.852	.0172	.8549	.845	.0099

			•				Shernded					
λ		Lab 1			Lab 2			Lab 3			Lab 4	
<u>1110</u>	ρ	Cal	Diff	ρ	Cal	Diff	ρ	Cal	Diff	ρ	Cal	Diff
900 910 920 930	.847	.845	.002	.8576	.852 .853	.0056	.8645 .8590 .8584 .8566 .8564	.852 .853 .853 .853 .853	.0125 .0060 .0054 .0036 .0034	.8560 .8566 .8575 .8574 .8585	.845 .845 .846 .846 .846	.0110 .0116 .0115 .0114 .0125
940 950 960 970	.84 <b>9</b>	.846	.003			.0114	.8551 .8576 .8578	.853 .853 .853	.0021 .0046 .0048	.8586 .8585 .8589	.846 .846 .847	.0126 .0125 .0119
980 990				.8685	.856	.0125	.8585 .8598	.856 .856	.0025 .0038	.8595 .8624	.849 .849	.0105 .0134
1000 1010 1020 1030 1040	.851	.851	.000 *	.8733	.860	.0133	.8604 .8618 .8652 .86 <b>61</b> .8692	.858 .859 .860 .860 .861	.0024 .0028 .0052 .0061 .0082	.8631 .8630 .8636 .8631 .8626	.851 .852 .853 .854 .855	.0121 .0110 .0106 .0091 .0076
1050 1060 1070 1080 <b>1090</b>	.856	.856	.000	.8708	.861	.0 <b>09</b> 8	.8703 .8692 .8709 .8708 .8698	.861 .861 .861 .862 .862	.0093 .0082 .0099 .0088 .0078	.8625 .8626 .8627 .8627 .8641	.856 .857 .857 .857 .857	.0065 .0056 .0057 .0057 .0071
1100 1110 1120 1130 1140	.860	.857	.003	.8698	.861 .869	.0088	.8686 .8670 .8707 .8742 .8748	.361 .862 .865 .868 .869	.0076 .0050 .0057 .0062 .0058	.8650 .8655 .8665 .8686 .8696	.857 .857 .860 .862 .864	.0080 .0085 .0065 .0066 .0056
11 <b>5</b> 0 1160 1170 118 <b>0</b>	.867	.865	.002	.87 <b>9</b> 8	.871	.0088	.8783 .8777 .8795 .8812	.870 .871 .871 .871	.0083 .0067 .0085 .0102	.8710 .8716 .8726 .8730	.865 .867 .867 .866	.0060 .0046 .0056 .0070
1190 1200 1210	.870	.867	.003				.8826 .8833 .8823	.871 .872 .872	.0116 .0113 .0103	.8740 .8740 .8732	.867 .867 .867	.0070 .0070 .0062
1220 1230 1240				.8809	.872	.0089	.8828 .8818 .8827	.872 .872 .871	.0103 .0108 .00 <b>9</b> 8 .0117	.8732 .8732 .8737 .8737	.867 .867 .867	.0062 .0067 .0067
1250 1260 1270 1280 1290	.872	.868	.004	.8806	.873	.0076	.8826 .8852 .8834 .8817 .8798	.872 .873 .871 .871 .871	.0106 .0122 .0124 .0107 .0088	.8741 .8737 .8736 .8723 .8722	.868 .368 .367 .367 .867	.0061 .0057 .0066 .0053 .0052
1300 1310 1320 1330 1340	.870	.866	.004	.8791	.870 .869	.0091 .0148	.3806 .8809 .8807 .8800 .8800	.870 .870 .870 .870 .869	.0106 .0109 .0107 .0100 .0116	.8732 .3726 .8731 .8726 .8726	.866 .866 .866 .866 .866	.0072 .0066 .0071 .0066 .0066
1350 1360 1370 1380 1390	.870	.865	.005	.8831	.867	.0161	.8837 .8875 .8916 .9116 .9074	.86 <b>9</b> .869 .869 .367 .866	.0147 .0185 .0226 .0446 .0414	.8727 .8710 .8710 .8699 .8711	.865 .865 .865 .864 .864	.0077 .0060 .0060 .0059 .0091
1400 1410 1420	.866	.862	.004	.8832	.867	.0162	.8939 .8859 .8848	.366 .867 .867	.0279 .0189 .0 <b>1</b> 78	.86 <b>9</b> 1 .3702 .8697	.862 .863 .863	.0071 .0072 .0067
1430 1440							.8848 .8830 .8846	.868 .867	.0150 .0176	.8707 .8712	.863	.0087
1450 1460 1470 1480 1490	.869	.863	.006	.8833	.867	.0163	.8836 .8826 .8831 .8825 .8806	.867 .867 .866 .865 .863	.0166 .0156 .0171 .0175 .0176	.8708 .8707 .8707 .8702 .8701	.363 .863 .862 .861 .859	.0078 .0077 .0087 .0092 .0111
1500 1510 1520 1530	.864	.859	.005	.8741	.863	.0111	.8770 .8818 .8848 .8851	.863 .867 .869 .871	.0140 .0148 .0158 .0141	.8700 .8722 .8747 .8747	.859 .864 .866 .866	.0110 .0082 .0087 .0087
1540				.8830	.871	.0120	.8858	.871	.0148	.3748	.866	.0088

TABLE 1 - continued

λ		Lab 1			Lab 2		L	ab 3		L	ab 4	
28.	ρ	Cal	Diff	ρ	Cal	Diff	ρ	Cal	Diff	ρ	Cal	Diff
1550 1560 1570 1580 1580	.874	.867	.007	.8887	.872	.0167	.8876 .8874 .8873 .8884 .8876	.871 .871 .871 .872 .872	.0166 .0164 .0163 .0164 .0156	.8748 .8752 .8752 .8766 .8761	.867 .867 .867 .867 .868	.0078 .0082 .0082 .0096 .0081
1600 1610 1620 1630 1640	.873	.868	.005	.8841	.872	.0121	.8877 .8875 .8877 .8873 .8884	.872 .872 .872 .871 .871	.0157 .0155 .0157 .0163 .0164	.8766 .8766 .8761 .8753 .8753	.868 .868 .868 .367 .867	.0086 .0086 .0081 .0083 .0083
1650 1660 1670 1680 1690	.873	.867	.006	.8836	.872	.0116	.8878 .8883 .8887 .8899 .8883	.872 .872 .872 .872 .871	.0158 .0163 .0167 .0179 .0173	.8753 .8757 .8758 .8757 .8749	.867 .867 .867 .867 .867	.0083 .0087 .0088 .0087 .0087
1700 1710 1720 1730 1740	.872	.867	.005	.8840 .8821	.871	.0130	.8900 .8892 .8894 .8885 .8871	.871 .871 .871 .870 .870	.0190 .0182 .0184 .0185 .0171	.8753 .8749 .8753 .8743 .8744	.867 .867 .866 .866 .865	.0083 .0079 .0093 .0083 .0094
1750 1760 1770 1780 1790	.871	.865	.006	•882 <b>6</b>	. 870	.0126	.8869 .8856 .8862 .8866 .8861	.870 .870 .870 .870 .870	.0169 .0156 .0162 .0166 .0161	.8733 .8743 .8742 .8738 .8738	.865 .365 .865 .865 .864	.0093 .0093 .0092 .0083 .0083
1800 1810 1820 1830 1840	.870	.865	.005	. 8807	.868	.0127	.8871 .8811 .8868 .8846 .8858	.870 .869 .868 .868 .867	.0171 .0121 .0188 .0166 .0188	.8739 .8738 .8738 .8732 .8724	.865 .864 .863 .863 .863	.0099 .0095 .0103 .0102 .0104
1 <b>850</b> 1860 1870 1880 1890	.869	.861	.008	.8772	.865	.0122	.8886 .8857 .8897 .8844 .8784	.866 .865 .865 .861 .856	.0226 .0207 .0247 .0234 .0224	.8709 .8723 .8672 .8671 .8659	.861 .861 .851 .356 .352	.0099 .0113 .0062 .0111 .0139
1900 1910 1920 1930 1940	.857	.852	.005	.8682	.854 .863	.0142	.8778 .8793 .8830 .8837 .8843	.854 .856 .860 .862 .863	.0238 .0233 .0230 .0217 .0213	.8606 .8601 .8631 .8650 .8686	.852 .854 .856 .858 .858	.0055 .0061 .0071 .0070 .0096
1950 1960 1970 1980 1990	.865	.859	.006	.8757	.867	.0087	.8836 8862 .8892 .8903 .8999	.863 .865 .867 .867 .868	.0206 .0212 .0222 .0233 .0319	.8705 .3701 .8729 .8733 .8744	.859 .361 .862 .863 .364	.0115 .0091 .0119 .0103 .0104
2000 2010 2020 2030 2040	.866	.863	.003	.8799	.870	•0099	.8921 .8954 .8980 .9041 .9038	.869 .869 .870 .870 .870	.0231 .0264 .0280 .0341 .0338	.8760 .8776 .8777 .8784 .8776	.863 .864 .865 .866 .866	.0130 .0136 .0127 .0124 .0116
2050 2060 2070 2080 2090	.870	.865	.005	.8787	.869	.0097	.9043 .9049 .9079 .9131 .9088	.870 .869 .870 .870 .870	.0343 .0359 .0379 .0431 .0378	.8791 .8781 .8792 .8798 .3808	.865 .865 .365 .866 .867	.0141 .0131 .0142 .0138 .0138
2100 2110 2120 2130 2140	.874	.867	.007	.8872	.871	.0162	.9148 .9122 .9156 .9146 .9145	.871 .873 .872 .870 .869	.0438 .0392 .0436 .0446 .0455	.3804 .8811 .8827 .3798 .8775	.867 .868 .868 .866	.0134 .0131 .0147 .0138
2150 2160 2170 2180 2190	.874	.863	.011	.8769	.860	.0169	.9128	.867	.0458	.8754 .8751 .8714 .8682 .8641	.864 .863 .861 .858 .855 .855	.0135 .0124 .0141 .0134 .0132 .0121

λ		Lab 1		Le	b 2			Lab 3		Lab	4	
1.00	ρ	Cal	Diff	ρ	Cal	Diff	ø	Cal	Diff	ρ	Cal	Diff
2200 2210 2220 2230 2240	.860	.851	.009	.8637	.858	.0057				.8650 .8639 .8610 .8671 .8649	.851 .851 .852 .854 .855	.0104 .0129 .0090 .0131 .0099
2250 2260 2270 2280 2290	.862	.856	.006	.8673	.859	.0083		•		.8668 .8653 .8665 .8674 .8657	.856 .856 .858 .858 .859	.0108 .0093 .0085 .0094 .0067
2300 2310 2320 2330 2340	.867	.859	.008	<b>.</b> 8668	•860 •863	.0068				.8678 .8744 .8732 .8723 .8723	.859 .860 .861 .863 .863	.0088 .0144 .0122 .0093 .0101
2350 2360 2370 2380 2390	.869	.862	.007	.8775	.859	.0185				.8721 .8715 .8702 .8669 .8671	.862 .861 .859 .859 .858	.0101 .0115 .0112 .0079 .0091
2400 2410 2420 2430 2440	.866	.852	.014	.8642	.851	.0132				.8656 .8623 .8667 .8611 .8556	.852 .854 .851 .848 .846	.0136 .0083 .0157 .0131 .0096
2450 2460 2470 2480 2490	.858	.843	.015	.8521	.842	.0101				.8570 .8513 .8496 .8518 .8478	.843 .842 .840 .839 .838	.0140 .0093 .0096 .0128 .0098
2500				.8534	.839	.0144						
N EX Ex2	٠	4	.7 .146 .001936			86 .9995 2.07414	3		185 3.5086 .09202732			220 1339 02284661
x σ			.003106 .005677			.01162 .01036			.01895 .01177			00970 00313

#### TABLE 1 - continued

## TABLE 2 - RESULTS OF ROUND ROBIN TEST ON BLACK SAMPLES

,		Lab 1			Lab 2	•		Lab 3			Lab 4	
λ						D/ 66			D4 66			D466
1100	ρ	Cal	Diff	٥	Cal	Diff	ρ	Cal	Diff	ρ	Cal	Diff
300	.049	.057	008	.0559	.057 .057	0011 0003	.0701 .0670	.057	.0131 .0100	.0534 .0534	.057 .057	0036 0036
310 320				.0567 .0573	.059	0017	.0658	.059	.0068	.0544	.059	0046
330				.0568	.059	0022	.0660	.059	.0070	.0574	.059	0016
340				.0598	.060	0002	.0668	.06 <b>0</b>	.0068	.0573	.060	0027
350	.053	.060	007	.0596	.060	0004	.0681	.060	.0021	.0563	.060	0037
360				.0613	.062	0007	.0694	.062	.0074	.0575	.062	0045 0025
370 380				.0632 .0653	.063 .065	.0002 .0003	.0712 .0741	.063 .065	.0082 .0091	.0634	.063 .065	0016
390				.0686 -	.068	.0006	.0769	.068	.0089	.0643	.068	0037
400	.064	.070	006	.0706	.070	.0006	.0770	.070	.0070	.0643	.070	0057
410				.0709	.071	0001	.0767	.071	.0057	.0652	.071	0058
420				.0710	.070	.0010	.0759	.070 .070	.0059	.0662	.070 .070	0038 0037
430 440				.0708 .0704	.070 .069	.0008 .0014	.0749	.069	.0059	.0644	.069	0046
450 460	.063	.069	006	.0701	.069 .069	.0011 .0012	.0745 .0744	.069 .069	.0055 .0054	.0645 .0645	.069 .069	0045 0045
470				.0698	.069	.0008	.0742	.069	.0052	.0635	.069	0055
480				.0702	.069	.0012	.0736	.069	.0046	.0645	.069	0045
490				.0691	.068	.0011	.0733	.068	.0053	.0634	.068	0046
500	.062	.068	006	.0689	.068	.0009	.0727	.068	.0047	.0634	.068	0046
510				.0687	.068	.0007	.0721	.068	.0041	.0644	.068	0036
520				.0680	.067	.0010	.0718	.067	.0048	.0612	.067	0058
530 540				.0681 .0681	.067 .066	.0011 .0021	.0714 .0711	.067	.0054 .0051	.0613	.067 .066	0057 0038
	.059	.066		.0676	.066	.0016	.0708	.066	.0048			0047
550 560	.0.59	.000	007	.0672	.066	.0010	.0708	.066	.0046	.0613 .0616	.066 .066	0047
570				.0662	.065	.0012	.0692	.065	.0042	.0614	.065	0036
580				.0663	.065	.0013	.0685	.065	.0035	.0593	.065	0057
590				.0653	.064	.0013	.0683	.064	.0043	.0603	.064	0037
600	.058	.064	006	.0650	.064	.0010	.0678	.064	.0038	.0614	.064	0026
610 620				.0650 .0647	.064 .064	.0010 .0007	.0677 .0675	.064 .064	.0037 .0 <b>0</b> 35	.0625	.064	0015 0025
630				.0640	.064	.0000	.0675	.064	.0035	.0605	.064	0035
640				.0645	.063	.0015	.0674	.063	.0044	.0605	.063	0025
650	.053	.063	005	.0642	.063	.0012	.0677	.063	.0047	.0624	.063	0006
660				.0644	.063	.0014	.0678	.063	.0048	.0614	.063	0016
670				.0651	.063	.0021	.0681	.063	.0051	.0624	.063	0006
680 690				.0645 .0655	.064 .064	.0005 .0015	.0685 .0687	.064 .064	.0045 .0047	.0633 .0634	.064	0007 0006
							.0007	.004	.0047	.0004	.004	0000
700 700	.059	.065 .065	006 006	.0653	.065	.0003	.0666	.065	0016	061/	065	0036
710	.057			.0000	.005	.0005	.0683	.065	.0016 .0033	.0614 .0671	.065	0035
720							.0673	.065	.0023	.0662	.065	.0012
730 740				.0660	.065	.0010	.0677	.065	.0027	.0642	.065	0008
				.0000	.065	.0010	.0673	.065	.0023	.0643	.065	0007
750 750	.059 .059	.065	006				0701	0.45	005-			
760	-019	.065	006				.0701 .0652	.065 .065	.0051 .0002	.0623	.065	0027 0026
770							.0677	.065	.0027	.0624	.065	0026
780				-0574	.065	0076	.0722	.065	.0072	.0624	.065	0026
790							.0716	.065	.0066	.0624	.065	0026
800	.060	.065	005									
800 810	.061	.065	004				.0740 .0762	.065	.0090	.0624	.065	0026
820				.0698	.066	.0038	.0762	.065	.0112 .0081	.0622	.065 .066	0028 0017
830							.0763	.067	.0093	.0654	.067	0016
840							.0759	.067	.0089	.0665	.067	0005
850	.065	.068	003				.0756	.068	.0076	.0675	.068	0005
860 870				.0676	.069	0014	.0745	.069	.0055	.0694	.069	.0004
880							.0729	.070	.0029	.0714	.070 .072	.0014 .0013
890							.0776	.073	.0046	.0753	.073	.0023

.

#### TABLE 2 - continued

λ		Lab 1			Lab 2		I	ab 3		L	ab 4	
nm	ρ	Cal	Diff	ρ	Cal	Diff	p	Cal	Diff	ρ	Cal	Diff
900 910 920 930 940	.071	.074	003	.0781 .0934	.074	.0041	.0781 .0811 .0805 .0873 .0920	.074 .077 .079 .082 .085	.0041 .0041 .0015 .0053 .0070	.0773 .0812 .0842 .0872 .0902	.074 .077 .079 .082 .085	.0033 .0042 .0052 .0052 .0052
950 960 970 980 990	.086	.089	003	.0966	.097	0010	.0902 .0933 .0968 .0968 .1001	.089 .092 .095 .097 .099	.0012 .0013 .0018 0002 .0011	.0922 .0931 .0941 .0923 .0923	.089 .092 .095 .097 .099	.0032 .0011 0009 0047 0067
1000 1010 1020 1030 1040	.095	.100	005	.1000	.100	.0000	.0984 .1010 .1014 .0961 .0943	.100 .100 .100 .099 .098	0016 .0010 .0014 0029 0039	.0912 .0902 .0892 .0872 .0852	.100 .100 .100 .099 .098	0038 0098 0108 0118 0128
1050 1060 1070 1080 1090	.091	.097	006	.0969	.095	.0019	.0960 .0953 .0901 .0904 .0903	.097 .095 .092 .090 .087	0010 .0003 0019 .0004 .0035	.0832 .0822 .0801 .0782 .0762	.097 .095 .092 .090 .087	0138 0128 0019 0108 0108
1100 1110 1120 1130 1140	.080	.085	005	.0855	.085	.0005	.0828 .0838 .0829 .0809 .0807	.085 .083 .081 .079 .077	0022 .0008 .0019 .0019 .0037	.0743 .0733 .0713 .0683 .0683	.085 .083 .081 .079 .077	0107 0097 0097 0107 0087
1150 1160 1170 1180 1190	.071	.076	005	.0763	.073	.0033	.0732 .0776 .0744 .0733 .0727	.076 .075 .074 .073 .072	0028 .0026 .0004 .0003 .0007	.0674 .0673 .0674 .0674 .0684	.076 .075 .074 .073 .072	00°6 0077 0066 0056 0036
1200 1210 1220 1230 1240	.067	.072	005	.0777	.072	.0057	.0736 .0720 .0722 .0734 .0708	.072 .072 .072 .072 .072	.0016 .0000 .0002 .0014 0012	.0684 .0682 .0682 .0682 .0682	.072 .072 .072 .072 .072	0036 0039 0038 0039 0038
1250 1260 1270 1280 1290	.069	.072	003	.0738	.072	.0018	.0762 .0697 .0719 .0739 .0733	.072 .072 .072 .073 .073	.0042 0023 0001 .0009 .0003	.0692 .0692 .0701 .0711 .0711	.072 .072 .072 .073 .073	0028 0028 0019 0019 0019
1300 1310 1320 1330 1340	.069	.074	005	.0774	.071	.0034	.0764 .0727 .0762 .0749 .0792	.074 .075 .075 .076 .078	.0024 0023 .0012 0011 .0012	.0721 .0723 .0733 .0742 .0762	.074 .075 .075 .076 .078	0019 0027 0017 0013 0018
1350 1360 1370 1380 1390	.075	.079	004	.0901	.083	.0071	.0807 .0810 .0871 .0899 .0916	.079 .080 .081 .083 .084	.0017 .0010 .0061 .0069 .0076	.0771 .0790 .0800 .0810 .0831	.079 .080 .081 .083 .084	0019 0010 0010 0020 0009
1400 1410 1420 1430 1440	.081	.086	005	.0954	.089	.0064	.0939 .0902 .0929 .0967 .0922	.086 .087 .089 .090 .092	.0079 .0032 .0039 .0067 .0002	.0840 .0861 .0881 .0900 .0910	.086 .087 .089 .090 .092	0029 0009 0009 .0000 0010
1450 1460 1470 1480 1490	.089	.094	005	.1005	.095	.0055	.1000 .0975 .0997 .1020 .1024	.094 .095 .097 .098 .100	.0060 .0025 .0027 .0040 .0024	.0919 .0919 .0929 .0940 .0949	.094 .095 .097 .098 .100	0021 0031 0041 0040 0051
1500 1510 1520 1530 1540	.096	.101	005	.1042	.101	.0032	.1029 .1016 .1065 .1080 .1083	.101 .102 .104 .105 .106	.0019 0004 .0025 .0030 .0023	.0969 .0989 .0999 .1009 .1029	.101 .102 .104 .105 .106	0041 0031 0041 0041 0031

.

•

λ		Lab 1		L	ab 2			Lab 3		L	ab 4	
2.00	ρ	Cal	Diff	ρ	Cal	Diff	ρ	Cal	Diff	ρ	Cal	Diff
1550 1560 1570 1580 1590	.102	.107	005	.1136	.109	.0046	.1071 .1077 .1107 .1152 .1103	.107 .108 .108 .109 .109	.00 <b>01</b> 0003 .0027 .0062 .0013	.1037 .1057 .1057 .1058 .1058	.107 .108 .108 .109 .109	0033 0023 0023 0032 0032
1600 1610 1620 1630 1640	.105	.110	005	.1183	.110	. <b>0</b> 083	.1124 .1126 .1144 .1158 .1149	.110 .110 .110 .111 .111	.0024 .0026 .0044 .0048 .0039	.1058 .1059 .1058 .1057 .1057	.110 .110 .110 .111 .111	0042 0041 0042 0052 0053
1650 1660 1670 1680 1690	.106	.111	005	.1186	.111	.0076	.1161 .1188 .1194 .1156 .1171	.111 .111 .111 .111 .111	.0051 .0078 .0084 .0046 .0061	.1057 .1057 .1067 .1067 .1066	.111 .111 .111 .111 .111	0053 0053 0043 0043 0044
1700 1710 1720 1730 1740	.107	.111	004	.1136	.111	.0026 .0045	.1183 .1198 .1174 .1154 .1132	.111 .111 .111 .111 .111	.0073 .0088 .0064 .0044 .0022	.1056 .1056 .1057 .1046 .1036	.111 .111 .111 .111 .111	0054 0054 0053 0064 0074
1750 1760 1770 1780 1790	. 1 <b>0</b> 5	.110	005	.1124	.109	.0034	.1143 .1133 .1118 .1067 .1123	.110 .11 <b>0</b> .110 .109 .109	.0043 .0023 .0018 0023 .0033	.1026 .1017 .1007 .1007 .0997	.110 .110 .110 .109 .109	0074 0083 0093 0083 0093
1800 1810 1820 1830 1840	.104	.109	005	.1083	.108	. <b>0</b> 003	.1103 .1093 .1050 .1159 .1113	.109 .108 .108 .107 .107	.0013 .0013 0030 .0089 .0043	.0996 .0987 .0978 .0978 .0978	.109 .108 .108 .107 .107	0094 0093 0092 0092 0092
1850 1860 1870 1880 1890	.101	.106	005	.1096	.106	.0036	.1052 .1083 .1032 .1074 .1015	.106 .106 .105 .105 .104	0008 .0023 0018 .0024 0025	.0969 .0960 .0956 .0966 .0965	.106 .106 .105 .105 .104	0091 0100 0094 0084 0075
1900 1910 1920 1930 1940	.098	.104	006	.1065	.104	.0025	.1085 .1083 .0984 .1109 .1085	.104 .104 .103 .103 .102	.0045 .0043 0046 .0079 .0065	.0963 .0961 .0953 .0943 .0933	.104 .104 .103 .103 .102	0077 0079 0077 0087 +.0087
1950 19 <b>60</b> 1970 1980 1990	.095	. 102	007	.1020	.101	.0010	.1041 .1076 .1083 .1077 .1017	.102 .102 .102 .101 .101	.0021 .0056 .0063 .0067 .0007	.0943 .0943 .0952 .0951 .0951	.102 .102 .102 .101 .101	+.0077 0077 0068 0059 0059
2000 2010 202 <b>0</b> 2030 2040	.094	.101	007	.1045	.100	.0045	.1045 .1032 .1087 .1058 .0966	.101 .1 <b>0</b> 1 .100 .100 .099	.0035 .0022 .0087 .0058 0024	.0930 .0929 .0918 .0926 .0935	.101 .101 .100 .100 .099	0080 0081 0082 0074 0055
2050 2060 2070 2080 2090	.093	.099	006	.1060	.099	.007 <b>0</b>	.1112 .0857 .1024 .0977 .1076	.099 .099 .099 .099 .099	.0122 0133 .0034 0013 .0096	.0934 .0933 .0932 .0932 .0932	.099 .099 .099 .099 .099	0056 0057 0058 0058 0048
2100 2110 2120 2130 2140	.092	.098	006	.1011	.098	.0 <b>0</b> 31 0004	.1012 .0960 .0988 .1051 .1000	.098 .098 .098 .097 .097	.0032 0020 .0008 .0081 .0030	.0922 .0920 .0920 .0908 .0907	.098 .098 .098 .097 .097	0058 0060 0060 0062 0063
2150 2160 2170 2180	.091	.097	006	.0991	.095	.0041	.1350	.097	.0380	.0890 .0894 .0895 .0897	.097 .096 .096 .095	0080 0066 0065 0053
2190										.0898	.095	0052

•

TABLE 2 - continued

`λ		Lab 1		L	ab 2			Lab 3			Lab 4	
<b>DM</b>	p	Cal	Diff	p	Cal	Diff	p	Cal	Diff	ρ	Cal	Diff
2200	.090	.096	006							.0889	.096	0071
2230										.0892	.096	0068
2220				.0947	.096	0013				.0894	.096	0066
2230										.0894	.096	0066
2240										.0895	.096	0065
2250	.089	.096	007							.0895	.096	0065
2260				.0970	.096	.0010				.0895	.096	0065
2270										.0904	.096	0056
2280										.0901	.096	0059
2290				-						.0908	.096	0052
2300	.089	.096	007	.1029	.096	.0069				.0908	.096	0052
2310										.0908	.096	0052
2320										.0889	.096	0071
2330										.0888	.096	0072
2340				.0902	.096	0040				.0885	.096	0075
2350	.088	.096	008							.0892	.096	0068
2360										.0902	.096	0058
2370										.0910	.096	0050
2380				.0961	.096	.0001				.0908	.096	0052
2390										.0918	.096	0042
2400	.088	.096	008							.0907	.096	0053
2410										.0896	.096	0064
2420				.1079	.096	.0119				.0881	.096	0079
2430										.0880	.096	0080
2440										.0879	.096	0081
2450	.088	.096	008							.0889	.096	0071
2460				.0979	.095	.0029				.0887	.095	0063
2470										.0888	.095	0062
2480										.0878	.095	0072
2490										.0898	.095	0052
2500				.0923	.0 <b>9</b> 5	0027						
ΞX		262				.1529			· .6641			-1.0458
ZX2		.001	538			.0095455			.005838	57		.00744336
N		47	•			86			186			220
X		005				.00178			.00357			00475
a		001	30			.00283			.00433			.00366

#### TABLE 3 - REPORTED VALUES FOR SOLAR PROPERTIES

Property	Lab l <sup>1</sup>	Lab 2 <sup>2</sup>	Lab 3 <sup>3</sup>	Lab 4	NBS
Reflectance	%	%	%	%	
of White Sample	82.5	87.1	84.3	81.06 <sup>4</sup> 83.665 83.777 80.40	81.94 <sup>8</sup> 83.33 <sup>9</sup>
Absorptance of Black Sample	93.1	92.9	92.5	93.04 <sup>4</sup> 92.98 <sup>5</sup> 92.07	92.44

1. Beckman 5270 Spectroreflectometer - Halon powder

2. Beckman 5270 double-beam grating-filter instrument with high intensity source

3. Varian Cary 17D integrating sphere reflectometer

4. Perkin-Elmer 340 - Halon standards

5. Beckman 5420 - Halon standards

6. Beckman 5420 - BaSO<sub>4</sub> standard

7. Cary 17 - Halon standard

8. For Comparison to Laboratories 1 and 4

9. For Comparison to Laboratories 2 and 3

TABLE 4 - PROCEDURES USED FOR COMPUTING SOLAR PROPERTIES

. . . . .

Laboratory 1 - 100 Selected Ordinate Method, based on Thekaekara AM 1.5 solar spectral irradiance.

Laboratory 2 - Weighted Ordinate Method, based on NASA AM 1.5 TDSSID.

- Laboratory 3 ASTM E-424, Method A, Weighted Ordinate Method based on Parry Moon data.
- Laboratory 4 Weighted Ordinate Method, based on proposed revision of ASTM E-424.

## NBS - 100 Selected Ordinate Method, based on ASTM E-44 direct solar irradiance.

TABLE 5 - DIFFERENCES AS FRACTION OF REPORTED SOLAR VALUES

••

•

	Lab l	Lab 2	Lab 3	Lab 4
		WHITE SAMPLES - SOLAN	R REFLECTANCE	
Diff	0.0038	0.0133	0.0225	0.0118
σ	0.0069	0.0119	0.0140	0.0038
		BLACK SAMPLES - SOL	AR ABSORPTANCE	
Diff	-0.0033	0.0019	0.0039	0.0057
đ	0,0060	0.0030	0.0047	0.0039

• •




Figure 1 - Difference between the data reported by Laboratory 1 and the calibration data for the white sample, plotted as a function of wavelength.




Figure 2 - Difference between data reported by Laboratory 2 and calibration data for the white sample, plotted as a function of wavelength.




Figure 3 - Difference between data reported by Laboratory 3 and calibration data for the white sample, plotted as a function of wavelength.




Figure 4 - Difference between data reported by Laboratory 4 and calibration data for the black sample, plotted as a function of wavelength.




Figure 5 - Difference between the data reported by Laboratory 1 and the calibration data for the black sample, plotted as a function of wavelength.




Figure 6 - Difference between the data reported by Laboratory 2 and calibration data for the black sample, plotted as a function of wavelength.

.




Figure 7 - Difference between the data reported by Laboratory 3 and the calibration data for the black sample, plotted as a function of wavelength.




Figure 8 - Difference between data reported by Laboratory 4 and the calibration data for the balck sample, plotted as a function of wavelength.



NBS-114A (REV. 2-80)			
U.S. DEPT. OF COMM.	1. PUBLICATION OR	2. Performing Organ. Report No. 3.	Publication Date
BIBLIOGRAPHIC DATA	REPORT NO.		7 1 1001
SHEET (See instructions)	NBSIR 81-2311		July 1981
4. TITLE AND SUBTITLE			
Analysis of Resul	lts of Mini Round Robi	n Reflectance Test	
5. AUTHOR(S)			
Joseph C. Richmon	nd		
Joseph C. Richmon			
6. PERFORMING ORGANIZA	TION (If joint or other than NBS	, see instructions) 7.	Contract/Grant No.
NATIONAL BUREAU OF		2	Type of Report & Period Covered
WASHINGTON, D.C. 2023		·-	Type of Report a Ferror Covered
9 SPONSORING ORGANIZAT	TION NAME AND COMPLETE A	DDRESS (Street, City, State, ZIP)	
and the second design of the second sec			
Solar Energy Res	earch Institute, Golde	21, CO 80401	
10. SUPPLEMENTARY NOTE	S		
Document describes a	a computer program; SF-185, FIP	S Software Summary, is attached.	
		significant information. If document	includes a significant
bibliography or literature	survey, mention it here)		
A mini round rob	in test was conducted	in which four national 1	aboratories partici-
pated. The spect	ral directional-hemis	oherical reflectance for	near-normal incidence
		te, was measured in the w	
		e of the white sample and	-
		. Each laboratory used a	
		sured samples had previou	
	-		-
		f the National Bureau of	
		s. The average difference	
		ied from +0.0116 +0.0634	
the white sample	s and from -0.00475 +	0.00366 to +0.00357 <u>+</u> 0.00	)433 for the black
samples. The di	fferences for the whi	te samples were ascribed	primarily to errors
		the reference standard us	
		were ascribed primarily	
12. KEY WORDS (Six to twelv	e entries; alphabetical order; ca	pitalize only proper names; and sepa	rate key words by semicolons)
diffuse reflecta	nce; directional-hemi	spherical reflectance; in	ntegrating spheres;
laboratory inter	comparisons; reflecta	nce measurements; reflect	ance standards; solar
properties of ma	terials; spectral ref	lectance measurements.	
13. AVAILABILITY			14. NO. OF
			PRINTED PAGES
🔄 Unlimited			
For Official Distribut	ion. Do Not Release to NTIS		
Order From Superinter	adent of Documents, U.S. Govern	ment Printing Office, Washington, D.	C
20402.			15. Price
Order From National	Technical Information Service (N	TIS), Springfield VA. 22161	



- - - -

# .