

NIST TIME AND FREQUENCY BULLETIN  
NISTIR 5057-3

NO. 472      MARCH 1997

|  |   |
|--|---|
| 1. GENERAL BACKGROUND INFORMATION .....                                      | 2 |
| 2. TIME SCALE INFORMATION .....  | 2 |
| 3. UT1 CORRECTIONS AND LEAP SECOND ADJUSTMENTS .....                         | 2 |
| 4. PHASE DEVIATIONS FOR WWVB AND LORAN-C .....                               | 3 |
| 5. GOES TIME CODE INFORMATION .....  | 4 |
| 6. BROADCAST OUTAGES OVER FIVE MINUTES AND WWVB<br>PHASE PERTURBATIONS ..... | 4 |
| 7. NOTES ON NIST TIME SCALES AND PRIMARY STANDARDS .....                     | 4 |
| 8. BIBLIOGRAPHY .....  | 4 |
| 9. SPECIAL ANNOUNCEMENTS .....   | 6 |

This bulletin is published monthly. Address correspondence to:

Gwen E. Bennett, Editor  
Time and Frequency Division  
National Institute of Standards and Technology  
325 Broadway  
Boulder, CO 80303-3328  
(303) 497-3295

NOTE TO SUBSCRIBERS: Please include your address label (or a copy)  
with any correspondence regarding this bulletin.



---

U.S. DEPARTMENT OF COMMERCE, William M. Daley, Secretary  
TECHNOLOGY ADMINISTRATION, Mary L. Good, Under Secretary for Technology  
NATIONAL INSTITUTE OF STANDARDS AND TECHNOLOGY, Arati Prabhakar, Director

## 1. GENERAL BACKGROUND INFORMATION

### ABBREVIATIONS AND ACRONYMS USED IN THIS BULLETIN

|       |   |         |               |
|-------|---|---------|---------------|
| BIPM  | - Bureau International des Poids et Mesures           |         |               |
| CCIR  | - International Radio Consultative Committee          |         |               |
| Cs    | - Cesium standard                                     |         |               |
| GOES  | - Geostationary Operational Environmental Satellite   |         |               |
| GPS   | - Global Positioning System                           |         |               |
| IERS  | - International Earth Rotation Service                |         |               |
| LORAN | - Long Range Navigation                               |         |               |
| MC    | - Master Clock  |         |               |
| MJD   | - Modified Julian Date                                |         |               |
| NVLAP | - National Voluntary Laboratory Accreditation Program |         |               |
| NIST  | - National Institute of Standards & Technology        |         |               |
| NOAA  | - National Oceanic and Atmospheric Administration     | ns      | - nanosecond  |
| SI    | - International System of Units                       | $\mu$ s | - microsecond |
| TA    | - Atomic Time   | ms      | - millisecond |
| TAI   | - International Atomic Time                           | s       | - second      |
| USNO  | - United States Naval Observatory                     | min     | - minute      |
| UTC   | - Coordinated Universal Time                          | h       | - hour        |
| VLF   | - very low frequency                                  | d       | - day         |

| FEB 1997 | MJD   | UT1 – UTC(NIST)<br>( $\pm 5$ ms) | UTC – UTC(NIST)<br>( $\pm 0.1 \mu$ s) | UTC(USNO,MC) – UTC(NIST)<br>( $\pm 20$ ns) |
|----------|-------|----------------------------------|---------------------------------------|--|
| 13       | 50492 | - 187 ms                         | 0.0 $\mu$ s                           | - 15 ns                                    |
| 20       | 50499 | - 198 ms                         | 0.0 $\mu$ s                           | - 17 ns                                    |
| 27       | 50506 | - 211 ms                         | 0.0 $\mu$ s                           | - 17 ns                                    |

### 3. UT1 CORRECTIONS AND LEAP SECOND ADJUSTMENTS

The master clock pulses used by the WWV, WWVH, WWVB, and GOES time code transmissions are referenced to the UTC(NIST) time scale. Occasionally, 1 s is added to the UTC time scale. This second is called a leap second. Its purpose is to keep the UTC time scale within  $\pm 0.9$  s of the UT1 astronomical time scale, which changes slightly due to variations in the rotation of the Earth.

Positive leap seconds, beginning at 23 h 59 min 60 s UTC and ending at 0 h 0 min 0 s UTC, were inserted in the UTC timescale on 30 June 1972, 31 December 1972-1979, 30 June 1981-1983, 30 June 1985, 31 December 1987, 1989, 1990, and 1995, 30 June 1992, 1993, and 1994. **There will be a leap second inserted 30 June 1997.**

The use of leap seconds ensures that UT1 – UTC will always be held within  $\pm 0.9$  s. The current value of UT1 – UTC is called the DUT1 correction. DUT1 corrections are broadcast by WWV, WWVH, WWVB, and GOES and are printed below. These corrections may be added to received UTC time signals in order to obtain UT1.

|                    |   |
|--------------------|---|
| DUT1 = UT1 – UTC = | -0.1 s beginning 0000 UTC 05 December 1996<br>-0.2 s beginning 0000 UTC 06 February 1997<br>-0.3 s beginning 0000 UTC 20 March 1997 |
|--------------------|---|

#### 4. PHASE DEVIATIONS FOR WWVB AND LORAN-C

WWVB - The values shown for WWVB are the time difference between the time markers of the UTC(NIST) time scale and the first positive-going zero voltage crossover measured at the transmitting antenna. The uncertainty of the individual measurements is  $\pm 0.5 \mu\text{s}$ . The values listed are for 1300 UTC.

LORAN-C - The values shown for Loran-C represent the daily accumulated phase shift (in nanoseconds). The phase shift is measured by comparing the output of a Loran receiver to the UTC(NIST) time scale for a period of 24 h. If data were not recorded on a particular day, the symbol (-) is printed.

The master stations monitored are Dana, IN (8970) and Fallon, NV (9940). The monitoring is done from the NIST laboratories in Boulder, CO.

**Note: The values shown for Loran-C are in nanoseconds.**

| DATE    | MJD   | UTC(NIST)-WWVB(60 kHz)             | UTC(NIST) - LORAN PHASE (ns) |                             |
|---------|-------|------------------------------------|------------------------------|-----------------------------|
|         |       | ANTENNA PHASE<br>( $\mu\text{s}$ ) | LORAN-C (DANA)<br>(8970)     | LORAN-C (FALLON)*<br>(9940) |
| 2/01/97 | 50480 | 5.74                               | +310                         | +214                        |
| 2/02/97 | 50481 | 5.75                               | -342                         | -279                        |
| 2/03/97 | 50482 | 5.78                               | (-)                          | (-)                         |
| 2/04/97 | 50483 | 5.74                               | -14                          | +255                        |
| 2/05/97 | 50484 | 5.72                               | +815                         | -186                        |
| 2/06/97 | 50485 | 5.75                               | -680                         | -206                        |
| 2/07/97 | 50486 | 5.78                               | +186                         | +86                         |
| 2/08/97 | 50487 | 5.75                               | +22                          | +0                          |
| 2/09/97 | 50488 | 5.74                               | -41                          | +128                        |
| 2/10/97 | 50489 | 5.74                               | -479                         | +141                        |
| 2/11/97 | 50490 | 5.79                               | (-)                          | +172                        |
| 2/12/97 | 50491 | 5.78                               | -99                          | -328                        |
| 2/13/97 | 50492 | 5.79                               | -15                          | +319                        |
| 2/14/97 | 50493 | 5.79                               | (-)                          | (-)                         |
| 2/15/97 | 50494 | 5.78                               | -289                         | +63                         |
| 2/16/97 | 50495 | 5.78                               | +234                         | +346                        |
| 2/17/97 | 50496 | 5.78                               | -33                          | -235                        |
| 2/18/97 | 50497 | 5.78                               | -77                          | -425                        |
| 2/19/97 | 50498 | 5.77                               | +506                         | -207                        |
| 2/20/97 | 50499 | 5.76                               | (-)                          | (-)                         |
| 2/21/97 | 50500 | 5.78                               | (-)                          | (-)                         |
| 2/22/97 | 50501 | 5.75                               | (-)                          | +348                        |
| 2/23/97 | 50502 | 5.72                               | (-)                          | -3                          |
| 2/24/97 | 50503 | 5.71                               | (-)                          | -326                        |
| 2/25/97 | 50504 | 5.66                               | +94                          | -377                        |
| 2/26/97 | 50505 | 5.72                               | -68                          | -13                         |
| 2/27/97 | 50506 | 5.70                               | +531                         | +238                        |
| 2/28/97 | 50507 | 5.68                               | -339                         | -591                        |

## 5. GOES TIME CODE INFORMATION

### A. TIME CODE PERFORMANCE (1-28 February 1997)

#### GOES/East:

Currently using the GOES-8 satellite at 75° west longitude. Timing uncertainty is  $\pm 100 \mu\text{s}$  with respect to UTC(NIST).

The spring eclipse season for GOES/East began on February 26th and will continue until April 13, 1997. The peak eclipse hours are from 0434 to 0546 UTC. The GOES/East signal may be intermittent or erratic during the peak eclipse hours.

A GOES/East stationkeeping maneuver was performed on February 25th, 1997 at approximately 1530 UTC.

#### GOES/West:

Currently using the GOES-9 satellite at 135° west longitude. Timing uncertainty is  $\pm 100 \mu\text{s}$  with respect to UTC(NIST).

The spring eclipse season for GOES/West began on February 25th and will continue until April 12, 1997. The peak eclipse hours are from 0832 to 0944 UTC. The GOES/West signal may be intermittent or erratic during the peak eclipse hours.

## 6. BROADCAST OUTAGES OVER 5 MINUTES AND WWVB PHASE PERTURBATIONS

| OUTAGES |             |     |              |              | PHASE PERTURBATIONS WWVB 60 KHz |             |     |              |            |
|---------|-------------|-----|--------------|--------------|---------------------------------|-------------|-----|--------------|------------|
| Station | FEB<br>1997 | MJD | Began<br>UTC | Ended<br>UTC | Freq.                           | FEB<br>1997 | MJD | Began<br>UTC | End<br>UTC |
| WWVB    |             |     |              |              |                                 |             |     |              |            |
| WWV     |             |     |              |              |                                 |             |     |              |            |
| WWVH    |             |     |              |              |                                 |             |     |              |            |

## 7. NOTES ON NIST TIME SCALES AND PRIMARY STANDARDS

Primary frequency standards developed and maintained by NIST are used to provide accuracy (rate) input to the BIPM. NBS-6, which served as the U.S. primary standard from 1975 through 1992, has been replaced by NIST-7, an optically pumped cesium-beam standard. The uncertainty of the new standard is currently 1 part in  $10^{14}$ .

Since 1981, TA(NIST) has been computed retrospectively each month using a Kalman algorithm. The purpose of TA(NIST) was to provide a flywheel that realized our best estimate of the SI second between calibrations of our primary frequency standard, but the algorithm we have been using is not optimum for this purpose and is particularly unsuited to our new higher-accuracy environment. We therefore stopped computing TA(NIST) on 31 October 1993. We are studying alternate methods for incorporating the rate accuracy of NIST-7 into our time-scale algorithm, but no changes are likely until a thorough evaluation of the new procedure has been completed.

The AT1 scale is run in real time using data from an ensemble of cesium standards and hydrogen masers. It is a free-running scale whose frequency is maintained as constant as possible by choosing the optimum weight for each clock that contributes to the computation.

UTC(NIST) is generated as an offset from our real-time scale AT1. It is steered in frequency towards UTC using data published by the BIPM in its Circular T. Changes in the steering frequency will be made only at 0000 UTC on the first day of any month, and the change in frequency in any month is limited to  $\pm 2 \text{ ns/day}$ . The frequency of UTC(NIST) is kept as stable as possible at other times.

UTC is generated at the BIPM using a post-processed time-scale algorithm and is not available in real-time. The parameters that we use to generate UTC(NIST) in real-time are therefore based on an extrapolation of UTC from the most recent data available.

## 8. BIBLIOGRAPHY

- Allan, D.W.; Hellwig, H.; and Glaze, D.J., "An accuracy algorithm for an atomic time scale," *Metrologia*, Vol.11, No.3, pp.133-138 (September 1975).
- Allan, D.W. and Weiss, M.A., "Accurate time and frequency transfer during common view of a GPS satellite," *Proc. 34th Annual Symposium on Frequency Control*, p.334 (1980).
- Allan, D.W. and Barnes, J.A., "Optimal time and frequency using GPS signals," *Proc. 36th Annual Symposium on Frequency Control*, p.378 (1982).
- Drullinger, R.E.; Glaze, D.J.; Lowe, J.P.; and Shirley, J.H., "The NIST optically pumped cesium frequency standard," *IEEE Trans. Instrum. Meas.*, IM-40, 162-164 (1991).
- Glaze, D.J.; Hellwig, H.; Allan, D.W.; and Jarvis, S., "NBS-4 and NBS-6: The NIST primary frequency standards," *Metrologia*, Vol.13, pp.17-28 (1977).
- Wineland, D.J.; Allan, D.W.; Glaze, D.J.; Hellwig, H.; and Jarvis, S., "Results on limitations in primary cesium standard operation," *IEEE Trans. Instrum. Meas.*, IM-25, pp.453-458 (December 1976).

Table 7.1 is a list of the parameters that are used to define UTC(NIST) with respect to our real-time scale AT1. To find the value of UTC(NIST) – AT1 at any time T (expressed as a Modified Julian Day, including a fraction if needed), the appropriate equation to use is the one for which the desired T is greater than or equal to the entry in the  $T_0$  column and less than the entry in the last column. The values of  $x_{ls}$ ,  $x$ , and  $y$  for that month are then used in the equation below to find the desired value. The parameters  $x$  and  $y$  represent the offset in time and in frequency, respectively, between UTC(NIST) and AT1; the parameter  $x_{ls}$  is the number of leap seconds applied to both UTC(NIST) and UTC as specified by the IERS. Leap seconds are not applied to AT1.

| Table 7.1<br>UTC(NIST) – AT1 = $x_{ls} + x + y \cdot (T - T_0)$ |                 |             |                 |                |                               |
|---|-----------------|-------------|-----------------|----------------|-------------------------------|
| Month   | $x_{ls}$<br>(s) | $x$<br>(ns) | $y$<br>(ns/day) | $T_0$<br>(MJD) | Valid until 0000 on:<br>(MJD) |
| May 95  | -29             | -128075     | -43.0           | 49838          | 49869                         |
| Jun 95  | -29             | -129408     | -43.0           | 49869          | 49899                         |
| Jul 95  | -29             | -130698     | -43.0           | 49899          | 49930                         |
| Aug 95  | -29             | -132031     | -43.0           | 49930          | 49961                         |
| Sep 95  | -29             | -133364     | -43.0           | 49961          | 49991                         |
| Oct 95  | -29             | -134654     | -43.5           | 49991          | 50022                         |
| Nov 95  | -29             | -136003     | -44.0           | 50022          | 50052                         |
| Dec 95  | -29             | -137323     | -44.0           | 50052          | 50083                         |
| Jan 96  | -30             | -138687     | -43.5           | 50083          | 50114                         |
| Feb 96  | -30             | -140035     | -43.5           | 50114          | 50143                         |
| Mar 96  | -30             | -141297     | -43.5           | 50143          | 50174                         |
| Apr 96  | -30             | -142645     | -43.5           | 50174          | 50204                         |
| May 96  | -30             | -143950     | -43.5           | 50204          | 50235                         |
| Jun 96  | -30             | -145299     | -43.5           | 50235          | 50265                         |
| Jul 96  | -30             | -146604     | -44.0           | 50265          | 50296                         |
| Aug 96  | -30             | -147968     | -44.5           | 50296          | 50327                         |
| Sep 96  | -30             | -149347     | -44.5           | 50327          | 50357                         |
| Oct 96  | -30             | -150682     | -44.0           | 50357          | 50388                         |
| Nov 96  | -30             | -152046     | -44.0           | 50388          | 50418                         |
| Dec 96 <sup>†</sup>   | -30             | -153366     | -43.8           | 50418          | 50434                         |
|   | -30             | -154066.8   | -42.6           | 50434          | 50449                         |
| Jan 97  | -30             | -154705.8   | -42.5           | 50449          | 50480                         |
| Feb 97  | -30             | -156023.3   | -42.5           | 50480          | 50508                         |
| Mar 97  | -30             | -157213.3   | -42.7           | 50508          | 50539                         |
| Apr 97  | -30             | -158537     | -43.0*          | 50539          | 50569                         |

\*Provisional rate

<sup>†</sup>Note rate change in mid-month

## 9. SPECIAL ANNOUNCEMENTS

### TRACEABLE FREQUENCY CALIBRATIONS (Now NVLAP Certified)

Anyone needing traceable frequency calibrations can get them by subscribing to the NIST Frequency Measurement and Analysis Service. This service is offered on a lease basis by NIST to provide an easy and inexpensive means to obtain traceability of a laboratory main oscillator and, in addition, to calibrate other devices in the lab. This service has been designed for ease of operation and as a practical lab calibration tool.

All the equipment and software needed are provided by NIST. Users must provide their own oscillator(s) and an ordinary telephone line so that NIST can access the system by modem. A total of four oscillators can be calibrated at the same time. Radio signals from either Loran-C or GPS satellite are used. Results for either are at about the same accuracy.

The calibration data are displayed in color and a graph is plotted daily for each oscillator connected. Data are also stored on disk. The user can call up any of the data and view them onscreen or in the form of plots. Many months of data can be plotted.

The system plots are easy to read and understand. The system manual is written for easy understanding and the NIST staff is available by telephone to assist. The modem connection allows NIST to access the data and to prepare a monthly traceability report which is mailed to the user.

Frequency sources of any accuracy can be calibrated. The FMAS is particularly useful at the highest levels of performance. This is because each user of the system contributes information and calibration data for the others. If an uncertainty arises, it is possible for NIST to call by modem to another user nearby. In this way problems in data interpretation can be resolved.

NVLAP certification requirements for frequency measurement are met by following the NIST-FMAS operating manual. This service does not eliminate the NVLAP audits but, when installed and operated per the NIST guidelines, audit requirements are easily met.

NIST retains title to the equipment and supplies any needed system spares. Equipment that fails is replaced by overnight shipment. Training for use of the system is available if requested by the user.

The NIST Frequency Measurement and Analysis Service provides a complete solution to nearly all frequency measurement and calibration problems. For a free information package, please call Michael Lombardi at (303) 497-3212, or write to: Michael Lombardi, NIST, Division 847, 325 Broadway, Boulder, CO 80303.

---