


NIST  
PUBLICATIONS

**NISTIR 5432**

# **Proficiency Tests for the NIST Airborne Asbestos Program - 1991**

**Shirley Turner  
Eric B. Steel  
Stacy S. Doom\*  
Steve B. Burris\***



chrysotile diffraction pattern

U.S. DEPARTMENT OF COMMERCE  
Technology Administration  
National Institute of Standards  
and Technology  
Microanalysis Research Group  
Surface and Microanalysis Science Division  
Chemical Science & Technology Laboratory  
Gaithersburg, MD 20899

\*Research Triangle Institute  
Center for Environmental Measurements and  
Quality Assurance  
Research Triangle Park, NC 27709-2194

May 1994

~~QC~~  
100  
.U56  
NO. 5432  
1994

**NIST**



# **Proficiency Tests for the NIST Airborne Asbestos Program - 1991**

**Shirley Turner  
Eric B. Steel  
Stacy S. Doom\*  
Steve B. Burris\***

U.S. DEPARTMENT OF COMMERCE  
Technology Administration  
National Institute of Standards  
and Technology  
Microanalysis Research Group  
Surface and Microanalysis Science Division  
Chemical Science & Technology Laboratory  
Gaithersburg, MD 20899

\*Research Triangle Institute  
Center for Environmental Measurements and  
Quality Assurance  
Research Triangle Park, NC 27709-2194



**U.S. DEPARTMENT OF COMMERCE  
Ronald H. Brown, Secretary**

**TECHNOLOGY ADMINISTRATION  
Mary L. Good, Under Secretary for Technology**

**NATIONAL INSTITUTE OF STANDARDS  
AND TECHNOLOGY  
Arati Prabhakar, Director**



## Preface

The National Voluntary Accreditation Program (NVLAP) at the National Institute of Standards and Technology (NIST) has since 1990 had a program to accredit those laboratories involved in the analysis of airborne asbestos by transmission electron microscopy. As a part of that program, laboratories are sent proficiency tests twice yearly to evaluate their ability to correctly analyze samples and to test the general knowledge of laboratory personnel. The results of the tests are sent to the participating laboratories in the form of a summary report. This NIST Internal Report (NISTIR) contains the instructions and summary reports issued for the proficiency tests in 1991 (PT91-1, PT91-2). This NISTIR is one of a series covering the years of proficiency testing in the airborne asbestos accreditation program. The NISTIRs provide a historical record of materials sent to the laboratories for proficiency testing so that they can be referenced in other publications and so that background material can be given to those laboratories entering the accreditation program. The materials can also be used as educational aids. The material in the IRs are copies of the instructions and summary reports sent to the laboratories - if comments are warranted they are given on the chapter title page for the instructions or summary report.



## Acknowledgements

We thank Diane J. Hues of NIST for her assistance in PT91-2.





## Table of Contents

Chapter	Page
I. Instructions for PT91-1 . . . . .	1
II. Summary Report for PT91-1 . . . . .	23
III. Instructions for PT91-2 . . . . .	35
IV. Summary Report for PT91-2 . . . . .	41



## I. Instructions for PT91-1

## Proficiency test 91-1 Part 1 - EDS analysis

### Instructions for evaluation of EDS spectra

Enclosed are EDS spectra labelled Figures 1-4 with associated tables labelled Tables 1-5. Descriptions of the spectra and tables are given below:

- Figure 1a - spectrum plotted on a linear scale with peaks labelled 1 - 14
- Figure 1b - spectrum of Figure 1a plotted on a log scale with peaks labelled 1 - 14
- Table 1 - listing of energy, peak counts and background counts for the peaks in Fig. 1a,b
  
- Figure 2 - spectrum showing a Mn  $K\alpha$  peak
- Table 2 - listing of the total counts and energy for each channel in Fig. 2
  
- Figure 3 - spectrum from Standard Reference Material (SRM) 2063
- Table 3 - listing of the chemical composition of SRM 2063
- Table 4 - listing of the energy, peak counts and background counts for the peaks in Fig. 3
  
- Figure 4a - spectrum from SRM 2063 (linear scale)
- Figure 4b - spectrum of Figure 4a plotted on a log scale
- Table 5 - listing of the energy, peak counts and background counts for the peaks in Fig. 4a,b

Carefully follow the instructions below for deriving and recording values from each spectrum.

#### Figure 1

For this spectrum, record on Form 3 the source of the x-ray peaks labelled from 1 - 14. Note: there are no peaks in the spectrum above 10 keV. Include the element and peak type (K, L...,  $\alpha$ ,  $\beta$ , etc.) - for example, Zn  $K\alpha$  or Y  $L\beta$ . If there is more than one possibility that is supported by the spectrum, list them in order of most likely to least likely.

#### Figure 2

From this spectrum and the information given in Table 2, determine the resolution (FWHM) of the detector used to collect the spectrum. Record any work done on Form 1. Record the value for the resolution on Form 3 (to the nearest eV).

#### Figure 3

From this spectrum of SRM 2063 and the information given in Tables 3 and 4, calculate the k-values for each element (relative to Si). Record the method used and any work done on Form 2. Record the k-values on Form 3 (to two decimal places).

#### Figure 4

This spectrum was also collected from SRM 2063. Answer the questions on Form 3 concerning this spectrum.




Figure 1a. Spectrum plotted on a linear scale with peaks labelled 1-14.




Figure 1b. Spectrum plotted on a log scale with peaks labelled 1-14.

Table 1. Listing of the energy, peak counts (background subtracted) and the background counts for the peaks in Figure 1

Peak	Energy	Peak Counts*	Background Counts
1	1.04	31524	46916
2	1.25	48190	53082
3	1.49	150093	68297
4	1.74	946893	64349
5	3.31	64010	40436
6	3.69	107404	44180
7	4.01	11060	30167
8	4.51	31161	37135
9	4.93	4499	26635
10	5.89	15780	33682
11	6.40	87816	40171
12	7.06	12304	32947
13	8.04	263856	43055
14	8.91	38523	36763

\*Note: Peak counts are the number of counts in the peak integral with the background counts subtracted out

# Manganese K-alpha Peak for X-ray Spectrometer Calibration




Figure 2. Spectrum showing a Mn K-alpha peak.



Table 2. Listing of the total counts and energy for each channel in Figure 2.

CHANNEL	COUNTS	ENERGY (keV)
547	905	5.639
548	911	5.649
549	1011	5.659
550	1040	5.669
551	1174	5.679
552	1299	5.689
553	1564	5.699
554	1824	5.709
555	2352	5.719
556	2959	5.729
557	4106	5.739
558	5635	5.749
559	7367	5.759
560	10040	5.769
561	12911	5.779
562	17030	5.789
563	21366	5.799
564	26742	5.809
565	32099	5.819
566	38060	5.829
567	43673	5.839
568	49027	5.849
569	53403	5.859
570	56663	5.869
571	58883	5.879
572	59780	5.889
573	58292	5.899
574	55822	5.909
575	51858	5.919
576	46989	5.929
577	41598	5.939
578	35760	5.949
579	30101	5.959
580	24424	5.969
581	19290	5.979
582	15422	5.989
583	11501	5.999
584	8615	6.009
585	6288	6.019
586	4638	6.029
587	3314	6.039
588	2455	6.049
589	1815	6.059
590	1310	6.069
591	1003	6.079
592	895	6.089
593	741	6.099
594	691	6.109
595	676	6.119
596	623	6.129
597	634	6.139
598	624	6.149
599	613	6.159
600	634	6.169

# K-value Test X-ray Spectrum




Figure 3. Spectrum from Standard Reference Material (SRM) 2063

Table 3. Listing of the chemical composition of SRM 2063

Element	Weight % oxide	Weight % element	Atom % element
O	----	42.4	60.3
Mg	14.7	8.8	8.2
Si	54.3	25.4	20.3
Ar	1.0	1.0	0.7
Ca	15.5	11.1	6.2
Fe	14.4	11.2	4.4
Total	99.9	99.9	100.1

Table 4. Listing of the energy, peak counts (background subtracted) and the background counts for the peaks in Figure 3

Peak	Energy	Peak Counts*	Background Counts
1	1.25	213242	87854
2	1.74	980256	76855
3	2.96	24078	57696
4	3.69	434036	55295
5	4.01	57779	44936
6	6.40	341445	56117
7	7.06	47234	39791
8	8.04	200347	52307
9	8.91	31255	45400

\*Note: Peak counts are the number of counts in the peak integral after the background counts have been removed




Figure 4a. Spectrum from SRM 2063 (linear scale).




Figure 4b. Spectrum from SRM 2063 (log scale).

Table 5. Listing of the energy, peak counts (background subtracted) and the background counts for the peaks in Figure 4.

Peak	Energy	Peak Counts*	Background Counts
1	3.69	18770	2399
2	4.01	3873	2493
3	6.40	85190	8306
4	7.06	14265	6559
5	8.04	53341	7332
6	8.91	9806	6542

\*Note: Peak counts are the number of counts in the peak integral after the background counts have been removed

Form 1

1. Briefly outline the method used to determine the resolution for the peak in Figure 2.

2. Show work involved in calculation of resolution (attach another page if necessary).



## Form 2

1. Describe briefly the method used for calculating the k-values for the peaks in Figure 3.

2. Show work done in calculation of k-values (attach another page if necessary).

Form 3

Figure 1

Peak number	Peak identification Choice #1	Peak identification Choice #2	Peak identification Choice #3
1			
2			
3			
9			
5			
9			
7			
9			
9			
10			
10			
12			
13			
14			

Form 3 (continued)

Figure 2

Resolution of the detector (eV)	
------------------------------------	--

Figure 3

Element	k-value
Mg	
Si	
Ca	
Fe	

Form 3 (continued)

Figure 4 (collected from SRM 2063)

1. What features are abnormal about this spectrum? (compare to Figure 3 and use information from Table 3)

2. How would one know that the spectrum was abnormal if the chemistry of the material was not known?

Give at least one possible explanation for the cause of the abnormal spectrum of Figure 4 for the following cases.

3. The abnormal spectrum is collected from all areas observed on an SRM 2063 sample.

4. The abnormal spectrum is collected from one area of an SRM 2063 sample - other areas observed before and after the abnormal spectrum give the "normal" spectrum as shown in Figure 3.

Lab Code \_\_\_\_\_

Proficiency Test 91-1

Part 2 - Magnification calibration

Form 4

- 1) Figure 5 is a color copy of a print of a diffraction grating replica which contains 2160 lines/mm. Determine the total magnification of the replica. In the space below, list any measured distances in mm and show the work done to calculate the magnification and the standard deviation of this value (if multiple measurements were made). Enter the total magnification and standard deviation in the spaces provided.


Total magnification	
Standard deviation	

Lab Code \_\_\_\_\_

Form 4 (cont'd)

- 2) In the space below, please describe the method used to calibrate the magnification of the transmission electron microscope(s) in your laboratory.

Figure 5



Diffraction grating replica





## II. Summary Report for PT91-1

This summary report states that 32 laboratories did not pass this proficiency test. There was a change in the grading after laboratories reported that an incorrect formula for k-values was given at a course for asbestos analysis. After eliminating the points deducted for incorrect k-values, only three laboratories did not pass this test.

The information presented in this report is a summary of the results and performance of laboratories on the proficiency test distributed in April of 1991 (designated as test 91-1). Discussion of the responses to the problems in this proficiency test is given in the main portion of the report. Suggestions for further reading are given in Appendix A. The results obtained by the laboratory receiving this report are given in Appendix B.

Part 1. Identification of peaks on a spectrum

From the spectrum given in Figure 1 and from the information given in Table 1 of the proficiency test, laboratories were asked to identify the peaks labelled 1-14. An identification of a peak included reporting the element and the peak type ( $K\alpha$ ,  $K\beta$ ,  $L\alpha$ , etc.). The correct responses are given in the following table.

Peak number	Peak identification Choice #1	Peak identification Choice #2	Peak identification Choice #3
1	Na $K\alpha$ ( $K\beta$ )	(Cu $L\alpha$ , $L\beta$ )	none
2	Mg $K\alpha$ ( $K\beta$ )	none	none
3	Al $K\alpha$ ( $K\beta$ )	none	none
4	Si $K\alpha$ ( $K\beta$ )	none	none
5	K $K\alpha$	none	none
6	Ca $K\alpha$	(K $K\beta$ )	none
7	Ca $K\beta$	none	none
8	Ti $K\alpha$	none	none
9	Ti $K\beta$	none	none
10	Mn $K\alpha$	none	none
11	Fe $K\alpha$	(Mn $K\beta$ )	none
12	Fe $K\beta$	none	none
13	Cu $K\alpha$	none	none
14	Cu $K\beta$	none	none

In this table, minor peaks are indicated by parentheses. It was stated in the instructions that no peaks were present in the spectrum that were greater than 10 keV. Therefore, no L lines other than for Cu should have been indicated by the laboratories for this spectrum.

- Evaluation

The identification of each peak type is worth one point. Partial credit was given if the element was correctly identified but the electron-transition designation (K, L or  $\alpha$ ,  $\beta$ ) was incorrect (-0.25 for each incorrect designation). For Na, Mg, Al and Si no points were deducted if a laboratory reported K as the peak type instead of  $K\alpha$ . Partial credit (0.5 points) was given to laboratories if they reported

a possible L or K line as their first choice and the correct K line as their second choice. The second and third choices of the laboratories were not otherwise evaluated.

Part 2. Determination of detector resolution

The laboratories were given a copy of a spectrum showing a Mn K $\alpha$  x-ray peak (Figure 2 in the proficiency-test instructions) and a table listing the channels, the counts per channel and the corresponding energies (Table 2) for each point in the spectrum. The laboratories were asked to determine to the nearest eV, the resolution at FWHM of the detector used to collect the spectrum. The correct answers include one of the following:

Conditions	Resolution
Without background subtraction	144 eV
With background subtraction	145 eV

Many laboratories obtained values other than those above because approximate methods for determining detector resolution were used. A range of values from 140-160 eV was accepted for the resolution because the method for resolution determination and the expected accuracy are not specified in the NVLAP Program handbook for airborne asbestos analysis (Steel et al., 1989). A discussion of two acceptable methods for obtaining peak resolutions are given below.

- Method 1 - Interpolation of values in Table 2

To determine the resolution of the detector using interpolation of the data in Table 2 of the proficiency-test instructions (without subtraction of background), the following steps can be taken:

- 1) Determine the maximum number of counts for a channel in the spectrum.
- 2) Divide this value by 2 to obtain the value for the half maximum.
- 3) Find on Table 2 the range of counts and energies within which the half maximum falls for both the low and high energy sides of the peak. Label the energies of the channels bracketing the half maximum on the low energy side as A and B (where B is the higher energy). Similarly, label the channels bracketing the half maximum on the high energy side as C and D (where D has the higher energy). The counts for the corresponding energies are labelled as a, b, c and d (see Figure 1 on the following page).
- 4) Use interpolation to obtain the energies corresponding to the half maximum values on the low and high energy sides of the peak. For the interpolation, the following formula can be used determine the resolution:

$$Resolution = \frac{b-hm}{b-a} \cdot 10 \text{ eV} + \frac{c-hm}{c-d} \cdot 10 \text{ eV} + (C-B) \text{ eV} \quad (1)$$

where a, b, c, d, B and C were defined in the step above and hm equals the number of counts at half maximum.

# Manganese K-alpha Peak for X-ray Spectrometer Calibration




Figure 1. Illustration of the location of the energies (A, B, C and D) and counts (a, b, c and d) discussed in Method 1 for the determination of detector resolution.

Following this procedure for the peak given in Figure 2 and the values given in Table 2 of the proficiency test instructions leads to the following:

- 1) The maximum number of counts in a channel in Table 2 is 59780 (in channel 572).
- 2) Half of this value is 29890.
- 3) On the low energy side of the peak the half maximum falls within the following values:

Channel	Counts	Energy
564	26742	5.809
565	32099	5.819

On the high energy side of the peak, the half maximum falls within the following values:

Channel	Counts	Energy
579	30101	5.959
580	24424	5.969

- 4) To determine the resolution, the values above are put into equation 1 as follows:

$$Resolution = \frac{32099 - 29890}{32099 - 26742} \cdot 10 \text{ eV} + \frac{30101 - 29890}{30101 - 24424} \cdot 10 \text{ eV} + 140 \text{ eV} = 144.49 \text{ eV}^{(2)}$$

Rounding this number to the nearest eV, a value for the detector resolution of 144 eV is obtained.

Note: This method can be easily used with background subtracted counts. The values for the maximum number of counts, the half maximum and a, b, c and d will be slightly reduced by the background subtraction.

#### Method 2 - Estimation from physical measurement of the peak

Some laboratories derived the actual resolution by measuring distances on the peak. The following steps are followed:

- 1) The height of the peak is measured on the spectrum in mm or cm.
- 2) Half the height measured above is determined and a line is drawn through the peak at that height.
- 3) At the two places where the line intersects the peak, perpendicular lines are drawn that intersect the x axis (containing the energy in eV).
- 4) The physical distance (d) between the two lines is noted in mm or cm. The physical distance (D) between two arbitrarily chosen energies on the x axis is measured. The difference between the energies is determined (E).
- 5) The resolution is determined by solving for x in the following equation:

$$\frac{d}{x} = \frac{D}{E} \quad (3)$$

This method can give acceptable results if applied carefully. However, method 1 is preferable.

Comments

The resolution of an x-ray detector affects the degree of overlap between adjacent peaks in a spectrum and the detection limit of minor peaks. One of the requirements of the NVLAP handbook (Steel et al., 1989) is that the detector shall have a resolution of 175 eV or better at Mn K $\alpha$ . For this requirement a reasonably accurate determination of the resolution value (to within  $\pm 2$  eV) is assumed. Without such accuracy, significant under estimation may occur, e.g., a detector with approximately 185 eV resolution may be improperly calculated to have a resolution of 170 eV. To obtain such accuracy, at least 5,000 counts for the peak height should be accumulated to obtain adequate counting statistics and an appropriate method for resolution determination should be used. Methods in which resolution is estimated to the nearest channel (10 eV) should not be used. Similarly, methods in which a theoretical resolution is calculated based on electronic noise and the peak energy should not be used.

There is software available (some of which is associated with detector-multichannel analyzer systems) that will calculate the resolution of a peak. For some programs, a Gaussian curve is fitted to the peak so that the top of the peak and the FWHM can be accurately determined. This software should be used to monitor detector resolution only if periodically verified by Method 1 above.

A range of values for the resolution of 140-160 eV was considered acceptable for this proficiency test. This range was chosen because resolutions of 140, 150 or 160 eV can be obtained by estimation to the nearest channel. As stated above, this method should not be used to determine the resolution of detectors used in the laboratory.

- Evaluation

The value determined for the resolution is worth one point.

Part 3. Determination of k-values

For this part of the proficiency test, laboratories were asked to determine k-values for Mg, Si, Ca, and Fe based on information given in Tables 3 and 4 and Figures 4a and b of the proficiency test. Correct values are given in the following table:

Element	k-value
Mg	1.59
Si	1.00
Ca	0.99
Fe	1.27

- Method

To determine k-values relative to Si, the following formula is used:

$$\frac{C_A}{C_{Si}} = k_{ASi} \cdot \frac{I_A}{I_{Si}} \quad (4)$$

where A represents the element of interest, C is the concentration of element A in weight percent, and I is the intensity of the characteristic peak corresponding to the element. Solving for  $k_{ASi}$ :

$$k_{ASi} = \frac{C_A}{C_{Si}} \cdot \frac{I_{Si}}{I_A} \quad (5)$$

As an example, the k-value for Mg is obtained as follows:

$$k_{MgSi} = \frac{8.8}{25.4} \cdot \frac{980256}{213242} = 1.59 \quad (6)$$

Common errors made by the laboratories include the use of atomic % element or weight % oxide for the concentration, use of no concentration information (just peak intensities), and determining the reciprocal of the value for the k-values.

Note: It is recommended that the k-values for Na, Mg, Al, and Si be determined using the total peak integral (the  $K\alpha + K\beta$ ) while the k-values for Ca and Fe be determined using the intensities of only the  $K\alpha$  peak. This approach of using the integral of a single peak is recommended because the separate handling of the  $K\beta$  peak takes more processing steps and has the potential to introduce errors to the measurement. Please note that this approach may cause slight differences in k-values compared to literature or theoretical values. For this proficiency test, we also accepted values for Ca and Fe obtained by summing the intensities of the  $K\alpha$  and  $K\beta$  peaks.

- Evaluation

Each of the k-values are worth one point.

Part 4. Interpretation of an "abnormal" spectrum

For this section of the proficiency test, laboratories were asked to respond to questions concerning an "abnormal" spectrum collected from SRM 2063. The spectrum was displayed with a linear and log scale (Figures 4a, b) and the peak counts were given in Table 5 of the proficiency-test instructions. The laboratories were asked to compare the spectrum to the "normal" spectrum of SRM 2063 given in Figure 3 of the proficiency-test instructions.

This section was put into this proficiency test for educational purposes. It has been noted that some analysts have difficulty recognizing or interpreting "abnormal" x-ray spectra. Problems were also noted in some of the laboratory count sheets from the first proficiency test that could have resulted from misinterpretation of x-ray spectra. This section was not graded and therefore was not used to evaluate the laboratories' proficiency. The laboratories should, however, compare their answers to those given below.

1. What features are abnormal about this spectrum?

Acceptable answers are as follows:

- 1) The Mg and Si peaks are essentially absent.
- 2) The Ca peak is significantly reduced relative to the Fe and Cu peaks.
- 3) The continuum radiation does not have the characteristic shape usually present in energy dispersive spectra. The x-ray background decreases noticeably from about 5.5 keV to below 1 keV.

2. How would one know that the spectrum was abnormal if the chemistry of the material was not known?

The unusual shape of the x-ray background is an indication that the spectrum is abnormal.

3. What are possible explanations for the abnormal spectrum if it is collected from all areas observed on an SRM 2063 sample?

The most general response to this question is that something is present between the sample and the detector that is causing abnormal absorption of x-rays. Possible answers include:

- 1) There is an absorbing material on the detector window. The window may be coated with oil or ice.
- 2) There is a problem with the detector-sample geometry. For example, the detector may not be fully inserted or the sample holder could be at an improper tilt angle.
- 3) There is an absorbing material on the sample itself.

The most likely cause of the spectrum is the presence of oil or ice on the detector window. The detector window can accumulate oil or ice over time because its cool temperature provides a site for condensation of contaminants. For this case, the x-ray spectra collected from different areas of the grid would show the same degree of absorption at the lower energies. For the case of a problem with the sample-holder tilt angle, however, the x-ray spectra collected from different areas of the sample would likely differ slightly. This is because in different areas of the grid, different thicknesses of the sample holder would be between the sample and the detector. The presence of an absorbing material on the sample itself is a less likely cause. A layer thick enough to cause such absorption of x rays would cause problems with imaging of the sample.

Many incorrect responses were given by the laboratories. Some laboratories stated that the SRM was manufactured incorrectly. However, the presence of an "abnormal" background would indicate a different problem. In addition, Standard Reference Materials are tested extensively prior to and after production and though there is a possibility of a problem, other options should be checked first. Some laboratories postulated that the detector window is



too thick. Although this could create such a spectrum, the problem should have been noted upon receipt of the detector. Other laboratories stated that air or water vapor may be present in the microscope column thereby causing absorption of x rays. However, the loss of vacuum would affect the operation of the TEM. Finally, several laboratories postulated problems with the electronics of the detector-mca system (preamplifier, pulse processor, presence of ground loops, software problems, etc.). No such problem has been identified that would result in the spectrum of Figures 4a,b.

4. What are possible explanations for the abnormal spectrum if it is collected from one area of an SRM 2063 sample if other areas observed before and after the abnormal spectrum give the "normal" spectrum?

As for the previous case, it is likely that something is present between the sample and the detector that is absorbing the low-energy x rays. However, as this was only a localized situation, contamination of the detector window is not the cause. Possible causes are:

- 1) There is a problem with the specimen-detector geometry. Examples include:
  - a) The sample holder is not at a correct angle to the detector and the sample holder is in the line of sight of the detector. X rays are absorbed by the sample holder.
  - b) The sample is at a correct angle to the detector. However, x-ray spectra obtained from some portions of a grid (near the edges) can still be partially blocked by the sample holder.
  - c) The sample is too close to a grid bar. X rays are absorbed because the grid bar is between the sample and detector. This is usually a problem only if the grid is oriented above the sample.
- 2) There are large topographic changes in the sample. A portion of the sample is absorbing x rays.
- 3) There is localized contamination of the sample.

The last two explanations are less likely as the TEM operator would probably notice an unusual image obtained under such conditions.

An incorrect explanation given by the laboratories is that the specimen is inhomogeneous. The presence of the abnormal background in the spectrum is an indication that absorption, not chemical inhomogeneities, is the problem. Another incorrect explanation is that there are problems with the electronics. As stated above, no such problem has been identified that would generate the spectrum of Figures 4a,b.

#### Part 5. Magnification calibration

The laboratories were given a color copy of a TEM micrograph of a carbon replica made from an optical grating and were asked to determine the magnification of the micrograph. A histogram of the values reported by the laboratories was made and the obvious outliers discarded. The mean value for the magnification and its standard deviation ( $s$ ) are given in the following table:

Mean	s
27,402x	158

The values obtained by NIST personnel fall within one standard deviation of the mean given above. The reported magnifications considered acceptable for this proficiency test ranged from 26,700x to 28,100x.

- Discussion of results

The majority of laboratories used the correct method for deriving the magnification. Problems in the magnification determination occurred when laboratories measured a small number of lines - some laboratories measured only one, two or three lines. When such a small number of lines are measured, errors are introduced due to variations in the optical grating, distortions in the replica and from the measuring process itself. As many lines as possible should be measured for each magnification determination.

The standard deviation of the measurements was determined incorrectly by some laboratories. The following formula should be applied:

$$s = \sqrt{\frac{\sum (X - \text{Mean})^2}{N - 1}} \quad (7)$$

where X = a measurement value, Mean = the mean of the measurement values and N = the number of measurements.

- Evaluation

The magnification determination is worth one point.

Discussion of overall results of the laboratories

The results of the laboratory receiving this report are given in Appendix B. A laboratory passed this proficiency test if it accumulated less than two errors. Thirty-two laboratories did not pass this proficiency test.

References

E.B. Steel, S. Turner, H.W. Berger, NVLAP Program Handbook for Airborne Asbestos Analysis, National Institute of Standards and Technology, NISTIR 89-4137, 1989.

Appendix A. Suggestions for further reading

k-factors

G. Cliff, G.W. Lorimer, The quantitative analysis of thin specimens, J. of Microscopy, vol. 103, pp 203-207.

see the NVLAP handbook p. F5, item 10h for additional references on k-factors.

EDS artifacts

J.L. Goldstein, D. E. Newbury, P. Echlin, D.C. Joy, C. Fiori, E. Lifshin, Scanning Electron Microscopy and X-ray Microanalysis, Plenum Press, New York, NY, 1981, pp 226-264.



### III. Instructions for PT91-2

## **NIST AIR PROGRAM PROFICIENCY TEST 91-2**

The major purpose of this proficiency test is to evaluate each laboratory's abilities to do the following:

- Obtain and record selected area electron diffraction patterns.
- Determine d-spacings from the diffraction patterns.
- Obtain, record and interpret EDXA spectra.
- Report results correctly.

Please read the following directions and examine the forms carefully before beginning the test.

1. The grid box included with the test contains a grid which is located in the center slot. This location is circled on the grid box cover. Place the grid in the electron microscope in the same manner and under the same instrument conditions used for asbestos analysis.
2. The grid has several materials deposited on its surface. Obtain a diffraction pattern and an EDXA spectrum from 3 different grains on the grid which have the following characteristics:
  - The composition of the grains consists of mainly aluminum and silicon, and no potassium or calcium. Minor iron is acceptable.
  - The grains are single crystals; i.e. they are not part of an agglomerate of several crystallites.
  - The grains are electron transparent; i.e. they produce clear diffraction patterns.
  - The diffraction patterns are zone-axis patterns that do not require orientation of the sample grain at 0° stage tilt.
  - All patterns should be of the same zone.
3. Photograph the diffraction patterns, and obtain hard copies of the EDXA spectra.
4. Label the diffraction patterns and the EDXA spectra with the following designations: Particle #1, Particle #2, and Particle #3. Diffraction patterns may be labelled by writing both the negative ID# and the Particle # (1-3) on the negative carrier.

**NIST AIR PROGRAM PROFICIENCY TEST 91-2**

5. Determine the three largest unique d-spacings on each of the 3 diffraction patterns. Record these d-spacings (using Å units) in the appropriate spaces on Form #3, and show the work done in calculating the spacings on Form #1. The d-spacing should be reported to two decimal places (to 0.01 Å).
6. Circle the measured diffraction spots on the negative of the diffraction pattern.
7. Calculate the mean and standard deviation of the largest d-spacings determined on the three diffraction patterns obtained from the three particles. Similarly, determine the mean and standard deviation for the second and third largest d-spacings. Record these values in the appropriate spaces on Form #3. Show the method used to calculate these values on Form #2.
8. Submit negatives of the three diffraction patterns and hard copies of the three EDXA spectra along with your results. Save the grid for training purposes. The grid may be recalled by RTI or NIST if necessary.

**Note:** The laboratories should be able to distinguish between reflections on a diffraction pattern that have d-values that differ by 0.1 Å or more. For example, if a material has d-spacings of 5.20 Å and 5.32 Å, the laboratory should be able to determine that these two spacings are different. The accuracy of d-spacing determination must be within  $\pm 5\%$  relative of the true value.

Form #1

Show the work done to determine the three largest d-spacings on each of the three diffraction patterns. Include the camera constant used and any measurements made directly from the diffraction patterns. Use additional sheets if necessary.



Form #2

In the space below, show the method used to calculate the means and standard deviations of the d-spacings.

A large, faint rectangular box intended for showing calculations. The box is mostly empty and serves as a workspace for the student to demonstrate their method for calculating means and standard deviations.

Form #3

	Largest d-spacing (Å)	Second largest d-spacing (Å)	Third largest d-spacing (Å)
Particle #1			
Particle #2			
Particle #3			
Mean			
Standard deviation			

#### IV. Summary Report for PT91-2

The information presented in this report is a summary of the analysis of materials and the performance of laboratories on the proficiency test sent to laboratories in the airborne asbestos program in October of 1991 (designated as test 91-2). Discussion of results for both the diffraction pattern analysis and of the structure print submission is given in the main portion of the report. The results obtained by the laboratory receiving this report are given in Appendix A of the report.

## **PART 1 - DIFFRACTION PATTERN ANALYSIS**

### Material sent to laboratories

Laboratories were sent a carbon film grid containing a slurry of silicate layer materials. The laboratories were asked to obtain diffraction patterns and EDXA spectra from three different grains on the grid. The grains analyzed were to have the following characteristics: 1) their composition is mainly Al, Si, 2) the grains are single crystals that give clear diffraction patterns, 3) no orientation is required at 0° stage tilt, and 4) the patterns are of the same zone. The material fitting these criteria corresponds to kaolinite. The source material was obtained from Fisher Scientific.

### Analysis of material

The laboratories were asked to report the three largest unique d-spacings from the three diffraction patterns. To determine those values, the material was analyzed by NIST and RTI by two methods: 1) x-ray diffraction using internal mica and silica standards and 2) by electron microscopy using an internal gold standard. As shown below, the results obtained from the two techniques are not in complete agreement.

#### *X-ray diffraction*

The material was analyzed on a Philips PW1800 automated diffractometer equipped with an automatic divergent slit, copper radiation, and a graphite monochromator. The receiving slit was set to 0.3 mm. The power settings used were 40 kV and 55 mA. Initially, a qualitative analysis was made of the material by scanning three slurry mounts from 5 - 90° 2θ with a count time of two seconds per 0.01° 2θ step. The diffraction patterns consist primarily of kaolinite peaks with some additional minor peaks attributable to mica.

The d-spacings for kaolinite were determined by spiking a portion of the bulk sample with both SRM 640b (silicon) and SRM 675 (fluorophlogopite) as internal standards. Twelve backloaded powder mounts of the spiked material were prepared and scanned at six seconds per 0.01° 2θ step over a range of at least 10 - 40° 2θ. A calibration curve was prepared from the internal standards and applied to the data to correct for any sample displacement. Profile fitting was used to determine peak locations.

A summary of the average d-spacings obtained for those reflections in the [001] zone of kaolinite and a comparison to those values found on JCPDS card 14-164 are given in Table 1:

Table 1. Comparison of d-spacings obtained from x-ray diffraction (this work) to those given on JCPDS card 14-164 for reflections in the [001] orientation of kaolinite.

hkl	d-spacing (14-164)	d-spacing (this work)	standard deviation
(020)	4.478 Å	4.46 Å	0.0036 Å
(-110)	4.366	4.36	0.0018
(130)	2.553	*	
(200)	2.495	2.496	0.0003
(040)	2.237	*	
(220)	2.173	*	

\* were not determined due to peak overlaps

### Electron microscopy

Two grid preparations from the same lot prepared for the proficiency test were covered with a gold film by sputtering. Diffraction patterns were obtained by two analysts on two instruments - a Hitachi H7000 and a JEOL 200CX.

A total of 21 diffraction patterns were obtained. The d-spacings for the six diffracted spots closest to the central spot in the pseudohexagonal diffraction pattern were determined by measuring the distance covered by two to ten of the diffracted spots in the appropriate row. The spacing of the fourth gold ring (corresponding to 1.23 Å) was measured in the six directions corresponding to the diffracted spots and used to derive camera constants for the diffraction patterns.

The average d-spacings and the range of d-spacings determined for the six diffracted spots in 21 [001] diffraction patterns are shown in Table 2:

Table 2. Average and range of d-spacings obtained by SAED analysis of 21 gold-coated specimens.

spot number	d-spacing	standard deviation	range
1	4.48 Å	0.02 Å	4.45 - 4.50 Å
2	4.46	0.01	4.44 - 4.50
3	4.45	0.02	4.43 - 4.49
4	2.59	0.01	2.57 - 2.61
5	2.58	0.01	2.56 - 2.59
6	2.57	0.01	2.55 - 2.59

The difference between spots 1-3 ranged from 0.01 to 0.05 Å and averaged 0.02 Å. The difference between spots 4-6 ranged from 0.01 to 0.04 Å and averaged 0.02 Å.

### *Discussion of analyses*

Surprisingly, the results from x-ray diffraction and electron diffraction differ. The results of x-ray analysis correspond to those given by JCPDS card 14-164. The two largest distinguishable d-spacings are given by 4.46 and 4.36 Å. The results of electron microscopy analysis, however, do not show a spacing that corresponds to 4.36 Å. We at present do not know why there is a discrepancy between the two techniques. One speculation is that a slight change in the structure occurs in the vacuum of the electron microscope or by heating in the electron beam. If this is true, it is possible that different microscopes or different beam doses would affect the structure slightly differently.

### Results submitted by laboratories

One-hundred and thirty labs (main and subfacilities) submitted results for this proficiency test. The laboratories were asked to report the three largest unique d-spacings for the diffraction patterns. A variety of d-spacings were reported by the laboratories as the three largest.

The laboratories were graded on the value submitted for the largest d-spacing. The average value obtained by the laboratories is 4.47 Å with a standard deviation of 0.12 Å (obvious outliers were discarded in derivation of this value). This value compares well with the average value of 4.46 Å obtained at NIST and RTI by both XRD and TEM analysis. As stated in the instructions for the proficiency test, the laboratories were expected to obtain a value within 5% of the true value (SAED is capable of 1% or better accuracy, but for most cases 5% accuracy is sufficient to distinguish asbestos minerals from nonasbestos minerals). Thus the laboratories should have obtained a value in the range of 4.46 Å ± 0.22 Å. Since this was the first time measurement of patterns obtained by the laboratory was tested, however, we accepted values in the range of 3.96 to 4.96 Å. Those laboratories that obtained values outside the range of 4.24 to 4.68 Å should review and improve their procedure for analyzing diffraction patterns. Future tests of diffraction will require 5% accuracy.

Errors made by the laboratories include: 1) choosing an incorrect diffracted spot as having the largest d-spacing, 2) using a "radius" measurement for the camera constant and a "diameter" measurement for the d-spacing measurement (or the reverse) resulting in a halving or doubling of the derived d-spacing and 3) reporting the inverse of the d-spacing.

The values submitted for the second and third-largest d-spacings were not graded. There were two reasons for not evaluating these results. Firstly, laboratories interpreted the phrase "three largest unique d-spacings" in a variety of ways. Secondly, as discussed in the previous section, it is possible that different microscope conditions could affect the structure slightly differently leading to a variety of d-spacings.

The EDXA spectra submitted by the laboratories were reviewed. If any features of concern were observed on the EDXA spectra or SAED patterns, they are described in a note given in the Appendix to the cover letter.

## **PART II - STRUCTURE PRINTS**

For the second part of this proficiency test, laboratories were asked to provide at least ten prints or negatives of air-collected samples that had been analyzed in their laboratory using the AHERA protocol. The laboratories were asked to identify the type of asbestos, the number of structures present and the loading of the filters from which the images were obtained. Although not required, the laboratories were additionally asked to provide an image of their calibration grating.

One-hundred and twenty-one laboratories submitted either prints or negatives of asbestos analyzed in their laboratories. One of the laboratories submitted images of bulk asbestos samples. A total of 1292 prints or negatives were submitted that contained airborne asbestos. The types of asbestos reported by the laboratories (on a print or negative basis) are summarized in Table 3:

Table 3. Types of asbestos found on structure prints submitted by laboratories.

Chrysotile	94%
Amosite	5%
Tremolite	0.2%
Anthophyllite	0.01%
"amphibole"	1%

We were interested in determining the types of structures present on the prints and negatives. The asbestos structures were examined to determine if they were contained within the border of the negative or print. Ninety-eight of the laboratories had at least one negative or print with a fiber that was only partially in the field of view. The images from these laboratories were not used for further study, because the determination of the structure type from these labs could be either incomplete or biased.

The images from 22 laboratories were examined to determine the types of structures present. The 223 images were examined independently by two analysts and discrepancies were resolved in many cases. The total number of structures determined by the two analysts were 265 and 273, respectively. The structures were categorized using the counting rules given by EPA as modified by NIST. Five categories were used - single fiber, double fiber, bundle, matrix and cluster (a double fiber consists of two parallel, touching fibers). The distribution of structures is given in Table 4:

Table 4. Summary of types of structures on structure prints submitted by laboratories.

	Analyst 1	Analyst 2
	-----	-----
Single fiber	45%	41%
Double fiber	9%	8%
Bundle	18%	18%
Matrix	21%	24%
Cluster	7%	9%

### Discussion

The submitted images are of use in several areas. Firstly, they provide a database of actual air-collected structures for future round robins that test the precision of counting rules. Secondly, they provide a database of structures for future proficiency tests of the laboratories' ability to apply counting rules. Thirdly, they provide a database from which the types of samples analyzed by the laboratories can be characterized. As shown in Tables 3 and 4, the relative distribution of asbestos types (chrysotile vs. amosite, etc.) and the distribution of structure types (single fibers vs. clusters, etc.) can be determined. We plan also to determine dimensional information from the prints. This information is useful background information for the production of testing materials used in the NVLAP proficiency testing program. It is also useful background information for the derivation of

counting rules. Finally, the images provide evidence that the laboratories are capable of satisfying item 29g in the checklist in the Handbook for Airborne Asbestos Analysis (Steel et al., 1989). This item requires that the laboratories are capable of recording bright field images on electron micrographs or other suitable media.

Tentative conclusions can be derived from the data. The conclusions are based on the assumptions that: 1) the data submitted by the laboratories are representative of those analyzed by all laboratories (it is a random sampling) and 2) the laboratories correctly analyze asbestos samples, i.e. amphiboles are identified if present. The images submitted show that chrysotile is by far the most predominant asbestos type analyzed by the laboratories. Amosite is the most predominant amphibole. Single fibers represent at least 40% of the structures analyzed and clusters and matrices represent over 25% of the structures analyzed.

### **PART III - GRADING OF THE PROFICIENCY TEST**

The laboratories were graded on the determination of the largest d-spacing on the kaolinite diffraction pattern and on the submission of ten structure prints. The subfacilities were evaluated solely on their determination of the largest d-spacing. Their evaluation is for informational purposes only. Seven main laboratories did not pass this proficiency test.

#### References

E.B. Steel, S. Turner, H.W. Berger, NVLAP Program Handbook for Airborne Asbestos Analysis, National Institute of Standards and Technology, NISTIR 89-4137, 1989.





