

NISTIR 7340

**NIST Intercomparison Exercise
Program for Organic Contaminants in
the Marine Environment:
Description and Results of 2005 Organic
Intercomparison Exercises**

**Michele M. Schantz
John R. Kucklick
Reenie M. Parris
Dianne L. Poster
Stephen A. Wise**

NIST

National Institute of Standards and Technology
Technology Administration, U.S. Department of Commerce

NISTIR 7340

NIST Intercomparison Program for Organic Contaminants in the Marine Environment: Description and Results of 2005 Organic Intercomparison Exercises

Michele M. Schantz, John R. Kucklick,¹ Reenie M. Parris,
Dianne L. Poster, and Stephen A. Wise
Analytical Chemistry Division
National Institute of Standards and Technology
Gaithersburg, MD 20899-8392 and
¹Charleston, SC 29412

September 2006



U.S. DEPARTMENT OF COMMERCE
Carlos M. Gutierrez, Secretary
TECHNOLOGY ADMINISTRATION
Robert Cresanti, Under Secretary of Commerce for Technology
NATIONAL INSTITUTE OF STANDARDS AND TECHNOLOGY
William A. Jeffrey, Director

Table of Contents

Abstract	7
Introduction	8
Sources and Preparation of Materials Used in 2005 Intercomparison Exercises	9
Storage and Distribution of Materials	9
Evaluation of Exercise Results	10
Establishment of the Assigned Values	10
Reported Results	11
Performance Scores	12
Accuracy Assessment (z-score)	12
Precision Assessment (p-score)	13
Discussion	13
Acknowledgments	16
Disclaimer and References	16
Tables	18
Table 1. Target analytes in NIST Intercomparison Exercise Program for Organic Contaminants in the Marine Environment	20
Table 2. Mussel Tissue XII (QA05TIS12): Laboratory means of three replicates and exercise assigned values - TEO, and PAHs	22
Table 3. Mussel Tissue XII (QA05TIS12): Laboratory means of three replicates and exercise assigned values - Pesticides	23
Table 4. Mussel Tissue XII (QA05TIS12): Laboratory means of three replicates and exercise assigned values - PCBs	24
Table 5. Mussel Tissue XII (QA05TIS12): Laboratory means of three replicates and exercise assigned values - PBDEs	25

Table 6.	SRM 2977: Laboratory means of three replicates and target values - TEO and PAHs	26
Table 7.	SRM 2977: Laboratory means of three replicates and target values - Pesticides.....	27
Table 8.	SRM 2977: Laboratory means of three replicates and target values - PCBs	28
Table 9.	SRM 2977: Laboratory means of three replicates and target values - PBDEs.....	29
Table 10.	Marine Sediment XIII (QA05SED13): Laboratory means of three replicates and exercise assigned values - Water, TOC, and PAHs	30
Table 11.	Marine Sediment XIII (QA05SED13): Laboratory means of three replicates and exercise assigned values - Pesticides.....	31
Table 12.	Marine Sediment XIII (QA05SED13): Laboratory means of three replicates and exercise assigned values - PCBs.....	32
Table 13.	Marine Sediment XIII (QA05SED13): Laboratory means of three replicates and exercise assigned values - PBDEs.....	33
Table 14.	SRM 1941b: Laboratory means of three replicates and target values - Water, TOC, and PAHs	34
Table 15.	SRM 1941b: Laboratory means of three replicates and target values - Pesticides.....	35
Table 16.	SRM 1941b: Laboratory means of three replicates and target values - PCBs	36
Table 17.	SRM 1941b: Laboratory means of three replicates and target values - PBDEs.....	37
Table 18.	Mussel Tissue XII (QA05TIS12): z-scores (25 %) by laboratory -TEO and PAHs.....	38
Table 19.	Mussel Tissue XII (QA05TIS12): z-scores (25 %) by laboratory - Pesticides	39
Table 20.	Mussel Tissue XII (QA05TIS12): z-scores (25 %) by laboratory - PCBs.....	40
Table 21.	Mussel Tissue XII (QA05TIS12): z-scores (25 %) by laboratory - PBDEs	41

Table 22. Marine Sediment XIII (QA05SED13): z-scores (25 %) by laboratory - PAHs, Water, and TOC	42
Table 23. Marine Sediment XIII (QA05SED13): z-scores (25 %) by laboratory - Pesticides	43
Table 24. Marine Sediment XIII (QA05SED13): z-scores (25 %) by laboratory - PCBs	44
Table 25. Marine Sediment XIII (QA05SED13): z-scores (25 %) by laboratory - PBDEs...	45
Table 26. Mussel Tissue XII (QA05TIS12) and SRM 2977: RSDs for three replicates - TEO and PAHs	46
Table 27. Mussel Tissue XII (QA05TIS12) and SRM 2977: RSDs for three replicates - Pesticides	48
Table 28. Mussel Tissue XII (QA05TIS12) and SRM 2977: RSDs for three replicates - PCBs	50
Table 29. Mussel Tissue XII (QA05TIS12) and SRM 2977: RSDs for three replicates - PBDEs	52
Table 30. Marine Sediment XIII (QA05SED13) and SRM 1941b: RSDs for three replicates - TEO and PAHs	54
Table 31. Marine Sediment XIII (QA05SED13) and SRM 1941b: RSDs for three replicates - Pesticides	56
Table 32. Marine Sediment XIII (QA05SED13) and SRM 1941b: RSDs for three replicates - PCBs	58
Table 33. Marine Sediment XIII (QA05SED13) and SRM 1941b: RSDs for three replicates - PBDEs	60
Table 34. Comparison of Concentrations for Marine Sediment XIII (QA05SED13) and SRM 1941b	62
Figure	65
Figure 1. Comparison of Concentrations for Mussel Tissue XII as determined in this study and in a CCQM pilot study for selected analytes	65

Appendices

Appendix A: Description, Storage, Use, and Reporting Instructions for Mussel Tissue XII (QA05TIS12)	66
Appendix B: Description, Storage, Use, and Reporting Instructions for Marine Sediment XIII (QA05SED13)	76
Appendix C: Laboratory Notes Accompanying Data, Mussel Tissue XII	86
Appendix D: Laboratory Notes Accompanying Data, Marine Sediment XIII	89
Appendix E: Laboratory Methods Used, Mussel Tissue XII	93
Appendix F: Laboratory Methods Used, Marine Sediment XIII	101
Appendix G: Charts of Mussel Tissue XII and SRM 2977, Results by Analyte	109
Appendix H: Charts of Marine Sediment XIII and SRM 1941b, Results by Analyte	193
Appendix I: Data Received after the first draft of this Report was Distributed	270
Appendix J: List of Laboratories Participating in 2005 Intercomparison Exercises	281

Abstract

In support of marine monitoring measurement programs, the National Institute of Standards and Technology (NIST) conducts interlaboratory comparison exercises. The intercomparability of data after participation in these exercises provides one mechanism for participating laboratories/monitoring programs to evaluate the quality and comparability of their performance in measuring selected organic contaminants in environmental samples. In this report, results of the 2005 exercises of the NIST Intercomparison Exercise Program for Organic Contaminants in the Marine Environment are described in which selected polycyclic aromatic hydrocarbons (PAHs), polychlorinated biphenyl (PCB) congeners, chlorinated pesticides, and polybrominated diphenyl ethers (PBDE) congeners were determined in Mussel Tissue XII and Marine Sediment XIII exercise materials. The analytical methods used by each participating laboratory in this performance-based program are also summarized.

Introduction

The preparation and distribution of two materials, Mussel Tissue XII (QA05TIS12) and Marine Sediment XIII (QA05SED13), used in interlaboratory comparison exercises in 2005 for the National Institute of Standards and Technology (NIST) Intercomparison Exercise Program for Organic Contaminants in the Marine Environment, and the results of these exercises are described in this report. The analytical methods used by each participating laboratory are also summarized.

Tools and mechanisms for the assessment of data produced by laboratories providing environmental analyses are critical because decision-making based on inaccurate results or data of unknown quality can have significant economic and health consequences. NIST provides a variety of activities in support of environmental monitoring programs for organic contaminants. The largest of these programs was initiated and funded in part for 12 years (until 1999) by the National Oceanic and Atmospheric Administration (NOAA) National Status and Trends (NS&T) Marine Monitoring Program [1,2,3]. The Environmental Protection Agency (EPA) Environmental Monitoring and Assessment Program (EMAP) also participated in the NIST/NOAA NS&T effort for a number of years. Private sector and other laboratories that could not be accommodated under the NOAA, EPA, and NIST funding have reimbursed NIST for participation costs and have participated in these exercises and workshops as part of the NIST Intercomparison Exercise Program for Organic Contaminants in the Marine Environment. NIST is now continuing this program on a pay-to-participate basis. Through this program, NIST provides mechanisms for assessing the interlaboratory and temporal comparability of data with the goal of improving measurements for the monitoring of organic contaminants such as polycyclic aromatic hydrocarbons (PAHs), polychlorinated biphenyl (PCB) congeners, chlorinated pesticides, and, as of the 2005 exercise, polybrominated diphenyl ethers (PBDE) congeners in bivalve, sediment, and fish samples. This program includes the development of improved analytical methods, production of needed NIST Standard Reference Materials (SRMs) and other control materials, conduct of annual interlaboratory comparison exercises, and the coordination of workshops to discuss the results of these exercises and to provide a forum for cooperative problem-solving efforts by participants. Current participants represent multi-laboratory monitoring programs as well as a number of individual programs, and include federal, state/municipal, university/college, private sector, and international laboratories. In this performance-based program, each participating laboratory uses its current methods for analysis of similar materials for its program customers.

For the annual intercomparison exercises, samples of two natural-matrix, homogeneous materials that are derived from the marine environment and that have not been fortified with any of the target analytes are analyzed by the participating laboratories. Typical materials, such as mussel or fish tissue homogenates or wetted marine sediment, have levels of target analytes in the 1 ng/g to 15000 ng/g range. The target analytes are listed in Table 1.

Numerical indices, z- and p-scores, are used to assess and track laboratory performance for accuracy and precision, respectively, and to provide a mechanism for assessing the comparability

of data being produced by the participating laboratories for over 75 target analytes, total organic carbon (TOC), percent total extractable organics (TEO), and percent moisture.

Sources and Preparation of Materials Used in 2005 Intercomparison Exercises

Mussel Tissue XII. Mussel Tissue XII was prepared by freeze-drying approximately 5.5 kg of SRM 1974b Organics in Mussel Tissue (*Mytilus edulis*) [4]. Following freeze-drying, the bulk material (580 g) was radiation sterilized and then sieved through 25, 45, and 60 mesh sieves. The material that passed through the 60 mesh sieve (<250 μm) was then homogenized and bottled with approximately 8 g of freeze-dried tissue per bottle. As the Mussel Tissue XII material was a different sieved subset of the original SRM, it would not be expected to have the same concentrations of the analytes of interest as the original SRM. Each participant received one bottle. This freeze-dried mussel tissue homogenate material had not been enriched or spiked.

Marine Sediment XIII. Marine Sediment XIII was prepared from SRM 2702 Inorganics in Marine Sediment. SRM 2702 was prepared from bulk dried sediment remaining from the preparation of SRM 1941b Organics in Marine Sediment [6]. However, the sediment bottled for SRM 2702 was sieved at 70 μm while the sediment bottled for SRM 1941b was sieved at 150 μm . The Marine Sediment XIII material was issued as a wet sediment to more closely match the matrix of wet sediments routinely analyzed by the laboratories. A calibrated toploader balance (resolution of 0.01 g) was used for weighing the SRM 2702 sediment and water. For each sample, 11.00 g of SRM 2702 sediment (as received) was weighed into a tared 2-oz, wide-mouth bottle. The bottle was then capped and stored in the dark at room temperature. Approximately four days before samples were to be shipped to laboratories participating in the intercomparison exercise, 9.0 g of HPLC-grade water were added by pipet to each tared bottle of sediment. (Preliminary trials had shown that a minimum of 9 g of water would moisten 11 g of this sediment.) The mass of sediment and water in each bottle were recorded. Each sample was tilted by hand until no dry sediment was visible. Only a very small amount of water was observed on the top of the wet sediment. After being held 24 h at room temperature (in the dark), followed by approximately 4 h at -20 °C, each bottle of material was stored at -80 °C until shipped. The bottles were never inverted until the wet samples had been frozen in the bottom of the bottles. The material was not enriched or spiked with any of the analytes of interest in this intercomparison exercise.

Storage and Distribution of Materials

Mussel Tissue XII material was stored at room temperature, and Marine Sediment XIII material was stored at -80 °C until shipped via overnight delivery to participating laboratories. Instructions for the storage and use of the exercise material and a diskette with files for electronic submission of data were included with each set of material shipped. These instructions are reproduced in Appendices A and B.

Each laboratory participating in these intercomparison exercises was sent the following by overnight delivery:

Exercise 1: Mussel Tissue XII (QA05TIS12)

One bottle of Mussel Tissue XII material (shipped on dry ice)
Description of the materials and storage/use/reporting instructions for the exercise (see Appendix A.)
Files for the reporting of results were sent as an e-mail attachment.

Exercise 2: Marine Sediment XIII (QA05SED13)

Three bottles of Marine Sediment XIII material (shipped on dry ice)
Description of the materials and storage/use/reporting instructions for the exercise (see Appendix B.)
Files for the reporting of results were sent as an e-mail attachment.

In the an e-mail message sent notifying the participants of the sample shipment, each participant was asked to analyze each of three replicate samples (three from one bottle for the mussel tissue and one from each jar for the sediment) to provide a more realistic assessment of laboratory precision and, if possible, to concurrently analyze the NIST SRM 2977 Mussel Tissue [7] with Mussel Tissue XII and NIST SRM 1941b Organics in Marine Sediment [6] with Marine Sediment XIII.

Evaluation of Exercise Results

Establishment of the Assigned Values

The following guidelines were used by the NIST exercise coordinators for the establishment of the exercise "Assigned Values" for these two exercises. Each laboratory's performance on concurrent Standard Reference Material (SRM) analyses was used to determine if that laboratory's results would be eligible for inclusion in the calculation of the exercise assigned value for the unknown material for a particular analyte. The results reported for the unknown materials from laboratories that did not report results for the SRMs were not used in these calculations. After the exercise assigned values, standard deviations, and 95% confidence limits had been calculated, all reported results for the Mussel Tissue XII and Marine Sediment XIII materials were evaluated relative to the exercise assigned values.

Laboratory data submission: Each participating laboratory was to submit data from three replicate determinations of the "unknown" materials (Mussel Tissue XII and Marine Sediment XIII) and was requested to report results of concurrent analyses of NIST SRM 2977, a freeze-dried mussel tissue SRM, and SRM 1941b, a marine sediment SRM. Laboratories were requested to report these results to three significant figures and to provide brief descriptions of their extraction, cleanup, and analytical procedures.

Determination of laboratory analyte means: For each laboratory, the laboratory analyte mean of the three sample results (S1, S2, and S3) was calculated for each analyte. Non-numerical data were treated as follows: A mean "<value" was used when three "<values" were reported; NA (not analyzed/determined) was used for three reported NA's; and, if the reported results were of mixed type, e.g., S1 and S2 were numerical values and S3 was reported as "<value", the two similar "types" were used to either determine the mean or to set a non-numerical descriptor.

Determination of assigned values: The assigned values are the means of the acceptable data as defined here. For a particular analyte, the performance on the reference material was deemed acceptable for the purpose of this exercise if the laboratory result was within 30 % of the upper and lower limits of the confidence interval for analytes listed in the Certificates of Analysis for SRM 2977 and SRM 1941b. For each analyte of interest for which a certified value is not provided in these materials, a “target” concentration and the associated uncertainty were calculated. The targets for SRM 2977 were based on reference concentrations for SRM 2977. The targets for SRM 1941b were based on results of the 1999 exercise in which SRM 1941b was used as the unknown material and for the PBDEs on an interlaboratory study coordinated in 2004 specifically for the determination of PBDE congeners in sediment [8]. Laboratory results within target upper and lower limits, typically 30 % to 40 %, of these concentrations were deemed acceptable for this exercise. If a laboratory demonstrated acceptable performance on a particular analyte in the reference material, that laboratory’s results for that analyte in the corresponding “unknown” exercise material was then used in the calculation of the analyte’s exercise assigned value, unless it was deemed an outlier. For evaluation of potential outliers, statistical tests and expert analyst judgement were used after viewing both normal and log normal plots of the data. This judgement utilized knowledge of potential coeluters based on the laboratory's reported methods. In instances in which the analyte concentration was below the detection limit of most participating laboratories, no exercise assigned value was calculated. In data sets where a number of laboratories report results as "not detected" at various detection limits, there is no consensus as to what numerical value should be assigned to these results in the computation of grand means, etc.; e.g., "0," half Detection Limit (DL), and the DL value itself have all been used and the choice is influenced by the particular data set.

Reported Results

Laboratories were assigned numerical identification codes in order of receipt of data with the exception of NIST, which is Laboratory 1 in these exercises. A laboratory was assigned the same code for each material. There are two results from NIST reported: 1a generated in the NIST Gaithersburg laboratory and 1c generated in the NIST Charleston laboratory. The laboratory mean replicate data are shown in Tables 2 to 5 and Tables 6 to 9 for the Mussel Tissue XII and SRM 2977, respectively, and in Tables 10 to 13 and 14 to 17 for Marine Sediment XIII and SRM 1941b, respectively. Included in the means tables for Mussel Tissue XII and Marine Sediment XIII are the exercise assigned values, the standard deviation of the assigned value, the percent relative standard deviation (% RSD), and the calculated 95 % confidence limit of the assigned value for the percent water (sediment), percent total extractable organics, TEO (mussel tissue), total organic carbon, TOC (sediment), PAHs, chlorinated pesticides, PCB congeners, and PBDE congeners. Notes included by a laboratory with its data are listed in Appendices C (Mussel Tissue XII) and D (Marine Sediment XIII). Summaries of the methods used by each laboratory are in Appendices E (Mussel Tissue XII) and F (Marine Sediment XIII). Tables 6 through 9 and 14 through 17 summarize the data received from the participating laboratories for SRM 2977 and 1941b, respectively. The certified and target values for the analytes of interest are also shown in these tables.

In Appendices G (Mussel Tissue XII) and H (Marine Sediment XIII), charts of the mean numerical results reported by each laboratory for each analyte are shown for the exercise material and the corresponding reference material.

Three laboratories reported data after the first draft of this report was distributed to the participants. The data from these laboratories are summarized in Appendix I but are not presented in the charts (Appendices G and H) and are not included in the calculation of the assigned values.

Performance Scores

The exercise coordinators recognize that different programs have different data quality needs. The acceptability of the results submitted by a particular laboratory will be decided by the individual program(s) for which the laboratory provides data. Typically, the program will use these exercise results in conjunction with the laboratory's performance in the analysis of certified reference materials and/or control materials, and of other quality assurance samples. These exercise results are exhibited in a number of ways in this report to facilitate their use by these programs in their acceptability assessments.

IUPAC guidelines [9] describe the use of z-scores and p-scores for assessment of accuracy and precision in intercomparison exercises such as those described in this report. These indices assess the difference between the result of the laboratory and the exercise assigned value and can be used, with caution, to compare performance on different analytes and on different materials.

Accuracy Assessment (z-score)

$$\text{z-score} = (\text{bias estimate})/(\text{performance criterion}) = (x - X)/\sigma$$

where x is the individual laboratory result, X is the "Exercise Assigned Value," and σ is the target value for standard deviation.

As described in the IUPAC guidelines, the choice of σ is dependent upon data quality objectives of a particular program. It can be "fixed" and arrived at by perception, prescription, or reference to validated methodology (e.g., $\sigma = 0.025 X$; X is the exercise assigned value), or it can be an estimate of the actual variation (e.g., the calculated sample standard deviation, s , from the exercise data). The "fixed" performance criterion is more useful in the comparison of a laboratory's performance on different materials while the use of the actual variation may be more useful within a given exercise, for example, if the determination of a particular analyte is exceptionally problematic.

We have calculated and reported z-scores using the fixed performance criterion for each analyte for each laboratory. At a previous workshop, it was decided to use "25 % of the exercise assigned value" as the fixed target value for standard deviation for this program. The z-scores calculated for these exercises can thus be interpreted as shown in the following examples:

z-score (25 % X):

+1 \Rightarrow laboratory result is 25 % higher than the assigned value

-2 \Rightarrow laboratory result is 50 % lower than the assigned value.

From a scientific point of view, IUPAC does not recommend the classification of z-scores but allows that a common classification is:

$ z \leq 2$	Satisfactory
$2 < z < 3$	Questionable
$ z \geq 3$	Unsatisfactory.

Tables 18 through 21 summarize the z-scores (25 %) for each laboratory for each reported analyte in Mussel Tissue XII while Tables 22 through 25 summarize the z-scores (25 %) for each laboratory for each reported analyte in Marine Sediment XIII.

Precision Assessment (p-score)

$$\text{p-score} = \sigma_{\text{lab}} / \sigma_{\text{target}}$$

Prior to the 1994 exercises, participating laboratories typically analyzed the three replicate samples for an exercise with the same sample set, i.e., one set of samples with the same blank, calibration curve, etc. applicable for each. Since the repeatability for replicates within a set generally shows better reproducibility than for replicates across different sets, this does not result in data that are very useful for realistic uncertainty assessment. Since 1994, laboratories have been requested to process each replicate in a different sample set for uncertainty assessment. For the calculation of p-scores for this program, the σ values used are coefficients of variation (CV calculated as relative standard deviations) with the current target σ (CV) for the three replicates being 15 %.

Tables 26 through 29 summarize the relative standard deviations (RSDs) calculated from the three concentrations reported by the laboratory for each analyte quantified in Mussel Tissue XII while Tables 30 through 33 summarize the RSDs calculated for each reported analyte by laboratory in Marine Sediment XIII. To calculate the p-scores (15 %), divide the RSDs reported in the tables by 15%. If a different criterion is chosen, follow the same procedure, and divide the RSD by that criterion.

Discussion

Laboratories were requested to quantify 26 PAHs, 25 chlorinated pesticides, 25 PCB congeners, and 34 PBDE congeners in this year's exercise. A total of 12 sets of results were submitted for Mussel Tissue XII, and 11 sets of results were submitted for Marine Sediment XIII. In the mussel tissue exercise, one laboratory (12) reported data for SRM 2978 Mussel Tissue as the control material for the pesticides, and in the sediment exercise, one laboratory (2) reported data for SRM 1944 New York/New Jersey Waterway Sediment as the control material. Their data were evaluated based on the certified and target values for these SRMs (see Evaluation of Exercise Results above).

The concentrations of the PAHs of interest in Mussel Tissue XII range from 2 ng/g dry-mass basis to 200 ng/g dry-mass basis, the concentrations of the pesticides of interest range from < 1 ng/g dry-mass basis to 35 ng/g dry-mass basis, and the concentrations of the PCB congeners range from < 1 ng/g dry-mass basis to 90 ng/g dry-mass basis. For the chlorinated pesticides, 9 of the 25 compounds were above the detection limits for the majority of the laboratories reporting, while 22 of the 25 PCB congeners were above the detection limits for the majority of the laboratories. There was poor agreement among the laboratories for total extractable organics (TEO), ranging from 0.3 % to 15.0 % even though the laboratories are reporting using similar methods for determining the TEOs (Appendix E). TEO is sometimes referred to as percent lipid but is typically determined by taking a known portion of the extract and evaporating to dryness and then weighing the dried residue. As one can imagine, the TEO value is then dependent on the extraction method and solvent used and the drying method used. It is, therefore typical to see the TEO values vary greatly from lab to lab particularly for relatively lean (non-fatty) materials.

Some of the bottles prepared as Mussel Tissue XII were labeled for use in a separate interlaboratory study as part of the Organic Working Group of the Comité Consultative pour la Quantité de Matière (CCQM). Seven National Metrology Institutes (NMIs) or designated NMIs participated in this study for a limited number of analytes. The exercise means and standard deviations are shown in Figure 1 and compared to the exercise assigned means and standard deviations from this study. The means from the two studies agreed within the uncertainties of the data for this limited analyte set.

The z-scores for the PAHs, pesticides, PCB congeners, and PBDE congeners in Mussel Tissue XII based on 25 % of the exercise assigned value are summarized in Tables 18 to 21, respectively. The majority of the z-scores based on 25 % are within ± 2 (± 50 % of the exercise assigned value). The RSDs for Mussel Tissue XII are summarized in Tables 26 to 29 for the PAHs, pesticides, PCB congeners, and PBDE congeners, respectively. Only five laboratories reported results for a limited number of PBDE congeners; however, for some of the congeners the agreement among the laboratories reporting is good particularly considering the low concentrations.

The PAH concentrations in Marine Sediment XIII range from 25 ng/g dry-mass basis to 800 ng/g dry-mass basis. The pesticide concentrations range from below the detection limits of the methods used to 4.5 ng/g dry-mass basis, while the PCB concentrations range from <1 ng/g dry-mass basis to 5 ng/g dry-mass basis. There was good agreement among the laboratories for percent water in the wet sediment. Only four laboratories returned data for the TOC with the values ranging from 1.8 % to 3.0 %.

As discussed above, the material used for Marine Sediment XIII, SRM 2702, was prepared from the same bulk sediment as SRM 1941b with the difference in the sieve fraction used. SRM 2702 used the material <70 μm while SRM 1941b used the material <150 μm . The concentrations determined in this study for Marine Sediment XIII are compared to the certified and reference concentrations for SRM 1941b in Table 34. For the majority of the analytes, the concentrations in Marine Sediment XIII (SRM 2702) are lower than those in SRM 1941b although the difference between the two materials is not consistent across all of the analytes even within the PAHs, PCBs, and pesticides.

The z-scores for the PAHs, pesticides, PCB congeners, and PBDE congeners based on 25 % of the exercise assigned value are summarized for Marine Sediment XIII in Tables 22 to 25, respectively. In general, the z-scores based on 25 % were within ± 2 (± 50 % of the exercise assigned value) for Marine Sediment XIII. The RSDs for the Marine Sediment XIII are summarized in Tables 30 to 33 for the PAHs, pesticides, PCB congeners, and PBDE congeners, respectively. As for the Mussel Tissue XI, only five laboratories reported data for the PBDE congeners.

As in the past exercises, a variety of methods were used for extraction, extract cleanup, and analysis. These are summarized in Appendix E for the mussel tissue and Appendix F for the marine sediment. For the PAHs in the mussel tissue and marine sediment, all of the laboratories used gas chromatography with mass spectrometry (GC/MS). For the chlorinated analytes in the mussel tissue, laboratories 8, 11, and 12 specified the use of high-resolution MS, and laboratories 5, 9, and 10 used GC-ECD for the PCB congeners and 5, 6, 9, and 10 for the pesticides. For the PBDE congeners, laboratories 7, 8, and 12 used GC with high-resolution MS while laboratory 1c used GC with low-resolution MS in the negative chemical ionization mode and laboratory 4 used GC with low-resolution MS in the electron ionization mode. There was no obvious correlation between z-scores and method used.

For the 2005 exercises, the data provided in the various figures and tables of this report can be used for assessing the comparability of results of over 100 analytes of interest in this program and the performance of individual laboratories. In these exercises, interlaboratory variability is a greater contributor to measurement incomparability than intralaboratory variability.

Subgroups of the exercise participants have demonstrated comparability of results for many analytes within the 0 to 2 z-range based on use of 25 % of the exercise assigned concentration as the performance criterion. This implies that this subgroup can distinguish between two samples that have an analyte concentration difference of 100 %. The reported accuracy and reproducibility indices (z- and p-scores, respectively) can be easily converted to conform to the acceptability requirements of a particular program. For example, a z-score based on 25 % can be multiplied by two to convert to a z-score based on 12.5 % of the analyte concentration.

It is important to evaluate the non-quantitative results reported by each laboratory as well. Although these results are not easily presented or numerically evaluated, they are included in the various tables of this report that list the mean and individual results of the laboratories. The laboratory and its data users should closely examine these non-quantitative results. Decisions based on false negative or false positive results from a laboratory can lead to significant environmental and/or economic consequences. Some laboratories reported detection limits in these “real” matrix materials that may be too high for the data quality needs of their program(s), and these issues should be assessed as well.

Intercomparison exercises provide an important mechanism for assessing the comparability, accuracy, precision, and reproducibility of data being produced by the participating laboratories. Exercise materials similar in matrix, form, and analyte concentration to typical samples routinely

analyzed by the laboratories are most useful for demonstrating the level of comparability and for revealing potential problem areas.

For the determination of the target compounds in these complex marine matrices with relatively low concentrations of these analytes, the levels of bias and reproducibility of many of the participating laboratories meet their current acceptability requirements; however, there is certainly room for improvement. Minimizing the among-laboratory biases so that the analytical variability is significantly less than the field sampling variability should be an achievable goal.

Acknowledgments

The time and effort of the analysts and management of the participating laboratories and the assistance of the NIST Standard Reference Materials Program with the procurement and preparation of the exercise materials are gratefully acknowledged.

Disclaimer

Certain commercial equipment, instruments, or materials are identified in this report to specify adequately the experimental procedure. Such identification does not imply recommendation or endorsement by the National Institute of Standards and Technology, nor does it imply that the materials or equipment identified are the best available for the purpose.

References

1. A. Y. Cantillo and R. M. Parris, "Evaluation of Trace Organic NOAA Status and Trends Quality Assurance Project Performance," in *Quality Assurance for Analytical Laboratories*, M. Parkany (ed.), Royal Society of Chemistry, Spec. Publ. No. 130 (1993).
2. A. Y. Cantillo and R. M. Parris, *National Status and Trends Program Quality Assurance Project: Trace Organic Intercomparison Exercise Results 1986-1990*, NOAA Tech. Memo. NOS/ORCA 69, Silver Spring, MD (1994).
3. A. Y. Cantillo, *NS&T Quality Assurance Project Intercomparison Exercise Results 1991-1993*, NOAA Tech. Memo. NOS/ORCA 79, Silver Spring, MD (1995).
4. Certificate of Analysis for Standard Reference Material (SRM) 1974b Organics in Mussel Tissue (*Mytilus edulis*), National Institute of Standards and Technology (NIST), Gaithersburg, MD, 2003. https://srmors.nist.gov/view_detail.cfm?srm=1974B
5. Certificate of Analysis for Standard Reference Material (SRM) 2702 Inorganics in Marine Sediment, National Institute of Standards and Technology (NIST), Gaithersburg, MD, 2004. https://srmors.nist.gov/view_detail.cfm?srm=2702

6. Certificate of Analysis for Standard Reference Material (SRM) 1941b Organics in Marine Sediment, National Institute of Standards and Technology (NIST), Gaithersburg, MD, 2002. https://srmors.nist.gov/view_detail.cfm?srm=1941B
7. Certificate of Analysis for Standard Reference Material (SRM) 2977 Mussel Tissue (Organic Contaminants and Trace Elements), National Institute of Standards and Technology (NIST), Gaithersburg, MD, 2000. https://srmors.nist.gov/view_detail.cfm?srm=2977
8. Stapleton, H.M., Keller, J.M., Schantz, M.M., Kucklick, J.R., and Wise, S.A., NIST Intercomparison Exercise Program for Polybrominated Diphenyl Ethers (PBDEs) in Marine Sediment: Description and Results of the 2004 Intercomparison Exercise, NISTIR 7278, Gaithersburg, MD (2005).
9. IUPAC "The International Harmonized Protocol for the Proficiency Testing of (Chemical) Analytical Laboratories," Pure Appl. Chem. 65 (9), 2123-2144 (1993).

List of Tables

- Table 1. Target Analytes in NIST Intercomparison Exercise Program for Organic Contaminants in the Marine Environment.
- Table 2. Mussel Tissue XII (QA05TIS12): Laboratory means of three replicates and exercise assigned values – TEO and PAHs
- Table 3. Mussel Tissue XII (QA05TIS12): Laboratory means of three replicates and exercise assigned values – Pesticides
- Table 4. Mussel Tissue XII (QA05TIS12): Laboratory means of three replicates and exercise assigned values – PCBs.
- Table 5. Mussel Tissue XII (QA05TIS12): Laboratory means of three replicates and exercise assigned values – PBDEs.
- Table 6. SRM 2977: Laboratory means of three replicates and target values – TEO and PAHs.
- Table 7. SRM 2977: Laboratory means of three replicates and target values – Pesticides.
- Table 8. SRM 2977: Laboratory means of three replicates and target values – PCBs.
- Table 9. SRM 2977: Laboratory means of three replicates and target values – PBDEs
- Table 10. Marine Sediment XIII (QA05SED13): Laboratory means of three replicates and exercise assigned values – Water, TOC, and PAHs
- Table 11. Marine Sediment XIII (QA05SED13): Laboratory means of three replicates and exercise assigned values – Pesticides
- Table 12. Marine Sediment XIII (QA05SED13): Laboratory means of three replicates and exercise assigned values – PCBs
- Table 13. Marine Sediment XIII (QA05SED13): Laboratory means of three replicates and exercise assigned values – PBDEs
- Table 14. SRM 1941b: Laboratory means of three replicates and target values – Water, TOC, and PAHs.
- Table 15. SRM 1941b: Laboratory means of three replicates and target values – Pesticides.
- Table 16. SRM 1941b: Laboratory means of three replicates and target values – PCBs.
- Table 17. SRM 1941b: Laboratory means of three replicates and target values – PBDEs.

Table 18. Mussel Tissue XII (QA05TIS12): z-scores (25 %) by laboratory – TEO and PAHs

Table 19. Mussel Tissue XII (QA05TIS12): z-scores (25 %) by laboratory – Pesticides

Table 20. Mussel Tissue XII (QA05TIS12): z-scores (25 %) by laboratory – PCBs

Table 21. Mussel Tissue XII (QA05TIS12): z-scores (25 %) by laboratory – PBDEs

Table 22. Marine Sediment XIII (QA05SED13): z-scores (25 %) by laboratory – PAHs, Water, and TOC

Table 23. Marine Sediment XIII (QA05SED13): z-scores (25 %) by laboratory – Pesticides

Table 24. Marine Sediment XIII (QA05SED13): z-scores (25 %) by laboratory – PCBs

Table 25. Marine Sediment XIII (QA05SED13): z-scores (25 %) by laboratory – PBDEs

Table 26. Mussel Tissue XII and SRM 2977: RSDs for three replicates – TEO and PAHs

Table 27. Mussel Tissue XII and SRM 2977: RSDs for three replicates – Pesticides

Table 28. Mussel Tissue XII and SRM 2977: RSDs for three replicates – PCBs

Table 29. Mussel Tissue XII and SRM 2977: RSDs for three replicates – PBDEs

Table 30. Marine Sediment XIII and SRM 1941b: RSDs for three replicates – Water, TOC, and PAHs

Table 31. Marine Sediment XIII and SRM 1941b: RSDs for three replicates – Pesticides

Table 32. Marine Sediment XIII and SRM 1941b: RSDs for three replicates – PCBs

Table 33. Marine Sediment XIII and SRM 1941b: RSDs for three replicates – PBDEs

Table 34. Comparison of concentrations for Marine Sediment XIII and SRM 1941b

Table 1. Target Analytes in NIST Intercomparison Exercise Program for Organic Contaminants in the Marine Environment

Chlorinated Pesticides

hexachlorobenzene	2,4'-DDE
alpha-HCH (alpha-BHC)	4,4'-DDE
gamma-HCH (gamma-BHC, Lindane)	2,4'-DDD
beta-HCH	4,4'-DDD
heptachlor	2,4'-DDT
heptachlor epoxide	4,4'-DDT
<i>cis</i> -chlordane (alpha-chlordane)	aldrin
<i>trans</i> -chlordane (gamma-chlordane)	dieldrin
oxychlordane	endrin
<i>cis</i> -nonachlor	endosulfan sulfate
<i>trans</i> -nonachlor	endosulfan I
mirex	endosulfan II
chlorpyrifos	

Polychlorinated Biphenyl Congeners

<i>PCB No.</i>	<i>Compound Name</i>
8	2,4'-dichlorobiphenyl
18	2,2',5-trichlorobiphenyl
28	2,4,4'-trichlorobiphenyl
31	2,4',5-trichlorobiphenyl
44	2,2',3,5'-tetrachlorobiphenyl
49	2,2',4,5'-tetrachlorobiphenyl
52	2,2',5,5'-tetrachlorobiphenyl
66	2,3',4,4'-tetrachlorobiphenyl
95	2,2',3,5',6-pentachlorobiphenyl
99	2,2',4,4',5-pentachlorobiphenyl
101	2,2',4,5,5'-pentachlorobiphenyl
105	2,3,3',4,4'-pentachlorobiphenyl
118	2,3',4,4',5-pentachlorobiphenyl
128	2,2',3,3',4,4'-hexachlorobiphenyl
138	2,2',3,4,4',5'-hexachlorobiphenyl
149	2,2',3,4',5',6-hexachlorobiphenyl
153	2,2',4,4',5,5'-hexachlorobiphenyl
156	2,3,3',4,4',5-hexachlorobiphenyl
170	2,2',3,3',4,4',5-heptachlorobiphenyl
180	2,2',3,4,4',5,5'-heptachlorobiphenyl
187	2,2',3,4',5,5',6-heptachlorobiphenyl
194	2,2',3,3',4,4',5,5'-octachlorobiphenyl
195	2,2',3,3',4,4',5,6-octachlorobiphenyl
206	2,2',3,3',4,4',5,5',6-nonachlorobiphenyl
209	decachlorobiphenyl

Table 1. (continued)

Polycyclic aromatic hydrocarbons (PAH)

naphthalene	benz[<i>a</i>]anthracene
2-methylnaphthalene	chrysene
1-methylnaphthalene	triphenylene
biphenyl	benzo[<i>b</i>]fluoranthene
2,6-dimethylnaphthalene	benzo[<i>j</i>]fluoranthene
acenaphthylene	benzo[<i>k</i>]fluoranthene
acenaphthene	benzo[<i>e</i>]pyrene
1,6,7-trimethylnaphthalene	benzo[<i>a</i>]pyrene
fluorene	perylene
phenanthrene	indeno[1,2,3- <i>cd</i>]pyrene
anthracene	dibenz[<i>a,h</i>]anthracene
1-methylphenanthrene	benzo[<i>ghi</i>]perylene
fluoranthene	
pyrene	

Polybrominated diphenyl ethers (PBDEs)

BDE 15 (4,4'-dibromo-)	BDE 138 (2,2',3,4,4',5'-hexabromo-)
BDE 17 (2,2',4-tribromo-)	BDE 153 (2,2',4,4',5,5'-hexabromo-)
BDE 25 (2,3',4-tribromo-)	BDE 154 (2,2',4,4',5,6'-hexabromo-)
BDE 28 (2,4,4'-tribromo-)	BDE 155 (2,2',4,4',6,6'-hexabromo-)
BDE 30 (2,4,6-tribromo-)	BDE 156 (2,3,3',4,4',5-hexabromo-)
BDE 33 (2',3,4-tribromo-)	BDE 181 (2,2',3,4,4',5,6-heptabromo-)
BDE 47 (2,2',4,4'-tetrabromo-)	BDE 183 (2,2',3,4,4',5',6-heptabromo-)
BDE 49 (2,2',4,5'-tetrabromo-)	BDE 190 (2,3,3',4,4',5,6-heptabromo-)
BDE 66 (2,3',4,4'-tetrabromo-)	BDE 191 (2,3,3',4,4',5,6'-heptabromo-)
BDE 71 (2,3',4',6-tetrabromo-)	BDE 196 (2,2',3,3',4,4',5,6'-octabromo-)
BDE 75 (2,4,4',6-tetrabromo-)	BDE 197 (2,2',3,3',4,4',6,6'-octabromo-)
BDE 85 (2,2',3,4,4'-pentabromo-)	BDE 203 (2,2',3,4,4',5,5',6-octabromo-)
BDE 99 (2,2',4,4',5-pentabromo-)	BDE 205 (2,3,3',4,4',5,5',6-octabromo-)
BDE 100 (2,2',4,4',6-pentabromo-)	BDE 206 (2,2',3,3',4,4',5,6,6'-nonabromo-)
BDE 116 (2,3,4,5,6-pentabromo-)	BDE 207 (2,2'3,3',4,4',5,6,6'-nonabromo-)
BDE 118 (2,3',4,4',5-pentabromo-)	BDE 208 (2,2',3,3',4,5,5',6,6'-nonabromo-)
BDE 119 (2,3',4,4',6-pentabromo-)	BDE 209 (decabromo-)

Table 2. Mussel TissueXII (QA05TIS12): Laboratory means of three replicates and exercise assigned values - TEO and PAHs
(reported as if three figures were significant)

Laboratory No.	1a	1c	3	4	5	6	7	8	9	10	11	12	Value	s	%RSD
TEO (percent)	NA	4.13	NA	NA	1.93	15.0	0.295	1.68	4.30	1.69	1.45	NA	3.81	4.73	124.1

PAHs (ng/g dry mass)	Laboratory No.	1a	1c	3	4	5	6	7	8	9	10	11	12	Value	s	%RSD
naphthalene		9.87	8.18	8.76	5.61	12.5	43.5	14.2	21.6	2.52	103	NA	NA	9.86	6.07	61.5
2-methylnaphthalene		4.77	2.05	3.19	3.82	<10	14.1	13.3	9.88	1.10	NA	NA	NA	8.00	5.16	64.5
1-methylnaphthalene		4.47	0.763	4.28	1.81	<8	6.96	4.15	NA	0.692	NA	NA	NA	3.66	2.44	66.6
biphenyl		2.87	2.02	2.11	1.81	<7	4.40	<0.7	NA	0.745	38.0	NA	NA	1.91	0.77	40.0
2,6-dimethylnaphthalene		coelution	<5	5.55	1.78	<8	8.24	<0.8	NA	3.10	NA	NA	<3	4.67	2.85	61.0
acenaphthylene		<2	8.45	2.93	0.491	<11	3.64	3.75	<12.5	1.94	3.02	NA	5.50	3.72	2.40	64.6
acenaphthene		2.39	1.44	6.48	0.896	<8	<0.78	3.42	<12.5	3.67	43.5	NA	<2	2.93	2.21	75.4
1,6,7-trimethylnaphthalene		coelution	6.44	NA	4.21	NA	5.72	NA	NA	2.74	NA	NA	coelution	4.78	1.65	34.5
fluorene		3.90	5.57	3.54	2.44	<7	4.46	6.50	<12.5	1.25	5.95	NA	3.40	3.64	1.64	45.1
phenanthrene		101	105	78.5	107	88.5	82.8	92.1	93.9	48.3	115	NA	96.5	88.7	17.9	20.2
anthracene		4.75	9.72	7.78	2.87	<10	4.32	2.60	<12.5	3.39	39.0	NA	10.9	5.79	3.24	55.9
1-methylphenanthrene		111	113	78.5	122	104	73.9	79.6	NA	54.0	43.4	NA	118	89.7	27.7	30.9
fluoranthene		165	150	119	172	129	127	119	146	69.7	253	NA	144	133	30	22.9
pyrene		213	210	177	241	179	209	172	210	100	352	NA	207	190	40	21.1
benz[a]anthracene		35.9	29.6	19.8	31.5	23.8	21.4	22.6	25.9	11.9	NA	NA	26.1	24.7	7.1	28.6
chrysene		coelution	48.0	81.3	coelution	94.5	106	46.9	110	48.9	coelution	NA	coelution	63.9	22.4	35.1
triphenylene		coelution	49.2	NA	NA	NA	NA	NA	NA	NA	coelution	NA	coelution	no target		
benzo[b]fluoranthene		53.5	coelution	56.7	52.0	52.9	39.0	56.9	64.6	22.5	162	NA	coelution	47.6	12.6	26.5
benzo[j]fluoranthene		13.4	coelution	NA	NA	NA	NA	NA	NA	16.8	NA	NA	coelution	no target		
benzo[k]fluoranthene		17.3	12.7	17.8	coelution	31.3	33.1	17.5	37.3	NA	43.1	NA	30.7	16.3	2.4	15.0
benzo[e]pyrene		82.5	85.8	68.9	86.6	77.1	75.3	68.7	66.6	41.7	166	NA	94.0	74.7	14.6	19.6
benzo[a]pyrene		8.66	12.2	4.66	7.26	8.86	7.72	6.55	<12.5	2.94	144	NA	6.36	7.25	2.65	36.6
perylene		4.26	4.74	<14.0	3.57	<5	17.1	NA	NA	1.47	26.2	NA	NA	3.51	1.44	41.2
indeno[1,2,3-cd]pyrene		12.2	coelution	11.0	15.7	19.5	12.0	15.7	26.4	5.15	23.3	NA	14.8	15.1	6.4	42.4
dibenz[a,h]anthracene		coelution	coelution	3.35	coelution	<11	<0.78	<0.7	<12.5	2.08	5.31	NA	coelution	no target		
benzo[ghi]perylene		23.3	25.9	19.3	28.5	23.9	23.1	34.9	43.1	11.5	21.9	NA	28.8	24.6	6.4	25.8

Note: Bolded values were not used in the calculation of the exercise assigned value; NA = not analyzed

Table 3. Mussel Tissue XII (QA05TISI12): Laboratory means of three replicates and exercise assigned values - Pesticides
(reported as if three figures were significant)

Laboratory No.	1a	1c	3	4	5	6	7	8	9	10	11	12	Exercise Assigned	
													Value	s
alpha-HCH (a-BHC)	<2	<1	NA	<1.05	<2.0	<1.8	<2.5	NA	<0.565	1.87	NA	0.247	no target	
hexachlorobenzene	<2	<1	1.51	<0.996	<2.5	<1.8	<2.5	NA	<0.355	NA	NA	0.131	no target	
gamma-HCH (g-BHC,lindane)	<2	<1	<2.72	4.08	<1.5	<1.8	<2.5	NA	<0.355	0.079	NA	0.094	no target	
beta-HCH (b-BHC)	<2	<1	NA	1.37	NA	2.50	<2.5	NA	<0.301	0.351	NA	<0.04	no target	
heptachlor	<2	<5	<2.72	<1.00	<2.0	<1.8	<2.5	NA	<0.419	2.36	NA	0.226	no target	
aldrin	<2	<1	<2.72	<1.00	1.63	<1.8	<4	NA	<0.428	0.658	NA	<0.03	no target	
heptachlor epoxide	<2	<1	<2.72	<1.00	<2.0	<1.8	<5	NA	<0.465	0.774	NA	0.297	no target	
oxychlordane	<2	<1	40.8	4.15	NA	<1.8	<5	NA	<0.52	2.89	NA	0.41057	no target	
gamma-chlordane	8.47	<1	7.22	8.43	5.80	8.30	6.59	NA	7.97	6.32	NA	7.99	7.45	1.00
2,4'-DDE	<2	<1	<2.72	1.43	<1.0	<1.8	<0.5	NA	<1.28	8.42	NA	1.15	no target	
endosulfan I	<2	<1	<2.72	<2.63	<1.5	24.3	<5	NA	<1.35	0.112	NA	<0.2	no target	
cis-chlordane (alpha-chlordane)	16.8	17.8	9.39	11.0	8.43	11.0	12.1	NA	5.98	5.40	NA	10.5	12.1	3.4
trans-nonachlor	9.43	9.22	7.94	9.92	9.17	10.0	6.40	NA	5.47	5.10	NA	9.84	9.00	1.24
dieldrin	<5	<5	<2.72	3.51	7.23	14.2	9.09	NA	3.28	1.37	NA	2.94	6.70	4.42
4,4'-DDE	41.1	46.9	27.6	34.7	23.0	30.5	25.9	NA	18.4	29.6	NA	30.8	33.9	7.6
2,4'-DDD	5.61	10.8	7.11	14.4	7.70	8.12	6.05	NA	5.69	coelution	NA	6.93	8.04	2.86
endrin	<2	<1	<2.72	NA	<2.0	<1.8	<4	NA	<0.502	coelution	NA	<0.07	no target	
endosulfan II	<2	<1	<2.72	NA	<3.4	<1.8	<5	NA	<0.392	0.130	NA	<0.4	no target	
4,4'-DDD	16.4	27.6	20.6	42.8	18.3	<1.8	14.8	NA	14.5	15.1	NA	30.0	21.7	10.8
2,4'-DDT	<2	<1	<2.72	1.85	<3.0	<1.8	<2.5	NA	<0.547	7.59	NA	0.664	no target	
cis-nonachlor	5.31	4.59	3.11	4.33	NA	3.84	3.86	NA	3.72	5.47	NA	4.22	4.27	0.76
4,4'-DDT	<2	<1	<2.72	1.22	2.50	10.2	<2	NA	1.15	4.58	NA	1.85	1.68	0.63
mirex	<2	<1	<2.72	3.89	<1.5	<1.8	<1.5	NA	<0.31	NA	NA	0.421	no target	
endosulfan sulfate	<2	<1	NA	NA	NA	<1.8	<4	NA	<0.775	NA	NA	<0.2	no target	
chlorpyrifos	<2	NA	NA	NA	NA	<1.8	NA	NA	<0.401	NA	NA	NA	no target	

Note: Bolded values were not used in the calculation of the exercise assigned value; NA = not analyzed

Table 4. Mussel Tissue XII (QA05TISI2): Laboratory means of three replicates and exercise assigned values - PCBs
(reported as if three figures were significant)

Laboratory No.	1a	1c	3	4	5	6	7	8	9	10	11	12	Exercise Assigned	
													Value	s
PCB 8	2.22	2.76	1.74	NA	<2.8	0.300	2.15	3.54	1.53	11.9	1.81	2.14	2.56	0.60
PCB 18	8.38	8.78	4.29	7.84	4.00	0.897	5.46	4.40	4.05	6.76	4.93	5.03	5.71	1.87
PCB 28	30.0	27.6	19.8	24.3	20.4	2.81	21.0	24.8	14.2	coelution	23.1	26.9	23.2	4.6
PCB 31	22.1	26.2	NA	21.0	NA	2.52	16.7	21.3	NA	coelution	16.6	22.9	21.7	3.1
PCB 44	38.8	37.3	21.9	33.8	20.2	3.73	24.1	23.4	16.8	19.1	45.3	50.4	31.4	11.2
PCB 49	49.7	52.3	NA	47.1	NA	5.53	27.9	41.5	23.9	21.4	34.4	36.5	37.2	11.3
PCB 52	57.0	66.6	38.7	55.1	37.7	6.32	42.0	40.7	31.6	22.3	45.0	51.5	46.6	10.7
PCB 66	52.8	57.4	43.9	51.5	74.0	6.82	33.3	39.1	30.7	50.7	42.0	57.1	48.4	12.4
PCB 95	59.4	64.7	NA	54.4	NA	4.68	36.9	47.0	NA	16.4	46.4	51.4	51.5	9.2
PCB 99	47.6	43.7	NA	48.0	NA	5.51	32.9	49.1	NA	14.9	55.5	52.1	47.0	7.2
PCB 101	97.7	113	82.9	85.5	98.5	13.1	70.0	66.3	52.2	26.1	89.1	90.4	88.1	14.4
PCB 105	34.9	36.5	25.6	34.3	37.1	4.39	25.6	31.8	21.3	9.80	31.6	32.7	31.1	5.3
PCB 118	86.3	79.7	84.7	93.3	101	12.0	51.1	79.7	54.1	28.5	79.6	89.4	79.9	15.9
PCB 128	14.5	14.7	15.0	15.8	13.3	2.24	9.21	11.0	8.36	4.57	11.5	12.0	13.0	2.2
PCB 138	coelution	61.7	68.8	coelution	91.8	11.3	57.8	coelution	50.8	54.0	coelution	coelution	64.1	14.9
PCB 149	66.4	80.7	NA	57.6	NA	6.28	47.6	57.5	NA	26.9	50.4	59.3	61.5	11.2
PCB 153	coelution	147	72.3	123	86.3	11.4	52.1	60.6	63.9	71.9	82.6	97.3	85.7	29.6
PCB 156	6.80	5.19	NA	6.17	NA	0.713	4.07	4.03	NA	NA	6.04	7.17	5.64	1.25
PCB 170	1.82	1.46	1.38	1.69	2.20	0.390	<2	1.98	<0.729	2.53	1.63	1.50	1.80	0.38
PCB 180	8.23	7.10	6.88	10.3	6.90	1.00	9.28	6.64	15.7	12.7	coelution	coelution	9.29	3.11
PCB 187	19.9	19.3	15.9	21.9	17.9	2.84	15.2	14.0	12.6	14.2	24.9	22.0	18.0	4.0
PCB 194	<2	0.593	NA	0.535	NA	<0.5	<2	0.406	NA	0.553	0.307	0.417	0.501	0.084
PCB 195	<2	<1	<2.72	<1.01	<1.8	<0.5	<2	<0.400	<0.265	0.687	0.050	0.0686	no target	
PCB 206	<2	<1	<2.72	<1.02	<1.7	<0.5	<2.2	<0.400	<0.256	0.220	0.040	0.050	no target	
PCB 209	<2	<1	<2.72	<1.01	<1.7	<0.5	<2.2	<0.400	<0.419	0.111	0.054	0.060	no target	

Note: Bolded values were not used in the calculation of the exercise assigned value; NA = not analyzed

Table 5. Mussel Tissue XII (QA05TISI12): Laboratory means of three replicates and exercise assigned values - PBDEs
(reported as if three figures were significant)
ng/g dry mass

	1a	1c	3	4	5	6	7	8	9	10	11	12	Exercise Assigned		
													Value	s	%RSD
BDE 15	NA	<1	NA	NA	NA	NA	NA	<2.00	NA	NA	NA	0.160	no target		
BDE 17	NA	2.98	NA	NA	NA	NA	NA	4.26	NA	NA	NA	3.92	3.72	0.66	17.8
BDE 25	NA	<1	NA	NA	NA	NA	NA	4.26	NA	NA	NA	other	no target		
BDE 28	NA	5.47	NA	3.76	NA	NA	1.38	2.39	NA	NA	NA	2.40	3.08	1.58	51.4
BDE 30	NA	<1	NA	NA	NA	NA	NA	<2.00	NA	NA	NA	<0.03	no target		
BDE 33	NA	<1	NA	NA	NA	NA	NA	2.39	NA	NA	NA	other	no target		
BDE 47	NA	23.5	NA	28.9	NA	NA	14.4	24.5	NA	NA	NA	25.2	23.3	5.4	23.1
BDE 49	NA	<1	NA	8.99	NA	NA	NA	6.11	NA	NA	NA	5.15	6.75	2.00	29.6
BDE 66	NA	<1	NA	1.54	NA	NA	0.566	0.895	NA	NA	NA	0.941	0.984	0.403	41.0
BDE 71	NA	3.23	NA	NA	NA	NA	NA	<2.00	NA	NA	NA	NA	no target		
BDE 75	NA	<1	NA	NA	NA	NA	NA	<2.00	NA	NA	NA	0.102	no target		
BDE 85	NA	<1	NA	<2.40	NA	NA	0.333	0.473	NA	NA	NA	0.447	0.418	0.074	17.8
BDE 99	NA	inf	NA	16.2	NA	NA	8.377	11.0	NA	NA	NA	10.230	11.5	3.4	29.3
BDE 100	NA	7.26	NA	9.55	NA	NA	4.128	6.55	NA	NA	NA	6.749	6.85	1.93	28.3
BDE 116	NA	<1	NA	NA	NA	NA	NA	<2.00	NA	NA	NA	<0.04	no target		
BDE 118	NA	<1	NA	NA	NA	NA	NA	<2.00	NA	NA	NA	NA	no target		
BDE 119	NA	<1	NA	NA	NA	NA	NA	<2.00	NA	NA	NA	0.123	no target		
BDE 138	NA	<1	NA	NA	NA	NA	0.049	<2.00	NA	NA	NA	0.101	no target		
BDE 153	NA	0.382	NA	<2.39	NA	NA	0.601	0.5705	NA	NA	NA	0.508	0.515	0.097	18.8
BDE 154	NA	0.399	NA	<2.41	NA	NA	0.621	0.610	NA	NA	NA	0.570	0.550	0.103	18.7
BDE 155	NA	1.37	NA	NA	NA	NA	NA	<2.00	NA	NA	NA	0.232	no target		
BDE 156	NA	<1	NA	NA	NA	NA	NA	NA	NA	NA	NA	NA	no target		
BDE 181	NA	<10	NA	NA	NA	NA	NA	<2.00	NA	NA	NA	<0.024	no target		
BDE 183	NA	<10	NA	<2.42	NA	NA	NA	<0.200	NA	NA	NA	0.062	no target		
BDE 190	NA	<10	NA	NA	NA	NA	NA	<2.00	NA	NA	NA	<0.035	no target		
BDE 191	NA	NA	NA	NA	NA	NA	NA	NA	NA	NA	NA	NA	no target		
BDE 196	NA	NA	NA	NA	NA	NA	NA	NA	NA	NA	NA	NA	no target		
BDE 197	NA	NA	NA	NA	NA	NA	NA	NA	NA	NA	NA	NA	no target		
BDE 203	NA	NA	NA	NA	NA	NA	NA	<2.00	NA	NA	NA	0.065	no target		
BDE 205	NA	NA	NA	NA	NA	NA	NA	<2.00	NA	NA	NA	NA	no target		
BDE 206	NA	NA	NA	NA	NA	NA	NA	<2.00	NA	NA	NA	<0.3	no target		
BDE 207	NA	NA	NA	NA	NA	NA	NA	<2.00	NA	NA	NA	<0.5	no target		
BDE 208	NA	NA	NA	NA	NA	NA	NA	<2.00	NA	NA	NA	<0.4	no target		
BDE 209	NA	NA	NA	NA	NA	NA	NA	<100	NA	NA	NA	<5	no target		

Note: Bolded values were not used in the calculation of the exercise assigned value; NA = not analyzed

Table 6. SRM 2977: Laboratory means of three replicates and target values - TEO and PAHs
(reported as if three figures were significant)

Laboratory No.		1a	1c	3	4	5	6	7	8	9	10	11	12
TEO (percent)		NA	6.30	NA	NA	7.23	40.40	3.47	2.78	7.50	NA	1.31	NA
PAHs (ng/g dry mass)													
Laboratory No.	1a	1c	3	4	5	6	7	8	9	10	11	12	Certificate Values
naphthalene	22.2	20.1	11.8	9.23	15.1	44.2	7.68	25.5	9.90	112	NA	NA	19 5 Reference
2-methylnaphthalene	16.3	17.6	6.43	11.5	11.0	14.2	8.59	17.4	6.76	NA	NA	NA	18 5 Reference
1-methylnaphthalene	16.0	14.9	7.82	7.05	<8	9.15	4.23	NA	5.81	NA	NA	NA	16 5 Reference
biphenyl	6.81	4.48	3.39	3.81	<7	0.630	<0.2	NA	3.27	163	NA	NA	6.8 0.6 Reference
2,6-dimethylnaphthalene	coelution	17.6	11.1	17.8	14.6	17.1	12.8	NA	14.3	NA	NA	18.4	no target Target
acenaphthylene	<2	<5	1.58	<0.765	<11	4.44	1.38	<6.25	3.00	3.39	NA	2.06	no target Target
acenaphthene	4.06	6.30	6.65	3.06	<8	<0.78	4.29	10.1	11.8	181	NA	3.76	4.2 0.4 Reference
1,6,7-trimethylnaphthalene	coelution	43.5	NA	33.1	NA	32.0	NA	NA	26.5	NA	NA	81.6	no target Target
fluorene	10.2	22.7	7.02	10.5	8.43	8.37	7.15	16.9	7.86	18.4	NA	10.1	10.24 0.43 Certified
phenanthrene	34.4	39.1	29.6	44.0	31.6	33.6	34.9	70.8	26.3	59.1	NA	39.5	35.1 3.8 Certified
anthracene	6.18	9.91	3.46	1.79	<10	9.14	2.32	7.77	4.74	26.7	NA	5.18	8 4 Reference
1-methylphenanthrene	39.8	37.4	27.7	57.8	42.1	33.2	31.6	NA	34.0	39.9	NA	47.9	44 2 Reference
fluoranthene	38.7	39.3	25.2	46.3	30.3	35.4	38.0	70.2	27.2	95.3	NA	38.2	38.7 1.0 Certified
pyrene	78.7	75.3	52.5	95.8	57.5	82.6	59.5	143	53.5	170	NA	77.4	78.9 3.5 Certified
benz[a]anthracene	21.0	19.2	14.3	25.4	17.8	16.1	17.1	38.1	13.0	NA	NA	22.7	20.34 0.78 Certified
chrysene	coelution	38.9	58.7	86.6	67.0	78.6	41.4	172	54.2	coelution	NA	95.6	49 2 Reference
triphenylene	coelution	35.1	NA	NA	NA	NA	NA	NA	NA	coelution	NA	other	39 1 Reference
benzo[b]fluoranthene	11.4	coelution	37.4	14.1	14.1	10.9	14.7	39.4	10.2	88.7	NA	13.2	11.01 0.28 Certified
benzo[j]fluoranthene	4.53	coelution	NA	NA	NA	NA	NA	NA	8.33	NA	NA	other	4.6 0.2 Reference
benzo[k]fluoranthene	4.77	3.32	3.36	12.4	7.30	10.0	4.49	24.7	NA	27.4	NA	9.39	4 1 Reference
benzo[e]pyrene	13.0	13.9	12.2	18.5	15.0	16.0	13.4	<6.25	14.3	61.6	NA	19.5	13.1 1.1 Certified
benzo[a]pyrene	8.71	10.8	5.04	6.61	7.13	6.06	5.01	<6.25	4.83	37.9	NA	10.5	8.35 0.72 Certified
perylene	3.56	4.02	<12.7	3.01	<5	3.47	NA	NA	2.02	44.0	NA	NA	3.50 0.76 Certified
indeno[1,2,3-cd]pyrene	4.83	coelution	3.34	4.38	8.90	4.44	4.73	<6.25	4.22	7.01	NA	5.27	4.84 0.81 Certified
dibenz[a,h]anthracene	coelution	coelution	1.15	1.93	<11	4.45	<0.2	<6.25	1.95	3.28	NA	1.84	1.41 0.19 Certified
benzo[ghi]perylene	9.68	9.14	5.44	10.9	<15	9.67	12.1	25.1	9.59	5.99	NA	10.1	9.53 0.43 Certified

Table 7. SRM 2977: Laboratory means of three replicates and target values - Pesticides
(reported as if three figures were significant)
ng/g dry mass

Laboratory No.	1a	1c	3	4	5	6	7	8	9	10	11	12	Certificate Values	
													conc.	95%CL type
alpha-HCH (a-BHC)	<2	<1	NA	<0.719	<2.0	<1.8	<1	NA	<0.565	0.203	NA	SRM 2978	no target	Target
hexachlorobenzene	<2	<1	1.72	<0.684	<2.5	<1.8	<1	NA	<0.355	NA	NA	SRM 2978	no target	Target
gamma-HCH (g-BHC,lindane)	<2	<1	<2.49	1.68	<1.5	<1.8	<1	NA	<0.355	0.381	NA	SRM 2978	no target	Target
beta-HCH (b-BHC)	<2	<1	NA	7.97	NA	9.94	<1	NA	<0.301	7.22	NA	SRM 2978	no target	Target
heptachlor	<2	<5	<2.49	<0.687	<2.0	<1.8	<1	NA	<0.419	1.10	NA	SRM 2978	no target	Target
aldrin	<2	<1	<2.49	<0.689	<1.5	<1.8	<1.5	NA	<0.428	0.872	NA	SRM 2978	no target	Target
heptachlor epoxide	<2	<1	<2.49	<0.687	<2.0	<1.8	<2	NA	<0.465	0.168	NA	SRM 2978	no target	Target
oxychlorodane	<2	<1	<2.49	4.90	NA	<1.8	<2	NA	<0.52	0.789	NA	SRM 2978	no target	Target
gamma-chlordane	2.20	<1	1.32	1.54	2.29	<1.8	<0.5	NA	<0.427	2.88	NA	SRM 2978	no target	Target
2,4'-DDE	<2	<1	<2.49	0.723	<1.0	<1.8	<0.2	NA	<1.3	1.04	NA	SRM 2978	no target	Target
endosulfan I	<2	<1	<2.49	<1.81	<1.5	2.45	<2	NA	<1.38	0.121	NA	SRM 2978	no target	Target
cis-chlordane (alpha-chlordane)	1.34	<2	0.648	0.673	1.10	<1.8	1.34	NA	<0.4	0.313	NA	SRM 2978	1.42	0.13 Certified
trans-nonachlor	1.23	<2	0.563	<0.690	<1.5	2.89	1.07	NA	0.744	0.365	NA	SRM 2978	1.43	0.10 Certified
dieldrin	<5	<5	<2.49	5.82	4.87	8.11	5.39	NA	4.20	2.05	NA	SRM 2978	6.04	0.52 Certified
4,4'-DDE	12.2	12.1	6.71	10.7	4.97	9.17	10.8	NA	6.73	6.66	NA	SRM 2978	12.5	1.6 Certified
2,4'-DDD	3.18	4.30	2.47	interference	<4.0	3.70	2.94	NA	2.48	coelution	NA	SRM 2978	3.32	0.29 Certified
endrin	<2	<1	<2.49	NA	<2.0	<1.8	<1.5	NA	<0.512	1/2,4'-DD	NA	SRM 2978	no target	Target
endosulfan II	<2	<1	<2.49	NA	<3.4	<1.8	<2	NA	<0.4	0.163	NA	SRM 2978	no target	Target
4,4'-DDD	4.22	11.8	1.25	interference	2.43	3.25	4.56	NA	2.61	2.53	NA	SRM 2978	4.30	0.38 Certified
2,4'-DDT	<2	<1	<2.49	<0.681	<3.0	<1.8	<1	NA	<0.558	13.8	NA	SRM 2978	no target	Target
cis-nonachlor	0.422	<2	<2.49	<0.703	NA	<1.8	<0.8	NA	<0.362	1.04	NA	SRM 2978	no target	Target
4,4'-DDT	<2	<1	<2.49	<0.683	<2.5	3.74	<0.7	NA	1.18	13.8	NA	SRM 2978	1.28	0.18 Certified
mirex	<2	<1	<2.49	2.08	<1.5	<1.8	<0.5	NA	<0.316	NA	NA	SRM 2978	no target	Target
endosulfan sulfate	<2	<1	NA	NA	NA	<1.8	<1.5	NA	<0.79	NA	NA	SRM 2978	no target	Target
chlorpyrifos	<2	NA	NA	NA	NA	<1.8	NA	NA	<0.408	NA	NA	SRM 2978	no target	Target

NA = not analyzed

Table 8. SRM 2977: Laboratory means of three replicates and target values - PCBs
(reported as if three figures were significant)

ng/g dry mass

Laboratory No.	1a	1c	3	4	5	6	7	8	9	10	11	12	Certificate Values	
													conc.	95%CL type
PCB 8	2.05	2.05	0.961	NA	<2.8	1.33	0.844	1.83	3.38	3.29	0.536	0.973	2.10	0.15 Certified
PCB 18	2.74	2.43	1.16	2.64	<2.7	5.16	1.70	1.25	2.30	1.35	0.731	1.21	2.65	0.30 Certified
PCB 28	5.28	5.53	4.75	6.82	4.00	5.60	4.24	6.76	4.85	coelution	4.36	6.30	5.37	0.44 Certified
PCB 31	4.05	3.77	NA	4.30	NA	3.61	3.30	3.97	NA	w/ PCB 28	1.83	3.39	3.92	0.24 Certified
PCB 44	3.26	3.15	1.95	2.44	<2.4	2.70	3.35	2.05	1.37	1.68	3.17	4.57	3.25	0.63 Certified
PCB 49	2.59	1.16	NA	2.13	NA	2.90	1.97	2.62	<0.371	3.01	0.889	1.41	no target	Target
PCB 52	8.33	8.67	4.98	8.86	4.50	8.10	7.79	8.20	7.27	3.76	4.60	6.99	8.37	0.54 Certified
PCB 66	3.57	3.44	5.87	4.15	2.65	4.86	3.51	3.79	3.64	4.26	2.32	3.73	3.64	0.32 Certified
PCB 95	5.61	6.01	NA	6.01	NA	4.03	4.58	6.69	NA	1.87	3.81	4.53	5.39	0.59 Certified
PCB 99	1.66	3.09	NA	4.81	NA	4.89	1.53	6.47	NA	1.73	4.66	4.18	no target	Target
PCB 101	11.7	11.3	7.09	10.2	6.27	11.8	9.46	9.46	<0.4	3.50	8.60	8.77	11.2	1.2 Certified
PCB 105	3.42	3.33	2.19	4.68	2.53	3.73	3.03	4.10	2.37	1.57	2.78	3.17	3.76	0.49 Certified
PCB 118	10.8	9.04	13.7	12.7	6.27	11.3	10.3	11.1	9.97	4.34	8.08	9.57	10.5	1.0 Certified
PCB 128	2.60	2.37	2.06	2.45	<1.9	3.52	2.38	2.11	1.60	0.755	1.35	1.63	2.49	0.28 Certified
PCB 138	coelution	7.50	8.40	coelution	7.83	14.3	9.42	coelution	15.5	8.98	coelution	coelution	no target	Target
PCB 149	10.2	9.49	NA	8.11	NA	6.48	8.22	10.0	NA	3.71	5.17	6.47	9.23	0.12 Certified
PCB 153	coelution	15.6	7.40	18.1	6.83	12.8	13.4	10.4	9.71	11.5	10.0	11.5	14.1	1.0 Certified
PCB 156	0.972	0.945	NA	0.772	NA	0.967	0.956	0.591	NA	NA	0.597	0.847	0.960	0.085 Certified
PCB 170	3.03	2.82	1.99	2.87	<1.8	3.04	2.58	2.44	2.13	3.34	2.56	2.40	2.95	0.23 Certified
PCB 180	6.74	7.26	3.62	6.83	3.27	5.46	4.49	5.41	8.10	10.2	coelution	coelution	6.79	0.67 Certified
PCB 187	4.59	4.13	2.63	4.41	2.60	4.89	4.75	4.63	3.22	3.06	4.61	4.05	4.76	0.38 Certified
PCB 194	<2	0.889	NA	0.857	NA	0.990	<0.8	0.884	NA	0.604	0.361	0.680	0.897	0.042 Certified
PCB 195	<2	<1	<2.49	<0.696	<1.8	<0.5	<0.8	<200	<0.265	0.362	0.096	0.159	no target	Target
PCB 206	<2	<1	<2.49	<0.699	<1.7	<0.5	<1	<200	<0.256	0.095	0.033	0.044	no target	Target
PCB 209	<2	<1	<2.49	<0.694	<1.7	<0.5	<1	<200	<0.419	0.033	0.005	<0.010	no target	Target

NA = not analyzed

Table 9. SRM 2977: Laboratory means of three replicates and target values - PBDEs
(reported as if three figures were significant)

ng/g dry mass

Laboratory No.	1a	1c	3	4	5	6	7	8	9	10	11	12	Target Value	
													conc.	std dev
BDE 15	NA	<1	NA	NA	NA	NA	NA	<1.00	NA	NA	NA	0.121	no target	
BDE 17	NA	0.876	NA	NA	NA	NA	NA	1.89	NA	NA	NA	1.26	no target	
BDE 25	NA	<1	NA	NA	NA	NA	NA	1.89	NA	NA	NA	other	no target	
BDE 28	NA	5.58	NA	4.42	NA	NA	1.42	2.78	NA	NA	NA	2.79	no target	
BDE 30	NA	<1	NA	NA	NA	NA	NA	<1.00	NA	NA	NA	<0.024	no target	
BDE 33	NA	<1	NA	NA	NA	NA	NA	2.78	NA	NA	NA	other	no target	
BDE 47	NA	28.1	NA	47.8	NA	NA	18.9	38.2	NA	NA	NA	37.9	no target	
BDE 49	NA	<1	NA	1.95	NA	NA	NA	1.25	NA	NA	NA	1.05	1.02	0.03
BDE 66	NA	<1	NA	<1.65	NA	NA	0.310	0.552	NA	NA	NA	0.564	0.375	0.062
BDE 71	NA	0.780	NA	NA	NA	NA	NA	<1.00	NA	NA	NA	0.135	no target	
BDE 75	NA	<1	NA	NA	NA	NA	NA	<1.00	NA	NA	NA	0.092	0.166	0.013
BDE 85	NA	<1	NA	<1.65	NA	NA	0.075	0.473	NA	NA	NA	0.057	no target	
BDE 99	NA	inf	NA	7.90	NA	NA	3.33	5.10	NA	NA	NA	4.67	4.11	0.4
BDE 100	NA	5.30	NA	3.38	NA	NA	1.33	2.19	NA	NA	NA	2.15	1.06	0.18
BDE 116	NA	<1	NA	NA	NA	NA	NA	<1.00	NA	NA	NA	< 0.072	no target	
BDE 118	NA	<1	NA	NA	NA	NA	NA	<1.00	NA	NA	NA	NA	no target	
BDE 119	NA	<1	NA	NA	NA	NA	NA	<1.00	NA	NA	NA	0.054	no target	
BDE 138	NA	<1	NA	NA	NA	NA	0.030	<1.00	NA	NA	NA	0.059	no target	
BDE 153	NA	<LOD	NA	<1.64	NA	NA	0.181	0.143	NA	NA	NA	0.147	no target	
BDE 154	NA	<LOD	NA	<1.66	NA	NA	0.161	0.165	NA	NA	NA	0.163	no target	
BDE 155	NA	<LOD	NA	NA	NA	NA	NA	<1.00	NA	NA	NA	0.032	no target	
BDE 156	NA	<1	NA	NA	NA	NA	NA	NA	NA	NA	NA	NA	no target	
BDE 181	NA	<10	NA	NA	NA	NA	NA	<1.00	NA	NA	NA	0.026	no target	
BDE 183	NA	<10	NA	<1.67	NA	NA	NA	<0.100	NA	NA	NA	0.057	no target	
BDE 190	NA	<10	NA	NA	NA	NA	NA	<1.00	NA	NA	NA	0.028	no target	
BDE 191	NA	NA	NA	NA	NA	NA	NA	NA	NA	NA	NA	NA	no target	
BDE 196	NA	NA	NA	NA	NA	NA	NA	NA	NA	NA	NA	NA	no target	
BDE 197	NA	NA	NA	NA	NA	NA	NA	NA	NA	NA	NA	NA	no target	
BDE 203	NA	NA	NA	NA	NA	NA	NA	<1.00	NA	NA	NA	0.036	no target	
BDE 205	NA	NA	NA	NA	NA	NA	NA	<1.00	NA	NA	NA	NA	no target	
BDE 206	NA	NA	NA	NA	NA	NA	NA	<1.00	NA	NA	NA	<0.3	no target	
BDE 207	NA	NA	NA	NA	NA	NA	NA	1.43	NA	NA	NA	<0.5	no target	
BDE 208	NA	NA	NA	NA	NA	NA	NA	<1.00	NA	NA	NA	<0.4	no target	
BDE 209	NA	NA	NA	NA	NA	NA	NA	<50.0	NA	NA	NA	<5	no target	

NA = not analyzed

Table 10. Marine Sediment XIII (QA05SED13): Laboratory means of three replicates and exercise assigned values - Water, TOC, and PAHs
(reported as if three figures were significant)

Laboratory No.	1a	1c	2	3	4	6	7	8	10	11	12	Value	s	%RSD
Water (percent)	45.8	46.8	49.6	46.2	46.5	NA	52.1	53.3	46.0	44.8	43.1	47.44	3.23	6.8
TOC (percent)	NA	NA	NA	2.83	NA	1.84	NA	3.07	3.04	NA	NA	2.70	0.58	21.5

Laboratory No.	1a	1c	2	3	4	6	7	8	10	11	12	Value	s	%RSD
naphthalene	864	604	614	174	928	557	994	1253	261	NA	937	785	186	23.7
2-methyl/naphthalene	244	136	157	36.3	323	174	255	438	NA	NA	244	219	66	30.1
1-methyl/naphthalene	106	76.8	80.2	65.2	126	92.2	106	NA	NA	NA	133	98.2	24.0	24.4
biphenyl	77.1	45.1	33.1	24.0	84.2	46.0	83.6	NA	27.6	NA	91.5	65.8	23.6	35.8
2,6-dimethylnaphthalene	coelution	79.0	22.0	22.3	131	91.9	114	NA	NA	NA	109	81.3	43.6	53.7
acenaphthylene	55.6	73.2	22.6	28.8	47.7	40.1	57.3	79.1	61.3	NA	47.0	45.1	13.8	30.6
acenaphthene	36.6	inf	18.7	16.0	38.5	24.6	29.7	47.7	35.3	NA	32.1	28.9	8.4	29.0
1,6,7-trimethylnaphthalene	coelution	19.7	NA	NA	37.3	51.2	NA	NA	NA	NA	coelution	no target		
fluorene	66.0	52.0	46.3	29.0	67.4	52.3	56.6	114	83.4	NA	51.4	56.1	12.2	21.7
phenanthrene	332	289	178	209	452	248	425	509	298	NA	326	306	89	29.1
anthracene	167	99.4	78.4	78.7	204	98.0	178	212	189	NA	140	137	47	34.7
1-methyl/phenanthrene	58.0	51.1	28.7	37.2	85.6	79.2	54.0	NA	43.4	NA	54.7	55.4	18.2	32.8
fluoranthene	559	455	273	370	714	389	609	669	605	NA	489	496	140	28.2
pyrene	479	391	225	301	601	259	513	588	601	NA	424	421	142	33.6
benz[a]anthracene	277	235	126	207	371	155	308	345	NA	NA	247	241	80	33.2
chrysene	coelution	189	134	248	coelution	225	299	498	coelution	coelution	coelution	219	62	28.4
triphenylene	coelution	84.0	NA	NA	NA	NA	NA	NA	w/ chrysen	NA	other	no target		
benzo[b]fluoranthene	441	coelution	197	511	489	210	661	579	630	NA	coelution	413	174	42.1
benzo[j]fluoranthene	216	coelution	NA	NA	NA	NA	NA	NA	NA	NA	coelution	no target		
benzo[k]fluoranthene	228	130	91	144	coelution	203	228	443	167	NA	250	180	56	31.1
benzo[e]pyrene	326	234	121	235	353	181	341	327	399	NA	388	286	97	34.0
benzo[a]pyrene	362	451	113	196	317	144	351	344	441	NA	245	282	120	42.7
perylene	354	327	111	171	486	219	NA	NA	384	NA	314	311	131	42.2
indeno[1,2,3-cd]pyrene	332	coelution	94.6	231	372	133	340	331	316	NA	243	258	101	39.3
dibenz[a,h]anthracene	coelution	coelution	11.1	49.0	coelution	47.8	59.7	97.2	96.7	NA	NA	41.9	21.2	50.6
benzo[ghi]perylene	316	196	73.5	136	321	154	328	301	329	NA	234	244	96	39.5

Note: Bolded values were not used in the calculation of the exercise assigned values; NA = not analyzed

Table 11. Marine Sediment XIII (QA05SED13): Laboratory means of three replicates and exercise assigned values - Pesticides
(reported as if three figures were significant)

Laboratory No.	Exercise Assigned											
	1a	1c	2	3	4	6	7	8	10	11	12	Value
alpha-HCH (a-BHC)	<3	<1	NA	NA	<0.571	<0.15	<1	NA	0.555	NA	0.031	no target
hexachlorobenzene	5.16	4.86	NA	4.12	7.01	2.40	7.48	NA	NA	NA	6.11	5.38
gamma-HCH (g-BHC, lindane)	<3	<1	NA	<0.736	16.4	<0.15	<1	NA	0.775	NA	0.030	no target
beta-HCH (b-BHC)	<3	<1	NA	NA	<0.566	<0.15	<1	NA	0.911	NA	0.019	no target
heptachlor	<3	<1	NA	<0.736	<0.545	<0.15	<1	NA	4.26	NA	<0.006	no target
aldrin	<3	<1	NA	<0.736	<0.547	<0.15	<1.5	NA	0.247	NA	0.013	no target
heptachlor epoxide	<3	<1	NA	<0.736	<0.545	<0.15	<2	NA	0.671	NA	0.036	no target
oxychlordan	<3	<1	NA	<0.736	3.99	<0.15	<2	NA	0.137	NA	<0.01	no target
gamma-chlordane	0.589	<1	NA	0.658	0.583	0.341	<0.5	NA	0.775	NA	0.486	0.572
2,4'-DDE	0.328	<2	NA	<0.736	0.499	<0.15	<0.2	NA	1.70	NA	0.314	0.380
endosulfan I	<3	<1	NA	<0.736	<0.548	<0.15	<2	NA	0.088	NA	<0.032	no target
cis-chlordane (alpha-chlordan)	0.542	<1	NA	0.451	0.54833	0.441	<0.25	NA	2.20	NA	0.430	0.482
trans-nonachlor	0.326	<1	NA	0.198	<0.548	0.383	0.19677	NA	0.382	NA	0.228	0.286
dieldrin	<3	<1	NA	<0.736	<0.544	0.424	<1	NA	0.360	NA	0.375	0.386
4,4'-DDE	3.38	3.61	NA	3.85	6.57	1.46	4.11	NA	4.88	NA	2.82	3.44
2,4'-DDD	<3	0.954	NA	<0.736	1.80	0.484	<0.5	NA	coelution	NA	0.471	0.927
endrin	<3	<1	NA	<0.736	NA	<0.15	<1.5	NA	coelution	NA	<0.016	no target
endosulfan II	<3	<1	NA	<0.736	NA	<0.15	<2	NA	0.162	NA	<0.059	no target
4,4'-DDD	4.46	4.07	NA	4.88	NA	1.86	14.8	NA	2.76	NA	5.64	4.18
2,4'-DDT	<3	<2	NA	<0.736	0.512	<0.15	<1	NA	3.10	NA	0.097	no target
cis-nonachlor	<3	<1	NA	0.172	<0.558	<0.15	3.86	NA	1.02	NA	0.167	0.454
4,4'-DDT	<3	<2	NA	<0.736	0.504	0.725	<0.7	NA	5.95	NA	0.381	0.537
mirex	<3	<2	NA	<0.736	4.63	1.28	<0.7	NA	NA	NA	0.015	no target
endosulfan sulfate	<3	<1	NA	NA	NA	<0.15	<0.5	NA	NA	NA	<0.018	no target
chlorpyrifos	<3	NA	NA	NA	NA	<0.15	<1.5	NA	NA	NA	NA	no target

Note: Bolded values were not used in the calculation of the exercise assigned values; NA = not analyzed

Table 12. Marine Sediment XIII (QA05SED13): Laboratory means of three replicates and exercise assigned values - PCBs
(reported as if three figures were significant)

Laboratory No.	1a	1c	2	3	4	6	7	8	10	11	12	Exercise Assigned		
												Value	s	%RSD
PCB 8	1.48	1.76	NA	1.18	NA	0.820	1.90	3.52	7.91	0.874	1.53	1.36	0.42	30.7
PCB 18	2.22	2.43	NA	1.49	2.55	1.25	2.46	1.99	2.40	0.447	1.61	2.04	0.48	23.6
PCB 28	3.76	4.71	NA	3.24	4.69	2.24	4.61	6.07	coelution	2.69	4.37	3.79	0.97	25.6
PCB 31	3.34	3.65	NA	NA	3.43	1.69	3.35	4.48	coelution	1.28	3.07	2.83	0.94	33.2
PCB 44	3.65	3.85	NA	2.70	4.46	2.06	3.95	4.88	4.50	1.93	4.17	3.47	0.99	28.6
PCB 49	4.00	4.51	NA	NA	4.63	2.45	4.65	6.81	4.38	1.35	3.16	3.64	1.22	33.4
PCB 52	5.38	5.45	NA	3.75	6.16	2.95	5.88	6.70	4.15	2.04	4.59	4.48	1.39	31.1
PCB 66	4.97	5.42	NA	4.01	5.50	2.96	4.96	6.11	9.44	1.55	5.20	4.32	1.41	32.5
PCB 95	3.98	3.70	NA	NA	5.36	1.76	4.40	5.52	2.91	1.73	3.67	3.44	1.25	36.5
PCB 99	2.91	3.50	NA	NA	3.22	1.55	3.11	4.08	1.76	1.15	2.40	2.45	0.87	35.6
PCB 101	5.15	6.30	NA	4.93	5.98	3.83	6.11	5.23	3.25	1.97	4.29	4.86	1.36	28.0
PCB 105	1.48	1.53	NA	1.04	1.92	0.893	1.50	1.77	1.17	0.578	1.21	1.26	0.40	31.7
PCB 118	4.19	4.21	NA	3.33	5.57	2.48	4.76	4.89	2.30	1.69	3.77	3.59	1.26	35.0
PCB 128	0.663	0.662	NA	0.787	0.914	0.660	1.25	0.743	0.389	0.256	0.516	0.684	0.275	40.3
PCB 138	coelution	4.77	NA	3.69	coelution	2.86	3.55	coelution	5.86	2.08	3.96	3.83	1.23	32.2
PCB 149	4.39	5.54	NA	NA	5.55	2.37	5.88	6.43	3.69	2.12	3.94	4.18	1.44	34.4
PCB 153	coelution	coelution	NA	3.79	8.58	2.89	5.84	3.99	8.59	2.11	4.14	4.99	2.46	49.4
PCB 156	0.517	0.565	NA	NA	0.554	0.390	0.443	0.448	NA	0.165	0.473	0.444	0.128	28.7
PCB 170	1.37	1.40	NA	1.24	1.79	0.877	1.08	1.41	3.05	0.522	1.05	1.19	0.36	30.5
PCB 180	3.23	3.06	NA	2.44	4.41	1.52	3.11	3.08	5.04	1.23	2.61	2.97	1.15	38.8
PCB 187	2.15	2.13	NA	1.51	2.54	3.36	2.69	1.80	2.89	1.05	1.84	2.20	0.69	31.3
PCB 194	<2	1.74	NA	NA	1.18	0.567	<0.8	1.18	1.31	0.230	0.863	1.01	0.50	49.7
PCB 195	<2	0.350	NA	0.227	0.576	0.245	<0.8	<0.428	4.49	0.076	0.272	0.291	0.166	57.0
PCB 206	2.38	2.68	NA	1.64	2.53	1.49	1.90	3.55	4.57	0.638	2.03	1.91	0.66	34.7
PCB 209	4.63	5.49	NA	3.14	5.23	NA	4.35	6.50	2.21	1.19	4.11	4.02	1.47	36.5

Note: Bolded values were not used in the calculation of the exercise assigned values; NA = not analyzed

Table 13. Marine Sediment XIII (QA05SED13): Laboratory means of three replicates and exercise assigned values - PBDEs
(reported as if three figures were significant)

ng/g dry mass

Laboratory No.	1a	1c	2	3	4	6	7	8	10	11	12	Exercise Assigned		
												Value	s	%RSD
BDE 15	NA	<1	NA	NA	NA	NA	NA	<2.14	NA	NA	0.185	no target		
BDE 17	NA	<1	NA	NA	NA	NA	NA	<2.14	NA	NA	0.099	no target		
BDE 25	NA	<1	NA	NA	NA	NA	NA	<2.14	NA	NA	other	no target		
BDE 28	NA	<1	NA	NA	<2.1	NA	0.053	<0.214	NA	NA	0.059	0.056	0.005	8.2
BDE 30	NA	<1	NA	NA	NA	NA	NA	<2.14	NA	NA	<0.002	no target		
BDE 33	NA	<1	NA	NA	NA	NA	NA	<2.14	NA	NA	other	no target		
BDE 47	NA	1.21	NA	NA	<2.1	NA	1.255	<10.7	NA	NA	0.635	1.03	0.35	33.4
BDE 49	NA	<1	NA	NA	<2.1	NA	NA	<2.14	NA	NA	0.154	no target		
BDE 66	NA	<1	NA	NA	<2.1	NA	0.044	<0.214	NA	NA	0.032	0.038	0.009	22.7
BDE 71	NA	<1	NA	NA	NA	NA	NA	<2.14	NA	NA	0.014	no target		
BDE 75	NA	<1	NA	NA	NA	NA	NA	<2.14	NA	NA	<0.002	no target		
BDE 85	NA	<1	NA	NA	<2.1	NA	0.066	<0.257	NA	NA	0.012	no target		
BDE 99	NA	<1	NA	NA	<2.1	NA	1.352	<10.7	NA	NA	0.452	no target		
BDE 100	NA	<1	NA	NA	<2.1	NA	0.294	<2.14	NA	NA	0.109	no target		
BDE 116	NA	<1	NA	NA	NA	NA	NA	<2.14	NA	NA	<0.014	no target		
BDE 118	NA	<1	NA	NA	NA	NA	NA	<2.14	NA	NA	NA	no target		
BDE 119	NA	<1	NA	NA	NA	NA	NA	<2.14	NA	NA	0.004	no target		
BDE 138	NA	<1	NA	NA	NA	NA	0.034	<2.14	NA	NA	0.008	no target		
BDE 153	NA	<1	NA	NA	<2.1	NA	0.177	<0.257	NA	NA	0.064	no target		
BDE 154	NA	<1	NA	NA	<2.1	NA	0.137	<0.214	NA	NA	0.069	no target		
BDE 155	NA	<1	NA	NA	NA	NA	NA	<2.14	NA	NA	0.010	no target		
BDE 156	NA	<1	NA	NA	NA	NA	NA	NA	NA	NA	NA	no target		
BDE 181	NA	<10	NA	NA	NA	NA	NA	<2.14	NA	NA	<0.141	no target		
BDE 183	NA	<10	NA	NA	<2.1	NA	NA	<0.214	NA	NA	<0.3	no target		
BDE 190	NA	<10	NA	NA	NA	NA	NA	<2.14	NA	NA	<0.206	no target		
BDE 191	NA	NA	NA	NA	NA	NA	NA	NA	NA	NA	NA	no target		
BDE 196	NA	NA	NA	NA	NA	NA	NA	NA	NA	NA	NA	no target		
BDE 197	NA	NA	NA	NA	NA	NA	NA	NA	NA	NA	NA	no target		
BDE 203	NA	NA	NA	NA	NA	NA	NA	<2.14	NA	NA	0.142	no target		
BDE 205	NA	NA	NA	NA	NA	NA	NA	<2.14	NA	NA	NA	no target		
BDE 206	NA	NA	NA	NA	NA	NA	NA	<2.14	NA	NA	NA	no target		
BDE 207	NA	NA	NA	NA	NA	NA	NA	<2.14	NA	NA	NA	no target		
BDE 208	NA	NA	NA	NA	NA	NA	NA	<2.14	NA	NA	NA	no target		
BDE 209	NA	NA	NA	NA	NA	NA	NA	NA	NA	NA	25.7	no target		

Note: Bolded values were not used in the calculation of the exercise assigned values; NA = not analyzed

Table 14. SRM 1941b: Laboratory means of three replicates and target values - Water, TOC, and PAHs
(reported as if three figures were significant)

Laboratory No.	1a	1c	2	3	4	6	7	8	10	11	12
Water (percent)	2.46	2.86	SRM 1944	NA	1.83	NA	92.70	36.20	NA	NA	NA
TOC (percent)	NA	NA	SRM 1944	3.03	NA	0.94	NA	2.88	NA	NA	NA

PAHs (ng/g dry mass)

Laboratory No.	1a	1c	2	3	4	6	7	8	10	11	12	conc.	95%CL	type
naphthalene	866	971	SRM 1944	226	877	941	933	1680	170	NA	939	848	95	Certified
2-methylnaphthalene	267	219	SRM 1944	40.9	314	308	244	445	NA	NA	266	276	53	Reference
1-methylnaphthalene	136	111	SRM 1944	75.8	119	158	96.6	NA	NA	NA	138	127	14	Reference
biphenyl	75.8	72.5	SRM 1944	23.6	82.7	80.0	76.3	NA	26.2	NA	95.0	74.0	8.0	Reference
2,6-dimethylnaphthalene	coelution	128.2	SRM 1944	23.4	126	163	108	NA	NA	NA	122	75.9	4.5	Reference
acenaphthylene	54.4	117	SRM 1944	30.4	48.8	74.2	61.5	168	54.8	NA	55.9	53.3	6.4	Reference
acenaphthene	33.9	inf	SRM 1944	16.0	40.7	44.7	31.3	132	31.3	NA	36.9	38.4	5.2	Reference
1,6,7-trimethylnaphthalene	coelution	31.0	SRM 1944	NA	37.2	83.3	NA	NA	NA	NA	coelution	25.5	5.1	Reference
fluorene	87.1	73.6	SRM 1944	28.1	71.0	91.3	66.1	225	82.1	NA	67.8	85	15	Certified
phenanthrene	427	466	SRM 1944	207	452	469	409	743	229	NA	389	406	44	Certified
anthracene	179	179	SRM 1944	75.6	208	179	184	345	157	NA	165	184	18	Certified
1-methylphenanthrene	71.7	77.4	SRM 1944	36.8	87.8	135	50.1	NA	35.3	NA	63.4	73.2	5.9	Certified
fluoranthene	656	716	SRM 1944	392	726	747	565	1047	477	NA	615	651	50	Certified
pyrene	571	596	SRM 1944	307	607	480	493	924	508	NA	501	581	39	Certified
benz[a]anthracene	373	358	SRM 1944	221	370	280	296	562	NA	NA	296	335	25	Certified
chrysene	coelution	278	SRM 1944	267	coelution	407	303	748	coelution	NA	coelution	291	31	Certified
triphenylene	coelution	104	SRM 1944	NA	NA	NA	NA	NA	w/ chrysen	NA	other	108	5	Certified
benzo[b]fluoranthene	445	coelution	SRM 1944	561	493	380	649	935	505	NA	coelution	453	21	Certified
benzo[k]fluoranthene	218	coelution	SRM 1944	NA	NA	NA	NA	NA	NA	NA	other	217	5	Reference
benzo[e]pyrene	226	210	SRM 1944	168	coelution	367	233	690	123	NA	307	225	18	Certified
benzo[a]pyrene	337	373	SRM 1944	256	362	328	328	547	331	NA	440	325	25	Certified
perylene	367	686	SRM 1944	197	303	241	329	535	276	NA	291	358	17	Certified
indeno[1,2,3-cd]pyrene	386	650	SRM 1944	148	453	305	NA	NA	259	NA	375	397	45	Certified
dibenzo[a,h]anthracene	348	coelution	SRM 1944	259	369	222	350	562	178	NA	322	341	57	Certified
benzo[ghi]perylene	coelution	coelution	SRM 1944	53.4	coelution	80.9	57.2	113	95.7	NA	NA	53	10	Certified
	312	353	SRM 1944	154	305	207	329	525	293	NA	308	307	45	Certified

NA = not analyzed

Table 15. SRM 1941b: Laboratory means of three replicates and target values - Pesticides
(reported as if three figures were significant)

Laboratory No.	1a	1c	2	3	4	6	7	8	10	11	12	Certificate Values	
												conc.	95%CL type
alpha-HCH (a-BHC)	<3	<1	NA	NA	<0.679	<0.15	<0.5	NA	0.438	NA	0.027	no target	Target
hexachlorobenzene	5.94	10.3	NA	2.88	7.84	6.33	5.94	NA	NA	NA	8.59	5.83	0.38 Certified
gamma-HCH (g-BHC,lindane)	<3	<1	NA	<0.816	15.8	<0.15	<0.5	NA	NA	NA	0.024	no target	Target
beta-HCH (b-BHC)	<3	<1	NA	NA	<0.673	<0.15	<0.5	NA	0.323	NA	0.019	no target	Target
heptachlor	<3	<1	NA	<0.816	<0.648	<0.15	<0.5	NA	3.70	NA	<0.006	no target	Target
aldrin	<3	<1	NA	<0.816	<0.650	<0.15	<0.8	NA	0.122	NA	0.011	no target	Target
heptachlor epoxide	<3	<1	NA	<0.816	<0.648	<0.15	<1	NA	0.421	NA	0.045	no target	Target
oxychlordan	<3	<1	NA	<0.816	3.85	<0.15	<1	NA	0.096	NA	<0.025	no target	Target
gamma-chlordane	0.688	<1	NA	0.488	0.586	0.674	0.543	NA	0.701	NA	0.514	0.566	0.093 Certified
2,4'-DDE	0.331	<2	NA	<0.816	0.526	<0.15	0.368	NA	1.48	NA	0.342	0.38	0.12 Reference
endosulfan I	<3	<1	NA	<0.816	0.833	<0.15	<1	NA	0.061	NA	<0.029	no target	Target
cis-chlordane (alpha-chlordan)	0.846	<1	NA	0.379	0.511	0.529	0.779	NA	1.84	NA	0.461	0.85	0.11 Certified
trans-nonachlor	0.436	<1	NA	0.191	<0.651	0.547	0.180	NA	0.291	NA	0.243	0.438	0.073 Certified
dieldrin	<3	<1	NA	<0.816	<0.647	<0.15	<0.5	NA	0.286	NA	0.403	no target	Target
4,4'-DDE	3.30	4.22	NA	3.35	5.72	2.83	2.52	NA	3.94	NA	3.16	3.22	0.28 Certified
2,4'-DDD	<3	1.57	NA	<0.816	1.88	<0.15	<0.3	NA	coelution	NA	0.506	no target	Target
endrin	<3	<1	NA	<0.816	NA	<0.15	<0.8	NA	2,4'-DD	NA	<0.014	no target	Target
endosulfan II	<3	<1	NA	<0.816	NA	<0.15	<1	NA	0.044	NA	<0.022	no target	Target
4,4'-DDD	4.55	4.97	NA	3.53	NA	<0.15	4.58	NA	1.88	NA	4.74	4.66	0.46 Certified
2,4'-DDT	<3	<2	NA	<0.816	0.575	<0.15	<0.5	NA	1.10	NA	0.089	no target	Target
cis-nonachlor	<3	0.756	NA	0.139	<0.664	<0.15	<0.3	NA	0.846	NA	0.183	0.378	0.053 Certified
4,4'-DDT	<3	<2	NA	<0.816	<0.645	<0.15	0.743	NA	0.990	NA	0.278	1.12	0.42 Reference
mirex	<3	<2	NA	<0.816	4.69	<0.15	<0.3	NA	NA	NA	<0.02	no target	Target
endosulfan sulfate	<3	<1	NA	NA	NA	<0.15	<0.8	NA	NA	NA	<0.023	no target	Target
chlorpyrifos	<3	NA	NA	NA	NA	<0.15	NA	NA	NA	NA	NA	no target	Target

NA = not analyzed

Table 16. SRM 1941b: Laboratory means of three replicates and target values - PCBs
(reported as if three figures were significant)

Laboratory No.	ng/g dry mass												Certificate Values	
	1a	1c	2	3	4	6	7	8	10	11	12	conc.	95%CL	type
PCB 8	1.56	1.25	NA	0.982	NA	1.27	1.75	4.32	4.78	1.17	1.56	1.65	0.19	Certified
PCB 18	2.31	1.90	NA	1.25	2.47	1.96	2.13	2.54	1.97	1.61	1.59	2.39	0.29	Certified
PCB 28	4.50	3.16	NA	2.93	4.72	3.65	4.01	8.17	coelution	4.72	4.56	4.52	0.57	Certified
PCB 31	3.22	2.73	NA	NA	3.30	2.59	2.87	5.48	coelution	2.34	3.09	3.18	0.41	Certified
PCB 44	3.69	3.10	NA	2.48	4.28	3.24	3.33	5.89	3.84	5.15	5.61	3.85	0.20	Certified
PCB 49	4.38	3.72	NA	NA	4.45	3.94	4.44	8.23	3.94	2.87	3.47	4.34	0.28	Certified
PCB 52	5.40	4.92	NA	3.24	5.68	4.48	4.98	7.78	3.60	4.34	4.86	5.24	0.28	Certified
PCB 66	4.97	4.35	NA	3.65	5.40	4.78	4.44	7.29	7.80	4.00	6.04	4.96	0.53	Certified
PCB 95	3.99	3.14	NA	NA	4.81	2.61	3.67	6.62	2.37	4.23	3.91	3.93	0.62	Certified
PCB 99	3.05	2.80	NA	NA	3.03	2.43	2.65	5.15	1.52	3.73	2.74	2.90	0.36	Certified
PCB 101	5.21	5.01	NA	4.19	5.43	5.87	5.30	6.46	2.69	5.91	5.29	5.11	0.34	Certified
PCB 105	1.42	1.17	NA	0.950	1.80	1.34	1.36	2.34	0.957	1.35	1.33	1.43	0.10	Certified
PCB 118	4.10	3.19	NA	2.90	5.17	3.84	3.77	6.33	2.20	3.94	4.08	4.23	0.19	Certified
PCB 128	0.669	0.452	NA	0.749	0.811	1.05	0.697	0.986	0.403	0.528	0.595	0.696	0.044	Certified
PCB 138	coelution	4.24	NA	3.26	coelution	4.47	3.41	coelution	4.99	4.74	4.36	3.6	0.28	Certified
PCB 149	4.42	4.22	NA	NA	4.90	3.59	5.25	8.37	3.32	4.38	4.30	4.35	0.26	Certified
PCB 153	coelution	coelution	NA	3.37	7.49	4.40	5.00	5.50	7.02	5.03	4.37	5.47	0.32	Certified
PCB 156	0.520	<LOD	NA	NA	0.485	0.533	0.402	0.575	NA	0.394	0.528	0.507	0.090	Certified
PCB 170	1.42	1.01	NA	1.07	1.44	1.41	1.44	1.81	2.76	1.33	1.24	1.35	0.09	Certified
PCB 180	3.37	2.62	NA	2.13	3.38	2.39	2.98	4.02	4.43	3.21	2.98	3.24	0.51	Certified
PCB 187	2.17	1.55	NA	1.34	2.07	2.04	2.09	3.20	2.16	2.58	2.04	2.17	0.22	Certified
PCB 194	<2	1.10	NA	NA	0.858	0.887	0.895	1.57	1.14	0.669	1.07	1.04	0.06	Certified
PCB 195	<2	<LOD	NA	0.207	<0.657	<05	0.409	<0.627	3.83	0.213	0.348	0.645	0.060	Certified
PCB 206	2.55	1.83	NA	1.56	2.44	<05	2.12	4.79	3.79	2.27	2.47	2.42	0.19	Certified
PCB 209	4.80	3.44	NA	3.08	4.95	NA	4.29	9.06	1.90	4.72	5.27	4.86	0.45	Certified

NA = not analyzed

Table 17. SRM 1941b: Laboratory means of three replicates and target values - PBDEs
(reported as if three figures were significant)

Laboratory No.	ng/g dry mass												Target Values	
	1a	1c	2	3	4	6	7	8	10	11	12	conc.	std dev	type
BDE 15	NA	<1	NA	NA	NA	NA	NA	<3.13	NA	NA	0.211	no target		Target
BDE 17	NA	<1	NA	NA	NA	NA	NA	<3.13	NA	NA	0.161	no target		Target
BDE 25	NA	<1	NA	NA	NA	NA	NA	<3.13	NA	NA	other	0.18	0.07	Target
BDE 28	NA	<1	NA	NA	<2.5	NA	0.106	<0.313	NA	NA	0.170	no target		Target
BDE 30	NA	<1	NA	NA	NA	NA	NA	<3.13	NA	NA	<0.002	no target		Target
BDE 33	NA	<1	NA	NA	NA	NA	NA	<3.13	NA	NA	other	w/ BDE 28		Target
BDE 47	NA	0.255	NA	NA	<2.5	NA	1.59	<15.7	NA	NA	1.61	1.48	0.51	Target
BDE 49	NA	<1	NA	NA	<2.5	NA	NA	<3.13	NA	NA	0.190	no target		Target
BDE 66	NA	<1	NA	NA	<2.5	NA	0.038	<0.313	NA	NA	0.045	no target		Target
BDE 71	NA	<1	NA	NA	NA	NA	NA	<3.13	NA	NA	0.017	no target		Target
BDE 75	NA	<1	NA	NA	NA	NA	NA	<3.13	NA	NA	0.003	no target		Target
BDE 85	NA	<1	NA	NA	<2.4	NA	0.070	<0.376	NA	NA	0.014	no target		Target
BDE 99	NA	<1	NA	NA	<2.4	NA	1.54	<15.7	NA	NA	0.575	0.62	0.19	Target
BDE 100	NA	<1	NA	NA	<2.4	NA	0.322	<3.13	NA	NA	0.146	0.15	0.06	Target
BDE 116	NA	<1	NA	NA	NA	NA	NA	<3.13	NA	NA	<0.003	no target		Target
BDE 118	NA	<1	NA	NA	NA	NA	NA	<3.13	NA	NA	NA	no target		Target
BDE 119	NA	<1	NA	NA	NA	NA	NA	<3.13	NA	NA	0.003	no target		Target
BDE 138	NA	<1	NA	NA	NA	NA	0.028	<3.13	NA	NA	0.009	no target		Target
BDE 153	NA	<1	NA	NA	<2.4	NA	0.200	<0.376	NA	NA	0.079	0.09	0.04	Target
BDE 154	NA	<1	NA	NA	<2.5	NA	0.158	<0.313	NA	NA	0.080	0.09	0.02	Target
BDE 155	NA	<1	NA	NA	NA	NA	NA	<3.13	NA	NA	0.011	no target		Target
BDE 156	NA	<1	NA	NA	NA	NA	NA	NA	NA	NA	NA	no target		Target
BDE 181	NA	<10	NA	NA	NA	NA	NA	<3.13	NA	NA	<0.055	no target		Target
BDE 183	NA	<10	NA	NA	<2.5	NA	NA	<0.313	NA	NA	<0.3	0.05	0.02	Target
BDE 190	NA	<10	NA	NA	NA	NA	NA	<3.13	NA	NA	<0.08	no target		Target
BDE 191	NA	NA	NA	NA	NA	NA	NA	NA	NA	NA	NA	no target		Target
BDE 196	NA	NA	NA	NA	NA	NA	NA	NA	NA	NA	NA	no target		Target
BDE 197	NA	NA	NA	NA	NA	NA	NA	NA	NA	NA	NA	no target		Target
BDE 203	NA	NA	NA	NA	NA	NA	NA	<3.13	NA	NA	0.157	no target		Target
BDE 205	NA	NA	NA	NA	NA	NA	NA	<3.13	NA	NA	NA	no target		Target
BDE 206	NA	NA	NA	NA	NA	NA	NA	<3.13	NA	NA	NA	no target		Target
BDE 207	NA	NA	NA	NA	NA	NA	NA	<3.13	NA	NA	NA	no target		Target
BDE 208	NA	NA	NA	NA	NA	NA	NA	<3.13	NA	NA	NA	no target		Target
BDE 209	NA	NA	NA	NA	NA	NA	NA	<157	NA	NA	27.6	24.1	15.0	Target

NA = not analyzed

Table 18. Mussel TissueXII (QA05TIS12): z scores (25%) by laboratory - TEO and PAHs

(z=+1 is 25% higher than the exercise assigned value; z=-1 is 25% lower than the exercise assigned value.)

Laboratory No.	1a	1c	3	4	5	6	7	8	9	10	11	12
TEO (percent)		0.3			-2.0	11.8	-3.7	-2.2	0.5	-2.2	-2.5	

PAHs

Laboratory No.	1a	1c	3	4	5	6	7	8	9	10	11	12
naphthalene	0.0	-0.7	-0.4	-1.7	1.1	13.7	1.7	4.8	-3.0	37.9		
2-methylnaphthalene	-1.6	-3.0	-2.4	-2.1		3.1	2.7	0.9	-3.4			
1-methylnaphthalene	0.9	-3.2	0.7	-2.0		3.6	0.5		-3.2			
biphenyl	2.0	0.2	0.4	-0.2		5.2			-2.4	75.5		
2,6-dimethylnaphthalene			0.8	-2.5		3.1			-1.3			
acenaphthylene		5.1	-0.8	-3.5		-0.1	0.0		-1.9	-0.8		1.9
acenaphthene	-0.7	-2.0	4.9	-2.8			0.7		1.0	55.4		
1,6,7-trimethylnaphthalene		1.4		-0.5		0.8			-1.7			
fluorene	0.3	2.1	-0.1	-1.3		0.9	3.1		-2.6	2.5		-0.3
phenanthrene	0.5	0.7	-0.5	0.8	0.0	-0.3	0.2	0.2	-1.8	1.2		0.3
anthracene	-0.7	2.7	1.4	-2.0		-1.0	-2.2		-1.7	22.9		3.5
1-methylphenanthrene	1.0	1.0	-0.5	1.4	0.6	-0.7	-0.4		-1.6	-2.1		1.3
fluoranthene	1.0	0.5	-0.4	1.2	-0.1	-0.2	-0.4	0.4	-1.9	3.6		0.3
pyrene	0.5	0.4	-0.3	1.1	-0.2	0.4	-0.4	0.4	-1.9	3.4		0.4
benz[a]anthracene	1.8	0.8	-0.8	1.1	-0.1	-0.5	-0.3	0.2	-2.1			0.2
chrysene		-1.0	1.1		1.9	2.6	-1.1	2.9	-0.9			
triphenylene												
benzo[b]fluoranthene	0.5		0.8	0.4	0.4	-0.7	0.8	1.4	-2.1	9.6		
benzo[j]fluoranthene												
benzo[k]fluoranthene	0.2	-0.9	0.4		3.7	4.1	0.3	5.1		6.6		3.5
benzo[e]pyrene	0.4	0.6	-0.3	0.6	0.1	0.0	-0.3	-0.4	-1.8	4.9		1.0
benzo[a]pyrene	0.8	2.8	-1.4	0.0	0.9	0.3	-0.4		-2.4	75.7		-0.5
perylene	0.9	1.4		0.1		15.5			-2.3	25.9		
indeno[1,2,3-cd]pyrene	-0.8		-1.1	0.2	1.1	-0.8	0.1	3.0	-2.6	2.2		-0.1
dibenz[a,h]anthracene												
benzo[ghi]perylene	-0.2	0.2	-0.9	0.6	-0.1	-0.3	1.7	3.0	-2.1	-0.4		0.7

Table 19. Mussel TissueXII (QA05TIS12): z scores (25%) by laboratory - Pesticides

(z=+1 is 25% higher than the exercise assigned value; z=-1 is 25% lower than the exercise assigned value.)

Laboratory No.	1a	1c	3	4	5	6	7	8	9	10	11	12
alpha-HCH (a-BHC)												
hexachlorobenzene												
gamma-HCH (g-BHC,lindane)												
beta-HCH (b-BHC)												
heptachlor												
aldrin												
heptachlor epoxide												
oxychlordane												
gamma-chlordane	0.5		-0.1	0.5	-0.9	0.5	-0.5		0.3	-0.6		0.3
2,4'-DDE												
endosulfan I												
cis-chlordane (alpha-chlordane)	1.5	1.9	-0.9	-0.4	-1.2	-0.4	0.0		-2.0	-2.2		-0.5
trans-nonachlor	0.2	0.1	-0.5	0.4	0.1	0.5	-1.2		-1.6	-1.7		0.4
dieldrin				-1.9	0.3	4.4	1.4		-2.0	-3.2		-2.2
4,4'-DDE	0.8	1.5	-0.7	0.1	-1.3	-0.4	-1.0		-1.8	-0.5		-0.4
2,4'-DDD	-1.2	1.4	-0.5	3.1	-0.2	0.0	-1.0		-1.2			-0.6
endrin												
endosulfan II												
4,4'-DDD	-1.0	1.1	-0.2	3.9	-0.6		-1.3		-1.3	-1.2		1.5
2,4'-DDT												
cis-nonachlor	1.0	0.3	-1.1	0.1		-0.4	-0.4		-0.5	1.1		0.0
4,4'-DDT				-1.1	2.0	20.4			-1.3	6.9		0.4
mirex												
endosulfan sulfate												
chlorpyrifos												

Table 20 Mussel TissueXII (QA05TIS12): z scores (25%) by laboratory - PCBs

(z=+1 is 25% higher than the exercise assigned value; z=-1 is 25% lower than the exercise assigned value.)

Laboratory No.	1a	1c	3	4	5	6	7	8	9	10	11	12
PCB 8	-0.5	0.3	-1.3			-3.5	-0.6	1.5	-1.6	14.5	-1.2	-0.7
PCB 18	1.9	2.1	-1.0	1.5	-1.2	-3.4	-0.2	-0.9	-1.2	0.7	-0.6	-0.5
PCB 28	1.2	0.8	-0.6	0.2	-0.5	-3.5	-0.4	0.3	-1.6		0.0	0.6
PCB 31	0.1	0.8		-0.1		-3.5	-0.9	-0.1			-0.9	0.2
PCB 44	0.9	0.7	-1.2	0.3	-1.4	-3.5	-0.9	-1.0	-1.9	-1.6	1.8	2.4
PCB 49	1.3	1.6		1.1		-3.4	-1.0	0.5	-1.4	-1.7	-0.3	-0.1
PCB 52	0.9	1.7	-0.7	0.7	-0.8	-3.5	-0.4	-0.5	-1.3	-2.1	-0.1	0.4
PCB 66	0.4	0.7	-0.4	0.3	2.1	-3.4	-1.3	-0.8	-1.5	0.2	-0.5	0.7
PCB 95	0.6	1.0		0.2		-3.6	-1.1	-0.3		-2.7	-0.4	0.0
PCB 99	0.0	-0.3		0.1		-3.5	-1.2	0.2		-2.7	0.7	0.4
PCB 101	0.4	1.1	-0.2	-0.1	0.5	-3.4	-0.8	-1.0	-1.6	-2.8	0.0	0.1
PCB 105	0.5	0.7	-0.7	0.4	0.8	-3.4	-0.7	0.1	-1.3	-2.7	0.1	0.2
PCB 118	0.3	0.0	0.2	0.7	1.1	-3.4	-1.4	0.0	-1.3	-2.6	0.0	0.5
PCB 128	0.5	0.5	0.6	0.9	0.1	-3.3	-1.2	-0.6	-1.4	-2.6	-0.5	-0.3
PCB 138		-0.2	0.3		1.7	-3.3	-0.4		-0.8	-0.6		
PCB 149	0.3	1.2		-0.3		-3.6	-0.9	-0.3		-2.3	-0.7	-0.1
PCB 153		2.8	-0.6	1.8	0.0	-3.5	-1.6	-1.2	-1.0	-0.6	-0.1	0.5
PCB 156	0.8	-0.3		0.4		-3.5	-1.1	-1.1			0.3	1.1
PCB 170	0.1	-0.7	-0.9	-0.2	0.9	-3.1		0.4		1.6	-0.4	-0.7
PCB 180	-0.5	-0.9	-1.0	0.4	-1.0	-3.6	0.0	-1.1	2.7	1.4		
PCB 187	0.4	0.3	-0.5	0.9	0.0	-3.4	-0.6	-0.9	-1.2	-0.8	1.5	0.9
PCB 194		0.7		0.3				-0.8		0.4	-1.5	-0.7
PCB 195												
PCB 206												
PCB 209												

Table 21 Mussel TissueXII (QA05TIS12): z scores (25%) by laboratory - PBDEs

(z=+1 is 25% higher than the exercise assigned value; z=-1 is 25% lower than the exercise assigned value.)

	1a	1c	3	4	5	6	7	8	9	10	11	12
BDE 15												
BDE 17		-0.8						0.6				0.2
BDE 25												
BDE 28		3.1		0.9			-2.2	-0.9				-0.9
BDE 30												
BDE 33												
BDE 47		0.0		1.0			-1.5	0.2				0.3
BDE 49				1.3				-0.4				-0.9
BDE 66				2.2			-1.7	-0.4				-0.2
BDE 71												
BDE 75												
BDE 85							-0.8	0.5				0.3
BDE 99				1.7			-1.1	-0.2				-0.4
BDE 100		0.2		1.6			-1.6	-0.2				-0.1
BDE 116												
BDE 118												
BDE 119												
BDE 138												
BDE 153		-1.0					0.7	0.4				-0.1
BDE 154		-1.1					0.5	0.4				0.1
BDE 155												
BDE 156												
BDE 181												
BDE 183												
BDE 190												
BDE 191												
BDE 196												
BDE 197												
BDE 203												
BDE 205												
BDE 206												
BDE 207												
BDE 208												
BDE 209												

Table 22. Marine Sediment XIII (QA05SED13): z scores (25% by laboratory)- water, TOC, and PAHs
(z=+1 is 25% higher than the exercise assigned value; z=-1 is 25% lower than the exercise assigned value.)

Laboratory No.	1a	1c	2	3	4	6	7	8	10	11	12
Water (percent)	-0.1	-0.1	0.2	-0.1	-0.1		0.4	0.5	-0.1	-0.2	-0.4
TOC				0.2		-1.3		0.6	0.5		

PAHs

Laboratory No.	1a	1c	2	3	4	6	7	8	10	11	12
naphthalene	0.4	-0.9	-0.9	-3.1	0.7	-1.2	1.1	2.4	-2.7		0.8
2-methylnaphthalene	0.5	-1.5	-1.1	-3.3	1.9	-0.8	0.7				0.5
1-methylnaphthalene	0.3	-0.9	-0.7	-1.3	1.1	-0.2	0.3				1.4
biphenyl	0.7	-1.3	-2.0	-2.5	1.1	-1.2	1.1		-2.3		1.6
2,6-dimethylnaphthalene		-0.1	-2.9	-2.9	2.4	0.5	1.6				1.4
acenaphthylene	0.9	2.5	-2.0	-1.4	0.2	-0.4	1.1	3.0	1.4		0.2
acenaphthene	1.1		-1.4	-1.8	1.3	-0.6	0.1	2.6	0.9		0.4
1,6,7-trimethylnaphthalene											
fluorene	0.7	-0.3	-0.7	-1.9	0.8	-0.3	0.0	4.1	2.0		-0.3
phenanthrene	0.3	-0.2	-1.7	-1.3	1.9	-0.8	1.6	2.6	-0.1		0.3
anthracene	0.9	-1.1	-1.7	-1.7	2.0	-1.1	1.2	2.2	1.5		0.1
1-methylphenanthrene	0.2	-0.3	-1.9	-1.3	2.2	1.7	-0.1		-0.9		0.0
fluoranthene	0.5	-0.3	-1.8	-1.0	1.8	-0.9	0.9	1.4	0.9		-0.1
pyrene	0.5	-0.3	-1.9	-1.1	1.7	-1.5	0.9	1.6	1.7		0.0
benz[a]anthracene	0.6	-0.1	-1.9	-0.6	2.2	-1.4	1.1	1.7			0.1
chrysene		-0.6	-1.6	0.5		0.1	1.5	5.1			
triphenylene											
benzo[b]fluoranthene	0.3		-2.1	1.0	0.7	-2.0	2.4	1.6	2.1		
benzo[j]fluoranthene											
benzo[k]fluoranthene	1.1	-1.1	-2.0	-0.8		0.5	1.1	5.8	-0.3		1.6
benzo[e]pyrene	0.6	-0.7	-2.3	-0.7	0.9	-1.5	0.8	0.6	1.6		1.4
benzo[a]pyrene	1.1	2.4	-2.4	-1.2	0.5	-2.0	1.0	0.9	2.3		-0.5
perylene	0.5	0.2	-2.6	-1.8	2.2	-1.2			0.9		0.0
indeno[1,2,3-cd]pyrene	1.1		-2.5	-0.4	1.8	-1.9	1.3	1.1	0.9		-0.2
dibenz[a,h]anthracene			-2.9	0.7		0.6	1.7	5.3	5.2		
benzo[ghi]perylene	1.2	-0.8	-2.8	-1.8	1.3	-1.5	1.4	0.9	1.4		-0.2

Table 23. Marine Sediment XIII (QA05SED13): z scores (25% by laboratory)- pesticides
(z=+1 is 25% higher than the exercise assigned value; z=-1 is 25% lower than the exercise assigned value.)

Laboratory No.	1a	1c	2	3	4	6	7	8	10	11	12
alpha-HCH (a-BHC)											
hexachlorobenzene	-0.2	-0.4		-0.9	1.2	-2.2	1.6				0.5
gamma-HCH (g-BHC,lindane)											
beta-HCH (b-BHC)											
heptachlor											
aldrin											
heptachlor epoxide											
oxychlordan											
gamma-chlordane	0.1			0.6	0.1	-1.6			1.4		-0.6
2,4'-DDE	-0.5				1.2				13.9		-0.7
endosulfan I											
cis-chlordane (alpha-chlordane)	0.5			-0.3	0.5	-0.3			14.2		-0.4
trans-nonachlor	0.6			-1.2		1.4	-1.2		1.3		-0.8
dieldrin						0.4			-0.3		-0.1
4,4'-DDE	-0.1	0.2		0.5	3.6	-2.3	0.8		1.7		-0.7
2,4'-DDD		0.1			3.8	-1.9					-2.0
endrin											
endosulfan II											
4,4'-DDD	0.3	-0.1		0.7		-2.2	10.2		-1.4		1.4
2,4'-DDT											
cis-nonachlor				-2.5			30.0		5.0		-2.5
4,4'-DDT					-0.2	1.4			40.3		-1.2
mirex											
endosulfan sulfate											
chlorpyrifos											

Table 24. Marine Sediment XIII (QA05SED13): z scores (25% by laboratory)- PCBs
(z=+1 is 25% higher than the exercise assigned value; z=-1 is 25% lower than the exercise assigned value.)

Laboratory No.	1a	1c	2	3	4	6	7	8	10	11	12
PCB 8	0.3	1.1		-0.5		-1.6	1.6	6.3	19.2	-1.4	0.5
PCB 18	0.3	0.8		-1.1	1.0	-1.6	0.8	-0.1	0.7	-3.1	-0.8
PCB 28	0.0	1.0		-0.6	0.9	-1.6	0.9	2.4		-1.2	0.6
PCB 31	0.7	1.2			0.9	-1.6	0.7	2.3		-2.2	0.3
PCB 44	0.2	0.4		-0.9	1.1	-1.6	0.5	1.6	1.2	-1.8	0.8
PCB 49	0.4	1.0			1.1	-1.3	1.1	3.5	0.8	-2.5	-0.5
PCB 52	0.8	0.9		-0.7	1.5	-1.4	1.2	2.0	-0.3	-2.2	0.1
PCB 66	0.6	1.0		-0.3	1.1	-1.3	0.6	1.7	4.7	-2.6	0.8
PCB 95	0.6	0.3			2.2	-1.9	1.1	2.4	-0.6	-2.0	0.3
PCB 99	0.8	1.7			1.3	-1.5	1.1	2.7	-1.1	-2.1	-0.1
PCB 101	0.2	1.2		0.1	0.9	-0.8	1.0	0.3	-1.3	-2.4	-0.5
PCB 105	0.7	0.8		-0.7	2.1	-1.2	0.8	1.6	-0.3	-2.2	-0.2
PCB 118	0.7	0.7		-0.3	2.2	-1.2	1.3	1.5	-1.4	-2.1	0.2
PCB 128	-0.1	-0.1		0.6	1.3	-0.1	3.3	0.3	-1.7	-2.5	-1.0
PCB 138		1.0		-0.1		-1.0	-0.3		2.1	-1.8	0.1
PCB 149	0.2	1.3			1.3	-1.7	1.6	2.2	-0.5	-2.0	-0.2
PCB 153				-1.0	2.9	-1.7	0.7	-0.8	2.9	-2.3	-0.7
PCB 156	0.7	1.1			1.0	-0.5	0.0	0.0		-2.5	0.3
PCB 170	0.6	0.7		0.2	2.0	-1.1	-0.4	0.7	6.2	-2.3	-0.5
PCB 180	0.3	0.1		-0.7	1.9	-2.0	0.2	0.1	2.8	-2.3	-0.5
PCB 187	-0.1	-0.1		-1.3	0.6	2.1	0.9	-0.7	1.3	-2.1	-0.7
PCB 194		2.9			0.7	-1.8		0.7	1.2	-3.1	-0.6
PCB 195		0.8		-0.9	3.9	-0.6			57.8	-3.0	-0.3
PCB 206	1.0	1.6		-0.6	1.3	-0.9	0.0	3.4	5.6	-2.7	0.3
PCB 209	0.6	1.5		-0.9	1.2		0.3	2.5	-1.8	-2.8	0.1

Table 25. Marine Sediment XIII (QA05SED13): z scores (25% by laboratory)- PBDEs
(z=+1 is 25% higher than the exercise assigned value; z=-1 is 25% lower than the exercise assigned value.)

Laboratory No.	1a	1c	2	3	4	6	7	8	10	11	12
BDE 15											
BDE 17											
BDE 25											
BDE 28							-0.2				0.2
BDE 30											
BDE 33											
BDE 47		0.7					0.9				-1.5
BDE 49											
BDE 66							0.6				-0.6
BDE 71											
BDE 75											
BDE 85											
BDE 99											
BDE 100											
BDE 116											
BDE 118											
BDE 119											
BDE 138											
BDE 153											
BDE 154											
BDE 155											
BDE 156											
BDE 181											
BDE 183											
BDE 190											
BDE 191											
BDE 196											
BDE 197											
BDE 203											
BDE 205											
BDE 206											
BDE 207											
BDE 208											
BDE 209											

Table 26. Mussel Tissue XII (QA05TISI12): RSDs for three replicates - TEO and PAHs													
	Lab 1a		Lab 1c		Lab 3		Lab 4		Lab 5		Lab 6		
	Tissue XII	SRM 2977	Tissue XII	SRM 2977	Tissue XII	SRM 2977	Tissue XII	SRM 2977	Tissue XII	SRM 2977	Tissue XII	SRM 2977	
	rsd	rsd	rsd	rsd	rsd	rsd	rsd	rsd	rsd	rsd	rsd	rsd	
TEO or lipid			7.9%	1.9%					15.8%	4.9%	10.2%	28.6%	
PAH ANALYSES													
	Lab 1a		Lab 1c		Lab 3		Lab 4		Lab 5		Lab 6		
	Tissue XII	SRM 2977	Tissue XII	SRM 2977	Tissue XII	SRM 2977	Tissue XII	SRM 2977	Tissue XII	SRM 2977	Tissue XII	SRM 2977	
	rsd	rsd	rsd	rsd	rsd	rsd	rsd	rsd	rsd	rsd	rsd	rsd	
naphthalene	1.7%	4.3%	7.5%	9.1%	10.3%	24.9%	0.9%	2.3%	23.8%	8.4%	26.1%	17.4%	
2-methylnaphthalene	5.8%	2.2%	70.7%	4.5%	18.3%	8.4%	4.2%	1.3%			11.6%	2.5%	
1-methylnaphthalene	4.8%	3.6%	89.8%	7.2%	12.5%	9.5%	7.9%	0.6%			6.0%	3.3%	
biphenyl	1.3%	2.3%	7.9%	6.8%	17.2%	36.3%	5.1%	25.9%			7.6%	5.7%	
2,6-dimethylnaphthalene				3.6%	7.0%	24.1%	4.7%	0.6%		19.8%	52.8%	4.0%	
acenaphthylene			8.7%		5.4%	8.7%	0.0%				6.0%	14.3%	
acenaphthene	7.6%	1.0%	17.3%	2.6%	17.7%	14.5%	13.6%	4.2%					
1,6,7-trimethylnaphthalene			7.8%	2.5%			3.2%	5.1%			14.3%	24.7%	
fluorene	3.2%	4.2%	7.3%	2.2%	18.1%	3.2%	1.4%	1.1%		7.2%	10.9%	7.6%	
phenanthrene	6.5%	3.3%	0.5%	1.9%	6.5%	15.8%	1.4%	1.0%	10.5%	6.0%	7.7%	20.0%	
anthracene	4.3%	1.5%	5.4%	6.2%	55.4%	25.4%	8.6%	25.8%			18.2%	7.2%	
1-methylphenanthrene	3.5%	2.0%	3.4%	2.2%	2.7%	11.7%	1.4%	1.1%	3.7%	5.0%	10.7%	14.1%	
fluoranthene	1.1%	2.4%	0.6%	2.7%	9.2%	8.3%	2.0%	0.9%	9.2%	4.7%	8.0%	7.3%	
pyrene	1.7%	1.2%	0.2%	3.5%	7.1%	8.8%	2.3%	0.4%	9.2%	5.7%	12.2%	8.3%	
benz[a]anthracene	9.2%	3.8%	3.2%	0.7%	11.4%	4.9%	3.1%	1.9%	11.7%	5.1%	5.8%	0.9%	
chrysene			2.4%	1.7%	7.1%	8.1%	2.0%	0.7%	10.5%	5.5%	7.4%	10.4%	
triphenylene			1.4%	2.4%									
benzo[b]fluoranthene	1.7%	4.4%			8.7%	32.7%	5.7%	0.7%	12.3%	8.9%	4.6%	1.3%	
benzo[j]fluoranthene	6.2%	1.1%											
benzo[k]fluoranthene	4.7%	2.4%	6.9%	4.3%	10.5%	16.7%	2.7%	1.2%	12.7%	3.6%	10.6%	6.0%	
benzo[e]pyrene	3.2%	3.2%	1.8%	2.2%	9.3%	8.6%	2.0%	0.6%	10.0%	4.8%	6.3%	4.2%	
benzo[a]pyrene	1.5%	2.2%	4.4%	8.0%	18.0%	6.4%	13.3%	1.9%	17.0%	4.3%	1.7%	12.8%	
perylene	3.9%	3.9%	4.7%	5.6%			8.0%	2.7%			3.1%	8.8%	
indeno[1,2,3-cd]pyrene	3.5%	1.7%			23.0%	14.5%	6.7%	0.8%	9.8%	4.1%	24.2%	19.5%	
dibenzo[a,h]anthracene					3.9%	38.2%	3.3%	1.5%				5.1%	
benzo[ghi]perylene	4.3%	3.2%	3.6%	0.8%	8.9%	12.3%	1.1%	0.0%	10.0%		13.8%	17.9%	

Table 26 (cont). Mussel Tissue XII (QA05TIS12): RSDs for three replicates - TEO and PAHs													
	Lab 7	SRM 2977	Lab 8	SRM 2977	Lab 9	SRM 2977	Lab 10	SRM 2977	Lab 11	SRM 2977	Lab 12	SRM 2977	SRM 2978
	rds	rds	rds	rds	rds	rds	rds	rds	rds	rds	rds	rds	rds
TEO or lipid	2.4%	6.2%	0.0%	0.0%	16.8%	7.1%	11.9%	0.0%	4.4%	0.0%			
PAH ANALYSES													
	Lab 7	SRM 2977	Lab 8	SRM 2977	Lab 9	SRM 2977	Lab 10	SRM 2977	Lab 11	SRM 2977	Lab 12	SRM 2977	SRM 2978
	rds	rds	rds	rds	rds	rds	rds	rds	rds	rds	rds	rds	rds
naphthalene	38.3%	28.9%	6.5%	19.4%	16.7%	15.0%	18.2%	15.3%					
2-methylnaphthalene	15.8%	18.3%	14.7%	30.6%	10.0%	3.2%							
1-methylnaphthalene	32.6%	11.4%			7.0%	5.2%							
biphenyl					13.2%	4.9%	21.7%	9.0%					
2,6-dimethylnaphthalene		8.8%			19.1%	4.7%						5.8%	
acenaphthylene	54.1%	52.6%			1.3%	9.3%	17.4%	3.7%		25.2%		9.3%	
acenaphthene	62.4%	10.6%		0.0%	26.6%	13.1%	23.1%	10.7%				13.7%	
1,6,7-trimethylnaphthalene					11.6%	38.9%					7.1%	14.1%	
fluorene	17.3%	6.5%		45.7%	19.3%	1.5%	19.2%	2.1%			31.4%	3.7%	
phenanthrene	21.5%	19.8%	22.7%	58.3%	6.1%	3.6%	16.3%	12.2%			0.8%	0.5%	
anthracene	67.9%	12.8%		0.0%	17.6%	21.1%	10.1%	0.4%			22.7%	6.1%	
1-methylphenanthrene	8.2%	18.7%			4.4%	1.5%	29.0%	24.2%			1.0%	14.1%	
fluoranthene	9.5%	29.4%	23.2%	53.4%	6.1%	2.5%	19.9%	8.4%			2.1%	3.4%	
pyrene	8.9%	8.6%	22.2%	53.0%	5.1%	3.8%	19.7%	16.6%			2.0%	2.6%	
benz[a]anthracene	15.1%	13.6%	18.9%	55.7%	5.7%	8.1%					0.9%	4.1%	
chrysene	10.0%	11.4%	19.3%	63.5%	5.6%	2.2%	22.4%	6.1%			2.2%	1.5%	
triphenylene													
benzo[b]fluoranthene	7.6%	12.0%	33.3%	0.0%	8.2%	6.4%	9.4%	11.1%			4.9%	10.2%	
benzo[k]fluoranthene					9.2%	5.3%							
benzo[e]pyrene	22.2%	20.0%	14.4%	71.6%			8.2%	1.6%			5.2%	6.0%	
benzo[a]pyrene	9.4%	10.6%	18.0%		10.1%	4.9%	18.0%	23.2%			5.5%	7.4%	
perylene	34.5%	20.9%			3.0%	9.2%	52.6%	3.0%			6.0%	23.9%	
indeno[1,2,3-cd]pyrene					3.1%	10.4%	24.7%	11.5%					
dibenz[a,h]anthracene	24.2%	33.1%	5.9%		2.6%	4.4%	38.2%	29.4%			3.6%	24.2%	
benzo[ghi]perylene	17.6%	17.3%	3.4%	37.7%	7.7%	3.9%	43.8%	26.5%			3.7%	5.9%	
					4.5%	6.4%	37.8%	12.7%			2.5%	2.5%	

Table 27. Mussel Tissue XII (QA05TISI2): RSDs for three replicates - Pesticides													
	Lab 1a		Lab 1c		Lab 3		Lab 4		Lab 5		Lab 6		
	Tissue XII	SRM 2977	Tissue XII	SRM 2977	Tissue XII	SRM 2977	Tissue XII	SRM 2977	Tissue XII	SRM 2977	Tissue XII	SRM 2977	
	rsd	rsd	rsd	rsd	rsd	rsd	rsd	rsd	rsd	rsd	rsd	rsd	
alpha-HCH (a-BHC)													
hexachlorobenzene					44.7%	24.2%							
gamma-HCH (g-BHC,lindane)							3.7%	12.4%					
beta-HCH (b-BHC)							8.6%	0.6%			35.2%	1.9%	
heptachlor													
aldrin									3.5%				
heptachlor epoxide													
oxychlorane					12.1%		8.8%	8.1%					
gamma-chlordane	1.3%	2.8%			3.0%	5.4%	2.0%	0.6%	6.2%	5.7%	3.9%		
2,4'-DDE							10.6%	6.8%					
endosulfan I											7.7%	7.6%	
cis-chlordane (alpha-chlordane)	2.5%	5.6%	2.5%		8.1%	15.3%	2.1%	10.1%	9.5%	0.0%	14.9%		
trans-nonachlor	2.7%	5.8%	1.0%		9.5%	20.6%	1.6%		8.9%		11.6%	9.4%	
dieldrin							14.3%	0.5%	13.8%	7.8%	9.2%	3.3%	
4,4'-DDE	2.2%	2.9%	0.6%	7.7%	7.9%	15.6%	3.0%	0.0%	2.3%	13.7%	5.1%	11.0%	
2,4'-DDD	5.4%	3.7%	0.5%	5.8%	13.3%	17.0%	4.9%		1.8%		18.1%	13.6%	
endrin													
endosulfan II													
4,4'-DDD	8.4%	3.7%	0.9%	2.5%	9.1%	31.6%	11.9%		5.7%	8.6%		13.3%	
2,4'-DDT							16.0%						
cis-nonachlor	3.5%	4.2%	3.3%		9.8%		4.3%				27.8%		
4,4'-DDT							11.1%		0.0%		14.7%	44.2%	
mirex							7.2%	6.3%					
endosulfan sulfate													
chlorpyrifos													

Table 27 (cont). Mussel Tissue XII (QA05TISI2): RSDs for three replicates - Pesticides													
	Lab 7		Lab 8		Lab 9		Lab 10		Lab 11		Lab 12		
	Tissue XII	SRM 2977	Tissue XII	SRM 2977	Tissue XII	SRM 2977	Tissue XII	SRM 2977	Tissue XII	SRM 2977	Tissue XII	SRM 2977	
	rsd	rsd	rsd	rsd	rsd	rsd	rsd	rsd	rsd	rsd	rsd	rsd	rsd
alpha-HCH (a-BHC)							6.3%	0.7%			13.8%		18.2%
hexachlorobenzene											7.4%		9.4%
gamma-HCH (g-BHC,lindane)							1.9%	44.8%			11.5%		12.3%
beta-HCH (b-BHC)							50.0%	0.6%					11.4%
heptachlor							52.9%	27.8%			21.4%		18.5%
aldrin							37.4%	19.3%					19.1%
heptachlor epoxide							10.2%	29.1%			1.7%		6.9%
oxychlorodane							2.4%	26.2%			4.2%		17.8%
gamma-chlordane	28.5%				22.7%		3.7%	24.0%			4.5%		1.1%
2,4'-DDE							8.7%	25.8%			6.6%		5.3%
endosulfan I							46.8%	40.5%					
cis-chlordane (alpha-chlordane)	22.7%	28.3%			3.3%		4.9%	11.3%			2.5%		0.1%
trans-nonachlor	31.8%	30.3%			3.9%	4.5%	12.3%	8.3%			1.9%		1.8%
dieldrin	5.1%	28.9%			8.6%	7.7%	10.3%	31.8%			3.8%		7.7%
4,4'-DDE	10.5%	11.2%			2.9%	1.8%	6.7%	21.8%			4.1%		1.9%
2,4'-DDD	4.7%	12.4%			8.2%	22.5%	10.8%	32.6%			17.4%		1.2%
endrin													
endosulfan II							100.8%	88.3%					
4,4'-DDD	12.0%	22.9%			6.2%	19.1%	5.8%	19.5%			10.4%		3.6%
2,4'-DDT							42.9%	57.9%			8.1%		10.5%
cis-nonachlor	23.3%				4.8%		6.2%	15.0%			4.5%		0.3%
4,4'-DDT					3.3%	10.3%	24.0%	7.7%			3.4%		1.8%
mirex											4.9%		2.7%
endosulfan sulfate													
chlorpyrifos													

Table 28. Mussel Tissue XII (QA05TISI2): RSDs for three replicates - PCBs													
	Lab 1a		Lab 1c		Lab 3		Lab 4		Lab 5		Lab 6		
	Tissue XII	rsd	SRM 2977	Tissue XII	rsd	SRM 2977	Tissue XII	rsd	SRM 2977	Tissue XII	rsd	SRM 2977	
PCB 8		6.2%	5.0%	4.8%	11.0%	7.1%					11.5%	13.2%	
PCB 18		3.7%	2.3%	3.3%	9.4%	15.5%		1.2%	3.1%	2.5%	11.2%	28.4%	
PCB 28		1.3%	2.3%	1.4%	10.7%	9.8%		2.6%	6.1%	5.9%	2.5%	10.5%	
PCB 31		2.2%	1.8%	1.8%				1.8%	2.0%		7.6%	5.4%	
PCB 44		2.9%	1.4%	0.7%	10.9%	20.0%		1.4%	2.3%	3.1%	4.5%	8.8%	
PCB 49		5.8%	3.4%	6.4%				1.4%	2.3%		5.2%	7.6%	
PCB 52		2.9%	0.7%	0.4%	11.3%	6.9%		1.5%	0.7%	4.6%	4.9%	1.7%	
PCB 66		4.2%	1.6%	2.5%	10.6%	25.2%		2.0%	0.8%	5.4%	6.1%	2.9%	
PCB 95		6.4%	3.3%	1.8%				1.3%	1.8%		6.3%	6.5%	
PCB 99		1.8%	3.7%	1.8%				1.5%	0.1%		4.8%	3.4%	
PCB 101		3.5%	5.0%	0.7%	12.2%	5.0%		1.3%	1.1%	3.0%	6.1%	1.7%	
PCB 105		7.3%	4.9%	0.2%	13.4%	9.7%		1.8%	1.7%	3.1%	5.1%	9.8%	
PCB 118		2.5%	7.2%	0.1%	12.8%	2.6%		1.0%	2.4%	4.0%	5.4%	3.1%	
PCB 128		1.0%	7.0%	0.9%	13.0%	14.1%		0.6%	1.6%	2.6%	8.4%	5.4%	
PCB 138				3.3%	9.7%	12.4%		1.1%	0.8%	3.0%	5.2%	2.4%	
PCB 149		2.8%	5.8%	1.2%				2.0%	2.0%		4.4%	4.1%	
PCB 153				0.3%	9.6%	14.1%		1.2%	1.1%	2.4%	5.6%	0.3%	
PCB 156		1.0%	3.9%	1.9%				1.4%	3.7%		8.6%	12.4%	
PCB 170		1.3%	3.9%	4.3%	8.6%	11.6%		7.3%	1.8%	0.0%	9.2%	3.3%	
PCB 180		7.4%	1.7%	1.1%	10.6%	14.4%		1.5%	1.1%	6.6%	1.0%	5.9%	
PCB 187		2.8%	2.3%	1.4%	12.5%	14.2%		0.9%	0.0%	8.1%	5.2%	2.5%	
PCB 194				14.6%				0.0%	7.3%			32.9%	
PCB 195													
PCB 206													
PCB 209													

Table 28 (cont). Mussel Tissue XII (QA05TISI2): RSDs for three replicates - PCBs															
	Lab 7		Lab 8		Lab 9		Lab 10		Lab 11		Lab 12				
	Tissue XII	SRM 2977	Tissue XII	SRM 2977	Tissue XII	SRM 2977	Tissue XII	SRM 2977	Tissue XII	SRM 2977	Tissue XII	SRM 2977	SRM 2977	SRM 2977	
	rsd	rsd	rsd	rsd	rsd	rsd	rsd	rsd	rsd	rsd	rsd	rsd	rsd	rsd	rsd
PCB 8	20.7%	22.2%	3.0%	30.9%	26.5%	27.1%	10.9%	26.0%	4.5%	50.0%	9.5%	10.6%			
PCB 18	38.2%	18.7%	0.3%	29.1%	7.1%	19.6%	8.2%	22.9%	10.3%	62.9%	9.8%	14.2%			
PCB 28	31.5%	15.4%	1.7%	26.4%	2.3%	10.2%	6.2%	23.1%	9.1%	14.9%	9.6%	9.3%			
PCB 31	7.7%	10.9%	6.6%	23.5%					17.5%	18.5%	12.9%	12.6%			
PCB 44	5.7%	25.5%	3.9%	41.1%	3.6%	19.1%	6.3%	28.3%	5.2%	25.9%	9.9%	10.8%			
PCB 49	41.7%	5.3%	11.1%	46.4%	4.0%		6.0%	23.2%	18.5%	27.3%	9.8%	10.3%			
PCB 52	4.4%	16.8%	5.6%	46.7%	4.7%	3.4%	5.6%	21.8%	9.5%	29.4%	9.8%	10.1%			
PCB 66	28.0%	16.8%	4.9%	32.7%	6.1%	11.0%	6.1%	23.0%	13.6%	8.9%	6.9%	8.2%			
PCB 95	13.3%	12.8%	8.7%	50.9%			7.3%	21.8%	11.3%	21.7%	9.2%	8.4%			
PCB 99	21.2%	24.8%	5.3%	48.3%			6.9%	22.9%	12.8%	5.6%	10.3%	5.8%			
PCB 101	10.4%	9.8%	10.0%	33.5%	4.5%		6.4%	21.7%	10.8%	6.4%	6.7%	9.0%			
PCB 105	16.9%	11.2%	2.4%	38.6%	3.1%	5.3%	6.6%	20.6%	8.0%	3.5%	1.4%	6.8%			
PCB 118	8.5%	2.9%	5.0%	36.5%	4.7%	7.5%	6.8%	22.2%	5.7%	3.6%	2.4%	6.1%			
PCB 128	7.2%	7.6%	3.2%	38.2%	2.9%	1.3%	9.0%	23.4%	12.2%	10.8%	4.3%	4.9%			
PCB 138	6.9%	29.4%	3.2%	39.0%	24.6%	28.8%	7.1%	22.6%	2.9%	1.0%	6.8%	5.2%			
PCB 149	6.7%	17.1%	16.5%	57.4%			6.3%	22.1%	3.9%	11.1%	13.2%	4.3%			
PCB 153	10.0%	7.2%	4.6%	37.6%	4.6%	4.8%	6.6%	20.6%	3.3%	2.2%	4.9%	5.3%			
PCB 156	16.7%	20.4%	3.5%	41.3%					4.0%	2.0%	3.3%	2.4%			
PCB 170		11.6%	4.7%	35.1%		3.4%	8.7%	24.4%	14.6%	3.1%	4.9%	5.1%			
PCB 180		19.2%	7.2%	36.6%	0.4%	15.3%	9.3%	24.2%	14.9%	1.9%	4.5%	6.5%			
PCB 187		6.9%	8.6%	12.1%	5.0%	9.3%	6.5%	17.9%	15.5%	10.2%	5.2%	2.4%			
PCB 194			3.8%	44.3%			6.5%	28.3%	27.4%	16.4%	13.9%	6.1%			
PCB 195							17.6%	24.6%	10.6%	5.2%	15.8%	10.0%			
PCB 206							44.3%	13.4%	15.1%	9.2%	6.5%	3.8%			
PCB 209							34.8%	51.4%	8.1%	65.5%	10.0%				

Table 29. Mussel Tissue XII (QA05TISI2): RSDs for three replicates - PBDEs													
	Lab 1a		Lab 1c		Lab 3		Lab 4		Lab 5		Lab 6		
	Tissue XII	SRM 2977	Tissue XII	SRM 2977	Tissue XII	SRM 2977	Tissue XII	SRM 2977	Tissue XII	SRM 2977	Tissue XII	SRM 2977	
	rsd	rsd	rsd	rsd	rsd	rsd	rsd	rsd	rsd	rsd	rsd	rsd	rsd
BDE 15													
BDE 17			1.4%	14.9%									
BDE 25													
BDE 28			1.0%	8.8%			3.5%	2.5%					
BDE 30													
BDE 33													
BDE 47			8.2%	6.6%			4.4%	1.8%					
BDE 49							3.7%	1.1%					
BDE 66							4.1%						
BDE 71			6.4%	6.3%									
BDE 75													
BDE 85													
BDE 99							8.0%	3.8%					
BDE 100			3.8%	4.2%			3.7%	3.4%					
BDE 116													
BDE 118													
BDE 119													
BDE 138													
BDE 153			5.7%										
BDE 154			12.4%										
BDE 155			10.8%										
BDE 156													
BDE 181													
BDE 183													
BDE 190													
BDE 191													
BDE 196													
BDE 197													
BDE 203													
BDE 205													
BDE 206													
BDE 207													
BDE 208													
BDE 209													

Table 29 (cont). Mussel Tissue XII (QA05TISI2): RSDs for three replicates - PBDEs													
	Lab 7		Lab 8		Lab 9		Lab 10		Lab 11		Lab 12		
	Tissue XII	SRM 2977	Tissue XII	SRM 2977	Tissue XII	SRM 2977	Tissue XII	SRM 2977	Tissue XII	SRM 2977	Tissue XII	SRM 2977	
	rsd	rsd	rsd	rsd	rsd	rsd	rsd	rsd	rsd	rsd	rsd	rsd	rsd
BDE 15											5.5%	10.7%	
BDE 17			19.8%	33.4%							11.7%	7.6%	
BDE 25			19.8%	33.4%									
BDE 28	3.2%	3.7%	5.0%	14.8%							12.2%	7.3%	
BDE 30													
BDE 33			5.0%	14.8%									
BDE 47	2.8%	0.4%	5.5%	17.2%							9.1%	9.6%	
BDE 49			4.7%	1.7%							14.7%	9.4%	
BDE 66	6.8%	8.2%	11.8%	11.9%							18.2%	19.1%	
BDE 71												5.4%	
BDE 75											14.7%	19.0%	
BDE 85	8.2%	35.0%	23.3%								6.8%	32.8%	
BDE 99	5.3%	11.0%	10.3%	12.9%							10.0%	8.7%	
BDE 100	2.3%	6.3%	6.4%	17.4%							10.1%	8.9%	
BDE 116													
BDE 118													
BDE 119											22.0%	61.9%	
BDE 138		57.0%									22.8%		
BDE 153	10.2%	37.8%	12.5%	14.4%							5.9%	7.9%	
BDE 154	15.4%	16.7%	11.8%	16.3%							11.8%	4.8%	
BDE 155											17.0%	45.0%	
BDE 156													
BDE 181												0.0%	
BDE 183											37.1%	25.5%	
BDE 190												0.0%	
BDE 191													
BDE 196													
BDE 197													
BDE 203											79.9%	71.5%	
BDE 205													
BDE 206													
BDE 207				0.0%									
BDE 208													
BDE 209													

Table 30. Marine Sediment XIII (QA05SED13): RSDs for three replicates - Water, TOC, and PAHs													
	Lab 1a		Lab 1c		Lab 2		Lab 3		Lab 4				
	Sed XIII	SRM 1941b	Sed XIII	SRM 1941b	Sed XIII	SRM 1944	Sed XIII	SRM 1941b	Sed XIII	SRM 1941b			
	rsd	rsd	rsd	rsd	rsd	rsd	rsd	rsd	rsd	rsd			
Water	1.5%	1.7%	2.0%	6.2%	10.2%		1.2%		0.4%				
TOC							4.5%	0.6%					
PAH ANALYSES													
	Lab 1a	SRM 1941b	Lab 1c	SRM 1941b	Lab 2	SRM 1944	Lab 3	SRM 1941b	Lab 4	SRM 1941b			
	Sed XIII	rsd	Sed XIII	rsd	Sed XIII	rsd	Sed XIII	rsd	Sed XIII	rsd			
naphthalene	1.1%	2.9%	28.7%	1.2%	8.7%	3.8%	29.5%	15.0%	2.1%	0.0%			
2-methylnaphthalene	2.5%	4.1%	23.8%	19.0%	15.0%	3.7%	11.8%	8.6%	2.1%	0.0%			
1-methylnaphthalene	3.9%	4.7%	18.0%	13.9%	13.0%	1.7%	14.9%	9.6%	0.5%	0.0%			
biphenyl	2.2%	1.2%	24.1%	3.6%	14.1%	2.9%	7.8%	6.0%	3.0%	0.0%			
2,6-dimethylnaphthalene			23.9%	3.4%	32.5%	42.3%	16.1%	3.8%	4.3%	0.0%			
acenaphthylene	3.6%	5.5%	16.1%	11.8%	8.3%	7.4%	6.4%	8.4%	5.2%	0.0%			
acenaphthene	3.3%	7.4%			17.6%	0.8%	2.5%	8.9%	2.8%	0.0%			
1,6,7-trimethylnaphthalene			14.4%	4.2%					6.5%	0.0%			
fluorene	2.1%	1.8%	38.8%	5.1%	20.9%	2.6%	5.5%	3.5%	5.2%	0.0%			
phenanthrene	4.0%	0.8%	29.4%	1.7%	18.0%	3.2%	5.0%	6.8%	3.0%	0.0%			
anthracene	3.0%	4.7%	24.8%	1.6%	19.5%	4.2%	12.0%	24.4%	3.1%	0.0%			
1-methylphenanthrene	1.8%	1.9%	12.6%	3.8%	29.8%	1.5%	2.9%	7.2%	2.6%	0.0%			
fluoranthene	0.7%	2.0%	21.4%	1.5%	16.1%	1.5%	5.9%	9.8%	5.1%	0.0%			
pyrene	2.0%	2.3%	21.0%	1.6%	16.3%	5.3%	8.9%	8.8%	4.4%	0.0%			
benz[a]anthracene	2.8%	6.2%	26.4%	2.4%	11.9%	3.3%	7.5%	15.4%	3.1%	0.0%			
chrysene			18.5%	2.1%	9.0%	1.5%	5.7%	12.0%	2.0%	0.0%			
triphenylene			17.3%	4.6%									
benzo[b]fluoranthene	1.3%	1.5%			31.2%	25.5%	5.5%	11.5%	1.1%	0.0%			
benzo[j]fluoranthene	2.5%	1.6%											
benzo[k]fluoranthene	3.9%	3.2%	16.2%	0.7%	4.3%	11.0%	6.5%	15.5%	2.2%	0.0%			
benzo[e]pyrene	1.5%	2.5%	17.9%	2.8%	32.4%	6.9%	5.9%	12.2%	1.7%	0.0%			
benzo[a]pyrene	1.6%	3.3%	14.6%	5.6%	21.6%	10.9%	7.2%	25.3%	1.9%	0.0%			
perylene	1.1%	2.5%	17.8%	3.3%	26.2%	5.5%	6.3%	21.2%	0.9%	0.0%			
indeno[1,2,3-cd]pyrene	1.5%	3.9%			10.9%	11.0%	8.2%	16.0%	1.6%	0.0%			
dibenz[a,h]anthracene					11.3%	13.3%	17.5%	26.0%	1.9%	0.0%			
benzo[ghi]perylene	1.6%	1.8%	8.3%	1.6%	6.4%	8.0%	13.4%	22.3%	1.6%	0.0%			

Table 30 (cont). Marine Sediment XIII (QA05SED13): RSDs for three replicates - Water, TOC, and PAHs													
	Lab 6		Lab 7		Lab 8		Lab 10		Lab 11		Lab 12		
	Sed XIII	SRM 1941b	Sed XIII	SRM 1941b	Sed XIII	SRM 1941b	Sed XIII	SRM 1941b	Sed XIII	SRM 1941b	Sed XIII	SRM 1941b	
	rsd	rsd	rsd	rsd	rsd	rsd	rsd	rsd	rsd	rsd	rsd	rsd	
Water													
TOC		8.7%		7.8%									
PAH ANALYSES													
	Lab 6		Lab 7		Lab 8		Lab 10		Lab 11		Lab 12		
	Sed XIII	SRM 1941b	Sed XIII	SRM 1941b	Sed XIII	SRM 1941b	Sed XIII	SRM 1941b	Sed XIII	SRM 1941b	Sed XIII	SRM 1941b	
	rsd	rsd	rsd	rsd	rsd	rsd	rsd	rsd	rsd	rsd	rsd	rsd	
naphthalene		11.0%		1.8%		19.8%		0.8%		3.5%		5.0%	9.1%
2-methylnaphthalene		14.7%		2.9%		16.6%		8.4%				1.3%	5.8%
1-methylnaphthalene		13.6%		4.0%		10.5%						4.2%	6.4%
biphenyl		11.2%		2.2%		11.4%				9.6%		9.3%	4.4%
2,6-dimethylnaphthalene		15.4%		8.3%		7.3%						3.7%	4.3%
acenaphthylene		12.6%		7.7%		4.0%		40.4%		2.7%		7.4%	4.3%
acenaphthene		10.5%		3.6%		0.0%		65.5%		3.9%		8.3%	8.4%
1,6,7-trimethylnaphthalene		34.0%		18.8%								10.6%	6.3%
fluorene		5.4%		4.9%		7.2%		33.4%		8.5%		0.4%	4.4%
phenanthrene		9.7%		3.7%		3.7%		8.1%		3.4%		6.5%	5.3%
anthracene		16.4%		4.8%		3.4%		3.4%		1.1%		3.6%	1.8%
1-methylphenanthrene		10.8%		4.1%		10.6%				0.5%		2.3%	1.5%
fluoranthene		6.0%		1.1%		3.5%		23.0%		3.9%		7.8%	1.0%
pyrene		5.2%		1.8%		4.7%		25.7%		3.0%		7.5%	6.2%
benz[a]anthracene		6.1%		7.5%		3.8%		33.0%				7.6%	5.0%
chrysene		6.1%		5.3%		4.6%		28.9%		0.3%		6.1%	9.2%
triphenylene													
benzo[b]fluoranthene		5.4%		2.9%		3.4%		45.8%		8.8%		7.4%	10.4%
benzo[j]fluoranthene													
benzo[k]fluoranthene		4.5%		6.1%		4.3%		8.3%		2.3%		15.1%	10.3%
benzo[e]pyrene		5.1%		2.6%		4.3%		23.9%		2.5%		3.5%	7.5%
benzo[a]pyrene		3.5%		17.1%		2.5%		33.1%		3.8%		8.3%	5.7%
perylene		5.0%		21.9%						4.9%		6.6%	5.2%
indeno[1,2,3-cd]pyrene		8.7%		23.2%		8.9%		32.0%		71.6%		8.3%	6.8%
dibenz[a,h]anthracene		7.1%		22.7%		14.5%		3.1%		8.2%			
benzo[ghi]perylene		12.3%		25.3%		6.3%		21.8%		5.0%		7.6%	5.3%

Table 31. Marine Sediment XIII (QA05SED13): RSDs for three replicates - Pesticides											
	Lab 1a		Lab 1c		Lab 2		Lab 3		Lab 4		
	Sed XIII	SRM 1941b	Sed XIII	SRM 1941b	Sed XIII	SRM 1944	Sed XIII	SRM 1941b	Sed XIII	SRM 1941b	
	rsd	rsd	rsd	rsd	rsd	rsd	rsd	rsd	rsd	rsd	
alpha-HCH (a-BHC)											
hexachlorobenzene	0.4%	0.8%	7.0%	2.6%			7.5%	74.2%	0.9%	6.0%	
gamma-HCH (g-BHC,lindane)									4.3%	6.0%	
beta-HCH (b-BHC)											
heptachlor											
aldrin											
heptachlor epoxide											
oxychlordane									14.7%	18.4%	
gamma-chlordane	0.5%	1.5%					50.4%	20.0%	4.4%	0.1%	
2,4'-DDE	4.1%	2.4%							0.0%	0.0%	
endosulfan I										0.0%	
cis-chlordane (alpha-chlordane)	2.4%	1.7%					24.9%	5.2%	3.1%	4.4%	
trans-nonachlor	1.1%	0.9%					3.5%	16.9%			
dieldrin											
4,4'-DDE	4.1%	3.3%	3.7%	4.9%			1.1%	2.4%	24.8%	8.5%	
2,4'-DDD			3.9%	30.7%					11.1%	19.3%	
endrin											
endosulfan II											
4,4'-DDD	1.9%	2.7%	6.5%	5.6%			12.9%	9.9%			
2,4'-DDT									20.6%	0.0%	
cis-nonachlor				14.3%			14.9%	2.1%			
4,4'-DDT									0.0%		
mirex									5.0%	9.6%	
endosulfan sulfate											
chlorpyrifos											

Table 31(cont). Marine Sediment XIII (QA05SED13): RSDs for three replicates - Pesticides													
	Lab 6		Lab 7		Lab 8		Lab 10		Lab 11		Lab 12		
	Sed XIII	rsd	SRM 1941b	rsd	Sed XIII	rsd	Sed XIII	rsd	Sed XIII	rsd	Sed XIII	rsd	
alpha-HCH (a-BHC)													SRM 1941b
hexachlorobenzene		40.4%		1.0%				2.8%				16.4%	rsd
gamma-HCH (g-BHC,lindane)								5.1%				4.4%	3.4%
beta-HCH (b-BHC)								45.4%				16.3%	9.6%
heptachlor								21.4%				23.5%	5.9%
aldrin								6.6%				0.0%	39.2%
heptachlor epoxide								14.0%				13.9%	1.7%
oxychlordane								10.3%					
gamma-chlordane								5.1%				6.1%	0.3%
2,4'-DDE		31.7%		52.0%		6.1%		0.8%				0.6%	6.7%
endosulfan I						35.8%		19.3%					
cis-chlordane (alpha-chlordane)								7.2%				7.2%	4.1%
trans-nonachlor		18.0%		37.8%		4.6%		35.4%				6.7%	4.9%
dieldrin		23.8%		22.4%		0.0%		10.4%				2.7%	2.5%
4,4'-DDE		30.6%						8.9%				2.1%	4.5%
2,4'-DDD		43.2%		6.8%		1.2%		141.3%				33.5%	30.9%
2,4'-DDD		52.3%											
endrin													
endosulfan II								8.7%					
4,4'-DDD		24.8%						1.7%				15.1%	7.8%
2,4'-DDT						32.3%		2.0%				11.3%	11.5%
cis-nonachlor								6.4%				10.3%	3.2%
4,4'-DDT		43.8%				0.0%		46.5%				9.8%	6.9%
mirex		32.4%										0.0%	
endosulfan sulfate													
chlorpyrifos													

Table 32. Marine Sediment XIII (QA05SED13): RSDs for three replicates - PCBs												
	Lab 1a		Lab 1c		Lab 2		Lab 3		Lab 4			
	Sed XIII	SRM 1941b	Sed XIII	SRM 1941b	Sed XIII	SRM 1944	Sed XIII	SRM 1941b	Sed XIII	SRM 1941b	SRM 1941b	SRM 1941b
	rsd	rsd	rsd	rsd	rsd	rsd	rsd	rsd	rsd	rsd	rsd	rsd
PCB 8	3.8%	6.9%	2.3%	1.9%			17.6%	13.0%				
PCB 18	1.4%	4.1%	8.5%	6.1%			7.1%	4.7%	1.0%			6.4%
PCB 28	1.1%	0.4%	6.1%	5.3%			13.1%	9.9%	0.6%			2.9%
PCB 31	2.4%	1.3%	2.4%	7.2%					1.9%			8.7%
PCB 44	5.7%	4.9%	2.0%	2.6%			8.0%	6.7%	1.4%			6.6%
PCB 49	2.3%	3.8%	0.9%	3.8%					0.8%			7.2%
PCB 52	2.6%	2.3%	5.1%	1.8%			10.3%	7.0%	1.8%			6.0%
PCB 66	0.9%	1.3%	4.4%	7.2%			13.5%	9.0%	0.7%			6.5%
PCB 95	2.1%	1.4%	9.0%	1.3%					4.0%			6.4%
PCB 99	4.7%	3.9%	1.5%	3.7%					5.0%			7.1%
PCB 101	2.8%	1.2%	7.0%	5.5%			11.9%	10.3%	4.7%			5.6%
PCB 105	1.7%	3.9%	2.9%	7.5%			9.3%	15.1%	9.8%			5.4%
PCB 118	0.9%	2.7%	3.2%	6.3%			13.6%	14.8%	6.1%			4.2%
PCB 128	1.7%	2.4%	1.9%	9.7%			8.8%	21.6%	7.9%			4.5%
PCB 138			7.6%	5.6%			4.5%	7.5%	2.6%			5.8%
PCB 149	0.6%	1.7%	7.4%	4.0%					4.4%			7.5%
PCB 153							5.1%	6.7%	4.0%			7.2%
PCB 156	1.5%	1.5%	3.4%						10.3%			
PCB 170	2.5%	4.2%	1.3%	3.9%			7.0%	16.4%	24.9%			9.8%
PCB 180	3.3%	2.4%	9.5%	9.5%			6.1%	7.3%	27.5%			7.8%
PCB 187	2.8%	4.3%	3.9%	4.3%			3.0%	5.5%	21.9%			8.7%
PCB 194			7.0%	6.5%					27.4%			10.5%
PCB 195			14.4%				7.5%	15.9%	0.0%			
PCB 206	5.5%	3.1%	10.7%	7.4%			5.3%	6.7%	4.4%			11.1%
PCB 209	4.8%	2.6%	4.1%	9.6%			3.1%	7.2%	0.1%			12.6%

Table 32 (cont). Marine Sediment XIII (QA05SED13): RSDs for three replicates - PCBs													
	Lab 6		Lab 7		Lab 8		Lab 10		Lab 11		Lab 12		
	Sed XIII	SRM 1941b	Sed XIII	SRM 1941b	Sed XIII	SRM 1941b	Sed XIII	SRM 1941b	Sed XIII	SRM 1941b	Sed XIII	SRM 1941b	
	rsd	rsd	rsd	rsd	rsd	rsd	rsd	rsd	rsd	rsd	rsd	rsd	
PCB 8	23.4%	3.2%	9.5%	2.4%	5.2%	10.3%	0.1%	2.1%	17.6%	2.0%	11.3%	2.6%	
PCB 18	20.1%	3.5%	7.1%	2.4%	0.2%	10.5%	0.7%	2.2%	23.5%	19.9%	11.2%	4.1%	
PCB 28	23.3%	1.6%	6.0%	1.5%	5.7%	5.4%	0.1%	5.7%	5.8%	4.6%	6.4%	2.3%	
PCB 31	24.8%	4.7%	4.9%	0.2%	0.9%	1.8%			12.4%	6.8%	5.4%	4.6%	
PCB 44	22.3%	1.6%	2.9%	5.6%	22.3%	34.4%	0.5%	2.1%	13.5%	20.5%	6.0%	3.2%	
PCB 49	25.6%	2.9%	7.0%	6.3%	34.5%	42.0%	0.2%	3.2%	16.0%	18.6%	7.5%	1.8%	
PCB 52	24.1%	7.3%	4.4%	3.1%	22.2%	31.7%	0.5%	1.6%	17.1%	20.3%	6.4%	1.2%	
PCB 66	23.7%	3.2%	8.1%	5.0%	10.6%	13.1%	5.9%	3.4%	2.7%	2.4%	6.1%	5.9%	
PCB 95	26.7%	3.7%	7.4%	5.2%	8.4%	18.0%	8.7%	12.3%	6.2%	21.1%	9.6%	4.0%	
PCB 99	25.4%	5.0%	5.6%	6.6%	8.9%	16.0%	0.0%	1.3%	3.1%	10.6%	7.9%	1.5%	
PCB 101	26.6%	4.4%	4.7%	0.9%	10.6%	2.4%	0.1%	1.4%	0.8%	10.9%	19.6%	1.8%	
PCB 105	27.9%	4.3%	3.2%	3.1%	2.9%	4.2%	0.4%	1.1%	8.7%	4.8%	4.6%	3.0%	
PCB 118	26.0%	5.0%	6.6%	5.1%	0.2%	7.4%	0.5%	0.4%	7.9%	5.7%	4.0%	2.7%	
PCB 128	31.3%	13.6%	20.8%	8.3%	0.9%	10.2%	1.3%	20.6%	11.8%	11.0%	4.9%	3.2%	
PCB 138	25.9%	4.3%	4.7%	0.3%	3.7%	13.8%	0.5%	0.4%	14.8%	5.1%	4.4%	2.7%	
PCB 149	26.3%	3.3%	8.3%	7.0%	24.8%	34.2%	1.6%	0.7%	17.5%	13.3%	3.4%	2.9%	
PCB 153	25.9%	5.2%	2.9%	10.3%	16.3%	13.7%	0.4%	1.1%	15.5%	5.3%	7.8%	7.2%	
PCB 156	26.8%	1.1%	18.4%	5.7%	2.7%	7.7%			10.5%	3.1%	5.1%	2.4%	
PCB 170	23.9%	6.8%	9.8%	4.5%	9.0%	3.7%	0.7%	1.6%	13.7%	4.6%	2.9%	7.9%	
PCB 180	26.3%	6.3%	9.5%	4.5%	7.8%	0.4%	0.2%	2.5%	12.0%	7.0%	2.5%	1.8%	
PCB 187	109.5%	7.5%	6.3%	0.0%	6.9%	28.8%	0.2%	4.2%	15.7%	8.4%	4.9%	3.3%	
PCB 194	27.0%	9.6%		2.8%	0.7%	6.5%	2.3%	3.2%	13.6%	7.1%	6.9%	1.1%	
PCB 195	8.7%			0.0%			4.3%	3.5%	13.8%	6.8%	3.9%	5.7%	
PCB 206	29.7%		7.1%	4.5%	37.3%	20.9%	7.9%	3.9%	12.9%	1.8%	2.7%	1.8%	
PCB 209			3.4%	6.3%	3.4%	8.4%	0.7%	4.8%	13.5%	4.0%	2.9%	0.5%	

Table 33. Marine Sediment XIII (QA05SED13): RSDs for three replicates - PBDEs													
	Lab 1a		Lab 1c		Lab 2		Lab 3		Lab 4				
	Sed XIII	SRM 1941b	Sed XIII	SRM 1941b	Sed XIII	SRM 1944	Sed XIII	SRM 1941b	Sed XIII	SRM 1941b	SRM 1941b	SRM 1941b	
	rsd	rsd	rsd	rsd	rsd	rsd	rsd	rsd	rsd	rsd	rsd	rsd	
BDE 15													
BDE 17													
BDE 25													
BDE 28													
BDE 30													
BDE 33													
BDE 47													
BDE 49													
BDE 66													
BDE 71													
BDE 75													
BDE 85													
BDE 99													
BDE 100													
BDE 116													
BDE 118													
BDE 119													
BDE 138													
BDE 153													
BDE 154													
BDE 155													
BDE 156													
BDE 181													
BDE 183													
BDE 190													
BDE 191													
BDE 196													
BDE 197													
BDE 203													
BDE 205													
BDE 206													
BDE 207													
BDE 208													
BDE 209													

Table 33 (cont). Marine Sediment XIII (QA05SED13): RSDs for three replicates - PBDEs													
	Lab 6		Lab 7		Lab 8		Lab 10		Lab 11		Lab 12		
	Sed XIII	SRM 1941b	Sed XIII	SRM 1941b	Sed XIII	SRM 1941b	Sed XIII	SRM 1941b	Sed XIII	SRM 1941b	Sed XIII	SRM 1941b	
	rsd	rsd	rsd	rsd	rsd	rsd	rsd	rsd	rsd	rsd	rsd	rsd	
BDE 15													
BDE 17											0.5%	9.3%	
BDE 25											5.6%	11.0%	
BDE 28													
BDE 30				13.0%		6.9%					1.9%	8.1%	
BDE 33													
BDE 47				32.8%		19.3%					4.9%	11.9%	
BDE 49											1.1%	6.8%	
BDE 66				13.2%		17.5%					8.3%	12.7%	
BDE 71											8.4%	5.9%	
BDE 75												0.0%	
BDE 85				26.3%		76.8%					9.4%	14.5%	
BDE 99				20.0%		47.4%					2.1%	5.1%	
BDE 100				17.0%		27.6%					0.0%	3.2%	
BDE 116													
BDE 118													
BDE 119													
BDE 138				56.0%		0.0%					35.4%	0.0%	
BDE 153				7.0%		55.0%					25.0%	29.4%	
BDE 154				14.9%		62.6%					4.0%	0.7%	
BDE 155											6.5%	2.9%	
BDE 156											6.0%	9.1%	
BDE 181													
BDE 183													
BDE 190													
BDE 191													
BDE 196													
BDE 197													
BDE 203													
BDE 205											92.0%	85.7%	
BDE 206													
BDE 207													
BDE 208													
BDE 209											3.9%	4.3%	

Table 34. Comparison of Concentrations for Marine Sediment XIII (QA05SED13) and SRM 1941b

	Marine Sediment XIII Exercise Assigned		SRM 1941b from Certificate		
	Value	s	Value	95% CI	% difference
Water (percent)	47.4	3.2			
TOC (percent)	2.70	0.58			
	Marine Sediment XIII Exercise Assigned		SRM 1941b from Certificate		
	Value	s	Value	95% CI	% difference
PAHs (ng/g dry mass)					
naphthalene	785	186	848	95	-7.4%
2-methylnaphthalene	219	66	276	53	-20.7%
1-methylnaphthalene	98.2	24.0	127	14	-22.7%
biphenyl	65.8	23.6	74	8	-11.1%
2,6-dimethylnaphthalene	81.3	43.6	75.9	4.5	7.1%
acenaphthylene	45.1	13.8	53.3	6.4	-15.4%
acenaphthene	28.9	8.4	38.4	5.2	-24.6%
1,6,7-trimethylnaphthalene	no target		25.5	5.1	
fluorene	56.1	12.2	85	15	-34.1%
phenanthrene	306	89	406	44	-24.5%
anthracene	137	47	184	18	-25.5%
1-methylphenanthrene	55.4	18.2	73.2	5.9	-24.4%
fluoranthene	496	140	651	50	-23.8%
pyrene	421	142	581	39	-27.5%
benz[a]anthracene	241	80	335	25	-28.2%
chrysene	219	62	291	31	-24.7%
triphenylene	no target		108	5	
benzo[b]fluoranthene	413	174	453	21	-8.8%
benzo[j]fluoranthene	no target		217	5	
benzo[k]fluoranthene	180	56	225	18	-20.0%
benzo[e]pyrene	286	97	325	25	-11.9%
benzo[a]pyrene	282	120	358	17	-21.3%
perylene	311	131	397	45	-21.6%
indeno[1,2,3-cd]pyrene	258	101	341	57	-24.4%
dibenz[a,h]anthracene	41.9	21.2	53	10	-20.9%
benzo[ghi]perylene	244	96	307	45	-20.5%

Note: Bolded values are certified concentrations while other values for SRM 1941b are reference values

Table 34 (cont). Comparison of Concentrations for Marine Sediment XIII (QA05SED13) and SRM 1941b

Pesticides (ng/g dry mass)	Marine Sediment XIII		SRM 1941b		
	Exercise Assigned		from Certificate		
	Value	s	Value	95% CI	% difference
alpha-HCH (a-BHC)	no target		no target		
hexachlorobenzene	5.38	1.90	5.83	0.38	-7.8%
gamma-HCH (g-BHC,lindane)	no target		no target		
beta-HCH (b-BHC)	no target		no target		
heptachlor	no target		no target		
aldrin	no target		no target		
heptachlor epoxide	no target		no target		
oxychlordane	no target		no target		
gamma-chlordane	0.572	0.148	0.566	0.093	1.1%
2,4'-DDE	0.380	0.103	0.38	0.12	0.1%
endosulfan I	no target		no target		
cis-chlordane (alpha-chlordane)	0.482	0.058	0.85	0.11	-43.2%
trans-nonachlor	0.286	0.089	0.438	0.073	-34.8%
dieldrin	0.386	0.034	no target		
4,4'-DDE	3.44	1.08	3.22	0.28	7.0%
2,4'-DDD	0.927	0.624	no target		
endrin	no target		no target		
endosulfan II	no target		no target		
4,4'-DDD	4.18	1.42	4.66	0.46	-10.3%
2,4'-DDT	no target		no target		
cis-nonachlor	0.454	0.493	0.378	0.053	20.0%
4,4'-DDT	0.537	0.174	1.12	0.42	-52.1%
mirex	no target		no target		
endosulfan sulfate	no target		no target		
chlorpyrifos	no target		no target		

Note: Bolded values are certified concentrations while other values for SRM 1941b are reference values

Table 34 (cont). Comparison of Concentrations for Marine Sediment XIII (QA05SED13) and SRM 1941b

PCBs (ng/g dry mass)	Marine Sediment XIII		SRM 1941b		
	Exercise Assigned		from Certificate		
	Value	s	Value	95% CI	% difference
PCB 8	1.36	0.42	1.65	0.19	-17.3%
PCB 18	2.04	0.48	2.39	0.29	-14.5%
PCB 28	3.79	0.97	4.52	0.57	-16.1%
PCB 31	2.83	0.94	3.18	0.41	-11.0%
PCB 44	3.47	0.99	3.85	0.2	-9.8%
PCB 49	3.64	1.22	4.34	0.28	-16.1%
PCB 52	4.48	1.39	5.24	0.28	-14.5%
PCB 66	4.32	1.41	4.96	0.53	-12.9%
PCB 95	3.44	1.25	3.93	0.62	-12.5%
PCB 99	2.45	0.87	2.9	0.36	-15.5%
PCB 101	4.86	1.36	5.11	0.34	-4.8%
PCB 105	1.26	0.40	1.43	0.1	-12.0%
PCB 118	3.59	1.26	4.23	0.19	-15.1%
PCB 128	0.684	0.275	0.696	0.044	-1.8%
PCB 138	3.83	1.23	3.6	0.28	6.3%
PCB 149	4.18	1.44	4.35	0.26	-3.8%
PCB 153	4.99	2.46	5.47	0.32	-8.8%
PCB 156	0.444	0.128	0.507	0.09	-12.3%
PCB 170	1.19	0.36	1.35	0.09	-11.5%
PCB 180	2.97	1.15	3.24	0.51	-8.2%
PCB 187	2.20	0.69	2.17	0.22	1.2%
PCB 194	1.01	0.50	1.04	0.06	-2.8%
PCB 195	0.291	0.166	0.645	0.06	-54.9%
PCB 206	1.91	0.66	2.42	0.19	-21.1%
PCB 209	4.02	1.47	4.86	0.45	-17.3%

Note: Bolded values are certified concentrations while other values for SRM 1941b are reference values




Figure 1. Comparison of Concentrations for Mussel Tissue XII as determined in this study and in a CCQM pilot study for selected analytes. Shown are the assigned values and associated uncertainties from each study.

**Appendix A: Description, Storage, Use, and Reporting Instructions
for Mussel Tissue XII (QA05TIS12)**

**NIST Intercomparison Exercise Program for
Organic Contaminants in the Marine Environment**

NIST QA Program

**Intercomparison Exercise: Mussel Tissue XII
Description of Materials and Instructions**

Intercomparison Exercise Materials:

QA05TIS12 (Mussel Tissue XII)

The one jar contains approximately 8 g (dry-mass basis) of Mussel Tissue XII. This freeze-dried material was prepared from mussels collected from an urban area. This material has not been enriched or spiked. Each 30-mL amber jar has a Teflon-lined screw cap and is labeled with an individual jar number as well as the above name.

It is requested that three concurrent analyses of SRM 2977 Mussel Tissue (Organic Contaminants and Trace Elements) are also performed. This material can be obtained from the NIST Standard Reference Materials Program (\$502/10 g (dry-mass basis) (phone: 301/975-6776; fax: 301/948-3730). See the following link for information on ordering on-line: https://srms.nist.gov/view_detail.cfm?srm=2977.

Storage of Materials:

Mussel Tissue Material. The tissue material should be stored in the dark at room temperature. If only a portion of the contents of a jar is used, the jar should be tightly closed immediately after removal of a subsample to preserve the integrity of the remaining material for later analysis.

Instructions for Use:

You are to analyze Mussel Tissue XII and SRM 2977, using **your** laboratory's and/or program's analytical protocols, for the concentrations (mass/mass [dry-mass basis]) of the 26 polycyclic aromatic hydrocarbon (PAH) compounds, 25 chlorinated pesticides, 25 polychlorinated biphenyl (PCB) congeners, and 34 polybrominated diphenyl ether (PBDE) congeners¹ of interest in the current NIST Intercomparison Exercise Program for Organic Contaminants in the Marine Environment. These compounds are listed in Table 1.

¹If your laboratory is not analyzing samples for all four chemical classes, you are expected to submit results only for those compounds currently being determined in your laboratory.

The percentage of total extractable organics (or lipid) in Mussel Tissue XII and SRM 2977 should also be determined. You should have received sufficient material for this purpose. The amount of material used for each analysis should correspond to the amount (dry-mass basis) of marine tissue that you would typically analyze as prescribed in your protocols.

You should analyze three samples of Mussel Tissue XII and at least one or more samples of SRM 2977 in three different batches using your protocol for tissue samples. Specifically, we are asking that you analyze one sample of Mussel Tissue XII and one sample of SRM 2977 with one batch of laboratory samples; analyze a second sample of each material with another batch; and the third sample with yet another batch. This will allow a more realistic assessment of laboratory precision over a longer term than the assessment obtained when a laboratory places all three samples in the same extraction and cleanup batch and the resulting extracts are analyzed using the same calibration curve, etc.

Reporting of Results:

Please report one result, as if three figures were significant, for each of the requested analytes in each of the three replicates of the Mussel Tissue XII and of SRM 2977. Report results in units of ng/g **dry-mass** basis. Report the date of measurement of each sample in the requested m/d/y format.

We recognize that the reported concentrations for some of the requested determinands will probably include concentrations of compounds reported to coelute with the determinand of interest with methods commonly in use in environmental laboratories. Please note at the bottom of your table of reported results if any coelution qualifiers are applicable to your data. Please note that any changes you make to the column or row headings **within** the tables will **not** be seen by the coordinators because only the table entries and comments at the bottom of the tables are automatically transferred to the exercise database.

We prefer that concentration values be reported for each analyte determined. If the measured concentration is below your typical reporting concentration for an analyte in a particular matrix, you can report the number and list the appropriate detection limit, quantification limit, etc. at the bottom of the data table. However, if you need to report non-numerical data please use the following conventions:

NA	"Not analyzed", "not determined"
<"value"	"Less than specified concentration", e.g., <8 ng/g
Other	"Other"; add note of explanation at end of data table, e.g., interference
DL	"Below detection limit" may be used, however, <"value" is preferable

Do not use negative numbers or parentheses to indicate "less than detection limits".

The attached file is an EXCEL file, TIS12.xls. If you have any software/hardware conversion problems, please contact Michele Schantz. The data file templates also include places for you to list the surrogate/internal standards and type of calibration curve used, and to provide a brief description of the analyses. Please **do not** add spaces before entering numbers in the table cells and enter them as "numbers" not as "labels". Please **do not** insert any columns or rows **within** the table in the data file. If you wish to include additional data and/or other information or comments, you may add it to the bottom of the data table in the diskette file or send it in hard copy. A printout of the data file format is shown in Table 2.

Submit your results by **December 15, 2005** as an attached file via e-mail to:

E-mail: michele.schantz@nist.gov

Further Information:

If you need further information, please contact Michele at the following address or phone numbers:

Michele M. Schantz
NIST
100 Bureau Drive Stop 8392
Gaithersburg, MD 20899-8392

Phone: (301)975-3106
FAX: (301)977-0685

Table 1: Analytes of Interest in NIST Intercomparison Exercise Program for Organic Contaminants in the Marine Environment

Chlorinated Pesticides

hexachlorobenzene	2,4'-DDE
alpha-HCH (alpha-BHC)	4,4'-DDE
beta-HCH (beta-BHC)	2,4'-DDD
gamma-HCH (gamma-BHC, Lindane)	4,4'-DDD
heptachlor	2,4'-DDT
heptachlor epoxide	4,4'-DDT
<i>cis</i> -chlordane (alpha-chlordane)	chlorpyrifos
<i>trans</i> -chlordane (gamma-chlordane)	aldrin
oxychlordane	dieldrin
<i>cis</i> -nonachlor	endrin
<i>trans</i> -nonachlor	endosulfan I
mirex	endosulfan II
	endosulfan sulfate

Polychlorinated Biphenyl Congeners

<i>PCB No.</i>	<i>Compound Name</i>
8	2,4'-dichlorobiphenyl
18	2,2',5-trichlorobiphenyl
28	2,4,4'-trichlorobiphenyl
31	2,4',5-trichlorobiphenyl
44	2,2',3,5'-tetrachlorobiphenyl
49	2,2',4,5'-tetrachlorobiphenyl
52	2,2',5,5'-tetrachlorobiphenyl
66	2,3',4,4'-tetrachlorobiphenyl
95	2,2',3,5',6-pentachlorobiphenyl
99	2,2',4,4',5-pentachlorobiphenyl
101	2,2',4,5,5'-pentachlorobiphenyl
105	2,3,3',4,4'-pentachlorobiphenyl
118	2,3',4,4',5-pentachlorobiphenyl
128	2,2',3,3',4,4'-hexachlorobiphenyl
138	2,2',3,4,4',5'-hexachlorobiphenyl
149	2,2',3,4',5',6-hexachlorobiphenyl
153	2,2',4,4',5,5'-hexachlorobiphenyl
156	2,3,3',4,4',5-hexachlorobiphenyl
170	2,2',3,3',4,4',5-heptachlorobiphenyl
180	2,2',3,4,4',5,5'-heptachlorobiphenyl
187	2,2',3,4',5,5',6-heptachlorobiphenyl
194	2,2',3,3',4,4',5,5'-octachlorobiphenyl
195	2,2',3,3',4,4',5,6-octachlorobiphenyl
206	2,2',3,3',4,4',5,5',6-nonachlorobiphenyl
209	decachlorobiphenyl

Table 1. (continued)

Polycyclic aromatic hydrocarbons (PAH)

naphthalene	benz[<i>a</i>]anthracene
2-methylnaphthalene	chrysene
1-methylnaphthalene	triphenylene
biphenyl	benzo[<i>b</i>]fluoranthene
2,6-dimethylnaphthalene	benzo[<i>j</i>]fluoranthene
acenaphthylene	benzo[<i>k</i>]fluoranthene
acenaphthene	benzo[<i>e</i>]pyrene
1,6,7-trimethylnaphthalene	benzo[<i>a</i>]pyrene
fluorene	perylene
phenanthrene	indeno[1,2,3- <i>cd</i>]pyrene
anthracene	dibenz[<i>a,h</i>]anthracene
1-methylphenanthrene	benzo[<i>ghi</i>]perylene
fluoranthene	
pyrene	

Polybrominated diphenyl ethers (PBDEs)

BDE 15 (4,4'-dibromo-)	BDE 138 (2,2',3,4,4',5'-hexabromo-)
BDE 17 (2,2',4-tribromo-)	BDE 153 (2,2',4,4',5,5'-hexabromo-)
BDE 25 (2,3',4-tribromo-)	BDE 154 (2,2',4,4',5,6'-hexabromo-)
BDE 28 (2,4,4'-tribromo-)	BDE 155 (2,2',4,4',6,6'-hexabromo-)
BDE 30 (2,4,6-tribromo-)	BDE 156 (2,3,3',4,4',5-hexabromo-)
BDE 33 (2',3,4-tribromo-)	BDE 181 (2,2',3,4,4',5,6-heptabromo-)
BDE 47 (2,2',4,4'-tetrabromo-)	BDE 183 (2,2',3,4,4',5',6-heptabromo-)
BDE 49 (2,2',4,5'-tetrabromo-)	BDE 190 (2,3,3',4,4',5,6-heptabromo-)
BDE 66 (2,3',4,4'-tetrabromo-)	BDE 191 (2,3,3',4,4',5,6'-heptabromo-)
BDE 71 (2,3',4',6-tetrabromo-)	BDE 196 (2,2',3,3',4,4',5,6'-octabromo-)
BDE 75 (2,4,4',6-tetrabromo-)	BDE 197 (2,2',3,3',4,4',6,6'-octabromo-)
BDE 85 (2,2',3,4,4'-pentabromo-)	BDE 203 (2,2',3,4,4',5,5',6-octabromo-)
BDE 99 (2,2',4,4',5-pentabromo-)	BDE 205 (2,3,3',4,4',5,5',6-octabromo-)
BDE 100 (2,2',4,4',6-pentabromo-)	BDE 206 (2,2',3,3',4,4',5,6,6'-nonabromo-)
BDE 116 (2,3,4,5,6-pentabromo-)	BDE 207 (2,2',3,3',4,4',5,6,6'-nonabromo-)
BDE 118 (2,3',4,4',5-pentabromo-)	BDE 208 (2,2',3,3',4,5,5',6,6'-nonabromo-)
BDE 119 (2,3',4,4',6-pentabromo-)	BDE 209 (decabromo-)

Table 2. Diskette Data File Format (File: TIS12.*)

NIST Intercomparison Exercise Program for Organics in the Marine Environment
NIST QA Program
Sample: QA05TIS12 - Mussel Tissue XII

Please fill in all blanks; Use requested units of concentration; Report results as if 3 figures were significant
DO NOT INSERT ROWS OR COLUMNS WITHIN THIS TABLE. DO NOT MOVE CELLS.

- If necessary, add additional data/information at the end of the table.
- Use one of the following if no concentration is reported for an analyte:
 NA = Not analyzed/determined; <"conc" = <detection limit conc.; Other = other, explain in a note at end of table
 (DL = "below detection limit" may be used, but <"conc", e.g., <8, is preferable.)
 Do not use parentheses or negative numbers to indicate "less than detection limit".

Reporting Date (m/d/y): _____
 Laboratory: _____
 Submitted by: _____

BRIEF DESCRIPTION OF PROCEDURES USED:

Approximate amount of sample extracted:
 Mussel XII _____ g, dry basis; SRM 2977 _____ g, dry basis

Method used for determining percentage Total Extractable Organics (TEO) or lipid:

Extraction method: _____
 Extraction solvent: _____
 Extraction time: _____
 Extraction - other: _____

Sample extract cleanup method: _____

Analytical method used (e.g., GC-FID, GC-ECD):

Analyt.	Instr.	Column Phase	Col. Length, m	Col. i.d., mm	Col. film thickness, µm
---------	--------	--------------	----------------	---------------	-------------------------

PAH	_____	_____	_____	_____	_____
Pesticides	_____	_____	_____	_____	_____
PCB Congeners	_____	_____	_____	_____	_____
BDE Congeners	_____	_____	_____	_____	_____

Method of quantitation (IS = internal standard, ES = external standard):

PAH	_____
Pesticides	_____
PCB Congeners	_____
BDE Congeners	_____

IF internal standard method was used, please complete the following section:

Identity of internal standards/surrogates used that were:

Added PRIOR to extraction of sample:

PAH	_____
Pesticides	_____
PCB Congeners	_____
BDE Congeners	_____

Added after extraction/cleanup and JUST PRIOR to chromatographic analysis:

PAH	_____
Pesticides	_____
PCB Congeners	_____
BDE Congeners	_____

Any others? Added at what point in analyses _____

PAH	_____
Pesticides	_____
PCB Congeners	_____
BDE Congeners	_____

IS/surrogate standards used for quantitation calculations were:

_____ those added prior to extraction
 _____ those added after extraction/cleanup and just prior to chromatographic analysis

If the IS/surrogates added after extraction/cleanup extraction were used for quantitation,
 were results corrected for percent recovery? _____

Percent recovery range:

PAH	_____
Pesticides	_____
PCB Congeners	_____
BDE Congeners	_____

Calibration Curve

	Points	Conc. Range	Analytes outside of calibration curve calibration range
PAH	_____	_____	_____
Pesticides	_____	_____	_____
PCB Congeners	_____	_____	_____
BDE Congeners	_____	_____	_____

Were PCB congeners separated from pesticides prior to GC? _____

Please note any differences in procedures used for SRM 2977 analyses from those for Mussel Tissue XII described above:

RESULTS:

PERCENT Total Extractable Organics (TEO) or Lipid (List each result if determined more than once. Enter results as a number, for example 90.0. DO NOT change format of cell to percent.)

	Tissue XII (percent)	Tissue XII (percent)	Tissue XII (percent)	SRM 2977 (percent)	SRM 2977 (percent)	SRM 2977 (percent)
TEO or lipid	_____	_____	_____	_____	_____	_____
PAH ANALYSES	Tissue XII Batch A Sample 1	Tissue XII Batch B Sample 2	Tissue XII Batch C Sample 3	SRM 2977 Batch A Sample 1	SRM 2977 Batch B Sample 2	SRM 2977 Batch C Sample 3
Analyst (Initials)	_____	_____	_____	_____	_____	_____
Date(s) of measurements (m/d/y)	_____	_____	_____	_____	_____	_____
Sample Jar number	_____	_____	_____	_____	_____	_____
	Tissue XII Sample 1 (ng/g dry mass)	Tissue XII Sample 2 (ng/g dry mass)	Tissue XII Sample 3 (ng/g dry mass)	SRM 2977 Sample 1 (ng/g dry mass)	SRM 2977 Sample 2 (ng/g dry mass)	SRM 2977 Sample 3 (ng/g dry mass)
naphthalene	_____	_____	_____	_____	_____	_____
2-methylnaphthalene	_____	_____	_____	_____	_____	_____
1-methylnaphthalene	_____	_____	_____	_____	_____	_____
biphenyl	_____	_____	_____	_____	_____	_____
2,6-dimethylnaphthalene	_____	_____	_____	_____	_____	_____
acenaphthylene	_____	_____	_____	_____	_____	_____
acenaphthene	_____	_____	_____	_____	_____	_____
1,6,7-trimethylnaphthalene	_____	_____	_____	_____	_____	_____
fluorene	_____	_____	_____	_____	_____	_____
phenanthrene	_____	_____	_____	_____	_____	_____
anthracene	_____	_____	_____	_____	_____	_____
1-methylphenanthrene	_____	_____	_____	_____	_____	_____
fluoranthene	_____	_____	_____	_____	_____	_____
pyrene	_____	_____	_____	_____	_____	_____
benz[a]anthracene	_____	_____	_____	_____	_____	_____
chrysene	_____	_____	_____	_____	_____	_____
triphenylene	_____	_____	_____	_____	_____	_____
benzo[b]fluoranthene	_____	_____	_____	_____	_____	_____
benzo[j]fluoranthene	_____	_____	_____	_____	_____	_____
benzo[k]fluoranthene	_____	_____	_____	_____	_____	_____
benzo[e]pyrene	_____	_____	_____	_____	_____	_____
benzo[a]pyrene	_____	_____	_____	_____	_____	_____
perylene	_____	_____	_____	_____	_____	_____
indeno[1,2,3-cd]pyrene	_____	_____	_____	_____	_____	_____
dibenz[a,h]anthracene	_____	_____	_____	_____	_____	_____
benzo[ghi]perylene	_____	_____	_____	_____	_____	_____

PESTICIDE ANALYSES	Tissue XII Batch A Sample 1	Tissue XII Batch B Sample 2	Tissue XII Batch C Sample 3	SRM 2977 Batch A Sample 1	SRM 2977 Batch B Sample 2	SRM 2977 Batch C Sample 3
Analyst (Initials)						
Date(s) of measurements (m/d/y)						
Sample Jar number						
	Tissue XII Sample 1 (ng/g dry mass)	Tissue XII Sample 2 (ng/g dry mass)	Tissue XII Sample 3 (ng/g dry mass)	SRM 2977 Sample 1 (ng/g dry mass)	SRM 2977 Sample 2 (ng/g dry mass)	SRM 2977 Sample 3 (ng/g dry mass)
alpha-HCH (a-BHC)						
hexachlorobenzene						
gamma-HCH (g-BHC,lindane)						
beta-HCH (b-BHC)						
heptachlor						
aldrin						
heptachlor epoxide						
oxychlordane						
gamma-chlordane						
2,4'-DDE						
endosulfan I						
cis-chlordane (alpha-chlordane)						
trans-nonachlor						
dieldrin						
4,4'-DDE						
2,4'-DDD						
endrin						
endosulfan II						
4,4'-DDD						
2,4'-DDT						
cis-nonachlor						
4,4'-DDT						
mirex						
endosulfan sulfate						
chlorpyrifos						
PCB CONGENER ANALYSES	Tissue XII Batch A Sample 1	Tissue XII Batch B Sample 2	Tissue XII Batch C Sample 3	SRM 2977 Batch A Sample 1	SRM 2977 Batch B Sample 2	SRM 2977 Batch C Sample 3
Analyst (Initials)						
Date(s) of measurements (m/d/y)						
Sample Jar number						
	Tissue XII Sample 1 (ng/g dry mass)	Tissue XII Sample 2 (ng/g dry mass)	Tissue XII Sample 3 (ng/g dry mass)	SRM 2977 Sample 1 (ng/g dry mass)	SRM 2977 Sample 2 (ng/g dry mass)	SRM 2977 Sample 3 (ng/g dry mass)
PCB 8						
PCB 18						
PCB 28						
PCB 31						
PCB 44						
PCB 49						
PCB 52						
PCB 66						
PCB 95						
PCB 99						
PCB 101						
PCB 105						
PCB 118						
PCB 128						
PCB 138						
PCB 149						
PCB 153						
PCB 156						
PCB 170						
PCB 180						
PCB 187						
PCB 194						
PCB 195						
PCB 206						
PCB 209						

BDE CONGENER ANALYSES

	Tissue XII Batch A Sample 1	Tissue XII Batch B Sample 2	Tissue XII Batch C Sample 3	SRM 2977 Batch A Sample 1	SRM 2977 Batch B Sample 2	SRM 2977 Batch C Sample 3
Analyst (Initials)						
Date(s) of measurements (m/d/y)						
Sample Jar number						
	Tissue XII Sample 1 (ng/g dry mass)	Tissue XII Sample 2 (ng/g dry mass)	Tissue XII Sample 3 (ng/g dry mass)	SRM 2977 Sample 1 (ng/g dry mass)	SRM 2977 Sample 2 (ng/g dry mass)	SRM 2977 Sample 3 (ng/g dry mass)
BDE 15						
BDE 17						
BDE 25						
BDE 28						
BDE 30						
BDE 33						
BDE 47						
BDE 49						
BDE 66						
BDE 71						
BDE 75						
BDE 85						
BDE 99						
BDE 100						
BDE 116						
BDE 118						
BDE 119						
BDE 138						
BDE 153						
BDE 154						
BDE 155						
BDE 156						
BDE 181						
BDE 183						
BDE 190						
BDE 191						
BDE 196						
BDE 197						
BDE 203						
BDE 205						
BDE 206						
BDE 207						
BDE 208						
BDE 209						

(Any additional data/information should be added here.)

Appendix B: Description, Storage, Use, and Reporting Instructions for Marine Sediment XIII (QA05SED13)

**NIST Intercomparison Exercise Program for
Organic Contaminants in the Marine Environment**

NIST QA Program

**Intercomparison Exercise: Marine Sediment XIII
Description of Materials and Instructions**

Intercomparison Exercise Materials:

QA05SED13 (Marine Sediment XIII)

Each of the three jars contains approximately 21 g (wet basis) of Marine Sediment XII. This wetted sediment was prepared from material that was collected from a harbor area in the northeastern section of the US coast and then freeze-dried, ground sieved, and radiation-sterilized. This material has not been enriched or spiked. Each 2-oz clear glass jar has a Teflon-lined screw cap and is labeled with an individual jar number as well as the above name.

It is requested that three concurrent analyses of SRM 1941b Organics in Marine Sediment are also performed. This material can be obtained from the NIST Standard Reference Materials Program (\$524/50 g (dry-mass basis) (phone: 301/975-6776; fax: 301/948-3730). See the following link for information on ordering on-line:
https://srmors.nist.gov/view_detail.cfm?srm=1941B.

Storage of Materials:

Marine Sediment Material. This Marine Sediment XIII material should be stored in the dark at temperatures of -15 °C or lower. If only a portion of the contents of a jar is used, that jar should be tightly closed immediately after removal of a subsample to preserve the integrity of the remaining material for later analysis.

Instructions for Use:

You are to analyze Marine Sediment XIII and SRM 1941b, using **your** laboratory's and/or program's analytical protocols, for the concentrations (mass/mass [dry-mass basis]) of the 26 polycyclic aromatic hydrocarbon (PAH) compounds, 25 chlorinated pesticides, 25 polychlorinated biphenyl (PCB) congeners, and 34 polybrominated diphenyl ether (PBDE) congeners² of interest in the current NIST Intercomparison Exercise Program for Organic Contaminants in the Marine Environment. These compounds are listed in Table 1.

²If your laboratory is not analyzing samples for all chemical classes, you are expected to submit results only for those compounds currently being determined in your laboratory.

The percentage of water in Sediment XIII should be determined so that the results can be reported on a dry basis. You should have received sufficient material so that you can perform separate determinations for the water content if you do not dry your sediment samples prior to analysis. In addition, the percentage of total organic carbon should be determined in Sediment XIII and SRM 1941b.

The amount of material used for each analysis should correspond to the amount (wet basis) of marine sediment that you would typically analyze as prescribed in your protocols. Prior to removing an aliquot of Sediment XIII, you should thaw the sample in the jar and then **stir or otherwise mix it thoroughly**.

You should analyze three samples of Marine Sediment XIII and at least one or more samples of SRM 1941b in three different batches using your protocol for marine sediment samples. Specifically, we are asking that you analyze one sample of Sediment XIII and one sample of SRM 1941b with one batch of laboratory samples; analyze a second sample of each material with another batch; and the third sample with yet another batch. This will allow a more realistic assessment of laboratory precision over a longer term than the assessment obtained when a laboratory places all three samples in the same extraction and cleanup batch and the resulting extracts are analyzed using the same calibration curve, etc.

Reporting of Results:

Please report one result, as if three figures were significant, for each of the requested analytes in each of the three replicates of the Marine Sediment XIII and of SRM 1941b. Report results in units of ng/g **dry-mass** basis. Report the date of measurement of each sample in the requested m/d/y format. Also, report the results of your percentage water determinations of Marine Sediment XIII.

We recognize that the reported concentrations for some of the requested determinands will probably include concentrations of compounds reported to coelute with the determinand of interest with methods commonly in use in environmental laboratories. Please note at the bottom of your table of reported results if any coelution qualifiers are applicable to your data. Please note that any changes you make to the column or row headings **within** the tables will **not** be seen by the coordinators because only the table entries and comments at the bottom of the tables are automatically transferred to the exercise database.

We prefer that concentration values be reported for each analyte determined. If the measured concentration is below your typical reporting concentration for an analyte in a particular matrix, you can report the number and list the appropriate detection limit, quantification limit, etc. at the bottom of the data table. However, if you need to report non-numerical data please use the following conventions:

NA	"Not analyzed", "not determined"
<"value"	"Less than specified concentration", e.g., <8 ng/g

Other	"Other"; add note of explanation at end of data table, e.g., interference
DL	"Below detection limit" may be used, however, <"value" is preferable

Do not use negative numbers or parentheses to indicate "less than detection limits".

The attached file is an EXCEL file, SED13.xls. If you have any software/hardware conversion problems, please contact Michele Schantz. The data file templates also include places for you to list the surrogate/internal standards and type of calibration curve used, and to provide a brief description of the analyses. Please **do not** add spaces before entering numbers in the table cells and enter them as "numbers" not as "labels". Please **do not** insert any columns or rows **within** the table in the data file. If you wish to include additional data and/or other information or comments, you may add it to the bottom of the data table in the diskette file or send it in hard copy. A printout of the data file format is shown in Table 2.

Submit your results by **December 15, 2005** as an attached file via e-mail to:

E-mail:
michele.schantz@nist.gov

Further Information:

If you need further information, please contact Michele at the following address or phone numbers:

Michele M. Schantz
NIST
100 Bureau Drive Stop 8392
Gaithersburg, MD 20899-8392

Phone: (301)975-3106
FAX: (301)977-0685

Table 1: Analytes of Interest in NIST Intercomparison Exercise Program for Organic Contaminants in the Marine Environment

Chlorinated Pesticides

hexachlorobenzene	2,4'-DDE
alpha-HCH (alpha-BHC)	4,4'-DDE
beta-HCH (beta-BHC)	2,4'-DDD
gamma-HCH (gamma-BHC, Lindane)	4,4'-DDD
heptachlor	2,4'-DDT
heptachlor epoxide	4,4'-DDT
<i>cis</i> -chlordane (alpha-chlordane)	chlorpyrifos
<i>trans</i> -chlordane (gamma-chlordane)	aldrin
oxychlordane	dieldrin
<i>cis</i> -nonachlor	endrin
<i>trans</i> -nonachlor	endosulfan I
mirex	endosulfan II
	endosulfan sulfate

Polychlorinated Biphenyl Congeners

<i>PCB No.</i>	<i>Compound Name</i>
8	2,4'-dichlorobiphenyl
18	2,2',5-trichlorobiphenyl
28	2,4,4'-trichlorobiphenyl
31	2,4',5-trichlorobiphenyl
44	2,2',3,5'-tetrachlorobiphenyl
49	2,2',4,5'-tetrachlorobiphenyl
52	2,2',5,5'-tetrachlorobiphenyl
66	2,3',4,4'-tetrachlorobiphenyl
95	2,2',3,5',6-pentachlorobiphenyl
99	2,2',4,4',5-pentachlorobiphenyl
101	2,2',4,5,5'-pentachlorobiphenyl
105	2,3,3',4,4'-pentachlorobiphenyl
118	2,3',4,4',5-pentachlorobiphenyl
128	2,2',3,3',4,4'-hexachlorobiphenyl
138	2,2',3,4,4',5'-hexachlorobiphenyl
149	2,2',3,4',5',6-hexachlorobiphenyl
153	2,2',4,4',5,5'-hexachlorobiphenyl
156	2,3,3',4,4',5-hexachlorobiphenyl
170	2,2',3,3',4,4',5-heptachlorobiphenyl
180	2,2',3,4,4',5,5'-heptachlorobiphenyl
187	2,2',3,4',5,5',6-heptachlorobiphenyl
194	2,2',3,3',4,4',5,5'-octachlorobiphenyl
195	2,2',3,3',4,4',5,6-octachlorobiphenyl
206	2,2',3,3',4,4',5,5',6-nonachlorobiphenyl
209	decachlorobiphenyl

Table 1. (continued)

Polycyclic aromatic hydrocarbons (PAH)

naphthalene	benz[<i>a</i>]anthracene
2-methylnaphthalene	chrysene
1-methylnaphthalene	triphenylene
biphenyl	benzo[<i>b</i>]fluoranthene
2,6-dimethylnaphthalene	benzo[<i>j</i>]fluoranthene
acenaphthylene	benzo[<i>k</i>]fluoranthene
acenaphthene	benzo[<i>e</i>]pyrene
1,6,7-trimethylnaphthalene	benzo[<i>a</i>]pyrene
fluorene	perylene
phenanthrene	indeno[1,2,3- <i>cd</i>]pyrene
anthracene	dibenz[<i>a,h</i>]anthracene
1-methylphenanthrene	benzo[<i>ghi</i>]perylene
fluoranthene	
pyrene	

Polybrominated diphenyl ethers (PBDEs)

BDE 15 (4,4'-dibromo-)	BDE 138 (2,2',3,4,4',5'-hexabromo-)
BDE 17 (2,2',4-tribromo-)	BDE 153 (2,2',4,4',5,5'-hexabromo-)
BDE 25 (2,3',4-tribromo-)	BDE 154 (2,2',4,4',5,6'-hexabromo-)
BDE 28 (2,4,4'-tribromo-)	BDE 155 (2,2',4,4',6,6'-hexabromo-)
BDE 30 (2,4,6-tribromo-)	BDE 156 (2,3,3',4,4',5-hexabromo-)
BDE 33 (2',3,4-tribromo-)	BDE 181 (2,2',3,4,4',5,6-heptabromo-)
BDE 47 (2,2',4,4'-tetrabromo-)	BDE 183 (2,2',3,4,4',5',6-heptabromo-)
BDE 49 (2,2',4,5'-tetrabromo-)	BDE 190 (2,3,3',4,4',5,6-heptabromo-)
BDE 66 (2,3',4,4'-tetrabromo-)	BDE 191 (2,3,3',4,4',5,6'-heptabromo-)
BDE 71 (2,3',4',6-tetrabromo-)	BDE 196 (2,2',3,3',4,4',5,6'-octabromo-)
BDE 75 (2,4,4',6-tetrabromo-)	BDE 197 (2,2',3,3',4,4',6,6'-octabromo-)
BDE 85 (2,2',3,4,4'-pentabromo-)	BDE 203 (2,2',3,4,4',5,5',6-octabromo-)
BDE 99 (2,2',4,4',5-pentabromo-)	BDE 205 (2,3,3',4,4',5,5',6-octabromo-)
BDE 100 (2,2',4,4',6-pentabromo-)	BDE 206 (2,2',3,3',4,4',5,6,6'-nonabromo-)
BDE 116 (2,3,4,5,6-pentabromo-)	BDE 207 (2,2',3,3',4,4',5,6,6'-nonabromo-)
BDE 118 (2,3',4,4',5-pentabromo-)	BDE 208 (2,2',3,3',4,5,5',6,6'-nonabromo-)
BDE 119 (2,3',4,4',6-pentabromo-)	BDE 209 (decabromo-)

Table 2. Diskette Data File Format (File: SED13.*)

NIST Intercomparison Exercise Program for Organics in the Marine Environment
NIST QA Program
Sample: QA05SED13 - Marine Sediment XIII

Please fill in all blanks; Use requested units of concentration; Report results as if 3 figures were significant
DO NOT INSERT ROWS OR COLUMNS WITHIN THIS TABLE. DO NOT MOVE CELLS.

- If necessary, add additional data/information at the end of the table.
- Use one of the following if no concentration is reported for an analyte:
 NA = Not analyzed/determined; <"conc" = <detection limit conc.; Other = other, explain in a note at end of table
 (DL = "below detection limit" may be used, but <"conc", e.g., <8, is preferable.)
 Do not use parentheses or negative numbers to indicate "less than detection limit".

Reporting Date (m/d/y): _____
 Laboratory: _____
 Submitted by: _____

BRIEF DESCRIPTION OF PROCEDURES USED:

Approximate amount of sample extracted:

Sediment XIII _____ g, wet basis; SRM 1941b _____ g, dry basis

Method used for determining percentage water: _____

Were "wet" or "dry" samples extracted?

Sediment XIII _____ SRM 1941b _____

Extraction method: _____
 Extraction solvent: _____
 Extraction time: _____
 Extraction - other: _____

Sample extract cleanup method: _____

Analytical method used (e.g., GC-FID, GC-ECD):

	Analyt.	Instr.	Column Phase	Col. Length, m	Col. i.d., mm	Col. film thickness, µm
PAH	_____	_____	_____	_____	_____	_____
Pesticides	_____	_____	_____	_____	_____	_____
PCB Congeners	_____	_____	_____	_____	_____	_____
BDE Congeners	_____	_____	_____	_____	_____	_____

Method of quantitation (IS = internal standard, ES = external standard):

PAH	_____
Pesticides	_____
PCB Congeners	_____
BDE Congeners	_____

IF internal standard method was used, please complete the following section:

Identity of internal standards/surrogates used that were:

Added PRIOR to extraction of sample:

PAH	_____
Pesticides	_____
PCB Congeners	_____
BDE Congeners	_____

Added after extraction/cleanup and JUST PRIOR to chromatographic analysis:

PAH	_____
Pesticides	_____
PCB Congeners	_____
BDE Congeners	_____

Any others? Added at what point in analyses _____

PAH	_____
Pesticides	_____
PCB Congeners	_____
BDE Congeners	_____

IS/surrogate standards used for quantitation calculations were:

_____ those added prior to extraction
 _____ those added after extraction/cleanup and just prior to chromatographic analysis

If the IS/surrogates added after extraction/cleanup extraction were used for quantitation,
 were results corrected for percent recovery? _____

Percent recovery range:

PAH	_____
Pesticides	_____
PCB Congeners	_____
BDE Congeners	_____

Calibration Curve

	Points	Conc. Range	Analytes outside of calibration curve calibration range
PAH	_____	_____	_____
Pesticides	_____	_____	_____
PCB Congeners	_____	_____	_____
BDE Congeners	_____	_____	_____

Were PCB congeners separated from pesticides prior to GC? _____

Please note any differences in procedures used for SRM 1941b analyses from those for Marine Sediment XIII described above:

RESULTS:

PERCENT WATER & total organic carbon, TOC (List each result if determined more than once. Enter results as a number, for example 90.0. DO NOT change format of cell to percent.)

	Sediment XIII (percent)	Sediment XIII (percent)	Sediment XIII (percent)	SRM 1941b (percent)	SRM 1941b (percent)	SRM 1941b (percent)
Water	_____	_____	_____	_____	_____	_____
TOC	_____	_____	_____	_____	_____	_____
PAH ANALYSES	Sediment XIII Batch A Sample 1	Sediment XIII Batch B Sample 2	Sediment XIII Batch C Sample 3	SRM 1941b Batch A Sample 1	SRM 1941b Batch B Sample 2	SRM 1941b Batch C Sample 3
Analyst (Initials)	_____	_____	_____	_____	_____	_____
Date(s) of measurements (m/d/y)	_____	_____	_____	_____	_____	_____
Sample Jar number	_____	_____	_____	_____	_____	_____

	Sediment XIII Sample 1 (ng/g dry mass)	Sediment XIII Sample 2 (ng/g dry mass)	Sediment XIII Sample 3 (ng/g dry mass)	SRM 1941b Sample 1 (ng/g dry mass)	SRM 1941b Sample 2 (ng/g dry mass)	SRM 1941b Sample 3 (ng/g dry mass)
naphthalene	_____	_____	_____	_____	_____	_____
2-methylnaphthalene	_____	_____	_____	_____	_____	_____
1-methylnaphthalene	_____	_____	_____	_____	_____	_____
biphenyl	_____	_____	_____	_____	_____	_____
2,6-dimethylnaphthalene	_____	_____	_____	_____	_____	_____
acenaphthylene	_____	_____	_____	_____	_____	_____
acenaphthene	_____	_____	_____	_____	_____	_____
1,6,7-trimethylnaphthalene	_____	_____	_____	_____	_____	_____
fluorene	_____	_____	_____	_____	_____	_____
phenanthrene	_____	_____	_____	_____	_____	_____
anthracene	_____	_____	_____	_____	_____	_____
1-methylphenanthrene	_____	_____	_____	_____	_____	_____
fluoranthene	_____	_____	_____	_____	_____	_____
pyrene	_____	_____	_____	_____	_____	_____
benz[a]anthracene	_____	_____	_____	_____	_____	_____
chrysene	_____	_____	_____	_____	_____	_____
triphenylene	_____	_____	_____	_____	_____	_____
benzo[b]fluoranthene	_____	_____	_____	_____	_____	_____
benzo[j]fluoranthene	_____	_____	_____	_____	_____	_____
benzo[k]fluoranthene	_____	_____	_____	_____	_____	_____
benzo[e]pyrene	_____	_____	_____	_____	_____	_____
benzo[a]pyrene	_____	_____	_____	_____	_____	_____
perylene	_____	_____	_____	_____	_____	_____
indeno[1,2,3-cd]pyrene	_____	_____	_____	_____	_____	_____
dibenz[a,h]anthracene	_____	_____	_____	_____	_____	_____
benzo[ghi]perylene	_____	_____	_____	_____	_____	_____

PESTICIDE ANALYSES	Sediment XIII Batch A Sample 1	Sediment XIII Batch B Sample 2	Sediment XIII Batch C Sample 3	SRM 1941b Batch A Sample 1	SRM 1941b Batch B Sample 2	SRM 1941b Batch C Sample 3
Analyst (Initials)						
Date(s) of measurements (m/d/y)						
Sample Jar number						
	Sediment XIII Sample 1 (ng/g dry mass)	Sediment XIII Sample 2 (ng/g dry mass)	Sediment XIII Sample 3 (ng/g dry mass)	SRM 1941b Sample 1 (ng/g dry mass)	SRM 1941b Sample 2 (ng/g dry mass)	SRM 1941b Sample 3 (ng/g dry mass)
alpha-HCH (a-BHC)						
hexachlorobenzene						
gamma-HCH (g-BHC,lindane)						
beta-HCH (b-BHC)						
heptachlor						
aldrin						
heptachlor epoxide						
oxychlordane						
gamma-chlordane						
2,4'-DDE						
endosulfan I						
cis-chlordane (alpha-chlordane)						
trans-nonachlor						
dieldrin						
4,4'-DDE						
2,4'-DDD						
endrin						
endosulfan II						
4,4'-DDD						
2,4'-DDT						
cis-nonachlor						
4,4'-DDT						
mirex						
endosulfan sulfate						
chlorpyrifos						
PCB CONGENER ANALYSES	Sediment XIII Batch A Sample 1	Sediment XIII Batch B Sample 2	Sediment XIII Batch C Sample 3	SRM 1941b Batch A Sample 1	SRM 1941b Batch B Sample 2	SRM 1941b Batch C Sample 3
Analyst (Initials)						
Date(s) of measurements (m/d/y)						
Sample Jar number						
	Sediment XIII Sample 1 (ng/g dry mass)	Sediment XIII Sample 2 (ng/g dry mass)	Sediment XIII Sample 3 (ng/g dry mass)	SRM 1941b Sample 1 (ng/g dry mass)	SRM 1941b Sample 2 (ng/g dry mass)	SRM 1941b Sample 3 (ng/g dry mass)
PCB 8						
PCB 18						
PCB 28						
PCB 31						
PCB 44						
PCB 49						
PCB 52						
PCB 66						
PCB 95						
PCB 99						
PCB 101						
PCB 105						
PCB 118						
PCB 128						
PCB 138						
PCB 149						
PCB 153						
PCB 156						
PCB 170						
PCB 180						
PCB 187						
PCB 194						
PCB 195						
PCB 206						
PCB 209						

BDE CONGENER ANALYSES	Sediment XIII Batch A Sample 1	Sediment XIII Batch B Sample 2	Sediment XIII Batch C Sample 3	SRM 1941b Batch A Sample 1	SRM 1941b Batch B Sample 2	SRM 1941b Batch C Sample 3
Analyst (Initials)	_____	_____	_____	_____	_____	_____
Date(s) of measurements (m/d/y)	_____	_____	_____	_____	_____	_____
Sample Jar number	_____	_____	_____			
	Sediment XIII Sample 1 (ng/g dry mass)	Sediment XIII Sample 2 (ng/g dry mass)	Sediment XIII Sample 3 (ng/g dry mass)	SRM 1941b Sample 1 (ng/g dry mass)	SRM 1941b Sample 2 (ng/g dry mass)	SRM 1941b Sample 3 (ng/g dry mass)
BDE 15	_____	_____	_____	_____	_____	_____
BDE 17	_____	_____	_____	_____	_____	_____
BDE 25	_____	_____	_____	_____	_____	_____
BDE 28	_____	_____	_____	_____	_____	_____
BDE 30	_____	_____	_____	_____	_____	_____
BDE 33	_____	_____	_____	_____	_____	_____
BDE 47	_____	_____	_____	_____	_____	_____
BDE 49	_____	_____	_____	_____	_____	_____
BDE 66	_____	_____	_____	_____	_____	_____
BDE 71	_____	_____	_____	_____	_____	_____
BDE 75	_____	_____	_____	_____	_____	_____
BDE 85	_____	_____	_____	_____	_____	_____
BDE 99	_____	_____	_____	_____	_____	_____
BDE 100	_____	_____	_____	_____	_____	_____
BDE 116	_____	_____	_____	_____	_____	_____
BDE 118	_____	_____	_____	_____	_____	_____
BDE 119	_____	_____	_____	_____	_____	_____
BDE 138	_____	_____	_____	_____	_____	_____
BDE 153	_____	_____	_____	_____	_____	_____
BDE 154	_____	_____	_____	_____	_____	_____
BDE 155	_____	_____	_____	_____	_____	_____
BDE 156	_____	_____	_____	_____	_____	_____
BDE 181	_____	_____	_____	_____	_____	_____
BDE 183	_____	_____	_____	_____	_____	_____
BDE 190	_____	_____	_____	_____	_____	_____
BDE 191	_____	_____	_____	_____	_____	_____
BDE 196	_____	_____	_____	_____	_____	_____
BDE 197	_____	_____	_____	_____	_____	_____
BDE 203	_____	_____	_____	_____	_____	_____
BDE 205	_____	_____	_____	_____	_____	_____
BDE 206	_____	_____	_____	_____	_____	_____
BDE 207	_____	_____	_____	_____	_____	_____
BDE 208	_____	_____	_____	_____	_____	_____
BDE 209	_____	_____	_____	_____	_____	_____

(Any additional data/information should be added here.)

**Appendix C: Laboratory Notes Accompanying Data,
Mussel Tissue XII**

Lab	Additional notes for Mussel Tissue XII					
1a		Tissue XII (percent)	Tissue XII (percent)	Tissue XII (percent)	SRM 2977 (percent)	SRM 2977 (percent)
	% water	6.28	5.34	6.44	4.71	4.65
		Tissue XII Sample 1	Tissue XII Sample 2	Tissue XII Sample 3	SRM 2977 Sample 1	SRM 2977 Sample 2
		(ng/g dry mas	1g/g dry mas	1g/g dry mas	1g/g dry mas	1g/g dry mas
		131	124	129	88.7	84.4
	chrysene/triphenylene					
	dibenz[a,h + a,c]anthracene	6.33	6.59	6.61	2.25	2.13
	PCB 138/163	84.9	82.5	81.3	11.9	12.5
	PCB 153/132	121	119	119	13.8	14.7
1c	* BDEs determined by GC-MS NCL. 13C BDE was in sample as internal standard and interferes with determination of native.					
4		Tissue XII Sample 1	Tissue XII Sample 2	Tissue XII Sample 3	SRM 2977 Sample 1	SRM 2977 Sample 2
	Percent water	8.0	7.8	7.8	12.8	13.0
		Tissue XII Sample 1	Tissue XII Sample 2	Tissue XII Sample 3	SRM 2977 Sample 1	SRM 2977 Sample 2
		(ng/g dry mas	1g/g dry mas	1g/g dry mas	1g/g dry mas	1g/g dry mas
	dibenzothiophene	34.3	33.1	32.9	30.5	30.3
	retene	10.5	10.3	10.5	3.88	3.30
	nonachlor III	<1.00	<0.397	<0.609	<0.690	<0.647
	PCB 17	8.66	7.94	8.37	2.15	2.23
	PDB 33	5.95	6.21	6.27	1.77	1.64
	PDB 70	62.0	60.2	59.6	3.57	3.40
	PDB 74	31.7	30.6	30.9	2.73	2.62
	PCB 82	10.4	9.94	10.0	0.667	0.542
	PCB 87	39.9	37.9	37.8	1.89	1.56
	PCB 110	81.7	80.0	79.3	6.65	6.24
	PCB 151	15.8	15.8	15.5	3.17	3.16
	PCB 158	9.47	9.10	9.15	<0.689	<0.646
	PCB 171	4.09	4.44	4.19	<0.692	<0.649
	PCB 177	9.71	9.87	9.72	1.95	1.97
	PCB 183	11.3	11.2	11.1	0.703	0.678
	PCB 191	<1.01	<0.400	<0.613	<0.695	<0.652
	PCB 199	<0.759	<0.300	<0.460	0.671	0.714
	PCB 205	<1.01	<0.398	<0.611	<0.693	<0.650
	PCB 208	<1.01	<0.400	<0.613	<0.695	<0.652
	Notes: I - cannot be reported due to analytical interference which is not apparent in Tissue XII. Chrysene includes triphenylene; BkF includes BjF; dibenz[a,h]anthracene includes dibenz[a,c]anthracene. PCB 101 includes PCB 90; PCB 138 includes PCB 163 and PCB 164; PCB 153 includes PCB 132; PCB 187 includes PCB 159 and PCB 182.					
5	INT = interference					
7	1. For 3rd PT sample (Tissue XII), residue of Na2SO4 or water may be present in the extract. 2. PCB 101 and PCB 90 are coeluted.					

8	<p>All results are in wet weight as %moisture analysis was not performed on the tissue unknown sample or SRM.</p> <p>Only a duplicate analysis was performed on the measurements for the unknown sample due to limited unknown sample volume for PAH, PCB and BDE analyses.</p> <p>Only a duplicate analysis was performed on the SRM due to limited SRM volume for PAH, PCB and BDD analyses.</p> <p>Reporting limit for BDE's are based on laboratory background levels and not calibration curve range.</p> <p>PCB Coelutions:</p> <p>PCB-8/PCB-5 PCB-43/PCB-49 PCB-52/PCB-73 PCB-66/PCB-80 PCB-89/PCB-90/PCB-101 PCB-93/PCB-95 PCB-105/PCB-127 PCB-106/PCB-118 PCB-138/PCB-163/PCB-164 PCB-139/PCB-149 PCB-170/PCB-190 PCB-182/PCB-187</p> <p>BDE Coelutions:</p> <p>BDE-17/BDE-25 BDE-28/BDE-33 BDE-119/BDE-120 BDE-198/BDE-203</p>
9	<p>The analyst notes that the following data are estimates due to chromatographic interferents:</p> <p>SRM 1 - PCB 180 Tissue XII Sample 1 - PCB 138 SRM 2 - PCB 138 & PCB 180 SRM 3 - PCB 180</p>
10	<p>PLEASE NOTE: Samples highlighted with color signify co-eluting congeners/compounds:</p> <p>PCB : PCB 28+31 PAH : Chrysene + Triphenylene PESTICIDE: 2,4' DDD + endrin</p> <p>Only completed duplicate of SRM 2977</p>
11	<p>PCB Co-eluters: PCB-18/30, PCB-28/20/21/33, PCB-44/47/65, PCB-49/69, PCB-52/43/73, PCB-99/83, PCB-101/90/113, PCB-128/166, PCB-138/163/129/160, PCB-149/147, PCB-153/168, PCB-156/157, PCB-180/193</p>
12	<p>NA = not analyzed</p> <p>"other" = congener co-elutes as follows:</p> <p>PCB co-elutions: PCB 18/30, 20/28, 44/47/65, 95/100/93/102/98, 99/83, 101/90/113, 128/166, 138/163/129/160, 149/147, 153/168, 156/157, 180/193</p> <p>BDE co-elutions: BDE 17/25, 28/33, 119/120, 138/166</p> <p>PAH co-elutions: triphenylene/chrysene, benzo[b]fluoranthene/benzo[j]fluoranthene, 1,2,6-trimethynaphthalene/1,2,7-trimethynaphthalene/1,6,7-trimethynaphthalene/2,3,5-trimethynaphthalene, dibenz[a,h]anthracene/dibenz[a,c]anthracene</p>

Appendix D: Laboratory Notes Accompanying Data, Marine Sediment XIII

Lab	Additional notes for Sediment XIII						
1a	Sediment XIII	Sediment XIII	Sediment XIII	SRM 1941b	SRM 1941b	SRM 1941b	
	Sample 1	Sample 2	Sample 3	Sample 1	Sample 2	Sample 3	
	(ng/g dry mass)	(ng/g dry mass)	(ng/g dry mass)	(ng/g dry mass)	(ng/g dry mass)	(ng/g dry mass)	
	chrysene/triphenylene	409	406	416	445	441	417
	dibenz[a,h + a,c]anthracene	87.3	86.8	87.4	84.7	87.5	86.3
	PCB 138/163	4.89	4.89	4.96	4.91	4.74	4.73
	PCB 153/132	6.56	6.41	5.53	6.27	6.53	6.56
1c	* BDEs determined by GC-MS NCI. 13C BDE was in sample as internal standard and interferes with determination of native.						
2	The SRM we used for the sediment analysis was SRM 1944.						
	The sample was extracted dry						
	About 0.3 g were extracted						
	RESULTS: (for SRM 1944)						
	PAH ANALYSES						
		SRM 1944	SRM 1944	SRM 1944			
		Batch A	Batch B	Batch C			
		Sample 1	Sample 2	Sample 3			
	Analyst (Initials)	JGL	JGL	JGL			
	Date(s) of measurements (m/d/y)	7/15/2005	7/22/2005	7/29/2005			
		(ng/g dry mass)	(ng/g dry mass)	(ng/g dry mass)			
	naphthalene	1620	1605	1510			
	2-methylnaphthalene	911	882	949			
	1-methylnaphthalene	586	568	571			
	biphenyl	257	259	271			
	2,6-dimethylnaphthalene	285	339	610			
	acenaphthylene	421	424	479			
	acenaphthene	490	494	486			
	1,6,7-trimethylnaphthalene	NA	NA	NA			
	fluorene	987	963	937			
	phenanthrene	5690	5870	6060			
	anthracene	1550	1510	1640			
	1-methylphenanthrene	1740	1690	1720			
	fluoranthene	10300	10100	10000			
	pyrene	9810	10900	10400			
	benz[a]anthracene	5080	5170	5410			
	chrysene	5590	5580	5440			
	triphenylene	NA	NA	NA			
	benzo[b]fluoranthene	5400	6350	3750			
benzo[j]fluoranthene	NA	NA	NA				
benzo[k]fluoranthene	2260	2080	2580				
benzo[e]pyrene	3750	3420	3280				
benzo[a]pyrene	4440	4560	3710				
perylene	1110	1220	1110				
indeno[1,2,3-cd]pyrene	2830	2510	3130				
dibenz[a,h]anthracene	458	359	456				
benzo[ghi]perylene	2930	2770	3240				

4		Sediment XIII	Sediment XIII	Sediment XIII	SRM 1941b		
		Batch A	Batch B	Batch C	Batch A		
		Sample 1	Sample 2	Sample 3	Sample 1		
	Analyst (Initials)	DarB	DarB	DarB	DarB		
	Date(s) of measurements (m/d/y)	7/19/2005	7/19/2005	7/19/2005	7/19/2005		
	Sample Jar number	110	134	158			
	dibenzothiophene	59.9	57.2	58.9	60.2		
	retene	23.7	20.5	21.5	25.0		
		Sediment XIII	Sediment XIII	Sediment XIII	SRM 1941b	SRM 1941b	SRM 1941b
		Batch A	Batch B	Batch C	Batch A	Batch B	Batch C
		Sample 1	Sample 2	Sample 3	Sample 1	Sample 2	Sample 3
	Analyst (Initials)	RHB	RHB	RHB	RHB	RHB	RHB
	Date(s) of measurements (m/d/y)	12/5/2005	12/5/2005	12/5/2005	12/5/2005	12/5/2005	12/5/2005
	Sample Jar number	110	134	158			
nonachlor III	<0.548	<0.479	<0.433	<0.474	<0.651	<0.505	
PCB 17	1.31	1.34	1.31	1.37	1.19	1.36	
PCB 33	2.39	2.37	2.48	2.50	2.22	2.78	
PCB 70	5.77	5.66	5.80	5.69	5.17	5.56	
PCB 74	2.23	2.20	2.22	2.29	2.08	2.21	
PCB 82	0.551	0.515	0.547	0.520	0.479	0.567	
PCB 87	1.74	1.59	1.90	1.55	1.45	1.57	
PCB 110	6.72	6.48	7.11	6.51	5.88	6.40	
PCB 151	1.26	1.48	1.28	1.20	1.07	1.20	
PCB 158	<0.546	0.549	0.574	<0.472	<0.650	<0.504	
PCB 171	<0.549	0.575	<0.434	<0.475	<0.653	<0.507	
PCB 177	0.928	1.42	0.948	0.912	0.807	0.912	
PCB 183	0.902	1.40	0.916	0.862	0.760	0.877	
PCB 191	<0.551	<0.482	<0.436	<0.477	<0.656	<0.509	
PCB 199	1.50	2.06	1.45	1.49	1.33	1.50	
PCB 205	<0.549	<0.480	<0.434	<0.475	<0.654	<0.507	
PCB 208	1.15	1.21	1.16	1.21	0.950	1.15	
Notes: Chrysene includes triphenylene; BkF includes BjF; dibenz[a,h]anthracene includes dibenz[a,c]anthracene. PCB 101 includes PCB 90; PCB 138 includes PCB 163 and PCB 164; PCB 153 includes PCB 132; PCB 187 includes PCB 159 and PCB 182							
7	1. PCB 101 and PCB 90 are coeluted. 2. The glassware containing the extract of the 2nd SRM sample was broken during operation and the sample was completely lost.						
8	Only a duplicate analysis was performed on the measurements for the unknown sample due laboratory oversight for PCB and BDE analyses. Only a duplicate analysis was performed on the SRM due to laboratory oversight for PAH, PCB and BDD analyses. Reporting limit for BDE's are based on laboratory background levels and not calibration curve range. BDE-209 could not be postively detected due to the loss of 13C-BDE-209 during analytical process possibly associated with matrix. PCB Coelutions: PCB-8/PCB-5 PCB-43/PCB-49 PCB-52/PCB-73 PCB-66/PCB-80 PCB-89/PCB-90/PCB-101 PCB-93/PCB-95 PCB-105/PCB-127 PCB-106/PCB-118 PCB-138/PCB-163/PCB-164 PCB-139/PCB-149 PCB-170/PCB-190 PCB-182/PCB-187 BDE Coelutions: BDE-17/BDE-25 BDE-28/BDE-33 BDE-119/BDE-120 BDE-198/BDE-203						

10	<p>PLEASE NOTE: Samples highlighted with color signify co-eluting congeners/compounds:</p> <p>PCB : PCB 28+31</p> <p>PAH : Chrysene + Triphenylene</p> <p>PESTICIDE: 2,4' DDD + endrin</p> <p>Only completed duplicate of SRM 1941b, and trial 1 of Marine Sediment XIII was LOST due to evaporation at GCMS stage.</p> <p>OTHER= LOST sample see note above</p>
11	<p>PCB Co-eluters: PCB-18/30, PCB-28/20/21/33, PCB-44/47/65, PCB-49/69, PCB-52/43/73, PCB-99/83, PCB-101/90/113, PCB-128/166, PCB Co-eluters: PCB-138/163/129/160, B93PCB-149/147, PCB-153/168, PCB-156/157, PCB-180/193</p>
12	<p>NA = not analyzed</p> <p>"other" = congener co-elutes as follows:</p> <p>PCB co-elutions: PCB 18/30, 20/28, 44/47/65, 95/100/93/102/98, 99/83, 101/90/113, 128/166, 138/163/129/160, 149/147, 153/168, 156/157, 180/193</p> <p>BDE co-elutions: BDE 17/25, 28/33, 119/120, 138/166</p> <p>PAH co-elutions: triphenylene/chrysene, benzo[b]fluoranthene/benzo[j]fluoranthene, 1,2,6-trimethynaphthalene/1,2,7-trimethynaphthalene/1,6,7-trimethynaphthalene/2,3,5-trimethynaphthalene, dibenz[a,h]anthracene/dibenz[a,c]anthracene</p>

Appendix E: Laboratory Methods Used, Mussel Tissue XII

Lab #	Reported	g extracted QA05TIS12	g extracted SRM 2977	% TEO Determination	Extraction Method	Extraction Solvent	Extraction Time	Extraction other
1a	11/28/2005	1 dry	3 dry	gravimetric using 100 µL of extract	PFE	dichloromethane	3 cycles each 5 min	temp = 100 °C; pressure 2000 psi; 3 static cycles / sample
1c	2/22/2006	2 dry	2 dry	gravimetric using portion of extract	PFE	dichloromethane	3 cycles each 5 min	temp = 100 °C; pressure 2000 psi; 3 static cycles / sample
3	12/15/2005	2 dry	2 dry	not analyzed	Sonication	dichloromethane	3 x 2.0 min each	
4	1/4/2006	0.9 dry	0.9 dry	gravimetric	PFE	dichloromethane	approx. 16 min	temp = 100 °C; pressure 2000 psi
5	1/10/2006	1.2 dry	1.2 dry	gravimetric using 1/6 of extract	polytron	dichloromethane (3 x 100 mL)	3 x 2.0 min each	filtered on glass fiber-filters (1.2 µm pore size) during extraction
6	1/13/2006	1 dry	1 dry	gravimetric using 1/10 of extract	microscale extraction 3570	acetone: dichloromethane	24 h	solvent changes at specified time intervals
7	1/13/2006	1 dry	3 dry	gravimetric using 1/10 of extract	Soxhlet EPA 3540	acetone:hexane (1:1, volume fraction)	24 h	
8	1/16/2006	0.5 dry	1 dry		Soxhlet	dichloromethane for PAHs; toluene for PCBs and PBDEs	16 h	
9	1/17/2006	2 dry	2 dry	gravimetric using portion of extract	PFE	dichloromethane	2 x 2 min high speed extractions followed by 30 min on shaker table	
10	1/25/2006	1 dry	1 dry	gravimetric	Soxhlet	dichloromethane	18 h	
11	2/6/2006	1.5 dry	1.5 dry	gravimetric using portion of extract	PFE	dichloromethane		
12	2/6/2006	1 dry	1 dry	gravimetric	Soxhlet	dichloromethane	16 h	

Lab #	Sample extract cleanup method	PCBs and Pesticides Separated?	Method of quantitation
1a	size exclusion chromatography (SEC); silica solid phase extraction (SPE) column, condition and elute with 15 mL of 10 % dichloromethane in hexane	no	IS
1c	1.8 g alumina column (5% deactivated) with 9 mL 35 % dichloromethane in hexane; 0.5 g aminopropyl SPE column using 10 mL of 10% dichloromethane in hexane	yes	IS
3	silica gel; activated copper; sulfuric acid	no	IS
4	Gravity flow column with silica gel and neutral alumina, followed by HPLC-SEC to elute fraction containing analytes of interest	no	IS
5	SEC; fractionated on 7.4% deactivated silica gel	yes	IS
6	SEC; silica gel cartridges	no	IS
7	SEC for PAH, PCB, Pesticide, and PBDE; Florisil for PCB, Pesticide, and PBDE	no	IS
8	silica gel only for PAH; silica gel and acid alumina for PCB and PBDE	some	IS
9	alumina gravity column; HPLC-SEC fractionation	no	IS
10	alumina for PAHs; Florisil with petroleum ether for PCBs and with 1:1 dichloromethane:petroleum ether for pesticides	yes	IS
11	alumina added to PFE extraction cells prior to extraction; SEC; acid/base silica column	yes	IS/ES
12	PAHs - SEC, silica; pesticides - SEC, Florisil; PCBs and PBDEs - SEC, Florisil, acid/base silica, alumina	no	IS

Lab #	PAHs		Calibration Curve		
	Instrument	Phase	Dimensions	# points	range
1a	GC/MS	HP-5MS	30m x 0.25 mm, 0.25µm filr	5	5 ng - 1500 ng extracted
1c	GC/MS	DB-XLB	30m x 0.18 mm, 0.18µm filr	6	1.72 ng/g - 983 ng/g
3	GC/MS	RTX-5 Sil MS	30m x 0.28 mm, 0.25µm filr	5	5 ng/mL - 2000 ng/mL
4	GC/MS	DB-5	60m x 0.25 mm, 0.25µm filr	7	0.011 ng/µL - 1.1 ng/µL
5	GC/MS	HP-5MS	30m x 0.25 mm, 0.25µm filr	5	10 ng/mL - 500 ng/mL
6	GC/MS	RTX-5	60m x 0.25 mm, 0.25µm filr	7	10 ppb - 10000 ppb
7	GC/MS	DB-XLB	60m x 0.25 mm, 0.25µm filr	1	50 ppb
8	GC/MS	DB-5MS	30m x 0.25 mm, 0.25µm filr	5	25 µg - 2500 µg
9	GC/MS	DB-5	60m x 0.25 mm, 0.25µm filr	8	0.005 ng/µL - 10 ng/µL
10	GC/MS	DB-5	30m x 0.25 mm, 0.25µm filr	5	5 ng - 100 ng
12	GC/MS	DB-5	30m x 0.25 mm, 0.25µm filr	5	50 ng/mL - 5000 ng/mL

Lab #	PBDE's			Calibration Curve	
	Instrument	Phase	Dimensions	# points	range
1c	GC/MS NCI	DB-XLB	30m x 0.18 mm, 0.18µm film	5	0.07 ng/g - 386 ng/g
4	GC/MS	DB-5	60m x 0.25 mm, 0.25µm film	4	0.0025 ng/µL - 1 ng/µL
7	GC/HRMS	DB-5MS	30m x 0.25 mm, 0.25µm film	7	0.05 ppb - 100 ppb
8	HRGC/MS	DB-5HT	30m x 0.25 mm, 0.1µm film	5	20 pg - 500000 pg
12	GC/HRMS	DB-5HT	30m x 0.25 mm, 0.1µm film	5	1 ng/mL - 2500 ng/mL

Lab #	PCBs					PESTICIDES				
	Instrument	Phase	Dimensions	Calibration Curve # points	range	Instrument	Phase	Dimensions	Calibration Curve # points	range
1a	GC/MS	HP-5MS	30m x 0.25 mm, 0.25um	5	5 ng - 300 ng extracted	GC/MS	HP-5MS	30m x 0.25 mm, 0.25um	5	5 ng - 300 ng extracted
1c	GC/MS	DB-XLB	30m x 0.18 mm, 0.18um film	6	0.29 ng/g - 3930 ng/g	GC/MS	DB-XLB	30m x 0.18 mm, 0.18um film	6	0.65 ng/g - 244 ng/g
3	GC/MS	RTX-5	60m x 0.25 mm, 0.25um film	5	2 ng/mL - 100 ng/mL	GC/MS	RTX-5	60m x 0.25 mm, 0.25um film	5	2 ng/mL - 100 ng/mL
4	GC/MS	DB-5	60m x 0.25 mm, 0.25um film	6	0.0012 ng/ μ L - 0.32 ng/ μ L	GC/MS	DB-5	60m x 0.25 mm, 0.25um film	6	0.0012 ng/ μ L - 0.32 ng/ μ L
5	GC-ECD	HP-5MS/DB-XLB	30m x 0.25 mm, 0.25um film	5	5 ng/mL - 50 ng/mL	GC-ECD	HP-5MS/DB-XLB	30m x 0.25 mm, 0.25um film	5	5 ng/mL - 50 ng/mL
6	GC/MS	RTX-5	60m x 0.25 mm, 0.25um film	8	0.25 ppb - 400 ppb	GC-ECD	RTX-5	60m x 0.25 mm, 0.25um film	7	0.2 ppb - 200 ppb
7	GC/MS	DB-XLB	30m x 0.18 mm, 0.18um film	1	20 ppb	GC/MS	DB-XLB	30m x 0.18 mm, 0.18um film	1	20 ppb
8	HRGC/MS	DB-5	60m x 0.32 mm, 0.25um 60m x 0.25 mm, 0.25um	5	20 pg - 20000 pg			60m x 0.25 mm, 0.25um film		
9	GC-ECD	DB-5	60m x 0.25 mm, 0.25um film	7	0.001 ng/ μ L - 1 ng/ μ L	GC-ECD	DB-5	60m x 0.25 mm, 0.25um film	7	0.001 ng/ μ L - 1 ng/ μ L
10	GC-ECD	DB-5	60m x 0.25 mm, 0.25um film	5	1 ng - 43 ng	GC-ECD	DB-5	60m x 0.25 mm, 0.25um film	5	1 ng - 10 ng
11	GC/HRMS	SPB-Octyl	30m x 0.25 mm, 0.25um film	6	0.2 ng/mL - 2000 ng/mL					
12	GC/HRMS	SPB-Octyl	30m x 0.25 mm, 0.1um film	5	1 ng/mL - 2000 ng/mL	GC/HRMS	DB-5	60m x 0.25 mm, 0.1um film	5	10 ng/mL - 4000 ng/mL

Lab #	IS/surrogate added prior to extraction	PAHs			corrected for recovery?	others?
		Used?	added prior to analysis	Used?		
1a	deuterated naphthalene, biphenyl, acenaphthene, phenanthrene, fluoranthene, pyrene, B[a]A, B[a]P, perylene, B[ghi]P, DB[a,h]A	x				
1c	deuterated naphthalene, biphenyl, acenaphthene, phenanthrene, fluoranthene, pyrene, B[a]A, B[a]P, perylene, B[ghi]P, DB[a,h]A	x				
3	deuterated naphthalene, phenanthrene, and chrysene		deuterated fluorene, acenaphthene, B[a]P	x	n	
4	deuterated naphthalene, acenaphthene, B[a]P	x	hexamethylbenzene			deuterated phenanthrene prior to clean-up
5	deuterated naphthalene, acenaphthene, phenanthrene, fluoranthene, chrysene, B[a]P		deuterated fluorene, pyrene, perylene	x	n	
6	deuterated 2-methyl naphthalene, pyrene, B[b]F		deuterated naphthalene, acenaphthene, phenanthrene, chrysene, perylene	x	n	
7	17 deuterated PAHs	x				
8	deuterated naphthalene, acenaphthylene, fluorene, phenanthrene, pyrene, B[a]A, chrysene, B[b]F, B[k]F, B[a]P, perylene, I[1,2,3-cd]P, DB[a,h]A, B[ghi]P	x	deuterated 2-methylnaphthalene, anthracene, terphenyl, B[c]P			
9	deuterated naphthalene, acenaphthene, phenanthrene, B[a]P		deuterated acenaphthylene and chrysene	x	n	
10	surrogates- deuterated naphthalene, acenaphthene, phenanthrene, chrysene, perylene		IS- deuterated fluorene, anthracene, fluoranthene, B[a]P	x		
12	deuterated naphthalene, 2-methylnaphthalene, biphenyl, 2,6-dimethylnaphthalene, acenaphthylene, phenanthrene, fluoranthene, B[a]A, chrysene, B[b,k]F, B[a]P, perylene, DB[a,h]A, I[1,2,3-cd]P, B[ghi]P	x	deuterated acenaphthene, pyrene, B[c]P used to quantify labeled surrogates			

Lab #	IS/surrogate added prior to extraction	PBDEs			corrected for recovery?	others?
		Used?	added prior to analysis	Used?		
1c	13C-PCB 194, 13C-t-chlodane, endosulfan-d4	x				
4	PCB 103	x	tetrachloro-o-xylene			tetrachloro-m-xylene prior to clean-up
7	13C-BDEs (3, 5, 28, 47, 99, 100, 118, 153, 183)	x				
8	13C-BDEs (28, 47, 99, 100, 153, 154, 183, 209)	x	13 C-PCBs (138, 202)			
12	13C BDEs (15, 28, 47, 77, 99, 100, 126, 153, 154, 183, 209)	x	13C PCBs (52, 138) used to quantify labeled surrogates			13C BDE 139 prior to clean-up

Lab #	IS/surrogate added prior to extraction	PCBs				corrected for recovery?	others?
		Used?	added prior to analysis	Used?			
1a	PCB 103 and PCB 198	x					
1c	13C-PCB 28, 52, 118, 153, 180, 194, 206; deuterated 4,4'-DDE, 4,4'-DDD, 4,4'-DDT	x					
3	2',3,5-Trichlorobiphenyl, 2,2',4,6,6'-Pentachlorobiphenyl, 2,3,3',4,5,5',6-Heptachlorobiphenyl		3-Chlorobiphenyl, 2,3,3',4,4',5,5',6-Octachlorobiphenyl (13C12 labelled)	x	n		
4	PCB 103	x	tetrachloro-o-xylene				tetrachloro-m-xylene prior to clean-up
5	g-chlordane, PCB 103, PCB 198		IS - 4,4'-dibromooctafluorobiphenyl	x	n		
6	13C-PCB 19 and 202		13C-PCB 15 and 180	x	n		
7	246/246-HBB	x	34/34-TBB				
8	13C-PCB 3, 15, 28, 77, 81, 105, 114, 118, 123, 126, 156, 157, 167, 169, 170, 180, 189, 194, 208, 209	x	13C-PCB 52, 101, 202				
9	Cl3(34); Cl6(152)		CL5(96); Cl6(161)	x	n		
10	Surrogates - PCB 14, 64, 166		IS - PCB 30 and 204	x			
11	13C-PCB 1,3,4,19,15,54,104,37,155,81,77,123,118,188,114,105,126,202,156,157,169,208,189,205,206,209	x	13C-PCB 9,52,138, 194				13C-PCB 28, 111,178 prior to clean-up
12	13C-PCBs 4,15,19,37,54,77,81,104,105,114,118,123,126,155,156,157,167,169,170,180,188,189,202,205,206,208,209	x	13C-PCBs 9,52,101,138,194, used to quantify labelled surrogates only.				13C-PCBs 28,111,178, used as cleanup standards.

Lab #	IS/surrogate added prior to extraction	Pesticides				corrected for	
		Used?	added prior to analysis	Used?	recovery?	others?	
1a	13C- lindane, trans-nonachlor, 4,4'-DDE, 4,4'-DDT	x					
1c	13C-PCB 52, trans-chlordane, PCB 118, PCB 153; deuterated endosulfan I	x					
3	13C-gamma-BHC and 4,4'-DDT		3-Chlorobiphenyl, 2,3,3',4,4',5,5',6-Octachlorobiphenyl (13C12 labelled)	x	n	tetrachloro-m-xylene prior to clean-up	
4	PCB 103	x	tetrachloro-o-xylene				
5	g-chlordane, PCB 103, PCB 198		IS - 4,4'-dibromooctafluorobiphenyl	x	n		
6	TMX and DCB		PCB 192	x	n		
7	246/246-HBB	x	34/34-TBB				
8							
9	Cl3(34); Cl6(152)		CL5(96); Cl6(161)	x	n		
10			IS - PCB 30 and 204	x			
11							
12	13C-HCB, b-HCH, g-HCH, d-HCH, Heptachlor, Aldrin, Oxychlordane, t-Chlordane, t-Nonachlor, c-nonachlor, DDE, DDT, Mirex, Heptachlor-epoxide, Dieldrin, Endrin, Endosulfan-I, Endosulfan-II	x	13C-PCBs 52, 138, 153, used to quantify labelled surrogates only.				

Appendix F: Laboratory Methods Used, Marine Sediment XIII

Lab #	Reported	g extracted QA05SED13	g extracted SRM 1941b	% water Determination	Extraction Method	Extraction Solvent	Extraction Time	Extraction other
1a	11/28/2005	9 wet	3 dry	freeze-dry until constant mass	PFE	dichloromethane	3 cycles each 5 min	temp = 100 °C; pressure 2000 psi; 3 static cycles / sample
1c	2/22/2006	5 wet	5 dry	freeze-drying	PFE	dichloromethane	3 cycles each 5 min	temp = 100 °C; pressure 2000 psi; 3 static cycles / sample
2	12/14/2005	10 wet	SRM 1944	gravimetric - oven 100 °C for 24 h	PFE	dichloromethane	10 min	temp = 100 °C; pressure 2000 psi
3	12/15/2005	10 wet	5 dry	standard method 2540G	Sonication	dichloromethane	3 x 2.0 min each	
4	1/4/2006	1 wet	0.5 dry	oven 120 °C overnight	PFE	dichloromethane	approx. 16 min	temp = 100 °C; pressure 2000 psi solvent changes at specified time intervals
6	1/13/2006	3 wet	3 dry	assumed 100%	microscale extraction 3570	acetone: dichloromethane	24 h	
7	1/13/2006	5 wet	5 dry	oven 105 °C for 24 h	Soxhlet EPA 3540	acetone:hexane (1:1, volume fraction) dichloromethane for PAHs; toluene for PCBs and PBBDEs	24 h	
8	1/16/2006	1 -2 wet	0.5 -1 dry	ASTM D2216-98	Soxhlet		16 h	
10	1/25/2006	1 wet	1 dry	oven 100 °C for 24 h	Soxhlet	dichloromethane	18 h	
11	2/6/2006	10 wet	1 dry	oven overnight	PFE	dichloromethane		
12	2/6/2006	8.5 wet	5 dry	gravimetric	Soxhlet	dichloromethane	16 h	

Lab #	Sample extract cleanup method	PCBs and Pesticides Separated?	Method of quantitation
1a	silica solid phase extraction (SPE) column; condition and elute with 15 mL of 10 % dichloromethane in hexane	no	IS
1c	Size exclusion chromatography (SEC); 1.8 g alumina column (5% deactivated) with 9 mL 35 % dichloromethane in hexane	yes	IS
2	silica gel, alumina, copper chromatography; SEC HPLC cleanup	no	IS
3	silica gel; activated copper; sulfuric acid	no	IS
4	Gravity flow column with silica gel and neutral alumina, followed by HPLC-SEC to elute fraction containing analytes of interest	no	IS
6	silica cartridge; PCB wer acid cleaned post silica	no	IS
7	SEC for PAH, PCB, Pesticide, and PBDE; Florisil for PCB, Pesticide, and PBDE	no	IS
8	silica gel only for PAH; silica gel and acid alumina for PCB and PBDE	some	IS
9	alumina gravity column; HPLC-SEC fractionation	no	IS
10	alumina for PAHs; Florisil with petroleum ether for PCBs and with 1:1 dichloromethane:petroleum ether for pesticides	yes	IS
11	alumina added to PFE extraction cells prior to extraction; SEC; acid/base silica column	yes	IS/ES
12	PAHs -silica; pesticides - Florisil; PCBs and PBDEs - Florisil, acid/base silica, alumina	no	IS

Lab #	PAHs				Calibration Curve	
	Instrument	Phase	Dimensions	# points	range	
1a	GC/MS	HP-5MS	30m x 0.25 mm, 0.25µm film	5	5 ng - 1500 ng extracted	
1c	GC/MS	DB-XLB	30m x 0.18 mm, 0.18µm film	6	1.72 ng/g - 983 ng/g	
2	GC/MS	5% phenyl	25m x 0.2 mm, 0.33µm film	5	6 ng/mL - 1800 ng/mL	
3	GC/MS	RTX-5 Sil MS	30m x 0.28 mm, 0.25µm film	5	5 ng/mL - 2000 ng/mL	
4	GC/MS	DB-5	60m x 0.25 mm, 0.25µm film	7	0.015 ng/µL - 10 ng/µL	
6	GC/MS	RTX-5	60m x 0.25 mm, 0.25µm film	7	10 ppb - 10000 ppb	
7	GC/MS	DB-XLB	60m x 0.25 mm, 0.25µm film	1	50 ppb	
8	GC/MS	DB-5MS	30m x 0.25 mm, 0.25µm film	5	25 µg - 2500 µg	
10	GC/MS	DB-5	30m x 0.25 mm, 0.25µm film	5	5 ng - 100 ng	
12	GC/MS	DB-5	30m x 0.25 mm, 0.25µm film	5	50 ng/mL - 5000 ng/mL	

Lab #	PBDEs				Calibration Curve	
	Instrument	Phase	Dimensions	# points	range	
1c	GC/MS NCI	DB-XLB	30m x 0.18 mm, 0.18µm film	5	0.07 ng/g - 386 ng/g	
4	GC/MS	DB-5	60m x 0.25 mm, 0.25µm film	4	0.0025 ng/µL - 1 ng/µL	
7	GC/HRMS	DB-5MS	30m x 0.25 mm, 0.25µm film	7	0.05 ppb - 100 ppb	
8	HRGC/MS	DB-5HT	30m x 0.25 mm, 0.1µm film	5	20 pg - 500000 pg	
12	GC/HRMS	DB-5HT	30m x 0.25 mm, 0.1µm film	5	1 ng/mL - 2500 ng/mL	

Lab #	PCBs			Calibration Curve			PESTICIDES			Calibration Curve		
	Instrument	Phase	Dimensions	# points	range		Instrument	Phase	Dimensions	# points	range	
1a	GC/MS	HP-5MS	30m x 0.25 mm, 0.25um	5	5 ng - 300 ng extracted		GC/MS	HP-5MS	30m x 0.25 mm, 0.25um	5	5 ng - 300 ng extracted	
1c	GC/MS	DB-XLB	30m x 0.18 mm, 0.18um film	6	0.29 ng/g - 3930 ng/g		GC/MS	DB-XLB	30m x 0.18 mm, 0.18um film	6	0.65 ng/g - 244 ng/g	
3	GC/MS	RTX-5	60m x 0.25 mm, 0.25um film	5	2 ng/mL - 100 ng/mL 0.0012 ng/ μ L - 0.32		GC/MS	RTX-5	60m x 0.25 mm, 0.25um film	5	2 ng/mL - 100 ng/mL 0.0012 ng/ μ L - 0.32	
4	GC/MS	DB-5	60m x 0.25 mm, 0.25um film	6	ng/ μ L		GC/MS	DB-5	60m x 0.25 mm, 0.25um film	6	ng/ μ L	
6	GC/MS	RTX-5	30m x 0.18 mm, 0.18um film	8	0.25 ppb - 400 ppb		GC-ECD	RTX-5	30m x 0.18 mm, 0.18um film	7	0.2 ppb - 200 ppb	
7	GC/MS	DB-XLB	60m x 0.32 mm, 0.25um film	5	20 pg - 20000 pg		GC/MS	DB-XLB	30m x 0.18 mm, 0.18um film	1	20 ppb	
8	HRGC/MS	DB-5	60m x 0.25 mm, 0.25um film	5	1 ng - 43 ng							
10	GC-ECD	DB-5	30m x 0.25 mm, 0.25um film	5	0.2 ng/mL - 2000 ng/mL		GC-ECD	DB-5	60m x 0.25 mm, 0.25um film	5	1 ng - 10 ng	
11	GC/HRMS	SPB-Octyl	30m x 0.25 mm, 0.1um film	6	1 ng/mL - 2000 ng/mL							
12	GC/HRMS	SPB-Octyl	30m x 0.25 mm, 0.1um film	5	1 ng/mL - 2000 ng/mL		GC/HRMS	DB-5	60m x 0.25 mm, 0.1um film	5	10 ng/mL - 4000 ng/mL	

Lab #	IS/surrogate added prior to extraction	PAHs			corrected for recovery?	others?
		Used?	added prior to analysis	Used?		
1a	deuterated naphthalene, biphenyl, acenaphthene, phenanthrene, fluoranthene, pyrene, B[a]A, B[a]P, perylene, B[ghi]P, DB[a,h]A	x				
1c	deuterated naphthalene, biphenyl, acenaphthene, phenanthrene, fluoranthene, pyrene, B[a]A, B[a]P, perylene, B[ghi]P, DB[a,h]A	x				
2	deuterated naphthalene, acenaphthene, chrysene, B[a]P, perylene	x	HMB			
3	deuterated naphthalene, phenanthrene, and chrysene		deuterated fluorene, acenaphthene, B[a]P	x	n	
4	deuterated naphthalene, acenaphthene, B[a]P	x	hexamethylbenzene			deuterated phenanthrene prior to clean-up
6	deuterated 2-methyl naphthalene, pyrene, B[b]F	x	deuterated naphthalene, acenaphthene, phenanthrene, chrysene, perylene		n	
7	17 deuterated PAHs	x				
8	deuterated naphthalene, acenaphthylene, acenaphthene, fluorene, phenanthrene, pyrene, B[a]A, chrysene, B[b]F, B[k]F, B[a]P, perylene, I[1,2,3-cd]P, DB[a,h]A, B[ghi]P	x	deuterated 2-methylnaphthalene, anthracene, terphenyl, B[c]P			
10	surrogates- deuterated naphthalene, acenaphthene, phenanthrene, chrysene, perylene		13c- deuterated fluorene, anthracene, fluoranthene, B[a]P	x		
12	deuterated naphthalene, 2-methylnaphthalene, biphenyl, 2,6-dimethylnaphthalene, acenaphthylene, phenanthrene, B[a]A, chrysene, B[b,k]F, B[a]P, perylene, DB[a,h]A, I[123-cd]P, B[ghi]P	x	deuterated acenaphthene, pyrene, B[c]P used to quantify labeled surrogates			

Lab #	IS/surrogate added prior to extraction	PBDEs			corrected for recovery?	others?
		Used?	added prior to analysis	Used?		
1c	13C-PCB 194, 13C-t-chlodane, endosulfan-d4	x				
4	PCB 103	x	tetrachloro-o-xylene			tetrachloro-m-xylene prior to clean-up
7	13C-BDEs (3, 5, 28, 47, 99, 100, 118, 153, 183)	x				
8	13C-BDEs (28, 47, 99, 100, 153, 154, 183, 209)	x	13 C-PCBs (138, 202)			
12	13C BDEs (15, 28, 47, 77, 99, 100, 126, 153, 154, 183, 209)	x	13C PCBs (52, 138) used to quantify labeled surrogates			13C BDE 139 prior to clean-up

Lab #	IS/surrogate added prior to extraction	Used?	added prior to analysis	Used?	corrected for recovery?	others?
1a	PCB 103 and PCB 198	x				
1c	13C-PCB 28, 52, 118, 153, 180, 194, 206; deuterated 4,4'-DDE, 4,4'-DDD, 4,4'-DDT	x				
3	2',3,5-Trichlorobiphenyl, 2,2',4,6,6'-Pentachlorobiphenyl, 2,3,3',4,5,5',6-Heptachlorobiphenyl		3-Chlorobiphenyl, 2,3,3',4,4',5,5',6-Octachlorobiphenyl (13C12 labelled)	x	n	tetrachloro-m-xylene prior to clean-up
4	PCB 103	x	tetrachloro-o-xylene			
6	13C-PCB 19 and 202	x	13C-PCB 15 and 180		n	
7	246/246-HBB	x	34/34-TBB			
8	13C-PCB 3, 15, 28, 77, 81, 105, 114, 118, 123, 126, 156, 157, 167, 169, 170, 180, 189, 194, 208, 209	x	13C-PCB 52, 101, 202			
10	Surrogates - PCB 14, 64, 166		IS - PCB 30 and 204	x		13C-PCB 28, 111,178 prior to clean-up
11	13C-PCB 1,3,4,19,15,54,104,37,155,81,77,123,118,188,114,105,126,202,156,157,169,208,189,205,206,209	x	13C-PCB 9,52,138, 194			
12	13C-PCBs 4,15,19,37,54,77,81,104,105,114,118,123,126,155,156,157,167,169,170,180,188,189,202,205,206,208,209	x	13C-PCBs 9,52,101,138,194, used to quantify labelled surrogates only.			13C-PCBs 28,111,178, used as cleanup standards.

Lab #	IS/surrogate added prior to extraction	Pesticides				corrected for	
		Used?	added prior to analysis	Used?	recovery?	others?	
1a	13C- lindane, trans-nonachlor, 4,4'-DDE, 4,4'-DDT	x					
1c	13C-PCB 52, trans-chlordane, PCB 118, PCB 153; deuterated endosulfan I	x					
3	13C-gamma-BHC and 4,4'-DDT		3-Chlorobiphenyl, 2,3,3',4,4',5,5',6-Octachlorobiphenyl (13C12 labelled)	x	n	tetrachloro-m-xylene prior to clean-up	
4	PCB 103	x	tetrachloro-o-xylene				
6	TMX and DCB	x	PCB 192		n		
7	246/246-HBB	x	34/34-TBB				
8							
10			IS - PCB 30 and 204	x			
11							
12	13C-HCB, b-HCH, g-HCH, d-HCH, Heptachlor, Aldrin, Oxychlordane, t-Chlordane, t-Nonachlor, c-nonachlor, DDE, DDT, Mirex, Heptachlor-epoxide, Dieldrin, Endrin, Endosulfan-I, Endosulfan-II	x	13C-PCBs 52, 138, 153, used to quantify labelled surrogates only.				

Appendix G: Charts of Mussel Tissue XII and SRM 2977 Results by Analyte

See Tables 2 through 9 for results reported as *<number*, detection limit, etc.

Charts for analytes with few reported numerical results are not included in this appendix.

Note: The numbers added to the charts are the values reported that are off the scale of the chart.

For Mussel Tissue XII plots:

Solid line: exercise assigned value

Dotted line: $z = \pm 1$, i. e., 25 % from assigned value

Dotted/dashed line: $z = \pm 2$, i. e., 50 % from assigned value

Dashed line: $z = \pm 3$, i. e., 75 % from assigned value

For SRM 2977 plots:

Solid line: material certified concentration or target value (see caption of each plot)

Dotted line: 95 % confidence interval (CI)


Dashed line: 30 % from 95 % confidence interval (CI)

naphthalene

Tissue XII (QA05TIS12)

Assigned value = 9.86 ng/g $s = 6.07$ ng/g 95% CL = 5.61 ng/g (dry basis)

Reported Results: 10 Quantitative Results: 10




Solid line : exercise assigned value (EA V); dotted line: $z=\pm 1$ (25% from EA V); dotted/dashed line: $z=\pm 2$ (50% from EA V); dashed line: $z=\pm 3$ (75% from EA V)

naphthalene

SRM 2977

Reference Value = 19 ± 5 ng/g (dry basis)

Reported Results: 10 Quantitative Results: 10




Solid line: value from Certificate of Analysis ; dotted line: 95% confidence limits; dashed line: 30% from 95% confidence limits

2-methylnaphthalene

Tissue XII (QA05TIS12)

Assigned value = 8.00 ng/g $s = 5.16$ ng/g 95% CL = 5.42 ng/g (dry basis)

Reported Results: 9 Quantitative Results: 8




Solid line : exercise assigned value (EAV); dotted line: ± 1 (25% from EAV); dotted/dashed line: ± 2 (50% from EAV); dashed line: ± 3 (75% from EAV)

2-methylnaphthalene

SRM 2977

Reference Value = 18 ± 5 ng/g (dry basis)

Reported Results: 9 Quantitative Results: 9




Solid line: value from Certificate of Analysis ; dotted line: 95% confidence limits; dashed line: 30% from 95% confidence limits

1-methylnaphthalene

Tissue XII (QA05TIS12)

Assigned value = 3.66 ng/g $s = 2.44$ ng/g 95% CL = 3.02 ng/g (dry basis)

Reported Results: 8 Quantitative Results: 7




Solid line : exercise assigned value (EAV); dotted line: ± 1 (25% from EAV); dotted/dashed line: ± 2 (50% from EAV); dashed line: ± 3 (75% from EAV)

1-methylnaphthalene

SRM 2977

Reference Value = 16 ± 5 ng/g (dry basis)

Reported Results: 8 Quantitative Results: 7




Solid line: value from Certificate of Analysis ; dotted line: 95% confidence limits; dashed line: 30% from 95% confidence limits

biphenyl**Tissue XII (QA05TIS12)**

Assigned value = 1.91 ng/g $s = 0.77$ ng/g 95% CL = 0.95 ng/g (dry basis)

Reported Results: 9 Quantitative Results: 7




Solid line : exercise assigned value (EAV); dotted line: ± 1 (25% from EAV); dotted/dashed line: ± 2 (50% from EAV); dashed line: ± 3 (75% from EAV)

biphenyl**SRM 2977**

Reference Value = 6.8 ± 0.6 ng/g (dry basis)

Reported Results: 9 Quantitative Results: 7




Solid line: value from Certificate of Analysis ; dotted line: 95% confidence limits; dashed line: 30% from 95% confidence limits

2,6-dimethylnaphthalene

Tissue XII (QA05TIS12)

Assigned value = 4.67 ng/g $s = 2.85$ ng/g 95% CL = 4.54 ng/g (dry basis)

Reported Results: 8 Quantitative Results: 4




Solid line : exercise assigned value (EAV); dotted line: ± 1 (25% from EAV); dotted/dashed line: ± 2 (50% from EAV); dashed line: ± 3 (75% from EAV)

2,6-dimethylnaphthalene

SRM 2977

Target Value = no target ng/g (dry basis)

Reported Results: 8 Quantitative Results: 8




Solid line: value from Certificate of Analysis ; dotted line: 95% confidence limits; dashed line: 30% from 95% confidence limits

acenaphthylene**Tissue XII (QA05TIS12)**

Assigned value = 3.72 ng/g $s = 2.40$ ng/g 95% CL = 2.01 ng/g (dry basis)

Reported Results: 11 Quantitative Results: 8




Solid line : exercise assigned value (EAV); dotted line: $z \pm 1$ (25% from EAV); dotted/dashed line: $z \pm 2$ (50% from EAV); dashed line: $z \pm 3$ (75% from EAV)

acenaphthylene**SRM 2977**

Target Value = no target ng/g (dry basis)

Reported Results: 11 Quantitative Results: 6




Solid line: value from Certificate of Analysis ; dotted line: 95% confidence limits; dashed line: 30% from 95% confidence limits

acenaphthene**Tissue XII (QA05TIS12)**

Assigned value = 2.93 ng/g $s = 2.21$ ng/g 95% CL = 2.74 ng/g (dry basis)

Reported Results: 11 Quantitative Results: 7




Solid line : exercise assigned value (EAV); dotted line: ± 1 (25% from EAV); dotted/dashed line: ± 2 (50% from EAV); dashed line: ± 3 (75% from EAV)

acenaphthene**SRM 2977**

Reference Value = 4.2 ± 0.4 ng/g (dry basis)

Reported Results: 11 Quantitative Results: 9




Solid line: value from Certificate of Analysis ; dotted line: 95% confidence limits; dashed line: 30% from 95% confidence limits

1,6,7-trimethylnaphthalene

Tissue XII (QA05TIS12)

Assigned value = 4.78 ng/g $s = 1.65$ ng/g 95% CL = 2.62 ng/g (dry basis)

Reported Results: 4 Quantitative Results: 4




Solid line : exercise assigned value (EAV); dotted line: ± 1 (25% from EAV); dotted/dashed line: ± 2 (50% from EAV); dashed line: ± 3 (75% from EAV)

1,6,7-trimethylnaphthalene

SRM 2977

Target Value = no target ng/g (dry basis)

Reported Results: 5 Quantitative Results: 5




Solid line: value from Certificate of Analysis ; dotted line: 95% confidence limits; dashed line: 30% from 95% confidence limits

fluorene

Tissue XII (QA05TIS12)

Assigned value = 3.64 ng/g $s = 1.64$ ng/g 95% CL = 1.52 ng/g (dry basis)

Reported Results: 11 Quantitative Results: 9




Solid line : exercise assigned value (EAV); dotted line: $z \pm 1$ (25% from EAV); dotted/dashed line: $z \pm 2$ (50% from EAV); dashed line: $z \pm 3$ (75% from EAV)

fluorene

SRM 2977

Certified Value = 10.24 ± 0.43 ng/g (dry basis)

Reported Results: 11 Quantitative Results: 11




Solid line: value from Certificate of Analysis ; dotted line: 95% confidence limits; dashed line: 30% from 95% confidence limits

phenanthrene

Tissue XII (QA05TIS12)

Assigned value = 88.7 ng/g $s = 17.9$ ng/g 95% CL = 13.8 ng/g (dry basis)

Reported Results: 11 Quantitative Results: 11




Solid line : exercise assigned value (EAV); dotted line: $z \pm 1$ (25% from EAV); dotted/dashed line: $z \pm 2$ (50% from EAV); dashed line: $z \pm 3$ (75% from EAV)

phenanthrene

SRM 2977

Certified Value = 35.1 \pm 3.8 ng/g (dry basis)

Reported Results: 11 Quantitative Results: 11




Solid line: value from Certificate of Analysis ; dotted line: 95% confidence limits; dashed line: 30% from 95% confidence limits

anthracene**Tissue XII (QA05TIS12)**

Assigned value = 5.79 ng/g $s = 3.24$ ng/g 95% CL = 2.70 ng/g (dry basis)

Reported Results: 11 Quantitative Results: 9




Solid line : exercise assigned value (EAV); dotted line: ± 1 (25% from EAV); dotted/dashed line: ± 2 (50% from EAV); dashed line: ± 3 (75% from EAV)

anthracene**SRM 2977**

Reference Value = 8 ± 4 ng/g (dry basis)

Reported Results: 11 Quantitative Results: 10




Solid line: value from Certificate of Analysis ; dotted line: 95% confidence limits; dashed line: 30% from 95% confidence limits

1-methylphenanthrene

Tissue XII (QA05TIS12)

Assigned value = 89.7 ng/g $s = 27.7$ ng/g 95% CL = 19.8 ng/g (dry basis)

Reported Results: 10 Quantitative Results: 10




Solid line : exercise assigned value (EAV); dotted line: ± 1 (25% from EAV); dotted/dashed line: ± 2 (50% from EAV); dashed line: ± 3 (75% from EAV)

1-methylphenanthrene

SRM 2977

Reference Value = 44 ± 2 ng/g (dry basis)

Reported Results: 10 Quantitative Results: 10




Solid line: value from Certificate of Analysis ; dotted line: 95% confidence limits; dashed line: 30% from 95% confidence limits

fluoranthene**Tissue XII (QA05TIS12)**

Assigned value = 133 ng/g $s = 30$ ng/g 95% CL = 23 ng/g (dry basis)

Reported Results: 11 Quantitative Results: 11




Solid line : exercise assigned value (EAV); dotted line: $z=\pm 1$ (25% from EAV); dotted/dashed line: $z=\pm 2$ (50% from EAV); dashed line: $z=\pm 3$ (75% from EAV)

fluoranthene**SRM 2977**

Certified Value = 38.7 ± 1.0 ng/g (dry basis)

Reported Results: 11 Quantitative Results: 11




Solid line: value from Certificate of Analysis ; dotted line: 95% confidence limits; dashed line: 30% from 95% confidence limits

pyrene**Tissue XII (QA05TIS12)**

Assigned value = 190 ng/g $s = 40$ ng/g 95% CL = 31 ng/g (dry basis)

Reported Results: 11 Quantitative Results: 11




Solid line : exercise assigned value (EAV); dotted line: $z=\pm 1$ (25% from EAV); dotted/dashed line: $z=\pm 2$ (50% from EAV); dashed line: $z=\pm 3$ (75% from EAV)

pyrene**SRM 2977**

Certified Value = 78.9 ± 3.5 ng/g (dry basis)

Reported Results: 11 Quantitative Results: 11




Solid line: value from Certificate of Analysis ; dotted line: 95% confidence limits; dashed line: 30% from 95% confidence limits

benz[a]anthracene**Tissue XII (QA05TIS12)**

Assigned value = 24.7 ng/g $s = 7.1$ ng/g 95% CL = 5.4 ng/g (dry basis)

Reported Results: 10 Quantitative Results: 10




Solid line : exercise assigned value (EAV); dotted line: $z \pm 1$ (25% from EAV); dotted/dashed line: $z \pm 2$ (50% from EAV); dashed line: $z \pm 3$ (75% from EAV)

benz[a]anthracene**SRM 2977**

Certified Value = 20.3 ± 0.8 ng/g (dry basis)

Reported Results: 10 Quantitative Results: 10




Solid line: value from Certificate of Analysis ; dotted line: 95% confidence limits; dashed line: 30% from 95% confidence limits

chrysene

Tissue XII (QA05TIS12)

Assigned value = 63.9 ng/g $s = 22.4$ ng/g 95% CL = 27.8 ng/g (dry basis)

Reported Results: 7 Quantitative Results: 7




Solid line : exercise assigned value (EAV); dotted line: ± 1 (25% from EAV); dotted/dashed line: ± 2 (50% from EAV); dashed line: ± 3 (75% from EAV)

chrysene

SRM 2977

Reference Value = 49 ± 2 ng/g (dry basis)


Reported Results: 10 Quantitative Results: 10



Solid line: value from Certificate of Analysis ; dotted line: 95% confidence limits; dashed line: 30% from 95% confidence limits

benzo[b]fluoranthene**Tissue XII (QA05TIS12)**Assigned value = 47.6 ng/g $s = 12.6$ ng/g 95% CL = 11.7 ng/g (dry basis)


Reported Results: 9 Quantitative Results: 9



Solid line : exercise assigned value (EAV); dotted line: $z \pm 1$ (25% from EAV); dotted/dashed line: $z \pm 2$ (50% from EAV); dashed line: $z \pm 3$ (75% from EAV)

benzo[b]fluoranthene**SRM 2977**Certified Value = 11.01 ± 0.28 ng/g (dry basis)


Reported Results: 10 Quantitative Results: 10



Solid line: value from Certificate of Analysis ; dotted line: 95% confidence limits; dashed line: 30% from 95% confidence limits

benzo[j]fluoranthene**Tissue XII (QA05TIS12)**


Assigned value = no target ng/g (dry basis)
Reported Results: 2 Quantitative Results: 2



Solid line : exercise assigned value (EAV); dotted line: $z \pm 1$ (25% from EAV); dotted/dashed line: $z \pm 2$ (50% from EAV); dashed line: $z \pm 3$ (75% from EAV)

benzo[j]fluoranthene**SRM 2977**


Reference Value = 4.6 ± 0.2 ng/g (dry basis)
Reported Results: 2 Quantitative Results: 2



Solid line: value from Certificate of Analysis ; dotted line: 95% confidence limits; dashed line: 30% from 95% confidence limits

benzo[k]fluoranthene**Tissue XII (QA05TIS12)**Assigned value = 16.3 ng/g $s = 2.4$ ng/g 95% CL = 3.0 ng/g (dry basis)


Reported Results: 9 Quantitative Results: 9



Solid line : exercise assigned value (EAV); dotted line: $z \pm 1$ (25% from EAV); dotted/dashed line: $z \pm 2$ (50% from EAV); dashed line: $z \pm 3$ (75% from EAV)

benzo[k]fluoranthene**SRM 2977**Reference Value = 4 ± 1 ng/g (dry basis)


Reported Results: 10 Quantitative Results: 10



Solid line: value from Certificate of Analysis ; dotted line: 95% confidence limits; dashed line: 30% from 95% confidence limits

benzo[e]pyrene**Tissue XII (QA05TIS12)**Assigned value = 74.7 ng/g $s = 14.6$ ng/g 95% CL = 10.5 ng/g (dry basis)


Reported Results: 11 Quantitative Results: 11



Solid line : exercise assigned value (EAV); dotted line: $z \pm 1$ (25% from EAV); dotted/dashed line: $z \pm 2$ (50% from EAV); dashed line: $z \pm 3$ (75% from EAV)

benzo[e]pyrene**SRM 2977**Certified Value = 13.1 ± 1.1 ng/g (dry basis)

Reported Results: 11 Quantitative Results: 10




Solid line: value from Certificate of Analysis ; dotted line: 95% confidence limits; dashed line: 30% from 95% confidence limits

benzo[a]pyrene**Tissue XII (QA05TIS12)**

Assigned value = 7.25 ng/g $s = 2.65$ ng/g 95% CL = 2.04 ng/g (dry basis)

Reported Results: 11 Quantitative Results: 10




Solid line : exercise assigned value (EAV); dotted line: $z \pm 1$ (25% from EAV); dotted/dashed line: $z \pm 2$ (50% from EAV); dashed line: $z \pm 3$ (75% from EAV)

benzo[a]pyrene**SRM 2977**

Certified Value = 8.35 ± 0.72 ng/g (dry basis)

Reported Results: 11 Quantitative Results: 10




Solid line: value from Certificate of Analysis ; dotted line: 95% confidence limits; dashed line: 30% from 95% confidence limits

perylene**Tissue XII (QA05TIS12)**

Assigned value = 3.51 ng/g $s = 1.44$ ng/g 95% CL = 2.30 ng/g (dry basis)

Reported Results: 8 Quantitative Results: 6




Solid line : exercise assigned value (EAV); dotted line: ± 1 (25% from EAV); dotted/dashed line: ± 2 (50% from EAV); dashed line: ± 3 (75% from EAV)

perylene**SRM 2977**

Certified Value = 3.50 ± 0.76 ng/g (dry basis)


Reported Results: 8 Quantitative Results: 6



Solid line: value from Certificate of Analysis ; dotted line: 95% confidence limits; dashed line: 30% from 95% confidence limits

indeno[1,2,3-cd]pyrene**Tissue XII (QA05TIS12)**Assigned value = 15.1 ng/g $s = 6.4$ ng/g 95% CL = 4.9 ng/g (dry basis)


Reported Results: 10 Quantitative Results: 10



Solid line : exercise assigned value (EAV); dotted line: $z \pm 1$ (25% from EAV); dotted/dashed line: $z \pm 2$ (50% from EAV); dashed line: $z \pm 3$ (75% from EAV)

indeno[1,2,3-cd]pyrene**SRM 2977**Certified Value = 4.84 ± 0.81 ng/g (dry basis)


Reported Results: 10 Quantitative Results: 9



Solid line: value from Certificate of Analysis ; dotted line: 95% confidence limits; dashed line: 30% from 95% confidence limits

dibenz[a,h]anthracene**Tissue XII (QA05TIS12)**


Assigned value = no target ng/g (dry basis)
Reported Results: 7 Quantitative Results: 3



Solid line : exercise assigned value (EAV); dotted line: $z=\pm 1$ (25% from EAV); dotted/dashed line: $z=\pm 2$ (50% from EAV); dashed line: $z=\pm 3$ (75% from EAV)

dibenz[a,h]anthracene**SRM 2977**


Certified Value = 1.41 ± 0.19 ng/g (dry basis)
Reported Results: 9 Quantitative Results: 6



Solid line: value from Certificate of Analysis ; dotted line: 95% confidence limits; dashed line: 30% from 95% confidence limits

benzo[ghi]perylene**Tissue XII (QA05TIS12)**Assigned value = 24.6 ng/g $s = 6.4$ ng/g 95% CL = 4.9 ng/g (dry basis)


Reported Results: 11 Quantitative Results: 11



Solid line : exercise assigned value (EAV); dotted line: $z=\pm 1$ (25% from EAV); dotted/dashed line: $z=\pm 2$ (50% from EAV); dashed line: $z=\pm 3$ (75% from EAV)

benzo[ghi]perylene**SRM 2977**Certified Value = 9.53 ± 0.43 ng/g (dry basis)


Reported Results: 11 Quantitative Results: 10



Solid line: value from Certificate of Analysis ; dotted line: 95% confidence limits; dashed line: 30% from 95% confidence limits

alpha-HCH (a-BHC)**Tissue XII (QA05TIS12)**


Assigned value = no target ng/g (dry basis)
Reported Results: 9 Quantitative Results: 2



Solid line : exercise assigned value (EAV); dotted line: $z=\pm 1$ (25% from EAV); dotted/dashed line: $z=\pm 2$ (50% from EAV); dashed line: $z=\pm 3$ (75% from EAV)

alpha-HCH (a-BHC)**SRM 2977**

Target Value = no target ng/g (dry basis)
Reported Results: 8 Quantitative Results: 1




Solid line: value from Certificate of Analysis ; dotted line: 95% confidence limits; dashed line: 30% from 95% confidence limits

hexachlorobenzene

Tissue XII (QA05TIS12)

Assigned value = no target ng/g (dry basis)
Reported Results: 9 Quantitative Results: 2




Solid line : exercise assigned value (EAV); dotted line: $z=\pm 1$ (25% from EAV); dotted/dashed line: $z=\pm 2$ (50% from EAV); dashed line: $z=\pm 3$ (75% from EAV)

hexachlorobenzene

SRM 2977


Target Value = no target ng/g (dry basis)
Reported Results: 8 Quantitative Results: 1



Solid line: value from Certificate of Analysis ; dotted line: 95% confidence limits; dashed line: 30% from 95% confidence limits

gamma-HCH (g-BHC,lindane)**Tissue XII (QA05TIS12)**


Assigned value = no target ng/g (dry basis)
Reported Results: 10 Quantitative Results: 3



Solid line : exercise assigned value (EAV); dotted line: $z=\pm 1$ (25% from EAV); dotted/dashed line: $z=\pm 2$ (50% from EAV); dashed line: $z=\pm 3$ (75% from EAV)

gamma-HCH (g-BHC,lindane)**SRM 2977**


Target Value = no target ng/g (dry basis)
Reported Results: 9 Quantitative Results: 2



Solid line: value from Certificate of Analysis ; dotted line: 95% confidence limits; dashed line: 30% from 95% confidence limits

beta-HCH (b-BHC)**Tissue XII (QA05TIS12)**


Assigned value = no target ng/g (dry basis)
Reported Results: 8 Quantitative Results: 3



Solid line : exercise assigned value (EAV); dotted line: $z=\pm 1$ (25% from EAV); dotted/dashed line: $z=\pm 2$ (50% from EAV); dashed line: $z=\pm 3$ (75% from EAV)

beta-HCH (b-BHC)**SRM 2977**

Target Value = no target ng/g (dry basis)
Reported Results: 7 Quantitative Results: 3




Solid line: value from Certificate of Analysis ; dotted line: 95% confidence limits; dashed line: 30% from 95% confidence limits

heptachlor

Tissue XII (QA05TIS12)

Assigned value = no target ng/g (dry basis)
Reported Results: 10 Quantitative Results: 2




Solid line : exercise assigned value (EAV); dotted line: $z=\pm 1$ (25% from EAV); dotted/dashed line: $z=\pm 2$ (50% from EAV); dashed line: $z=\pm 3$ (75% from EAV)

heptachlor

SRM 2977


Target Value = no target ng/g (dry basis)
Reported Results: 9 Quantitative Results: 1



Solid line: value from Certificate of Analysis ; dotted line: 95% confidence limits; dashed line: 30% from 95% confidence limits

aldrin**Tissue XII (QA05TIS12)**


Assigned value = no target ng/g (dry basis)
Reported Results: 10 Quantitative Results: 2



Solid line : exercise assigned value (EAV); dotted line: $z=\pm 1$ (25% from EAV); dotted/dashed line: $z=\pm 2$ (50% from EAV); dashed line: $z=\pm 3$ (75% from EAV)

aldrin**SRM 2977**

Target Value = no target ng/g (dry basis)
Reported Results: 9 Quantitative Results: 1




Solid line: value from Certificate of Analysis ; dotted line: 95% confidence limits; dashed line: 30% from 95% confidence limits

heptachlor epoxide

Tissue XII (QA05TIS12)

Assigned value = no target ng/g (dry basis)
Reported Results: 10 Quantitative Results: 2




Solid line : exercise assigned value (EAV); dotted line: $z=\pm 1$ (25% from EAV); dotted/dashed line: $z=\pm 2$ (50% from EAV); dashed line: $z=\pm 3$ (75% from EAV)

heptachlor epoxide

SRM 2977

Target Value = no target ng/g (dry basis)
Reported Results: 9 Quantitative Results: 1




Solid line: value from Certificate of Analysis ; dotted line: 95% confidence limits; dashed line: 30% from 95% confidence limits

oxychlordan

Tissue XII (QA05TIS12)

Assigned value = no target ng/g (dry basis)
Reported Results: 9 Quantitative Results: 4




Solid line : exercise assigned value (EAV); dotted line: $z=\pm 1$ (25% from EAV); dotted/dashed line: $z=\pm 2$ (50% from EAV); dashed line: $z=\pm 3$ (75% from EAV)

oxychlordan

SRM 2977

Target Value = no target ng/g (dry basis)
Reported Results: 8 Quantitative Results: 2




Solid line: value from Certificate of Analysis ; dotted line: 95% confidence limits; dashed line: 30% from 95% confidence limits

gamma-chlordane

Tissue XII (QA05TIS12)

Assigned value = 7.45 ng/g $s = 1.00$ ng/g 95% CL = 0.77 ng/g (dry basis)

Reported Results: 10 Quantitative Results: 9




Solid line : exercise assigned value (EAV); dotted line: ± 1 (25% from EAV); dotted/dashed line: ± 2 (50% from EAV); dashed line: ± 3 (75% from EAV)

gamma-chlordane

SRM 2977

Target Value = no target ng/g (dry basis)

Reported Results: 9 Quantitative Results: 5




Solid line: value from Certificate of Analysis ; dotted line: 95% confidence limits; dashed line: 30% from 95% confidence limits

2,4'-DDE

Tissue XII (QA05TIS12)

Assigned value = no target ng/g (dry basis)
Reported Results: 10 Quantitative Results: 3




Solid line : exercise assigned value (EAV); dotted line: $z \pm 1$ (25% from EAV); dotted/dashed line: $z \pm 2$ (50% from EAV); dashed line: $z \pm 3$ (75% from EAV)

2,4'-DDE

SRM 2977


Target Value = no target ng/g (dry basis)
Reported Results: 9 Quantitative Results: 2



Solid line: value from Certificate of Analysis ; dotted line: 95% confidence limits; dashed line: 30% from 95% confidence limits

endosulfan I**Tissue XII (QA05TIS12)**


Assigned value = no target ng/g (dry basis)
Reported Results: 10 Quantitative Results: 2



Solid line : exercise assigned value (EAV); dotted line: $z=\pm 1$ (25% from EAV); dotted/dashed line: $z=\pm 2$ (50% from EAV); dashed line: $z=\pm 3$ (75% from EAV)

endosulfan I**SRM 2977**


Target Value = no target ng/g (dry basis)
Reported Results: 9 Quantitative Results: 2



Solid line: value from Certificate of Analysis ; dotted line: 95% confidence limits; dashed line: 30% from 95% confidence limits

cis-chlordane (alpha-chlordane)**Tissue XII (QA05TIS12)**Assigned value = 12.1 ng/g $s = 3.4$ ng/g 95% CL = 2.8 ng/g (dry basis)


Reported Results: 10 Quantitative Results: 10



Solid line : exercise assigned value (EA V); dotted line: $z=\pm 1$ (25% from EA V); dotted/dashed line: $z=\pm 2$ (50% from EA V); dashed line: $z=\pm 3$ (75% from EA V)

cis-chlordane (alpha-chlordane)**SRM 2977**Certified Value = 1.42 ± 0.13 ng/g (dry basis)

Reported Results: 9 Quantitative Results: 6




Solid line: value from Certificate of Analysis ; dotted line: 95% confidence limits; dashed line: 30% from 95% confidence limits

trans-nonachlor**Tissue XII (QA05TIS12)**

Assigned value = 9.00 ng/g $s = 1.24$ ng/g 95% CL = 1.04 ng/g (dry basis)

Reported Results: 10 Quantitative Results: 10




Solid line : exercise assigned value (EAV); dotted line: ± 1 (25% from EAV); dotted/dashed line: ± 2 (50% from EAV); dashed line: ± 3 (75% from EAV)

trans-nonachlor**SRM 2977**

Certified Value = 1.43 ± 0.10 ng/g (dry basis)

Reported Results: 9 Quantitative Results: 6




Solid line: value from Certificate of Analysis ; dotted line: 95% confidence limits; dashed line: 30% from 95% confidence limits

dieldrin**Tissue XII (QA05TIS12)**

Assigned value = 6.70 ng/g $s = 4.42$ ng/g 95% CL = 4.64 ng/g (dry basis)

Reported Results: 10 Quantitative Results: 7




Solid line : exercise assigned value (EAV); dotted line: ± 1 (25% from EAV); dotted/dashed line: ± 2 (50% from EAV); dashed line: ± 3 (75% from EAV)

dieldrin**SRM 2977**

Certified Value = 6.04 ± 0.52 ng/g (dry basis)

Reported Results: 9 Quantitative Results: 6




Solid line: value from Certificate of Analysis ; dotted line: 95% confidence limits; dashed line: 30% from 95% confidence limits

4,4'-DDE

Tissue XII (QA05TIS12)

Assigned value = 33.9 ng/g $s = 7.6$ ng/g 95% CL = 7.0 ng/g (dry basis)

Reported Results: 10 Quantitative Results: 10




Solid line : exercise assigned value (EAV); dotted line: $z \pm 1$ (25% from EAV); dotted/dashed line: $z \pm 2$ (50% from EAV); dashed line: $z \pm 3$ (75% from EAV)

4,4'-DDE

SRM 2977

Certified Value = 12.5 ± 1.6 ng/g (dry basis)

Reported Results: 9 Quantitative Results: 9




Solid line: value from Certificate of Analysis ; dotted line: 95% confidence limits; dashed line: 30% from 95% confidence limits

2,4'-DDD

Tissue XII (QA05TIS12)

Assigned value = 8.04 ng/g $s = 2.86$ ng/g 95% CL = 2.20 ng/g (dry basis)

Reported Results: 9 Quantitative Results: 9




Solid line : exercise assigned value (EAV); dotted line: ± 1 (25% from EAV); dotted/dashed line: ± 2 (50% from EAV); dashed line: ± 3 (75% from EAV)

2,4'-DDD

SRM 2977

Certified Value = 3.32 ± 0.29 ng/g (dry basis)

Reported Results: 7 Quantitative Results: 6




Solid line: value from Certificate of Analysis ; dotted line: 95% confidence limits; dashed line: 30% from 95% confidence limits

4,4'-DDD

Tissue XII (QA05TIS12)

Assigned value = 21.7 ng/g $s = 10.8$ ng/g 95% CL = 10.0 ng/g (dry basis)

Reported Results: 10 Quantitative Results: 9




Solid line : exercise assigned value (EAV); dotted line: ± 1 (25% from EAV); dotted/dashed line: ± 2 (50% from EAV); dashed line: ± 3 (75% from EAV)

4,4'-DDD

SRM 2977

Certified Value = 4.30 ± 0.38 ng/g (dry basis)

Reported Results: 8 Quantitative Results: 8




Solid line: value from Certificate of Analysis ; dotted line: 95% confidence limits; dashed line: 30% from 95% confidence limits

2,4'-DDT

Tissue XII (QA05TIS12)

Assigned value = no target ng/g (dry basis)
Reported Results: 10 Quantitative Results: 3




Solid line : exercise assigned value (EAV); dotted line: $z=\pm 1$ (25% from EAV); dotted/dashed line: $z=\pm 2$ (50% from EAV); dashed line: $z=\pm 3$ (75% from EAV)

2,4'-DDT

SRM 2977

Target Value = no target ng/g (dry basis)
Reported Results: 9 Quantitative Results: 1




Solid line: value from Certificate of Analysis ; dotted line: 95% confidence limits; dashed line: 30% from 95% confidence limits

cis-nonachlor**Tissue XII (QA05TIS12)**

Assigned value = 4.27 ng/g $s = 0.76$ ng/g 95% CL = 0.58 ng/g (dry basis)

Reported Results: 9 Quantitative Results: 9




Solid line : exercise assigned value (EA V); dotted line: $z=\pm 1$ (25% from EA V); dotted/dashed line: $z=\pm 2$ (50% from EA V); dashed line: $z=\pm 3$ (75% from EA V)

cis-nonachlor**SRM 2977**

Target Value = no target ng/g (dry basis)

Reported Results: 8 Quantitative Results: 2




Solid line: value from Certificate of Analysis ; dotted line: 95% confidence limits; dashed line: 30% from 95% confidence limits

4,4'-DDT

Tissue XII (QA05TIS12)

Assigned value = 1.68 ng/g $s = 0.63$ ng/g 95% CL = 1.00 ng/g (dry basis)

Reported Results: 10 Quantitative Results: 6




Solid line : exercise assigned value (EAV); dotted line: $z \pm 1$ (25% from EAV); dotted/dashed line: $z \pm 2$ (50% from EAV); dashed line: $z \pm 3$ (75% from EAV)

4,4'-DDT

SRM 2977

Certified Value = 1.28 ± 0.18 ng/g (dry basis)


Reported Results: 9 Quantitative Results: 3



Solid line: value from Certificate of Analysis ; dotted line: 95% confidence limits; dashed line: 30% from 95% confidence limits

mirex**Tissue XII (QA05TIS12)**


Assigned value = no target ng/g (dry basis)
Reported Results: 9 Quantitative Results: 2



Solid line : exercise assigned value (EAV); dotted line: $z=\pm 1$ (25% from EAV); dotted/dashed line: $z=\pm 2$ (50% from EAV); dashed line: $z=\pm 3$ (75% from EAV)

mirex**SRM 2977**

Target Value = no target ng/g (dry basis)
Reported Results: 8 Quantitative Results: 1




Solid line: value from Certificate of Analysis ; dotted line: 95% confidence limits; dashed line: 30% from 95% confidence limits

PCB 8**Tissue XII (QA05TIS12)**

Assigned value = 2.56 ng/g $s = 0.60$ ng/g 95% CL = 0.75 ng/g (dry basis)

Reported Results: 11 Quantitative Results: 10




Solid line : exercise assigned value (EAV); dotted line: ± 1 (25% from EAV); dotted/dashed line: ± 2 (50% from EAV); dashed line: ± 3 (75% from EAV)

PCB 8**SRM 2977**

Certified Value = 2.10 ± 0.15 ng/g (dry basis)

Reported Results: 11 Quantitative Results: 10




Solid line: value from Certificate of Analysis ; dotted line: 95% confidence limits; dashed line: 30% from 95% confidence limits

PCB 18**Tissue XII (QA05TIS12)**

Assigned value = 5.71 ng/g $s = 1.87$ ng/g 95% CL = 1.34 ng/g (dry basis)

Reported Results: 12 Quantitative Results: 12




Solid line : exercise assigned value (EAV); dotted line: $z \pm 1$ (25% from EAV); dotted/dashed line: $z \pm 2$ (50% from EAV); dashed line: $z \pm 3$ (75% from EAV)

PCB 18**SRM 2977**

Certified Value = 2.65 ± 0.30 ng/g (dry basis)

Reported Results: 12 Quantitative Results: 11




Solid line: value from Certificate of Analysis ; dotted line: 95% confidence limits; dashed line: 30% from 95% confidence limits

PCB 28**Tissue XII (QA05TIS12)**

Assigned value = 23.2 ng/g $s = 4.6$ ng/g 95% CL = 3.3 ng/g (dry basis)

Reported Results: 11 Quantitative Results: 11




Solid line : exercise assigned value (EAV); dotted line: $z=\pm 1$ (25% from EAV); dotted/dashed line: $z=\pm 2$ (50% from EAV); dashed line: $z=\pm 3$ (75% from EAV)

PCB 28**SRM 2977**

Certified Value = 5.37 ± 0.44 ng/g (dry basis)

Reported Results: 11 Quantitative Results: 11




Solid line: value from Certificate of Analysis ; dotted line: 95% confidence limits; dashed line: 30% from 95% confidence limits

PCB 31**Tissue XII (QA05TIS12)**

Assigned value = 21.7 ng/g $s = 3.1$ ng/g 95% CL = 3.2 ng/g (dry basis)

Reported Results: 8 Quantitative Results: 8




Solid line : exercise assigned value (EAV); dotted line: $z=\pm 1$ (25% from EAV); dotted/dashed line: $z=\pm 2$ (50% from EAV); dashed line: $z=\pm 3$ (75% from EAV)

PCB 31**SRM 2977**

Certified Value = 3.92 ± 0.24 ng/g (dry basis)

Reported Results: 8 Quantitative Results: 8




Solid line: value from Certificate of Analysis ; dotted line: 95% confidence limits; dashed line: 30% from 95% confidence limits

PCB 44**Tissue XII (QA05TIS12)**

Assigned value = 31.4 ng/g $s = 11.2$ ng/g 95% CL = 8.0 ng/g (dry basis)

Reported Results: 12 Quantitative Results: 12




Solid line : exercise assigned value (EAV); dotted line: $z=\pm 1$ (25% from EAV); dotted/dashed line: $z=\pm 2$ (50% from EAV); dashed line: $z=\pm 3$ (75% from EAV)

PCB 44**SRM 2977**

Certified Value = 3.25 ± 0.63 ng/g (dry basis)

Reported Results: 12 Quantitative Results: 11




Solid line: value from Certificate of Analysis ; dotted line: 95% confidence limits; dashed line: 30% from 95% confidence limits

PCB 49**Tissue XII (QA05TIS12)**

Assigned value = 37.2 ng/g $s = 11.3$ ng/g 95% CL = 8.7 ng/g (dry basis)

Reported Results: 10 Quantitative Results: 10




Solid line : exercise assigned value (EAV); dotted line: $z=\pm 1$ (25% from EAV); dotted/dashed line: $z=\pm 2$ (50% from EAV); dashed line: $z=\pm 3$ (75% from EAV)

PCB 49**SRM 2977**

Target Value = no target ng/g (dry basis)

Reported Results: 10 Quantitative Results: 9




Solid line: value from Certificate of Analysis ; dotted line: 95% confidence limits; dashed line: 30% from 95% confidence limits

PCB 52**Tissue XII (QA05TIS12)**

Assigned value = 46.6 ng/g $s = 10.7$ ng/g 95% CL = 7.6 ng/g (dry basis)

Reported Results: 12 Quantitative Results: 12




Solid line : exercise assigned value (EAV); dotted line: $z=\pm 1$ (25% from EAV); dotted/dashed line: $z=\pm 2$ (50% from EAV); dashed line: $z=\pm 3$ (75% from EAV)

PCB 52**SRM 2977**

Certified Value = 8.37 ± 0.54 ng/g (dry basis)

Reported Results: 12 Quantitative Results: 12




Solid line: value from Certificate of Analysis ; dotted line: 95% confidence limits; dashed line: 30% from 95% confidence limits

PCB 66**Tissue XII (QA05TIS12)**

Assigned value = 48.4 ng/g $s = 12.4$ ng/g 95% CL = 8.3 ng/g (dry basis)

Reported Results: 12 Quantitative Results: 12




Solid line : exercise assigned value (EAV); dotted line: $z=\pm 1$ (25% from EAV); dotted/dashed line: $z=\pm 2$ (50% from EAV); dashed line: $z=\pm 3$ (75% from EAV)

PCB 66**SRM 2977**

Certified Value = 3.64 ± 0.32 ng/g (dry basis)

Reported Results: 12 Quantitative Results: 12




Solid line: value from Certificate of Analysis ; dotted line: 95% confidence limits; dashed line: 30% from 95% confidence limits

PCB 95**Tissue XII (QA05TIS12)**

Assigned value = 51.5 ng/g $s = 9.2$ ng/g 95% CL = 8.5 ng/g (dry basis)

Reported Results: 9 Quantitative Results: 9




Solid line : exercise assigned value (EA V); dotted line: $z=\pm 1$ (25% from EA V); dotted/dashed line: $z=\pm 2$ (50% from EA V); dashed line: $z=\pm 3$ (75% from EA V)

PCB 95**SRM 2977**

Certified Value = 5.39 ± 0.59 ng/g (dry basis)

Reported Results: 9 Quantitative Results: 9




Solid line: value from Certificate of Analysis ; dotted line: 95% confidence limits; dashed line: 30% from 95% confidence limits

PCB 99**Tissue XII (QA05TIS12)**

Assigned value = 47.0 ng/g $s = 7.2$ ng/g 95% CL = 6.7 ng/g (dry basis)

Reported Results: 9 Quantitative Results: 9




Solid line : exercise assigned value (EAV); dotted line: $z=\pm 1$ (25% from EAV); dotted/dashed line: $z=\pm 2$ (50% from EAV); dashed line: $z=\pm 3$ (75% from EAV)

PCB 99**SRM 2977**

Target Value = no target ng/g (dry basis)

Reported Results: 9 Quantitative Results: 9




Solid line: value from Certificate of Analysis ; dotted line: 95% confidence limits; dashed line: 30% from 95% confidence limits

PCB 101**Tissue XII (QA05TIS12)**

Assigned value = 88.1 ng/g $s = 14.4$ ng/g 95% CL = 11.1 ng/g (dry basis)

Reported Results: 12 Quantitative Results: 12




Solid line : exercise assigned value (EAV); dotted line: $z=\pm 1$ (25% from EAV); dotted/dashed line: $z=\pm 2$ (50% from EAV); dashed line: $z=\pm 3$ (75% from EAV)

PCB 101**SRM 2977**

Certified Value = 11.2 ± 1.2 ng/g (dry basis)

Reported Results: 12 Quantitative Results: 11




Solid line: value from Certificate of Analysis ; dotted line: 95% confidence limits; dashed line: 30% from 95% confidence limits

PCB 105**Tissue XII (QA05TIS12)**

Assigned value = 31.1 ng/g $s = 5.3$ ng/g 95% CL = 3.8 ng/g (dry basis)

Reported Results: 12 Quantitative Results: 12




Solid line : exercise assigned value (EAV); dotted line: $z=\pm 1$ (25% from EAV); dotted/dashed line: $z=\pm 2$ (50% from EAV); dashed line: $z=\pm 3$ (75% from EAV)

PCB 105**SRM 2977**

Certified Value = 3.76 ± 0.49 ng/g (dry basis)

Reported Results: 12 Quantitative Results: 12




Solid line: value from Certificate of Analysis ; dotted line: 95% confidence limits; dashed line: 30% from 95% confidence limits

PCB 118**Tissue XII (QA05TIS12)**

Assigned value = 79.9 ng/g $s = 15.9$ ng/g 95% CL = 11.4 ng/g (dry basis)

Reported Results: 12 Quantitative Results: 12




Solid line : exercise assigned value (EAV); dotted line: $z=\pm 1$ (25% from EAV); dotted/dashed line: $z=\pm 2$ (50% from EAV); dashed line: $z=\pm 3$ (75% from EAV)

PCB 118**SRM 2977**

Certified Value = 10.5 ± 1.0 ng/g (dry basis)

Reported Results: 12 Quantitative Results: 12




Solid line: value from Certificate of Analysis ; dotted line: 95% confidence limits; dashed line: 30% from 95% confidence limits

PCB 128**Tissue XII (QA05TIS12)**

Assigned value = 13.0 ng/g $s = 2.2$ ng/g 95% CL = 1.7 ng/g (dry basis)

Reported Results: 12 Quantitative Results: 12




Solid line : exercise assigned value (EAV); dotted line: $z=\pm 1$ (25% from EAV); dotted/dashed line: $z=\pm 2$ (50% from EAV); dashed line: $z=\pm 3$ (75% from EAV)

PCB 128**SRM 2977**

Certified Value = 2.49 ± 0.28 ng/g (dry basis)

Reported Results: 12 Quantitative Results: 11




Solid line: value from Certificate of Analysis ; dotted line: 95% confidence limits; dashed line: 30% from 95% confidence limits

PCB 138**Tissue XII (QA05TIS12)**

Assigned value = 64.1 ng/g $s = 14.9$ ng/g 95% CL = 15.7 ng/g (dry basis)

Reported Results: 7 Quantitative Results: 7




Solid line : exercise assigned value (EAV); dotted line: $z=\pm 1$ (25% from EAV); dotted/dashed line: $z=\pm 2$ (50% from EAV); dashed line: $z=\pm 3$ (75% from EAV)

PCB 138**SRM 2977**

Target Value = no target ng/g (dry basis)

Reported Results: 7 Quantitative Results: 7




Solid line: value from Certificate of Analysis ; dotted line: 95% confidence limits; dashed line: 30% from 95% confidence limits

PCB 149**Tissue XII (QA05TIS12)**

Assigned value = 61.5 ng/g $s = 11.2$ ng/g 95% CL = 11.7 ng/g (dry basis)

Reported Results: 9 Quantitative Results: 9




Solid line : exercise assigned value (EAV); dotted line: $z=\pm 1$ (25% from EAV); dotted/dashed line: $z=\pm 2$ (50% from EAV); dashed line: $z=\pm 3$ (75% from EAV)

PCB 149**SRM 2977**

Certified Value = 9.23 ± 0.12 ng/g (dry basis)

Reported Results: 9 Quantitative Results: 9




Solid line: value from Certificate of Analysis ; dotted line: 95% confidence limits; dashed line: 30% from 95% confidence limits

PCB 153**Tissue XII (QA05TIS12)**

Assigned value = 85.7 ng/g $s = 29.6$ ng/g 95% CL = 21.2 ng/g (dry basis)

Reported Results: 11 Quantitative Results: 11




Solid line : exercise assigned value (EAV); dotted line: $z=\pm 1$ (25% from EAV); dotted/dashed line: $z=\pm 2$ (50% from EAV); dashed line: $z=\pm 3$ (75% from EAV)

PCB 153**SRM 2977**

Certified Value = 14.1 ± 1.0 ng/g (dry basis)

Reported Results: 11 Quantitative Results: 11




Solid line: value from Certificate of Analysis ; dotted line: 95% confidence limits; dashed line: 30% from 95% confidence limits

PCB 156**Tissue XII (QA05TIS12)**

Assigned value = 5.64 ng/g $s = 1.25$ ng/g 95% CL = 1.16 ng/g (dry basis)

Reported Results: 8 Quantitative Results: 8




Solid line : exercise assigned value (EAV); dotted line: $z=\pm 1$ (25% from EAV); dotted/dashed line: $z=\pm 2$ (50% from EAV); dashed line: $z=\pm 3$ (75% from EAV)

PCB 156**SRM 2977**

Certified Value = 0.960 ± 0.085 ng/g (dry basis)

Reported Results: 8 Quantitative Results: 8




Solid line: value from Certificate of Analysis ; dotted line: 95% confidence limits; dashed line: 30% from 95% confidence limits

PCB 170**Tissue XII (QA05TIS12)**

Assigned value = 1.80 ng/g $s = 0.38$ ng/g 95% CL = 0.29 ng/g (dry basis)

Reported Results: 12 Quantitative Results: 10




Solid line : exercise assigned value (EAV); dotted line: $z=\pm 1$ (25% from EAV); dotted/dashed line: $z=\pm 2$ (50% from EAV); dashed line: $z=\pm 3$ (75% from EAV)

PCB 170**SRM 2977**

Certified Value = 2.95 ± 0.23 ng/g (dry basis)

Reported Results: 12 Quantitative Results: 11




Solid line: value from Certificate of Analysis ; dotted line: 95% confidence limits; dashed line: 30% from 95% confidence limits

PCB 180**Tissue XII (QA05TIS12)**

Assigned value = 9.29 ng/g $s = 3.11$ ng/g 95% CL = 2.39 ng/g (dry basis)

Reported Results: 10 Quantitative Results: 10




Solid line : exercise assigned value (EAV); dotted line: $z=\pm 1$ (25% from EAV); dotted/dashed line: $z=\pm 2$ (50% from EAV); dashed line: $z=\pm 3$ (75% from EAV)

PCB 180**SRM 2977**

Certified Value = 6.79 ± 0.67 ng/g (dry basis)

Reported Results: 10 Quantitative Results: 10




Solid line: value from Certificate of Analysis ; dotted line: 95% confidence limits; dashed line: 30% from 95% confidence limits

PCB 187**Tissue XII (QA05TIS12)**

Assigned value = 18.0 ng/g $s = 4.0$ ng/g 95% CL = 2.7 ng/g (dry basis)

Reported Results: 12 Quantitative Results: 12




Solid line : exercise assigned value (EAV); dotted line: $z=\pm 1$ (25% from EAV); dotted/dashed line: $z=\pm 2$ (50% from EAV); dashed line: $z=\pm 3$ (75% from EAV)

PCB 187**SRM 2977**

Certified Value = 4.76 ± 0.38 ng/g (dry basis)

Reported Results: 12 Quantitative Results: 12




Solid line: value from Certificate of Analysis ; dotted line: 95% confidence limits; dashed line: 30% from 95% confidence limits

PCB 194**Tissue XII (QA05TIS12)**

Assigned value = 0.501 ng/g $s = 0.084$ ng/g 95% CL = 0.105 ng/g (dry basis)

Reported Results: 9 Quantitative Results: 6




Solid line : exercise assigned value (EAV); dotted line: $z=\pm 1$ (25% from EAV); dotted/dashed line: $z=\pm 2$ (50% from EAV); dashed line: $z=\pm 3$ (75% from EAV)

PCB 194**SRM 2977**

Certified Value = 0.897 ± 0.042 ng/g (dry basis)


Reported Results: 9 Quantitative Results: 7



Solid line: value from Certificate of Analysis ; dotted line: 95% confidence limits; dashed line: 30% from 95% confidence limits

PCB 195**Tissue XII (QA05TIS12)**


Assigned value = no target ng/g (dry basis)
Reported Results: 12 Quantitative Results: 3



Solid line : exercise assigned value (EAV); dotted line: $z=\pm 1$ (25% from EAV); dotted/dashed line: $z=\pm 2$ (50% from EAV); dashed line: $z=\pm 3$ (75% from EAV)

PCB 195**SRM 2977**


Target Value = no target ng/g (dry basis)
Reported Results: 12 Quantitative Results: 3



Solid line: value from Certificate of Analysis ; dotted line: 95% confidence limits; dashed line: 30% from 95% confidence limits

PCB 206**Tissue XII (QA05TIS12)**


Assigned value = no target ng/g (dry basis)
Reported Results: 12 Quantitative Results: 3



Solid line : exercise assigned value (EAV); dotted line: $z=\pm 1$ (25% from EAV); dotted/dashed line: $z=\pm 2$ (50% from EAV); dashed line: $z=\pm 3$ (75% from EAV)

PCB 206**SRM 2977**


Target Value = no target ng/g (dry basis)
Reported Results: 12 Quantitative Results: 3



Solid line: value from Certificate of Analysis ; dotted line: 95% confidence limits; dashed line: 30% from 95% confidence limits

PCB 209**Tissue XII (QA05TIS12)**


Assigned value = no target ng/g (dry basis)
Reported Results: 12 Quantitative Results: 3



Solid line : exercise assigned value (EAV); dotted line: $z=\pm 1$ (25% from EAV); dotted/dashed line: $z=\pm 2$ (50% from EAV); dashed line: $z=\pm 3$ (75% from EAV)

PCB 209**SRM 2977**

Target Value = no target ng/g (dry basis)
Reported Results: 12 Quantitative Results: 2




Solid line: value from Certificate of Analysis ; dotted line: 95% confidence limits; dashed line: 30% from 95% confidence limits

BDE 17**Tissue XII (QA05TIS12)**

Assigned value = 3.72 ng/g $s = 0.66$ ng/g 95% CL = 1.64 ng/g (dry basis)

Reported Results: 3 Quantitative Results: 3




Solid line : exercise assigned value (EA V); dotted line: ± 1 (25% from EA V); dotted/dashed line: ± 2 (50% from EA V); dashed line: ± 3 (75% from EA V)

BDE 17**SRM 2977**

Target Value = no target ng/g (dry basis)

Reported Results: 3 Quantitative Results: 3




Solid line: value from Certificate of Analysis ; dotted line: 95% confidence limits; dashed line: 30% from 95% confidence limits

BDE 28**Tissue XII (QA05TIS12)**

Assigned value = 3.08 ng/g $s = 1.58$ ng/g 95% CL = 1.97 ng/g (dry basis)

Reported Results: 5 Quantitative Results: 5




Solid line : exercise assigned value (EAV); dotted line: $z=\pm 1$ (25% from EAV); dotted/dashed line: $z=\pm 2$ (50% from EAV); dashed line: $z=\pm 3$ (75% from EAV)

BDE 28**SRM 2977**

Target Value = no target ng/g (dry basis)

Reported Results: 5 Quantitative Results: 5




Solid line: value from Certificate of Analysis ; dotted line: 95% confidence limits; dashed line: 30% from 95% confidence limits

BDE 47**Tissue XII (QA05TIS12)**

Assigned value = 23.3 ng/g $s = 5.4$ ng/g 95% CL = 6.7 ng/g (dry basis)

Reported Results: 5 Quantitative Results: 5




Solid line : exercise assigned value (EAV); dotted line: $z=\pm 1$ (25% from EAV); dotted/dashed line: $z=\pm 2$ (50% from EAV); dashed line: $z=\pm 3$ (75% from EAV)

BDE 47**SRM 2977**

Target Value = no target ng/g (dry basis)

Reported Results: 5 Quantitative Results: 5




Solid line: value from Certificate of Analysis ; dotted line: 95% confidence limits; dashed line: 30% from 95% confidence limits

BDE 49**Tissue XII (QA05TIS12)**

Assigned value = 6.75 ng/g $s = 2.00$ ng/g 95% CL = 4.97 ng/g (dry basis)

Reported Results: 4 Quantitative Results: 3




Solid line : exercise assigned value (EAV); dotted line: $z=\pm 1$ (25% from EAV); dotted/dashed line: $z=\pm 2$ (50% from EAV); dashed line: $z=\pm 3$ (75% from EAV)

BDE 49**SRM 2977**

Target Value = 1.02 ± 0.03 ng/g (dry basis)

Reported Results: 4 Quantitative Results: 3




Solid line: value from Certificate of Analysis ; dotted line: 95% confidence limits; dashed line: 30% from 95% confidence limits

BDE 66**Tissue XII (QA05TIS12)**

Assigned value = 0.984 ng/g $s = 0.403$ ng/g 95% CL = 0.642 ng/g (dry basis)

Reported Results: 5 Quantitative Results: 4




Solid line : exercise assigned value (EAV); dotted line: $z \pm 1$ (25% from EAV); dotted/dashed line: $z \pm 2$ (50% from EAV); dashed line: $z \pm 3$ (75% from EAV)

BDE 66**SRM 2977**

Target Value = 0.375 ± 0.062 ng/g (dry basis)

Reported Results: 5 Quantitative Results: 3




Solid line: value from Certificate of Analysis ; dotted line: 95% confidence limits; dashed line: 30% from 95% confidence limits

BDE 85**Tissue XII (QA05TIS12)**

Assigned value = 0.418 ng/g $s = 0.074$ ng/g 95% CL = 0.185 ng/g (dry basis)

Reported Results: 5 Quantitative Results: 3




Solid line : exercise assigned value (EAV); dotted line: $z=\pm 1$ (25% from EAV); dotted/dashed line: $z=\pm 2$ (50% from EAV); dashed line: $z=\pm 3$ (75% from EAV)

BDE 85**SRM 2977**

Target Value = no target ng/g (dry basis)

Reported Results: 5 Quantitative Results: 3




Solid line: value from Certificate of Analysis ; dotted line: 95% confidence limits; dashed line: 30% from 95% confidence limits

BDE 99**Tissue XII (QA05TIS12)**

Assigned value = 11.5 ng/g $s = 3.4$ ng/g 95% CL = 5.3 ng/g (dry basis)

Reported Results: 4 Quantitative Results: 4




Solid line : exercise assigned value (EAV); dotted line: $z=\pm 1$ (25% from EAV); dotted/dashed line: $z=\pm 2$ (50% from EAV); dashed line: $z=\pm 3$ (75% from EAV)

BDE 99**SRM 2977**

Target Value = 4.11 \pm 0.40 ng/g (dry basis)

Reported Results: 4 Quantitative Results: 4




Solid line: value from Certificate of Analysis ; dotted line: 95% confidence limits; dashed line: 30% from 95% confidence limits

BDE 100**Tissue XII (QA05TIS12)**

Assigned value = 6.85 ng/g $s = 1.93$ ng/g 95% CL = 2.40 ng/g (dry basis)

Reported Results: 5 Quantitative Results: 5




Solid line : exercise assigned value (EAV); dotted line: $z=\pm 1$ (25% from EAV); dotted/dashed line: $z=\pm 2$ (50% from EAV); dashed line: $z=\pm 3$ (75% from EAV)

BDE 100**SRM 2977**

Target Value = 1.06 ± 0.18 ng/g (dry basis)


Reported Results: 5 Quantitative Results: 5



Solid line: value from Certificate of Analysis ; dotted line: 95% confidence limits; dashed line: 30% from 95% confidence limits

BDE 138**Tissue XII (QA05TIS12)**


Assigned value = no target ng/g (dry basis)
Reported Results: 4 Quantitative Results: 2



Solid line : exercise assigned value (EAV); dotted line: $z=\pm 1$ (25% from EAV); dotted/dashed line: $z=\pm 2$ (50% from EAV); dashed line: $z=\pm 3$ (75% from EAV)

BDE 138**SRM 2977**

Target Value = no target ng/g (dry basis)
Reported Results: 13 Quantitative Results: 2




Solid line: value from Certificate of Analysis ; dotted line: 95% confidence limits; dashed line: 30% from 95% confidence limits

BDE 153**Tissue XII (QA05TIS12)**

Assigned value = 0.515 ng/g $s = 0.097$ ng/g 95% CL = 0.155 ng/g (dry basis)

Reported Results: 5 Quantitative Results: 4




Solid line : exercise assigned value (EAV); dotted line: $z=\pm 1$ (25% from EAV); dotted/dashed line: $z=\pm 2$ (50% from EAV); dashed line: $z=\pm 3$ (75% from EAV)

BDE 153**SRM 2977**

Target Value = no target ng/g (dry basis)

Reported Results: 5 Quantitative Results: 3




Solid line: value from Certificate of Analysis ; dotted line: 95% confidence limits; dashed line: 30% from 95% confidence limits

BDE 154**Tissue XII (QA05TIS12)**

Assigned value = 0.550 ng/g $s = 0.103$ ng/g 95% CL = 0.164 ng/g (dry basis)

Reported Results: 5 Quantitative Results: 4




Solid line : exercise assigned value (EA V); dotted line: $z=\pm 1$ (25% from EA V); dotted/dashed line: $z=\pm 2$ (50% from EA V); dashed line: $z=\pm 3$ (75% from EA V)

BDE 154**SRM 2977**

Target Value = no target ng/g (dry basis)


Reported Results: 5 Quantitative Results: 3



Solid line: value from Certificate of Analysis ; dotted line: 95% confidence limits; dashed line: 30% from 95% confidence limits

BDE 155**Tissue XII (QA05TIS12)**


Assigned value = no target ng/g (dry basis)
Reported Results: 3 Quantitative Results: 2



Solid line : exercise assigned value (EAV); dotted line: $z=\pm 1$ (25% from EAV); dotted/dashed line: $z=\pm 2$ (50% from EAV); dashed line: $z=\pm 3$ (75% from EAV)

BDE 155**SRM 2977**

Target Value = no target ng/g (dry basis)
Reported Results: 3 Quantitative Results: 1



Solid line: value from Certificate of Analysis ; dotted line: 95% confidence limits; dashed line: 30% from 95% confidence limits

Appendix H: Charts of Marine Sediment XIII and SRM 1941b Results by Analyte

See Tables 10 through 17 for results reported as *<number*, detection limit, etc.

Charts for analytes with few reported numerical results are not included in this appendix.

Note: The numbers added to the charts are the values reported that are off the scale of the chart.

For Marine Sediment XIII plots:

Solid line: exercise assigned value

Dotted line: $z = \pm 1$, i. e., 25 % from assigned value

Dotted/dashed line: $z = \pm 2$, i. e., 50 % from assigned value

Dashed line: $z = \pm 3$, i. e., 75 % from assigned value

For SRM 1941b plots:

Solid line: material certified concentration or target value (see caption of each plot)

Dotted line: 95 % confidence interval (CI)


Dashed line: 30 % from 95 % confidence interval (CI)

naphthalene

Sediment XIII (QA05SED13)

Assigned value = 785 ng/g $s = 186$ ng/g 95% CL = 172 ng/g (dry basis)

Reported Results: 10 Quantitative Results: 10




Solid line : exercise assigned value (EAV); dotted line: $z=\pm 1$ (25% from EAV); dotted/dashed line: $z=\pm 2$ (50% from EAV); dashed line: $z=\pm 3$ (75% from EAV)

naphthalene

SRM 1941b

Certified Value = 848 ± 95 ng/g (dry basis)
Reported Results: 9 Quantitative Results: 9




Solid line: value from Certificate of Analysis ; dotted line: 95% confidence limits; dashed line: 30% from 95% confidence limits

2-methylnaphthalene

Sediment XIII (QA05SED13)

Assigned value = 219 ng/g $s = 66$ ng/g 95% CL = 61 ng/g (dry basis)

Reported Results: 9 Quantitative Results: 9




Solid line : exercise assigned value (EAV); dotted line: $z=\pm 1$ (25% from EAV); dotted/dashed line: $z=\pm 2$ (50% from EAV); dashed line: $z=\pm 3$ (75% from EAV)

2-methylnaphthalene

SRM 1941b

Reference Value = 276 ± 53 ng/g (dry basis)

Reported Results: 8 Quantitative Results: 8




Solid line: value from Certificate of Analysis ; dotted line: 95% confidence limits; dashed line: 30% from 95% confidence limits

1-methylnaphthalene

Sediment XIII (QA05SED13)

Assigned value = 98.2 ng/g $s = 24.0$ ng/g 95% CL = 20.0 ng/g (dry basis)

Reported Results: 8 Quantitative Results: 8




Solid line : exercise assigned value (EAV); dotted line: $z=\pm 1$ (25% from EAV); dotted/dashed line: $z=\pm 2$ (50% from EAV); dashed line: $z=\pm 3$ (75% from EAV)

1-methylnaphthalene

SRM 1941b

Reference Value = 127 ± 14 ng/g (dry basis)

Reported Results: 7 Quantitative Results: 7




Solid line: value from Certificate of Analysis ; dotted line: 95% confidence limits; dashed line: 30% from 95% confidence limits

biphenyl**Sediment XIII (QA05SED13)**

Assigned value = 65.8 ng/g $s = 23.6$ ng/g 95% CL = 21.8 ng/g (dry basis)


Reported Results: 9 Quantitative Results: 9



Solid line : exercise assigned value (EAV); dotted line: $z=\pm 1$ (25% from EAV); dotted/dashed line: $z=\pm 2$ (50% from EAV); dashed line: $z=\pm 3$ (75% from EAV)

biphenyl**SRM 1941b**

Reference Value = 74 ± 8 ng/g (dry basis)
Reported Results: 8 Quantitative Results: 8




Solid line: value from Certificate of Analysis ; dotted line: 95% confidence limits; dashed line: 30% from 95% confidence limits

2,6-dimethylnaphthalene

Sediment XIII (QA05SED13)

Assigned value = 81.3 ng/g $s = 43.6$ ng/g 95% CL = 40.3 ng/g (dry basis)

Reported Results: 7 Quantitative Results: 7




Solid line : exercise assigned value (EAV); dotted line: $z=\pm 1$ (25% from EAV); dotted/dashed line: $z=\pm 2$ (50% from EAV); dashed line: $z=\pm 3$ (75% from EAV)

2,6-dimethylnaphthalene

SRM 1941b

Reference Value = 75.9 ± 4.5 ng/g (dry basis)


Reported Results: 6 Quantitative Results: 6



Solid line: value from Certificate of Analysis ; dotted line: 95% confidence limits; dashed line: 30% from 95% confidence limits

acenaphthylene**Sediment XIII (QA05SED13)**Assigned value = 45.1 ng/g $s = 13.8$ ng/g 95% CL = 11.5 ng/g (dry basis)


Reported Results: 10 Quantitative Results: 10



Solid line : exercise assigned value (EAV); dotted line: $z=\pm 1$ (25% from EAV); dotted/dashed line: $z=\pm 2$ (50% from EAV); dashed line: $z=\pm 3$ (75% from EAV)

acenaphthylene**SRM 1941b**Reference Value = 53.3 ± 6.4 ng/g (dry basis)

Reported Results: 9 Quantitative Results: 9




Solid line: value from Certificate of Analysis ; dotted line: 95% confidence limits; dashed line: 30% from 95% confidence limits

1,6,7-trimethylnaphthalene

Sediment XIII (QA05SED13)

Assigned value = no target ng/g (dry basis)

Reported Results: 3 Quantitative Results: 3




Solid line : exercise assigned value (EAV); dotted line: $z=\pm 1$ (25% from EAV); dotted/dashed line: $z=\pm 2$ (50% from EAV); dashed line: $z=\pm 3$ (75% from EAV)

1,6,7-trimethylnaphthalene

SRM 1941b

Reference Value = 25.5 ± 5.1 ng/g (dry basis)

Reported Results: 4 Quantitative Results: 4




Solid line: value from Certificate of Analysis ; dotted line: 95% confidence limits; dashed line: 30% from 95% confidence limits

fluorene**Sediment XIII (QA05SED13)**

Assigned value = 56.1 ng/g $s = 12.2$ ng/g 95% CL = 10.2 ng/g (dry basis)

Reported Results: 10 Quantitative Results: 10




Solid line : exercise assigned value (EAV); dotted line: $z=\pm 1$ (25% from EAV); dotted/dashed line: $z=\pm 2$ (50% from EAV); dashed line: $z=\pm 3$ (75% from EAV)

fluorene**SRM 1941b**

Certified Value = 85.0 ± 15.0 ng/g (dry basis)

Reported Results: 9 Quantitative Results: 9




Solid line: value from Certificate of Analysis ; dotted line: 95% confidence limits; dashed line: 30% from 95% confidence limits

phenanthrene

Sediment XIII (QA05SED13)

Assigned value = 306 ng/g $s = 89$ ng/g 95% CL = 75 ng/g (dry basis)

Reported Results: 10 Quantitative Results: 10




Solid line : exercise assigned value (EAV); dotted line: $z=\pm 1$ (25% from EAV); dotted/dashed line: $z=\pm 2$ (50% from EAV); dashed line: $z=\pm 3$ (75% from EAV)

phenanthrene

SRM 1941b

Certified Value = 406 ± 44 ng/g (dry basis)
Reported Results: 9 Quantitative Results: 9




Solid line: value from Certificate of Analysis ; dotted line: 95% confidence limits; dashed line: 30% from 95% confidence limits

anthracene**Sediment XIII (QA05SED13)**

Assigned value = 137 ng/g $s = 47$ ng/g 95% CL = 40 ng/g (dry basis)


Reported Results: 10 Quantitative Results: 10



Solid line : exercise assigned value (EAV); dotted line: $z=\pm 1$ (25% from EAV); dotted/dashed line: $z=\pm 2$ (50% from EAV); dashed line: $z=\pm 3$ (75% from EAV)

anthracene**SRM 1941b**

Certified Value = 184 ± 18 ng/g (dry basis)
Reported Results: 9 Quantitative Results: 9




Solid line: value from Certificate of Analysis ; dotted line: 95% confidence limits; dashed line: 30% from 95% confidence limits

1-methylphenanthrene

Sediment XIII (QA05SED13)

Assigned value = 55.4 ng/g $s = 18.2$ ng/g 95% CL = 19.1 ng/g (dry basis)

Reported Results: 9 Quantitative Results: 9




Solid line : exercise assigned value (EAV); dotted line: $z=\pm 1$ (25% from EAV); dotted/dashed line: $z=\pm 2$ (50% from EAV); dashed line: $z=\pm 3$ (75% from EAV)

1-methylphenanthrene

SRM 1941b

Certified Value = 73.2 ± 5.9 ng/g (dry basis)

Reported Results: 8 Quantitative Results: 8




Solid line: value from Certificate of Analysis ; dotted line: 95% confidence limits; dashed line: 30% from 95% confidence limits

fluoranthene**Sediment XIII (QA05SED13)**

Assigned value = 496 ng/g $s = 140$ ng/g 95% CL = 117 ng/g (dry basis)

Reported Results: 10 Quantitative Results: 10




Solid line : exercise assigned value (EAV); dotted line: $z=\pm 1$ (25% from EAV); dotted/dashed line: $z=\pm 2$ (50% from EAV); dashed line: $z=\pm 3$ (75% from EAV)

fluoranthene**SRM 1941b**

Certified Value = 651 ± 50 ng/g (dry basis)

Reported Results: 9 Quantitative Results: 9




Solid line: value from Certificate of Analysis ; dotted line: 95% confidence limits; dashed line: 30% from 95% confidence limits

pyrene**Sediment XIII (QA05SED13)**

Assigned value = 421 ng/g $s = 142$ ng/g 95% CL = 118 ng/g (dry basis)

Reported Results: 10 Quantitative Results: 10




Solid line : exercise assigned value (EAV); dotted line: $z=\pm 1$ (25% from EAV); dotted/dashed line: $z=\pm 2$ (50% from EAV); dashed line: $z=\pm 3$ (75% from EAV)

pyrene**SRM 1941b**

Certified Value = 581 ± 39 ng/g (dry basis)

Reported Results: 9 Quantitative Results: 9




Solid line: value from Certificate of Analysis ; dotted line: 95% confidence limits; dashed line: 30% from 95% confidence limits

benz[a]anthracene**Sediment XIII (QA05SED13)**

Assigned value = 241 ng/g $s = 80$ ng/g 95% CL = 67 ng/g (dry basis)

Reported Results: 9 Quantitative Results: 9




Solid line : exercise assigned value (EAV); dotted line: $z=\pm 1$ (25% from EAV); dotted/dashed line: $z=\pm 2$ (50% from EAV); dashed line: $z=\pm 3$ (75% from EAV)

benz[a]anthracene**SRM 1941b**

Certified Value = 335 ± 25 ng/g (dry basis)

Reported Results: 8 Quantitative Results: 8




Solid line: value from Certificate of Analysis ; dotted line: 95% confidence limits; dashed line: 30% from 95% confidence limits

chrysene**Sediment XIII (QA05SED13)**

Assigned value = 219 ng/g $s = 62$ ng/g 95% CL = 65 ng/g (dry basis)

Reported Results: 6 Quantitative Results: 6




Solid line : exercise assigned value (EAV); dotted line: $z=\pm 1$ (25% from EAV); dotted/dashed line: $z=\pm 2$ (50% from EAV); dashed line: $z=\pm 3$ (75% from EAV)

chrysene**SRM 1941b**

Certified Value = 291 ± 31 ng/g (dry basis)

Reported Results: 5 Quantitative Results: 5




Solid line: value from Certificate of Analysis ; dotted line: 95% confidence limits; dashed line: 30% from 95% confidence limits

benzo[b]fluoranthene**Sediment XIII (QA05SED13)**

Assigned value = 413 ng/g $s = 174$ ng/g 95% CL = 183 ng/g (dry basis)

Reported Results: 8 Quantitative Results: 8




Solid line : exercise assigned value (EAV); dotted line: $z=\pm 1$ (25% from EAV); dotted/dashed line: $z=\pm 2$ (50% from EAV); dashed line: $z=\pm 3$ (75% from EAV)

benzo[b]fluoranthene**SRM 1941b**

Certified Value = 453 ± 21 ng/g (dry basis)

Reported Results: 7 Quantitative Results: 7




Solid line: value from Certificate of Analysis ; dotted line: 95% confidence limits; dashed line: 30% from 95% confidence limits

benzo[k]fluoranthene**Sediment XIII (QA05SED13)**

Assigned value = 180 ng/g $s = 56$ ng/g 95% CL = 47 ng/g (dry basis)

Reported Results: 9 Quantitative Results: 9




Solid line : exercise assigned value (EAV); dotted line: $z=\pm 1$ (25% from EAV); dotted/dashed line: $z=\pm 2$ (50% from EAV); dashed line: $z=\pm 3$ (75% from EAV)

benzo[k]fluoranthene**SRM 1941b**

Certified Value = 225 ± 18 ng/g (dry basis)

Reported Results: 8 Quantitative Results: 8




Solid line: value from Certificate of Analysis ; dotted line: 95% confidence limits; dashed line: 30% from 95% confidence limits

benzo[e]pyrene**Sediment XIII (QA05SED13)**

Assigned value = 286 ng/g $s = 97$ ng/g 95% CL = 75 ng/g (dry basis)

Reported Results: 10 Quantitative Results: 10




Solid line : exercise assigned value (EAV); dotted line: $z=\pm 1$ (25% from EAV); dotted/dashed line: $z=\pm 2$ (50% from EAV); dashed line: $z=\pm 3$ (75% from EAV)

benzo[e]pyrene**SRM 1941b**

Certified Value = 325 ± 25 ng/g (dry basis)

Reported Results: 9 Quantitative Results: 9




Solid line: value from Certificate of Analysis ; dotted line: 95% confidence limits; dashed line: 30% from 95% confidence limits

benzo[a]pyrene**Sediment XIII (QA05SED13)**

Assigned value = 282 ng/g $s = 120$ ng/g 95% CL = 111 ng/g (dry basis)

Reported Results: 10 Quantitative Results: 10




Solid line : exercise assigned value (EAV); dotted line: $z=\pm 1$ (25% from EAV); dotted/dashed line: $z=\pm 2$ (50% from EAV); dashed line: $z=\pm 3$ (75% from EAV)

benzo[a]pyrene**SRM 1941b**

Certified Value = 358 ± 17 ng/g (dry basis)

Reported Results: 9 Quantitative Results: 9




Solid line: value from Certificate of Analysis ; dotted line: 95% confidence limits; dashed line: 30% from 95% confidence limits

perylene

Sediment XIII (QA05SED13)

Assigned value = 311 ng/g s = 131 ng/g 95% CL = 138 ng/g (dry basis)

Reported Results: 8 Quantitative Results: 8




Solid line : exercise assigned value (EAV); dotted line: $z=\pm 1$ (25% from EAV); dotted/dashed line: $z=\pm 2$ (50% from EAV); dashed line: $z=\pm 3$ (75% from EAV)

perylene

SRM 1941b

Certified Value = 397 ± 45 ng/g (dry basis)

Reported Results: 7 Quantitative Results: 7




Solid line: value from Certificate of Analysis ; dotted line: 95% confidence limits; dashed line: 30% from 95% confidence limits

indeno[1,2,3-cd]pyrene**Sediment XIII (QA05SED13)**

Assigned value = 258 ng/g $s = 101$ ng/g 95% CL = 85 ng/g (dry basis)

Reported Results: 9 Quantitative Results: 9




Solid line : exercise assigned value (EAV); dotted line: $z=\pm 1$ (25% from EAV); dotted/dashed line: $z=\pm 2$ (50% from EAV); dashed line: $z=\pm 3$ (75% from EAV)

indeno[1,2,3-cd]pyrene**SRM 1941b**

Certified Value = 341 ± 57 ng/g (dry basis)

Reported Results: 8 Quantitative Results: 8




Solid line: value from Certificate of Analysis ; dotted line: 95% confidence limits; dashed line: 30% from 95% confidence limits

dibenz[a,h]anthracene**Sediment XIII (QA05SED13)**

Assigned value = 42 ng/g $s = 21$ ng/g 95% CL = 34 ng/g (dry basis)

Reported Results: 6 Quantitative Results: 6




Solid line : exercise assigned value (EAV); dotted line: $z=\pm 1$ (25% from EAV); dotted/dashed line: $z=\pm 2$ (50% from EAV); dashed line: $z=\pm 3$ (75% from EAV)

dibenz[a,h]anthracene**SRM 1941b**

Certified Value = 53 ± 10 ng/g (dry basis)

Reported Results: 5 Quantitative Results: 5




Solid line: value from Certificate of Analysis ; dotted line: 95% confidence limits; dashed line: 30% from 95% confidence limits

benzo[ghi]perylene**Sediment XIII (QA05SED13)**

Assigned value = 244 ng/g $s = 96$ ng/g 95% CL = 81 ng/g (dry basis)

Reported Results: 10 Quantitative Results: 10




Solid line : exercise assigned value (EAV); dotted line: $z=\pm 1$ (25% from EAV); dotted/dashed line: $z=\pm 2$ (50% from EAV); dashed line: $z=\pm 3$ (75% from EAV)

benzo[ghi]perylene**SRM 1941b**

Certified Value = 307 ± 45 ng/g (dry basis)

Reported Results: 9 Quantitative Results: 9




Solid line: value from Certificate of Analysis ; dotted line: 95% confidence limits; dashed line: 30% from 95% confidence limits

alpha-HCH (a-BHC)**Sediment XIII (QA05SED13)**

Assigned value = no target ng/g (dry basis)

Reported Results: 7 Quantitative Results: 2




Solid line : exercise assigned value (EAV); dotted line: $z=\pm 1$ (25% from EAV); dotted/dashed line: $z=\pm 2$ (50% from EAV); dashed line: $z=\pm 3$ (75% from EAV)

alpha-HCH (a-BHC)**SRM 1941b**

Target Value = no target ng/g (dry basis)

Reported Results: 7 Quantitative Results: 2




Solid line: value from Certificate of Analysis ; dotted line: 95% confidence limits; dashed line: 30% from 95% confidence limits

hexachlorobenzene**Sediment XIII (QA05SED13)**

Assigned value = 5.38 ng/g $s = 1.90$ ng/g 95% CL = 2.00 ng/g (dry basis)

Reported Results: 7 Quantitative Results: 7




Solid line : exercise assigned value (EAV); dotted line: $z=\pm 1$ (25% from EAV); dotted/dashed line: $z=\pm 2$ (50% from EAV); dashed line: $z=\pm 3$ (75% from EAV)

hexachlorobenzene**SRM 1941b**

Certified Value = 5.83 ± 0.38 ng/g (dry basis)

Reported Results: 7 Quantitative Results: 7




Solid line: value from Certificate of Analysis ; dotted line: 95% confidence limits; dashed line: 30% from 95% confidence limits

gamma-HCH (g-BHC,lindane)**Sediment XIII (QA05SED13)**

Assigned value = no target ng/g (dry basis)

Reported Results: 8 Quantitative Results: 3




Solid line : exercise assigned value (EAV); dotted line: $z=\pm 1$ (25% from EAV); dotted/dashed line: $z=\pm 2$ (50% from EAV); dashed line: $z=\pm 3$ (75% from EAV)

gamma-HCH (g-BHC,lindane)**SRM 1941b**

Target Value = no target ng/g (dry basis)

Reported Results: 7 Quantitative Results: 2




Solid line: value from Certificate of Analysis ; dotted line: 95% confidence limits; dashed line: 30% from 95% confidence limits

beta-HCH (b-BHC)**Sediment XIII (QA05SED13)**

Assigned value = no target ng/g (dry basis)

Reported Results: 7 Quantitative Results: 2




Solid line : exercise assigned value (EAV); dotted line: $z=\pm 1$ (25% from EAV); dotted/dashed line: $z=\pm 2$ (50% from EAV); dashed line: $z=\pm 3$ (75% from EAV)

beta-HCH (b-BHC)**SRM 1941b**

Target Value = no target ng/g (dry basis)


Reported Results: 7 Quantitative Results: 2



Solid line: value from Certificate of Analysis ; dotted line: 95% confidence limits; dashed line: 30% from 95% confidence limits

aldrin**Sediment XIII (QA05SED13)**


Assigned value = no target ng/g (dry basis)
Reported Results: 8 Quantitative Results: 2



Solid line : exercise assigned value (EAV); dotted line: $z=\pm 1$ (25% from EAV); dotted/dashed line: $z=\pm 2$ (50% from EAV); dashed line: $z=\pm 3$ (75% from EAV)

aldrin**SRM 1941b**

Target Value = no target ng/g (dry basis)
Reported Results: 8 Quantitative Results: 2




Solid line: value from Certificate of Analysis ; dotted line: 95% confidence limits; dashed line: 30% from 95% confidence limits

heptachlor epoxide

Sediment XIII (QA05SED13)

Assigned value = no target ng/g (dry basis)
Reported Results: 8 Quantitative Results: 2




Solid line : exercise assigned value (EAV); dotted line: $z=\pm 1$ (25% from EAV); dotted/dashed line: $z=\pm 2$ (50% from EAV); dashed line: $z=\pm 3$ (75% from EAV)

heptachlor epoxide

SRM 1941b

Target Value = no target ng/g (dry basis)
Reported Results: 8 Quantitative Results: 2




Solid line: value from Certificate of Analysis ; dotted line: 95% confidence limits; dashed line: 30% from 95% confidence limits

oxychlordan

Sediment XIII (QA05SED13)

Assigned value = no target ng/g (dry basis)
Reported Results: 8 Quantitative Results: 2




Solid line : exercise assigned value (EAV); dotted line: $z=\pm 1$ (25% from EAV); dotted/dashed line: $z=\pm 2$ (50% from EAV); dashed line: $z=\pm 3$ (75% from EAV)

oxychlordan

SRM 1941b

Target Value = no target ng/g (dry basis)
Reported Results: 8 Quantitative Results: 2




Solid line: value from Certificate of Analysis ; dotted line: 95% confidence limits; dashed line: 30% from 95% confidence limits

gamma-chlordane**Sediment XIII (QA05SED13)**

Assigned value = 0.572 ng/g $s = 0.148$ ng/g 95% CL = 0.155 ng/g (dry basis)

Reported Results: 8 Quantitative Results: 6




Solid line : exercise assigned value (EAV); dotted line: $z=\pm 1$ (25% from EAV); dotted/dashed line: $z=\pm 2$ (50% from EAV); dashed line: $z=\pm 3$ (75% from EAV)

gamma-chlordane**SRM 1941b**

Certified Value = 0.566 ± 0.093 ng/g (dry basis)

Reported Results: 8 Quantitative Results: 7




Solid line: value from Certificate of Analysis ; dotted line: 95% confidence limits; dashed line: 30% from 95% confidence limits

2,4'-DDE

Sediment XIII (QA05SED13)

Assigned value = 0.380 ng/g $s = 0.103$ ng/g 95% CL = 0.256 ng/g (dry basis)

Reported Results: 8 Quantitative Results: 4




Solid line : exercise assigned value (EAV); dotted line: $z=\pm 1$ (25% from EAV); dotted/dashed line: $z=\pm 2$ (50% from EAV); dashed line: $z=\pm 3$ (75% from EAV)

2,4'-DDE

SRM 1941b

Reference Value = 0.380 ± 0.120 ng/g (dry basis)

Reported Results: 8 Quantitative Results: 5




Solid line: value from Certificate of Analysis ; dotted line: 95% confidence limits; dashed line: 30% from 95% confidence limits

cis-chlordane (alpha-chlordane)**Sediment XIII (QA05SED13)**

Assigned value = 0.482 ng/g $s = 0.058$ ng/g 95% CL = 0.072 ng/g (dry basis)

Reported Results: 8 Quantitative Results: 6




Solid line : exercise assigned value (EAV); dotted line: $z=\pm 1$ (25% from EAV); dotted/dashed line: $z=\pm 2$ (50% from EAV); dashed line: $z=\pm 3$ (75% from EAV)

cis-chlordane (alpha-chlordane)**SRM 1941b**

Certified Value = 0.85 ± 0.11 ng/g (dry basis)

Reported Results: 8 Quantitative Results: 7




Solid line: value from Certificate of Analysis ; dotted line: 95% confidence limits; dashed line: 30% from 95% confidence limits

trans-nonachlor**Sediment XIII (QA05SED13)**

Assigned value = 0.286 ng/g $s = 0.089$ ng/g 95% CL = 0.093 ng/g (dry basis)

Reported Results: 8 Quantitative Results: 6




Solid line : exercise assigned value (EAV); dotted line: $z=\pm 1$ (25% from EAV); dotted/dashed line: $z=\pm 2$ (50% from EAV); dashed line: $z=\pm 3$ (75% from EAV)

trans-nonachlor**SRM 1941b**

Certified Value = 0.438 ± 0.073 ng/g (dry basis)

Reported Results: 8 Quantitative Results: 6




Solid line: value from Certificate of Analysis ; dotted line: 95% confidence limits; dashed line: 30% from 95% confidence limits

dieldrin**Sediment XIII (QA05SED13)**

Assigned value = 0.386 ng/g $s = 0.034$ ng/g 95% CL = 0.084 ng/g (dry basis)

Reported Results: 8 Quantitative Results: 3




Solid line : exercise assigned value (EAV); dotted line: $z=\pm 1$ (25% from EAV); dotted/dashed line: $z=\pm 2$ (50% from EAV); dashed line: $z=\pm 3$ (75% from EAV)

dieldrin**SRM 1941b**

Target Value = no target ng/g (dry basis)

Reported Results: 8 Quantitative Results: 2




Solid line: value from Certificate of Analysis ; dotted line: 95% confidence limits; dashed line: 30% from 95% confidence limits

4,4'-DDE**Sediment XIII (QA05SED13)**

Assigned value = 3.44 ng/g $s = 1.08$ ng/g 95% CL = 1.00 ng/g (dry basis)

Reported Results: 8 Quantitative Results: 8




Solid line : exercise assigned value (EAV); dotted line: $z=\pm 1$ (25% from EAV); dotted/dashed line: $z=\pm 2$ (50% from EAV); dashed line: $z=\pm 3$ (75% from EAV)

4,4'-DDE**SRM 1941b**

Certified Value = 3.22 ± 0.28 ng/g (dry basis)

Reported Results: 8 Quantitative Results: 8




Solid line: value from Certificate of Analysis ; dotted line: 95% confidence limits; dashed line: 30% from 95% confidence limits

2,4'-DDD

Sediment XIII (QA05SED13)

Assigned value = 0.927 ng/g $s = 0.624$ ng/g 95% CL = 0.992 ng/g (dry basis)

Reported Results: 7 Quantitative Results: 4




Solid line : exercise assigned value (EAV); dotted line: $z=\pm 1$ (25% from EAV); dotted/dashed line: $z=\pm 2$ (50% from EAV); dashed line: $z=\pm 3$ (75% from EAV)

2,4'-DDD

SRM 1941b

Target Value = no target ng/g (dry basis)

Reported Results: 3 Quantitative Results: 3




Solid line: value from Certificate of Analysis ; dotted line: 95% confidence limits; dashed line: 30% from 95% confidence limits

4,4'-DDD

Sediment XIII (QA05SED13)

Assigned value = 4.18 ng/g $s = 1.42$ ng/g 95% CL = 1.77 ng/g (dry basis)

Reported Results: 7 Quantitative Results: 7




Solid line : exercise assigned value (EAV); dotted line: $z=\pm 1$ (25% from EAV); dotted/dashed line: $z=\pm 2$ (50% from EAV); dashed line: $z=\pm 3$ (75% from EAV)

4,4'-DDD

SRM 1941b

Certified Value = 4.66 ± 0.46 ng/g (dry basis)

Reported Results: 7 Quantitative Results: 6




Solid line: value from Certificate of Analysis ; dotted line: 95% confidence limits; dashed line: 30% from 95% confidence limits

2,4'-DDT

Sediment XIII (QA05SED13)

Assigned value = no target ng/g (dry basis)
Reported Results: 8 Quantitative Results: 3




Solid line : exercise assigned value (EAV); dotted line: $z=\pm 1$ (25% from EAV); dotted/dashed line: $z=\pm 2$ (50% from EAV); dashed line: $z=\pm 3$ (75% from EAV)

2,4'-DDT

SRM 1941b

Target Value = no target ng/g (dry basis)
Reported Results: 8 Quantitative Results: 3




Solid line: value from Certificate of Analysis ; dotted line: 95% confidence limits; dashed line: 30% from 95% confidence limits

cis-nonachlor**Sediment XIII (QA05SED13)**

Assigned value = 0.454 ng/g $s = 0.493$ ng/g 95% CL = 1.22 ng/g (dry basis)

Reported Results: 8 Quantitative Results: 4




Solid line : exercise assigned value (EAV); dotted line: $z=\pm 1$ (25% from EAV); dotted/dashed line: $z=\pm 2$ (50% from EAV); dashed line: $z=\pm 3$ (75% from EAV)

cis-nonachlor**SRM 1941b**

Certified Value = 0.378 ± 0.053 ng/g (dry basis)

Reported Results: 8 Quantitative Results: 4




Solid line: value from Certificate of Analysis ; dotted line: 95% confidence limits; dashed line: 30% from 95% confidence limits

4,4'-DDT**Sediment XIII (QA05SED13)**

Assigned value = 0.537 ng/g $s = 0.174$ ng/g 95% CL = 0.433 ng/g (dry basis)

Reported Results: 8 Quantitative Results: 4




Solid line : exercise assigned value (EAV); dotted line: $z=\pm 1$ (25% from EAV); dotted/dashed line: $z=\pm 2$ (50% from EAV); dashed line: $z=\pm 3$ (75% from EAV)

4,4'-DDT**SRM 1941b**

Reference Value = 1.12 ± 0.42 ng/g (dry basis)

Reported Results: 8 Quantitative Results: 3




Solid line: value from Certificate of Analysis ; dotted line: 95% confidence limits; dashed line: 30% from 95% confidence limits

mirex

Sediment XIII (QA05SED13)

Assigned value = no target ng/g (dry basis)
Reported Results: 7 Quantitative Results: 3




Solid line : exercise assigned value (EAV); dotted line: $z=\pm 1$ (25% from EAV); dotted/dashed line: $z=\pm 2$ (50% from EAV); dashed line: $z=\pm 3$ (75% from EAV)

mirex

SRM 1941b

Target Value = no target ng/g (dry basis)
Reported Results: 7 Quantitative Results: 1




Solid line: value from Certificate of Analysis ; dotted line: 95% confidence limits; dashed line: 30% from 95% confidence limits

PCB 8**Sediment XIII (QA05SED13)**

Assigned value = 1.36 ng/g $s = 0.42$ ng/g 95% CL = 0.39 ng/g (dry basis)

Reported Results: 9 Quantitative Results: 9




Solid line : exercise assigned value (EAV); dotted line: $z=\pm 1$ (25% from EAV); dotted/dashed line: $z=\pm 2$ (50% from EAV); dashed line: $z=\pm 3$ (75% from EAV)

PCB 8**SRM 1941b**

Certified Value = 1.65 ± 0.19 ng/g (dry basis)

Reported Results: 9 Quantitative Results: 9




Solid line: value from Certificate of Analysis ; dotted line: 95% confidence limits; dashed line: 30% from 95% confidence limits

PCB 18**Sediment XIII (QA05SED13)**

Assigned value = 2.04 ng/g $s = 0.48$ ng/g 95% CL = 0.37 ng/g (dry basis)

Reported Results: 10 Quantitative Results: 10




Solid line : exercise assigned value (EAV); dotted line: $z=\pm 1$ (25% from EAV); dotted/dashed line: $z=\pm 2$ (50% from EAV); dashed line: $z=\pm 3$ (75% from EAV)

PCB 18**SRM 1941b**

Certified Value = 2.39 ± 0.29 ng/g (dry basis)

Reported Results: 10 Quantitative Results: 10




Solid line: value from Certificate of Analysis ; dotted line: 95% confidence limits; dashed line: 30% from 95% confidence limits

PCB 28**Sediment XIII (QA05SED13)**

Assigned value = 3.79 ng/g $s = 0.97$ ng/g 95% CL = 0.81 ng/g (dry basis)

Reported Results: 9 Quantitative Results: 9




Solid line : exercise assigned value (EAV); dotted line: $z=\pm 1$ (25% from EAV); dotted/dashed line: $z=\pm 2$ (50% from EAV); dashed line: $z=\pm 3$ (75% from EAV)

PCB 28**SRM 1941b**

Certified Value = 4.52 ± 0.57 ng/g (dry basis)

Reported Results: 9 Quantitative Results: 9




Solid line: value from Certificate of Analysis ; dotted line: 95% confidence limits; dashed line: 30% from 95% confidence limits

PCB 31**Sediment XIII (QA05SED13)**

Assigned value = 2.83 ng/g $s = 0.94$ ng/g 95% CL = 0.87 ng/g (dry basis)

Reported Results: 8 Quantitative Results: 8




Solid line : exercise assigned value (EAV); dotted line: $z=\pm 1$ (25% from EAV); dotted/dashed line: $z=\pm 2$ (50% from EAV); dashed line: $z=\pm 3$ (75% from EAV)

PCB 31**SRM 1941b**

Certified Value = 3.18 ± 0.41 ng/g (dry basis)

Reported Results: 8 Quantitative Results: 8




Solid line: value from Certificate of Analysis ; dotted line: 95% confidence limits; dashed line: 30% from 95% confidence limits

PCB 44**Sediment XIII (QA05SED13)**

Assigned value = 3.47 ng/g $s = 0.99$ ng/g 95% CL = 0.76 ng/g (dry basis)

Reported Results: 10 Quantitative Results: 10




Solid line : exercise assigned value (EAV); dotted line: $z=\pm 1$ (25% from EAV); dotted/dashed line: $z=\pm 2$ (50% from EAV); dashed line: $z=\pm 3$ (75% from EAV)

PCB 44**SRM 1941b**

Certified Value = 3.85 ± 0.20 ng/g (dry basis)

Reported Results: 10 Quantitative Results: 10




Solid line: value from Certificate of Analysis ; dotted line: 95% confidence limits; dashed line: 30% from 95% confidence limits

PCB 49**Sediment XIII (QA05SED13)**

Assigned value = 3.64 ng/g $s = 1.22$ ng/g 95% CL = 1.02 ng/g (dry basis)

Reported Results: 9 Quantitative Results: 9




Solid line : exercise assigned value (EAV); dotted line: $z=\pm 1$ (25% from EAV); dotted/dashed line: $z=\pm 2$ (50% from EAV); dashed line: $z=\pm 3$ (75% from EAV)

PCB 49**SRM 1941b**

Certified Value = 4.34 ± 0.28 ng/g (dry basis)

Reported Results: 9 Quantitative Results: 9




Solid line: value from Certificate of Analysis ; dotted line: 95% confidence limits; dashed line: 30% from 95% confidence limits

PCB 52**Sediment XIII (QA05SED13)**

Assigned value = 4.48 ng/g $s = 1.39$ ng/g 95% CL = 1.07 ng/g (dry basis)

Reported Results: 10 Quantitative Results: 10




Solid line : exercise assigned value (EAV); dotted line: $z=\pm 1$ (25% from EAV); dotted/dashed line: $z=\pm 2$ (50% from EAV); dashed line: $z=\pm 3$ (75% from EAV)

PCB 52**SRM 1941b**

Certified Value = 5.24 ± 0.28 ng/g (dry basis)

Reported Results: 10 Quantitative Results: 10




Solid line: value from Certificate of Analysis ; dotted line: 95% confidence limits; dashed line: 30% from 95% confidence limits

PCB 66**Sediment XIII (QA05SED13)**

Assigned value = 4.32 ng/g $s = 1.41$ ng/g 95% CL = 1.18 ng/g (dry basis)

Reported Results: 10 Quantitative Results: 10




Solid line : exercise assigned value (EAV); dotted line: $z \pm 1$ (25% from EAV); dotted/dashed line: $z \pm 2$ (50% from EAV); dashed line: $z \pm 3$ (75% from EAV)

PCB 66**SRM 1941b**

Certified Value = 4.96 ± 0.53 ng/g (dry basis)

Reported Results: 10 Quantitative Results: 10




Solid line: value from Certificate of Analysis ; dotted line: 95% confidence limits; dashed line: 30% from 95% confidence limits

PCB 95**Sediment XIII (QA05SED13)**

Assigned value = 3.44 ng/g $s = 1.25$ ng/g 95% CL = 1.05 ng/g (dry basis)

Reported Results: 9 Quantitative Results: 9




Solid line : exercise assigned value (EAV); dotted line: $z=\pm 1$ (25% from EAV); dotted/dashed line: $z=\pm 2$ (50% from EAV); dashed line: $z=\pm 3$ (75% from EAV)

PCB 95**SRM 1941b**

Certified Value = 3.93 ± 0.62 ng/g (dry basis)

Reported Results: 9 Quantitative Results: 9




Solid line: value from Certificate of Analysis ; dotted line: 95% confidence limits; dashed line: 30% from 95% confidence limits

PCB 99**Sediment XIII (QA05SED13)**

Assigned value = 2.45 ng/g $s = 0.87$ ng/g 95% CL = 0.73 ng/g (dry basis)

Reported Results: 9 Quantitative Results: 9




Solid line : exercise assigned value (EAV); dotted line: $z=\pm 1$ (25% from EAV); dotted/dashed line: $z=\pm 2$ (50% from EAV); dashed line: $z=\pm 3$ (75% from EAV)

PCB 99**SRM 1941b**

Certified Value = 2.90 ± 0.36 ng/g (dry basis)

Reported Results: 9 Quantitative Results: 9




Solid line: value from Certificate of Analysis ; dotted line: 95% confidence limits; dashed line: 30% from 95% confidence limits

PCB 101**Sediment XIII (QA05SED13)**

Assigned value = 4.86 ng/g $s = 1.36$ ng/g 95% CL = 1.05 ng/g (dry basis)

Reported Results: 10 Quantitative Results: 10




Solid line : exercise assigned value (EAV); dotted line: $z=\pm 1$ (25% from EAV); dotted/dashed line: $z=\pm 2$ (50% from EAV); dashed line: $z=\pm 3$ (75% from EAV)

PCB 101**SRM 1941b**

Certified Value = 5.11 ± 0.34 ng/g (dry basis)

Reported Results: 10 Quantitative Results: 10




Solid line: value from Certificate of Analysis ; dotted line: 95% confidence limits; dashed line: 30% from 95% confidence limits

PCB 105**Sediment XIII (QA05SED13)**

Assigned value = 1.26 ng/g $s = 0.40$ ng/g 95% CL = 0.31 ng/g (dry basis)

Reported Results: 10 Quantitative Results: 10




Solid line : exercise assigned value (EAV); dotted line: $z=\pm 1$ (25% from EAV); dotted/dashed line: $z=\pm 2$ (50% from EAV); dashed line: $z=\pm 3$ (75% from EAV)

PCB 105**SRM 1941b**

Certified Value = 1.43 ± 0.10 ng/g (dry basis)

Reported Results: 10 Quantitative Results: 10




Solid line: value from Certificate of Analysis ; dotted line: 95% confidence limits; dashed line: 30% from 95% confidence limits

PCB 118**Sediment XIII (QA05SED13)**

Assigned value = 3.59 ng/g $s = 1.26$ ng/g 95% CL = 0.96 ng/g (dry basis)

Reported Results: 10 Quantitative Results: 10




Solid line : exercise assigned value (EAV); dotted line: $z=\pm 1$ (25% from EAV); dotted/dashed line: $z=\pm 2$ (50% from EAV); dashed line: $z=\pm 3$ (75% from EAV)

PCB 118**SRM 1941b**

Certified Value = 4.23 ± 0.19 ng/g (dry basis)

Reported Results: 10 Quantitative Results: 10




Solid line: value from Certificate of Analysis ; dotted line: 95% confidence limits; dashed line: 30% from 95% confidence limits

PCB 128**Sediment XIII (QA05SED13)**

Assigned value = 0.684 ng/g $s = 0.275$ ng/g 95% CL = 0.197 ng/g (dry basis)

Reported Results: 10 Quantitative Results: 10




Solid line : exercise assigned value (EAV); dotted line: $z=\pm 1$ (25% from EAV); dotted/dashed line: $z=\pm 2$ (50% from EAV); dashed line: $z=\pm 3$ (75% from EAV)

PCB 128**SRM 1941b**

Certified Value = 0.696 ± 0.044 ng/g (dry basis)

Reported Results: 10 Quantitative Results: 10




Solid line: value from Certificate of Analysis ; dotted line: 95% confidence limits; dashed line: 30% from 95% confidence limits

PCB 138**Sediment XIII (QA05SED13)**

Assigned value = 3.83 ng/g $s = 1.23$ ng/g 95% CL = 1.14 ng/g (dry basis)

Reported Results: 7 Quantitative Results: 7




Solid line : exercise assigned value (EAV); dotted line: $z=\pm 1$ (25% from EAV); dotted/dashed line: $z=\pm 2$ (50% from EAV); dashed line: $z=\pm 3$ (75% from EAV)

PCB 138**SRM 1941b**

Certified Value = 3.60 ± 0.28 ng/g (dry basis)

Reported Results: 7 Quantitative Results: 7




Solid line: value from Certificate of Analysis ; dotted line: 95% confidence limits; dashed line: 30% from 95% confidence limits

PCB 149**Sediment XIII (QA05SED13)**

Assigned value = 4.18 ng/g $s = 1.44$ ng/g 95% CL = 1.20 ng/g (dry basis)

Reported Results: 9 Quantitative Results: 9




Solid line : exercise assigned value (EAV); dotted line: $z=\pm 1$ (25% from EAV); dotted/dashed line: $z=\pm 2$ (50% from EAV); dashed line: $z=\pm 3$ (75% from EAV)

PCB 149**SRM 1941b**

Certified Value = 4.35 ± 0.26 ng/g (dry basis)

Reported Results: 9 Quantitative Results: 9




Solid line: value from Certificate of Analysis ; dotted line: 95% confidence limits; dashed line: 30% from 95% confidence limits

PCB 153**Sediment XIII (QA05SED13)**

Assigned value = 4.99 ng/g $s = 2.46$ ng/g 95% CL = 2.06 ng/g (dry basis)

Reported Results: 8 Quantitative Results: 8




Solid line : exercise assigned value (EAV); dotted line: $z=\pm 1$ (25% from EAV); dotted/dashed line: $z=\pm 2$ (50% from EAV); dashed line: $z=\pm 3$ (75% from EAV)

PCB 153**SRM 1941b**

Certified Value = 5.47 ± 0.32 ng/g (dry basis)

Reported Results: 8 Quantitative Results: 8




Solid line: value from Certificate of Analysis ; dotted line: 95% confidence limits; dashed line: 30% from 95% confidence limits

PCB 156**Sediment XIII (QA05SED13)**

Assigned value = 0.444 ng/g $s = 0.128$ ng/g 95% CL = 0.107 ng/g (dry basis)

Reported Results: 8 Quantitative Results: 8




Solid line : exercise assigned value (EAV); dotted line: $z=\pm 1$ (25% from EAV); dotted/dashed line: $z=\pm 2$ (50% from EAV); dashed line: $z=\pm 3$ (75% from EAV)

PCB 156**SRM 1941b**

Certified Value = 0.507 ± 0.090 ng/g (dry basis)

Reported Results: 7 Quantitative Results: 7




Solid line: value from Certificate of Analysis ; dotted line: 95% confidence limits; dashed line: 30% from 95% confidence limits

PCB 170**Sediment XIII (QA05SED13)**

Assigned value = 1.19 ng/g $s = 0.36$ ng/g 95% CL = 0.28 ng/g (dry basis)

Reported Results: 10 Quantitative Results: 10




Solid line : exercise assigned value (EAV); dotted line: $z=\pm 1$ (25% from EAV); dotted/dashed line: $z=\pm 2$ (50% from EAV); dashed line: $z=\pm 3$ (75% from EAV)

PCB 170**SRM 1941b**

Certified Value = 1.35 ± 0.09 ng/g (dry basis)

Reported Results: 10 Quantitative Results: 10




Solid line: value from Certificate of Analysis ; dotted line: 95% confidence limits; dashed line: 30% from 95% confidence limits

PCB 180**Sediment XIII (QA05SED13)**

Assigned value = 2.97 ng/g $s = 1.15$ ng/g 95% CL = 0.83 ng/g (dry basis)

Reported Results: 10 Quantitative Results: 10




Solid line : exercise assigned value (EAV); dotted line: $z=\pm 1$ (25% from EAV); dotted/dashed line: $z=\pm 2$ (50% from EAV); dashed line: $z=\pm 3$ (75% from EAV)

PCB 180**SRM 1941b**

Certified Value = 3.24 ± 0.51 ng/g (dry basis)

Reported Results: 10 Quantitative Results: 10




Solid line: value from Certificate of Analysis ; dotted line: 95% confidence limits; dashed line: 30% from 95% confidence limits

PCB 187**Sediment XIII (QA05SED13)**

Assigned value = 2.20 ng/g $s = 0.69$ ng/g 95% CL = 0.49 ng/g (dry basis)

Reported Results: 10 Quantitative Results: 10




Solid line : exercise assigned value (EAV); dotted line: $z=\pm 1$ (25% from EAV); dotted/dashed line: $z=\pm 2$ (50% from EAV); dashed line: $z=\pm 3$ (75% from EAV)

PCB 187**SRM 1941b**

Certified Value = 2.17 ± 0.22 ng/g (dry basis)

Reported Results: 10 Quantitative Results: 10




Solid line: value from Certificate of Analysis ; dotted line: 95% confidence limits; dashed line: 30% from 95% confidence limits

PCB 194**Sediment XIII (QA05SED13)**

Assigned value = 1.01 ng/g $s = 0.50$ ng/g 95% CL = 0.46 ng/g (dry basis)

Reported Results: 9 Quantitative Results: 7




Solid line : exercise assigned value (EAV); dotted line: $z=\pm 1$ (25% from EAV); dotted/dashed line: $z=\pm 2$ (50% from EAV); dashed line: $z=\pm 3$ (75% from EAV)

PCB 194**SRM 1941b**

Certified Value = 1.04 ± 0.06 ng/g (dry basis)

Reported Results: 9 Quantitative Results: 8




Solid line: value from Certificate of Analysis ; dotted line: 95% confidence limits; dashed line: 30% from 95% confidence limits

PCB 195**Sediment XIII (QA05SED13)**

Assigned value = 0.291 ng/g $s = 0.166$ ng/g 95% CL = 0.174 ng/g (dry basis)

Reported Results: 10 Quantitative Results: 7




Solid line : exercise assigned value (EAV); dotted line: $z=\pm 1$ (25% from EAV); dotted/dashed line: $z=\pm 2$ (50% from EAV); dashed line: $z=\pm 3$ (75% from EAV)

PCB 195**SRM 1941b**

Certified Value = 0.645 ± 0.060 ng/g (dry basis)

Reported Results: 10 Quantitative Results: 5




Solid line: value from Certificate of Analysis ; dotted line: 95% confidence limits; dashed line: 30% from 95% confidence limits

PCB 206**Sediment XIII (QA05SED13)**

Assigned value = 1.91 ng/g $s = 0.66$ ng/g 95% CL = 0.55 ng/g (dry basis)


Reported Results: 10 Quantitative Results: 10



Solid line : exercise assigned value (EAV); dotted line: $z=\pm 1$ (25% from EAV); dotted/dashed line: $z=\pm 2$ (50% from EAV); dashed line: $z=\pm 3$ (75% from EAV)

PCB 206**SRM 1941b**

Certified Value = 2.42 ± 0.19 ng/g (dry basis)
Reported Results: 9 Quantitative Results: 9




Solid line: value from Certificate of Analysis ; dotted line: 95% confidence limits; dashed line: 30% from 95% confidence limits

PCB 209**Sediment XIII (QA05SED13)**

Assigned value = 4.02 ng/g $s = 1.47$ ng/g 95% CL = 1.36 ng/g (dry basis)

Reported Results: 9 Quantitative Results: 9




Solid line : exercise assigned value (EAV); dotted line: $z=\pm 1$ (25% from EAV); dotted/dashed line: $z=\pm 2$ (50% from EAV); dashed line: $z=\pm 3$ (75% from EAV)

PCB 209**SRM 1941b**

Certified Value = 4.86 ± 0.45 ng/g (dry basis)

Reported Results: 9 Quantitative Results: 9




Solid line: value from Certificate of Analysis ; dotted line: 95% confidence limits; dashed line: 30% from 95% confidence limits

BDE 28**Sediment XIII (QA05SED13)**

Assigned value = 0.056 ng/g $s = 0.005$ ng/g 95% CL = 0.042 ng/g (dry basis)

Reported Results: 5 Quantitative Results: 2




Solid line : exercise assigned value (EAV); dotted line: $z=\pm 1$ (25% from EAV); dotted/dashed line: $z=\pm 2$ (50% from EAV); dashed line: $z=\pm 3$ (75% from EAV)

BDE 28**SRM 1941b**

Target Value = 0.18 ± 0.07 ng/g (dry basis)

Reported Results: 5 Quantitative Results: 2




Solid line: value from Certificate of Analysis ; dotted line: 95% confidence limits; dashed line: 30% from 95% confidence limits

BDE 47**Sediment XIII (QA05SED13)**

Assigned value = 1.03 ng/g $s = 0.35$ ng/g 95% CL = 0.86 ng/g (dry basis)

Reported Results: 5 Quantitative Results: 3




Solid line : exercise assigned value (EAV); dotted line: $z=\pm 1$ (25% from EAV); dotted/dashed line: $z=\pm 2$ (50% from EAV); dashed line: $z=\pm 3$ (75% from EAV)

BDE 47**SRM 1941b**

Target Value = 1.48 ± 0.51 ng/g (dry basis)

Reported Results: 5 Quantitative Results: 3




Solid line: value from Certificate of Analysis ; dotted line: 95% confidence limits; dashed line: 30% from 95% confidence limits

BDE 66**Sediment XIII (QA05SED13)**

Assigned value = 0.038 ng/g $s = 0.009$ ng/g 95% CL = 0.078 ng/g (dry basis)

Reported Results: 5 Quantitative Results: 2




Solid line : exercise assigned value (EAV); dotted line: $z=\pm 1$ (25% from EAV); dotted/dashed line: $z=\pm 2$ (50% from EAV); dashed line: $z=\pm 3$ (75% from EAV)

BDE 66**SRM 1941b**

Target Value = no target ng/g (dry basis)


Reported Results: 5 Quantitative Results: 2



Solid line: value from Certificate of Analysis ; dotted line: 95% confidence limits; dashed line: 30% from 95% confidence limits

BDE 85**Sediment XIII (QA05SED13)**


Assigned value = no target ng/g (dry basis)
Reported Results: 5 Quantitative Results: 2



Solid line : exercise assigned value (EAV); dotted line: $z=\pm 1$ (25% from EAV); dotted/dashed line: $z=\pm 2$ (50% from EAV); dashed line: $z=\pm 3$ (75% from EAV)

BDE 85**SRM 1941b**


Target Value = no target ng/g (dry basis)
Reported Results: 5 Quantitative Results: 2



Solid line: value from Certificate of Analysis ; dotted line: 95% confidence limits; dashed line: 30% from 95% confidence limits

BDE 99**Sediment XIII (QA05SED13)**


Assigned value = no target ng/g (dry basis)
Reported Results: 5 Quantitative Results: 2



Solid line : exercise assigned value (EAV); dotted line: $z=\pm 1$ (25% from EAV); dotted/dashed line: $z=\pm 2$ (50% from EAV); dashed line: $z=\pm 3$ (75% from EAV)

BDE 99**SRM 1941b**


Target Value = 0.62 ± 0.19 ng/g (dry basis)
Reported Results: 5 Quantitative Results: 2



Solid line: value from Certificate of Analysis ; dotted line: 95% confidence limits; dashed line: 30% from 95% confidence limits

BDE 100**Sediment XIII (QA05SED13)**


Assigned value = no target ng/g (dry basis)
Reported Results: 5 Quantitative Results: 2



Solid line : exercise assigned value (EAV); dotted line: $z=\pm 1$ (25% from EAV); dotted/dashed line: $z=\pm 2$ (50% from EAV); dashed line: $z=\pm 3$ (75% from EAV)

BDE 100**SRM 1941b**


Target Value = 0.15 ± 0.06 ng/g (dry basis)
Reported Results: 5 Quantitative Results: 2



Solid line: value from Certificate of Analysis ; dotted line: 95% confidence limits; dashed line: 30% from 95% confidence limits

BDE 138**Sediment XIII (QA05SED13)**


Assigned value = no target ng/g (dry basis)
Reported Results: 4 Quantitative Results: 2



Solid line : exercise assigned value (EAV); dotted line: $z=\pm 1$ (25% from EAV); dotted/dashed line: $z=\pm 2$ (50% from EAV); dashed line: $z=\pm 3$ (75% from EAV)

BDE 138**SRM 1941b**

Target Value = no target ng/g (dry basis)
Reported Results: 4 Quantitative Results: 2




Solid line: value from Certificate of Analysis ; dotted line: 95% confidence limits; dashed line: 30% from 95% confidence limits

BDE 153**Sediment XIII (QA05SED13)**

Assigned value = no target ng/g (dry basis)

Reported Results: 5 Quantitative Results: 2




Solid line : exercise assigned value (EAV); dotted line: $z=\pm 1$ (25% from EAV); dotted/dashed line: $z=\pm 2$ (50% from EAV); dashed line: $z=\pm 3$ (75% from EAV)

BDE 153**SRM 1941b**

Target Value = 0.09 ± 0.04 ng/g (dry basis)


Reported Results: 5 Quantitative Results: 2



Solid line: value from Certificate of Analysis ; dotted line: 95% confidence limits; dashed line: 30% from 95% confidence limits

BDE 154**Sediment XIII (QA05SED13)**


Assigned value = no target ng/g (dry basis)
Reported Results: 5 Quantitative Results: 2



Solid line : exercise assigned value (EAV); dotted line: $z=\pm 1$ (25% from EAV); dotted/dashed line: $z=\pm 2$ (50% from EAV); dashed line: $z=\pm 3$ (75% from EAV)

BDE 154**SRM 1941b**

Target Value = 0.09 ± 0.02 ng/g (dry basis)
Reported Results: 5 Quantitative Results: 2



Solid line: value from Certificate of Analysis ; dotted line: 95% confidence limits; dashed line: 30% from 95% confidence limits

**Appendix I: Data Submitted After the First Draft of this Report
Was Distributed to the Participants**

Two Laboratories Submitted Data after the first draft of this report for Mussel TissueXII (QA05TIS12) and the associated control sample.

Laboratory No.	QA05TIS12			QA05TIS12			SRM 2977		
	Lab 14	Lab 15	RSD	Lab 14	Lab 15	RSD	Lab 14	Lab 15	RSD
	mean	mean		mean	mean		mean	mean	
TEO (percent)	5.7	6.5	3.46%	3.81	4.73	3.96	SRM 1974b	6.51	3.46%
PAHs (ng/g dry mass)	Lab 14	Lab 15	RSD	Value	s	95% CL	Lab 14	Lab 15	RSD
naphthalene	23.0	18.1	15.65%	9.86	6.07	5.61	SRM 1974b	11.8	70.85%
2-methylnaphthalene	13.9	9.51	7.10%	8.00	5.16	5.42	SRM 1974b	9.37	52.15%
1-methylnaphthalene	6.42	5.70	11.95%	3.66	2.44	3.02	SRM 1974b	7.43	42.91%
biphenyl	6.71	8.68	14.00%	1.91	0.77	0.95	SRM 1974b	5.78	69.96%
2,6-dimethylnaphthalene	3.59	4.07	11.36%	4.67	2.85	4.54	SRM 1974b	13.2	13.21%
acenaphthylene	5.09	0.605	5.51%	3.72	2.40	2.01	SRM 1974b	2.20	16.22%
acenaphthene		1.11		2.93	2.21	2.74	SRM 1974b	2.63	9.16%
1,6,7-trimethylnaphthalene	7.21	3.30	15.51%	4.78	1.65	2.62	SRM 1974b	16.9	0.90%
fluorene	4.84	2.95	11.39%	3.64	1.64	1.52	SRM 1974b	6.87	3.04%
phenanthrene	25.8	74.8	8.10%	88.7	17.9	13.8	SRM 1974b	31.3	0.99%
anthracene	5.76	0.981	10.15%	5.79	3.24	2.70	SRM 1974b	2.29	11.06%
1-methylphenanthrene	10.4	55.8	25.88%	89.7	27.7	19.8	SRM 1974b	44.6	10.42%
fluoranthene	167	126	17.79%	133	30	23	SRM 1974b	28.8	4.71%
pyrene	198	126	4.51%	190	40	31	SRM 1974b	58.1	11.54%
benz[a]anthracene	47.5	27.3	13.86%	24.7	7.1	5.4	SRM 1974b	14.0	8.17%
chrysene	90.0	104	11.54%	63.9	22.4	27.8	SRM 1974b	69.0	4.60%
triphenylene		NA		no target			SRM 1974b	NA	
benzo[b]fluoranthene	70.4	87.8	4.26%	47.6	12.6	11.7	SRM 1974b	17.8	6.33%
benzo[j]fluoranthene	52.5	NA	3.36%	no target			SRM 1974b	NA	
benzo[k]fluoranthene		16.6		16.3	2.4	3.0	SRM 1974b	5.39	12.47%
benzo[e]pyrene	95.3	124	1.82%	74.7	14.6	10.5	SRM 1974b	17.4	6.15%
benzo[a]pyrene	28.9	4.28	25.18%	7.25	2.65	2.04	SRM 1974b	5.08	5.94%
perylene	10.3	3.08	3.95%	3.51	1.44	2.30	SRM 1974b	2.70	26.78%
indeno[1,2,3-cd]pyrene	24.1	13.3	8.27%	15.1	6.4	4.9	SRM 1974b	2.78	14.08%
dibenz[a,h]anthracene		4.22		no target			SRM 1974b	2.19	82.20%
benzo[ghi]perylene	33.2	26.5	12.18%	24.6	6.4	4.9	SRM 1974b	8.17	11.64%

From 2977 Certificate of Analysis			
19	5	Reference	
18	5	Reference	
16	5	Reference	
6.8	0.6	Reference	
no target		Target	
no target		Target	
4.2	0.4	Reference	
no target		Target	
10.24	0.43	Certified	
35.1	3.8	Certified	
8	4	Reference	
44	2	Reference	
38.7	1	Certified	
78.9	3.5	Certified	
20.34	0.78	Certified	
49	2	Reference	
39	1	Reference	
11.01	0.28	Certified	
4.6	0.2	Reference	
4	1	Reference	
13.1	1.1	Certified	
8.35	0.72	Certified	
3.5	0.76	Certified	
4.84	0.81	Certified	
1.41	0.19	Certified	
9.53	0.43	Certified	

Pesticides (ng/g dry mass)	QA05TIS12			QA05TIS12			QA05TIS12			SRM 2977		
	Lab 14 mean	RSD	Lab 15 mean	RSD	Value	s	Exercise Assigned	95% CL	Lab 14	Lab 15 mean	RSD	From 2977 Certificate of Analysis
alpha-HCH (a-BHC)	NA		<0.69		no target				SRM 1974b	<0.69		no target
hexachlorobenzene	NA		0.093	50.10%	no target				SRM 1974b	0.266	28.08%	no target
gamma-HCH (g-BHC,lindane)	NA		<0.66		no target				SRM 1974b	<0.66		no target
beta-HCH (b-BHC)	NA		<0.68		no target				SRM 1974b	7.49	7.70%	no target
heptachlor	NA		<0.75		no target				SRM 1974b	<0.75		no target
aldrin	NA		<0.73		no target				SRM 1974b	<0.73		no target
heptachlor epoxide	NA		<0.68		no target				SRM 1974b	<0.68		no target
oxychlorodane	NA		<0.83		no target				SRM 1974b	<0.83		no target
gamma-chlordane	NA		6.13	12.15%	7.45	1.00		0.77	SRM 1974b	0.870	3.93%	no target
2,4'-DDE	7.89	13.43%	1.84	2.44%	no target				SRM 1974b	0.264	7.09%	no target
endosulfan I	NA		<0.74		no target				SRM 1974b	<0.74		no target
cis-chlordane (alpha-chlordane)	NA		8.42	5.58%	12.1	3.4		2.8	SRM 1974b	0.588	9.93%	0.13
trans-nonachlor	NA		8.95	4.10%	9.00	1.24		1.04	SRM 1974b	1.04	10.11%	0.1
dieldrin	NA		2.55	23.53%	6.70	4.42		4.64	SRM 1974b	5.33	3.71%	0.52
4,4'-DDE	38.9	2.80%	33.3	1.22%	33.9	7.6		7.0	SRM 1974b	8.87	3.97%	1.6
2,4'-DDD	7.33	12.68%	10.7	7.35%	8.04	2.86		2.20	SRM 1974b	2.96	3.91%	0.29
endrin	NA		<0.61		no target				SRM 1974b	1.22	11.79%	no target
endosulfan II	NA		<0.74		no target				SRM 1974b	<0.74		no target
4,4'-DDD	21.6	4.75%	21.9	5.87%	21.7	10.8		10.0	SRM 1974b	3.04	14.56%	0.38
2,4'-DDT	<2.98		4.31	0.58%	no target				SRM 1974b	1.44	38.58%	no target
cis-nonachlor	NA		6.03	2.07%	4.27	0.76		0.58	SRM 1974b	1.91	5.46%	no target
4,4'-DDT	1.35	24.79%	2.55	36.81%	1.68	0.63		1.00	SRM 1974b	0.897	8.17%	0.18
mirex	NA		0.345	2.68%	no target				SRM 1974b	0.363	43.50%	no target
endosulfan sulfate	NA		<0.81		no target				SRM 1974b	<0.81		no target
chlorpyrifos	NA		<0.74		no target				SRM 1974b	<0.74		no target

QA05TIS12				QA05TIS12				SRM 2977					
PCBs (ng/g dry mass)	Lab 14		Lab 15		Exercise Assigned		Lab 14	Lab 15 mean	RSD	From 2977 Certificate of Analysis			
	mean	RSD	mean	RSD	Value	s				95% CL			
PCB 8	3.02	7.99%	3.10	5.65%	2.56	0.60	0.75	SRM 1974b	1.84	22.70%	2.1	0.15	Certified
PCB 18	8.22	17.05%	6.68	2.40%	5.71	1.87	1.34	SRM 1974b	2.81	24.14%	2.65	0.3	Certified
PCB 28	26.1	4.87%	30.4	8.21%	23.2	4.6	3.3	SRM 1974b	5.92	9.23%	5.37	0.44	Certified
PCB 31	19.8	3.86%	29.4	6.54%	21.7	3.1	3.2	SRM 1974b	3.47	12.17%	3.92	0.24	Certified
PCB 44	35.7	1.65%	36.8	6.00%	31.4	11.2	8.0	SRM 1974b	3.70	5.73%	3.25	0.63	Certified
PCB 49	NA		50.5	4.12%	37.2	11.3	8.7	SRM 1974b	<0.71		no target		Target
PCB 52	65.0	1.78%	58.7	4.24%	46.6	10.7	7.6	SRM 1974b	10.8	5.52%	8.37	0.54	Certified
PCB 66	58.4	1.96%	56.0	1.07%	48.4	12.4	8.3	SRM 1974b	2.79	11.27%	3.64	0.32	Certified
PCB 95	66.1	7.41%	49.9	7.38%	51.5	9.2	8.5	SRM 1974b	6.47	10.40%	5.39	0.59	Certified
PCB 99	59.5	5.30%	57.6	2.42%	47.0	7.2	6.7	SRM 1974b	2.31	9.50%	no target		Target
PCB 101	123	5.71%	102	1.73%	88.1	14.4	11.1	SRM 1974b	11.0	7.71%	11.2	1.2	Certified
PCB 105	39.1	2.31%	28.5	0.76%	31.1	5.3	3.8	SRM 1974b	2.69	7.09%	3.76	0.49	Certified
PCB 118	112	3.77%	106	1.03%	79.9	15.9	11.4	SRM 1974b	11.1	3.32%	10.5	1	Certified
PCB 128	17.2	3.07%	13.7	2.52%	13.0	2.2	1.7	SRM 1974b	1.71	14.94%	2.49	0.28	Certified
PCB 138	102	3.48%	106	2.17%	64.1	14.9	15.7	SRM 1974b	12.4	1.90%	no target		Target
PCB 149	65.3	5.05%	43.9	3.30%	61.5	11.2	11.7	SRM 1974b	5.94	8.94%	9.23	0.12	Certified
PCB 153	110	3.14%	152	2.03%	85.7	29.6	21.2	SRM 1974b	17.7	10.24%	14.1	1	Certified
PCB 156	5.45	2.86%	12.2	1.86%	5.64	1.25	1.16	SRM 1974b	1.11	13.01%	0.96	0.085	Certified
PCB 170	2.50	0.88%	2.13	42.55%	1.80	0.38	0.29	SRM 1974b	2.97	10.75%	2.95	0.23	Certified
PCB 180	9.57	0.86%	7.97	13.34%	9.29	3.11	2.39	SRM 1974b	5.70	10.45%	6.79	0.67	Certified
PCB 187	25.7	2.29%	24.8	3.51%	18.0	4.0	2.7	SRM 1974b	4.05	6.78%	4.76	0.38	Certified
PCB 194	<0.732		0.967	5.44%	0.501	0.084	0.105	SRM 1974b	0.657	3.16%	0.897	0.042	Certified
PCB 195	<2.13		<0.8		no target	no target		SRM 1974b	<0.8		no target		Target
PCB 206	<0.866		<0.87		no target	no target		SRM 1974b	<0.87		no target		Target
PCB 209	<0.973		<0.73		no target	no target		SRM 1974b	<0.73		no target		Target

BDEs (ng/g dry mass)	QA05TIS12		QA05TIS12		QA05TIS12		SRM 2977		From 2977 Certificate of Analysis	
	Lab 14 mean	RSD	Lab 15 mean	RSD	Lab 14	Lab 15 mean	RSD			
BDE 15	NA		0.433	48.04%	SRM 1974b	0.5		no target	type	
BDE 17	6.23	22.63%	0.767	19.92%	3.72	0.66	1.64	SRM 1974b	0.2	no target
BDE 25	NA		3.93	5.87%	no target			SRM 1974b	1.5	no target
BDE 28	2.05	14.35%	0.500	20.00%	3.08	1.58	1.97	SRM 1974b	0.2	no target
BDE 30	NA		<8.1		no target			SRM 1974b	<8.1	no target
BDE 33	NA		0.533	75.78%	no target			SRM 1974b	0.2	no target
BDE 47	30.4	2.64%	9.63	10.08%	23.3	5.4	6.7	SRM 1974b	1.7	no target
BDE 49	NA		0.500	34.64%	6.75	2.00	4.97	SRM 1974b	0.4	0.03
BDE 66	<1.13		36.2	11.35%	0.984	0.403	0.642	SRM 1974b	48.6	0.062
BDE 71	<1.18		NA		no target			SRM 1974b	NA	no target
BDE 75	NA		0.633	32.87%	no target			SRM 1974b	<8.5	0.013
BDE 85	<6.15		1.03	11.17%	0.418	0.074	0.185	SRM 1974b	0.6	0.166
BDE 99	9.66	5.49%	0.767	27.15%	11.5	3.4	5.3	SRM 1974b	0.5	no target
BDE 100	7.12	8.69%	0.400	25.00%	6.85	1.93	2.40	SRM 1974b	0.4	4.11
BDE 116	NA		16.3	17.30%	no target			SRM 1974b	6.4	1.06
BDE 118	NA		1.13	51.70%	no target			SRM 1974b	0.4	no target
BDE 119	NA		9.17	40.82%	no target			SRM 1974b	1.5	no target
BDE 138	<3.13		4.40	47.73%	no target			SRM 1974b	0.7	no target
BDE 153	<1.00		2.67	31.89%	0.515	0.097	0.155	SRM 1974b	0.6	no target
BDE 154	<5.75		3.57	41.62%	0.550	0.103	0.164	SRM 1974b	0.7	no target
BDE 155	NA		1.17	40.51%	no target			SRM 1974b	0.8	no target
BDE 156	NA		NA	#DIV/0!	no target			SRM 1974b	NA	no target
BDE 181	NA		1.47	43.83%	no target			SRM 1974b	0.6	no target
BDE 183	<1.48		2.90	32.89%	no target			SRM 1974b	0.7	no target
BDE 190	6.17	0.00%	1.53	42.43%	no target			SRM 1974b	0.8	no target
BDE 191	NA		NA		no target			SRM 1974b	NA	no target
BDE 196	NA		NA		no target			SRM 1974b	NA	no target
BDE 197	NA		NA		no target			SRM 1974b	NA	no target
BDE 203	NA		NA		no target			SRM 1974b	NA	no target
BDE 205	NA		NA		no target			SRM 1974b	NA	no target
BDE 206	NA		NA		no target			SRM 1974b	NA	no target
BDE 207	NA		NA		no target			SRM 1974b	NA	no target
BDE 208	NA		NA		no target			SRM 1974b	NA	no target
BDE 209	NA		NA		no target			SRM 1974b	NA	no target

Note only one sample of SRM 2977 analyzed by Lab 15 for BDE congeners

LAB NOTES RECEIVED WITH THE MUSSEL TISSUE DATA:

Lab 14 notes:

SRM 1974b analyzed instead of SRM 2977.

Chrysene and triphenylene co-elute

Benzo(j)fluoranthene and benzo(k)fluoranthene co-elute

PCB 8 was measured as cong 8+5.

PCB 101 was measured as 101+90+89.

PCB 118 measured as 118+106

PCB 128 measured as 128+167

PCB 138 measured as 164+163+138

PCB 170 measured as 170+190

Lab 15 notes:

PCB101 coelutes with PCB90

PCB153 coelutes with PCB132 and 168

PCB170 coelutes with PCB190

PCB 8 coelutes with PCB 5

PCB 195 coelutes with PCB208

<symbol refers to values less than our MDL

the chrysene number is the sum of chrysene and triphenylene, we cannot resolve the two

we do not analyze for benzo(j)fluoranthene because it coelutes with benzo(b)fluoranthene

Tissue MDL values are high due to limited sample volume and reporting data on a wet weight basis.

Our MDLs are based on 13.0g wet weight of mussel. Therefore MDLs reported here are approximately 2X higher than those on 13 g wet. Additionally, sample PAH concentrations are low which is compounded by the small sample volume.

BDE71 coelutes with BDE47

Three laboratories submitted data for QA05SED13 and the associated control sample after the first draft of this report

	QA05SED13			QA05SED13			QA05SED13			QA05SED13		
	Lab 13		RSD	Lab 14		RSD	Lab 15		RSD	Exercise Assigned		95% CL
	mean	RSD		mean	RSD		mean	RSD		Value	s	
Water (percent)	35.0	21.57%		45.7			46.6	0.99%		47.4	3.2	2.3
TOC (percent)	NA						2.88	0.36%		2.70	0.58	0.92
PAHs												
ug/g dry mass	QA05SED13		RSD	QA05SED13		RSD	QA05SED13		RSD	Exercise Assigned		95% CL
	Lab 13	mean		Lab 14	mean		Lab 15	mean		Value	s	
	609	22.15%		881	6.61%		709	18.02%		785	186	172
	203	24.73%		322	3.12%		210	19.24%		219	66	61
	82.6	19.68%		131	5.04%		104	17.61%		98.2	24.0	20.0
	34.3	16.98%		82.7	2.22%		60.9	15.08%		65.8	23.6	21.8
	58.3	20.06%		87.4	11.83%		50.6	19.48%		81.3	43.6	40.3
	42.1	19.26%		59.3	4.89%		52.9	17.09%		45.1	13.8	11.5
	20.5	17.33%		42.3	11.46%		25.8	12.36%		28.9	8.4	7.0
	NA			73.3	4.81%		16.8	20.24%		no target		
1,6,7-trimethylnaphthalene	41.4	27.33%		79.9	10.19%		53.8	16.71%		56.1	12.2	10.2
fluorene	262	22.42%		440	5.38%		350	13.34%		306	89	75
phenanthrene	125	17.46%		209	3.92%		141	17.75%		137	47	40
1-methylphenanthrene	55.5	26.51%		84.4	2.53%		53.8	10.19%		55.4	18.2	19.1
fluoranthene	521	12.10%		721	3.64%		560	8.24%		496	140	117
pyrene	445	18.36%		598	2.80%		451	9.97%		421	142	118
benz[a]anthracene	198	14.48%		432	1.75%		295	9.58%		241	80	67
chrysene	267	19.16%		377	5.12%		410	9.68%		219	62	65
triphenylene	NA						NA			no target		
benzo[b]fluoranthene	279	14.07%		403	8.48%		567	8.38%		413	174	183
benzo[k]fluoranthene	NA			432	6.02%		NA			no target		
benzo[a]pyrene	265	12.82%		281	4.49%		301	7.48%		180	56	47
benzo[e]pyrene	205	13.71%		350	9.62%		287	6.20%		286	97	75
benzo[a]pyrene	188	14.32%		350	9.62%		399	8.66%		282	120	111
perylene	277	13.01%		416	3.30%		342	13.09%		311	131	138
indeno[1,2,3-cd]pyrene	153	9.90%		353	6.36%		342	15.22%		258	101	85
dibenz[a,h]anthracene	40.7	26.35%					78.5	18.14%		41.9	21.2	33.8
benzo[ghi]perylene	167	10.15%		276	3.02%		277	12.41%		244	96	81

SRM 1941b	SRM 1941b			SRM 1941b			SRM 1941b		
	Lab 13		RSD	Lab 14		RSD	Lab 15		RSD
	mean	RSD		mean	RSD		mean	RSD	
NA	NA			0.0239					
NA	NA								
SRM 1941b	SRM 1941b		RSD	SRM 1941b		RSD	SRM 1941b		RSD
	mean	RSD		mean	RSD		mean	RSD	
	859	5.27%		1006	9.23%		811	9.23%	
	265	1.57%		282	9.63%		264	9.63%	
	113	2.56%		134	12.41%		135	12.41%	
	56.5	14.42%		83.5	2.76%		57.4	2.76%	
	94.4	6.70%		96.6	18.81%		101	18.81%	
	57.5	9.37%		59.0	11.86%		69.7	11.86%	
	28.6	15.06%		43.2	3.94%		28.5	3.94%	
	NA			62.2	3.22%		17.8	3.22%	
Lab 13	SRM 1941b		RSD	SRM 1941b		RSD	SRM 1941b		RSD
	mean	RSD		mean	RSD		mean	RSD	
	80.3	11.94%		76.5	11.50%		61.4	11.50%	
	430	4.42%		458	7.45%		376	7.45%	
	184	12.89%		209	3.65%		185	3.65%	
	70.8	8.40%		83.9	4.43%		69.8	4.43%	
	841	0.89%		711	8.35%		567	8.35%	
	693	4.41%		585	9.61%		491	9.61%	
	287	8.26%		454	1.60%		334	1.60%	
	415	4.94%		368	9.12%		393	9.12%	
Lab 14	SRM 1941b		RSD	SRM 1941b		RSD	SRM 1941b		RSD
	mean	RSD		mean	RSD		mean	RSD	
	NA			387	5.92%		444	5.92%	
	488	2.38%		409	9.89%		NA	9.89%	
	NA			378	3.21%		204	3.21%	
	330	3.74%		311	12.68%		336	12.68%	
	278	5.89%		454	5.27%		339	5.27%	
	364	2.33%		410	1.76%		345	1.76%	
	241	13.62%		360	4.98%		436	4.98%	
	47.9	10.74%		274	7.69%		76.9	7.69%	
Lab 15	SRM 1941b		RSD	SRM 1941b		RSD	SRM 1941b		RSD
	mean	RSD		mean	RSD		mean	RSD	
	278	10.38%		274	7.69%		312	7.69%	
	278	10.38%		274	7.69%		312	7.69%	
	278	10.38%		274	7.69%		312	7.69%	
	278	10.38%		274	7.69%		312	7.69%	
	278	10.38%		274	7.69%		312	7.69%	
	278	10.38%		274	7.69%		312	7.69%	
	278	10.38%		274	7.69%		312	7.69%	
	278	10.38%		274	7.69%		312	7.69%	
From 1941b Certif.									
95%CL									
type									
Certified									
Reference									
Reference									
Reference									
Reference									
Reference									
Reference									
Reference									
Reference									
Reference									
Reference									
Reference									
Reference									
Reference									
Reference									
Reference									
Reference									
Reference									
Reference									
Reference									
Reference									
Reference									
Reference									
Reference									
Reference									
Reference									
Reference									
Reference									
Reference									
Reference									
Reference									
Reference									
Reference									
Reference									
Reference									
Reference									
Reference									
Reference									
Reference									
Reference									
Reference									
Reference									
Reference									
Reference									
Reference									
Reference									
Reference									
Reference									
Reference									
Reference									
Reference									
Reference									
Reference									
Reference									
Reference									
Reference									
Reference									
Reference									
Reference									
Reference									
Reference									
Reference									
Reference									
Reference									
Reference									
Reference									
Reference									
Reference									
Reference									
Reference									
Reference									
Reference									
Reference									
Reference									
Reference									
Reference									
Reference									
Reference									
Reference									
Reference									
Reference									
Reference									
Reference									
Reference									
Reference									
Reference									
Reference									
Reference									
Reference									
Reference									
Reference									
Reference									
Reference									
Reference									
Reference									
Reference									
Reference									
Reference									
Reference									
Reference									
Reference									
Reference									
Reference									
Reference									
Reference									
Reference									
Reference									
Reference									
Reference									
Reference									
Reference									
Reference									
Reference									
Reference									
Reference									
Reference									
Reference									
Reference									
Reference									
Reference									
Reference									
Reference									
Reference									
Reference									
Reference									
Reference									
Reference									
Reference									
Reference									
Reference									
Reference									
Reference									
Reference									
Reference									
Reference									
Reference									
Reference									
Reference									
Reference									
Reference									
Reference									
Reference									
Reference									
Reference									
Reference									
Reference									
Reference									
Reference									
Reference									
Reference									
Reference									
Reference									
Reference									
Reference									
Reference									
Reference									
Reference									
Reference									
Reference									
Reference									
Reference									
Reference									
Reference									
Reference									
Reference									
Reference									
Reference									
Reference									
Reference									
Reference									
Reference									
Reference									
Reference									
Reference									
Reference									
Reference									
Reference									
Reference									
Reference									
Reference									
Reference									
Reference									
Reference									
Reference									
Reference									
Reference									
Reference									
Reference									
Reference									
Reference									
Reference									
Reference									
Reference									
Reference									
Reference									
Reference									
Reference									
Reference									
Reference									
Reference									
Reference									
Reference									
Reference									
Reference									
Reference									
Reference									
Reference									
Reference									
Reference									
Reference									
Reference									
Reference									
Reference									
Reference									
Reference									
Reference									
Reference									
Reference									
Reference									
Reference									
Reference									
Reference									
Reference									
Reference									
Reference									
Reference									
Reference									
Reference									
Reference									
Reference									
Reference									

Note: Only one sample of SRM 1941b was analyzed by Lab 15

Pesticides ng/g dry mass	QA05SED13		QA05SED13		QA05SED13		QA05SED13	
	Lab 13	RSD	Lab 14	RSD	Lab 15	RSD	Value	Exercise Assigned s
alpha-HCH (α-BHC)	0.032	26.90%	NA	NA	<0.93	10.24%	no target	1.90
hexachlorobenzene	4.42	14.33%	NA	NA	5.83	5.28%	5.38	2.00
gamma-HCH (γ-BHC, lindane)	0.035	20.44%	NA	NA	1.08	NA	no target	NA
beta-HCH (β-BHC)	0.025	28.28%	NA	NA	<0.76	NA	no target	NA
heptachlor	0.014	57.08%	NA	NA	<0.68	NA	no target	NA
aldrin	0.033	18.48%	NA	NA	<0.54	NA	no target	NA
heptachlor epoxide	0.025	38.91%	NA	NA	<0.68	NA	no target	NA
oxychlorodane	0.003	4.68%	NA	NA	<0.91	NA	no target	NA
gamma-chlordane	0.321	19.14%	NA	NA	<0.5	NA	0.572	0.148
2,4'-DDE	0.184	15.72%	<0.264	NA	0.064	19.03%	0.380	0.103
endosulfan I	NA	NA	NA	NA	<1.06	NA	no target	NA
cis-chlordane (alpha-chlordane)	0.329	24.21%	NA	NA	0.252	19.53%	0.482	0.058
trans-nonachlor	0.137	20.82%	NA	NA	0.175	6.47%	0.286	0.089
dieldrin	0.266	19.62%	NA	NA	0.728	8.73%	0.386	0.034
4,4'-DDE	1.76	18.86%	5.63	5.90%	3.71	9.26%	3.44	1.08
2,4'-DDD	0.237	11.07%	<1.06	NA	1.31	16.57%	0.927	0.624
endrin	0.037	NA	NA	NA	<2	NA	no target	NA
endosulfan II	NA	NA	NA	NA	<1.43	NA	no target	NA
4,4'-DDD	2.29	17.94%	3.10	4.17%	5.49	11.95%	4.18	1.42
2,4'-DDT	0.052	17.25%	<1.19	NA	NA	NA	no target	NA
cis-nonachlor	0.082	21.26%	NA	NA	0.177	5.79%	0.454	0.493
4,4'-DDT	0.273	2.65%	<2.65	NA	0.825	7.43%	0.537	0.174
mirex	0.010	35.07%	NA	NA	0.063	6.48%	no target	NA
endosulfan sulfate	NA	NA	NA	NA	<0.69	NA	no target	NA
chlorpyrifos	NA	NA	NA	NA	0.461	26.44%	no target	NA

SRM 1941b	SRM 1941b		SRM 1941b		SRM 1941b		SRM 1941b	
	Lab 13	RSD	Lab 14	RSD	Lab 15	RSD	From 1941b Certif	type
mean	0.036	1.89%	NA	NA	<0.33	NA	no target	Target
	9.80	NA	NA	NA	7.96	NA	5.83	0.38
	0.042	NA	NA	NA	<0.18	NA	no target	Certified
	0.033	NA	NA	NA	<0.27	NA	no target	Target
	0.018	NA	NA	NA	<0.24	NA	no target	Target
	0.042	NA	NA	NA	<0.19	NA	no target	Target
	0.041	NA	NA	NA	1.66	NA	no target	Target
	0.009	NA	NA	NA	0.458	NA	no target	Target
	0.561	2.52%	NA	NA	<0.18	NA	0.566	0.093
	0.288	15.71%	<0.420	NA	0.247	NA	0.38	0.12
			NA	NA	<0.37	NA	no target	Reference
	0.527	6.35%	NA	NA	0.822	NA	0.85	0.11
	0.238	1.99%	NA	NA	0.341	NA	0.438	0.073
	0.408	10.71%	NA	NA	0.810	NA	no target	Certified
	2.95	6.62%	4.84	17.28%	3.45	NA	3.22	0.28
	0.412	2.81%	<0.786	NA	2.21	NA	no target	Certified
	<0.907	NA	NA	NA	<0.7	NA	no target	Target
			NA	NA	<0.5	NA	no target	Target
	3.51	6.35%	3.08	11.11%	4.01	NA	4.66	0.46
	0.079	NA	<1.90	NA	0.223	NA	no target	Target
	0.166	11.36%	NA	NA	0.493	NA	0.378	0.053
	0.492	61.32%	<4.22	NA	1.69	NA	1.12	0.42
	0.037	68.01%	NA	NA	<0.27	NA	no target	Reference
	NA	NA	NA	NA	<0.24	NA	no target	Target
	NA	NA	NA	NA	2.07	NA	no target	Target

Note: Only one sample of SRM 1941b was analyzed by Lab 15

PCBs µg/g dry mass	QA05SED13		QA05SED13		QA05SED13		QA05SED13		QA05SED13	
	Lab 13	RSD	Lab 14	RSD	Lab 15	RSD	Value	s	Exercise Assigned	95% CL
	mean		mean		mean					
PCB 8	0.697	20.61%	2.34	11.16%	<1.09	15.71%	1.36	0.42		0.39
PCB 18	1.18	18.55%	3.01	6.72%	2.00	9.36%	2.04	0.48		0.37
PCB 28	2.30	32.13%	6.00	2.27%	5.40	10.91%	3.79	0.97		0.81
PCB 31	2.04	35.01%	3.56	5.26%	2.96	10.91%	2.83	0.94		0.87
PCB 44	2.75	16.23%	4.15	6.30%	3.75	6.34%	3.47	0.99		0.76
PCB 49	1.87	19.11%	NA		4.00	0.32%	3.64	1.22		1.02
PCB 52	2.80	24.26%	7.28	4.99%	5.26	0.63%	4.48	1.39		1.07
PCB 66	2.93	29.68%	4.10	2.13%	5.19	3.96%	4.32	1.41		1.18
PCB 95	1.68	15.33%	5.61	7.03%	3.93	12.81%	3.44	1.25		1.05
PCB 99	1.41	16.62%	3.59	13.44%	2.54	8.27%	2.45	0.87		0.73
PCB 101	2.42	17.44%	7.12	6.18%	6.01	12.97%	4.86	1.36		1.05
PCB 105	0.572	14.99%	1.37	2.48%	1.93	19.91%	1.26	0.40		0.31
PCB 118	1.81	16.14%	4.52	10.86%	4.23	15.38%	3.59	1.26		0.96
PCB 128	0.361	17.58%	0.832	14.73%	0.433	14.12%	0.684	0.275		0.197
PCB 138	2.37	17.51%	5.47	4.67%	3.80	7.69%	3.83	1.23		1.14
PCB 149	1.94	15.08%	5.32	5.55%	1.83	5.40%	4.18	1.44		1.20
PCB 153	2.50	20.27%	6.04	4.34%	6.23	6.20%	4.99	2.46		2.06
PCB 156	0.231	18.71%	0.421	11.59%	2.12	8.72%	0.128	0.128		0.107
PCB 170	0.553	19.78%	1.35	6.57%	2.23	3.50%	1.19	0.36		0.28
PCB 180	1.50	17.22%	3.43	7.54%	4.35	4.16%	2.97	1.15		0.83
PCB 187	0.761	18.78%	2.51	3.62%	2.49	23.24%	2.20	0.69		0.49
PCB 194	0.413	21.17%	0.972	8.30%	<0.43		1.01	0.50		0.46
PCB 195	0.141	22.97%	<0.850		1.06	20.14%	0.291	0.166		0.174
PCB 206	NA		2.73	6.89%	2.95	7.31%	1.91	0.66		0.55
PCB 209	NA		5.92	7.59%	7.35	6.77%	4.02	1.47		1.36

SRM 1941b Lab 13	SRM 1941b Lab 14		SRM 1941b Lab 15		From 1941b Certif	
	mean	RSD	mean	RSD	conc	95%CL type
	0.920	2.78%	2.01	14.43%	1.65	0.19 Certified
	1.64	6.93%	2.93	25.53%	2.39	0.29 Certified
	3.15	2.40%	5.66	4.80%	4.52	0.57 Certified
	2.71	2.16%	3.23	8.66%	3.18	0.41 Certified
	5.51	2.46%	4.09	1.94%	3.85	0.2 Certified
	3.01	2.53%	NA		4.34	0.28 Certified
	4.22	2.05%	6.23	10.97%	5.24	0.28 Certified
	4.51	3.92%	4.27	5.47%	4.96	0.53 Certified
	2.63	1.66%	5.28	5.12%	3.93	0.62 Certified
	2.28	4.94%	3.62	3.33%	2.9	0.36 Certified
	3.78	0.55%	6.83	6.55%	5.11	0.34 Certified
	0.95	2.20%	1.39	9.80%	1.43	0.1 Certified
	2.94	1.61%	4.59	12.89%	4.23	0.19 Certified
	0.63	3.87%	0.776	14.46%	0.696	0.044 Certified
	4.01	4.64%	5.42	5.77%	3.6	0.28 Certified
	3.20	1.90%	5.10	3.76%	4.35	0.26 Certified
	4.05	2.34%	5.89	3.07%	5.47	0.32 Certified
	0.374	0.77%	0.369	3.10%	0.507	0.09 Certified
	0.926	3.09%	1.36	9.51%	1.35	0.09 Certified
	2.57	2.72%	3.34	1.77%	3.24	0.51 Certified
	1.33	4.39%	2.45	0.92%	2.17	0.22 Certified
	0.693	1.52%	0.982	2.55%	1.04	0.06 Certified
	0.244	8.13%	<1.36		0.645	0.06 Certified
	NA		2.83	1.85%	2.42	0.19 Certified
	NA		6.05	1.82%	4.86	0.45 Certified

Note: Only one sample of SRM 1941b was analyzed by Lab 15

BDEs ng/g dry mass	QA05SED13			QA05SED13			QA05SED13			QA05SED13		
	Lab 13	mean	RSD	Lab 14	mean	RSD	Lab 15	mean	RSD	Value	Exercise Assigned	95% CL
BDE 15		0.126	15.62%	NA	0.967	11.66%				no target		
BDE 17		0.039	10.24%	<0.357	<0.44					no target		
BDE 25		0.036	8.83%	NA	<0.44					no target		
BDE 28		0.036	8.83%	<0.532	<0.27					0.056	0.005	0.042
BDE 30		0.028	55.16%	NA	<0.44					no target		
BDE 33		0.380	23.61%	NA	<0.44					no target		
BDE 47		0.091	24.31%	0.890	1.09	9.53%				1.03	0.35	0.86
BDE 49		0.023	9.58%	NA	<0.74					no target		
BDE 66		0.008	1.34%	<0.450	<1.41					0.038	0.009	0.078
BDE 71		0.008	13.92%	<0.471	NA					no target		
BDE 75		0.012	13.92%	NA	<0.74					no target		
BDE 85		0.011	94.73%	<2.46	<0.58					no target		
BDE 99		0.252	26.15%	0.710	0.833	22.72%				no target		
BDE 100		0.060	23.85%	<0.418	<0.78					no target		
BDE 116		<0.00291	NA	NA	<0.78					no target		
BDE 118		NA	NA	NA	<0.78					no target		
BDE 119		0.003	18.40%	NA	<0.78					no target		
BDE 138		0.009	47.80%	<1.25	<0.65					no target		
BDE 153		0.044	19.85%	0.121	<1.93	19.78%				no target		
BDE 154		0.034	41.31%	0.207	8.27%					no target		
BDE 155		0.007	NA	NA	<0.95					no target		
BDE 156		NA	NA	NA	NA					no target		
BDE 181		<0.00961	NA	NA	<1.41					no target		
BDE 183		0.064	132.57%	<0.384	<1.41					no target		
BDE 190		0.011	62.39%	<2.47	<1.65					no target		
BDE 191		NA	NA	NA	NA					no target		
BDE 196		NA	NA	NA	NA					no target		
BDE 197		NA	83.65%	NA	NA					no target		
BDE 203		0.044	NA	NA	NA					no target		
BDE 205		<0.0524	NA	NA	NA					no target		
BDE 206		0.256	28.83%	NA	NA					no target		
BDE 207		0.153	34.92%	NA	NA					no target		
BDE 208		0.119	28.78%	NA	NA					no target		
BDE 209		11.4	16.11%	NA	NA					no target		

SRM 1941b Lab 13	SRM 1941b Lab 14			SRM 1941b Lab 15			From 1941b Certif 95%CL	type
	mean	RSD	mean	RSD	mean	conc.		
coelute (17/25)	0.069	34.37%	<0.586		NA	no target	no target	Target
	0.135	5.47%	<0.847		NA	no target	0.18	Target
	0.047	10.47%	NA		NA	w/ BDE 28	0.07	Target
coelute (28/33)	1.34	7.79%	1.87	6.21%	NA	no target	0.51	Target
	0.133	74.26%	NA		NA	no target		Target
	0.045	4.56%	<0.716		NA	no target		Target
	0.167	32.97%	<0.750		NA	no target		Target
	0.019	30.29%	NA		NA	no target		Target
	0.025		<3.91		NA	no target		Target
	0.398	12.88%	0.667	54.24%	NA	0.62	0.19	Target
	0.103	13.83%	<0.665		NA	0.15	0.06	Target
	<0.00309		NA		NA	no target		Target
	0.035	79.52%	NA		NA	no target		Target
	0.007	3.24%	<1.99		NA	no target		Target
	0.064	9.30%	0.114	35.29%	NA	0.09	0.04	Target
	0.058	7.27%	0.199	11.19%	NA	0.09	0.02	Target
	0.009		NA		NA	no target		Target
	NA		NA		NA	no target		Target
	<0.00325		NA		NA	no target		Target
	0.026	20.98%	<0.943		NA	0.05	0.02	Target
	0.002		<3.93		NA	no target		Target
	NA		NA		NA	no target		Target
	NA		NA		NA	no target		Target
	0.065	37.74%	NA		NA	no target		Target
	<0.0236		NA		NA	no target		Target
	0.374	14.59%	NA		NA	no target		Target
	0.215	26.25%	NA		NA	no target		Target
	0.200	45.62%	NA		NA	no target		Target
	21.2	14.24%	NA		NA	24.11	14.97	Target

Note: No data for Lab 15 on SRM 1941b

LAB NOTES SUBMITTED WITH THE SEDIMENT DATA

Lab 13 Notes:

other= ion ratio did not meet method criteria

other= coelution

for PBDEs, PCBs and pesticides, samples were concentrated down to 100 uL

	QA05SED13		SRM 1941b	
	Lab 13		Lab 13	
	mean	RSD	mean	RSD
C2-NAPHTHALENES	277	23.56%	246	49.62%
C1-NAPHTHALENES	290	24.87%	296	53.63%
2,3,5-TRIMETHYLNAPHTHALENE	22.0	21.66%	23.6	22.67%
C1-CHRYSENES	141	4.52%	188	14.71%
C1-DIBENZOTHIOPHENES	< 3.5		23.9	
C1-FLUORANTHENES/PYRENES	244	20.77%	312	2.72%
C1-FLUORENES	< 3.5		116	
C1-PHENANTHRENES/ANTHRACENES	308	1.95%	312	52.09%
C2-CHRYSENES	< 3.5		<10	
C2-DIBENZOTHIOPHENES	< 3.5		<10	
C2-FLUORENES	< 3.5		773	
C2-PHENANTHRENES/ANTHRACENES	327	9.88%	357	
C3-CHRYSENES	< 3.5		<10	
C3-DIBENZOTHIOPHENES	< 3.5		85.3	
C3-FLUORENES	< 3.5		<10	
C3-NAPHTHALENES	190	23.91%	94.6	61.75%
C3-PHENANTHRENES/ANTHRACENES	< 3.5		<10	
C4-CHRYSENES	< 3.5		10.9	
C4-NAPHTHALENES	< 3.5		110	
C4-PHENANTHRENES/ANTHRACENES	< 3.5		23.6	
DIBENZOTHIOPHENE	30.3	16.95%	45.8	26.25%
IUPAC# 33	0.975	35.52%	1.35	1.96%
IUPAC# 56	1.20	30.86%	1.76	2.56%
IUPAC# 60	0.264	34.31%	0.381	1.24%
IUPAC# 70/74	4.15	32.35%	6.06	3.07%
IUPAC# 87/97	1.27	13.59%	2.20	0.94%
IUPAC# 110	2.67	12.36%	4.53	1.96%
IUPAC# 132	0.588	16.54%	1.00	3.11%
IUPAC# 141	0.345	20.29%	0.588	7.43%
IUPAC# 151	0.896	15.68%	1.47	3.74%
IUPAC# 158	0.196	17.00%	0.322	3.11%
IUPAC# 174	0.657	20.50%	1.07	3.88%
IUPAC# 177	0.302	16.82%	0.536	3.17%
IUPAC# 183	0.442	18.51%	0.741	2.51%
IUPAC# 201	0.098	19.35%	0.174	2.07%
IUPAC# 203	0.344	18.72%	0.619	2.18%
DELTA BHC	0.016	63.46%	<0.44	
PBDE # 7	0.031	11.98%	0.043	3.93%
PBDE # 8/11	0.032	16.85%	0.043	15.20%
PBDE # 10	0.002	23.60%	0.002	12.86%
PBDE # 12/13	0.012	6.22%	0.015	1.72%
PBDE # 32	0.007	27.24%	0.046	5.58%
PBDE # 35	0.008	16.42%	0.103	154.61%
PBDE # 37	0.006	18.82%	0.007	22.39%
PBDE # 77	0.003	11.45%	0.001	
PBDE # 79	0.002		0.005	
PBDE # 105	< 0.00723		<0.00588	
PBDE # 126	< 0.00232		0.011	4.00%
PBDE # 128	0.004		<0.00543	
PBDE # 140	0.003		0.003	
PBDE # 204	0.056	124.34%	0.047	47.04%

Lab 14 notes:

Chrysene and triphenylene co-elute

Benzo(j)fluoranthene and benzo(k)fluoranthene co-elute

PCB 8 was measured as cong 8+5.

PCB 101 was measured as 101+90+89.

PCB 118 measured as 118+106

PCB 128 measured as 128+167

PCB 138 measured as 164+163+138

PCB 170 measured as 170+190

Lab 15 notes:

the chrysene number is the sum of chrysene and triphenylene, we cannot resolve the two

we can not separate benzo(j)fluoranthene because under our conditions it coelutes with

benzo(b)fluoranthene

<symbol refers to values less than our MDL

PCB101 coelutes with PCB90

PCB153 coelutes with PCB132 and 168

PCB170 coelutes with PCB190

PCB8 coelutes with PCB5

PCB195 coelutes with PCB208

Total carbon and total organic carbon are measured independently in oven-dried sediments

and soils using a LECO CR-412 Carbon Determinator

BDE71 coelutes with BDE47

Appendix J: List of Laboratories Participating in 2005 Intercomparison Exercises

For this exercise, data were received from the following laboratories within the required timeframe. (This listing does NOT correspond to the laboratory number identification codes used in this report which were assigned in order of receipt of data with the exception of NIST which is Laboratory #1 in this exercise. The same code was used for both exercises.)

Academy of Natural Sciences
1900 Benjamin Franklin Parkway
Philadelphia, PA 19103
Jeffrey Ashley, Linda Zaoudeh, and Mike Schafer

Alpha Woods Hole Laboratories
375 Paramount Dr, Suite B
Raynham, MA 02767
Pete Kane and Elizabeth Porta

AXYS Analytical
2045 Mills Rd West / PO Box 2219
Sidney, BC V8L 3S8
Canada
Dale Hoover

Battelle Columbus
505 King Ave
Columbus, OH 43201
Karen Tracy and Mary Schrock

Battelle Duxbury Operations
397 Washington Street
Duxbury, MA 02332
Carole-Sue Peven McCarthy

East Bay Municipal Utility District
2020 Wake Avenue
Oakland, CA 94607
Saskai van Bergen and Francois Rodigari

Environment Canada
Environmental Science Center
Corner Morton & Université Ave
Moncton, NB E1A3E9 Canada
Jamie Aubé

Massachusetts Water Resources Authority
100 Tafts Ave.
Winthrop, MA 02152
Jennifer Prasse

NIST
100 Bureau Drive, Stop 8392
Gaithersburg, MD 20899-8392
Michele Schantz

NIST-Charleston Laboratory
331 Fort Johnson Road
Charleston, SC 29412-9110
John Kucklick, Stacy Vander Pol, and Aurore Guichard

NOAA Fisheries / ABL
11305 Glacier Hwy
Juneau, AK 99801
Marie Larsen

NOAA-NMFS
2725 Montlake Boulevard, East
Seattle, WA 98112
Donald Brown / Jennie Bolton

NOAA-NOS
Hollings Marine Laboratory
331 Fort Johnson Road
Charleston, SC 29412
Ed Wirth

STL Sacramento
880 Riverside Pkwy
West Sacramento, CA 95605
Michael Flournoy

TDI-Brooks International
B&B Laboratories
1902 Pinon
College Station, TX 77845
Juan Ramirez

Wadsworth Center, NYSDOH
Empire State Plaza
P-1 North Dock (Rm D520)
Albany, NY 12237
Chia-Swee Hong