

Time & Frequency Bulletin

No. 394 September 1990
(NISTIR 90-3940-9)

National
Institute of
Standards &
Technology



NIST TIME & FREQUENCY BULLETIN
(Supersedes No. 393 August 1990)

NO. 394

SEPTEMBER 1990

1. GENERAL BACKGROUND INFORMATION	1
2. TIME SCALE INFORMATION	1
International Timing Center comparisons via GPS common-view	2
3. UT1 CORRECTIONS AND LEAP SECOND ADJUSTMENTS	2
4. PHASE DEVIATIONS FOR WWVB AND LORAN-C	3
5. GOES TIME CODE INFORMATION	4
6. BROADCAST OUTAGES OVER FIVE MINUTES AND WWVB PHASE PERTURBATIONS	4
7. NOTES ON NIST TIME SCALES AND PRIMARY STANDARDS	5
8. SPECIAL ANNOUNCEMENTS	7

This bulletin is published monthly. Address correspondence to:

Gwen E. Bennett, Editor
Time and Frequency Division
National Institute of Standards & Technology
325 Broadway
Boulder, CO 80303
(303) 497-3295

NOTE TO SUBSCRIBERS: Please include your address label (or a copy)
with any correspondence regarding this bulletin.

1. GENERAL BACKGROUND INFORMATION

ABBREVIATIONS AND ACRONYMS USED IN THIS BULLETIN

APL	-	John Hopkins University Applied Physics Laboratory	
BIH	-	International Time Bureau, France	
CCIR	-	International Radio Consultative Committee	
CRL	-	Communications Research Laboratories, Japan	
Cs	-	Cesium standard	
CSIRO	-	Commonwealth Scientific and Industrial Research Organization, Australia	
GOES	-	Geostationary Operational Environmental Satellite	
GPS	-	Global Positioning System	
IEN	-	National Institute of Electronics, Italy	
INPL	-	National Physical Laboratory, Israel	
LORAN	-	Long Range Navigation	
MC	-	Master Clock	
MJD	-	Modified Julian Date	
NIST	-	National Institute of Standards & Technology	
NPL	-	National Physical Laboratory, England	
NRC	-	National Research Council, Canada	
NOAA	-	National Oceanic and Atmospheric Administration	
OP	-	Paris Observatory, France	
PTB	-	Physical Technical Federal Laboratory, Germany	
SI	-	International System of Units	ns - nanosecond
SV	-	Space vehicle	us - microsecond
TA	-	Atomic Time	ms - millisecond
TAI	-	International Atomic Time	s - second
TAO	-	Tokyo Astronomical Observatory, Japan	min - minute
TUG	-	Technical University of Graz, Austria	h - hour
USNO	-	United States Naval Observatory	d - day
UTC	-	Coordinated Universal Time	
VLF	-	very low frequency	
VSL	-	Van Swinden Laboratory, Netherlands	

2. TIME SCALE INFORMATION

The values listed below are based on data from the BIH, the USNO, and the NIST. The UTC - UTC(NIST) values are extrapolations since UTC is computed more than two months after the fact. The UTC(USNO,MC) - UTC(NIST) values are averaged measurements from NAVSTAR satellites 3,4,6, and 8 (see references on page 6).

0000 HOURS COORDINATED UNIVERSAL TIME

AUGUST		UT1 - UTC(NIST)	UTC - UTC(NIST)	UTC(USNO,MC) - UTC(NIST)	
1990	MJD	(± 5 ms)	(± 0.2 us)	(± 0.04 us)	
	2	48105	-77 ms	0.3 us	0.60 us
	9	48112	-86 ms	0.3 us	0.52 us
	16	48119	-98 ms	0.2 us	0.44 us
	23	48126	-110 ms	0.2 us	0.35 us
	30	48133	-121 ms	0.2 us	0.29 us

INTERNATIONAL TIMING CENTER COMPARISONS VIA GPS COMMON-VIEW

The table below is a weighted average of the indicated GPS satellites used as transfer standards to measure the time difference of Timing Center (i) - UTC(NIST) by the simultaneous common-view approach (see references, page 6). The day-to-day variations of this technique are a few nanoseconds and the accuracy is about 10 ns. The time of the measurement is interpolated to 0000 UTC for the particular MJD ending in 9. These data are prepared for the BI PM for the computation on TAI and of UTC. All differential delays are 0 unless otherwise noted.

UTC(i)	UTC(i) - UTC(NIST) (ns)				MJD			
	PRN NUMBERS				48099	48109	48119	48129
APL	3,6,9,	12,13,14,	16,17,	20	-6	29	51	81
CRL	6,9	12			-207	-323	-503	-608
CSIRO	++				26076	26343	26590	26825
IEN		12,13		20	974	811	579	359
NPL	2,	12,13			1920	1608	1396	1167
NRC+++	2, 6,9,	12,13,14,	16,17,	20	-1228	-1385*	-1316	-1241
OP		12,13,14,	16,	20	577	517	462	395
PTB		12,	16,	20	-3540	-3625	-3701	-3778
TAO	6,9,	12,		17, 20	2322	2041	1742	1513
TUG	2,	12,13,14,	16,	20	-548	-866	-1203	-1523
USNO,MC	3,6,9,	12,13	16,17,	20	673	563	444	327
VSL	2,	12,13,14,	16,	20	-2939	-3196	-3390	-3553

+ These values have been updated from those printed in last month's Bulletin.

++ UTC(CSIRO) - UTC(NIST) is computed from the average via CRL, TAO, & WWVH.

+++ UTC(NRC) - UTC(NIST) has a differential delay of 41.2 ns; all other comparisons are computed using zero (0).

* NRC experienced a frequency shift of approximately +28.0 E-14 (NET) between MJD's 48110 - 48114. Therefore, the value for MJD 48109 was computed from weighted and unfiltered data.

3. UT1 CORRECTIONS AND LEAP SECOND ADJUSTMENTS

The master clock pulses used by the WWV, WWVH, WWVB, and GOES time code transmissions are referenced to the UTC(NIST) time scale. Occasionally, 1 second is added to the UTC time scale. This second is called a leap second. Its purpose is to keep the UTC time scale within ± 0.9 s of the UT1 astronomical time scale, which changes slightly due to variations in the rotation of the earth.

Positive leap seconds, beginning at 23 h 59 min 60 s UTC and ending at 0 h 0 min 0 s UTC, were inserted in the UTC timescale on 30 June 1972, 31 December 1972-1979, 30 June 1981-1983, 30 June 1985, 31 December 1987 and 1989. When future leap seconds are scheduled, advance notice will be provided in this bulletin.

The use of leap seconds ensures that UT1 - UTC will always be held within ± 0.9 s. The current value of UT1 - UTC is called the DUT1 correction. DUT1 corrections are broadcast by WWV, WWVH, WWVB, and GOES and are printed below. These corrections may be added to received UTC time signals in order to obtain UT1.

DUT1 = UT1 - UTC	= 0.0 s beginning 0000 UTC on 10 May 1990
DUT1 = UT1 - UTC	= -0.1 s beginning 0000 UTC on 26 July 1990
	= -0.2 s beginning 0000 UTC on 20 September 1990

SPECIAL ANNOUNCEMENT: The International Earth Rotation Service has announced that a leap second will be inserted into the UTC time scale at the end of December 31, 1990.

4. PHASE DEVIATIONS FOR WWVB AND LORAN-C

WWVB - The values shown for WWVB are the time difference between the time markers of the UTC(NIST) time scale and the first positive-going zero voltage crossover measured at the transmitting antenna. The uncertainty of the individual measurements is ± 0.5 μ s. The values listed are for 1300 UTC.

LORAN-C - The values shown for Loran-C represent the daily accumulated phase shift (in microseconds). The phase shift is measured by comparing the output of a Loran receiver to the UTC(NIST) time scale for a period of 24 hours. If data were not recorded on a particular day, the symbol (-) is printed.

The stations monitored are Dana, Indiana (8970 M) and Fallon, Nevada (9940 M). The monitoring is done from the NIST laboratories in Boulder, Colorado.

DATE	MJD	UTC(NIST) - WWVB(60 kHz) ANTENNA PHASE (μ s)	UTC(NIST) - LORAN PHASE (in μ s)	
			LORAN-C (DANA) (8970 M)	LORAN-C (FALLON) (9940 M)
08/01/90	48104	5.64	+0.00	-0.19
08/02/90	48105	5.75	-0.29	-0.36
08/03/90	48106	5.74	(-)	+0.23
08/04/90	48107	5.73	(-)	+0.08
08/05/90	48108	5.71	(-)	-0.09
08/06/90	48109	5.70	-0.09	-0.18
08/07/90	48110	5.70	-0.01	-0.08
08/08/90	48111	5.70	-0.04	-0.17
08/09/90	48112	5.73	+0.02	+0.04
08/10/90	48113	5.72	+0.12	+0.24
08/11/90	48114	5.73	+0.07	(-)
08/12/90	48115	5.74	-0.03	(-)
08/13/90	48116	5.74	-0.05	+0.02
08/14/90	48117	5.76	+0.02	+0.07
08/15/90	48118	5.75	-0.09	+0.07
08/16/90	48119	5.78	(-)	(-)
08/17/90	48120	5.68	+0.17	+0.27
08/18/90	48121	5.66	+0.11	-0.01
08/19/90	48122	5.64	-0.31	-0.21
08/20/90	48123	5.62	-0.22	-0.15
08/21/90	48124	5.67	+0.24	+0.08
08/22/90	48125	5.64	-0.20	+0.06
08/23/90	48126	5.73	+0.02	+0.22
08/24/90	48127	5.73	(-)	(-)
08/25/90	48128	5.74	+0.16	+0.05
08/26/90	48129	5.75	-0.21	-0.50
08/27/90	48130	5.77	-0.24	-0.14
08/28/90	48131	5.76	-0.03	-0.17
08/29/90	48132	5.75	-0.07	-0.08
08/30/90	48133	5.78	-0.07	-0.01
08/31/90	48134	5.66	(-)	(-)

5. GOES TIME CODE INFORMATION

A. TIME CODE PERFORMANCE (1 - 31 August 1990)

GOES/East: Performance within normal limits during this period except for the following: During eclipse periods from 0300-0500 UTC each day the GOES/East time code is transferred to GOES-7. The transmitted position data during these periods may produce deviations exceeding the normal 100 microsecond limit.

GOES/West: Performance within normal limits during this period.

B. SPECIAL REMINDER: Current satellite locations are about 60° West longitude for GOES/East and 135° West longitude for GOES/West.

C. FALL 1990 ECLIPSE OPERATIONS FOR GOES/EAST:

Due to the current use of an older satellite, GOES-2, as GOES/East, the GOES/East time code will have to be shifted from GOES-2 to GOES-7 temporarily for two hours each day during the Eclipse season. Beginning on 7 August, the GOES/East time code will be shifted to GOES-7 at 98° West longitude from 0300-0500 UTC each day until 29 September. Satellite position data transmitted from GOES/East during these periods will be approximately correct for GOES-7 and should not cause large errors in receivers that correct for satellite position. Some users may experience loss of signal during these periods, however, unless antennas are repointed appropriately.

D. FUTURE SATELLITE LAUNCHES: NOAA's present launch schedule for replacement of the current East and West satellites has been moved back to February 1992 at the earliest.

E. GOES STATUS REPORTS

A brief message from the NIST giving current GOES time code status information is available from the U.S. Naval Observatory's Automated Data Service computer system in Washington, DC. The message may be accessed 24 hours per day without charge by using a variety of terminals operating at 300, 1200, or 2400 Baud and even parity. Two different sets of telephone access numbers are available: (1) for 300 or 1200 Baud and the Bell 103 standard use (202) 653-1079 (commercial), 653-1079 (FTS), or 294-1079 (Autovon); (2) for 1200 or 2400 Baud with either the CCITT V.22 standard or the Bell standard use (202) 653-1783 (commercial), 653-1783 (FTS), or 294-1783 (Autovon). To receive the GOES status message, use the following procedure:

1. Access the USNO computer database by dialing one of the appropriate telephone numbers above;
2. In response to the prompt for identification, type your name and the name of your organization, followed by a carriage return;
3. Type "@NBSGO" followed by a carriage return to receive the status message at your terminal;
4. Disconnect by typing Control-D.

6. BROADCAST OUTAGES OVER 5 MINUTES AND WWVB PHASE PERTURBATIONS

OUTAGES					PHASE PERTURBATIONS WWVB 60 kHz				
STATION	JULY 1990	MJD	BEGAN (UTC)	ENDED (UTC)	FREQUENCY	JULY 1990	MJD	BEGAN (UTC)	ENDED (UTC)
WWVB	NONE					NONE			
WWV	NONE					NONE			
WWVH	NONE					NONE			

7. NOTES ON NIST TIME SCALES AND PRIMARY STANDARDS

The frequencies of the time scales, TA(NIST) and UTC(NIST), are calibrated with the NIST primary frequency standards. The UTC(NIST) scale is coordinated within a microsecond of the internationally coordinated time scale, UTC, generated at the BIH. It is used to control all of the NIST time and frequency services. The last calibration of the relative frequency offset, y , of UTC(NIST) as generated in Boulder, Colorado, gave:

$$1) \quad y_{UTC(NIST)} (\text{July 1987}) - y_{NBS-6} (\text{July 1987}) = (-0.6 \pm 2 (1 \text{ sigma})) \times 10^{-13}$$

for the date shown. This calibration includes a correction for the systematic offset due to room temperature blackbody radiation, which is approximately $(\Delta y_{BB}) = -1.7 \times 10^{-14}$. Using GPS¹, the frequency of TAI for the dates shown were measured to be:

$$2) \quad y_{TAI} (\text{July 1987}) - y_{NBS-6} (\text{July 1987 on geoid}) = (+1.7 \pm 2 (1 \text{ sigma})) \times 10^{-13}$$

where account has been taken of the gravitational "red shift."

Starting 1 January 1975, an accuracy algorithm was implemented to bring the second used in the generation of TA(NIST) closer to the NIST "best estimate" of the SI second (see references, p.6). The relative frequency associated with this "best estimate" is denoted $y_{Cs(NIST)}$. The last calibration (July 1987) covered the period from October 1986 through July 1987.

$$3) \quad y_{Cs(NIST)} - y_{NBS-6} = (+1.4 \pm 2) \times 10^{-13} (\text{July 1987})$$

and

$$4) \quad y_{TAI} - y_{Cs(NIST)} \text{ on geoid} = (+0.3 \pm 0.7) \times 10^{-13} (\text{July 1987})$$

This algorithm should provide nearly optimum accuracy and stability for TA(NIST) since it uses all past frequency calibrations with the NIST primary standards. These calibrations are weighted proportionately to the frequency memory of the clock ensemble that generates atomic time. This algorithm, therefore, capitalizes on a weighted combination of all the frequency calibrations with the primary standards in order to gain a "best estimate" of the SI Second while simultaneously obtaining the best uniformity available from the ensemble of working clocks in the atomic time scale system. The relative frequency of TA(NIST) is steered toward $y_{Cs(NIST)}$ by slight frequency drift corrections of the order of 1 part in 10^{13} /yr.

TA(NIST) and UTC(NIST) are no longer simply related by an equation. TA(NIST) is now computed each month using a Kalman algorithm which minimizes the mean square time dispersion. UTC(NIST) is now independently computed using a different algorithm and is steered in frequency to keep its time within a microsecond of UTC(BIH). Table 7.1 lists monthly values of the time difference between UTC(NIST) and TA(NIST). A linear interpolation between monthly values will typically be within 10 ns of the actual time difference, TA(NIST) - UTC(NIST).

The primary standards of NIST (NBS-4 and NBS-6) are used in either of two modes: as calibrators of frequency to provide a reference for the SI second; or as member clocks of the NIST clock ensemble, to help keep the proper time for TA(NIST) and the coordinated time for UTC(NIST). Operating in the clock mode, NBS-4 and NBS-6 are only used and weighted according to their stability performance. Accuracy enters only when they are used as frequency calibrators, in which case clock operation is necessarily interrupted.

¹GPS is the Global Positioning System, a network of navigation satellites.

Table 7.1 is a list of changes in the time scale frequencies of both TA(NIST) and UTC(NIST) as well as, a list of the time and frequency differences between TA(NIST) and UTC(NIST) at the dates of leap seconds, and/or frequency or frequency drift changes.

TABLE 7.1

FREQUENCY CHANGES					
DATE	(MJD)	TA(NIST)	UTC(NIST)	TA(NIST) - UTC(NIST)	y[UTC(NIST)] - y[TA(NIST)]
1 Mar 89	47586	0	-1.25 ns/d	24.045 121 622 s	-4.58 E-13
1 Apr 89	47617	0	-1.50 ns/d	24.045 122 871 s	-4.66 E-13
1 May 89	47647	0	-1.50 ns/d	24.045 124 078 s	-4.75 E-13
1 Jun 89	47678	0	-1.00 ns/d	24.045 125 375 s	-4.92 E-13
1 Jul 89	47708	0	-1.00 ns/d	24.045 126 670 s	-5.09 E-13
1 Aug 89	47739	0	-1.00 ns/d	24.045 128 060 s	-5.35 E-13
1 Sep 89	47770	0	-1.00 ns/d	24.045 129 538 s	-5.58 E-13
1 Oct 89	47800	0	1.00 ns/d	24.045 131 001 s	-5.68 E-13
1 Nov 89	47831	0	1.00 ns/d	24.045 132 534 s	-5.85 E-13
1 Dec 89	47861	0	1.00 ns/d	24.045 134 082 s	-6.05 E-13
1 Jan 90	47892	0	1.00 ns/d	25.045 135 724 s	-6.16 E-13
1 Feb 90	47923	0	1.00 ns/d	25.045 137 382 s*	-6.21 E-13
1 Mar 90	47951	0	1.00 ns/d	25.045 138 888 s*	-6.23 E-13
1 Apr 90	47982	0	1.00 ns/d	25.045 140 560 s*	-6.36 E-13
1 May 90	48012	0	1.00 ns/d	25.045 142 241 s*	-6.42 E-13
1 June 90	48043	0	1.00 ns/d	25.045 143 942 s*	-6.34 E-13
1 July 90	48073	0	0.50 ns/d	25.045 145 580 s*	-6.46 E-13
1 Aug 90	48104	0	0.00 ns/d	25.045 147 350 s	-6.62 E-13

UTC(NIST) is steered in time toward UTC by occasional frequency changes of the order of a few nanoseconds per day; 1 ns/d is approximately 1.16E-14. Otherwise, y[UTC(NIST)] is maintained as stable as possible.

*Due to the omission of the Dec. 31, 1989 Leap Second there has been an error of 1 second in the listed values of TA(NIST) - UTC(NIST) in Bulletins No. 387 through No. 392, page 6.

REFERENCES

- Allan, D.W., Hellwig, H. and Glaze, D.J., "An accuracy algorithm for an atomic time scale," *Metrologia*, Vol.11, No.3, pp.133-138 (September 1975).
- Glaze, D.J., Hellwig, H., Allan, D.W., and Jarvis, S., "NBS-4 and NBS-6: The NIST primary frequency standards," *Metrologia*, Vol.13, pp.17-28 (1977).
- Wineland, D.J., Allan, D.W., Glaze, D.J., Hellwig, H., and Jarvis, S., "Results on limitations in primary cesium standard operation," *IEEE Trans. on Instr. and Meas.*, Vol.IM-25, No.4, pp.453-458 (December 1976).
- Allan, D.W. and Weiss, M.A., "Accurate Time and Frequency Transfer During Common View of a GPS Satellite," *Proc. 34th Annual Symposium on Frequency Control*, p.334 (1980).
- Allan, D.W. and Barnes, J.A., "Optimal Time & Frequency using GPS signals," *Proc. 36th Annual Symposium on Frequency Control*, p.378 (1982).

8. SPECIAL ANNOUNCEMENTS

AUTOMATED COMPUTER TIME SERVICE (ACTS)

On March 9, 1988, NIST initiated operation of a telephone time service designed to provide computers with telephone access to NIST time at accuracies approaching 1 ms. Features of the service include automated compensation for telephone-line delay, advanced alert for changes to and from daylight savings time and advanced notice of insertion of leap seconds. The ASCII-character time code should operate with standard modems and most computer systems. While the system can be used to set computer time-of-day clocks, simple hardware can also be developed to set other clock systems.

The test phase for this service is now complete and NIST is committed to long-term operation of the service. Additional lines will be added as use expands. NIST requests that calling times be spread out so that the system is not heavily taxed in some narrow time frame (e.g., midnight). The service telephone number is (303) 494-4774. The number may be changed at a later date. A help message can be obtained by returning a ? during the first 6 s of transmission.

With appropriate user software, the NIST-ACTS service provides three modes for checking and/or setting computer time-of-day clocks.

1. In the simplest form of the (1200 Baud) service, the user receives the time code and an on-time marker/character which has been advanced a fixed period to nominally account for modem and telephone-line delays. Accuracy in this mode should be no worse than 0.1 s unless the connection is routed through a satellite.
2. At 1200 Baud, if the user's system echoes all characters to NIST, the round-trip line delay will be measured and the on-time marker advanced to compensate for that delay. The accuracy in this mode should be better than 10 ms. Our experience to date indicates that the asymmetry in conventional, 1200-Baud modems limits the accuracy at this level. Repeatability is about 1 ms.
3. At 300 Baud the user can obtain the same type of service as described in item 2 above, but there is generally less problem with modem asymmetry at this rate and our experience indicates that the accuracy is about 1 ms.

The accuracy statements here are based upon the assumption that the telephone connection is reciprocal, that is, that both directions of communication follow the same path with the same delay. Discussions with telephone carriers indicate that this is the general mode of operation and our tests to date indicate that the lines are both stable and reciprocal.

In order to assist users of the service, NIST has developed documentation of the features of the service, some example software which can be used in conjunction with certain popular personal computers and simple circuitry which can be used to extract an on-time pulse. This material is available on a 5¼-in, 360-kbyte DOS diskette with instructions for \$35.00 from the NIST Office of Standard Reference Materials, B311-Chemistry Bldg, NIST, Gaithersburg, MD, 20899, (301) 975-6776. Specify the Automated Computer Time Service, RM8101. Further technical questions and comments should be directed to NIST-ACTS, NIST Time and Frequency Division, 325 Broadway, Boulder, CO 80303.

FREQUENCY CONTROL SYMPOSIUM ANNOUNCEMENT

The 45th Annual Frequency Control Symposium will be held May 29-31, 1991 in Los Angeles, CA. This symposium is the leading technical conference addressing all aspects of frequency control and precision timekeeping. Authors are invited to submit papers dealing with recent progress in research, development and applications in areas represented by the following topics:

- Fundamental properties of piezoelectric crystals
- Theory and design of piezoelectric resonators
- Resonator processing techniques
- Surface acoustic wave devices (SAW)
- Quartz crystal oscillators
- Microwave and millimeter wave oscillators
- Signal processing and frequency control circuitry
- Atomic and molecular frequency standards
- Frequency and time coordination and distribution
- Sensors and transducers
- Applications of frequency control
- Measurement and specifications

Sponsors: IEEE-UFFC and the US Army Electronics Technology and Devices Laboratory (LABCOM).
Place: Los Angeles Airport Marriott, CA
Contact: Clark Wardrip, P. O. Box 6147, Vandenberg AFB, CA 93437; (805) 865-3214.

4. TITLE AND SUBTITLE

Time & Frequency Bulletin

5. AUTHOR(S)

6. PERFORMING ORGANIZATION (If joint or dual agency NBS, see instructions)

National Institute of Standards & Technology
325 Broadway
Boulder, CO 80303

7. Contract/Grant No.

8. Type of Report & Period Covered

9. SPONSORING ORGANIZATION NAME AND COMPLETE ADDRESS (Street, City, State, ZIP)

10. SUPPLEMENTARY NOTES

Document describes a computer program. SF-185, PIPS Software Summary, is attached.

11. ABSTRACT (A 200-word or less factual summary of most significant information. If document includes a significant bibliography or literature survey, mention it here)

The Time & Frequency Bulletin provides information on performance of time scales and a variety of broadcasts (and related information) to users of the NIST services.

12. KEY WORDS (5 to twelve entries, alphabetical order; capitalize only proper names; and separate key words by semicolons)

clocks; dissemination; frequency; GPS; oscillators; time

13. AVAILABILITY

- Unlimited
- For Official Distribution. Do Not Release to NTIS
- Order From Superintendent of Documents, U.S. Government Printing Office, Washington, D.C. 20402.
- Order From National Technical Information Service (NTIS), Springfield, VA. 22161

14. NO. OF PRINTED PAGES

8

15. Price

AN EQUAL OPPORTUNITY EMPLOYER

U.S. Department of Commerce
National Institute of Standards and Technology
Time & Frequency Division
Boulder, Colorado 80303

FIRST-CLASS MAIL
POSTAGE & FEES PAID
NIST
Permit No. G195

OFFICIAL BUSINESS
Penalty for Private Use, \$300

GWEN BENNETT
DIVISION 524 COPY 11
NATL BUREAU OF STANDARDS
325 BROADWAY
BOULDER, CO 80303