

# NATIONAL BUREAU OF STANDARDS REPORT

5317

REPORT  
ON  
THE LUMINOUS INTENSITY OF 8 LAMPS  
AND  
THE LUMINOUS FLUX OF 8 LAMPS  
USED IN  
INTERCOMPARISONS MADE  
AT  
BUREAU INTERNATIONAL DES POIDS ET MESURES

By  
Velma I. Burns  
Ray P. Teele



U. S. DEPARTMENT OF COMMERCE  
NATIONAL BUREAU OF STANDARDS

## THE NATIONAL BUREAU OF STANDARDS

### Functions and Activities

The functions of the National Bureau of Standards are set forth in the Act of Congress, March 3, 1901, as amended by Congress in Public Law 619, 1950. These include the development and maintenance of the national standards of measurement and the provision of means and methods for making measurements consistent with these standards; the determination of physical constants and properties of materials; the development of methods and instruments for testing materials, devices, and structures; advisory services to Government Agencies on scientific and technical problems; invention and development of devices to serve special needs of the Government; and the development of standard practices, codes, and specifications. The work includes basic and applied research, development, engineering, instrumentation, testing, evaluation, calibration services, and various consultation and information services. A major portion of the Bureau's work is performed for other Government Agencies, particularly the Department of Defense and the Atomic Energy Commission. The scope of activities is suggested by the listing of divisions and sections on the inside of the back cover.

### Reports and Publications

The results of the Bureau's work take the form of either actual equipment and devices or published papers and reports. Reports are issued to the sponsoring agency of a particular project or program. Published papers appear either in the Bureau's own series of publications or in the journals of professional and scientific societies. The Bureau itself publishes three monthly periodicals, available from the Government Printing Office: The Journal of Research, which presents complete papers reporting technical investigations; the Technical News Bulletin, which presents summary and preliminary reports on work in progress; and Basic Radio Propagation Predictions, which provides data for determining the best frequencies to use for radio communications throughout the world. There are also five series of nonperiodical publications: The Applied Mathematics Series, Circulars, Handbooks, Building Materials and Structures Reports, and Miscellaneous Publications.

Information on the Bureau's publications can be found in NBS Circular 160, Publications of the National Bureau of Standards (\$1.25) and its Supplement (\$0.75), available from the Superintendent of Documents, Government Printing Office, Washington 25, D. C.

Inquiries regarding the Bureau's reports should be addressed to the Office of Technical Information, National Bureau of Standards, Washington 25, D. C.

# NATIONAL BUREAU OF STANDARDS REPORT

NBS PROJECT

NBS REPORT

0201-20-0205

June 10, 1957

5317

REPORT  
ON  
THE LUMINOUS INTENSITY OF 8 LAMPS  
AND  
THE LUMINOUS FLUX OF 8 LAMPS  
USED IN  
INTERCOMPARISONS MADE  
AT  
BUREAU INTERNATIONAL DES POIDS ET MESURES

By  
Velma I. Burns  
Ray P. Teele  
Photometry and Colorimetry Section  
Optics and Metrology Division  
(File Ref. 1100)

## IMPORTANT NOTICE

NATIONAL BUREAU OF STANDARDS  
intended for use within the  
to additional evaluation and  
listing of this Report, either  
the Office of the Director, Na  
however, by the Government  
to reproduce additional copie

Approved for public release by the  
director of the National Institute of  
Standards and Technology (NIST)  
on October 9, 2015

Progress accounting documents  
publically published it is subjected  
reproduction, or open-literature  
ion is obtained in writing from  
Such permission is not needed,  
prepared if that agency wishes



U. S. DEPARTMENT OF COMMERCE  
NATIONAL BUREAU OF STANDARDS



THE LUMINOUS INTENSITY OF 8 LAMPS  
AND  
THE LUMINOUS FLUX OF 8 LAMPS  
USED IN  
INTERCOMPARISONS MADE  
AT  
BUREAU INTERNATIONAL DES POIDS ET MESURES

By  
Velma I. Burns  
Ray P. Teele

This report gives the results of measurements made on 16 lamps after their use in intercomparisons made at the Bureau International des Poids et Mesures. These lamps represent the photometric units as maintained at the National Bureau of Standards, these units being, (1) the candela at  $2042^{\circ}\text{K}$ , (2) the candela at  $2353^{\circ}\text{K}$ , (3) the lumen at  $2353^{\circ}\text{K}$ , and (4) the lumen at  $2788^{\circ}\text{K}$ . These "after" measurements are made to ascertain the stability of the lamps by the same methods and in terms of the same standards used to obtain the values given in National Bureau of Standards Report 4341. "Report on the Luminous Intensity of 8 Lamps and The Luminous Flux of 8 Lamps for Intercomparisons to be Carried Out at Bureau International des Poids et Mesures" In addition to measurement against the various groups of National Bureau of Standards standards the lamps were evaluated in terms of lamps retained at the NBS which had been measured originally with the lamps sent to the BIPM.

## 1. PURPOSE

This report gives the values found in 1957 which serve to determine the stability of the lamps measured in 1954 and 1955.

## 2. MATERIAL

The lamps are all of special construction in accordance with the desires of the BIPM, which supervised the manufacture of the lamps in France.



### 3. STANDARDS

The lamps were calibrated in terms of groups of standards representing the four photometric units maintained at the National Bureau of Standards. These groups of standards are as follows:

(a) Candela at 2042°K. Group consisting of lamps BS2395, BS2398, BS2399, BS2400, BS2401, BS2402, BS2407, and BS2270. The mean of the eight lamps is 16.817 candelas. This group of lamps was originally calibrated against the primary black body in 1937 and their stability was checked by reference to other standards prior to the measurements reported herein. The lamps were shown to have remained stable, the intercomparisons agreeing to better than 0.1 percent.

(b) Candela at 2353°K. Group consisting of lamps BS2987, BS2990, BS2991, BS2992, and BS2993. The mean of the five lamps is 33.608 candelas. This group of standards is the current reference group maintained at this Bureau.

(c) Flux at 2353°K. Group consisting of lamps BS5470, BS5472, BS5473, BS5477, BS5478, and BS5485. The mean of the six lamps is 444.0 lumens. This group of standards is the current reference group.

(d) Flux at 2788°K. Group consisting of lamps BS5872A, BS5873, BS5875, BS5876, BS5877, BS5878. The mean of the six lamps is 7564 lumens. This group of standards is the current reference group.

### 4. METHOD

(1) Luminous Intensity. The luminous intensity measurements were made by a substitution method on a horizontal bar photometer with the lamps operating in a base down position. The orientation was such that the plane of the filament was perpendicular to the photometer axis and the glass supporting structure was turned away from the photometer. The perpendicularity of the plane of the filament was determined by projecting a beam of light at right angles to the photometer axis and aligning the shadow of the filament on a target card.





The photometric distance (measured to the plane of the filament) for the lamps at 2042°K was 1.56 meters. The photometric distance for the lamps at 2353°K was 1.25 meters.

(2) Luminous Flux. The luminous flux measurements were made by a substitution method in an 88-inch integrating sphere with the lamps operating in a base up position.

(3) Photometric Measurements. All measurements were made photoelectrically. With the voltage held constant at the designated value, readings were taken of the current and luminous intensity or luminous flux. Measurements were made in the following order: three standards, the test lamps, and the remaining standards. Three or more sets of measurements were made, the order of taking the readings being reversed in alternate sets. The values given in 5. Results are the averages of all readings taken.

## 5. RESULTS

The results are given in Tables A, B, C, and D. The uncertainty of the luminous value (intensity or flux) for each individual lamp was estimated by calculation at the 0.1 percent confidence level from the variation in the results in the several sets of measurements. The uncertainty of the average was also computed in this way, and, in addition, was estimated as one-fourth of the square root of the sum of the squares of the uncertainties of the individual lamps. The latter uncertainty estimates are shown in parentheses.

Table A. Luminous Intensity at 2042°K (BIPM 1951 Scale) (2039°K NBS Scale). Measured at 1.56 meters.

Lamp No.	Volts	Amperes	Candela	Uncertainty
NBS 3757	97.200	.5543	11.67	±0.05
NBS 3759	97.900	.5599	11.99	±0.05
NBS 3760	98.600	.5581	11.90	±0.00
NBS 3762	98.200	.5587	11.96	±0.08
			11.88 <sub>0</sub>	±0.01 <sub>0</sub>
				(±0.027)
Lamps retained at NBS				
NBS 3758	96.000	.5551	11.59	± .04
NBS 3761	98.200	.5381	11.53	± .06



Table B. Luminous Intensity at 2353°K (BIPM Scale)  
(2352°K NBS Scale). Measured at 1.25 meters.

Lamp No.	Volts	Amperes	Candela	Uncertainty
NBS 3764	92.300	0.3403	22.30	±.33
NBS 3765	91.700	Loose base		
NBS 3769	91.800	.3407	22.04	±.04
NBS 3770	97.800	.3539	26.98*	±.14
Average			<u>23.77</u>	<u>±.17</u> (±.10)

Lamps retained at NBS

NBS 3767	90.900	0.3437	21.61	±.09
NBS 3768	91.300	.3483	23.00	±.12
NBS 3771	90.900	.3464	21.99	±.09
Average			<u>22.20</u>	<u>±.08</u> (±.05)

\*In terms of retained lamps before damage to NBS 3770.

Table C Luminous Flux at 2353°K (BIPM Scale)  
(2356°K NBS Scale)

Lamp No.	Volts	Amperes	Lumens	Uncertainty
NBS 3780	99.000	0.3208	232.7	±0.8
NBS 3782	98.500	.3208	231.7	±0.5
NBS 3783	98.200	.3198	225.8	±0.5
NBS 3784	99.100	.3217	232.8	±0.3
Average			<u>230.8</u>	<u>±0.5</u> (±0.3)

Lamps retained at NBS

NBS 3785	98.600	0.3201	230.2	± .6
NBS 3786	98.500	.3202	229.6	± .3
NBS 3787	98.200	.3192	226.6	± .3
Average			<u>228.8</u>	<u>± .0</u> (± .1)

Table D Luminous Flux at 2788°K (BIPM 1951 Scale)  
(2811°K NBS Scale)

Lamp No.	Volts	Amperes	Lumens	Uncertainty
NBS 3772	108.900	1.7373	2720	±23
NBS 3773	110.000	1.7574	2780	±19
NBS 3775	109.000	1.7338	2727	±29
NBS 3776	108.900	1.7432	2732	±35
Average			<u>2740</u>	<u>±16</u> (±.14)

Table Continued



Lamps retained at NBS

NBS 3777	110.100	1.7608	2782	±17
NBS 3778	108.400	1.7246	2729	±19
NBS 3779	110.300	1.7614	2794	±24
			<u>2768</u>	<u>±19</u>
				(±12)

Table E. Comparison of Luminous Intensity at 2042°K "Before" and "After" BIPM Measurements

Lamp No.	Amperes		Candela		Mean
	"Before"	"After"	"Before"	"After"	
NBS 3757	0.5541 <sub>9</sub>	0.5543	11.85±0.10	11.67±0.05	11.76
NBS 3759	.5593 <sub>9</sub>	.5599	11.84±0.10	11.99±0.05	11.92
NBS 3760	.5577 <sub>0</sub>	.5581	11.88±0.17	11.90±0.00	11.89
NBS 3762	.5590 <sub>3</sub>	.5587	12.14±0.13	11.96±0.08	12.05
Average			<u>11.928±0.021</u> (±0.06 <sub>4</sub> )	<u>11.880±0.010</u> (±0.02 <sub>7</sub> )	11.90 <sub>5</sub>

Lamps retained at NBS

NBS 3758	0.5551 <sub>6</sub>	0.5551	11.63±.31	11.59±.04
NBS 3761	.5381 <sub>3</sub>	.5581	<u>11.56±.23</u>	<u>11.53±.06</u>
Average			<u>11.60±.04</u> (±.19)	<u>11.56±.05</u> (±.04)

If the lamps retained at the NBS were used as a base the "After" values would be increased by .04 candela bringing the "After" values for the average of the 4 lamps into close agreement with the "Before" average. It is concluded that there have been real changes in three of the four lamps. We would like to have the mean of the "Before" and "After" values used to represent our unit of luminous intensity at 2042°K.

Table F. Comparison of Luminous Intensity at 2353°K "Before" and "After" BIPM Measurements.

Lamp No.	Amperes		Candela		Mean
	"Before"	"After"	"Before"	"After"	
NBS 3764	0.3402 <sub>0</sub>	0.3403	22.38±0.12	22.30±0.33	22.34
NBS 3765	loose base		loose base		
NBS 3769	.3402 <sub>3</sub>	.3407	22.03±0.15	22.04±.04	22.04
NBS 3770	.3543 <sub>1</sub>	.3539	<u>26.93±0.11</u>	<u>26.98±.14</u>	23.96
Average			<u>23.78±0.08</u> (±0.07)	<u>23.77±.17</u> (±.10)	23.78

Table Continued



Lamps retained at NBS

NBS 3767	0.3434 <sub>3</sub>	0.3437	21.74±0.22	21.61±.09
NBS 3768	.3482 <sub>8</sub>	.3483	22.82±0.33	23.00±.12
NBS 3771	.3462 <sub>3</sub>	.3464	21.92±0.21	21.99±.09
Average			<u>22.16±0.25</u> (±0.06)	<u>22.20±.08</u> (±.05)

If the lamps retained at the NBS were used as a base the "After" values would be decreased by 0.04 candela. The estimated uncertainty of the individual lamps or of the mean does not justify any changes

Table G. Comparison of Luminous Flux at 2353°K "Before" and "After" BIPM Measurements

Lamp No.	Amperes		Lumens		Mean
	"Before"	"After"	"Before"	"After"	
NBS 3780	0.3205 <sub>3</sub>	0.3208	232.7±0.2	232.7±0.8	232.7
NBS 3782	.3205 <sub>6</sub>	.3208	231.6±0.6	231.7±0.5	231.6
NBS 3783	.3196 <sub>3</sub>	.3198	226.3±0.7	225.8±0.5	226.0
NBS 3784	.3215 <sub>1</sub>	.3217	233.7±0.2	232.8±0.3	233.2
Average			<u>231.1±0.3</u> (±0.2)	<u>230.8±0.5</u> (±0.3)	230.9

Lamps retained at NBS

NBS 3785	0.3200 <sub>0</sub>	0.3201	230.1±0.9	230.2±0.6
NBS 3786	.3201 <sub>2</sub>	.3202	229.8±1.1	229.6±0.3
NBS 3787	.3191 <sub>5</sub>	.3192	226.6±1.5	226.6±0.3
			<u>228.8±1.1</u> (±0.7)	<u>228.8±0.0</u> (±0.2)

No significant changes have taken place in these lamps.

Table H. Comparison of Luminous Flux at 2788°K "Before" and "After" BIPM Measurements

Lamp No.	Amperes		Lumens		Mean
	"Before"	"After"	"Before"	"After"	
NBS 3772	1.7369	1.7373	2719±12	2720±23	2720
NBS 3773	1.7580	1.7574	2778±15	2780±19	2779
NBS 3775	1.7336	1.7338	2730±27	2727±29	2728
NBS 3776	1.7437	1.7432	2733±11	2732±35	2732
Average			<u>2740±16</u> (±9)	<u>2740±16</u> (±14)	2740

Lamps retained at NBS

NBS 3777	1.7611	1.7608	2780±25	2782±17
NBS 3778	1.7237	1.7246	2721±25	2729±19
NBS 3779	1.7594	1.7614	2778±18	2794±24
			<u>2760±23</u> (±12)	<u>2768±19</u> (±8)





No significant changes have taken place in these lamps. If the lamps retained at the NBS were used as a base the "After" values would be decreased by 8 lumens. The larger uncertainties estimated for the "After" values do not justify such a change.

Conclusions. Although there is some indication of change in a few of the lamps, the estimated uncertainties do not justify discarding any lamps. For the sake of uniformity of treatment it is suggested that the mean of the "Before" and "After" values be used for all lamps to represent their value at the time of the comparisons carried out at the BIPM.



U. S. DEPARTMENT OF COMMERCE

Sinclair Weeks, *Secretary*

NATIONAL BUREAU OF STANDARDS

A. V. Astin, *Director*



## THE NATIONAL BUREAU OF STANDARDS

The scope of activities of the National Bureau of Standards at its headquarters in Washington, D. C., and its major field laboratories in Boulder, Colorado, is suggested in the following listing of the divisions and sections engaged in technical work. In general, each section carries out specialized research, development, and engineering in the field indicated by its title. A brief description of the activities, and of the resultant reports and publications, appears on the inside front cover of this report.

### WASHINGTON, D. C.

**Electricity and Electronics.** Resistance and Reactance. Electron Tubes. Electrical Instruments. Magnetic Measurements. Dielectrics. Engineering Electronics. Electronic Instrumentation. Electrochemistry.

**Optics and Metrology.** Photometry and Colorimetry. Optical Instruments. Photographic Technology. Length. Engineering Metrology.

**Heat and Power.** Temperature Physics. Thermodynamics. Cryogenic Physics. Rheology and Lubrication. Engine Fuels.

**Atomic and Radiation Physics.** Spectroscopy. Radiometry. Mass Spectrometry. Solid State Physics. Electron Physics. Atomic Physics. Nuclear Physics. Radioactivity. X-rays. Betatron. Nucleonic Instrumentation. Radiological Equipment. AEC Radiation Instruments.

**Chemistry.** Organic Coatings. Surface Chemistry. Organic Chemistry. Analytical Chemistry. Inorganic Chemistry. Electrodeposition. Gas Chemistry. Physical Chemistry. Thermochemistry. Spectrochemistry. Pure Substances.

**Mechanics.** Sound. Mechanical Instruments. Fluid Mechanics. Engineering Mechanics. Mass and Scale. Capacity, Density, and Fluid Meters. Combustion Controls.

**Organic and Fibrous Materials.** Rubber. Textiles. Paper. Leather. Testing and Specifications. Polymer Structure. Organic Plastics. Dental Research.

**Metallurgy.** Thermal Metallurgy. Chemical Metallurgy. Mechanical Metallurgy. Corrosion. Metal Physics.

**Mineral Products.** Engineering Ceramics. Glass. Refractories. Enameled Metals. Concreting Materials. Constitution and Microstructure.

**Building Technology.** Structural Engineering. Fire Protection. Heating and Air Conditioning. Floor, Roof, and Wall Coverings. Codes and Specifications.

**Applied Mathematics.** Numerical Analysis. Computation. Statistical Engineering. Mathematical Physics.

**Data Processing Systems.** SEAC Engineering Group. Components and Techniques. Digital Circuitry. Digital Systems. Analogue Systems. Application Engineering.

• Office of Basic Instrumentation

• Office of Weights and Measures

### BOULDER, COLORADO

**Cryogenic Engineering.** Cryogenic Equipment. Cryogenic Processes. Properties of Materials. Gas Liquefaction.

**Radio Propagation Physics.** Upper Atmosphere Research. Ionospheric Research. Regular Propagation Services. Sun-Earth Relationships.

**Radio Propagation Engineering.** Data Reduction Instrumentation. Modulation Systems. Navigation Systems. Radio Noise. Tropospheric Measurements. Tropospheric Analysis. Radio Systems Application Engineering.

**Radio Standards.** Radio Frequencies. Microwave Frequencies. High Frequency Electrical Standards. Radio Broadcast Service. High Frequency Impedance Standards. Calibration Center. Microwave Physics. Microwave Circuit Standards.

