Reference

NBS PUBLICATIONS

N327



NBS TECHNICAL NOTE 1089

U.S. DEPARTMENT OF COMMERCE / National Bureau of Standards

Site Attenuation

R. G. FitzGerrell

-QC 100 .U5753 No. 1089 1985 he National Bureau of Standards¹ was established by an act of Congress on March 3, 1901. The Bureau's overall goal is to strengthen and advance the nation's science and technology and facilitate their effective application for public benefit. To this end, the Bureau conducts research and provides: (1) a basis for the nation's physical measurement system, (2) scientific and technological services for industry and government, (3) a technical basis for equity in trade, and (4) technical services to promote public safety. The Bureau's technical work is performed by the National Measurement Laboratory, the National Engineering Laboratory, the Institute for Computer Sciences and Technology, and the Institute for Materials Science and Engineering.

The National Measurement Laboratory

Provides the national system of physical and chemical measurement; coordinates the system with measurement systems of other nations and furnishes essential services leading to accurate and uniform physical and chemical measurement throughout the Nation's scientific community, industry, and commerce; provides advisory and research services to other Government agencies; conducts physical and chemical research; develops, produces, and distributes Standard Reference Materials; and provides calibration services. The Laboratory consists of the following centers:

The National Engineering Laboratory

Provides technology and technical services to the public and private sectors to address national needs and to solve national problems; conducts research in engineering and applied science in support of these efforts; builds and maintains competence in the necessary disciplines required to carry out this research and technical service; develops engineering data and measurement capabilities; provides engineering measurement traceability services; develops test methods and proposes engineering standards and code changes; develops and proposes new engineering practices; and develops and improves mechanisms to transfer results of its research to the ultimate user. The Laboratory consists of the following centers:

The Institute for Computer Sciences and Technology

Conducts research and provides scientific and technical services to aid Federal agencies in the selection, acquisition, application, and use of computer technology to improve effectiveness and economy in Government operations in accordance with Public Law 89-306 (40 U.S.C. 759), relevant Executive Orders, and other directives; carries out this mission by managing the Federal Information Processing Standards Program, developing Federal ADP standards guidelines, and managing Federal participation in ADP voluntary standardization activities; provides scientific and technological advisory services and assistance to Federal agencies; and provides the technical foundation for computer-related policies of the Federal Government. The Institute consists of the following centers:

The Institute for Materials Science and Engineering

Conducts research and provides measurements, data, standards, reference materials, quantitative understanding and other technical information fundamental to the processing, structure, properties and performance of materials; addresses the scientific basis for new advanced materials technologies; plans research around cross-country scientific themes such as nondestructive evaluation and phase diagram development; oversees Bureau-wide technical programs in nuclear reactor radiation research and nondestructive evaluation; and broadly disseminates generic technical information resulting from its programs. The Institute consists of the following Divisions:

- Basic Standards²
- Radiation Research
- Chemical Physics
- Analytical Chemistry

- Applied Mathematics
- Electronics and Electrical Engineering²
- Manufacturing Engineering
- Building Technology
- Fire Research
- Chemical Engineering²
- Programming Science and Technology
- Computer Systems Engineering

- Inorganic Materials
- Fracture and Deformation³
- Polymers
- Metallurgy
- Reactor Radiation

¹Headquarters and Laboratories at Gaithersburg, MD, unless otherwise noted; mailing address Gaithersburg, MD 20899. ²Some divisions within the center are located at Boulder, CO 80303.

³Located at Boulder, CO, with some elements at Gaithersburg, MD



NES RESEARCH INFORMATION CENTER NBS-Keg. QC100 USTS3 NO. 1089 1985

Site Attenuation

R. G. FitzGerrell

Electromagnetic Fields Division Center for Electronics and Electrical Engineering National Engineering Laboratory National Bureau of Standards Boulder, Colorado 80303



U.S. DEPARTMENT OF COMMERCE, Malcolm Baldrige, Secretary

NATIONAL BUREAU OF STANDARDS, Ernest Ambler, Director

Issued November 1985

National Bureau of Standards Technical Note 1089 Natl. Bur. Stand. (U.S.), Tech Note 1089, 48 pages (Nov. 1985) CODEN: NBTNAE

> U.S. GOVERNMENT PRINTING OFFICE WASHINGTON: 1985

For sale by the Superintendent of Documents, U.S. Government Printing Office, Washington, DC 20402

Contents

									Page
1.	Introduct	on							1
2.	Calculati	ns							2
3.	Measureme	ts						• • • • • • • •	4
4.	Error Est	mate		• • • • • • •					6
5.	Conclusio	s		•••••	••••••				7
6.	Reference	• • • • • • • •		• • • • • • •	• • • • • • •				7
Арр	endix I	The Dipo	le Antenn	inas					13
		1. Phys 2. Misma 3. Inpu 4. Dipo 5. Mutu	ical Char atch Loss t Impedan le Droop. al Impeda	acterist	tics Dipoles	in Ech	elon	• • • • • • • • •	13 14 16 17 19
Ref	erences fo	Appendi	< I				• • • • • • •		21
Арр	endix II	Computer	Programs	; HVD3 aı	nd ZMM1.		• • • • • • • •		29

SITE ATTENUATION

R. G. FitzGerrell Electromagnetic Fields Division National Bureau of Standards Boulder, Colorado 80303

Site attenuation is a measure of performance of an open test site used at frequencies below about 1 GHz for antenna calibration and equipment emission and susceptibility testing. These sites typically consist of a large, obstruction-free ground plane and the hemisphere above it. Site attenuation of an ideal site is calculated and compared to data measured using the 30 m by 60 m NBS ground screen.

Key words: antenna measurements; dipole antennas; site attenuation.

1. Introduction

Site attenuation is a measure of performance of an open test site used to determine the levels of emissions from electromagnetic sources or to calibrate monopole and dipole antennas used for field strength measurements. An ideal site consists of obstruction-free plane ground and the hemisphere above it -- both infinite in extent. Ground constants may range from typical earth values to values for good conductors. Site attenuation measurements are used to compare performance of real sites to the ideal or reference site. Calculations presented here provide a reference for the measurement results obtained from a 30 m by 60 m wire mesh ground screen.

Site attenuation is defined as the minimum relative insertion loss measured between the terminals of two polarization-matched antennas located on the test site when one antenna is moved vertically over a specified height range. The following parameters must also be specified: separation distance between the two antennas, measured horizontally; antenna type and polarization; and impedance of the receiving and transmitting systems attached to the antenna terminals. Calculated and measured site attenuation includes mismatch losses but not balun or cable losses.

2. Calculations

Insertion loss between two thin dipoles in echelon located in free space may be calculated using the mutual impedance between the dipoles [1] and self impedance of the dipoles [2]. The mutual impedance is given as follows:

$$Z_{21} = -\frac{V_{21}}{I_{1b}} .$$
 (1)

Assuming the dipoles are identical, the self impedance, Z_i , of each dipole is calculated as described in Appendix I. An arbitrary power level, P_{inc} , is assumed to be delivered to the transmitting dipole and, with Z_i and the transmitter impedance, is used to calculate the base current, I_{1b} . The mutual impedance Z_{21} , is calculated based upon the antenna geometry, and V_{21} , the receiving dipole open circuit voltage, is obtained using (1). The power at the receiving dipole terminals, P_{rec} , is calculated using V_{21} , Z_i and the receiver impedance assuming the receiving dipole current has no effect on the transmitting dipole impedance. That is, any additional mutual impedance effects are ignored. The ratio, P_{inc}/P_{rec} , is the insertion loss between the dipoles.

Insertion loss between two thin dipoles in echelon over perfectly conducting ground may be calculated if they are vertically or horizontally polarized (a restriction occurring only because of the mutual impedance formulations used in this document) and their positions are specified. The test range geometry and multiple mutual impedances required for the calculation are shown in figure 1 for horizontal polarization. The value of I_{1b} is calculated assuming that 1 W is delivered to a transmitting dipole that has an input impedance, Z_{in} , equal to Z_i properly combined with the mutual impedance between it and its image in the perfectly conducting ground. A total mutual impedance consisting of Z_{21} and Z_{23} is calculated and used with I_{1b} to obtain V_{oc} , the open circuit voltage in the receiving dipole. Input impedance of the receiving dipole is calculated and used with V_{oc} to calculate P_{rec} . Again, the impedance of the transmitting dipole is assumed to be unaffected by the current

in the receiving dipole's finite load impedance. This assumption results in an approximation to the true value of P_{rec} and reduces the number of mutual impedance equations required to the four shown in figure 1.

The input impedance for dipole antennas over perfect ground may be written using the notation in figure 1 as follows:

$$Z_{in} = Z_i \pm Z_{13}$$
 for the transmitting dipole (2)

and

$$Z_{in}$$
 (h) = $Z_i \pm Z_{24}$ for the receiving dipole . (3)

The sign of the mutual impedance terms depend upon polarization, and the fact that the receiving dipole impedance is a function of height, h, is emphasized in the notation. (The transmitting dipole is always at a fixed height.) The mutual impedances are calculated using equations in [1] and Z_i is calculated using equation (4A) given in Appendix I.

The power from the transmitting dipole is given by

$$P_{trans} = P_{inc} (1 - | (Z_{in} - Z_s)/(Z_{in} + Z_s) |),$$
 (4)

where P_{inc} is the power delivered to the dipole terminals and Z_s is the transmitter impedance presented to the dipole terminals. The current required in (1) can now be written as

$$I_{1b} = (P_{trans}/Rea1 (Z_{in}))^{1/2}$$
 (5)

The total mutual impedance between the dipole antennas is

$$Z_{m}(h) = Z_{21}(h) \pm Z_{23}(h)$$
, (6)

with the height dependance again emphasized. The receiving dipole open-circuit voltage is obtained as a result of the definition of mutual impedance as

$$v_{oc}(h) = I_{1b} | Z_m(h) | ,$$
 (7)

and the received power is approximately given by

$$P_{rec}(h) = Z_R (V_{oc}(h) / | Z_R + Z_{in}(h) |)^2,$$
 (8)

where Z_R is the receiver impedance presented to the dipole terminals.

Finally, site attenuation is obtained as

$$S_{NBS} = 10 \log_{10} (P_{inc}/P_{rec}(H)),$$
 (9)

where P_{rec} (H) is the maximum value of $P_{rec}(h)$ which occurs at a height H in the range of heights considered. For the calculations presented in this paper, $Z_R = Z_S = 100$ ohms and P_{inc} , an arbitrary value, is set to 1 W.

Insertion loss is the ratio of the power delivered to the transmitting dipole terminals to the power received at the receiving dipole terminals and is a positive quantity. Site attenuation is the minimum insertion loss occurring when the receiving dipole is scanned in height. The height-scan patterns in figure 2 show calculated insertion loss for horizontally polarized (HP) and vertically polarized (VP) dipoles with a 3 m separation distance. The transmitting dipole is 2 m above perfectly conducting ground. Site attenuation is the single minimum value of each pattern. Figure 3 shows site attenuation versus frequency for separation distances (S) of 3 m, 10 m, and 30 m. Heights scanned are 1 m to 4 m at the 3 m and 10 m separation distances and 2 m to 6 m at the 30 m separation distance except for vertical polarization when the half-length of the receiving dipole is always positioned 5 cm, or more, above the ground.

Calculated site attenuation provides an ideal reference. Deviations of measured data from the reference indicate site imperfections or dipole failure.

3. Measurements

Figure 4 shows the measurement procedure used for determining site attenuation. The reference insertion loss, figure 4(a), is a received signal level, expressed in dBm, dependent upon the output level of the signal

generator and balun, cable, and attenuator losses. The output level remains fixed during the measurement procedure and is somewhat arbitrary. It is less than the maximum level acceptable by the receiver but great enough so the signal level in figure 4(b) is well above the ambient noise level.

Implicit in the measurement technique is the assumption that the signal generator output level is constant for both the reference insertion loss site attenuation, measurement at the measurement and the measurement frequency. That is, it is assumed that the incident power at the 3 dB attenuator (pad) port is constant for the two measurement conditions. The signal generator used for these measurements contains an automatic level control, ALC, and it is assumed that this circuit works perfectly for the small load VSWR variation (estimated to be 5 percent or less) occurring between the two measurement conditions [3]. The incident power may be monitored using directional couplers and a power meter but these devices are not considered necessary when using high quality signal generators with ALC.

The dipole antennas have hybrid junctions for antenna baluns. Equal length coaxial cables from the dipole terminals to the hybrid junction form a 100 ohm, balanced, shielded transmission line. As a result, the receiving and transmitting system impedances are 100 ohms. This value is used to calculate the reference site attenuation data shown in figure 3 as the solid lines for very thin dipoles and (+) points for dipoles with dimensions of those actually used for the measurements. To insure that this impedance is 100 ohms, miniature 3 dB attenuators are permanently installed in the four separate cables at the point of attachment to the dipole terminals.

The receiving dipole is moved up and down over the specified height range by a person below the ground screen who observes the receiver display. The maximum received signal level is recorded by a peak sample-and-hold circuit in the receiver. The magnitude of the difference between the reference insertion loss value and this maximum measured value gives site attenuation directly. No additional measurements or corrections are required. The measured data are shown as the circles in figure 3.

4. Error Estimate

Test range antenna heights and separation distances are set within ± 1 cm assuming the ground screen surface is perfectly flat. Dipole height is measured at the dipole feedpoint. Calculated errors in site attenuation are at most ± 0.09 dB as a result of possible positioning errors. The extensible dipoles used at 97 MHz and lower frequencies droop about 16 cm at the tips at 30 MHz. No effort is made to keep these dipoles straight and the resulting error, discussed in Appendix I, is negligible.

The stability of the combination of the receiver, signal generator, hybrid junctions, and cables is determined by the repeatability of the reference insertion loss measurements, figure 4(a), performed before and after each subset of measurements. (The three subsets of measurement frequencies are determined by the frequency ranges of the three sets of hybrid junctions). Over these typically two hour time periods, the difference between the initial and final reference insertion loss data is at most ± 0.11 dB. This variability appears to be predominantly caused by cable handling, connector mating, and moving the signal generator. Since this is a relative measurement, not an absolute one, the primary accuracy limitation is the manufacturer's specified "cumulative fidelity" for the receiver (spectrum analyzer) of "< ±1.0 dB over 0 to 80 dB display, 20-30° C." Errors in calculated mismatch loss at the 3 dB attenuators is discussed in Appendix I. The inability of the 3 dB attenuators to force the signal source and receiving system reflection coefficients to O results in estimated errors of 0.27 dB, 0.17 dB due to source system VSWR and 0.1 dB due to receiving system VSWR. Therefore, a simple worst case error estimate gives $(\pm 0.09 \pm 0.11 \pm 1.0 \pm 0.27) = \pm 1.47$ dB. The largest single error, that of the receiver cumulative fidelity, can be reduced by carefully calibrating the receiver.

The average difference between the measured and calculated (calculated - measured, dB) data for horizontal polarization is -0.2 dB with a standard deviation of 0.4 dB calculated using the dB values. For vertical polarization, these statistics are 0.03 dB and 0.86 dB respectively. The greatest difference between measured and calculated site attenuation is 1.02 dB for horizontal polarization, where the effect of feed cable reflections is negligible, and 1.94 dB for vertical polarization.

The data for the single measurement set upon which this paper is based are shown in Table 1. This was the first set of data measured after the preliminary set used to determine the suitability of the various test components. Failure of the air-supported fabric cover over the NBS ground screen facility has temporarily halted further measurements.

5. Conclusions

The agreement between measured and calculated site attenuation data is good, even for vertical polarization. The good agreement for vertical polarization is believed to result from the use of commercial hybrid junctions as dipole antenna baluns. These devices split the E port power $180^{\circ} \pm 1^{\circ}$ at the 1 and 2 ports. Little or no unbalanced (common mode) current flows on the feed cables and produces unwanted radiation.

Implementation of the deceptively simple mutual impedance equations is actually a fairly straightforward computer programming exercise. To relieve interested readers of this exercise, Appendix II contains Program ZMM1, the listing of the FORTRAN 4 code used to calculate site attenuation. No claim is made for the efficiency or structure of the code or its accuracy beyond the comparison of measured and calculated results presented here.

The actual measurement technique is certainly not new. It is basically a "two identical antenna" gain measurement that stops short of the last step -subtracting the path loss and mismatch loss from the measured data to obtain the product of the gain of two identical antennas over ground.

6. References

- [1] King, H.E., "Mutual impedance of unequal length antennas in echelon," IEEE T-AP-5, pp. 306-313, July 1957.
- [2] Schelkunoff, S.A., and Friis, H.T., <u>Antennas, Theory and Practice</u>, New York: Wiley, 1952, Ch. 13, pp. 431-434.
- [3] Engen, G.F., "Amplitude stabilization of a microwave signal source," IEEE T-MTT-6, pp. 202-206, April 1958.

Frequency, MHz	3 m Dis	tance	10 m D	istance	30 m D	ristance
	meas.	calc.	meas.	calc.	meas.	calc.
30*0	10.3	10.182	22.1	21.238	34.3	33.461
44.0	10.8	10.357	20.9	20.912	35 .7	35.850
65.0	11.6	11.389	22.6	22.332	35.6	35.350
97 ° 0	13.1	13.124	22.1	22.446	34 .6	34.689
143.0	17.0	17.142	25.9	26.261	36.2	36.630
210.0	20.9	20.954	29.3	29.527	39 • 0	39.057
311.0	23.9	23.903	32.7	32.455	42.3	41.977
459.0	27.7	27.401	36.4	36.370	45.5	45.435
677.0	31.1	30.720	40.8	39.777	49.4	49 °078
1000.0	35.0	34.045	43.5	43.112	53.4	52.439
Frequency, MHz	3 m Dis	V e tance	ertical Polariz 10 m D	ation istance	30 m D	istance
	meas.	calc.	meas.	calc.	meas.	calc.
44 .0	12.3	12.312	17.9	18.507	25.2	26.014
65 ° 0	14.7	15.286	20.1	20.782	28.0	28.508
97 ° 0	19.1	19.438	23.1	24.295	31.6	32.530
143.0	20.9	20.268	29.0	27.909	36 . 7	36.152
210.0	22.8	23.412	30.7	31.529	38.5	40.437
311.0	26.4	26.798	34 .9	33 . 998	46.1	44.459
459.0	30.2	30.212	37.5	36.998	46.3	46.068
677.0	33.4	33.531	41.0	40.202	50.6	49 °230
1000.0	37.7	36.961	44.2	43.779	51.7	52.576

Horizontal Polarization

Table 1. Measured and Calculated Site Attenuation Data, dB







Figure 2. Selected relative insertion loss data calculated for horizontally polarized, HP, and vertically polarized, VP, dipoles over perfect ground. Site attenuation is the single minimum relative insertion loss value of each curve. Separation distance is 3 m; transmitting dipole height is 2 m.



Figure 3. Site attenuation for horizontally polarized, HP, and vertically polarized, VP, dipoles over perfect ground. Solid curves are calculated for half-wave dipoles with radius = 1.E-30 m; + + points are calculated for dimensions of actual dipoles. Measured data are shown as o o.



(a) Measure reference insertion loss



(b) Measure reference insertion loss plus site attenuation



1. Physical Characteristics

Figures 1A through 4A are photographs showing the dipole antenna fabrication details. The 2.54 cm diameter white teflon feedpoint assembly is machined from a 8.89 cm length of rod stock. A cylindrical hole 2.22 cm long, 1.27 cm wide, and 1.31 cm deep provides room for the SMA panel jacks which form the dipole feedpoint. A flat surface opposite this hole supports the panel jack nuts. These dimensions are chosen to allow a nut driver to fit on the panel jacks for attaching SMA connectors, and give a sufficient thickness of teflon to support the panel jacks yet allow the SMA connectors to seat full depth on the jack body. Another flat surface, 90 deg around the feedpoint assembly with a centered hole, provides the mounting surface for a 2.54 cm diameter wood dowel used as the antenna mast.

The ends of the feedpoint assembly are drilled to a 1 cm diameter to form holes 2.64 cm deep. These holes provide a press fit for the extensible dipole elements used from 30 MHz to 97 MHz. Small holes are drilled in the center of the bottom of these 1 cm holes into the feedpoint cylinder and tapped with 6-32 threads. Stock banana plugs (13/16 in x 6-32, threaded shank) are screwed into place and the threaded shank is soldered to the panel jack center pin. A solder bridge is formed to join the outer conductor hexagonal surfaces of the panel jacks as shown in figure 3A; figure 4A shows the banana plug in the 1 cm diameter hole. The ends of the extensible dipole elements are drilled with a hole for a snug fit over the banana plug. These extensible elements are four section stainless steel whips with a 68.58 cm minimum length and a 243.84 cm maximum length.

None of the dimensions are critical to the electrical performance of the dipole. Basically, the feedpoint assembly is made from a low loss dielectric material that is strong enough to support the 30 MHz half-wave elements and provide a feedpoint gap that is much smaller than a wavelength at the highest frequency of interest. Dipole elements for the frequencies above 97 MHz are fixed lengths of thin wall brass tubing having a 0.48 cm outer diameter. These elements are supported by the banana plug and a 1 cm diameter dielectric washer positioned at the end of the feedpoint assembly.

Antenna half-length, measured from the center of the feedpoint assembly, is approximately one-half wavelength. It is chosen so that the calculated freespace input impedance is purely real. These lengths are determined using Program HVD3 which is included in Appendix II.

The 50 ohm coaxial hybrid junctions, serving as antenna baluns, are fourport devices having the following minimum performance specifications:

- 1. Isolation: 30 dB
- 2. Phase balance, E (diff) port feed: 180 deg ± 1 deg
- 3. Amplitude balance: 0.2 dB
- 4. VSWR: 1.3
- 5. Insertion loss, E(diff) port feed: 0.75 dB.

Equal length coaxial cables, 2 m long, separate the hybrid junctions from the feedpoint assembly. These cables are cut physically the same length from a single spool of cable and the connectors are carefully attached according to the manufacturer's instructions. Insertion loss versus frequency of two cable-hybrid junction assemblies is measured between the E ports. The H ports are terminated with 50 ohm loads. The cable pair attached to the colinear ports of the hybrid junctions form a shielded, balanced, 100 ohm transmission line. Satisfactory performance is observed when the measured insertion loss is a smooth function of frequency and approximately equal to the cable loss plus the hybrid junction insertion loss.

2. Mismatch Loss

As mentioned in the main body of the text, high performance, miniature 3 dB attenuators having maximum VSWR specifications of 1.07 + 0.015f(GHz), DC to 8 GHz are permanently attached to the equal length cables at the dipole feedpoint. The attenuators reduce the errors incurred in assuming the source and receiving system impedances are exactly 100 ohms in the site attenuation calculation. These impedances are not exactly 100 ohms because the source signal generator is specified to have a VSWR < 2 and the receiver is specified to have a VSWR < 1.5.

A commercial computer software package, "General Microwave Circuit Analysis Program" [A1], is used to determine the input VSWR at the attenuator ports attached via one of the equal length cables to the hybrid junction and the source signal generator or receiver as shown in figure 5A. The attenuator and hybrid

junction are modeled with resistive "T" networks to match manufacturer specifications. The signal generator is modeled with 25 ohm and 100 ohm resistors and the receiver is modeled with 33.3 ohm and 75 ohm resistors. Worst case input VSWR values occur at 44 MHz (calculations use the 10 test frequencies) and the smaller of the two resistive values representing receiver and source signal generator impedances. Without the 3 dB attenuators, the source circuit input VSWR is 2.01 and the receiving circuit input VSWR is 1.64. With the attenuators, these VSWR values are 1.5 and 1.4 respectively. The mismatch losses resulting from these non-perfect input impedances are the magnitudes of the errors which may occur if the worst case conditions are realized; 0.17 dB and 0.1 dB respectively. Therefore, a total worst case error of 0.27 dB may occur because the impedances of the source and receiving circuitry attached to the dipole terminals are not exactly 50 ohms at each attenuator. Larger attenuators will reduce these errors even more if the reduced signal levels can be tolerated during the site attenuation measurements.

The conjugate mismatch loss, M_L (referred to simply as mismatch loss in this paper) is calculated as follows [A2]:

$$M_{L} = 10 \log_{10} \frac{|1 - \Gamma_{s} \Gamma_{a}|^{2}}{(1 - |\Gamma_{s}|^{2})(1 - |\Gamma_{a}|^{2})}, \qquad (1A)$$

where Γ_s = complex reflection coefficient of the source circuitry, and

 Γ_a = complex reflection coefficient of the antenna.

Forcing the source VSWR to 1 with large enough attenuators, i.e., forcing $\Gamma_s = 0$, gives

$$M_{L} = 10 \log_{10} \frac{1}{1 - |\Gamma_{a}|^{2}}$$
 (2A)

Computer Program HVD3 evaluates this expression using

$$\Gamma_{a} \equiv R = (ZR - ZCHAR)/(ZR + ZCHAR), \qquad (3A)$$

where ZR = dipole input impedance, and

ZCHAR = 100 ohms, the receiving or source system impedance.

Tuning devices may be used at fixed frequencies to force Γ_s to zero but this technique is not normally used for antenna measurements over a wide range of frequencies.

3. Input Impedance

Dipole input impedance is calculated using well documented equations. "Aside from the shunt admittance of the input region, which may or may not be significant, the input impedance of the antenna is" [A3]:

$$Z_{i} = K_{a} \frac{(K_{a} - M) \cos \beta \ell + j(Z_{a} + j\omega C_{t} K_{a}^{2} - jN) \sin \beta \ell}{(Z_{a} + j\omega C_{t} K_{a}^{2} + jN) \cos \beta \ell + j(K_{a} + M) \sin \beta \ell}, \quad (4A)$$

where
$$K_a = 120 \ln \frac{2l}{a_b} + \frac{120 a_t}{a_b - a_t} \ln \frac{a_t}{a_b}$$

a_t = radius of tip of linearly tapered antenna ,

a_b = radius of base of linearly tapered antenna ,

or
$$K_a = 120 (\ln \frac{2\ell}{a_t} - 1)$$
 for $a_t = a_b$,

 $M = 60 (Cin 2\beta \ell - 1 + \cos 2\beta \ell)$,

 $N = 60 (Si 2\beta\ell - sin 2\beta\ell) ,$

$$\beta = 2\pi/\lambda$$
,

$$Cin x = \int_{0}^{x} \frac{1 - cost}{t} dt,$$

Si x = $\int_{0}^{x} \frac{sin t}{t} dt,$

$$j\omega C_{t} = \frac{ja_{t}}{30\lambda},$$

$$Z_{a} = R_{a} + j X_{a},$$

$$R_{a} = 60 \operatorname{Cin} 2\beta\ell + 30(2 \operatorname{Cin} 2\beta\ell - \operatorname{Cin} 4\beta\ell) \cos 2\beta\ell + 30(\operatorname{Si} 4\beta\ell - 2\operatorname{Si} 2\beta\ell) \sin 2\beta\ell ,$$

and $X_{2} = 60$ Si 2Bl -30(Cin 4Bl - ln4)sin 2Bl -30 Si 4Bl cos 2Bl.

These equations appear in lines 1070 - 1180, Program HVD3 and lines 1600 - 1710, Program ZMM1, both in Appendix II, and apply only to straight dipoles.

4. Dipole Droop

Unfortunately, real dipoles droop at the lower frequencies considered in this paper. A relatively new computer program does enable the calculation of gain pattern and input impedance of arbitrary wire antennas of finite dimensions over perfect ground. Drooping dipole input impedance can be calculated using this program, called MININEC, which is a small version of a larger numerical electromagnetics code [A4, A5].

The profile of the actual 30 MHz dipole was measured and the resulting MININEC model utilizes eight wires, one for each telescoping section comprising the antenna. Each of the 8 wires is represented by 10 equal length segments. Currents on the wire are expanded in pulses centered at the junctions of adjacent segments.

MININEC must be used carefully and enough segments selected to insure near convergence of the input impedance data. Table 1A shows input impedance data calculated for the 30 MHz tapered dipole used for the site attenuation measurement. Input impedance data for the straight dipole calculated using Schelkunoff's equations are included for interest and entitled HVD3, straight. This is the "worst case" dipole because it is physically the longest, has the most droop, and is electrically closest to the ground. Since MININEC and HVD3 give slightly different input impedances for the straight dipoles over perfect ground, the effect of droop is approximated by comparing the difference in

Table 1A. Calculated Input Impedance for Horizontal, 30 MHz, Half-Wave Dipoles Over Perfect Ground

Half-length = 2.4 m

Tip radius = 0.0013 m

Base radius = 0.005 m

Calculation, Geometry	Height, m	Zin, ohms	Mismatch Loss, dB
HVD3, straight	2	60.1 + j33.1	0.46
, ,	4	82.1 - j11.0	0.06
	6	52.2 - j 2.0	0.45
	ω	65.7 - j 0.3	0.19
MININEC, straight	2	62.8 + j 2.3	0.23
	4	79.9 - j44.8	0.32
	6	51.5 - j32.7	0.67
	ω	65.0 - j32.3	0.36
MININEC, drooping	2	62.9 + j 6.8	0.24
	4	81.2 - j41.6	0.27
	6	51.7 - j30.0	0.63
	ω	65.5 - j29.1	0.33

Note: It would require an uneconomical amount of computer time to determine Zin continuously versus height above ground for both straight and drooping 30 MHz dipoles. The 0.05 dB mismatch loss difference, attributable to dipole droop, is assumed to be the worst case value. impedances, more specifically the difference in mismatch loss, between the MININEC results. This difference is at most 0.05 dB and is considered negligible.

Dipole droop will slightly change the effective height of the dipole above ground. However, the current distribution is sinusoidal, not linear, and the least drooped sections support the greatest current density. This, together with the fact that the site attenuation measurement utilizes height scanning, means that the exact height is not important. Figures 6A and 7A show the height versus frequency at which the site attenuation data are calculated.

5. The Mutual Impedance of Dipoles in Echelon

In contrast to the input impedance calculations, King [A6] states that "the mutual impedance equations apply accurately to infinitely thin antennas only, but nonetheless serve as a practical approximation for real antennas of finite thickness". Half-wave dipoles were chosen for the site attenuation measurements because their performance parameters are calculable. Professor Jordan states, "Inasmuch as the actual current distribution is known to be very nearly sinusoidal for thin transmitting antennas (dipoles and monopoles), the (calculated radiated field) obtained is a good approximation to the true power radiated" [A7]. For receiving dipoles he states, "For a resonant-length half-wave receiving dipole, the current amplitude is independent of the thickness of the antenna" [A7, p. 549]. Finally he assigns a problem which completes these comments on the current distribution of half-wave dipoles as follows:

"The current distribution in a half-wave transmitting antenna is known to be nearly sinusoidal. The current distribution of a half-wave dipole receiving antenna is also nearly sinusoidal when the terminals are short-circuited. Using the superposition and compensation theorems, verify that the current distribution of the half-wave dipole as a receiving antenna must be approximately sinusoidal when the terminals are connected to any load impedance Z" [A7, p. 338].

Therefore, calculated site attenuation based almost entirely upon King's mutual impedance formulations and sinusoidal current distribution gives a good reference model when half-wave dipoles are employed.

Subroutine ZZ in Program ZMM1, Appendix II, is used to calculate 18 arguments for sine and cosine integrals required by the mutual impedance equations. Ninety-six of these integral terms are used in the general formulations to obtain the mutual impedance referred to the loop current. The mutual impedance referred to the base current is the quantity needed to evaluate dipole input impedance, the power transmitted, and the power received. Professor Jordan also discusses the relations between the base and loop current references for input and mutual impedance calculations [A7,p. 345-351]. This change of reference is implemented in Program ZMM1 line 03970 as follows:

Z12B (K) = Z12L(K)/((SIN(BETA*L))**2)

where Z12B is the mutual impedance referred to the base current and Z12L is the mutual impedance referred to the loop current.

Since Program ZMM1 calculates site attenuation for both vertical and horizontal dipoles, the reversal of the apparent current direction in the image dipole for horizontal polarization is handled in lines 03940 to 03960.

Subroutine MUTUAL, used to compute the mutual impedance between the dipoles and their images in perfect ground, uses a subset of the general formulations given in Subroutine ZZ. These two subroutines were developed separately and left that way for clarity. Subroutine MUTUAL could be replaced with a call to Subroutine ZZ if ZZ were modified to handle the call. This would shorten the program substantially.

The sine and cosine integrals are evaluated using series expansion equations given in Abramowitz and Stegun [A8]. The computer implementation of these equations was developed many years ago and different ones may be available in mathematical software packages. As noted in Program ZMM1, a linear interpolation is imposed to smooth the functions in the transition region occurring between the large argument and small argument approximations to these integrals.

References for Appendix I

- [A1] Tektronix Product Number 4050A06, Tektronix, Inc., Beaverton, OR; 1975.
- [A2] Beatty, R.W., "Insertion loss concepts," Proc. IEEE, Vol. 52: 663-671; 1964 June.
- [A3] Schelkunoff, S.A., Friis, H.T., <u>Antennas, Theory and Practice</u>, New York, NY: John Wiley & Sons; 1952. 431-434.
- [A4] Julian, A.J., Logan, J.C., Rockway, J.W., "MININEC: A mini-numerical electromagnetic code," Tech. Doc. 516, Naval Ocean Systems Center, San Diego, CA; 1982 6 Sept.
- [A5] Campbell, D.V., "Personal computer applications of MININEC," IEEE AP-S Newsletter, Vol. 26; 1984 February.
- [A6] King, H.E., "Mutual impedance of unequal length antennas in echelon," IEEE T-AP-5, pp. 306-313; July 1957.
- [A7] Jordan, E.C., <u>Electromagnetic Waves and Radiating Systems</u>, Englewood Cliffs, NJ: Prentice-Hall; 1950. p. 365.
- [A8] Abramowitz, M., Stegun, I.A., <u>Handbook of Mathematical Functions With</u> <u>Formulas, Graphs, and Mathematical Tables</u>, Washington, DC; Nat. Bur. of Stand. (U.S.) Applied Mathematics Series 55; 1964; pp. 231-233.















Figure 6A. Heights versus frequency for which site attenuation is calculated for horizontal dipoles.

Solid line	-	3 m separation distance
Dashed line	-	10 m separation distance
Dotted line	-	30 m separation distance.



Figure 7A. Heights versus frequency for which site attenuation is calculated for vertical dipoles.

Solid line	-	3 m separation distance
Dashed line	-	10 m separation distance
Dotted line	_	30 m separation distance.

.

APPENDIX II

Computer Program HVD3

and

Computer Program ZMM1

```
00100 PROGRAM HVD3(INPUT, OUTPUT)
00110C******
                               001200*
       GAIN VS. ELEVATION ANGLE FOR VERTICAL AND HORIZONTAL DIPOLES
        OVER GROUND AND FOR MONOPOLES OVER PERFECT GROUND. NOTE:
00130C*
        SCHELKUNDEF'S MODE THEORY IS USED TO CALCULATE FREE SPACE
001400*
       ANTENNA IMPEDANCE. IF THE DIPOLE FEEDPOINT HEIGHT IS WITHIN
00150C*
00160C* 0.1*LAMBDA OF THE GROUND, THE SELF IMPEDANCE OF THE ANTENNA
00170C* IS CALCULATED USING THE MUTUAL IMPEDANCE WITH 0.0001*LAMBDA
       SEPARATION BECAUSE USE OF SCHELKUNDER'S SELF IMPEDANCE
CC180C*
       MAY RESULT IN NEGATIVE INPUT IMPEDANCES. ANTENNA RADII MAY
001900*
                                                               榆
002000*
       BE ZERD OR THE ANTENNA MAY HAVE A LINEAR TAPER FROM THE
00210C* FEEDPOINT TO THE END.
00230 DIMENSION SI(2), CIN(2), S(21), C(21), T(44), Z(2)
00240 COMPLEX CT, E, J, NRL, R, RH, RV, RH90, RV90, TS0, ZA, ZIND, ZINN, ZM, ZR, ZI, ZM1
OC250 REAL KA, LAMBDA, L, L2, LOSS, MBL, MOND
00270C* THE USER SPECIFIES THE ANTENNA GEOMETRY
00280C*
       NOTE: FOR A MONOPOLE (FED AGAINST PERFECT*
C0290C*
* 200500
       GROUND ONLY) THE CONDUCTIVITY MUST BE
00310C* SET TO ZERO.
00330 1 CONTINUE
00340 2 FORMAT(*TYPE THE REQUESTED DATA AFTER EACH ?
00350+
              FREQUENCY, MHZ*)
00360 PRINT 2
00370 3 EORMAT(F12.4)
00380 READ 3, FRED
00390 4 FORMAT(*ANTENNA HALF-LENGTH, M, LAMBDA/4 OR LESS*)
00400 PRINT 4
00410 READ 3.L
00420 5 FORMAT(#ANTENNA RADII, END THEN FEEDPOINT, M#)
00430 PPINT 5
00440 READ 3,RT
00450 PEAD 3, PB
00460 IF (RT.EQ.0.AND.RB.EQ.0) RT=1.E-30
00470 IF(RT.EQ.1.E-30)RB=1.E-30
00480 6 FORMAT(*GROUND CONDUCTIVITY, MILLIMHOS/M*)
OC490 PPINT 6
00500 READ 3, SIGMA
00520C* PERFECT GROUND AND FREE-SPACE APPROXIMATIONS ARE MADE. *
00540 IF(SIGMA.GE.10000000.0) SIGMA = 1.E30
00550 IF(SIGMA.GE.1000000.0) EPSILON = 1.0
00560 IF(SIGMA.GE.1000000.0) GD TO 8
00570 IF(SIGMA.EQ.0.) EPSILON = 1.
00580 IF(SIGMA.EQ.0.) H=1.E-30
00590 IF(SIGMA.E0.0.) GO TO 10
00600 7 FORMAT (*GPOUND RELATIVE DIELECTRIC CONSTANT*)
00610 PPINT 7
00620 READ 3, EPSILON
00630 8 CONTINUE
00640 9 FORMAT(*DIPOLE FEED-POINT HEIGHT ABOVE GROUND, M*)
00650 PRINT 9
00660 RFAD 3,H
00670 10 CONTINUE
00680 11 FORMAT(*IMPEDANCE OF T-LINE ATTACHED TO ANTENNA, OHMS*)
00690 PRINT 11
00700 READ 3, ZCHAR
00710 J = CMPLX(0.,1.)
00720 LAMADA = 299.792458/FREQ
00730 BETA = 6.2831853/LAMBDA
```

```
00750C* THE CALCULATIONS BEGIN HERE *
00770 12 B = 2. +BETA+L
00780 BL = BETA+L
00790 BH = BETA+H
00800 X = 18. + SIGMA/FREQ
00810 Y . EPSILON
00830C# COMPUTE SINE AND COSINE INTEGRALS FOR SELF #
00840C* IMPEDANCE CALCULATION, SI(M) AND CIN(M)
00860 20 K = 1
00870 T(1) = 1.0
00980 DO 30 M = 1,42
00890 K = K
          + 1
00900 \ 30 \ T(K) = M + T(K-1)
00910 DD 60 M=1,2
00920 Z(M) = M + B
00930 DD 40 N = 1,20
00940 40 S(N) = ((-1)**N)*(Z(M)**(2*N+1))/((2*N+1)*T(2*N+2))
00950 SI(M) = Z(M) + S(1) + S(2) + S(3) + S(4) + S(5) + S(6) + S(7) + S(8) + S(9) + S(10)
00960+
          +S(11)+S(12)+S(13)+S(14)+S(15)+S(16)+S(17)+S(18)+S(19)+S(20)
00970 DD 50 N = 1,20
00980 50 C(N) = ((-1)**N)*(Z(M)**(2*N))/((2*N)*T(2*N+1))
00990 \ 60 \ CIN(M) = -(C(1)+C(2)+C(3)+C(4)+C(5)+C(6)+C(7)+C(8)+C(9)+C(10)
01000+
        +C(11)+C(12)+C(13)+C(14)+C(15)+C(16)+C(17)+C(18)+C(19)+C(20))
01020C* NO LARGE ARGUMENT APPROXIMATION IS REQUIRED FOR SI AND CIN
                                                           *
01030C* BECAUSE OF THE RESTRICTION ON THE DIPOLE LENGTH IN LINE 260
                                                          - 34
01050C* COMPUTE SELF IMPEDANCE, ZI, USING EQN. 108 DF REFERENCE [3]
01070 IF(RT.EQ.RB)KA = 120.*(ALDG(2.*L/RT)-1.)
01080 IF(RT.NE.RB)KA=120.*(ALDG(2*L/RB)+(RT/(RB-RT))*ALDG(RT/RB))
01090 \text{ MBL} = 60.*(CIN(1)-1.+COS(B))
01100 NBL = J*60.*(SI(1)-SIN(B))
01110 ZAR = 60.*CIN(1)+30.*(2.*CIN(1)-CIN(2))*CDS(B)+30.*(SI(2)
01120 +
         -2.*SI(1))*SIN(B)
01130 ZAI = 60.*SI(1)-30.*(CIN(2)-ALDG(4.))*SIN(B)-30.*SI(2)*CDS(B)
01140 ZA = ZAR+J+ZAI
01150 CT = J + RT / (30 . + LAMBDA)
01160 ZINN = KA+((KA-MBL)+COS(BL)+J+(ZA+CT+KA+KA-NBL)+SIN(BL))
01170 ZIND = (ZA+CT+KA+KA+NBL)+COS(BL)+J+(KA+MBL)+SIN(BL)
01180 ZI = ZINN/ZIND
01190 62 CONTINUE
01200 64 FORMAT(+TYPE 1. FOR HORIZONTAL POLARIZATION, O FOR VERTICAL*)
01210 PRINT 64
01220 READ 3, POLAR
01230 MONC = 1.
01240 66 FORMAT (* TYPE 1. FOR DIPOLE OR 0 FOR MONOPOLE*)
01250 IF(POLAR.EQ.O.AND.SIGMA.EQ.O.) PRINT 66
01260 IF(POLAR.EQ.C.AND.SIGMA.EQ.O.) READ 3, MOND
01280C* COMPUTE ANTENNA INPUT IMPEDANCE, ZR. SUBROUTINE MUTUAL *
                                                       *
       RETURNS THE CORRECT MUTUAL IMPEDANCE, ZM, BETWEEN TWO
01290C*
       DIPOLES IN FREE SPACE FOR THE SPECIFIED POLARIZATION
013000+
01320 CALL MUTUAL (BETA, H, J, L, RH90, RV90, ZM, X, Y, POLAR, T)
01330 ZM1=ZM
01340 67 FORMAT(#INPUT IMPEDANCE =#F12.4*+J[*F12.4*]0HMS*)
01350 IF(H.GE.O.I*LAMBDA.OR.SIGMA.EQ.O) GD TD 68
01360 H1=0.0001*LAMBDA
01370 POLAR1=1.
```

```
01380 CALL MUTUAL (BETA, H1, J,L,DUMMY1,DUMMY2,ZM,X,Y,POLAR1,T)
01390 ZJ=ZM/(SIN(BL)++2)
01400 68 CONTINUE
01420C* THE ABOVE 6 LINES HELP PREVENT NEGATIVE INPUT IMPEDANCES *
01430C* FOR HORIZONTAL DIPOLES NEAR GROUND BECAUSE SCHELKUNDFF'S *
01440C*__INPUT_IMPEDANCE DDES_NOT_CONVERGE TO THAT COMPUTED BY
01450C# SUBROUTINE MUTUAL.
01470 IF(PDLAR.EQ.1.) 7R=ZI+PH90*ZM1/(SIN(BL)**2)
01480 IF(PDLAP.EQ.O.) ZR=ZI+RV90+Z41/(SIN(BL)++2)
01490 IF(MOND.EQ.D.) ZR=ZR/2.
01500 PRINT 67, ZR
01510 L2 = L
01520 81 FORMAT (*ENTER NEW HALF-LENGTH IF DESIRED, OTHERWISE ENTER 0*)
01530 PRINT 81
01540 READ 3,L
01550 IF(L.GT.0.) GO TO 12
01560 L = L2
01580C* COMPUTE ANTENNA FACTOR, AF, AND OTHER IMPEDANCE RELATED
                                                       *
01590C* PARAMETERS: MUTUAL IMPEDANCE, VSWR, AND MISMATCH LOSS
                                                       ±
***********************
01610 IF(MONO.E0.0.)AF=(CABS(ZCHAR+ZR)/ZCHAR)/((1./BETA)+TAN(BL/2.))
01620 IF(MOND.EQ.1.)AF=(CABS(ZCHAR+ZR)/ZCHAP)/((2./BETA)+TAN(BL/2.))
01640C* ANTENNA FACTOR = VOLTAGE DIVIDER FACTOR/EFFECTIVE LENGTH *
01660 82 FORMAT(*ANTENNA FACTOR ASSUMING A *F6.2*-OHM RECEIVER*)
01670 83 FORMAT(*AND A 1:1 BALUN ARE EMPLOYED =*F8.2* DB*)
01680 84 FORMAT (*AND TRANSMISSION LINE ARE EMPLOYED =*F8.2* DB*)
01690 PRINT 82, ZCHAR
01700 IF(MOND.EQ.1.) PRINT 83, 20.*ALOG10(AF)
01710 IF(MOND.EQ.O.) PRINT 84, 20. # ALDG10(AF)
01720 90 FORMAT (*MUTUAL IMPEDANCE DUE TO THE ANTENNA'S IMAGE
01730 +
                        IN THE GROUND = + F12.4 + + J[+ F12.4 + ]]HMS+)
01740 IF(MOND.NE.O) PRINT 90, ZR-ZI
01750 R = (ZR - ZCHAR)/(ZR + ZCHAR)
01760 VSWR = (1.+CABS(R))/(1.-CABS(R))
01770 110 FORMAT( +VSWR =+F12.4)
017BO PRINT 110,VSWR
01790 LOSS = 10.0*ALOG10(1.0/(1.0-(CABS(R))*(CABS(R))))
01800 200 FORMAT(*MISMATCH LDSS =*F12.4* DB*)
01810 PRINT 200,LDSS
01830C* COMPUTE REFLECTION COEFFICIENTS, AND THEN *
       GAIN VERSUS ELEVATION ANGLE
018400*
01860 300 FORMAT(*PRESS RETURN FOR GAIN VS PHI, OTHERWISE TYPE CNTL T*)
01870 PRINT 300
01880 PAUSE
01890 350 FORMAT(*ELEV ANGLE INCREMENT FOR BELOW 10 DEG AND THE
01900 +
                         ADDITIONAL INCREMENT FOR ABOVE 10 DEG*)
01910 PRINT 350
01920 READ 3, ENC
01930 RFAD 3, ENC2
01940 355 FORMAT( #HORIZONTAL POLARIZATION #)
01950 356 FORMAT( *VERTICAL POLARIZATION*)
01960 IF(POLAR.EQ.0.) GO TO 660
01970 360 FORMAT(+H-PLANE ELEVATION ANGLE IN DEGREES AND GAIN IN DB+)
01980 PRINT 360
01990 PRINT 355
02000 PSI = 0
02010C**********************
```

```
02020C* HORIZONTAL POLARIZATION *
02C40 400 PSI = PSI + ENC
02050 IF (PSI .GT. 90.0) GD TO 640
02060 IF (PSI .LT. 10.1) GD TD 450
02070 PSI = PSI + ENC2
02080 450 PSIRAD = 0.01745329252*PSI
02090 SNPSI = SIN(PSIRAD)
02100 CNPSI = COS(PSIRAD)
0?110 TSO = CSORT(Y-J*X-(CNPSI**2))
02120 \text{ RH} = (\text{SNPSI} - \text{TSQ})/(\text{SNPSI} + \text{TSQ})
02140C* RH & RV ARE HORIZONTAL AND VERTICAL REFLECTION COEFFICIENTS *
02160 E = TAN(BL/2.)*(CFXP(J*BH*SNPSI)+RH*CONJG(CEXP(J*BH*SNPSI)))
02170 E2 = 120.*CABS(E*E)/REAL(ZR)
02190C* THE GAIN EQUATION IS BASICALLY JORDAN'S EQN. 12-18 USING *
022COC* EFFECTIVE LENGTH = (2/BETA)*TAN(BL/2) FOR THE H-PLANE
02210C* PATTERN OF A HORIZONTAL DIPOLE
                                                      *
                                                      *
02230 \text{ GAIN} = 10.0 \pm \text{ALOG10(E2)}
02240 500 FORMAT(2F15.3)
02250 PRINT 500, PSI, GAIN
02260 GD TD 400
02270 600 FORMAT(*PRESS RETURN FOR E-PLANE PATTERN OR TYPE CNTL T*)
02280 640 PRINT 600
02290 PAUSE
02300 650 FORMAT(*E-PLANE ELEVATION ANGLE IN DEGREES AND GAIN IN DB*)
02310 660 PRINT 650
02320 IF(POLAR.E0.1.) PRINT 355
02330 IF(PDLAR.E0.0.) PRINT 356
02340 PSI = -ENC + 0.0001
02350C***********************
02360C* VERTICAL POLARIZATION *
02370C**********************
02380 700 PSI = PSI + ENC
02390 IF (PSI .GT. 90.0) GO TO 900
02400 IF (PSI .LT. 10.1) GD TD 800
02410 PSI = PSI + ENC2
02420 800 PSIRAD = 0.01745329252*PSI
02430 SNPSI = SIN(PSIRAD)
02440 CNPSI = COS(PSIRAD)
02450 TS0 = CSORT(Y-J+X-(CNPSI++2))
02460 \text{ RV} = (((Y-J+X)+SNPSI)-TSO)/(((Y-J+X)+SNPSI)+TSO)
02470 IF(POLAR.EQ.1.) GO TO 850
02480 E = (CEXP(J*BH*SNPSI)+RV*CONJG(CEXP(J*BH*SNPSI)))*(COS(BL)
O2490+ -CDS(BL*SNPSI))/(CNPSI*SIN(BL))
THE TERMS IN THE PREVIOUS LINE FOLLOWING THE SIXTH * ARE *
02510C*
      A GENERAL FORM OF THE EFFECTIVE LENGTH INCORPORATING
02520C*
                                                      *
02530C* FLEVATION ANGLE FOR A VERTICAL DIPOLE
02550 E2 = 120.*CABS(E*E)/REAL(ZR)
02560 IF(E2.E0.0.) E2=1.E-12
02570 GAIN = 10.0*ALOG10(E2)
02580 PRINT 500, PSI, GAIN
02590 GD TD 700
02610C* VERTICAL POLARIZATION FOR A HORIZONTAL DIPOLE *
02630 850 E = (CEXP(J*BH*SNPSI)-RV*CONJG(CEXP(J*BH*SNPSI)))*(COS(BL)
          -COS(BL*CNPSI))/(SNPSI*SIN(BL))
02640+
```

```
THE TERMS IN THE PREVIOUS LINE FOLLOWING THE SIXTH * ARE A
026600*
* 107650
      GENERAL FORM OF THE EFFECTIVE LENGTH INCORPORATING ELEVATION *
026800* ANGLE FOR THE E-PLANE PATTERN OF A HORIZONTAL DIPOLE.
02690(**********************************
                          02700 F2 = 120.0*CABS(E*E)/REAL(ZR)
02710 \text{ GAIN} = 10.0 \pm 4 \text{LOG10}(\text{E2})
02720 PRINT 500, PSI, GAIN
02730 60 17 700
02750C* CALCULATIONS END HERE. USER MAY START OVER OR CONTINUE WITH *
02750C* SAME DIPOLE ANTENNA GEOMETRY BUT A DIFFERENT HEIGHT, OR STOP *
02770(************************
02790 900 PRINT 870
02800 910 FORMAT(*ENTER DIPPLE HEIGHT IN METERS,*)
02810 920 FORMAT(*TYPE O TO START OVER, OR TYPE ONTL T*)
02830C* NOTE: CONTROL T IS THE BREAK MODE FOR THE CYBER 170/750 *
028400*
           MPERATING SYSTEM.
02860 IF(MOND.E0.0.) PRINT 920
02870 IF(MOND. 50.1.) PRINT 910
02880 IF(MOND.E0.1.) PRINT 920
02890 READ 3,4
02900 IF(H.GT.0.) GO TO 12
02910 IF(H.E0.0.) AD TO 1
02920 END
SUBROUTINE MUTUAL (BETA, H, J, L, RH90, RV90, ZM, X, Y, POLAR, T)
02940
*************
02960 DIMENSION S(21), C(21), T(44), ARG(5), SI(5), CI(5)
02970 REAL L.L. 1.2. L.3. L4
02980 COMPLEY J,RH90, FV90, TSC, 74
03000C* COMPUTE REFLECTION CREEFICIENTS FOR PSI = 90 DEG
C3C2O RV9C = ((Y-J*X)-CSQRT(Y-J*X))/((Y-J*X)+CSQRT(Y-J*X))
03030 PH90 = (1.-CSCRT(Y-J*X))/(1.+CSCRT(Y-J*X))
03040 IF(X.EQ.0.) GD TO 30
DB060C* COMPUTE ARGUMENTS FOR SINE AND COSINE INTEGRALS *
C3080 Y0 = 0.
03090 X0 = 2.*H
03100 IF(POLAR.EQ.0.) GD TO 1
03110 YO = 2.*H
03120 X0 = 0.
031400* XO IS THE SEPARATION BETWEEN THE FEEDPOINTS OF THE VERTICAL
031FOC* DIPOLE AND ITS COLINEAR IMAGE. YO IS THE SEPARATION BETWEEN *
03160C* THE HORIZONTAL DIPOLE AND ITS IMAGE.
03180 1 B = BETA
03190 \ Y02 = YC + 2
03200 4L1 = X0-2.*L
03210 HL2 = XG-L
03220 HL3 = YC
03230 HL4 = X0+L
03240 HL5 = X0+2.*L
03250 ARG(1) = B*(SQRT(Y02+HL1**2)+HL1)
03260 ARG(2) = B*(SORT(Y02+HL2**2)+HL2)
03270 ARG(3) = B*(SCRT(Y02+HL3**2)+HL3)
03280 ARG(4) = R*(SORT(Y02+HL4**2)+HL4)
03290 ARG(5) = B*(SQFT(YC2+4L5**2)+4L5)
```

```
03700 DD 9+ I=1+5
03310 \ Z = A^{p}G(I)
03320 IF(7.GT.1.) 60 TO 7
03340C* COMPUTE SI(APG(I)) AND CI(APG(I)) FOR OCAPG(I)<1 *
03361 DD 5 N=1,20
03370 5 S(N)=((-1)**N)*(7**(2*N+1))/((2*N+1)*T(2*N+2))
03360 \quad SI(I) = Z+S(1)+S(2)+S(3)+S(4)+S(5)+S(6)+S(7)+S(8)+S(9)+S(10)+S(11)
03390+
           +$(12)+$(13)+$(14)+$(15)+$(16)+$(17)+$(18)+$(19)+$(20)
03400 PD 6 N=1,20
03410 6 C(N)=((-1)**N)*(7**(2*N))/((2*N)*T(2*N+1))
03420 SA = C(1)+C(2)+C(3)+C(4)+C(5)+C(6)+C(7)+C(8)+C(9)+C(10)+C(11)
02633+
        +C(12)+C(13)+C(14)+C(15)+C(15)+C(17)+C(18)+C(19)+C(20)
03440 CI(I) = 0.577215665 + ALOG(Z) + SA
03450 GD TD 9
03470°* COMPUTE SI(ARG(I)) AND CI(APG(I)) FOF 1<ARG(I)<INFINITY
03490 7 FZ=(Z**9+32.027264*(Z**6)+25*.187033*(Z**4)+335.67732*(7**2)
03500+
        +3 P.102495)/(7*((2**9)+40.021433*(7**5)+322.62491)*(2**4)
03510+
         +570.23628*(2**2)+157.105423))
03520 G7 = (2**8+42.242855*(7**6)+302.757865*(2**4)+352.018498*(2**2)
03520+
        +21.321899)/((2**2)*((2**8)+48.196927*(2**6)+492.485984
03540+
        *(Z**4)+1114.978885*(Z**2)+449.690326))
03550 SI(I) = 1.570796327-FZ*COS(Z)-GZ*SIN(Z)
03560 \text{ CI(I)} = FZ + SIN(Z) + GZ + COS(Z)
03570 9 CONTINUE
C3580 IF(POLAP.E0.0.) GO TO 15
026000* COMPUTE MUTUAL IMPEDANCE USING EQUATIONS FROM H.E. KING *
03610C* [5] FOR HORIZONTAL POLARIZATION
03630 R12 = 30.*(COS(2.*P*L)*(CI(1)+CI(5)-2.*CI(2)-2.*CI(4)+2.*CI(3))
            +SIN(2.*P*L)*(-SI(1)+SI(5)+2.*SI(2)-2.*SI(4))
03640+
03650+
            +4.*CI(3)+2.*CT(2)-?.*CI(4))
03650 X12 = 30.*(COS(2.*R*L)*(-SI(1)-SI(5)+2.*SI(2)+2.*SI(4)-2.*SI(3))
03470+
            +SIN(2.*P#L)*(-CI(1)+CI(5)+2.*CI(2)-2.*CI(4))
03580+
            -4.*ST(3)+2.*SI(2)+2.*SI(4))
03690 7M = R12+J*X12
03700 RETURN
03720C* COMPUTE MUTUAL IMPEDANCE USING EQUATIONS FROM 4.5. KING *
03730C* FOR VERTICAL POLARIZATION.
03750 15 L1 = ALDG(HL2/HL1)
03760 L2 = ALOG(HL4/HL3)
03770 L3 = ALDG(HL2/HL3)
03780 L4 = ALDG (HL4/HL5)
03790 P1 = B*HL1
03800 82 = 9*413
03810 P3 = B*HL5
03927 R4 = R+4L2
C3830 B5 = B+HL4
03840 P12=15.*(CCS(B1)*(CI(1)-CI(2)+L1)+SIN(B1)*(SI(1)-SI(2))+COS(B2)
        *(CI(3)-CI(4)+L2)+SIN(92)*(SI(3)-SI(4))+CUS(82)*(-CI(2)+CI(3)
03850+
        +L3)+SIN(R2)*(-SI(2)+SI(3))+COS(B3)*(-CI(4)+CI(5)+L4)+SIN(R3)
03860+
        *(-ST(4)+SI(5))+2.*COS(R*L)*COS(B4)*(-CT(2)+CI(3)+L3)+2.
03870+
03880+
        *CDS(P*L)*SIN(P4)*(-SI(2)+SI(3))+2.*CDS(B*L)*CDS(B5)*(CI(3)
03890+
        -CI(4)+L2)+2.*COS(P*L)*SIN(B5)*(SI(3)-SI(4)))
039C0 X12=15.*(CDS(B1)*(-SI(1)+SI(2))+SIN(B1)*(CI(1)-CI(2)-L1)+CDS(B2)
        *(-ST(3)+ST(4))+SIN(B2)*(CI(3)-CI(4)-L2)+CDS(B2)*(SI(2))
03910+
03920+
        -SI(3))+SIN(R2)*(-CI(2)+CI(3)-L3)+CDS(B3)*(SI(4)-SI(5))
        +STN(33)*(-CI(4)+CI(5)+L4)+2*COS(8*L)*COS(P4)*(SI(2)
03930+
```

```
03940+ -SI(3))+2.*CDS(B*L)*SIN(B4)*(-CI(2)+CI(3)-L3)+2.*CCS(B*L)

03950+ *(-SI(3)+SI(4))+2.*CDS(B*L)*SIN(B5)*(CI(3)-CI(4)-L2))

03960 ZM = F12+J*X12

03970 PETURN

03980 3C ZM = 0.0+J*0.0

03990 RETURN

04000 END
```

00100 PROGRAM 7MM1(INPUT, DUTPUT, TAPE1) 00120C* THIS PROGRAM CALCULATES SITE ATTENUATION BETWEEN A PARTICULAR SET OF HALF-WAVE DIPCLES POSITIONED OVER PLANE 00130C* - 38 PERFECT GROUND. THE OPEN CIRCUIT PROBE VOLTAGE IS OBTAINED* 00140C* 00150C# USING MUTUAL IMPEDANCES BETWEEN PROBE AND SOURCE ANTENNAS AND SOURCE ANTENNA IMAGES. INPUT IMPEDANCES ARE CALCULATED* 00160C* 00170C* TAKING MUTUAL IMPEDANCES IN GROUND INTO ACCOUNT. 001800* THE INCIDENT POWER FROM A SIGNAL SOURCE IS 1 W WHICH IS USED WITH THE SOUPCE DIPOLE INPUT IMPEDANCE, REFERRED TO* 001900* 002000* THE BASE, TO DETERMINE THE BASE CURRENT USED IN THE MUTUAL * IMPEDANCE CALCULATIONS. 00210C* *202200 TAPE1 CONTAINS THE SITE ATTENUATION AMPLITUDE DATA. \$00230C* NOTESI 1. THE PROBE IS ASSUMED TO BE ATTACHED DIRECTLY TO A * 00240C* PECEIVER. THAT IS, A LOSSLESS LINE IS ASSUMED. 00250C* * ?. A LOSSLESS BALUN WITH A 1:1 TRANSFORMER RATIO IS - 20 00260C* 00270C* ASSUMED. 3. PROC. FILES D7 OR DZM ARE USED TO PLOT THE DATA. 002800* 00290C* 4. SAVE TAPES WITH REPLACE COMMAND AFTER RUN COMPLETE* 5. IN THE DATA FILES: LEN IS DIPOLE HALF-LENGTH. 003000* 00310C* RT IS THE RADIUS OF THE DIPOLE TIP IN METERS. 003200* RB IS THE RADIUS OF THE BASE OF THE DIPOLE, (AT * THE FEEDPOINT). RT = R8 FOR A STRAIGHT ELEMENT. 00330C* * 003400* OBVIOUSLY, A LINEAR TAPER IS ASSUMED. sk. 00360 DIMENSION T(44),712(1000),71(1000), PREC(1000), VOC(1000), 00370+FRE0(10), LEN(10), RT(10), RB(10) 00380 REAL II, LAMBDA, LEN 00390 COMPLEX J, Z12, Z1, ZP, ZR, ZS1 00400 DATA FREC/30.,44.,65.,97.,143.,210.,311.,459.,677.,1000./ 00410 DATA LEN/2.4, 1.625, 1.082, .717, .49, .331, .221, .147, .09P, .0644/ 00420 DATA RT/.0013,.0026,.0036,.005,.00235,.00235,.00235,.00235, 00430+.00235..00235/ 00450+.00235/ 00460 10 FORMAT(F12.4) 00470 15 FORMAT(2F10.3) 00480 17 FORMAT(215) 00490 20 FORMAT(*INPUT DATA AFTER EACH 7*) 00500 PRINT 20 00510 30 FORMAT(#SOURCE DIPOLE FEEDPOINT HEIGHT OVER PERFECT GROUND, M*) 00520 PRINT 30 00530 READ 10,H2 00540 40 FORMAT(*PROBE DIPOLE NORMAL DISTANCE FROM SOURCE PLANE, M*) 00550 PRINT 40 00560 READ 10, RS 00570 IF(RS.LT.25.) HPSS=1. 00580 IF(RS.GT.25.) HPSS=2. 00590 IF(RS.LT.25.) HPF=4. 00600 IF(RS.GT.25.) HPF=6. 00610 HPI=0.01 00630C* THE ABOVE LINE SETS THE HEIGHT SCAN INCREMENT TO 0.01 M ★ 00650 45 CONTINUE 00660 50 FORMAT(*TYPE 1. FOR HORIZONTAL POLARIZATION, O FOR VERTICAL*) 00670 PRINT 50 00680 READ 10, POLAR 00700C* RECEIVER AND TRANSMITTER IMPEDANCES ARE 100 DHMS * 00720 RECZ=10C. 00730 TXZ =100.

```
00740 INC=0
00750 IF(PDLAR.EQ.0) INC=1

        O0770C*
        INC IS ADJUSTED TO SKIP THE 30 MHZ DATA SET BECAUSE THE 30 MHZ *

        00780C*
        DIPOLE HALF-LENGTH EXCLUDES ITS USE VERTICALLY POLARIZED *

00800 60 INC=INC+1
00810 F=FREQ(INC)
00820 L=LEN(INC)
00830 RT=RT(INC)
00840 RB=RB(INC)
00850 HPS=HPSS
00860 IF(POLAR.EQ.O.AND.HPS.LE.L) HPS=L+0.05
00870 N=INT(0.0001+(HPF-HPS)/HPI)+1
00880 J=CMPLX(0.,1.)
00890 PI=3.14159265
00900 LAMEDA= 299.792458/F
00910 BETA=2.*PI/LAMEDA
00920 K=1
00930 T(1)=1.
00940 DD 200 MM=1,42
00950 K=K+1
00960 200 T(K)=MM+T(K-1)
00970C***********
                    *************
       ALL INPUT AND MUTUAL IMPEDANCES ARE PEFERRED TO THE
00980C*
00990C* BASE OR TERMINAL CURRENT.
01000C ******
                                 ********
01010 CALL IMPED(H2, RETA, J, L, LAMPDA, RT, PB, 7R, POLAR, T)
01020 ZS1=ZR
01030 CALL ZZ (BETA, HPS, HPF, HPI, H2, J, LAMPDA, RS, Z12, POLAR, T)
01040 DD 300 I=1, N
01050 \ 300 \ Z1(I) = Z12(I)
01060 PINC = 1.0
01070 PTR=PINC*(1,-(CABS((ZS1-TXZ)/(ZS1+TXZ)))**2)
01080 I1=SQRT(PTR/REAL(ZS1))
01090 HP=HPS-HPI
01100 DD 500 I=1,N
01110 VOC(I)=CABS(I1*Z1(I))
01120 HP=HP+HPI
01130 CALL IMPED(HP,BETA,J,L,LAM3DA,RT,RB,ZR,POLAR,T)
01140 ZP=ZR
01150 500 PREC(I) = ((VOC(I)/CABS(REC7+ZP)) ++2) + RECZ
01160 PMAX=PREC(1)
01170 DD 600 I=2,N
01180 600 PMAX=AMAX1(PMAX, PREC(I))
01190 SMIN=10. * ALOG10(PINC/PMAX)
THE ABOVE 5 LINES OF CODE FIND THE MAXIMUM RECEIVED *
012100*
        POWER INTO A RECEIVER ATTACHED TO THE PROBE
01220C*
        DIPOLE AND USE IT TO CALCULATE THE MINIMUM RELATIVE
01230C*
                                                        *
01240C* INSERTION LOSS WHICH IS THE SITE ATTENUATION
01260 WRITE(1,15)SMIN,F
01270 800 FORMAT(*EXECUTINGI*)
01280 PRINT 800
01290 IF(FREQ(INC).LT.1000.) GD TD 60
01300 IF(POLAR.EQ.1.) GD TO 45
01310 END
SUBROUTINE IMPED(A, BETA, J, L, LAMBDA, RT, PP, ZR, PDLAR, T)
01330
        IS USED TO OPTAIN THE SELF IMPEDANCE OF THE SOURCE
01340C*
                                                        *
        AND PROBE ANTENNAS USING SCHELKUNDEF'S EQUATIONS
01350C*
                                                        *
       WE START BY COMPUTING THE SINE
01360C*
        AND COSINE INTEGRALS SI(M) AND CIN(M). SUBPOUTINE
01370C*
                                                        *
```

```
MUTUAL IS FINALLY CALLED TO ENABLE THE CALCULATION
01380C*
01390C* OF THE INPUT IMPEDANCE OVER GROUND.
01400C****
                                         **********
01410 DIMENSION SI(2), CIN(2), S(21), C(21), T(44), Z(2)
01420 REAL KA, L, LAMBDA, MBL
01430 CONPLEX CT, J, NBL, R, ZA, ZIND, ZINN, ZH, ZR, ZI
01440 BL = BETA*L
01450 B = 2.+BETA+L
01460 DD 60 M=1,2
01470 Z(H) = H = H
01480 DD 40 N = 1,20
01490 40 S(N) = ((-1)**N)*(Z(M)**(2*N+1))/((2*N+1)*T(2*N+2))
01500 SI(M) = Z(M) + S(1) + S(2) + S(3) + S(4) + S(5) + S(6) + S(7) + S(8) + S(9) + S(10)
01510+
          +S(11)+S(12)+S(13)+S(14)+S(15)+S(16)+S(17)+S(18)+S(19)+S(20)
01520 DD 50 N = 1,20
01530 50 C(N) = ((-1)**N)*(Z(M)**(2*N))/((2*N)*T(2*N+1))
01540 60 CIN(M)=-(C(1)+C(2)+C(3)+C(4)+C(5)+C(6)+C(7)+C(8)+C(9)+C(10)
        +C(11)+C(12)+C(13)+C(14)+C(15)+C(16)+C(17)+C(18)+C(19)+C(20))
01550 +
01570C* COMPUTE SELF IMPEDANCE, ZI, USING EQN 108 DF SCHELKUNDFF
                                                               *
01580C+ AND FRIIS.
01600 IF(RT.EQ.RB) KA = 120.*(ALDG(2.*L/RT)-1.)
01610 IF(RT.NE.RB) KA=120.*(ALOG(2+L/RB)+(RT/(RB-RT))*ALOG(RT/RB))
01620 \text{ MBL} = 60.*(CIN(1)-1.+COS(B))
01630 NBL = J*60.*(SI(1)-SIN(B))
01640 ZAR = 60.*CIN(1)+30.*(2.*CIN(1)-CIN(2))*CUS(B)+30.*(SI(2)
01650+
          -2.*SI(1))*SIN(B)
01660 ZAI = 60.*SI(1)-30.*(CIN(2)-ALOG(4.))*SIN(B)-30.*SI(2)*COS(B)
01670 ZA = ZAR+J+ZAI
01680 CT = J*RT/(30.*LAMBDA)
01690 ZINN = KA+((KA-HBL)+COS(BL)+J+(ZA+CT+KA+KA-NBL)+SIN(BL))
01700 \text{ ZIND} = (\text{ZA+CT} \times \text{KA} \times \text{KA} + \text{NBL}) \times \text{CDS(BL}) + J \times (\text{KA} + \text{MBL}) \times \text{SIN(BL})
01710 ZI = ZINN/ZIND
01720 IF(A.GT.0.066+LAMBDA) GD TD 70
01730 H1=0.0001+LAMBDA
01740 POL=1.
01750 CALL MUTUAL (BETA, H1, J,L,ZM, POL, T)
01760 ZI=ZM/(SIN(BL)**2)
01770 70 CONTINUE
01790C* THE ABOVE 6 LINES HELP PREVENT NEGATIVE INPUT IMPEDANCES *
O1800C* FOR HORIZONTAL DIPOLES NEAR GROUND BECAUSE SCHELKUNDFF'S *
O1810C* INPUT IMPEDANCE DOES NOT CONVERGE TO THAT COMPUTED BY
                                                           *
01820C* SUBROUTINE MUTUAL.
01840 CALL MUTUAL (BETA, A, J, L, ZM, POLAR, T)
01850 IF(POLAR.EQ.1.) ZR=ZI-ZM/((SIN(BL)**2))
01860 IF(POLAR.EQ.O.) ZR=ZI+ZM/((SIN(BL)**2))
01870 RETURN
01880 END
SUBROUTINE MUTUAL (BETA, H, J, L, ZM, POLAR, T)
01900
        USES EQUATIONS FROM H.E. KING, "MUTUAL IMPEDANCE OF
01910C*
        UNEQUAL LENGTH ANTENNAS IN ECHELON", IRE TRANS. ANT.
                                                          *
01920C*
v1930C*
        AND PROP., JULY 1957.
                      **************
01940C**********
01950 DIMENSION S(21), C(21), T(44), ARG(5), SI(5), CI(5)
01960 REAL L, L1, L2, L3, L4
01970 COMPLEX J,ZM
J1990C* CUMPUTE ARGUMENTS FOR SINE AND COSINE INTEGRALS *
02010 Y0 = 0.
```

```
02020 X0 = 2. *H
02030 IF(POLAR.EQ.0.) GD TO 1
02040 Y0 = 2.*H
02050 \times 0 = 0.
02070C* XO IS THE SEPARATION BETWEEN THE FEEPPOINTS OF THE VERTICAL *
       DIPOLE AND ITS COLINEAR IMAGE. YO IS THE SEPARATION BETWEEN *
02080C*
02090C+ THE HORIZONTAL DIPOLE AND ITS IMAGE.
02110 \ 1 \ B = BETA
02120 Y02 = Y0 + + 2
02130 HL1 = X0-2.*L
02140 HL2 = X0-L
02150 HL3 = X0
02160 HL4 = X0+L
02170 HL5 = X0+2.*L
02180 ARG(1) = B*(SQRT(Y02+HL1**2)+HL1)
02190 ARG(2) = B*(SQRT(Y02+HL2**2)+HL2)
02200 ARG(3) = B*(SORT(Y02+HL3**2)+HL3)
02210 \ ARG(4) = B*(SQRT(Y02+HL4*+2)+HL4)
02220 ARG(5) = B*(SQRT(Y02+HL5**2)+HL5)
02230 DO 9, I=1,5
02240 Z = ARG(I)
02250 IF(Z.GT.1.) GD TD 7
02270C* COMPUTE SI(ARG(I)) AND CI(APG(I)) FOR O<ARG(I)<1 *
02290 DO 5 N=1,20
02300 \ 5 \ S(N) = ((-1) + N) + (Z + (2 + N + 1)) / ((2 + N + 1) + T(2 + N + 2))
02310 SI(I) = Z+S(1)+S(2)+S(3)+S(4)+S(5)+S(6)+S(7)+S(8)+S(9)+S(10)+S(11)
02320+
           +$(12)+$(13)+$(14)+$(15)+$(16)+$(17)+$(18)+$(19)+$(20)
02330 DD 6 N=1,20
02340 6 C(N)=((-1)++N)+(Z++(2+N))/((2+N)+T(2+N+1))
02350 SA = C(1)+C(2)+C(3)+C(4)+C(5)+C(6)+C(7)+C(8)+C(9)+C(10)+C(11)
02360+
        +C(12)+C(13)+C(14)+C(15)+C(16)+C(17)+C(18)+C(19)+C(20)
02370 CI(I) = 0.577215665 + ALOG(Z) + SA
02380 60 70 9
02400C* COMPUTE SI(APG(I)) AND CI(ARG(I)) FDR 1<ARG(I)<INFINITY *
02420 7 FZ=(Z**8+38.027264*(Z**6)+265.187033*(Z**4)+335.67732*(Z**2)
02430+
         +38.102495)/(2*((2**8)+40.021433*(2**6)+322.624911*(2**4)
02440+
         +570.23628*(Z**2)+157.105423))
02450 GZ = (2**8+42.242855*(2**6)+302.757865*(2**4)+352.018498*(2**2)
         +21.821899)/((Z++2)+((Z++8)+48.196927+(Z++6)+482.485984
02460+
         +(Z**4)+1114.978885*(Z**2)+449.690326))
02470 +
02480 SI(I) = 1.570796327 - FZ + COS(Z) - GZ + SIN(Z)
02490 CI(I) = FZ + SIN(Z) - GZ + CDS(Z)
02500 9 CONTINUE
02510 IF(POLAR.E0.0.) GD TO 15
02530C* COMPUTE MUTUAL IMPEDANCE USING EQUATIONS FROM H.E. KING *
02540C* FOR HORIZONTAL POLARIZATION *
02560 R12= 30.*(COS(2.*B*L)*(CI(1)+CI(5)-2.*CI(2)-2.*CI(4)+2.*CI(3))
02570+
            +SIN(2.*B*L)*(-SI(1)+SI(5)+2.*SI(2)-2.*SI(4))
            +4 *CI(3) -2 *CI(2) -2 *CI(4))
02580+
02590 X12 = 30.*(CDS(2.*B*L)*(-SI(1)-SI(5)+2.*SI(2)+2.*SI(4)-2.*SI(3))
02600+
            +SIN(2.*B+L)+(-CI(1)+CI(5)+2.*CI(2)-2.*CI(4))
02610+
            -4.*SI(3)+2.*SI(2)+2.*SI(4))
02620 ZM = R12+J+X12
02630 RETURN
02650C* COMPUTE MUTUAL IMPEDANCE USING EQUATIONS FROM H.E. KING *
```

```
02660C* FOR VERTICAL POLARIZATION
02670C*****
             02680 15 L1 = ALOG(HL2/HL1)
02690 L2 = ALOG(HL4/HL3)
02700 L3 = ALDG(HL2/HL3)
02710 L4 = ALOG(HL4/HL5)
02720 B1 = B + HL1
02730 B2 = B + HL3
02740 B3 = B+HL5
02750 B4 = 9 \pm HL2
02760 B5 = B + HL4
02770 R12=15.*(CDS(B1)*(CI(1)-CI(2)+L1)+SIN(B1)*(SI(1)-SI(2))+CDS(B2)
02780 +
         *(CI(3)-CI(4)+L2)+SIN(B2)*(SI(3)-SI(4))+COS(B2)*(-CI(2)+CI(3)
         +L3)+SIN(B2)*(-SI(2)+SI(3))+CDS(B3)*(-CI(4)+CI(5)+L4)+SIN(B3)
02790+
02800+
         *(-SI(4)+SI(5))+2.*COS(B*L)*COS(B4)*(-CI(2)+CI(3)+L3)+2.
         *COS(B*L)*SIN(B4)*(-SI(2)+SI(3))+2.*COS(B*L)*COS(B5)*(CI(3)
02810 +
         -CI(4)+L2)+2.*CDS(B*L)*SIN(B5)*(SI(3)-SI(4)))
02820+
02830 X12 = 15.*(CDS(B1)*(-SI(1)+SI(2))+SIN(B1)*(CI(1)-CI(2)-L1)+COS(B2)
02840+
         *(-SI(3)+SI(4))+SIN(B2)*(CI(3)-CI(4)-L2)+COS(B2)*(SI(2)
02850+
         -SI(3))+SIN(B2)*(-CI(2)+CI(3)-L3)+COS(B3)*(SI(4)-SI(5))
02860 +
         +SIN(B3)*(-CI(4)+CI(5)-L4)+2.*COS(B*L)*COS(B4)*(SI(2)
02870 +
         -SI(3))+2.*COS(B*L)*SIN(B4)*(-CI(2)+CI(3)-L3)+2.*COS(B*L)
02880+
         *(-SI(3)+SI(4))+2.*CDS(B*L)*SIN(B5)*(CI(3)-CI(4)-L2))
02890 ZM = R12+J*X12
02900 RETURN
02910 END
02930
        SUBROUTINE ZZ(BETA, HPS, HPF, HPI, HS, J, L, LAMBDA, RS, Z12B, POLAR, T)
02940C*
        COMPUTES THE MUTUAL IMPEDANCE BETWEEN EACH SOURCE AND SOURCE
        IMAGE AND THE PROBE ANTENNA. M=O FOR THE PROBE AND DNE SOURCE *
02950C*
02960C* CALCULATION AND M=1 FOR THE PROBE TO IMAGE CALCULATIONS.
02980 DIMENSION ARG(18), S(21), C(21), T(44), SI(18), CI(18), P12(1000),
02990+X12(1000),712L(1000),712B(1000),SIS(18),SIL(18),CIS(18),CIL(18)
03000 REAL LAMBDA JL
03010 COMPLEX J,Z12L,Z12B
03020 B=BETA
03030 M=0
03040 10 HP=HPS-HPI
03050 K=0
03060 20 HP=HP+HPI
03070 K=K+1
03080 IF(PDLAR.EQ.0.) D=RS
03090 IF(PDLAR.E0.1.AND.H.E0.0.) D=SORT(RS*RS+(HS-HP)**2)
03100 IF(POLAR.E0.1.AND.M.E0.1.) D=SORT(RS*RS+(HS+HP)**2)
03110 IF(POLAR.EQ.O.AND.M.EQ.O.) H=HS-HP-L
03120 IF(POLAR.E0.1.) H=-L
03130 IF(POLAR.EQ.O.AND.M.EQ.1.) H=HS+HP-L
03140 ARG(1)=B*(SORT(D*D+(H-L)**2)+(H-L))
03150 ARG(2)=B+(SQRT(D+D+(H-L)++2)-(H-L))
03160 ARG(3)=B*(SORT(D*D+(H+L)**2)-(H+L))
03170 ARG(4)=B*(SQRT(D*D+(H+L)**2)+(H+L))
03180 ARG(5)=B*(SQRT(D*D+(H)**2)+(H))
03190 ARG(6)=B*(SQRT(D+D+(H)++2)-(H))
03200 ARG(7)=B*(SORT(D*D+(H+2*L)**2)-(H+2*L))
03210 ARG(8)=B*(SQRT(D*D+(H+2*L)**2)+(H+2*L))
03220 ARG(9)=B*(SORT(D*D+(H+L)**2)+(H+L))
03230 ARG(10)=B*(SQRT(D*D+(H+L)**2)-(H+L))
03240 ARG(11)=B*(SORT(D*D+(H+3*L)**2)-(H+3*L))
03250 ARG(12)=B*(SQRT(D*D+(H+3*L)**2)+(H+3*L))
03260 APG(13)=B*(SORT(D*D+H+H)-H)
03270 ARG(14)=B*(SQRT(D*D+H*H)+H)
03280 ARG(15)=B*(SQRT(D*D+(H+L)**2)-(H+L))
03290 ARG(16)=B*(SQRT(D*D+(H+L)**2)+(H+L))
```

```
03300 ARG(17) = B + (SQRT(D+D+(H+2,+L)++2)-(H+2,+L))
03310 ARG(18)=B*(SQRT(D*D+(H+2.*L)**2)+(H+2.*L))
03320 DD 100, I=1, 18
03330 Z=ARG(I)
03340 IF(Z.GT.2.) GD TD 70
03360C* COMPUTE SI(ARG(I)) AND CI(ARG(I)) FOR O<ARG(I)<1 *
03380 DD 50 N=1,20
03390 50 S(N)=((-1)**N)*(Z**(2*N+1))/((2*N+1)*T(2*N+2))
03400 SIS(I) = Z + S(1) + S(2) + S(3) + S(4) + S(5) + S(6) + S(7) + S(B) + S(9) + S(10) + S(11)
            +S(12)+S(13)+S(14)+S(15)+S(16)+S(17)+S(18)+S(19)+S(20)
03410 +
03420 DD 60 N=1,20
03430 60 C(N) = ((-1) **N) * (Z**(2*N)) / ((2*N) *T(2*N+1))
03440 SA = C(1)+C(2)+C(3)+C(4)+C(5)+C(6)+C(7)+C(8)+C(9)+C(10)+C(11)
03450+
         +C(12)+C(13)+C(14)+C(15)+C(16)+C(17)+C(18)+C(19)+C(20)
03460 CIS(I) = 0.577215665 + ALOG(Z) + SA
034BOC* COMPUTE SI(ARG(I)) AND CI(ARG(I)) FOR 1<ARG(I)<INFINITY *
03500 IF(Z.LE.1.) GD TO 80
03510 70 FZ=(Z**B+3B.027264*(Z**6)+265.187033*(Z**4)+335.67732*(Z**2)
03520+
         +38.102495)/(Z*((Z**8)+40.021433*(Z**6)+322.624911*(Z**4)
03530+
         +570.23628*(Z**2)+157.105423))
03540 GZ = (Z**8+42.242855*(Z**6)+302.757865*(Z**4)+352.018498*(Z**2)
03550+
         +21.B21899)/((Z**2)*((Z**8)+4B.196927*(Z**6)+4B2.4B5984
03560+
         *(Z**4)+1114.978885*(Z**2)+449.690326))
03570 SIL(I) = 1.570796327-FZ*CDS(Z)-GZ*SIN(Z)
03580 CIL(I)=F7*SIN(Z)-GZ*CDS(Z)
036COC* THE FOLLOWING SIX LINES ARE AN INTERPOLATION ROUTINE TO *
03610C*
       SMOOTH THE TRANSITION REGION BETWEEN THE LARGE AND SMALL *
03620C*
       ARGUMENT APPROXIMATIONS TO THE SINE AND COSINE INTEGRALS *
03640 80 IF(Z.LE.1.) SI(I)=SIS(I)
03650 IF(Z.LE.1.) CI(I)=CIS(I)
03660 IF(Z.GT.1.AND.Z.LE.2.) SI(I)=SIS(I)+(Z-1.)*(SIL(I)-SIS(I))
03670 IF(Z.GT.1.AND.Z.LE.2.) CI(I)=CIS(I)+(Z-1.)*(CIL(I)-CIS(I))
036B0 IF(Z.GT.2.) SI(I)=SIL(I)
03690 IF(Z.GT.2.) CI(I)=CIL(I)
03700 100 CONTINUE
03720C* COMPUTE MUTUAL IMPEDANCE USING EQUATIONS FROM H.E. KING
03740 R12(K)=15.*(COS(B*(L-H))*(CI(1)+CI(2)-CI(5)-CI(6))+SIN(B*(L-H))
03750+*(-SI(1)+SI(2)+SI(5)-SI(6))+CDS(B*(L+H))*(CI(3)+CI(4)-CI(7)-CI(B))
03760++$IN(B*(L+H))*(-SI(3)+SI(4)+SI(7)-SI(8))+COS(B*(L-2.*L-H))*
03770+(-CI(5)-CI(6)+CI(9)+CI(10))+SIN(B*(L-2*L-H))*(SI(5)-SI(6)-SI(9)
037B0++SI(10))+COS(B*(L+2.*L+H))*(-CI(7)-CI(B)+CI(11)+CI(12))+SIN(B*
03790+(L+2.*L+H))*(SI(7)-SI(B)-SI(11)+SI(12))+2.*COS(B*L)*COS(B*H)*
03900+(-CI(13)-CI(14)+CI(15)+CI(16))+2.*COS(B*L)*SIN(B*H)*(SI(13)-SI(14))
03B10+-SI(15)+SI(16))+2.*COS(B*L)*COS(B*(2.*L+H))*(CI(15)+CI(16)-CI(17)
03820+-CI(18))+2.*COS(B*L)*SIN(B*(2.*L+H))*(-SI(15)+SI(16)+SI(17)
03830 + - SI(18)))
03840 X12(K)=15.*(CDS(B*(L-H))*(-SI(1)-SI(2)+SI(5)+SI(6))+SIN(B*(L-H))
03B50+*(-CI(1)+CI(2)+CI(5)-CI(6))+CDS(B*(L+H))*(-SI(3)-SI(4)+SI(7)+SI(8))
O3B60++SIN(B*(L+H))*(-CI(3)+CI(4)+CI(7)-CI(B))+COS(B*(L-2.*L-H))*
03870+(+SI(5)+SI(6)-SI(9)-SI(10))+SIN(B*(L-2.*L-H))*(CI(5)-CI(6)-CI(9)
038B0++CI(10))+COS(B*(L+2.*L+H))*(+SI(7)+SI(B)-SI(11)-SI(12))+SIN(B*
03B90+(L+2.*L+H))*(CI(7)-CI(B)-CI(11)+CI(12))+2.*COS(B*L)*COS(B*H)*
03900+(+SI(13)+SI(14)-SI(15)-SI(16))+2.*COS(B*L)*SIN(B*H)*(CI(13)-CI(14)
03910+-CI(15)+CI(16))+2.*COS(8*L)*COS(8*(2.*L+H))*(-SI(15)-SI(16)+SI(17)
03920++SI(1B))+2.*COS(B*L)*SIN(B*(2.*L+H))*(-CI(15)+CI(16)+CI(17)
03930+-CI(18)))
```

43

NBS-114A (REV. 2-80)								
U.S. DEPT. OF COMM.	1. PUBLICATION OR REPORT NO.	2. Performing Organ. Report No.	3. Publication Date					
SHEET (See instructions)	NBS TN-1089		November 1985					
4. TITLE AND SUBTITLE	1	I	,					
Site Attenuat	ion							
5. AUTHOR(S)								
R. G. FitzGer	rell							
6. PERFORMING ORGANIZA	TION (If joint or other than NBS	, see instructions)	7. Contract/Grant No.					
NATIONAL BUREAU OF	STANDARDS							
DEPARTMENT OF COMM	ERCE		8. Type of Report & Period Covered					
HASHINGTON, D.C. 2023	14							
9. SPONSORING ORGANIZA	TION NAME AND COMPLETE A	DDRESS (Street, City, State, ZIP)					
10. SUPPLEMENTARY NOTE	ΞS		·····					
Document describes	a computer program: SE-185 EIE	S Software Summary is attached						
11. ABSTRACT (A 200-word	or less factual summary of most	significant information. If docum	ent includes a significant					
bibliography or literature	survey, mention it here)							
Site attenuation is a measure of performance of an open test site used								
at trequencie emission and	emission and susceptibility testing. These sites typically consist of							
a large, obstruction-free ground plane and the hemisphere above it.								
Site attenuation of an ideal site is calculated and compared to data								
measured using the 30 m by 60 m NBS ground screen.								
12. KEY WORDS (Six to twelv	ve entries; alphabetical order; co	pitalize only proper names; and s	eparate key words by semicolons)					
antenna measu	rements; dipole anten	nas; site attenuation						
13. AVAILABILITY		· · · · · · · · · · · · · · · · · · ·	14. NO. OF					
XX Unlimited			PRINTED PAGES					
For Official Distribut	tion. Do Not Release to NTIS		48					
Order From Superinte	ndent of Documents, U.S. Govern	ment Printing Office, Washington,	D.C. 15. Price					
20702.								
Urder From National	lechnical Information Service (N	(115), Springfield, VA. 22161						

U.S. GOVERNMENT PRINTING OFFICE: 1985-676-877 45044 REGION NO. 8



Periodical

Journal of Research—The Journal of Research of the National Bureau of Standards reports NBS research and development in those disciplines of the physical and engineering sciences in which the Bureau is active. These include physics, chemistry, engineering, mathematics, and computer sciences. Papers cover a broad range of subjects, with major emphasis on measurement methodology and the basic technology underlying standardization. Also included from time to time are survey articles on topics closely related to the Bureau's technical and scientific programs. Issued six times a year.

Nonperiodicals

Monographs—Major contributions to the technical literature on various subjects related to the Bureau's scientific and technical activities.

Handbooks—Recommended codes of engineering and industrial practice (including safety codes) developed in cooperation with interested industries, professional organizations, and regulatory bodies.

Special Publications—Include proceedings of conferences sponsored by NBS, NBS annual reports, and other special publications appropriate to this grouping such as wall charts, pocket cards, and bibliographies.

Applied Mathematics Series—Mathematical tables, manuals, and studies of special interest to physicists, engineers, chemists, biologists, mathematicians, computer programmers, and others engaged in scientific and technical work.

National Standard Reference Data Series—Provides quantitative data on the physical and chemical properties of materials, compiled from the world's literature and critically evaluated. Developed under a worldwide program coordinated by NBS under the authority of the National Standard Data Act (Public Law 90-396). NOTE: The Journal of Physical and Chemical Reference Data (JPCRD) is published quarterly for NBS by the American Chemical Society (ACS) and the American Institute of Physics (AIP). Subscriptions, reprints, and supplements are available from ACS, 1155 Sixteenth St., NW, Washington, DC 20056.

Building Science Series—Disseminates technical information developed at the Bureau on building materials, components, systems, and whole structures. The series presents research results, test methods, and performance criteria related to the structural and environmental functions and the durability and safety characteristics of building elements and systems.

Technical Notes—Studies or reports which are complete in themselves but restrictive in their treatment of a subject. Analogous to monographs but not so comprehensive in scope or definitive in treatment of the subject area. Often serve as a vehicle for final reports of work performed at NBS under the sponsorship of other government agencies.

Voluntary Product Standards—Developed under procedures published by the Department of Commerce in Part 10, Title 15, of the Code of Federal Regulations. The standards establish nationally recognized requirements for products, and provide all concerned interests with a basis for common understanding of the characteristics of the products. NBS administers this program as a supplement to the activities of the private sector standardizing organizations.

Consumer Information Series—Practical information, based on NBS research and experience, covering areas of interest to the consumer. Easily understandable language and illustrations provide useful background knowledge for shopping in today's technological marketplace.

Order the above NBS publications from: Superintendent of Documents, Government Printing Office, Washington, DC 20402.

Order the following NBS publications—FIPS and NBSIR's—from the National Technical Information Service, Springfield, VA 22161.

Federal Information Processing Standards Publications (FIPS PUB)—Publications in this series collectively constitute the Federal Information Processing Standards Register. The Register serves as the official source of information in the Federal Government regarding standards issued by NBS pursuant to the Federal Property and Administrative Services Act of 1949 as amended, Public Law 89-306 (79 Stat. 1127), and as implemented by Executive Order 11717 (38 FR 12315, dated May 11, 1973) and Part 6 of Title 15 CFR (Code of Federal Regulations).

NBS Interagency Reports (NBSIR)—A special series of interim or final reports on work performed by NBS for outside sponsors (both government and non-government). In general, initial distribution is handled by the sponsor; public distribution is by the National Technical Information Service, Springfield, VA 22161, in paper copy or microfiche form.

U.S. Department of Commerce National Bureau of Standards Gaithersburg, MD 20899

Official Business Penalty for Private Use \$300



U.S.MAIL

POSTAGE AND FEES PAID U.S. DEPARTMENT OF COMMERCE COM-215

FIRST CLASS