




A11105 982906

REFERENCE

NIST
PUBLICATIONS



NBS TECHNICAL NOTE **1248**

U.S. DEPARTMENT OF COMMERCE/National Bureau of Standards

Traceability of Laser Interferometric Length Measurements

Howard P. Layer and W. Tyler Estler

QC
100
U5753
NO.1248
1988

T

he National Bureau of Standards¹ was established by an act of Congress on March 3, 1901. The Bureau's overall goal is to strengthen and advance the Nation's science and technology and facilitate their effective application for public benefit. To this end, the Bureau conducts research to assure international competitiveness and leadership of U.S. industry, science and technology. NBS work involves development and transfer of measurements, standards and related science and technology, in support of continually improving U.S. productivity, product quality and reliability, innovation and underlying science and engineering. The Bureau's technical work is performed by the National Measurement Laboratory, the National Engineering Laboratory, the Institute for Computer Sciences and Technology, and the Institute for Materials Science and Engineering.

The National Measurement Laboratory

Provides the national system of physical and chemical measurement; coordinates the system with measurement systems of other nations and furnishes essential services leading to accurate and uniform physical and chemical measurement throughout the Nation's scientific, community, industry, and commerce; provides advisory and research services to other Government agencies; conducts physical and chemical research; develops, produces, and distributes Standard Reference Materials; provides calibration services; and manages the National Standard Reference Data System. The Laboratory consists of the following centers:

- Basic Standards²
- Radiation Research
- Chemical Physics
- Analytical Chemistry

The National Engineering Laboratory

Provides technology and technical services to the public and private sectors to address national needs and to solve national problems; conducts research in engineering and applied science in support of these efforts; builds and maintains competence in the necessary disciplines required to carry out this research and technical service; develops engineering data and measurement capabilities; provides engineering measurement traceability services; develops test methods and proposes engineering standards and code changes; develops and proposes new engineering practices; and develops and improves mechanisms to transfer results of its research to the ultimate user. The Laboratory consists of the following centers:

- Applied Mathematics
- Electronics and Electrical Engineering²
- Manufacturing Engineering
- Building Technology
- Fire Research
- Chemical Engineering³

The Institute for Computer Sciences and Technology

Conducts research and provides scientific and technical services to aid Federal agencies in the selection, acquisition, application, and use of computer technology to improve effectiveness and economy in Government operations in accordance with Public Law 89-306 (40 U.S.C. 759), relevant Executive Orders, and other directives; carries out this mission by managing the Federal Information Processing Standards Program, developing Federal ADP standards guidelines, and managing Federal participation in ADP voluntary standardization activities; provides scientific and technological advisory services and assistance to Federal agencies; and provides the technical foundation for computer-related policies of the Federal Government. The Institute consists of the following divisions:

- Information Systems Engineering
- Systems and Software Technology
- Computer Security
- Systems and Network Architecture
- Advanced Systems

The Institute for Materials Science and Engineering

Conducts research and provides measurements, data, standards, reference materials, quantitative understanding and other technical information fundamental to the processing, structure, properties and performance of materials; addresses the scientific basis for new advanced materials technologies; plans research around cross-cutting scientific themes such as nondestructive evaluation and phase diagram development; oversees Bureau-wide technical programs in nuclear reactor radiation research and nondestructive evaluation; and broadly disseminates generic technical information resulting from its programs. The Institute consists of the following divisions:

- Ceramics
- Fracture and Deformation³
- Polymers
- Metallurgy
- Reactor Radiation

¹Headquarters and Laboratories at Gaithersburg, MD, unless otherwise noted; mailing address Gaithersburg, MD 20899.

²Some divisions within the center are located at Boulder, CO 80303.

³Located at Boulder, CO, with some elements at Gaithersburg, MD

Traceability of Laser Interferometric Length Measurements

Howard P. Layer

Center for Basic Standards
National Measurement Laboratory
National Bureau of Standards
Gaithersburg, MD 20899

W. Tyler Estler

Center for Manufacturing Engineering
National Engineering Laboratory
National Bureau of Standards
Gaithersburg, MD 20899

Issued June 1988



U.S. DEPARTMENT OF COMMERCE, C. William Verity, Secretary
NATIONAL BUREAU OF STANDARDS, Ernest Ambler, Director

National Bureau of Standards Technical Note 1248
Natl. Bur. Stand. (U.S.), Tech. Note 1248, 5 pages (June 1988)
CODEN: NBTNAE

U.S. GOVERNMENT PRINTING OFFICE
WASHINGTON: 1988

For sale by the Superintendent of Documents, U.S. Government Printing Office, Washington DC 20402

TRACEABILITY OF LASER INTERFEROMETRIC LENGTH MEASUREMENTS

Howard P. Layer
Center for Basic Standards
National Measurement Laboratory

W. Tyler Estler
Center for Manufacturing Engineering
National Engineering Laboratory

National Bureau of Standards, Gaithersburg, MD 20899

The National Bureau of Standards maintains the standard of length in accordance with the definition of the meter adopted by the General Conference of Weights and Measures (1983).¹ This standard is used within NBS to form a consistent system of units which are used in NBS calibration services.

LENGTH STANDARD: The meter is defined as "the length of the path travelled by light in a vacuum during the time interval of $1/c = 1/299\,792\,458$ of a second."² This definition establishes a fixed value for the speed of light, c , in a vacuum. Neither this definition nor the document which describes it discusses length measurements when the index of refraction is other than unity. Such information must come from other sources.

The realization of the meter for industrial and scientific measurements relies on several types of stabilized lasers operating under specified conditions, for which vacuum wavelengths have been recommended by the International Bureau of Weights and Measures.³ The wavelengths of these lasers are intrinsic standards. The iodine stabilized Helium-Neon laser operating at 633 nm is the instrument most suitable for the practical realization of the meter because of its high accuracy (about 2.3 parts in 10^{10}), its portability, and the ease with which it can be used to calibrate working lasers.⁴ The calibration of a working laser yields its vacuum wavelength and is accomplished by measuring the frequency difference between the working and the standard lasers using a heterodyne photodiode and a frequency counter. This measurement is independent of the index of refraction of air.

LENGTH MEASUREMENTS: For length measurements in air, the wavelength of the calibrated working laser must be adjusted for the difference in the index of refraction, n , between air and vacuum.⁵ The wavelength of laser light in air is equal to its vacuum wavelength divided by the index of refraction of air and is smaller by about 2.7 parts in 10^4 for standard atmospheric conditions. For accurate length measurements, the value of the index of refraction must either be measured directly or calculated from an empirically derived formula based on such measurements.⁶ The variables which most strongly affect the value of n are temperature, relative humidity, barometric pressure, and CO_2 abundance. The accuracy of the calculation using the empirical formula is limited to about 1 part in 10^7 when state-of-the-art technology is used for measuring these four parameters.

LASER INTERFEROMETER SYSTEMS: Relating practical laser interferometric translation measurements to the length standard is difficult because the meter is defined in terms of a vacuum wavelength whereas length and translation

measurements in the laboratory and shop are made in air. To account for this difference, laser interferometer systems incorporate transducers for measuring atmospheric conditions and the algorithm for computing the wavelength in air, all of which must be correct in order for measurements to be accurate to within stated uncertainties.⁷ Consequently, meaningful traceability of displacement measurements using laser interferometer systems to the NBS must be concerned with the total accuracy of the length measurement and not with laser calibration alone.

TRACEABILITY: The National Bureau of Standards supports laser interferometer system manufacturers to ensure the traceability of length measurements by providing technical information and assistance. In particular, NBS has developed a portable Iodine Stabilized HeNe Laser (633 nm) to ensure that manufacturers have access to the length standard at an accuracy of 2.3 parts in 10^{10} . In addition, NBS provides technical support and information to interferometer users when they encounter unusual problems which cannot be solved by using commercial equipment or services.

Calibration of individual laser interferometer systems at NBS is insufficient to ensure that subsequent measurements made using that system in other environments will be correct. Unlike gage blocks and other simple artifacts, the operating characteristics of laser interferometer systems are known to change after calibration, especially when the systems are transported and are turned on and off frequently. The long term accuracy of laser interferometer systems is determined by the manufacturer based on experience gained by comparing their system with one of higher accuracy over extended periods of time. The laser wavelength accuracy and stability can be measured by comparison with an iodine stabilized laser and the index of refraction correction accuracy by comparison to laboratory grade temperature, pressure, relative humidity, and CO₂ measuring instruments. This information should be incorporated into the specifications provided with the instrument.

In consideration of the above information, NBS believes that the traceability of stabilized lasers and laser interferometer systems to the National Bureau of Standards must proceed from the user through the manufacturer to NBS. We know that the calibration of individual lasers or laser systems does not characterize their long term accuracy and, thus, this service is not provided by NBS.

-
1. NBS Special Publication 330 (1986).
 2. Comptes Rendus des Scéances de la 17e CGPM, 1983.
 3. Metrologia 19, 163-177 (1984).
 4. H.P. Layer, IEEE Trans. Instrum. and Meas., IM-39, 358 (1980).
 5. F.E. Jones, J. Res. Nat. Bur. Stand. (U.S), 86, 27-32 (1981).
 6. W.T. Estler, Appl. Opt., 24, 808-815 (1985).
 7. Hewlett-Packard Jol., 34, No. 4 (Apr. 1983).

U.S. DEPT. OF COMM. BIBLIOGRAPHIC DATA SHEET (See instructions)	1. PUBLICATION OR REPORT NO. NBS/TN-1248	2. Performing Organ. Report No.	3. Publication Date June 1988
4. TITLE AND SUBTITLE Traceability of Laser Interferometric Length Measurements			
5. AUTHOR(S) Howard P. Layer and W. Tyler Estler			
6. PERFORMING ORGANIZATION (If joint or other than NBS, see instructions) NATIONAL BUREAU OF STANDARDS U.S. DEPARTMENT OF COMMERCE GAITHERSBURG, MD 20899		7. Contract/Grant No.	8. Type of Report & Period Covered Final
9. SPONSORING ORGANIZATION NAME AND COMPLETE ADDRESS (Street, City, State, ZIP) SAME AS ITEM #6 ABOVE.			
10. SUPPLEMENTARY NOTES <input type="checkbox"/> Document describes a computer program; SF-185, FIPS Software Summary, is attached.			
11. ABSTRACT (A 200-word or less factual summary of most significant information. If document includes a significant bibliography or literature survey, mention it here) The National Bureau of Standards maintains the standard of length in accordance with the definition of the meter, adopted by the General Conference of Weights and Measures (1983). This standard is used within NBS to form a consistent system of units which are used in NBS calibration services.			
12. KEY WORDS (Six to twelve entries; alphabetical order; capitalize only proper names; and separate key words by semicolons) calibration; length; measurements; meter; standards; traceability			
13. AVAILABILITY <input checked="" type="checkbox"/> Unlimited <input type="checkbox"/> For Official Distribution. Do Not Release to NTIS <input type="checkbox"/> Order From Superintendent of Documents, U.S. Government Printing Office, Washington, D.C. 20402. <input checked="" type="checkbox"/> Order From National Technical Information Service (NTIS), Springfield, VA. 22161		14. NO. OF PRINTED PAGES 5	15. Price

NBS *Technical Publications*

Periodical

Journal of Research—The Journal of Research of the National Bureau of Standards reports NBS research and development in those disciplines of the physical and engineering sciences in which the Bureau is active. These include physics, chemistry, engineering, mathematics, and computer sciences. Papers cover a broad range of subjects, with major emphasis on measurement methodology and the basic technology underlying standardization. Also included from time to time are survey articles on topics closely related to the Bureau's technical and scientific programs. Issued six times a year.

Nonperiodicals

Monographs—Major contributions to the technical literature on various subjects related to the Bureau's scientific and technical activities.

Handbooks—Recommended codes of engineering and industrial practice (including safety codes) developed in cooperation with interested industries, professional organizations, and regulatory bodies.

Special Publications—Include proceedings of conferences sponsored by NBS, NBS annual reports, and other special publications appropriate to this grouping such as wall charts, pocket cards, and bibliographies.

Applied Mathematics Series—Mathematical tables, manuals, and studies of special interest to physicists, engineers, chemists, biologists, mathematicians, computer programmers, and others engaged in scientific and technical work.

National Standard Reference Data Series—Provides quantitative data on the physical and chemical properties of materials, compiled from the world's literature and critically evaluated. Developed under a worldwide program coordinated by NBS under the authority of the National Standard Data Act (Public Law 90-396).

NOTE: The Journal of Physical and Chemical Reference Data (JPCRD) is published quarterly for NBS by the American Chemical Society (ACS) and the American Institute of Physics (AIP). Subscriptions, reprints, and supplements are available from ACS, 1155 Sixteenth St., NW, Washington, DC 20056.

Building Science Series—Disseminates technical information developed at the Bureau on building materials, components, systems, and whole structures. The series presents research results, test methods, and performance criteria related to the structural and environmental functions and the durability and safety characteristics of building elements and systems.

Technical Notes—Studies or reports which are complete in themselves but restrictive in their treatment of a subject. Analogous to monographs but not so comprehensive in scope or definitive in treatment of the subject area. Often serve as a vehicle for final reports of work performed at NBS under the sponsorship of other government agencies.

Voluntary Product Standards—Developed under procedures published by the Department of Commerce in Part 10, Title 15, of the Code of Federal Regulations. The standards establish nationally recognized requirements for products, and provide all concerned interests with a basis for common understanding of the characteristics of the products. NBS administers this program as a supplement to the activities of the private sector standardizing organizations.

Consumer Information Series—Practical information, based on NBS research and experience, covering areas of interest to the consumer. Easily understandable language and illustrations provide useful background knowledge for shopping in today's technological marketplace.

Order the above NBS publications from: Superintendent of Documents, Government Printing Office, Washington, DC 20402.

Order the following NBS publications—FIPS and NBSIR's—from the National Technical Information Service, Springfield, VA 22161.

Federal Information Processing Standards Publications (FIPS PUB)—Publications in this series collectively constitute the Federal Information Processing Standards Register. The Register serves as the official source of information in the Federal Government regarding standards issued by NBS pursuant to the Federal Property and Administrative Services Act of 1949 as amended, Public Law 89-306 (79 Stat. 1127), and as implemented by Executive Order 11717 (38 FR 12315, dated May 11, 1973) and Part 6 of Title 15 CFR (Code of Federal Regulations).

NBS Interagency Reports (NBSIR)—A special series of interim or final reports on work performed by NBS for outside sponsors (both government and non-government). In general, initial distribution is handled by the sponsor; public distribution is by the National Technical Information Service, Springfield, VA 22161, in paper copy or microfiche form.

U.S. Department of Commerce
National Bureau of Standards
Gaithersburg, MD 20899

Official Business
Penalty for Private Use \$300



Stimulating America's Progress
1913-1988