

A11102 621959

NBS  
PUBLICATIONS

NAT'L INST OF STANDARDS & TECH R.I.C.



A1102621959

Camell, D. G/NBS calibration procedures  
QC100 .U5753 NO.1309 1987 V198 C.1 NBS-P



# NBS TECHNICAL NOTE 1309

U.S. DEPARTMENT OF COMMERCE/National Bureau of Standards

## NBS Calibration Procedures for Horizontal Dipole Antennas (25 to 1000 MHz)

D.G. Camell  
E.B. Larsen  
J.E. Cruz

QC

100

.U5753

No.1309

1987

c.2

**T**he National Bureau of Standards<sup>1</sup> was established by an act of Congress on March 3, 1901. The Bureau's overall goal is to strengthen and advance the nation's science and technology and facilitate their effective application for public benefit. To this end, the Bureau conducts research to assure international competitiveness and leadership of U.S. industry, science and technology. NBS work involves development and transfer of measurements, standards and related science and technology, in support of continually improving U.S. productivity, product quality and reliability, innovation and underlying science and engineering. The Bureau's technical work is performed by the National Measurement Laboratory, the National Engineering Laboratory, the Institute for Computer Sciences and Technology, and the Institute for Materials Science and Engineering.

### *The National Measurement Laboratory*

---

Provides the national system of physical and chemical measurement; coordinates the system with measurement systems of other nations and furnishes essential services leading to accurate and uniform physical and chemical measurement throughout the Nation's scientific community, industry, and commerce; provides advisory and research services to other Government agencies; conducts physical and chemical research; develops, produces, and distributes Standard Reference Materials; provides calibration services; and manages the National Standard Reference Data System. The Laboratory consists of the following centers:

- Basic Standards<sup>2</sup>
- Radiation Research
- Chemical Physics
- Analytical Chemistry

### *The National Engineering Laboratory*

---

Provides technology and technical services to the public and private sectors to address national needs and to solve national problems; conducts research in engineering and applied science in support of these efforts; builds and maintains competence in the necessary disciplines required to carry out this research and technical service; develops engineering data and measurement capabilities; provides engineering measurement traceability services; develops test methods and proposes engineering standards and code changes; develops and proposes new engineering practices; and develops and improves mechanisms to transfer results of its research to the ultimate user. The Laboratory consists of the following centers:

- Applied Mathematics
- Electronics and Electrical Engineering<sup>2</sup>
- Manufacturing Engineering
- Building Technology
- Fire Research
- Chemical Engineering<sup>3</sup>

### *The Institute for Computer Sciences and Technology*

---

Conducts research and provides scientific and technical services to aid Federal agencies in the selection, acquisition, application, and use of computer technology to improve effectiveness and economy in Government operations in accordance with Public Law 89-306 (40 U.S.C. 759), relevant Executive Orders, and other directives; carries out this mission by managing the Federal Information Processing Standards Program, developing Federal ADP standards guidelines, and managing Federal participation in ADP voluntary standardization activities; provides scientific and technological advisory services and assistance to Federal agencies; and provides the technical foundation for computer-related policies of the Federal Government. The Institute consists of the following divisions:

- Information Systems Engineering
- Systems and Software Technology
- Computer Security
- Systems and Network Architecture
- Advanced Computer Systems

### *The Institute for Materials Science and Engineering*

---

Conducts research and provides measurements, data, standards, reference materials, quantitative understanding and other technical information fundamental to the processing, structure, properties and performance of materials; addresses the scientific basis for new advanced materials technologies; plans research around cross-cutting scientific themes such as nondestructive evaluation and phase diagram development; oversees Bureau-wide technical programs in nuclear reactor radiation research and nondestructive evaluation; and broadly disseminates generic technical information resulting from its programs. The Institute consists of the following Divisions:

- Ceramics
- Fracture and Deformation<sup>3</sup>
- Polymers
- Metallurgy
- Reactor Radiation

---

<sup>1</sup>Headquarters and Laboratories at Gaithersburg, MD, unless otherwise noted; mailing address Gaithersburg, MD 20899.

<sup>2</sup>Some divisions within the center are located at Boulder, CO 80303.

<sup>3</sup>Located at Boulder, CO, with some elements at Gaithersburg, MD

# NBS Calibration Procedures for Horizontal Dipole Antennas (25 to 1000 MHz)

D.G. Camell  
E.B. Larsen  
J.E. Cruz

Electromagnetic Fields Division  
Center for Electronics and Electrical Engineering  
National Engineering Laboratory  
National Bureau of Standards  
Boulder, Colorado 80303-3328



---

U.S. DEPARTMENT OF COMMERCE, Malcolm Baldrige, Secretary  
NATIONAL BUREAU OF STANDARDS, Ernest Ambler, Director

Issued April 1987

National Bureau of Standards Technical Note 1309  
Natl. Bur. Stand. (U.S.), Tech Note 1309, 48 pages (Apr. 1987)  
CODEN:NBTNAE

U.S. GOVERNMENT PRINTING OFFICE  
WASHINGTON: 1987

## CONTENTS

	Page
List of Tables . . . . .	v
List of Figures . . . . .	v
Abstract . . . . .	1
1. Introduction . . . . .	1
2. Standard Receiving Antenna Method . . . . .	3
2.1 Description of the Equipment . . . . .	3
2.2 Test Procedure . . . . .	4
2.3 Theoretical Basis of the Test . . . . .	7
2.4 Estimate of Calibration Uncertainty . . . . .	9
3. Standard Field Method . . . . .	15
3.1 Description of the Equipment . . . . .	15
3.2 Test Procedure . . . . .	16
3.3 Theoretical Basis of the Test . . . . .	18
3.3.1 Gain of Open-Ended Guides, 200 to 500 MHz. . . . .	19
3.3.2 Gain of Pyramidal Horns, 450 MHz to 1 GHz. . . . .	20
3.3.3 Net Power. . . . .	21
3.4 Estimate of Calibration Uncertainty . . . . .	22
4. References . . . . .	29
5. Appendix A: Sample Test Report for Standard Antenna Method . . .	33
6. Appendix B: Sample Test Report for Standard Field Method . . .	37





## List of Tables

	Page
1. List of equipment used for the open-field site test. . . . .	10
2. Pyramidal horns and OEGs for the anechoic chamber. . . . .	23
3. List of equipment for the NBS anechoic chamber . . . . .	23

## List of Figures

	Page
1. Field site instrumentation for calibrating horizontal dipoles, at frequencies of 25 to 1000 MHz . . . . .	11
2. Standard receiving dipole mount with diode detector, filter, and resistive transmission line . . . . .	12
3. Instrumentation for measuring $V_{oc}$ versus $V_{dc}$ of the NBS standard dipole . . . . .	13
4. Typical calibration curve of a standard dipole mount, showing the dc voltage versus rf voltage relation . . . . .	13
5. Sample data sheet for dipole measurements at the NBS open field site . . . . .	14
6. Side view of the NBS anechoic chamber . . . . .	24
7. Geometry of open-ended rectangular waveguide . . . . .	25
8. Sketch of a pyramidal horn showing the dimensions . . . . .	25
9. Instrumentation for generating a standard field in an anechoic chamber . . . . .	26
10. Automated system for the NBS anechoic chamber . . . . .	27
11. Graph of WR-2100 OEG gain versus distance at 500 MHz . . . . .	28
12. Graph of SA 12-0.5 horn gain versus distance at 500 MHz . . . . .	28





NBS Calibration Procedures for Horizontal Dipole Antennas  
(25 to 1000 MHz)

D. G. Camell, E. B. Larsen, and J. E. Cruz

Electromagnetic Fields Division  
National Bureau of Standards  
Boulder, Colorado 80303

This publication describes the theoretical basis and test procedures for horizontally polarized dipole calibrations at the National Bureau of Standards. Two different techniques and two different test sites are used. The standard antenna method uses the calculation of a field strength level, from the response of a simple half-wave dipole, to calibrate an antenna. This method is used at an open field site in the frequency range of 25 to 1000 MHz. The standard field method applies the theoretical gain equations of waveguides to determine the field strength level. This latter method is used in an anechoic chamber in the frequency range of 200 to 1000 MHz. Procedures for both techniques are explained and measurement setups are illustrated. Measurement uncertainties are discussed. Sample reports are included for both methods.

Keywords: anechoic chamber, calibration, dipole antenna, electromagnetic field, open field site, standard antenna, standard field.

## 1. Introduction

The purpose of this Technical Note is to provide documentation for the methods, equipment, and facilities employed by the National Bureau of Standards (NBS) for the calibration of horizontally polarized dipole antennas over the frequency range of 25 to 1000 MHz. The facilities are located in the Dept. of Commerce, Boulder Laboratories, Boulder, Colorado. The references cited are in the open literature and are available from such sources as libraries maintaining collections of scientific and technical

books and archival journals, and from the National Technical Information Service, which maintains archival collections of U. S. Government publications. Thus the material presented in this Technical Note in conjunction with the cited references provides the information necessary for the replication of the calibration procedures.

Two methods are used at NBS for horizontally polarized dipole calibrations. The standard receiving antenna method is used to calibrate continuous wave (CW) field strength levels with half-wave dipole receiving antennas at the NBS open-field site. This measurement site is used over a frequency range of 25 to 1000 MHz. The standard field method is used to establish reference CW field strength levels in an anechoic chamber. The anechoic chamber can be used over the frequency range of 200 to 1000 MHz.

At the NBS open-field site, two sets of standard half-wave dipole antennas are used to measure the CW electric field strength. The field strength is calculated from the open-circuit induced voltage in the standard antenna, the transmission frequency, and the dimensions of the standard receiving dipole.

In the NBS anechoic chamber, a series of standard-gain open-ended guides (OEGs) and pyramidal horns is used for transmitting known CW electromagnetic (EM) fields. The field level is calculated from the gain of the transmitting antenna, net power delivered to the transmitting antenna, the distance from the transmitting antenna to the field point, and the effect of chamber reflections. Most of the test equipment used is commercially available. A computer-aided system can be used to automate the data collection process.

This paper presents a description of the equipment, the test method, theoretical basis of the test, and estimated errors for these standard-antenna and standard-field calibrations.

## 2. Standard Receiving Antenna Method [1-11]

### 2.1 Description of the Equipment

The NBS open-field site has two facilities for calibrating dipoles. One of these is located on a flat area of unimproved rocky soil adjacent to the wood building used for calibrating loop antennas. The other facility is designed around a grounded flat screen mesh on a slab of concrete 30 m wide and 60 m long. A tunnel under this slab leads to an underground room, approximately 3 m by 3 m, where the rf source and receiving equipment are located. Several small (15 cm diameter) tunnels leading from this room to various locations on the ground screen convey rf transmission cables and ac power cords to devices operating on or above the ground screen. The cables, being underground, do not interfere with the EM fields. The transmitting antennas are (1) horizontally polarized half-wave dipole antennas for the frequency range of 25 to 80 MHz, and (2) a horizontally polarized log-periodic antenna for frequencies from 80 to 1000 MHz. A wooden ladder with a maximum height of 6 m is used to mount the standard receiving dipole or the antenna under test (AUT) above the ground screen. Figure 1 shows the open-field site with the common locations of the transmitting and receiving antennas.

The NBS standard receiving antennas are self-resonant, half-wave dipoles with a high-impedance balanced voltmeter built across the center gap of the dipole antenna. The filtered output of the shunt-diode detector is a dc voltage that is measured by a high-impedance dc voltmeter. A photo of one of the antenna mounts (frequency range of 25 to 300 MHz) is given in figure 2. The Schottky semiconductor diode used has a low turn-on voltage and a high peak-reverse voltage rating of 70 V [5]. This type of diode provides a very high shunt impedance ( $>100\text{ M}\Omega$ ) for the signal levels used in NBS calibrations. The filter consists of four resistors and two capacitors.

The high-resistance lines below the filter in figure 2 are made of carbon-impregnated plastic, coated with a nylon jacket. Each line has a resistance of about 600  $\Omega$ /cm.

The transmitted signal is produced by a combination of signal generators, amplifiers, low-pass filters, low-loss transmission lines, and a transmitting antenna. Figure 1 shows the equipment used for a measurement test at the open field site. The rf voltmeter is used to monitor the level of the transmitter voltage. A generic list of equipment used on the ground screen is given in table 1. The cables are 50  $\Omega$  coaxial with type N connectors.

The open-circuit rf voltage of the standard dipole can be determined if the relation of input rf voltage to output dc voltage is known. The rf-to-dc voltage curve of a standard dipole antenna is obtained by injecting known rf voltages across the dipole mount at 50 MHz and measuring the detected dc responses from the output of the mount. The response of each dipole mount is essentially independent of frequency over the frequency range used. The frequency of 50 MHz was chosen because it is high enough to give the correct calibration, yet low enough to avoid error caused by transformation of voltage in the short "transmission line" between the modified tee connector and the dipole mount (see figure 3). Also, 50 MHz is the frequency of the internal calibrating source used in the 50  $\Omega$  power meter shown for measuring rf voltage at the tee connector. Figure 3 shows the test setup for this measurement. The equation for the curve of rf input voltage to dc output voltage is used to calculate the open circuit rf voltage across the center gap of the standard dipole. A typical curve is given in figure 4.

## 2.2 Test Procedure

NBS tests dipoles and other field strength measuring devices at specified frequencies, field strength levels, heights above ground, and distances between transmitting and receiving antennas. When a device to be



tested arrives, the serial numbers are logged, a test number is assigned, and a test folder is issued. The unit is inspected to see if it is operating correctly. Before the actual testing is done, a review of the anticipated data and the procedure details are decided upon. The test data are usually reported in terms of the antenna factor or antenna output, such as detected dc voltage.

An EM field is generated by a transmitting antenna at the open field site. An NBS standard dipole antenna is used as the receiving antenna, and the detected dc voltage is measured with a high impedance (100 M $\Omega$ ) voltmeter. The transmitter voltage level is monitored with an rf voltmeter to ensure that it does not change during the test. The AUT is then substituted at the same position as the NBS standard dipole. The rf output of the AUT is measured with a calibrated 50  $\Omega$  spectrum analyzer or 50  $\Omega$  power meter. Data for the rf-to-dc curve of the standard dipole mount are measured before and after each test. The insertion loss of the cables used is also measured before and after each test to insure that the cable and connector losses are repeatable. The antenna factor is determined from these measurements, as described in section 2.3.

The transmitting antenna is mounted on a tripod and the height set to a specified value, usually 3 m. The vertical wooden ladder for supporting the receiving antennas is mounted directly over the underground room. If the distance between the transmitting and receiving antennas is not specified, the distance is usually set between 15 and 20 m. In order to produce a uniform wavefront at the receiving antenna, this separation distance should be at least two wavelengths at the operating frequency. If the antenna height above ground is not specified, the height of both receiving antennas (standard dipole and AUT) is set to 3 m. This height is usually sufficient to reduce calibration errors caused by impedance changes in the AUT, which are due to the proximity of the natural ground or conducting ground screen beneath the antenna [2-3]. If the AUT is adjustable in length, its length

is set to the value specified by the customer (usually an electrical half wavelength).

The equipment used is connected as shown in figure 1. The power meters with their appropriate sensors are calibrated by the NBS power calibration services. The spectrum analyzer is calibrated by the NBS microwave calibration service (test 61190S)[40], and verified before each test by means of calibrated power meters.

All of the serial numbers on the AUT and equipment used are recorded, and a sketch is made of the calibration setup. If the test requires an unusual arrangement, a photograph of the system is taken. Header titles are written for each column of data to be taken (along with units), and any other pertinent information is noted on the data sheet. A sample data sheet is given in figure 5.

The NBS standard dipoles used to measure the incident field ( $E_{inc}$ ) lack frequency selectivity, so it is not possible to perform a calibration in the presence of strong interfering signals. Therefore the detected ambient signal level is recorded. That is, the level of the dc voltage output of the standard dipole with no NBS test signal applied is noted on the data sheet. This level is usually less than 5 mV. The range of dc output voltages used at NBS for field strength calibrations is normally 0.5 to 1.5 V, which is sufficiently high to avoid calibration uncertainty caused by ambient signals and noise or by changes in ambient temperature. The rf-to-dc calibration of the dipole mount depends slightly on the diode temperature, especially at low signal levels. This uncertainty is negligible for detected voltages greater than 0.5 V, if the temperature is within  $\pm 10^{\circ}\text{C}$  of that for which the diode was calibrated.



### 2.3 Theoretical Basis of the Test

An NBS standard half-wave dipole is used to determine the calibrating field strength level from the relation

$$E_{inc} = \frac{V_{oc}}{L_{eff}} \quad (1)$$

where  $E_{inc}$  = electric field strength of the locally generated field, V/m,  
 $V_{oc}$  = open-circuit rf voltage induced in the standard dipole, V, and  
 $L_{eff}$  = effective length of the standard dipole, m.

The theoretical effective length of an infinitesimally thin half-wave dipole in free space is given by

$$L_{eff} = \lambda / \pi \quad (2)$$

where  $\lambda$  = free space wavelength, m.

The NBS standard dipoles have a cylindrical shape. An approximate solution for their effective length and required shortening for self resonance was derived by Schelkunoff [1]. The required length depends on the dipole length-to-diameter ratio, and is given approximately by

$$\text{Required } L \approx \left(\frac{\lambda}{2}\right) \left[1 - \frac{0.2257}{\ln(\lambda/D) - 1}\right] \quad (3)$$

where Required  $L$  = length of the cylindrical standard dipole  
required for self resonance, m, and  
 $D$  = diameter of the standard dipole, m.

The NBS standard dipoles consist of a center mount, as shown in figure 2, to which extension rods are attached. One pair of rods is available to

achieve a self-resonant length at each calibration frequency [2-4]. The effective length of a thin cylindrical dipole having a length near self resonance is given by

$$L_{\text{eff}} = \left(\frac{\lambda}{\pi}\right) \tan (\pi L/2\lambda) \quad (4)$$

where  $L_{\text{eff}}$  = effective length of the dipole, m, and  
 $L$  = measured tip-to-tip length of the dipole, m.

The open-circuit rf voltage of the NBS standard dipole is determined from the measured dc voltage by applying the equation

$$V_{\text{oc}} = M(V_{\text{det}}) + B \quad (5)$$

where  $V_{\text{oc}}$  = open-circuit rf voltage across the dipole center gap, V,  
 $M$  = slope of the rf input to dc output curve, measured over the voltage range of 0.5 to 2 V (see figure 4),  
 $V_{\text{det}}$  = detected dc voltage of the standard dipole, V, and  
 $B$  = y intercept of the rf-to-dc calibration curve.

The value of the calibrating electric field,  $E_{\text{inc}}$  is now obtained from eq (1) using the solutions of eqs (4) and (5). Once the electric field strength is known, the antenna factor of the AUT is calculated from the expression

$$K = (E_{\text{inc}}) - (V_{50\Omega}) - A \quad (6)$$

where  $K$  = antenna factor of the AUT, dB,  
 $E_{\text{inc}}$  = calibrating field strength level, dBmV/m,  
 $V_{50\Omega}$  = AUT voltage pickup, dBmV, and  
 $A$  = cable attenuation between the AUT and the 50  $\Omega$  receiver, dB.

## 2.4 Estimate of Calibration Uncertainty [6,7]

For calibrations at the NBS open field site, at frequencies of 25 to 1000 MHz, the following statements give the possible errors of a measurement

- (1) One source of error in the calibration is uncertainty of the calibrating field value at the open field site. This is due mainly to uncertainty in the dc-to-rf voltage transfer. Also, signals from radio stations and 2-way radio communications may cause extra errors. The estimated error is  $\pm 0.4$  dB.
- (2) Another source of error is uncertainty in the disturbance of the field caused by interconnecting cables and nearby buildings. This error is estimated to be less than  $\pm 0.2$  dB.
- (3) Other sources of uncertainty are associated with antenna alignment, measurement of antenna separation distance, and NBS calibrations of the various instruments used. These include spectrum analyzers, rf voltmeters, and power monitors. The error due to these last sources is estimated to be less than  $\pm 0.4$  dB.

The overall worst-case uncertainty of an antenna calibration is the simple sum of those listed, or  $\pm 1$  dB.

Table 1. List of equipment used for the open-field site test.

- synthesized signal generator, +3 dBm maximum output, 10 kHz-1000 MHz
- amplifier, 1-1000 MHz, 50 W output
- amplifier, 1-1000 MHz, 3 W output
- ac voltmeter, with coaxial tee connector
- dc voltmeter, high input impedance (100 M $\Omega$ )
- calibrated power meter, single sensor capability
- power sensors: 100 mW capacity
- calibrated spectrum analyzer, covering the frequency range of 25 to 1000 MHz
- assorted filters, attenuator pads, and low loss cables with type N connectors.




Figure 1. Field site instrumentation for calibrating horizontal dipoles, at frequencies of 25 to 1000 MHz.




Figure 2. Standard receiving dipole mount with diode detector, filter, and resistive transmission line.






Figure 3. Instrumentation for measuring  $V_{oc}$  versus  $V_{dc}$  of the NBS standard dipole.




Figure 4. Typical calibration curve of a standard dipole mount, showing the dc voltage versus rf voltage relation.



### 3. Standard Field Method [7-39]

#### 3.1 Description of the Equipment

The NBS anechoic chamber is constructed within a metal shielded room 5 m high, 6.7 m wide, and 8.5 m long. The chamber is lined with absorbing material on all of the inside walls, ceiling, and floor. An equipment cart placed on a precision track is movable over the entire length of the chamber. The cart is 1.9 m wide by 0.75 m deep, with pieces of absorber mounted on the front side. The cart is driven by a stepping motor that steps in minimum increments of 2 mm. A rotator/positioner can be placed on the cart for axial rotation and positioning of the AUT. A doorway 1.2 m wide, in line with the track, provides access into the chamber. Figure 6 shows a side view of the anechoic chamber, cart, and doorway.

Fiber optic cables are connected from the equipment cart through the chamber doorway to a computer in the laboratory. The fiber optic lines are essentially transparent to the rf energy and provide negligible interference with the generated fields. On the cart is a single-board computer which is used to process and transfer information through the fiber optic lines. This information consists of data from the AUT, instructions for movement of the cart, and instructions for turning the rotator.

A pyramidal horn or rectangular OEG is used as a transmitting antenna, and positioned near the chamber entry region on the measurement axis. A series of two pyramidal horns and two OEGs is available to establish known EM fields in the chamber. These horns and OEGs cover the frequency range of 200 to 1000 MHz. A sketch of an OEG is given in figure 7 and a sketch of a horn is given in figure 8. Table 2 lists each horn or OEG with its frequency range and dimensions.

The transmitted signal is produced from a combination of signal generators, amplifiers, low-pass filters, and low-loss transmission lines.

Figure 9 shows the equipment used for a measurement test in the anechoic chamber. Measurement of the net transmitting power is done with a dual directional coupler and two power meters with thermistor sensors. When measurements of high power are done, calibrated attenuators can be inserted in the coupler secondary ports to assure the power level is within the sensors' range. A generic list of equipment used in the anechoic chamber is given in table 3. The cables are 50  $\Omega$  coaxial with type N connectors.

### 3.2 Test Procedure

For anechoic chamber measurements, the AUT is mounted on the movable cart and aligned along the center axis of the transmitting antenna. Dielectric supports are used to mount the AUT in line with the axis of the transmitting antenna. A laser alignment system located outside the chamber is used to align the AUT and transmitting antenna on the measurement axis. If a rotator/positioner is used, it is placed on the cart and the AUT is mounted on the rotator and aligned on axis of the transmitting antenna.

The following steps are used to align the transmitting horn or OEG with the AUT

- (1) The laser system is leveled and placed on the measurement axis.
- (2) The AUT is placed on the cart and the laser is used to align the AUT on axis.
- (3) The appropriate transmitting horn or OEG is placed on a cart outside of the chamber doorway.
- (4) The transmitting antenna is aligned using the laser, with the AUT positioned in the chamber.

The equipment for each frequency is connected as shown in figure 9. The power meters and appropriate sensors are calibrated units, having been calibrated by the NBS microwave calibration service (test 61190S)[40]. All



of the serial numbers on the AUT and the equipment used are recorded and a sketch is made of the setup. If the test is an uncommon arrangement, photographs are taken.

There are two techniques available to test the AUT. One is to set the field strength at a prescribed level and then record the response of the AUT. The output of the AUT is recorded as a function of applied field strength. Most of the tests performed at NBS are done using this first technique. The other technique is to adjust the field level until the AUT is at the desired response and then record the field level necessary to achieve that response. The level of the standard EM field necessary for this indication is calculated with the help of a computer in terms of the known gain of the transmitting horn and the net delivered power, as described in section 3.3. To determine the value of the net power, the incident and reflected power are measured at the secondary ports of the dual directional coupler. The calibrated values of the coupler losses are also used to calculate the net power, as described in section 3.3.3.

The measurement data can be collected and processed automatically. Figure 10 shows the NBS system that is under computer control. Frequency and field level are selected as needed for the test. The computer controls the movement of the cart and rotation of the rotator. The amplifiers and filters are selected by a series of rf switches. The IEEE-488 bus is used for most of the data and command transfer. The computer is also used to print out the data and all pertinent information. The amount of automation used depends on the amount of data and the format in which the data is requested. If only a few data points are needed, the manual method may be more efficient.

The manual method of collecting data uses the computer to calculate the power necessary to create a desired field strength level in the anechoic chamber at the AUT, as described in section 3.3. The operator connects the

system as shown in figure 9. Power is applied at the test frequency as calculated by the computer. The response of the AUT is recorded from the appropriate metering unit. The metering unit may be a dc voltmeter, spectrum analyzer, power meter, etc. depending on the output of the AUT.

On completion of the test, the data are compared with theoretical computations, the data history of the AUT (if any), and/or the data history of the anechoic chamber. If the data do not agree with anticipated results, the AUT and/or the NBS measurement system are checked. If the AUT is suspected of being defective, arrangements are made to check and repair it, if possible. Most data are plotted with graphic routines available on the computer system. A sample of an NBS test report is given in appendix B. Different report formats are written for each type of calibration depending on the requested type of measurement.

### 3.3 Theoretical Basis of the Test

The standard field method used at NBS in the anechoic chamber involves calculating the radiated intensity in the near zone of standard gain antennas. This method consists of generating and computing the desired component of a field in terms of the dimensions and shape of a transmitting antenna, net delivered power, the distance from the transmitting antenna to the AUT, and the effect of chamber reflections. A set of two rectangular OEGs is available to cover the 200 to 500 MHz frequency range, while a set of two rectangular pyramidal horns is used from 450 to 1000 MHz.

The on-axis field intensity is calculated in terms of the net power delivered to the transmitting antenna and the calibrated gain of the pyramidal horn or OEG "launcher". The field intensity in the center of the beam is calculated, at each frequency, using power equation techniques. All of the instrumentation is readily available commercially except for the



large OEG launchers. These OEGs were made of sheet metal to the dimensions specified in table 2. The value of standard field is given by [32]

$$E = \frac{\sqrt{30 PG}}{d} \quad (7)$$

where  $E$  = on-axis magnitude of the radiated field, V/m,  
 $P$  = net power delivered to the transmitting horn or OEG, W,  
 $G$  = calibrated gain of the transmitting antenna, including appropriate near-zone correction factors, and  
 $d$  = distance from the horn or OEG aperture to the calibrating field point (center of dipole), m.

The effects of multipath reflections in the anechoic chamber have been analyzed and taken into account for each setup and frequency [32,34].

### 3.3.1 Gain of Open-Ended Guides, 200 to 500 MHz

Early work to determine the field pattern and gain of OEG radiators, both theoretically and experimentally, is described in [6], and later derivations are given in [20]. An equation giving the gain of an open-ended waveguide as a function of frequency and aperture dimensions has been determined experimentally at NBS. The original data for this equation came from a two-antenna calibration technique using two "identical" open-ended guides [13]. Further calibrations have been made with the two specially-fabricated OEGs, each having a length of about 2 m. All of the OEGs used at NBS have a 2-to-1 aspect ratio. In this case, the equation for calculating the antenna gain is [32]

$$\text{GAIN} = 21.6 Fw, \text{ or} \quad (8a)$$

$$\text{GAIN, dB} = 10 \log (Fw) + 13.34, \quad (8b)$$

where  $F$  = frequency, GHz, and  
 $w$  = width (larger side) of the 2-to-1 OEG, m.

Equation (8) is accurate to  $\pm 0.5$  dB if the distance,  $d$ , from the OEG aperture to the field point is greater than double the width,  $w$ . Figure 7 shows the dimensions used in the equation. Figure 11 is a plot of the theoretical gain from eq (10) and the measured gain values in the anechoic chamber as a function of distance from the OEG aperture.

### 3.3.2 Gain of Pyramidal Horns, 450 MHz to 1 GHz

Standard fields above 450 MHz are produced in the anechoic chamber by a series of standard-gain pyramidal horns. A complication known as near-zone gain reduction applies to the calculation of field strength when the AUT is very close to a transmitting antenna. Simple polynomial expressions have been derived at NBS to determine the near-zone gain-reduction factors,  $R_H$  and  $R_E$ , for pyramidal horns [29,32]. The pertinent horn dimensions used in the equations are shown in figure 8. The procedure involves a computation of the intensity produced by an in-phase aperture and then applying two near-zone correction factors. The values of these gain-reduction factors depend on frequency, horn dimensions, and distance to the on-axis field point. The two gain-reduction factors,  $R_H$  and  $R_E$ , are given by

$$R_H = (0.01\alpha) (1 + 10.19\alpha + 0.51\alpha^2 - 0.097\alpha^3), \quad (9a)$$

$$R_E = (0.1\beta^2) (2.31 + 0.053\beta), \quad (9b)$$

$$\text{where } \alpha = \left( \frac{a^2 F}{0.3} \right) \left( \frac{1}{l_H} + \frac{1}{d} \right) \text{ and } \beta = \left( \frac{b^2 F}{0.3} \right) \left( \frac{1}{l_E} + \frac{1}{d} \right),$$

$a, b, l_H$  and  $l_E$  = horn dimensions of table 2, m,

$F$  = frequency, GHz, and

$d$  = distance from the horn aperture to the field point, m.

The theoretical gain of the horn, near zone or far zone, is given by

$$\text{GAIN} = (113.3 \text{ a b F}^2) [10^{(-R_H - R_E)}], \text{ or} \quad (10a)$$

$$\text{GAIN, dB} = 10 \log (ab) + 20 \log F + 20.54 - R_H - R_E. \quad (10b)$$

These equations have been checked experimentally using several different standard-gain horns covering the frequency range of 450 MHz to 4 GHz. The horns were also calibrated at NBS by the three-antenna method [18,23,26,28,30,39] and these gains were compared with that given by eq (10). For distances greater than 0.5 m, the difference between the experimentally calibrated gain and the value calculated from eq (10b) was less than 0.5 dB. A plot of horn gain vs. distance is shown in figure 12, comparing the theoretical curve with measurements made in the NBS anechoic chamber.

### 3.3.3 Net Power

The calculation of the net power delivered to the pyramidal horn or OEG depends on the values of incident power and reflected power measurements in the dual-directional coupler secondary ports. This net power is given by

$$P_{\text{net}} = P_{\text{inc}} [10^{C_i/10}] - P_{\text{refl}} [10^{C_r/10}] \quad (11)$$

where  $P_{\text{net}}$  = net power delivered to the transmitting horn or OEG, W,  
 $P_{\text{inc}}$  = power measured in the incident arm of the coupler, W,  
 $C_i$  = coupling ratio from the output port to the incident arm of the coupler, dB,  
 $P_{\text{refl}}$  = power measured in the reflected arm of the coupler, W, and  
 $C_r$  = coupling ratio from the output port to the reflected arm of the coupler, dB.

### 3.4 Estimate of Calibration Uncertainty

For frequencies from 200 to 1000 MHz, horizontally polarized dipoles can be calibrated in the NBS anechoic chamber. The following statements give the possible errors of a measurement

- (1) The largest source of error is caused by uncertainty in the near-zone gain value of the transmitting antenna. The estimated possible error is  $\pm 0.5$  dB.
- (2) Another source of error is uncertainty in the magnitude of multi-path reflections within the anechoic chamber. Possible errors caused by chamber reflections are estimated to be less than  $\pm 0.2$  dB.
- (3) Other sources of uncertainty are associated with antenna alignment, measurement of antenna separation distance, and NBS calibrations of the various instruments used. These include directional couplers, and incident and reflected power monitors. The overall error due to these latter sources is estimated to be less than  $\pm 0.3$  dB.

The overall worst-case uncertainty of the measurement is the simple sum of those listed, or  $\pm 1$  dB.

Table 2. Pyramidal horns and OEGs for the anechoic chamber  
(200 to 1000 MHz)

ANTENNA	FREQUENCY RANGE (MHz)	a (m)	b (m)	$l_H$ (m)	$l_E$ (m)
WR-3600 OEG	180 to 310	0.9144	0.4572	NA	NA
WR-2100 OEG	310 to 530	0.5334	0.2567	NA	NA
SA 12-0.5 #48	450 to 750	1.2250	0.9075	1.420	1.213
SA 12-0.75 #57	700 to 1100	0.8278	0.6118	0.943	0.812

Table 3. List of equipment for the NBS anechoic chamber.

- computer with disk drives
- synthesized signal generator, +3 dBm maximum output, 100-1000 MHz
- amplifier, 1-1000 MHz, 50 W output
- amplifier, 1-1000 MHz, 3 W output
- dual directional coupler, 0.1-2 GHz, 50 W capacity
- calibrated power meter, single sensor capability
- calibrated power meter, dual sensor capability
- power sensors: 10  $\mu$ W, 100 mW, 3 W, maximum capacity
- calibrated spectrum analyzer, covering the frequency range of 200 to 1000 MHz
- assorted filters, attenuators, and type N cables




Figure 6. Side view of the NBS anechoic chamber.






Figure 7. Geometry of open-ended rectangular waveguide.




Figure 8. Sketch of a pyramidal horn showing the dimensions.




Figure 9. Instrumentation for generating a standard field in an anechoic chamber.




Figure 10. Automated system for the NBS anechoic chamber.




Figure 11. Graph of WR-2100 OEG gain versus distance at 500 MHz.




Figure 12. Graph of SA 12-0.5 horn gain versus distance at 500 MHz.



#### 4. References

- [1] Schelkunoff, S.A.; Friis, H.T. Antennas, theory and practice. Wiley; 306 p; 1952.
- [2] Greene, F.M. NBS field-strength standards and measurements (30 Hz to 1000 MHz). Proc. IEEE 55(6); 970-981; 1967 June.
- [3] Taggart, H.E.; Workman, J.L. Calibration principles for field strength meters (30 Hz to 1 GHz). Nat. Bur. Stand. (U.S.) Tech. Note 370; 151 p; 1969 March.
- [4] IEEE standard methods for measuring electromagnetic field strength for frequencies below 1000 MHz in radio wave propagation. IEEE no. 302; 15 p; 1969 August.
- [5] Anand, Y. X-band high burnout silicon Schottky barrier diodes. Microwave Journal; 56-61; 1979 March.
- [6] Stubenrauch, C.F. International comparison of electric-field strength at 100 MHz. IEEE Trans. Instr. and Meas.; IM-32(1); 235-237; 1983 March.
- [7] Nahman, N.S.; Kanda, M.; Larsen, E.B.; Crawford, M.L. Methodology for standard electromagnetic field measurements. IEEE Trans. Instr. and Meas.; IM-34(4); 490-503; 1985 December.
- [8] Kanda, M.; Larsen, E.B.; Borsero, M.; Galliano, P.G.; Yokoshima, I.; Nahman, N.S. Standards for electromagnetic field measurements. Proc. IEEE 74(1); 120-128; 1986 January.
- [9] Larsen, E.B. Calibration and meaning of antenna factor and gain for EMI antennas. ITEM 1986; Interference Technology Engineers Master; R & B Enterprises; 384 p; 1986.
- [10] FitzGerrell, R.G. Linear gain-standard antennas below 1000 MHz. Nat. Bur. Stand. (U.S.) Tech. Note 1098; 39 p; 1986 May.
- [11] Ma, M.T.; Kanda, M. Electromagnetic compatibility and interference metrology. Nat. Bur. Stand. (U.S.) Tech Note 1099; 176 p; 1986 July.
- [12] Barrow, W.L.; Greene, F.M. Rectangular hollow-pipe radiators. Proc. IRE, 26(12); 1498-1519; 1938 December.
- [13] Slayton, W.T. Design and calibration of microwave antenna gain standards. Naval Research Lab., Wash., DC; NRL Report 4433; 39 p; 1954 November 9.



- [14] Chu, T.S.; Semplak, R.A. Gain of electromagnetic horns. Bell Syst. Tech. J.; 527-537; 1965 March.
- [15] Beatty, R.W. Discussion of errors in gain measurements of standard electromagnetic horns. Nat. Bur. Stand. (U.S.) Tech. Note 351; 45 p; 1967 March.
- [16] Rumfelt, A.Y.; Elwell, L.B. Radio frequency power measurements. Proc. IEEE 55(6); 837-850; 1967 June.
- [17] Bowman, R.R. Field strength above 1 GHz : measurement procedures for standard antennas. Proc. IEEE 55(6); 981-990; 1967 June.
- [18] Bowman, R.R. Absolute gain measurements for horn antennas (2nd edition). Rome Air Dev. Center Tech. Report RADC-TR-68-349; Griffiss AFB, NY; 83 p; 1968 November.
- [19] Bowman, R.R. Transmitting and receiving formulas for antennas and field intensity meters (2nd edition). Rome Air Dev. Center Tech. Report RADC-TR-68-445; Griffiss AFB, NY; 48 p; 1969 April.
- [20] Woods, D. Standard intensity electromagnetic field installation for calibration of radiation hazard monitors from 400 MHz to 40 GHz. Non-ionizing Radiation; 1(1); 9-17; 1969 June.
- [21] Edited by Hollis, J.S.; Lyon, T.J.; Clayton, L., Jr. Microwave antenna measurements; Scientific Atlanta, Inc.; Chapter 8; "Measurement of Gain"; 8-1 to 8-29; 1970 July.
- [22] Jull, E.V. Finite-range gain of sectoral and pyramidal horns. Electronics Letters; 6(21); 680-681; 1970 October 15.
- [23] Newell, A.C.; Baird, R.C.; Wacker, P.F. Accurate measurement of antenna gain and polarization at reduced distances by an extrapolation technique. IEEE Trans. Ant. & Prop. AP-21; 418-431; 1973 July.
- [24] Bowman, R.R. Calibration techniques for electromagnetic hazard meters: 500 MHz to 20 GHz. Nat. Bur. Stand. (U.S.) NBSIR 75-805; 37 p; 1976 April.
- [25] Edited by Love, A.W. Electromagnetic horn antennas. IEEE Press; New York, NY; 1976.
- [26] Bassen, H.I.; Herman, W.A. Precise calibration of plane-wave microwave power density using power equation techniques. IEEE Trans. Mic. Theory & Tech. MTT-25 (8); 701-706; 1977 August.

- [27] EMC antennas and antenna factors : how to use them. Soc. of Automotive Engineers, Inc. AIR 1509; Prepared by SAE Committee AE-4; 22 p; 1978 January.
- [28] Kummer, W.H.; Gillespie, E.S. Antenna measurements-1978. Proc. IEEE 66(4); 483-507; 1978 January.
- [29] Larsen, E.B. Techniques for producing standard EM fields from 10 kHz to 10 GHz for evaluating radiation monitors; Proc. Symp. Electromag. Fields in Biolog. Systems; Ottawa, Canada; 96-112; 1978 June.
- [30] IEEE standard test procedures for antennas. ANSI/IEEE Std. 149-1979; 153 p; 1979.
- [31] Hunter, J.D. Final report on the international comparison of power density at 2.45 GHz (72-11). CIPM Consultative Committee on Electricity, Radio Frequency Working Group; 7 p; 1980 October.
- [32] Larsen, E.B.; Ries, F.X. Design and calibration of the isotropic electric-field monitor (EFM-5), 0.2 to 1000 MHz. Nat. Bur. Stand. (U.S.) Tech. Note 1033; 103 p; 1981 March.
- [33] American national standard, recommended practice for the measurement of hazardous electromagnetic fields-rf and microwave. ANSI C95.5-1981; Amer. Nat. Stand. Inst.; Inst. of Elec. & Electronics Engrs., Inc.; 35 p; 1981.
- [34] FitzGerrell, R.G. Using free-space transmission loss for evaluating anechoic chamber performance. IEEE Trans. EMC EMC-24(3); 356-358; 1982 August.
- [35] Yaghjian, A.D. Approximate formulas for the far fields and gain of open-ended rectangular waveguide. Nat. Bur. Stand. (U.S.) NBSIR 83-1689; 37 p; 1983 May.
- [36] Ma, M.T.; Kanda, M.; Crawford, M.L.; Larsen, E.B. A review of electromagnetic compatibility/interference measurement methodologies. Proc. IEEE 73(3); 388-411; 1985 March.
- [37] Kanda, M.; Orr, R.D. A radio-frequency power delivery system: procedures for error analysis and self-calibration. Nat. Bur. Stand. (U.S.) Tech. Note 1083; 24 p; 1985 August.

- [38] Clark, R.F. The microcalorimeter as a national microwave power standard. Proc IEEE 74(1); 102-104; 1986 January.
- [39] Kanda, M.; Orr, R.D. Near-field gain of a horn and an open-ended waveguide; comparison between theory and experiment. IEEE Trans. Ant. and Prop.; to be published; 1987 February.
- [40] NBS calibration services users guide 1986-88. Nat. Bur. Stand. (U.S.) Spec. Publ. 250, 1986 Ed.; 204 p; 1986 July.

U.S. DEPARTMENT OF COMMERCE  
NATIONAL BUREAU OF STANDARDS  
NATIONAL ENGINEERING LABORATORY  
Boulder, Colorado 80303

SPECIAL TEST

DIPOLE ANTENNA  
Manufacturing Company Name \_\_\_\_\_  
Model \_\_\_\_\_, Serial No. \_\_\_\_\_

Submitted by:

Requesting Company/Agency Name \_\_\_\_\_  
City, State, Zip code \_\_\_\_\_

I. Introduction

The tunable dipole antenna from the requesting company is for measuring horizontally polarized electric fields. The antenna, Model \_\_\_\_\_, is intended for use in the 20 to 300 MHz frequency range. The calibration data given in this report are not necessarily applicable to measurement of impulsive fields.

II. Description of the Tests

At 300 MHz and lower frequencies, horizontally polarized dipole antennas are calibrated at the ground screen site using the standard antenna method described by Taggart and Workman [1].

The calibration is performed by radiating a horizontally-polarized field of convenient magnitude from a dipole or log periodic antenna. A standard receiving antenna capable of accurately measuring the magnitude of this field is placed in the far-field at a fixed height above the ground, and the magnitude of the electric field is measured. The transmitting antenna power is held at a constant level, and the antenna being calibrated is substituted in place of the standard antenna. Figure 1 (see figure 1 in this Tech. Note) shows a diagram of the instrumentation used.

The NBS standard receiving antenna consists of a self-resonant, half-wave dipole antenna with a high impedance balanced voltmeter built into the center of the antenna. The balanced voltmeter is composed of a diode detector and an R-C filter network. The dc output of the filter is measured with a high impedance voltmeter. The antenna voltmeter dc-to-rf voltage transfer function is measured, and from the dc output of the antenna the rf voltage at the essentially open-circuit dipole terminals is determined.

Dipole Antenna  
Manufacturing Company Name  
Model \_\_\_\_\_, Serial No. \_\_\_\_\_

The resulting incident field strength at the measurement location is obtained using the following two equations:

$$L_{\text{eff}} = (\lambda/\pi) \tan (\pi L/2\lambda) \quad (1)$$

$$E_{\text{inc}} = V_{\text{oc}}/L_{\text{eff}} \quad (2)$$

where

$L_{\text{eff}}$  = dipole effective length, m,

$\lambda$  = free-space wavelength, m,

$L$  = length of standard dipole, m,

$V_{\text{oc}}$  = open-circuit standard dipole voltage, V,

$E_{\text{inc}}$  = calibrating incident field strength, V/m.

The antenna factor is the ratio of the field strength to the rf voltage from the antenna being calibrated, delivered to a 50-ohm receiver. Note that a lossless 50-ohm transmission line to the antenna being calibrated is implied in this definition of antenna factor.

### III. Experimental Data

An electric field strength can be determined with the calibrated dipole antenna by measuring the output voltage of the antenna when connected to a 50-ohm impedance and using the following equation:

$$E = K + V \quad (3)$$

where

$E$  = unknown electric field strength, dBV/m,

$K$  = antenna factor, dB, and



Dipole Antenna  
Manufacturing Company Name  
Model \_\_\_\_\_, Serial No. \_\_\_\_\_

V = output voltage of the antenna, dBV.

Note: Units of dB $\mu$ V/m and dB $\mu$ V may also be used in the above equation for E and V, respectively, if desired.

The measured values of antenna factor for the dipole antenna are given in the following table. These values are for the antenna alone, independent of any interconnecting cable loss. The height above ground was 3 meters.

Table 1. Antenna Factor for Dipole Antenna, Model \_\_\_\_\_, SN \_\_\_\_\_.

Frequency MHz	Antenna Factor dB
25	9.0
30	10.5
40	12.0
50	14.0
60	15.5
70	17.0
80	18.2
90	19.3
100	21.1
125	22.4
150	23.1
175	24.3
200	25.7
250	26.4
300	27.0

#### IV. Calibration Uncertainties

For frequencies up to 300 MHz, the dipole antenna is calibrated at the ground screen site. The electromagnetic field levels are sufficient to indicate the requested calibrating field strength at the antenna.

1. The largest source of error in the calibration is uncertainty of the standard field value at the ground screen site. This is due partly to uncertainty in the dc-to-rf voltage transfer. Also, signals from radio stations, CB's, and 2-way radio communications may cause extra errors. The estimated possible error is  $\pm 0.4$  dB.

Page 3 of 4  
Test No. \_\_\_\_\_  
Date of Test: \_\_\_\_\_

Dipole Antenna  
Manufacturing Company Name \_\_\_\_\_  
Model \_\_\_\_\_, Serial No. \_\_\_\_\_

2. Another source of error is uncertainty in the disturbance of the field caused by interconnecting cables and nearby buildings. This error is estimated to be less than  $\pm 0.3$  dB.
3. Other sources of uncertainty are associated with antenna alignment, measurement of antenna separation distance, and NBS calibrations of the various instruments used. These include spectrum analyzers, rf voltmeters, and power monitors. The error due to these latter sources is estimated to be less than  $\pm 0.3$  dB.

The overall worst-case uncertainty of the calibration is the simple sum of those listed, or  $\pm 1$  dB.

When an antenna is used in an actual measurement situation, an additional error may occur. The value of  $K$  is measured for the antenna at a specified height above ground. Use of the calibrated antenna later at a different height or over ground having different constants may cause an error, for heights less than about  $1\lambda$ . This is because the source impedance of the customer's antenna depends on its height above ground, which may lead to measurement error for the usual case where the antenna is loaded by a 50-ohm receiver. A change of impedance doesn't affect the measurement of  $V_{oc}$  by the open-circuit standard dipole, but it changes the mismatch between the customer's antenna and 50-ohm receiver. This error is generally less than 0.5 dB for heights greater than  $\lambda/2$ .

#### V. References

- [1] Taggart, H. E. and Workman, J. L., "Calibration Principles and Procedures for Field Strength Meters (30 Hz to 1 GHz)," NBS Technical Note 370, National Bureau of Standards, Boulder, Colorado (March 1969).

For the Director,  
National Engineering Laboratory

Motohisa Kanda, Ph.D.  
Group Leader, 723.03  
Fields Characterization Group  
Electromagnetic Fields Division

Page 4 of 4  
Test No. \_\_\_\_\_  
Date of Test: \_\_\_\_\_  
Reference: P. O. No. \_\_\_\_\_

U.S. DEPARTMENT OF COMMERCE  
NATIONAL BUREAU OF STANDARDS  
NATIONAL ENGINEERING LABORATORY  
Boulder, Colorado 80303

SPECIAL TEST

DIPOLE ANTENNA  
Manufacturing Company Name \_\_\_\_\_  
Model No. \_\_\_\_\_, Serial No. \_\_\_\_\_

Submitted by:

Requesting Company/Agency Name \_\_\_\_\_  
City, State, Zip code \_\_\_\_\_

I. Introduction

The tunable dipole antenna from the requesting company is for measuring horizontally polarized electric fields. The antenna, Model \_\_\_\_\_, is intended for use in the 200 to 1000 MHz frequency range. The calibration data given in this report are not necessarily applicable for measurement of impulsive fields.

II. Description of the Test

The response of the tunable dipole antenna was calibrated in a standard field in the National Bureau of Standards (NBS) anechoic chamber. This field was generated by a transmitter and calibrated standard-gain antenna. The radiated electric field in the center of the beam was calculated, at the requested frequencies, in terms of the measured value of net transmitted power and the known value of the antenna gain. The effects of multipath reflections within the anechoic chamber have been analyzed and taken into account.

The tunable dipole antenna was mounted on a movable cart in the anechoic chamber and located on the axis (beam center) of the transmitting horn. A block diagram of the instrumentation used is given in figure 1 (see figure 6 in this Tech. Note). The value of the standard field used in the calibration was calculated from the equation

$$E = \frac{\sqrt{30 PG}}{d} \quad (1)$$

where E = on-axis magnitude of the radiated field, volts/meter  
P = net power delivered to the transmitting horn or OEG, watts  
G = calibrated gain of the transmitting antenna, including appropriate near-zone correction factors, and  
d = distance from the horn or OEG aperture to the calibrating field point (center of dipole) meters.

The antenna factor is the ratio of the field strength to the rf voltage from the antenna being calibrated, delivered to a 50-ohm receiver. Note that a lossless 50-ohm transmission line to the antenna being calibrated is implied in this definition of antenna factor.

The value of an unknown electric field strength can be determined with the calibrated tunable dipole antenna by measuring the output voltage of the antenna when connected to a 50-ohm impedance and using the following equation:

$$E = K + V \quad (2)$$

where E = electric field strength, dBV/m  
K = antenna factor, dB, and  
V = 50-ohm output voltage of the antenna, dBV.

Note: Units of dB $\mu$ V/m and dB $\mu$ V may also be used in the above equation for E and V respectively, if desired.

### III. Experimental Data

The measured values of antenna factor for the dipole antenna are given in table 1. As requested, the antenna was calibrated in both the horizontal and vertical polarizations in the anechoic chamber. The measurements are normally done with horizontal polarization and those data are more reliable. The values given are for the antenna alone, independent of any interconnecting cable loss.

Table 1. Antenna factor for horizontal and vertical polarizations for Tunable Dipole Antenna, Requesting Company/Agency Name, Model \_\_\_\_, Serial No. \_\_\_\_.

Frequency (MHz)	Antenna Factor (dB)	
	horizontal	vertical
200	14.0	14.3
250	16.0	15.9
300	17.5	17.7
400	20.0	19.9
500	22.0	22.2
600	23.6	23.5
700	24.9	25.0
800	26.1	26.0
900	27.1	27.3
1000	28.0	28.0

#### IV. Overall Calibration Uncertainty

For frequencies equal to or greater than 200 MHz, the dipole antenna was calibrated in an anechoic chamber. The electromagnetic field level was sufficient to determine the antenna factor of the antenna.

1. The largest source of error is uncertainty in the near-zone gain values of the transmitting antennas. The estimated possible error is  $\pm 0.5$  dB.
2. Another source of error is uncertainty in the magnitude of multi-path reflections within the anechoic chamber. Possible errors caused by chamber reflections are estimated to be less than  $\pm 0.2$  dB.



3. Other sources of uncertainty are associated with antenna alignment, measurement of antenna separation distance, and NBS calibrations of the various instruments used. These include directional couplers, and incident and reflected power monitors. The overall error due to these latter sources is estimated to be less than  $\pm 0.3$  dB.

The overall worst-case uncertainty of the calibration is the simple sum of those listed, or  $\pm 1.0$  dB.

For the Director,  
National Engineering Laboratory

Motohisa Kanda, Ph.D.  
Group Leader, 723.03  
Fields Characterization Group  
Electromagnetic Fields Division

Page 4 of 4  
Test No. \_\_\_\_\_  
Date of Test: \_\_\_\_\_  
Reference: \_\_\_\_\_

U.S. DEPT. OF COMM. <b>BIBLIOGRAPHIC DATA SHEET</b> <i>(See instructions)</i>	<b>1. PUBLICATION OR REPORT NO.</b> NBS/TN-1309	<b>2. Performing Organ. Report No.</b>	<b>3. Publication Date</b> April 1987
<b>4. TITLE AND SUBTITLE</b> NBS Calibration Procedures for Horizontal Dipole Antennas (25 to 1000 MHz)			
<b>5. AUTHOR(S)</b> D.G. Camell, E.B. Larsen, and J.E. Cruz			
<b>6. PERFORMING ORGANIZATION</b> <i>(If joint or other than NBS, see instructions)</i> NATIONAL BUREAU OF STANDARDS DEPARTMENT OF COMMERCE WASHINGTON, D.C. 20234		<b>7. Contract/Grant No.</b>	<b>8. Type of Report &amp; Period Covered</b>
<b>9. SPONSORING ORGANIZATION NAME AND COMPLETE ADDRESS</b> <i>(Street, City, State, ZIP)</i>			
<b>10. SUPPLEMENTARY NOTES</b>  <input type="checkbox"/> Document describes a computer program; SF-185, FIPS Software Summary, is attached.			
<b>11. ABSTRACT</b> <i>(A 200-word or less factual summary of most significant information. If document includes a significant bibliography or literature survey, mention it here)</i> <p>This publication describes the theoretical basis and test procedures for horizontally polarized dipole calibrations at the National Bureau of Standards. Two different techniques and two different test sites are used. The standard antenna method uses the calculation of a field strength level, from the response of a simple half-wave dipole, to calibrate an antenna. This method is used at an open field site in the frequency range of 25 to 1000 MHz. The standard field method applies the theoretical gain equations of waveguides to determine the field strength level. This latter method is used in an anechoic chamber in the frequency range of 200 to 1000 MHz. Procedures for both techniques are explained and measurement setups are illustrated. Measurement uncertainties are discussed. Sample reports are included for both methods.</p>			
<b>12. KEY WORDS</b> <i>(Six to twelve entries; alphabetical order; capitalize only proper names; and separate key words by semicolons)</i> anechoic chamber; calibration; dipole antenna; electromagnetic field; open field site; standard antenna; standard field			
<b>13. AVAILABILITY</b> <input checked="" type="checkbox"/> Unlimited <input type="checkbox"/> For Official Distribution. Do Not Release to NTIS <input checked="" type="checkbox"/> Order From Superintendent of Documents, U.S. Government Printing Office, Washington, D.C. 20402. <input type="checkbox"/> Order From National Technical Information Service (NTIS), Springfield, VA. 22161		<b>14. NO. OF PRINTED PAGES</b> 48	<b>15. Price</b>



# NBS *Technical Publications*

## *Periodical*

---

**Journal of Research**—The Journal of Research of the National Bureau of Standards reports NBS research and development in those disciplines of the physical and engineering sciences in which the Bureau is active. These include physics, chemistry, engineering, mathematics, and computer sciences. Papers cover a broad range of subjects, with major emphasis on measurement methodology and the basic technology underlying standardization. Also included from time to time are survey articles on topics closely related to the Bureau's technical and scientific programs. Issued six times a year.

## *Nonperiodicals*

---

**Monographs**—Major contributions to the technical literature on various subjects related to the Bureau's scientific and technical activities.

**Handbooks**—Recommended codes of engineering and industrial practice (including safety codes) developed in cooperation with interested industries, professional organizations, and regulatory bodies.

**Special Publications**—Include proceedings of conferences sponsored by NBS, NBS annual reports, and other special publications appropriate to this grouping such as wall charts, pocket cards, and bibliographies.

**Applied Mathematics Series**—Mathematical tables, manuals, and studies of special interest to physicists, engineers, chemists, biologists, mathematicians, computer programmers, and others engaged in scientific and technical work.

**National Standard Reference Data Series**—Provides quantitative data on the physical and chemical properties of materials, compiled from the world's literature and critically evaluated. Developed under a worldwide program coordinated by NBS under the authority of the National Standard Data Act (Public Law 90-396).

NOTE: The Journal of Physical and Chemical Reference Data (JPCRD) is published quarterly for NBS by the American Chemical Society (ACS) and the American Institute of Physics (AIP). Subscriptions, reprints, and supplements are available from ACS, 1155 Sixteenth St., NW, Washington, DC 20056.

**Building Science Series**—Disseminates technical information developed at the Bureau on building materials, components, systems, and whole structures. The series presents research results, test methods, and performance criteria related to the structural and environmental functions and the durability and safety characteristics of building elements and systems.

**Technical Notes**—Studies or reports which are complete in themselves but restrictive in their treatment of a subject. Analogous to monographs but not so comprehensive in scope or definitive in treatment of the subject area. Often serve as a vehicle for final reports of work performed at NBS under the sponsorship of other government agencies.

**Voluntary Product Standards**—Developed under procedures published by the Department of Commerce in Part 10, Title 15, of the Code of Federal Regulations. The standards establish nationally recognized requirements for products, and provide all concerned interests with a basis for common understanding of the characteristics of the products. NBS administers this program as a supplement to the activities of the private sector standardizing organizations.

**Consumer Information Series**—Practical information, based on NBS research and experience, covering areas of interest to the consumer. Easily understandable language and illustrations provide useful background knowledge for shopping in today's technological marketplace.

Order the above NBS publications from: Superintendent of Documents, Government Printing Office, Washington, DC 20402.

Order the following NBS publications—FIPS and NBSIR's—from the National Technical Information Service, Springfield, VA 22161.

**Federal Information Processing Standards Publications (FIPS PUB)**—Publications in this series collectively constitute the Federal Information Processing Standards Register. The Register serves as the official source of information in the Federal Government regarding standards issued by NBS pursuant to the Federal Property and Administrative Services Act of 1949 as amended, Public Law 89-306 (79 Stat. 1127), and as implemented by Executive Order 11717 (38 FR 12315, dated May 11, 1973) and Part 6 of Title 15 CFR (Code of Federal Regulations).

**NBS Interagency Reports (NBSIR)**—A special series of interim or final reports on work performed by NBS for outside sponsors (both government and non-government). In general, initial distribution is handled by the sponsor; public distribution is by the National Technical Information Service, Springfield, VA 22161, in paper copy or microfiche form.

**U.S. Department of Commerce**  
National Bureau of Standards  
Gaithersburg, MD 20899

Official Business  
Penalty for Private Use \$300