

National Bureau of Standards

Library, N.W. Bldg

AUG 31 1960

Reference book not to be  
taken from the library.

PB 161564



*Technical Note*

63

---

# SINGLE SCATTERED NEUTRONS FROM AN ISOTROPIC POINT SOURCE



---

U. S. DEPARTMENT OF COMMERCE  
NATIONAL BUREAU OF STANDARDS

## THE NATIONAL BUREAU OF STANDARDS

### Functions and Activities

The functions of the National Bureau of Standards are set forth in the Act of Congress, March 3, 1901, as amended by Congress in Public Law 619, 1950. These include the development and maintenance of the national standards of measurement and the provision of means and methods for making measurements consistent with these standards; the determination of physical constants and properties of materials; the development of methods and instruments for testing materials, devices, and structures; advisory services to government agencies on scientific and technical problems; invention and development of devices to serve special needs of the Government; and the development of standard practices, codes, and specifications. The work includes basic and applied research, development, engineering, instrumentation, testing, evaluation, calibration services, and various consultation and information services. Research projects are also performed for other government agencies when the work relates to and supplements the basic program of the Bureau or when the Bureau's unique competence is required. The scope of activities is suggested by the listing of divisions and sections on the inside of the back cover.

### Publications

The results of the Bureau's work take the form of either actual equipment and devices or published papers. These papers appear either in the Bureau's own series of publications or in the journals of professional and scientific societies. The Bureau itself publishes three periodicals available from the Government Printing Office: The Journal of Research, published in four separate sections, presents complete scientific and technical papers; the Technical News Bulletin presents summary and preliminary reports on work in progress; and Basic Radio Propagation Predictions provides data for determining the best frequencies to use for radio communications throughout the world. There are also five series of nonperiodical publications: Monographs, Applied Mathematics Series, Handbooks, Miscellaneous Publications, and Technical Notes.

Information on the Bureau's publications can be found in NBS Circular 460, Publications of the National Bureau of Standards (\$1.25) and its Supplement (\$1.50), available from the Superintendent of Documents, Government Printing Office, Washington 25, D.C.

# NATIONAL BUREAU OF STANDARDS

## *Technical Note*

63

JULY 1960

### **SINGLE SCATTERED NEUTRONS FROM AN ISOTROPIC POINT SOURCE**

E. R. Mosburg, Jr., and W. M. Murphey

NBS Technical Notes are designed to supplement the Bureau's regular publications program. They provide a means for making available scientific data that are of transient or limited interest. Technical Notes may be listed or referred to in the open literature. They are for sale by the Office of Technical Services, U. S. Department of Commerce, Washington 25, D. C.

DISTRIBUTED BY

**UNITED STATES DEPARTMENT OF COMMERCE**

**OFFICE OF TECHNICAL SERVICES**

WASHINGTON 25, D. C.

Price 50 cents



# Single Scattered Neutrons From an Isotropic Point Source

E. R. Mosburg, Jr., and W. M. Murphey

A calculation was made to determine the number of particles from an isotropic source which will reach a detector after scattering once in the medium surrounding the detector and source. The simple result obtained is applicable when the separation of the detector and the source is much less than the mean free path of the particle in the medium. The particular case of 14.1 Mev neutrons in air was considered. It was found that 44 percent of the single scattered neutrons had scattering angles of less than 30 degrees.

In a time-of-flight scattering cross section measurement using an isotropic source, the number of air scattered neutrons reaching the detector may be comparable to the number of neutrons from the scattering sample. Since rather thick shields are needed to appreciably reduce the background flux from the air scattered neutrons, it is desirable to have information about the angular distribution of the air scattered neutrons. The general problem of air scattering has been studied for various situations previously<sup>1,2,3</sup>. Here, for cases in which only the first scattering is important, a particularly simple form is obtained.

The number of neutrons reaching the detector from a specific infinitesimal volume element,  $dV$ , in space is given by the solid angle,  $g$ , subtended by the detector at  $dV$ , times the probability per unit volume,  $P$ , of scattering in the direction of the detector from the position of the volume element, times,  $n$ , the number of neutrons reaching the volume element from the source, times the size of the volume element,  $dV$ . This product is then integrated over all space, to give the total number,  $N$ , of neutrons reaching the detector;

$$N = \int gPn dV. \quad (1)$$

The probability per unit volume of a neutron scattering in the direction of the detector from the position of the volume element is

the cross section  $\frac{\partial\sigma(\varphi)}{\partial\Omega}$  in millibarns per steradian as a function of scattering angle,  $\varphi$ , divided by the area element  $dA$  in  $\text{cm}^2$  subtended by the volume element as seen from the source, times the number of target atoms per unit volume,  $n_t$ . So,

$$P = \frac{\partial\sigma(\varphi)}{\partial\Omega} \frac{1}{dA} n_t. \quad (2)$$

For convenience the geometry as shown in fig. 1 will be used. All points with the same scattering angle lie on a circle drawn through the source, detector, and one of the scattering points with that angle (see note 1). The sum over all space will be carried out by using the cylindrical symmetry about the axis connecting the source and the detector and by integrating over all points with a common scattering angle and then integrating over scattering angle. Since there is cylindrical symmetry about the axis connecting the source and the detector the volume element,  $dV$ , can be defined as a ring of volume  $dV = (2\pi r_1 \sin\theta_1) r_1 d\theta_1 dr_1$ .

The solid angle,  $g$ , for a detector which is a sphere and has a cross sectional area of  $1 \text{ cm}^2$  is  $g = \frac{1}{r_2^2}$  steradians.

The number of neutrons reaching the volume element is  $n = \frac{SdA}{4\pi r_1^2}$

where  $S$  is the total emission of the isotropic source in number per second. Since the total cross sections at 14.1 Mev for nitrogen and oxygen are about 1.5 barns ( $1 \text{ barn} = 10^{-24} \text{ cm}^2$ ), and since there are about  $5.3 \times 10^{19}$  atoms per  $\text{cm}^3$  in air, the mean free path of neutrons in air is about 86 meters. The exponential attenuation of the unscattered flux may be neglected for distances under about 10 meters.

---

<sup>1/</sup> U. Fano, L. V. Spencer, and M. J. Berger, Penetration and Diffusion of X-rays, Encyclopedia of Physics Vol. 38/2 p.661.

<sup>2/</sup> A. Langsdorf, Neutron Collimation and Shielding, Chapter IVE of Fast Neutron Physics, Part I by J. B. Marion and J. L. Fowler, Interscience Publishers (1960).

<sup>3/</sup> R. E. Beissner and D. R. Smith, "Studies in Shielding", Convair Report FZK-9-117, 1958.

Making the above substitutions we have:

$$N = \int \frac{\partial \sigma(\varphi)}{\partial \Omega} \frac{1}{dA} n_t dV \frac{S}{4\pi r_1^2} dA \frac{1}{r_2^2} = \int_{r_1} \int_{\theta_1} \frac{\partial \sigma(\varphi)}{\partial \Omega} \frac{n_t S 2\pi r_1^2 \sin \theta_1 dr_1 d\theta_1}{4\pi r_1^2 r_2^2}, \quad (3)$$

$$N = \int_{r_1} \int_{\theta_1} \frac{\partial \sigma(\varphi)}{\partial \Omega} \frac{n_t S \sin \theta_1 d\theta_1 dr_1}{2 r_2^2}. \quad (4)$$

From fig. 1:

$$\alpha = 180 - 2(\theta_1 + 90^\circ - \varphi)$$

$$\text{so } \beta = 2\varphi - \alpha = 2\varphi - 180 + 2\theta_1 + 180^\circ - 2\varphi = 2\theta_1 \quad (5)$$

$$\text{so } R \sin \theta_1 = \frac{r_2}{2}; \quad r_2^2 = 4R^2 \sin^2 \theta_1.$$

$$N = \int_{r_1} \int_{\theta_1} \frac{\partial \sigma(\varphi)}{\partial \Omega} \frac{n_t S d\theta_1 dr_1}{8 R^2 \sin \theta_1}. \quad (6)$$

Now we transform from integrating over  $\theta_1$  and  $r_1$  to integrating over  $\theta_1$  and  $\varphi$ .

$$d\theta_1 dr_1 = J\left(\frac{\theta_1 r_1}{\theta_1 \varphi}\right) d\theta_1 d\varphi \quad (7)$$

where

$$J\left(\frac{\theta_1 r_1}{\theta_1 \varphi}\right) = \begin{vmatrix} \frac{\partial \theta_1}{\partial \theta_1} & \frac{\partial r_1}{\partial \theta_1} \\ \frac{\partial \theta_1}{\partial \varphi} & \frac{\partial r_1}{\partial \varphi} \end{vmatrix} = \begin{vmatrix} 1 & \frac{\partial r_1}{\partial \theta_1} \\ 0 & \frac{\partial r_1}{\partial \varphi} \end{vmatrix} = \frac{\partial r_1}{\partial \varphi} \quad (8)$$

since  $\theta_1$  and  $\varphi$  are independent variables. Again from fig. 1 we have

$$r_1 = -\frac{D}{\sin \varphi} \sin (\theta_1 - \varphi), \quad (9)$$

where  $D$  is the distance from the source to the detector,

$$\frac{\partial r_1}{\partial \varphi} = D \frac{\sin \theta_1}{\sin^2 \varphi}, \quad (10)$$

$$N = \int_{\varphi} \int_{\theta_1} \frac{\partial \sigma(\varphi)}{\partial \Omega} \frac{n_t S}{8 R^2} \frac{D}{\sin^2 \varphi} d\theta_1 d\varphi. \quad (11)$$

Since  $R = \frac{D}{2 \sin \varphi}$  ;

$$N = \frac{n_t S}{2 D} \int_{\varphi} \int_{\theta_1=0}^{\theta_1=\varphi} \frac{\partial \sigma(\varphi)}{\partial \Omega} d\theta_1 d\varphi \quad (12)$$

Integrating over  $\theta_1$  we have the simple result,

$$N = \frac{n_t S}{2 D} \int \varphi \frac{\partial \sigma(\varphi)}{\partial \Omega} d\varphi. \quad (13)$$

This integral may be used for any single scattering where exponential attenuation of the number of particles may be neglected. Usually  $\frac{\partial \sigma(\varphi)}{\partial \Omega}$  is not known as an analytic function and one must integrate numerically. Equation (13) actually gives the number of neutrons for one energy. If  $\frac{\partial \sigma(\varphi)}{\partial \Omega}$  varies greatly over the energies of neutrons being considered then an integration of (13) must be made for each energy.

In our case  $N$  was determined for a 14.1 Mev neutron source. Since the total cross sections of nitrogen and oxygen are about equal at 14 Mev and exact results were not required it was assumed, for the purpose of  $\sigma(\varphi)$ , that the air was all nitrogen.  $D$  was 100 cm and the angular distribution (fig. 2) was that given in BNL-400<sup>4/</sup>. The results of the integration of (13) are plotted versus scattering angle,  $\varphi$ , in fig. 3. Fig. 4 shows the fraction of single scattering angle less than a given angle. Reading from the graph of fig. 4 one notices that 44 percent of the single scattered neutrons have scattering angles of less than 30 degrees. It was concluded, therefore, that the detector shielding for scattering angles less than 30° should be designed to give about three times as much attenuation as the shielding for other angles.

NOTE:

$$\varphi = \theta_1 + \theta_2$$

$\alpha$  is a constant

$$\theta_1 = (90 - \frac{\delta}{2}) - (90 - \frac{\alpha}{2})$$

$$\theta_1 = -\frac{\delta}{2} + \frac{\alpha}{2}$$

$$\theta_2 = 90 - (\frac{\alpha}{2} - \frac{\delta}{2}) - \left[ 90 - (\frac{\alpha}{2}) \right]$$

$$\theta_2 = -\frac{\alpha}{2} + \frac{\alpha}{2} + \frac{\delta}{2} = +\frac{\delta}{2}$$

$$\text{Therefore } \theta_1 + \theta_2 = -\frac{\delta}{2} = +\frac{\delta}{2} + \frac{\alpha}{2} = \text{a constant.}$$

Therefore  $\varphi$  is a constant independent of  $\theta_1$  and  $\theta_2$ .

---

<sup>4/</sup> D. J. Hughes and R. S. Carter, Neutron Cross Sections, Angular Distributions. BNL 400, June 1956.




Figure 1. Geometry used in the calculation.

$$R^2 = R^2 + r_2^2 - 2r_2 R \cos (\theta_1 + 90 - \theta_2)$$

$$\phi = \theta_1 + \theta_2$$

$$r_2 = 2R \sin \theta_1$$

$$R = \frac{D}{2 \sin \phi}$$

$r_1$  = distance from source to scattering point

$r_2$  = distance from scattering point to detector

$\phi$  = scattering angle

$R$  = radius of a circle through source, scattering point and detector.




Figure 2. Elastic scattering cross section of 14.1 Mev neutrons on nitrogen (BNL-400).




Figure 3. Number of single scattered neutrons as a function of scattering angle.




Figure 4. Relative number of neutrons reaching the detector with scattering angles less than a given angle.

$$\frac{\sum_{\phi=0}^{180} \frac{v \sigma(\phi)}{8 \Omega} \Delta \phi}{\sum_{\phi=0}^{\phi} \frac{v \sigma(\phi)}{8 \Omega} \Delta \phi}$$




Figure 5. Geometry for note.



## THE NATIONAL BUREAU OF STANDARDS

The scope of activities of the National Bureau of Standards at its major laboratories in Washington, D.C., and Boulder, Colorado, is suggested in the following listing of the divisions and sections engaged in technical work. In general, each section carries out specialized research, development, and engineering in the field indicated by its title. A brief description of the activities, and of the resultant publications, appears on the inside of the front cover.

### WASHINGTON, D.C.

**Electricity and Electronics.** Resistance and Reactance. Electron Devices. Electrical Instruments. Magnetic Measurements. Dielectrics. Engineering Electronics. Electronic Instrumentation. Electrochemistry.

**Optics and Metrology.** Photometry and Colorimetry. Photographic Technology. Length. Engineering Metrology.

**Heat.** Temperature Physics. Thermodynamics. Cryogenic Physics. Rheology. Molecular Kinetics. Free Radicals Research.

**Atomic and Radiation Physics.** Spectroscopy. Radiometry. Mass Spectrometry. Solid State Physics. Electron Physics. Atomic Physics. Neutron Physics. Radiation Theory. Radioactivity. X-rays. High Energy Radiation. Nucleonic Instrumentation. Radiological Equipment.

**Chemistry.** Organic Coatings. Surface Chemistry. Organic Chemistry. Analytical Chemistry. Inorganic Chemistry. Electrodeposition. Molecular Structure and Properties of Gases. Physical Chemistry. Thermochemistry. Spectrochemistry. Pure Substances.

**Mechanics.** Sound. Mechanical Instruments. Fluid Mechanics. Engineering Mechanics. Mass and Scale. Capacity, Density, and Fluid Meters. Combustion Controls.

**Organic and Fibrous Materials.** Rubber. Textiles. Paper. Leather. Testing and Specifications. Polymer Structure. Plastics. Dental Research.

**Metallurgy.** Thermal Metallurgy. Chemical Metallurgy. Mechanical Metallurgy. Corrosion. Metal Physics.

**Mineral Products.** Engineering Ceramics. Glass. Refractories. Enameled Metals. Constitution and Microstructure.

**Building Technology.** Structural Engineering. Fire Protection. Air Conditioning, Heating, and Refrigeration. Floor, Roof, and Wall Coverings. Codes and Safety Standards. Heat Transfer. Concreting Materials.

**Applied Mathematics.** Numerical Analysis. Computation. Statistical Engineering. Mathematical Physics.

**Data Processing Systems.** SEAC Engineering Group. Components and Techniques. Digital Circuitry. Digital Systems. Analog Systems. Application Engineering.

• Office of Basic Instrumentation.

• Office of Weights and Measures.

### BOULDER, COLORADO

**Cryogenic Engineering.** Cryogenic Equipment. Cryogenic Processes. Properties of Materials. Gas Liquefaction.

**Radio Propagation Physics.** Upper Atmosphere Research. Ionospheric Research. Regular Propagation Services. Sun-Earth Relationships. VHF Research. Radio Warning Services. Airglow and Aurora. Radio Astronomy and Arctic Propagation.

**Radio Propagation Engineering.** Data Reduction Instrumentation. Modulation Research. Radio Noise. Tropospheric Measurements. Tropospheric Analysis. Propagation Obstacles Engineering. Radio-Meteorology. Lower Atmosphere Physics.

**Radio Standards.** High Frequency Electrical Standards. Radio Broadcast Service. High Frequency Impedance Standards. Electronic Calibration Center. Microwave Physics. Microwave Circuit Standards.

**Radio Communication and Systems.** Low Frequency and Very Low Frequency Research. High Frequency and Very High Frequency Research. Ultra High Frequency and Super High Frequency Research. Modulation Research. Antenna Research. Navigation Systems. Systems Analysis. Field Operations.













