

NIST Special Publication 250 NIST SP 250-99

Diffuse Visual Transmission Density

Clarence J. Zarobila Robert E. Vest Maria E. Nadal

This publication is available free of charge from: https://doi.org/10.6028/NIST.SP.250-99



NIST Special Publication 250 NIST SP 250-99

Diffuse Visual Transmission Density

Clarence J. Zarobila Robert E. Vest Maria E. Nadal Sensor Science Division Physical Measurement Laboratory

This publication is available free of charge from: https://doi.org/10.6028/ NIST.SP.250-99

August 2022



U.S. Department of Commerce Gina M. Raimondo, Secretary

National Institute of Standards and Technology Laurie E. Locascio, NIST Director and Under Secretary of Commerce for Standards and Technology

Certain commercial entities, equipment, or materials may be identified in this document in order to describe an experimental procedure or concept adequately. Such identification is not intended to imply recommendation or endorsement by the National Institute of Standards and Technology, nor is it intended to imply that the entities, materials, or equipment are necessarily the best available for the purpose.

NIST Technical Series Policies

Copyright, Fair Use, and Licensing Statements NIST Technical Series Publication Identifier Syntax

Publication History Approved by the NIST Editorial Review Board on 2022-08-11

How to Cite this NIST Technical Series Publication

Zarobila CJ, Vest RE, Nadal ME (2022) Diffuse Visual Transmission Density. (National Institute of Standards and Technology, Gaithersburg, MD), NIST Special Publication 250 (NIST SP 250) NIST SP 250-99. https://doi.org/10.6028/NIST.SP.250-99.

NIST Author ORCID iDs

Clarence J. Zarobila: 0000-0003-4547-0754 Robert E. Vest: 0000-0001-6193-9589 Maria E. Nadal: 0000-0002-8287-366X

Contact Information zarobila@nist.gov

Abstract

This document, SP250-99, NIST Measurement Services: Diffuse Visual Transmission Density, describes the instrument used to measure diffuse visual transmission density of both x-ray and photographic film step tablets. These step tablets replace former Standard Reference Material (SRM) offerings SRM1001 and SRM1008. The instrument is fully automated and is designed and characterized to comply with the international standards for measuring transmission density using the diffuse influx mode.

Keywords

densitometer; diffuse influx; diffuse transmission; transmission density.

Table of Contents

1. I	Introduction	1
2. 1	Theory and Measurement Equations	1
2.1.	. Overview	2
2.2.	. Measurement Equation Derivation	3
3. I	Description of the Instrument	8
3.1.	. Source System	8
3.2.	. Film Transport System	10
3.3.	Detector System	10
3.4.	System Assembly	12
4. I	Instrument Characterization and Validation	14
5. 0	Calibration Procedure	23
6. l	Uncertainties	25
7. (Conclusions	29
Refer	rences	29
Appe	endix A. Sample Report of Calibration	31

List of Tables

3
8
21
24
27
28

List of Figures

Fig. 1. Important Components of the Diffuse Transmittance Densiometer	3
Fig. 2. Cross-Sectional View of the Source System	9
Fig. 3. Top View of Film Transport System	10
Fig. 4. Cross-Section of the Film Holder and Opal Assembly	11
Fig. 5. Cross-Section of the Detector System	12
Fig. 6. Schematic Diagram of the Components in the Diffuse Transmission System	13
Fig. 7. DI/O Connections to the Various Components	14
Fig. 8. Relative Flux Distributions	16
Fig. 9. 6° - Hemispherical Reflectance of the Opal as a Function of Wavelength	17
Fig. 10. Normalized Signal from the Efflux of the Opal	18
Fig. 11. Spectral Response of the Detector System	19
Fig. 12. Relative Responsivity vs. Gain Settings	20
Fig. 13. Reflectance and Transmittance of the Indicated Films	22

Acknowledgments

The text presented herein has been edited and revised from the original work of Edward A. Early, Thomas R. O'Brian, Robert D. Saunders, and Albert C. Parr in NIST Special Publication SP260-135.

1. Introduction

Transmission density is an important physical property for exposed films in the fields of medicine, non-destructive testing, photography, and graphic arts. Standards for transmission density are provided by the National Institute of Standards and Technology (NIST) in the form of film step tablets, 254 mm long by 35 mm wide, with steps extending the width of each film and equally spaced along its length. Double emulsion x-ray film and single emulsion photographic film are used. The steps have increasing transmission densities from approximately 0.1 to 4 from one end of the film to the other. The X-ray film has 17 steps, while the photographic film has 23. The film step tablets described herein fulfill the requirements found in Section 5 of ASTM E1079 [1], and in Section T-262 of ASME V [2].

The transmission densities of film step tablets are determined with an instrument using the diffuse influx mode. Diffuse illumination is achieved with a flash opal, while directional detection is accomplished with a lens system. The transmitted radiant flux is detected by a temperature-controlled silicon (Si) photodiode with amplifier electronics capable of measuring signals spanning seven orders of magnitude. The instrument is designed to automatically measure films using computerized data acquisition and control.

This document is intended to provide customers and potential customers with a detailed overview of the capabilities and procedures used in performing the measurements of density standards. The NIST Calibration Services Program offers Service ID Numbers (also known as SKU) 38100C and 38120C for X-Ray Film Step Tablet Transmission, and Photographic Film Step Tablet Transmission, respectively and 38110C and 38130C for the recalibration of X-Ray Film Step Tablet Transmission, respectively. They are found in the Surface Color and Appearance section of the Optical Radiation Measurements category of the NIST Storefront [3]. The measurement service quality system is based on the International Organization for Standardization (ISO)17025 standard [4].

The organization of this Special Publication is in sections as follows. Section 1 presents the motivation for establishing the instrument. Section 2 describes the relevant theory and measurement equations. Section 3 describes the instrument. Section 4 discusses the instrument characterization and measurement validation. Section 5 describes the operation of the instrument. Section 6 discusses the sources of uncertainty and their associated values. Conclusions are drawn in Sec. 7. Finally, a sample calibration report is presented in the Appendix.

2. Theory and Measurement Equations

The purpose of the measurement equations derived in this section is to describe the measurement process in terms of all the quantities that contribute to the uncertainty of the transmission density values. These quantities are the signals measured by the Si photodiode detector, the spectral properties of the influx and detected radiant flux, and the spatial properties of the radiant flux. Therefore, three measurement equations are derived and are used both for calculating transmission densities from experimental results and for determining the uncertainty in these transmission densities.

2.1. Overview

Two radiant fluxes are important for determining transmission density. Following the nomenclature given in [5, 6], these fluxes are the aperture flux Φ_j , which is the flux emerging from the sampling aperture, and the transmitted flux Φ_{τ} , which is the flux that passes through the specimen. In terms of these fluxes, the transmittance factor, *T*, is given by

$$T = \frac{\phi_{\tau}}{\phi_{j}} \tag{1}$$

and the transmission density, $D_{\rm T}$, is given by

$$D_{\rm T} = -log_{10} T. \tag{2}$$

To this point, neither the geometrical nor spectral properties of the fluxes have been specified. For diffuse transmission density, either the incident flux (influx) must be diffuse or detection must include both the regular and diffuse components of the transmitted flux (efflux) [7]. Therefore, two equivalent experimental modes are possible for measuring diffuse transmission density: the diffuse influx mode, in which the illumination is diffuse and the detection is directional, and the diffuse efflux mode, in which the illumination is directional and the detection is diffuse.

Likewise, the spectral properties of the fluxes are defined for diffuse transmission density [8]. The relative spectral flux distribution of the incident flux is denoted by $S_{\rm H}$ and is based on the spectral flux distribution of the CIE Standard Illuminant A modified in the infrared region to protect the sample and optical elements from excessive heat. For visual transmission density, the combined spectral response of the detector and the spectral characteristics of the optical elements collecting the transmitted flux is denoted by $V_{\rm T}$. The spectral product is obtained by multiplying, at each wavelength, the spectral power of the influx spectrum by the spectral response of the receiver. The spectral product is specified to be the same as the product of CIE Standard Illuminant A, $S_{\rm A}$, with the spectral luminous efficiency function for photopic vision, V_{λ} . Therefore, at each wavelength,

$$S_{\rm H} \cdot V_{\rm T} = S_{\rm A} \cdot V_{\lambda} \,. \tag{3}$$

In terms of the definitions and concepts presented in the previous paragraphs, the NIST instrument measures diffuse visual transmission density using the diffuse influx mode. The important components of the instrument are shown in Fig. 1, as well as a spherical coordinate system. The film step tablet is in contact with the opal, which provides an influx with angles of incidence from 0° to 90° and a relative spectral flux distribution $S_{\rm H}$. The efflux is collected within an acceptance cone having a half-angle κ less than 10° and with a spectral response $V_{\rm T}$. Therefore, using the functional notation specified in [5, 6, 7], the measured transmission density is described by

$$D_{\rm T}(90^{\circ} \, opal; \, S_{\rm H} \le 10^{\circ}; \, V_{\rm T}).$$
 (4)



Fig. 1. Important Components of the Diffuse Transmittance Densiometer

2.2. Measurement Equation Derivation

The properties of the various components used in deriving the measurement equations are given in Table 1. The influx has radiance $L_j(\lambda)$, where λ is the wavelength. The opal has reflectance $\rho_0(\lambda)$ and diffusion coefficient *d* (defined below).

Property	Symbol
Influx Radiance	$L_{ m j}(\lambda)$
Opal reflectance	$ ho_{ m o}(\lambda)$
Opal diffusion coefficient	d
Step transmittance	$ au_{ m s}(\lambda)$
Step reflectance	$ ho_{ m s}(\lambda)$
Efflux radiance	$L_{ au}(\lambda)$
Acceptance cone half angle	κ
Filter transmittance	$ au_{ m f}(\lambda)$
Photodiode responsivity	$R(\lambda)$
Amplifier gain	G

Table 1. Properties and Symbols of Components in the Transmission Density Instrument

The differential radiant flux $d\Phi$ from the opal or the step to the collecting lens is given by

$$d\Phi(\lambda, a, \omega) = L(\lambda, a, \omega) \cdot da \cdot d\omega, \tag{5}$$

where *a* is a point on the opal or step and ω is a direction with angles 0 and ϕ . Assume that the radiance $L(\lambda, a, \omega)$ is constant within the acceptance cone. Then, by integrating Eq. (5) over the spatial variables the spectral flux $\Phi(\lambda)$ is given by

$$\Phi(\lambda) = \iint d\Phi(\lambda, a, \omega) \cdot da \cdot d\omega$$

= $L(\lambda) \cdot A \cdot \Omega,$ (6)

which is the spectral flux incident on the photopic filter. The photopic filter modifies this flux so that the spectral flux incident on the photodiode is given by

$$\Phi(\lambda) = L(\lambda) \cdot \tau_{\rm f}(\lambda) \cdot A \cdot \Omega. \tag{7}$$

For the opal, the spectral radiance is

$$L(\lambda) = L_{j}(\lambda), \tag{8}$$

while for the step the spectral radiance is

$$L(\lambda) = L_{\tau}(\lambda)$$

= $L_{j}(\lambda) \cdot \frac{\tau_{s}(\lambda)}{1 - \rho_{o}(\lambda) \cdot \rho_{s}(\lambda)},$ (9)

showing that the radiance from the opal is modified by the transmittance of the step and by interreflections between the opal and the step. Combining Eq. (7) with Eqs. (8) and (9) yields the spectral aperture and transmitted fluxes, respectively. Thus, the spectral aperture flux is

$$\Phi_{j}(\lambda) = L_{j}(\lambda, \omega) \cdot \tau_{f}(\lambda) \cdot A \cdot \Omega$$
(10)

and the spectral differential transmitted flux is

$$\Phi_{\tau}(\lambda) = L_{j}(\lambda, \omega) \cdot \frac{\tau_{s}(\lambda, \omega)}{1 - \rho_{o}(\lambda) \cdot \rho_{s}(\lambda)} \cdot \tau_{f}(\lambda) \cdot A \cdot \Omega .$$
(11)

The spectral output current $I(\lambda)$ from the photodiode, for either the spectral aperture flux or the spectral transmitted flux, is given by

$$I(\lambda) = \Phi(\lambda) \cdot R(\lambda). \tag{12}$$

The total current *I* is obtained by integrating Eq. (12) over wavelength, yielding

$$I = \int dI(\lambda) = \int \Phi(\lambda) \cdot R(\lambda) \cdot d\lambda.$$
(13)

From Eq. (13) and Eqs. (10) and (11), the current from the aperture flux I_j is given by

$$I_{j} = A \cdot \Omega \cdot \int L_{j}(\lambda) \cdot \tau_{f}(\lambda) \cdot R(\lambda) d\lambda$$
(14)

and the current from the transmitted flux I_{τ} by

$$I_{\tau} = A \cdot \Omega \cdot \int L_{j}(\lambda) \cdot \frac{\tau_{s}(\lambda)}{1 - \rho_{o}(\lambda) \cdot \rho_{s}(\lambda)} \cdot \tau_{f}(\lambda) \cdot R(\lambda) \cdot d(\lambda) .$$
(15)

Note that the currents in Eqs. (14) and (15) are proportional to the radiant flux and contain both the spectral and geometrical information specified by Eq. (2.4). In terms of the nomenclature,

$$S_{\rm H} \propto L_{\rm j}(\lambda,\omega)$$
 (16)

and

$$V_{\rm T} = \tau_{\rm f}(\lambda) \cdot R(\lambda) \,. \tag{17}$$

Finally, the measured signal S, a voltage, is given by

$$S_{j} = I_{j} \cdot G_{j} \tag{18}$$

for the aperture flux and by

$$S_{\tau} = I_{\tau} \cdot G_{\tau} \tag{19}$$

for the transmitted flux.

Re-expressing Eq. (2) in terms of the aperture and transmitted fluxes, the transmission density is

$$D_{\rm T} = \log_{10} \left(\frac{\phi_{\rm j}}{\phi_{\rm \tau}}\right). \tag{20}$$

The first two measurement equations derived below depend upon the currents being proportional to the fluxes, so that Eq. (20) is equivalent to

$$D_{\rm T} = \log_{10} \left(\frac{l_{\rm j}}{l_{\rm \tau}} \right). \tag{21}$$

For transmission density as a function of the measured signals, the first measurement equation is obtained by substituting Eqs. (18) and (19) into Eq. (21) to yield

$$D_{\rm T} = \log_{10} \left(\frac{s_{\rm j}}{s_{\rm \tau}} \cdot \frac{G_{\rm \tau}}{G_{\rm j}} \right). \tag{22}$$

The second measurement equation expresses the transmission density as a function of the spectral variables. Using Eqs. (14) and (15) and considering only the wavelength dependence of the variables, Eq. (21) becomes

$$D_{\rm T} = \log_{10} \left(\frac{\int L_{\rm j}(\lambda) \cdot \tau_{\rm f} \cdot R(\lambda) \cdot d(\lambda)}{\int L_{\rm j}(\lambda) \cdot \frac{\tau_{\rm s}(\lambda)}{1 - \rho_{\rm o}(\lambda) \cdot \rho_{\rm s}(\lambda)} \cdot \tau_{\rm f}(\lambda) \cdot R(\lambda) \cdot d(\lambda)} \right).$$
(23)

The transmission density as a function of the spatial variables is given by the third measurement equation. The derivation which follows is adapted from [8]. Considering only the spatial variables, the radiant flux incident on the photodiode is a fraction of the total efflux. Integrating Eq. (5) over the spatial variables and ignoring the wavelength yields a flux of

$$\Phi = \iint L(a,\omega) \cdot da \cdot d\omega \,. \tag{24}$$

Assuming that the radiance $L(a, \omega)$ is constant over the area of the opal and over all azimuthal angles φ , Eq. (24) becomes

$$\Phi = 2\pi \cdot A \cdot \int L(\theta) \cdot \cos\theta \sin\theta \cdot d\theta \,. \tag{25}$$

For the measured aperture and transmitted fluxes Φ_j and Φ_{τ} , respectively, the limits of

integration in Eq. (25) are 0 to κ , while for the total aperture and transmitted fluxes $\Phi_{T,j}$ and $\Phi_{T,\tau}$, respectively, the limits are 0 to $\pi/2$. The ratios of these fluxes are given by

$$\frac{\Phi_{j}}{\Phi_{T,j}} = \frac{\int_{0}^{\kappa} L_{j}(\theta) \cdot \cos\theta \sin\theta \cdot d\theta}{\int_{0}^{\pi/2} L_{j}(\theta) \cdot \cos\theta \sin\theta \cdot d\theta}$$
(26)

and

$$\frac{\Phi_{\tau}}{\Phi_{T,\tau}} = \frac{\int_{0}^{\kappa} L_{\tau}(\theta) \cdot \cos\theta \sin\theta \cdot \mathrm{d}\theta}{\int_{0}^{\pi/2} L_{\tau}(\theta) \cdot \cos\theta \sin\theta \cdot \mathrm{d}\theta}.$$
(27)

Substituting Eqs. (26) and (27) into the definition of transmission density given by Eq. (20) yields

$$D_{\rm T} = log_{10} \left(\frac{\Phi_{\rm T,j} \frac{\int_0^{\kappa} L_j(\theta) \cdot \cos\theta \sin\theta \cdot d\theta}{\int_0^{\pi/2} L_j(\theta) \cdot \cos\theta \sin\theta \cdot d\theta}}{\Phi_{\rm T,\tau} \frac{\int_0^{\kappa} L_{\tau}(\theta) \cdot \cos\theta \sin\theta \cdot d\theta}{\int_0^{\pi/2} L_{\tau}(\theta) \cdot \cos\theta \sin\theta \cdot d\theta}} \right).$$
(28)

Assuming that the opal is Lambertian for $0 < \theta < \kappa$, $L_{j,\theta}$ is constant for these angles. Furthermore, assume that the scattering properties of the step make it Lambertian for $0 < \theta < \pi/2$. Therefore, Eq. (28) reduces to

$$D_{\rm T} = \log_{10} \left(\frac{\Phi_{\rm T,j} \cdot \int_{0}^{\pi/2} \cos\theta \sin\theta \cdot \mathrm{d}\theta}{\Phi_{\rm T,\tau} \cdot \int_{0}^{\pi/2} \frac{L_{\rm j}(\theta)}{L_{\rm j}(0)} \cdot \cos\theta \sin\theta \cdot \mathrm{d}\theta} \right), \tag{29}$$

where $L_{i}(0)$ is the radiance from the opal at 0 = 0. Defining the diffusion coefficient d to be

$$d = \frac{\int_{0}^{\pi/2} \frac{L_{j}(\theta)}{L_{j}(0)} \cos\theta \sin\theta \cdot d\theta}{\int_{0}^{\pi/2} \cos\theta \sin\theta \cdot d\theta},$$
(30)

Eq. (29) becomes

$$D_{\rm T} = \log_{10} \left(\frac{\Phi_{\rm T,j}}{\Phi_{\rm T,\tau}} \cdot \frac{1}{d} \right). \tag{31}$$

The three measurement equations are given by Eqs. (22), (23), and (31). The first is the method for calculating D_T from the measured signals and the gains, while the second relates D_T to the spectral properties of the influx, opal, step, photopic filter, and photodiode. The third measurement equation expresses the dependence of D_T on the spatial properties of the aperture flux, accounting for the non-Lambertian quality of the opal.

3. Description of the Instrument

The instrument was designed and built to automatically measure the diffuse visual transmission density of film step tablets using the diffuse influx mode. The major optical and electronic components are those shown in Fig. 1. A computer controls the entire measurement sequence. This section provides a detailed description of the instrument, which is conveniently divided into three systems: the source, the film transport, and the detector. A list of the major components of the instrument is given in Table 2.¹

Component	Description	Component	Description
Lamp	100 W quartz-tungsten- halogen	Horizontal and Vertical Stages	Screw-driven
Lamp Reflector	Ellipsoidal, F/2	Horizontal Stage Gearbox	Inline, NEMA 34
Lamp Power Supply	20.0 V, 10.0 A	Horizontal and Vertical Stage Motors	Stepper
Optical Filter	Infrared cut-off	Solenoid Valve	
Water Chiller	Thermal control to +/- 0.05°C	Digital Input-Output (DIO)	USB driven
Flow Meter	0.4 L/min to 5.3 L/min	Collecting Lens	Bi-convex, 75.6 mm focal length
Shutter	35 mm aperture	Collimating and Focusing Lens	Bi-convex, 38.1 mm focal length
Shutter Controller	Single channel shutter driver	Photopic Filter	Replicates the Luminous Efficiency Function
Motion Control Power	Industrial DIN Rail	Photodiode	Si detector
Photodiode Amplifier		Digital Voltmeter	6 ¹ / ₂ digit multimeter

 Table 2. Table of Major Components Used in the Instrument

3.1. Source System

The source system provides a diffuse illumination to the film step tablet with the correct spectral flux distribution. The system consists of a lamp and housing, an infrared filter assembly, a shutter, and an opal assembly. A cross section of the source system is shown in Fig. 3.1. Starting from the bottom of the figure, a 100 W Quartz-Tungsten-Halogen (QTH) lamp is contained in a lamp housing. The lamp is burned in at 8.0 A for 24 h and then at 7.8 A for an additional 48 h prior to being used for measurements. An elliptical reflector focuses the light from this lamp at approximately the position of the opal. The lamp housing is cooled with distilled water at a temperature of 25 °C pumped through a chiller, and a constant dc current is run through the lamp from a power supply. If the flow of water is interrupted, a flow meter closes a switch and disables the power supply. This prevents the infrared filter assembly, described next, from overheating.

¹ Certain commercial entities, equipment, or materials may be identified in this document in order to describe an experimental procedure or concept adequately. Such identification is not intended to imply recommendation or endorsement by the National Institute of Standards and Technology, nor is it intended to imply that the entities, materials, or equipment are necessarily the best available for the purpose.



Fig. 2. Cross-Sectional View of the Source System

The infrared filter assembly consists of aluminum plates for mounting to the lamp housing and the shutter, and an infrared filter. Infrared filtering is accomplished with a combination of water and an optical filter. Distilled water is circulated, using the same chiller as for the lamp housing, through the stainless-steel water filter in the direction indicated in Fig. 3.2. This water absorbs most of the light at infrared wavelengths, as well as cooling the plates which seal, with rubber gaskets, the top and bottom of the water filter. The bottom plate is simply a piece of BK-7 glass with a diameter of 100 mm and a thickness of 4 mm. The top plate is an optical filter, Hoya LP-15 (HA15 equivalent), with the same diameter and thickness as the glass plate. The optical filter was chosen so that its transmittance would modify the spectral flux distribution from the lamp to one approximating S_H. When the shutter is closed, it blocks all light from reaching the opal so that a background signal can be measured. A controller operates the shutter and is interfaced to the computer.

The opal assembly consists of a plate for mounting to the shutter and two large pieces: one for holding the opal and mounting to the film plate and the other to provide a vacuum for bringing the film step tablet in contact with the opal (described below). Several smaller pieces are also used including a sleeve and a spacer. All pieces are aluminum and black-anodized. A flash opal with a diameter of 3 mm and a thickness of 1.5 mm is the source of diffuse illumination. The opal is fit partially into the sleeve and held in place with black tape. The depth of the opal in the sleeve and the height of the spacer are adjusted so that the top of the opal is aligned with the vacuum plate when the opal assembly is complete. Black enamel paint is used so that all the light incident on the film step tablet originates from the top of the opal and between the opal and the sleeve. The sleeve and spacer are inserted in the opal plate and a mounting plate attached to the opal plate holds the sleeve in place.

3.2. Film Transport System

The film transport system picks up a film from a tray, positions the film to measure the transmission density of each step on the film, and drops the film in another tray. A top view of this system is shown in Fig. 3. The film trays consist of flat aluminum plates with vertical posts to hold the films in the proper position. The film tray on the left side in Fig. 3 holds the films to be measured, while the one on the right side holds those that have been measured. The film holder, made of black-anodized aluminum, is attached to a vertical spring-loaded stage on a vertical translation stage, which in turn is attached to a horizontal translation stage. Both translation stages have encoders on the motor shafts. The horizontal motion stage motor is geared down by a 3:1 gearbox. The film holder and translation stages move the films to the appropriate location; the sequence is described in Section 5. The film plate is black-anodized aluminum.



Fig. 3. Top View of Film Transport System

A vacuum both holds the film on the film holder and brings the film into direct contact with the opal. Refer to Fig. 4. The vacuum is applied through solenoid valves controlled by the computer. The film holder has a groove along its outer edge. When the film holder is in contact with a film and a vacuum is applied to this groove, the film is attached to the film holder and can be moved vertically and horizontally with the translation stages. There are small holes in the vacuum plate of the opal assembly, shown in cross-section in Fig. 4. When a film is on the vacuum plate and a vacuum is applied between the opal and vacuum plates, the film is pulled down onto the opal.

3.3. Detector System

The detector system collects the efflux within an acceptance cone and focuses this radiant flux onto the Si photodiode for detection. A cross section of this system is shown in Fig. 5. All the lenses have diameters of 5 cm and are made of BK-7 glass. The lenses perform the functions indicated by their designations: collecting the light within the acceptance cone, collimating it through the photopic filter, and focusing it onto the detector. A baffle with a diameter of 10 mm reduces scattered light in the detector system. The spectral transmittance of the photopic filter is such that, in combination with the spectral response of the Si photodiode, the spectral response of the detector system closely approximates the photopic spectral luminous efficiency function. The photopic filter is slightly tilted so that the light reflected from it does not



Fig. 4. Cross-Section of the Film Holder and Opal Assembly

travel back down to the opal. The lenses, baffle, and filter are housed in a black-anodized aluminum cylinder. The 4 mm by 4 mm Si photodiode is contained in a package that provides both thermoelectric temperature control at 25 °C and current-to-voltage conversion with a gain that can be selected either manually or automatically. Light reflected from the Si photodiode radiant flux onto the Si photodiode for detection. A cross section of this system is shown in Fig. 5. All the lenses have diameters of 5 cm and are made of BK-7 glass. The lenses perform the functions indicated by their designations: collecting the light within the acceptance cone, collimating it through the photopic filter, and focusing it onto the detector. A baffle with a diameter of 10 mm reduces scattered light in the detector system. The spectral transmittance of the photopic filter is such that, in combination with the spectral response of the Si photodiode, the spectral response of the detector system closely approximates the photopic spectral luminous efficiency function. The photopic filter is slightly tilted so that the light reflected from it does not travel back down to the opal. The lenses, baffle, and filter are housed in a black-anodized aluminum cylinder. The 4 mm by 4 mm Si photodiode is contained in a package that provides both thermoelectric temperature control at 25 °C and current-to-voltage conversion with a gain that can be selected either manually or automatically. Light reflected from the Si photodiode back onto the opal has no statistically significant effect on the values of visual transmission density.



Fig. 5. Cross-Section of the Detector System

The aperture stop of the optical system is the collecting lens, which defines an acceptance cone with a half-angle $\kappa = 9.5^{\circ}$ between it and the opal. If the opal is included in the optical system, it is the field stop. However, the step of a film diffuses the transmitted light, making the area from which light exits the film ill-defined. Therefore, if the opal is not included in the optical system, the photodiode is the field stop. The field of view is about 10 mm x 10 mm at the position of the opal, which is large enough to capture the entire area from which light exits the film.

3.4. System Assembly

The systems of the instrument described above - source, transport, and detector - are located on an optical table. The horizontal translation stage and film trays are attached directly to the table, as is a vertical post of 80/20 aluminum to which the lamp housing, film plate, and optic housing attach. The mounting plate of the source attaches to the film plate, as shown in Fig. 2, and the photodiode package attaches to the optic housing, as shown in Fig. 5. The solenoid valves that control the vacuum are also attached to the vertical post.

A schematic diagram showing all the components of the instrument and connections is shown in Fig. 6. A computer performs all the data acquisition and control. The GPIB interface communicates with the lamp power supply and the digital voltmeter. The shutter, detector system gain, and AC relays are controlled by TTL signals generated from a USB Digital Input/Output device (DI/O). The computer communicates with the motors through a USB-to-RS485 converter.



Fig. 6. Schematic Diagram of the Components in the Diffuse Transmission System

Most of the automatic experimental control is achieved with digital signals from the DI/O. One digital channel drives the shutter controller, which in turn opens or closes the shutter. The two

solenoid valves on the vacuum lines and the one on the exhaust line are also opened and closed using AC relays controlled by digital signals from the DI/O The gain on the photodiode amplifier is set using three of the digital signals for the gain setting and one signal to trigger the gain setting action. The remainder of the automatic experimental control is achieved by the linear translation stages, which are operated over the RS-485 bus. A diagram of the electrical connections is shown in Fig. 7.



Fig. 7. DI/O Connections to the Various Components

Separate power supplies provide the dc voltages needed by the photodiode amplifier and the motion stage motor, while a temperature controller maintains the photodiode at a constant temperature. A digital voltmeter measures the signal from the photodiode amplifier.

4. Instrument Characterization and Validation

The instrument was thoroughly characterized not only to ensure proper operation but to also verify compliance with the applicable standards for measuring transmission density [6,7,8].

These standards specify both the geometrical and spectral conditions for this measurement. The characterization is detailed for the source and detector systems and for the step tablet films.

The relative spectral flux distribution of the source, denoted by $S_{\rm H}$, is specified in [8]. This distribution depends upon the opal, infrared optical filter, and current through the lamp. With the type of opal, lamp, and filter fixed, as well as the thickness of the filter, the only adjustable parameter is the current. Therefore, the optimal current to achieve a close approximation of $S_{\rm H}$ was determined experimentally.

The relative spectral flux distribution of the source at different lamp currents was measured by a spectroradiometer in the Low-Level Radiance facility [10]. The source system was placed on its side so that the opal was imaged onto the entrance slit of the monochromator. The spectral radiant flux emerging from the exit slit of the monochromator was measured with a Si photodiode. The spectral radiance responsivity of the spectroradiometer - imaging optics, monochromator, and photodiode – was determined by measuring the signal $S_s(\lambda)$ when the output port of an integrating sphere with known radiance $L_s(\lambda)$ was imaged onto the monochromator entrance slit. The signal from the opal $S_o(\lambda)$ was then measured for different lamp currents. The radiance of the opal, $L_o(\lambda)$, was calculated using

$$L_{\rm o}(\lambda) = \frac{S_{\rm o}(\lambda)}{S_{\rm s}(\lambda)} \cdot L_{\rm s}(\lambda)$$
(32)

and normalized to have a value of 100 at 560 nm to compare with $S_{\rm H}$.

The best agreement between $S_{\rm H}$ and the measured relative spectral flux distribution was obtained with a lamp current of 7.8 A. These two distributions are shown as a function of wavelength in Fig. 8, with the distribution of the instrument denoted by $S_{\rm I}$. This distribution was relatively insensitive to changes in lamp current of 0.1 A. For greater changes in current, the measured distribution at wavelengths less than 600 nm was either obviously greater or less than $S_{\rm H}$ for increased or decreased currents, respectively. Since the lamp power supply maintains a constant current to within 1 mA, the distribution $S_{\rm I}$ is not expected to change during measurements of step tablet films.

The spectral reflectance $\rho_0(\lambda)$ is specified to be 0.55 ± 0.05 [8]. The 6° - hemispherical reflectance of flash opals of the same type used in the instrument was measured in the Spectral Tri-function Automated Reference Reflectometer (STARR) facility [11]. A flash opal with a diameter of 6 mm was mounted in the vacuum plate of the source system, which was then placed at the sample port of the integrating sphere of the reflectometer. A converging lens reduced the monochromatic incident beam diameter to approximately 4 mm. Since this diameter was larger than the 3 mm diameter of the opal used in the instrument, a 6 mm diameter opal was measured instead. The angle of incidence of the beam on the opal was 6°, and both the specular and diffuse



Fig. 8. Relative Flux Distributions

components of reflection were included in the integrating sphere. The radiant flux at the detector port of the integrating sphere was measured with a Si photodiode. At each wavelength of the incident beam, the signal from reflection from the opal $S_0(\lambda)$ and the signal from reflection from the integrating sphere wall $S_w(\lambda)$ were measured. Since the spectral reflectance of the wall $\rho_w(\lambda)$ is known, the spectral reflectance of the opal $\rho_0(\lambda)$ is calculated from

$$\rho_{\rm o}(\lambda) = \frac{S_{\rm o}(\lambda)}{S_{\rm w}(\lambda)} \cdot \rho_{\rm w}(\lambda). \tag{33}$$

The reflectance of a flash opal as a function of wavelength is shown in Fig. 9. The reflectance of all three 6 mm diameter opals cut from the same large piece were the same, and this reflectance is expected to apply to the 3 mm diameter opal used in the instrument. The reflectance decreases monotonically with wavelength and is within the values specified by the standard for wavelengths longer than 490 nm. The effect of this discrepancy between the actual opal reflectance and the standard reflectance is discussed in Section 6.



Fig. 9. 6° - Hemispherical Reflectance of the Opal as a Function of Wavelength

The diffusion coefficient d of the opal is specified to be greater than or equal to 0.9 [8]. The radiant flux from the opal was measured as a function of polar angle θ using the bi-directional reflectance goniometer of STARR. The entire source system was placed on its side so that the opal was centered in the sample holder of the goniometer and the face of the opal was on the axis of rotation of the detector arm. The lamp was operated at a current of 7.8 A, and water was circulated through the lamp housing and water filter. A green filter was placed in front of the Si photodiode on the detector arm to simulate the spectral conditions of the transmission density instrument detector system.

The signal $S(\theta)$ was measured at polar angles of the detector from -80° to $+80^{\circ}$ in 5° steps. The signals were normalized by the signal at 0°, $S(\theta)$, to yield $s(\theta)$. These normalized signals are proportional to the factor $\frac{L_j(\theta)}{L_j(0)} \cdot \cos\theta$ in Eq. (30). The normalized signal divided by $\cos\theta$ as a function of polar angle is shown in Fig. 10. For an ideal, Lambertian diffuser, this ratio is one at all angles. The decrease of this ratio as the angle increases indicates that the opal is not an ideal diffuser.



Fig. 10. Normalized Signal from the Efflux of the Opal

There are two methods for calculating the diffusion coefficient. The first is given in [7],

$$d = \frac{\sum s(\theta)}{\sum \cos\theta},\tag{34}$$

where the summation is over all the angles at which signals were measured. The second method is adapted from Eq. (2.30), which was originally derived in [9], and is given by

$$d = \frac{\sum s(\theta) \cdot \sin\theta}{\sum \cos\theta \cdot \sin\theta},$$
(35)

The diffusion coefficient calculated using Eq. (34) is d = 0.95, while using Eq. (35) d = 0.91. Both of these values are greater than 0.90.

The spectral response of the detector system, V_{T} , is specified in [7] and is given by Eq. (3), namely

$$S_{\rm A} \cdot V_{\lambda} = S_{\rm H} \cdot V_{\rm T} \,. \tag{36}$$

Here, S_A is the relative spectral flux distribution of Illuminant A, V_λ is the photopic spectral luminous efficiency function, and S_H is the relative spectral flux distribution specified in [6]. The spectral response of the detector system is the product of the spectral responsivity of the photodiode and the transmittance of the filter, as indicated in Eq. (17). The transmittance of the

lenses is not included since it is constant with wavelength in the visible spectral region for BK-7 glass.

The spectral responsivity of the Si photodiode was measured in the Spectral Comparator Facility [12]. Monochromatic, collimated light was incident on a calibrated detector with spectral responsivity $R_c(\lambda)$ yielding a signal $S_c(\lambda)$. This light was then incident on the photodiode, yielding a signal $S(\lambda)$. Assuming the same incident spectral radiant flux, the spectral responsivity of the photodiode $R(\lambda)$ is given by

$$R(\lambda) = \frac{s(\lambda)}{s_{c}(\lambda)} \cdot R_{c}(\lambda).$$
(37)

The transmittance of the filter was measured using the Reference Transmittance Spectrophotometer [13]. The signals of a Si photodiode from a collimated, monochromatic beam with and without the photopic filter present in the beam are $S_f(\lambda)$ and $S(\lambda)$, respectively. The transmittance $\tau_f(\lambda)$ of the photopic filter is then

$$\tau_{\rm f}(\lambda) = \frac{s_{\rm f}(\lambda)}{s(\lambda)}.$$
(38)

The spectral response of the detector system, normalized by its value at 550 nm and denoted by $V_{\rm I}$ is shown as a function of wavelength in Fig. 11, along with the specified spectral response $V_{\rm T}$. There is an obvious discrepancy between the two spectral responses for wavelengths longer than 570 nm. However, $V_{\rm I}$ closely approximates V_{λ} , which is expected since the filter was designed so that the spectral response obtained in combination with the Si photodiode is V_{λ} . Because the photopic spectral luminous efficiency function is widely used and no filter could be readily found to modify $V_{\rm I}$ to $V_{\rm T}$, no attempt was made to obtain $V_{\rm T}$. The effect of the discrepancy between $V_{\rm T}$ and $V_{\rm I}$ is discussed in Section 6.



Fig. 11. Spectral Response of the Detector System

To use Eq. (22) to calculate transmission density from the signals and gains, the linearity of the photodiode and the ratio of gains G_{τ}/G_{j} must be determined accurately. The linearity of the photodiode-amplifier combination was measured in the Beam conjoiner facility [14]. This facility uses the beam addition method with sets of filters on each beam path to automatically vary the radiant flux on the detector by three decades. The source system of the transmission density instrument provided the input flux. The linearity was determined at each gain setting of the amplifier, the maximum radiant flux incident on the detector being controlled by the lamp current and neutral density filters in front of the photodiode. The relative responsivity, defined as the ratio of the measured signal to the incident radiant flux as a function of current from the photodiode is shown in Fig. 12 at several gain settings. A relative responsivity of one corresponds to ideal linearity. The relative responsivity of the photodiode – amplifier combination is very close to one for gain settings 5 to 8. Deviations from one occur at gain setting 9, mostly due to the small currents at this gain setting.



Fig. 12. Relative Responsivity vs. Gain Settings

The gain setting of the amplifier is the nominal power of ten by which the current from the photodiode is multiplied to convert it to an output voltage. For example, at a gain setting of 7 the gain is 10^7 V/A. From Eq. (22), the ratios of gains are required, not the actual gains. These were determined using a constant current source in place of the photodiode as the input to the amplifier and a voltmeter to measure the output signal. At each gain setting, the current was set to obtain a 10 V signal and then reduced in ten steps to obtain 1 V and then by another ten steps to obtain 0.1 V. The output signal was measured at each current. Since this procedure yielded currents that were common for successive gain settings, the average of the ratio of the signals at the same currents was the gain ratio. For example, currents from 0.1 μ A to 1 μ A yielded voltages from 1 V to 10 V at a gain setting of 7, while the same currents resulted in voltages from 0.1 V to 1 V at a gain setting of 6. The measured gain ratios for successive gain settings are given in Table 3, where the gain settings are denoted by subscripts on the gain G. Note that the gain ratios are not exactly 10.

Gain Ratio	Value	Relative Standard Uncertainty
G_{5}/G_{4}	10.00032	4.13 x 10 ⁻⁶
G_{6}/G_{5}	10.00493	6.31 x 10 ⁻⁵
G_{7}/G_{6}	10.00311	4.62 x 10 ⁻⁵
G_8/G_7	10.00098	5.29 x 10 ⁻⁵
G_9/G_8	9.97574	1.09 x 10 ⁻⁵
G_{10}/G_{9}	10.01515	2.36 x 10 ⁻⁴

Table 3. Values of Successive Gain Ratios

The spectral transmittance and reflectance of the film step tablets are parameters in Eq. (23). These properties were therefore measured for both the x-ray and photographic step tablets at the step with the lowest transmission density. This is the step that is not exposed during processing and so is representative of the properties of the base film. The regular transmittance and directional - hemispherical reflectance were measured using a commercial Varian-Cary 5E spectrophotometer. The transmittance and reflectance as a function of wavelength are shown in Fig. 13 for both the x-ray film step tablets and the photographic film step tablets. There is spectral structure in these properties for the x-ray film, giving it a blue tint, while the neutral photographic film has no such spectral structure.



Fig. 13. Reflectance and Transmittance of the Indicated Films

Several other geometrical conditions are specified by [7] in addition to the diffusion coefficient of the opal. The half-angle of the acceptance cone must be less than 10° . From the description of the lens system in the previous section, the half-angle of the instrument is 9.5° . The film must be in contact with the opal when the efflux is measured, which is achieved by pulling the film onto the opal with a vacuum.

The transmission density is to be measured at the center of each step. Centering the steps on the opal is accomplished by optical means in the transverse direction of the steps and by mechanical means in the longitudinal direction. Note that the transverse direction of a step corresponds to the longitudinal direction of the film step tablet. The step widths are equal for each type of film and the position along the longitudinal direction of the film is continuously adjustable with the horizontal translation stage.

Therefore, all that is required to center the opal in the transverse direction of each step is to determine the location of a reference position on each film. This is done by measuring the signal from the opal as the film is moved in the longitudinal direction in increments of 0.08 mm from the first to the second step. The signal decreases at the boundary between the steps; the location at which the signal has decreased by 5 % from its value in the first step is the reference position. Knowing this position sets the locations of the centers of each step along their transverse direction. The center along the longitudinal direction is set mechanically. The positions of the horizontal translation stage and the film holder are adjusted so that the film holder is centered on the opal over its entire length. This fixes the positions of the stage and holder. The posts on the film tray are then adjusted to give the minimum possible clearance between them and the film holder. This centers the films in the tray and thereby the longitudinal direction of the steps.

The instrument was designed to measure diffuse visual transmission density according to spectral and geometrical conditions specified by the applicable standards. The conditions are satisfied by the instrument, with the exception of the opal reflectance and the spectral response. The effects of these discrepancies between the standards and the instrument are detailed in Section 6.

5. Calibration Procedure

The instrument automatically measures the transmission density of each step of an individual or batch of step tablet films. The setup for the measurements is simple. The film(s) are loaded into the film holder in the order in which they are to be measured, with the serial numbers face-down so that the correct side of the film will be in contact with the opal. The serial numbers of the films are entered in a separate text file, which is used by the program during the measurements to keep track of the films. The digital voltmeter is set to auto-zero and to use ten power line cycles for each reading. The lamp is operated at 7.8 A. After a 30 minute warm-up period, the room lights are turned off, and the measurements are begun.

For each film, the film holder is moved horizontally over the film tray and then vertically until the holder is in contact with a film. The vertical spring-loaded stage allows the film holder to remain stationary and in contact with the film while the vertical translation stage continues to move downward, preventing unnecessary torques on the holder. The vacuum is applied, attaching the film to the holder. The signal from the aperture flux is measured at a gain setting of 5 by closing the shutter, recording five readings from the voltmeter, opening the shutter, and recording five more readings. The signals measured with the shutter closed are referred to as dark signals. The holder is then translated vertically and horizontally until the first step of the film is 1.6 mm above the opal. The signal is measured to verify that a film is attached to the film holder; if not, the process is repeated until a film is attached. The boundary between the first and second steps is found, as described in the previous section. This sets the location of the horizontal translation stage so that the centers of the steps are measured.

To measure the signal from the transmitted flux of a step, the film is moved horizontally to center the step on the opal, the film is lowered until it is on top of the opal, and a vacuum is applied with the vacuum plate while the vacuum on the film holder is released. This pulls the film into close contact with the opal, and five readings from the voltmeter are recorded. A

vacuum is applied to the film holder while it is released from the vacuum plate, which attaches the film to the holder. The holder is then raised and the process is repeated for the next step. If the signal from a step is less than 1.3 V, the gain setting for the next step is increased by one and a dark signal is measured prior to measuring the signal from the transmitted flux. After the last step is measured, the film is moved over the other film tray, the vacuum on the film holder is released and air is introduced into the line, and the film falls onto the tray. The signal from the aperture flux is then measured as before.

The transmission density is calculated from the signals and ratios of the gain settings. Each set of five signal readings are averaged, and the dark signals at each gain setting are subtracted from those signals obtained at that setting to yield the net signals. The two net signals for the aperture flux are averaged to yield the final aperture signal S_y . For each step, the transmission density is calculated using Eq. (22),

$$D_{\rm T} = \log_{10} \left(\frac{s_{\rm j}}{s_{\rm \tau}} \cdot \frac{G_{\rm \tau}}{G_{\rm j}} \right), \tag{39}$$

where S_{τ} is the net signal from the transmitted flux at each step and G_{τ}/G_j is the gain ratio. The signal from the aperture flux is approximately 1.5 V on gain setting 5. For $D_T = 4$, this corresponds to a signal of 1.5 V at gain setting 9. Therefore, only gain settings 5 to 9 are used for calibrating films. The gain ratios, relative to gain setting 5, are listed in Table 4, along with the relative standard uncertainties.

Gain Ratio	Value	Relative Standard Uncertainty
G_{6}/G_{5}	1.000 493 x 10 ¹	6.31 x 10 ⁻⁵
G_{7}/G_{5}	1.000 804 x 10 ²	7.82 x 10 ⁻⁵
G_8/G_5	1.000 903 x 10 ³	9.44 x 10 ⁻⁵
G9/G5	0.998 475 x 10 ⁴	9.50 x 10 ⁻⁵

Table 4. Values of the Gain Ratios used to Calculate Transmission Density

A batch of films is measured on three separate occasions in both ascending and descending serial number order. The average of the three determinations is reported on the calibration certificate as the transmission density of a step. The standard deviation of the three determinations must be less than the expanded uncertainty due to random effects, detailed in the next section. A sample calibration certificate is given in the Appendix. Proper operation of the instrument is verified by including in each batch at least one check standard film that were measured previously. These check standard films were randomly chosen from each lot of films received from the manufacturer.

6. Uncertainties

Each measurement of transmission density has an uncertainty associated with it. This section details the components of uncertainty, their evaluation, and the resulting uncertainty in transmission density. This uncertainty analysis follows the guidelines given in [15].

In general, the purpose of a measurement is to determine the value of a measurand y, which is obtained from n other quantities x_i through the functional relationship given by

$$y = f(x_1, x_2, \dots x_i, \dots x_n).$$
 (40)

The standard uncertainty of an input quantity x_i is the estimated standard deviation associated with this quantity and is denoted by $u(x_i)$. The relative standard uncertainty is given by $u(x_i)/x_i$. The standard uncertainties may be classified either by the effect of their source or by their method of evaluation. The effects are either random or systematic, the former arising from stochastic temporal or spatial variations in the measurement and the latter from recognized effects on a measurement. The method of evaluation is either Type A, which is based on statistical analysis, or Type B, which is based on other means.

To first order, the estimated standard uncertainty u(y) in the measurand due to a standard uncertainty $u(x_i)$ is

$$u(y) = \frac{\partial f}{\partial x_{i}} u(x_{i}), \tag{41}$$

where $\partial f/\partial x_i$ is the sensitivity coefficient. The combined standard uncertainty $u_c(y)$ in the measurand is the root-sum-square of the standard uncertainties associated with each quantity x_i , assuming that these standard uncertainties are uncorrelated. The expanded uncertainty U is given by $k \cdot u_c(y)$, where k is the coverage factor and is chosen on the basis of the desired level of confidence to be associated with the interval defined by U. For the films, k = 2, which defines an interval having a level of confidence of approximately 95 %.

The components of uncertainty for transmission density are conveniently divided into those arising from the signal measurements, the uniformity of each step, and agreement with the appropriate standards [6,7]. In the first case, the appropriate measurement equation is Eq. (22), while for the last case Eqs. (23) and (31) are applicable. Expressing Eq. (22) as

$$D_{\mathrm{T}} = \log_{10}(x) \tag{42}$$

where x can be a signal or the gain ratio, the sensitivity coefficient is

$$\frac{\partial D_{\rm T}}{\partial x} = 0.434 \cdot \frac{1}{x} \,. \tag{43}$$

Therefore, the standard uncertainty $u(D_T)$ due to the standard uncertainty u(x) is

$$u(D_{\rm T}) = 0.434 \cdot \frac{u(x)}{x}.$$
 (44)

Note that the standard uncertainty of D_{T} is proportional to the relative standard uncertainty of x.

There are several components of uncertainty associated with calculating transmission density from Eq. (22), which assumes that the signals are directly proportional to the fluxes. In addition, the same influx is assumed when measuring S_j and S_{τ} . These components are the accuracy of the digital voltmeter, signal noise, lamp stability, detector linearity, and the gain ratio.

The uncertainty arising from the accuracy of the digital voltmeter is a systematic effect with a Type B evaluation and assuming a normal probability distribution. Using the manufacturer's specifications, the relative standard uncertainty of S_j is 23 x 10⁻⁶, while the maximum relative standard uncertainty of S_{τ} is 77 x 10⁻⁶. The combined relative standard uncertainty of S_j/S_{τ} is therefore 80 x 10⁻⁶ which, using Eq. (44), results in a standard uncertainty $u(D_T) \ll 0.001$.

Signal noise and lamp stability result in uncertainties from random effects with Type A evaluations. A typical relative standard deviation of the signal calculated from measurements of ten different films is 0.001, which results in a standard uncertainty $u(D_T) = 0.001$. Likewise, a typical relative standard deviation of the signal from the aperture flux monitored over a period of 15 min is 0.001.

The uncertainties arising from the detector linearity and the gain ratio are systematic effects with Type A evaluations. The signals are always between 0.1 V and 12 V at each gain setting, except for gain setting 9, for which the minimum signal is 1 V. The maximum relative standard deviation of the relative responsivity of the photodiode - amplifier combination, subject to the ranges of the signals, is 0.0028, resulting in a standard uncertainty $u(D_T) < 0.001$. The relative standard uncertainty $u(D_T) < 0.001$.

The transmission density is supposed to be determined at the center of each step. However, if the transmission density of a step is not uniform, an uncertainty in centering the opal results in an uncertainty in the measured transmission density. The standard uncertainty in centering the opal along the transverse direction of a step depends upon finding the boundary between the first and second steps of a film and knowing the spacing between the centers of the steps, and is 0.5 mm. The standard uncertainty for the longitudinal direction depends upon the mechanical adjustments of the film holder and the posts on the film tray so that the transverse direction of the film is centered on the opal along the entire length of the film. The standard uncertainty in centering in this dimension is 1 mm.

The uniformity of the transmission densities of the steps of five representative samples of each type of film was determined by measuring the transmission density every 0.17 mm along three lines running the length of each film. One line was centered along the width of the films, the other two were displaced by -1.5 mm and +1.5 mm from the center line. All the transmission densities varied by much less than 0.003 within 0.5 mm of the center in the transverse direction of the steps and by approximately 0.003 within 1 mm of the center in the longitudinal direction. Therefore, the standard uncertainty $u(D_T)$ due to the uniformity of the steps is 0.003.

As mentioned in Section 4, neither the opal reflectance nor the spectral product are in agreement with the standards specified in [5]. The uncertainties arising from these two components are systematic effects evaluated using Type B methods and a normal probability distribution. The standard uncertainties are estimated from Eq. (23) using the measured transmittance $\tau_s(\lambda)$ and reflectance $\rho_s(\lambda)$ of the films and separating the effects of the two components by assuming either an ideal spectral product or opal reflectance.

The diffusion coefficient d = 0.91 is within the range of values specified in [4]. Therefore, the standard uncertainty arising from this component is less than 0.001. However, instruments using opals with different diffusion coefficients could measure significantly different transmission densities. Using Eq. (31), the difference in transmission density ΔD_T due to different diffusion coefficients is

$$\Delta D_{\rm T} = \log_{10}(\frac{d_{\rm o}}{d}), \qquad (45)$$

where d_0 is the diffusion coefficient of the opal used in this instrument and *d* is the diffusion coefficient of another instrument. From [6], the minimum acceptable value of *d* is 0.9, while from [8] the physically maximum value is 0.95. Therefore, from Eq. (45) $\Delta D_{\rm T}$ can range from + 0.005 to - 0.019.

As shown in Fig. 9, the opal reflectance $\rho_0(\lambda)$ is greater than the acceptable maximum of 0.6 [6] for wavelengths shorter than 490 nm. The standard uncertainty $u(D_T)$ is estimated from Eq. (23) by calculating D_T for different values of $\rho_0(\lambda)$ using

$$D_{\rm T} = \log_{10} \left(\frac{\int S_{\rm H} \cdot V_{\rm T} \cdot d\lambda}{\int S_{\rm H}(\lambda) \cdot \frac{\tau_{\rm S}(\lambda)}{1 - \rho_{\rm O}(\lambda) \cdot \rho_{\rm S}(\lambda)} \cdot V_{\rm T}(\lambda) \cdot d\lambda} \right).$$
(46)

Values of D_T for different opal reflectances and both types of films are given in Table 5. Since the transmission densities calculated using the actual opal reflectance are within the ranges obtained from the acceptable opal reflectances, the standard uncertainty $u(D_T)$ associated with the opal reflectance is < 0.001.

Opal Reflectance	X-Ray	Photographic
0.50	0.173	0.121
0.55	0.172	0.118
0.60	0.171	0.117
actual	0.171	0.118

Table 5. Calculated Values of Transmission Density for Both Types of Step Films

Finally, the spectral product of the instrument $S_{\rm I} \cdot V_{\rm I}$ is not $S_{\rm H} \cdot V_{\rm T}$, but approximately $S_{\rm H} \cdot V_{\lambda}$ since the spectral response of the detector is nearly the photopic spectral luminous efficiency function. The standard uncertainty $u(D_{\rm T})$ is estimated from Eq. (33) by calculating the transmission densities obtained with the ideal and actual spectral products, using the ideal opal reflectance $\rho_0(\lambda) = 0.55$. The difference in transmission density $\Delta D_{\rm T}$ is given by

$$\Delta D_{\rm T} = \log_{10} \left(\frac{\int S_{\rm I}(\lambda) \cdot V_{\rm T}(\lambda) \cdot d\lambda}{\int S_{\rm I}(\lambda) \cdot \frac{\tau_{\rm S}(\lambda)}{1 - 0.55 \cdot \rho_{\rm S}(\lambda)} \cdot V_{\rm T}(\lambda) \cdot d\lambda} \right) - \log_{10} \left(\frac{\int S_{\rm H}(\lambda) \cdot V_{\rm T}(\lambda) \cdot d\lambda}{\int S_{\rm H}(\lambda) \cdot \frac{\tau_{\rm S}(\lambda)}{1 - 0.55 \cdot \rho_{\rm S}(\lambda)} \cdot V_{\rm T}(\lambda) \cdot d\lambda} \right).$$

$$(47)$$

Using Eq. (47), $\Delta D_{\rm T} = 0.000\ 282$ and 0.000 002 for the x-ray and photographic films, respectively. This results in a standard uncertainty $u(D_{\rm T}) < 0.001$ for both types of films. Since the reflectance and transmittance of the photographic films are relatively constant over the wavelength range for which the spectral product is appreciable, $\Delta D_{\rm T}$ is expected to be nearly zero for these films.

The standard uncertainty $u(D_T)$ resulting from each component of uncertainty is given in Table 6, along with the combined and expanded uncertainties. The only components that contribute to the combined uncertainty are the signal noise, lamp stability, detector linearity, and step non-uniformity, all of which except the detector linearity are caused by random effects. The expanded uncertainty from these random effects (i.e., neglecting the detector linearity which is systematic) is 0.005. Therefore, as mentioned at the end of the previous section, the standard deviation of the three determinations of the transmission densities of a step must be less than 0.005 to accept the average as the reported transmission density of the step. Otherwise, the instrument is assumed to have malfunctioned in some manner during one of the runs and the film is measured again. Note that even though the opal reflectance and the spectral product do not perfectly comply with those specified in the standards [7], the resulting standard uncertainty in transmission density is less than 0.001 in both cases.

Source of Uncertainty ^a	Туре	Degrees of Freedom	Uncertainty Contribution (Transmission Density)
Source Stability	А	4	0.001
Signal Noise	А	4	0.001
Amplifier Gain	А	4	< < 0.001
Detector Linearity	А	4	0.001

Table 6. Components of Uncertainty

Step non-uniformity	А	4	0.003
Spectral Product	В		$<< 0.001^{b}$
Combined Standard Uncertainty, <i>u</i> c			0.003
Expanded Uncertainty U(k=2)			0.007

^a Other sources of uncertainty including voltmeter accuracy, diffusion coefficient, and opal reflectance were determined insignificant.

^b The value shown is the equivalent value for a normal distribution.

7. Conclusions

The instrument detailed in this Special Publication measures the diffuse visual transmission density of both x-ray film and photographic film step tablets using the diffuse influx mode. It is fully automated so that single, or multiple films can be measured in one batch run. Comprehensive characterizations of the instrument were performed to ensure that it complies with the relevant international standards for these measurements. The expanded uncertainty (k = 2) for transmission density is 0.007.

References

- [1] ASTM-E1079 (2021) Standard Practice for Calibration of Transmission Densitometers. ASTM, 100 Barr Harbor Drive, West Conshohocken, PA 19428.
- [2] ASME BPVC.V (2019) Densitometer and Step Wedge Comparison Film. T-262:41.
- [3] NIST Storefront. Available at <u>https://shop.nist.gov/ccrz_CCPage?pageKey=CI</u>.
- [4] International Organization for Standardization (2022) *ISO/IEC17025 Testing and Calibration Laboratories*. Available at /web/20220331183601/https://www.iso.org/ISO-IEC-17025-testing-and-calibration-laboratories.html
- [5] McCamy CS (1966) Concepts, Terminology, and Notation for Optical Modulation. *Photogr. Sci. Eng.* 10: 314.
- [6] ISO 5-1 (1984) Photography Density Measurements Part 1: Terms, Symbols, and Notations.
- [7] ISO 5-2 (1991) Photography Density Measurements Part2: Geometric Conditions for Transmission Density.
- [8] ISO 5-3 (1984) Photography Density Measurements Part 3: Spectral Conditions.
- [9] Buhr E *et al* (1995) The Measurement of Diffuse Optical Densities. Part I: The Diffusion Coefficient. *J. Imag. Sci. and Technol.* 39:453.
- [10] Walker JH and Thompson A (1995) Spectral Radiance of a Large-Area Integrating Sphere Source. J. Res. Natl. Inst. Stand. Technol. 100:37.
- [11] Barnes PY et al (1998) NIST Measurement Services: Spectral Reflectance. Natl. Inst. Stand. Technol., Spec. Publ. 250-48.

- [12] National Institute of Standards and Technology (2019) *Spectral Comparator Facilities*. Available at /web/20220331175321/https://www.nist.gov/laboratories/toolsinstruments/spectral-comparator-facilities
- [13] National Institute of Standards and Technology (2021) Reference Transmittance Spectrophotometer. Available at /web/20220331180923/https://www.nist.gov/laboratories/tools-instruments/referencetransmittance-spectrophotometer-rts
- [14] Thompson A and Chen HM (1994) Beamcon III, a Linearity Measurement Instrument for Optical Detectors. J. Res. Natl. Inst. Stand. Technol. 99:751.
- [15] Taylor BN and Kuyatt CE (2007) Guidelines for Evaluating and Expressing the Uncertainty of NIST Measurement Results. *Natl. Inst. Stand. Technol.* Tech. Note 1297. Available at <u>https://www.nist.gov/pml/nist-technical-note-1297</u>. National Institute of Standards and Technology, Gaithersburg, MD.

Appendix A. Sample Report of Calibration

SAMPLE REPORT OF CALIBRATION

38100C Diffuse Transmittance

for

One X-Ray Step Tablet

Issued to: XXX Company Attn: Customer Name Customer Address

Issue Date: XX/XX/XXXX

1. Description of Calibration Items

The item calibrated is a double emulsion x-ray film step tablet 25.4 cm (10 in) long by 3.5 cm (1.375 in) wide with 17 steps 1.3 cm (0.5 in) wide perpendicular to the long edge of the film and with transmission densities ranging from less than 0.200 to greater than 4.000. The step tablet contains an inscribed serial number XXXXXXX.

2. Description of Calibration

The transmission density of each step of each tablet was measured using the NIST Reference Diffuse Transmittance facility (RDT) which is described in Reference 1. The instrument and measurements conform to the conditions specified for ISO Standard Diffuse Visual Transmission Density, D_T (90° Opal; S_H: $\leq 10^\circ$; V_T) in References 2 and 3.

The RDT consists of an illumination system, a film transport system, and a detector system. The illuminator provides diffuse illumination to the film step tablet with the recommended spectral flux distribution. The illumination system consists of a lamp, a filter, a shutter and an opal. The shutter assembly blocks all light from reaching the opal assembly, allowing a background, dark signals, to be taken. The film transport system transports a standard film from the film tray to the flash opal, positions the film above the flash opal in order to measure the transmission density on each step of the film and then transports the film to the film tray on the opposite side of the instrument. The flash opal is the aperture of the detector system. The detector system detects the efflux from the acceptance cone and focuses the radiant flux onto a silicon diode. The system consists of a silicon photodiode, a photopic filter, a collecting lens, a collimating lens, a focusing lens, a temperature controller, an amplifier, and a voltmeter.

measurement was 24 °C, the ambient photometer temperature was 25 °C, and the relative humidity during the measurement was 30 %.

The transmission density, $D_{\rm T}$, is calculated from the ratios of the signal and the gain. A set of five signal readings are averaged and the dark signals at each gain setting are subtracted from those signals at that gain setting to yield the net of the signals. For each step on the film, the transmission density is calculated by

$$D_{T} = Log_{10} \left(\frac{S_{j}}{S_{\tau}} \cdot \frac{G_{\tau}}{G_{j}} \right)$$

where S_j is the signal with no sample, S_{τ} is the transmitted signal, G_{τ} is the amplifier gain for the transmitted signal, and G_j is the amplifier gain for the signal with no sample.

3. Results of Calibration

The calibrated transmission densities of the step tablet are listed in Table 1. Measurements were made on a 3 mm diameter circle at the center of each step and apply only to that area. When measured, the side of the tablet with the serial number was facing away from the opal diffuser of the instrument. The expanded uncertainty (k = 2) of the reported transmission density is 0.004. Table 2 summarizes the uncertainty budget for the measurements. The uncertainty associated with the measured values is discussed thoroughly in Reference 1. Refer to Reference 4 for the NIST policy on statement of uncertainties.

The optical densities of step tablets may change with time. To minimize such changes, step tablet films should be stored in a cool, dry place where they will not be exposed to light or other radiant energy or to chemical fumes, or to dust in the air. Scratches, abrasion marks, or foreign matter on the films can change the density values. Fingerprints are a common source of contamination. Fingerprints on film surfaces can be avoided by handling step tablets only by the edges and by wearing clean gloves. Any attempt to clean a film step tablet, other than to remove dust with a soft camel-hair brush, is not recommended as it is likely to alter the calibrated values. Calibration will be nullified if a step tablet is damaged or contaminated. The laboratory shall have the necessary intermediate check standards and internal quality system to maintain confidence in the calibration status of their standard.

Tabla 1	Transmission	domaiting of y ro	v stan tablat	corrigion number	vvvvvvv
	1141151111551011	uclisities of x-ra	y step tablet,	serial number	ΛΛΛΛΛΛΛ.

Step Number	Transmission Density
1	0.131
2	0.456
3	0.678
4	0.924

5	1.247
6	1.501
7	1.677
8	1.979
9	2.190
10	2.434
11	2.703
12	2.930
13	3.191
14	3.394
15	3.660
16	3.903
17	4.131

Table 2. Uncertainty contributions and expanded uncertainty (k = 2) of the transmission density for x-ray step tablet, serial number XXXXXXX.

Source of uncertainty ^a	Туре	Degrees of Freedom	Uncertainty contribution in Transmission Density
Source Stability	А	4	0.001
Signal Noise	А	4	0.001
Amplifier Gain	А	4	<< 0.001
Detector Linearity	А	4	0.001
Step non- uniformity	A	4	0.003
Spectral Product	В		<< 0.001 ^b

Combined Standard Uncertainty, <i>u</i> c		0.003
Expanded Uncertainty		0.007
U(k=2)		

^a Other sources of uncertainty including voltmeter accuracy, diffusion coefficient, and opal reflectance were determined insignificant.

^b The value shown is the equivalent value for a normal distribution.

4. General Information

- 1) The only preparation for measurement was the removal of dust with either an air bulb or a camel hairbrush.
- 2) The values listed in Table 1 correspond only to the 3 mm diameter circle at the center of each step.
- 3) This calibration report may not be reproduced except in full without the written consent of this laboratory.

Prepared by:

Approved by:

Gerald T. Fraser
For the Director,
National Institute of
Standards and Technology
(301) 975-3797

5. References

- [1] Zarobila, C.J.; Vest, R.E.; Nadal, M.E; *NIST Measurement Services: Diffuse Visual Transmission Density*; Natl. Inst. Stand. Technol. Spec. Publ. 250-99, U.S. Government Printing Office: Washington, DC (2022).
- [2] ISO 5-2: 2001, Photography Density Measurements Part 2: Geometric Conditions for Transmission Density; International Organization for Standardization: Geneva, Switzerland (2001).
- [3] ISO 5-3: 1995, *Photography Density Measurements Part 3: Spectral Conditions*; International Organization for Standardization: Geneva, Switzerland (1995).
- [4] Taylor, B.N.; Kuyatt, C.E.; Guidelines for Evaluating and Expressing the Uncertainty of NIST Measurement Results; NIST Technical Note 1297, U.S. Government Printing Office: Washington, DC (1994).