

**NISTIR 8104-02**

# **NIST Time and Frequency Bulletin**

Karen Hansen Lichtfuss, Editor

<http://dx.doi.org/10.6028/NIST.IR.8104-02>

**NIST**  
**National Institute of  
Standards and Technology**  
U.S. Department of Commerce

**NISTIR 8104-02**

# **NIST Time and Frequency Bulletin**

*Karen Hansen Lichtfuss, Editor  
Time and Frequency Division  
Physical Measurement Laboratory*

<http://dx.doi.org/10.6028/NIST.IR.8104-02>

February 2016



U.S. Department of Commerce  
*Penny Pritzker, Secretary*

National Institute of Standards and Technology  
*Willie E. May, Under Secretary of Commerce for Standards and Technology and Director*

**NIST TIME AND FREQUENCY BULLETIN  
NIST IR 8104-02**

**No. 699 February 2016**

1. GENERAL BACKGROUND INFORMATION.....	2
2. TIME SCALE INFORMATION .....	2
3. BROADCAST OUTAGES OVER FIVE MINUTES AND WWVB PHASE PERTURBATIONS.....	4
4. NOTES ON NIST TIME SCALES AND PRIMARY STANDARDS.....	4
5. UTC (NIST) – AT1 PARAMETERS .....	5

This bulletin is published monthly. Address correspondence to:

Karen Hansen Lichtfuss, Editor  
Time and Frequency Division  
National Institute of Standards and Technology  
325 Broadway  
Boulder, CO 80305-3328  
(303) 497-3295  
Email: [karen.hansenlichtfuss@nist.gov](mailto:karen.hansenlichtfuss@nist.gov)



---

U.S. DEPARTMENT OF COMMERCE, Penny Pritzker, Secretary  
NATIONAL INSTITUTE OF STANDARDS AND TECHNOLOGY, Willie E. May, Under Secretary of Commerce for  
Standards and Technology and Director

# 1. GENERAL BACKGROUND INFORMATION

## ACRONYMS AND ABBREVIATIONS USED IN THIS BULLETIN

ACTS	- Automated Computer Time Service		
BIPM	- Bureau International des Poids et Mesures		
GPS	- Global Positioning System		
IERS	- International Earth Rotation Service		
MC	- Master Clock		
MJD	- Modified Julian Date		
NIST	- National Institute of Standards and Technology	ns	- nanosecond
SI	- International System of Units	µs	- microsecond
TA	- Atomic Time	ms	- millisecond
TAI	- International Atomic Time	s	- second
USNO	- United States Naval Observatory	min	- minute
UT1	- Universal Time (Astronomical)		
UTC	- Coordinated Universal Time		

## 2. TIME SCALE INFORMATION

The values listed below are based on data from the IERS, the USNO, and NIST. The UTC(USNO,MC) - UTC(NIST) values are averaged measurements from all available common-view GPS satellites (see bibliography on page 5). UTC - UTC(NIST) data are on page 3.

0000 HOURS COORDINATED UNIVERSAL TIME			
January 2016	MJD	UT1-UTC(NIST) (±5 ms)	UTC(USNO,MC) - UTC(NIST) (±20 ns)
7	57394	+70 ms	- 4 ns
14	57401	+56 ms	- 6 ns
21	57408	+43 ms	- 9 ns
28	57415	+31 ms	- 7 ns

The master clock pulses used by the WWV, WWVH, and WWVB time-code transmissions are referenced to the UTC (NIST) time scale. Occasionally, 1 s is added to the UTC time scale. This second is called a leap second. Its purpose is to keep the UTC time scale within ±0.9 s of the UT1 astronomical time scale, which changes slightly due to variations in the Earth's period of rotation.

**NOTE:** No leap second will be added at the end of June 2016.

Positive leap seconds, beginning at 23 h 59 min 60 s UTC and ending at 0 h 0 min 0 s UTC, were inserted in the UTC time scale on 30 June 1972, 1981-1983, 1985, 1992-1994, 1997, 2012, 2015 and on 31 December 1972-1979, 1987, 1989, 1990, 1995, 1998, 2005, 2008.

The use of leap seconds ensures that UT1 - UTC will always be held within ±0.9 s. The current value of UT1 - UTC is called the DUT1 correction. DUT1 corrections are broadcast by WWV, WWVH, WWVB, and ACTS and are printed below. These corrections may be added to the received UTC time signals in order to obtain UT1.

DUT1 = UT1 - UTC =	+0.0 s beginning 0000 UTC 31 January 2016 +0.1 s beginning 0000 UTC 26 November 2015 +0.2 s beginning 0000 UTC 11 September 2015 +0.3 s beginning 0000 UTC 01 July 2015 -0.7 s beginning 0000 UTC 28 May 2015 -0.6 s beginning 0000 UTC 19 March 2015 -0.5 s beginning 0000 UTC 25 December 2014 -0.4 s beginning 0000 UTC 25 September 2014 -0.3 s beginning 0000 UTC 08 May 2014 - 0.2 s beginning 0000 UTC 20 February 2014 - 0.1 s beginning 0000 UTC 21 November 2013 +0.0 s beginning 0000 UTC 22 August 2013 +0.1 s beginning 0000 UTC 11 April 2013 +0.2 s beginning 0000 UTC 31 January 2013 +0.3 s beginning 0000 UTC 25 October 2012 +0.4 s beginning 0000 UTC 01 July 2012
--------------------	---

The difference between UTC(NIST) and UTC has been within  $\pm 100$  ns since July 6, 1994. The table below shows values of UTC - UTC(NIST) as supplied by the BIPM in their *Circular T* publication for the most recent 310-day period in which data are available. Data are given at ten-day intervals. Five-day interval data are available in *Circular T*.

0000 Hours Coordinated Universal Time		
DATE	MJD	UTC-UTC(NIST), ns
Nov. 23, 2015	57379	3.5
Nov. 13, 2015	57369	4.3
Nov. 03, 2015	57359	6.4
Nov. 23, 2015	57349	5.8
Nov. 13, 2015	57339	4.8
Nov. 3, 2015	57329	5.3
Oct. 24, 2015	57319	9.6
Oct. 14, 2015	57309	13.9
Oct. 4, 2015	57299	16.1
Sep. 24, 2015	57289	15.8
Sep. 14, 2015	57279	14.3
Set. 4, 2015	57269	13.6
Aug. 25, 2015	57259	13.5
Aug. 15, 2105	57249	13.0
Aug. 5, 2015	57239	11.2
Jul. 26, 2015	57229	9.4
Jul. 16, 2015	57219	8.8
Jul. 6, 2015	57209	7.7
Jun. 26, 2015	57199	6.6
Jun. 16, 2015	57189	6.3
Jun. 6, 2015	57179	5.8
May 27, 2015	57169	5.6
May 17, 2015	57159	4.7
May 7, 2015	57149	4.8
Apr. 27, 2015	57139	4.4
Apr. 17, 2015	57129	4.4
Apr. 7, 2015	57119	4.5
Mar. 28, 2015	57109	4.3
Mar. 18, 2015	57099	4.3
Mar. 8, 2015	57089	4.2

### 3. BROADCAST OUTAGES OVER FIVE MINUTES AND WWVB PHASE PERTURBATIONS

OUTAGES OF 5 MINUTES OR MORE						PHASE PERTURBATIONS 2 ms			
Station	Jan. 2016	MJD	Began UTC	Ended UTC	Freq.	Jan. 2016	MJD	Began UTC	End UTC
WWVB	None					None			
WWV	None					None			
WWVH	01/09/16	57396	1426	1709	5 MHz	None			

### 4. NOTES ON NIST TIME SCALES AND PRIMARY STANDARDS

Primary frequency standards developed and operated by NIST are used to provide accuracy (rate) input to the BIPM and to provide the best possible realization of the SI second. NIST-F1 and NIST F-2, cold-atom cesium fountain frequency standards, have served as the U.S. primary standards of time and frequency since 1999. The uncertainty of NIST-F2 is currently about 1 part in  $10^{16}$ .

The AT1 scale is run in real-time by use of data from an ensemble of cesium standards and hydrogen masers. It is a free-running scale whose frequency is maintained as nearly constant as possible by choosing the optimum weight for each clock that contributes to the computation.

UTC(NIST) is generated as an offset from our real-time scale AT1. It is steered in frequency towards UTC by use of data published by the BIPM in its *Circular T*. Changes in the steering frequency will be made, if necessary, at 0000 UTC on the first day of the month, and occasionally at mid-month. A change in frequency is limited to no more than  $\pm 2$  ns/day. The frequency of UTC(NIST) is kept as stable as possible at other times.

UTC is generated at the BIPM by use of a post-processed time-scale algorithm and is not available in real-time. The parameters that we use to generate UTC(NIST) in real-time are therefore based on an extrapolation of UTC from the most recent available data.

#### References:

Allan, D.W.; Hellwig, H.; and Glaze, D.J., "An accuracy algorithm for an atomic time scale," *Metrologia*, Vol.11, No.3, pp. 133-138 (1975).

Allan, D.W.; Davis, D.D.; Weiss, M.A.; Clements, A.; Guinot, B.; Granveaud, M.; Dorenwendt, K.; Fischer, B.; Hetzel, P.; Aoki, S.; Fujimoto, M.; Charron, L.; and Ashby, N., "Accuracy of international time and frequency comparisons via global positioning system satellites in common-view," *IEEE Transactions on Instrumentation and Measurement*, Vol. IM-34, pp.118-125 (1985).

Heavner, T.P.; Jefferts, S.R.; Donley, E.A.; Shirley, J.H. and Parker, T.E., "NIST F1; recent improvements and accuracy evaluations," *Metrologia*, Vol. 42, pp. 411-422 (2005).

Jefferts, S.R.; Shirley, J.; Parker, T.E.; Heavner, T.P.; Meekhof, D.M.; Nelson, C., Levi, F.; Costanza, G.; De Marchi, A.; Drullinger, R.; Hollberg, L.; Lee, W.D.; and Walls, F.L., "Accuracy evaluation of NIST-F1," *Metrologia*, Vol. 39, pp. 321-336 (2002).

Lewandowski, W. and Thomas, C., "GPS Time transfer," *Proceedings of the IEEE*, Vol. 79, pp. 991-1000 (1991).

Parker, T.E.; Jefferts, S.R.; Heavner, T.P.; and Donley, E.A., "Operation of the NIST-F1 caesium fountain primary frequency standard with a maser ensemble, including the impact of frequency transfer noise," *Metrologia*, Vol. 42, pp. 423-430 (2005).

Weiss, M.A.; Allan, D.W., "An NBS Calibration Procedure for Providing Time and Frequency at a Remote Site by Weighting and Smoothing of GPS Common View Data," *IEEE Transactions on Instrumentation and Measurement*, Vol. IM-36, pp. 572-578 (1987).

## 5. UTC(NIST) – AT1 PARAMETERS

The table below lists parameters that are used to define UTC(NIST) with respect to our real-time scale AT1. To find the value of UTC(NIST) - AT1 at any time T (expressed as a Modified Julian Date, including a fraction if needed), the appropriate equation to use is the one for which the desired T is greater than or equal to the entry in the  $T_0$  column and less than the entry in the last column. The values of  $x_{ls}$ , x, and y for that month are then used in the equation below to find the desired value. The parameters x and y represent the offsets in time and frequency, respectively, between UTC(NIST) and AT1; the parameter  $x_{ls}$  is the number of leap seconds applied to both UTC(NIST) and UTC, as specified by the IERS. Leap seconds are not applied to AT1.

$UTC(NIST) - AT1 = x_{ls} + x + y*(T - T_0)$					
Month	$x_{ls}$ (s)	x (ns)	y (ns/d)	$T_0$ (MJD)	Valid until 0000 on: (MJD)
Mar 16	-36	-428521.05	-37.3*	57448	57451
Feb 16	-36	-427816.15	-37.3	57429	57448*
Feb 16	-36	-427556.45	-37.1	57422	57429†
Feb 16	-36	-427446.05	-36.8	57419	57422†
Jan 16	-36	-427014.25	-36.8	57408	57419
Jan 16	-36	-426313.25	-36.4	57388	57408†
Dec 15	-36	-426056.65	-36.7	57381	57388
Dec 15	-36	-425761.45	-36.9	57373	57381†
Dec 15	-36	-425285.65	-36.6	57360	57373†
Dec 15	-36	-425162.45	-36.9	57357	57360†
Nov 15	-36	-424162.45	-36.9	57332	57357
Nov 15	-36	-424070.45	-36.4	57327	57332†
Oct 15	-36	-423378.95	-36.4	57305	57327
Oct 15	-36	-422939.65	-36.6	57296	57308†
Sep 15	-36	-421832.65	-36.9	57266	57296
Aug 15	-36	-421057.75	-36.9	57245	57266
Aug 15	-36	-420686.75	-37.1	57235	57245†
Jul 15	-36	-419536.65	-37.1	57204	57235
Jun 15	-35	-418423.65	-37.1	57174	57204
May 15	-35	-417273.55	-37.1	57143	57174
Apr 15	-35	-416161.55	-37.1	57113	57143
Mar 15	-35	-415010.45	-37.1	57082	57113
Feb 15	-35	-413971.65	-37.1	57054	57082
Jan 15	-35	-412824.65	-37.0	57023	57054

† Rate change in mid-month

\*Provisional value