

**NISTIR 7880-37**

**NIST Micronutrients Measurement  
Quality Assurance Program  
Spring and Fall 1987  
Comparability Studies**

Results for Round Robins IX and XI  
Fat-Soluble Vitamins and Carotenoids in Human Serum

David L. Duewer  
Margaret C. Kline  
Willie E. May (Retired)  
Robert Schaffer (Deceased)  
Emil Schonberger (Retired)  
Jeanice B. Thomas

This publication is available free of charge from  
<https://doi.org/10.6028/NIST.IR.7880-37>



# **NIST Micronutrients Measurement Quality Assurance Program Spring and Fall 1987 Comparability Studies**

**Results for Round Robins IX and XI  
Fat-Soluble Vitamins and Carotenoids in Human Serum**

David L. Duewer  
Wille E. May (Retired)  
Robert Schaffer (Deceased)  
Emil Schonberger, Visiting Scientist (Retired)  
Jeanice B. Thomas  
*Chemical Sciences Division  
Material Measurement Laboratory*

Margaret C. Kline  
*Biomolecular Measurement Division  
Material Measurement Laboratory*

This publication is available free of charge from  
<https://doi.org/10.6028/NIST.IR.7880-37>

May 2018



U.S. Department of Commerce  
*Wilbur L. Ross, Jr., Secretary*

National Institute of Standards and Technology  
*Walter Copan, NIST Director and Under Secretary of Commerce for Standards and Technology*

(This page intentionally blank)

## **Abstract**

From 1984 to 2017, the National Institute of Standards and Technology coordinated the Micronutrients Measurement Quality Assurance Program (MMQAP) for laboratories that measure fat-soluble vitamins and carotenoids in human serum and plasma. This report describes the design of and results for the Spring and Fall 1987 MMQAP measurement comparability improvement studies: 1) Round Robin IX Fat-Soluble Vitamins and Carotenoids in Human Serum and 2) Round Robin XI Fat-Soluble Vitamins and Carotenoids in Human Serum. The first participant results for Round Robin IX were received April 22, 1987; the last results were received June 23, 1987. The first participant results for Round Robin XI were received July 8, 1987; the last results were received September 4, 1987. The analytes in Round Robin X were selenium and zinc and are not discussed in this report.

## **Keywords**

Human Serum  
Retinol,  $\alpha$ -Tocopherol,  $\beta$ -Carotene

## Table of Contents

<b>Abstract.....</b>	<b>iii</b>
<b>Keywords .....</b>	<b>iii</b>
<b>Table of Contents .....</b>	<b>iv</b>
<b>Introduction.....</b>	<b>1</b>
<b>Round Robin IX: Fat-Soluble Vitamins and Carotenoids in Human Serum.....</b>	<b>1</b>
<b>Round Robin X: Selenium and Zinc in Human Serum.....</b>	<b>1</b>
<b>Round Robin XI: Fat-Soluble Vitamins and Carotenoids in Human Serum.....</b>	<b>2</b>
<b>Reference .....</b>	<b>2</b>
<b>Appendix A. Shipping Package Inserts for RR09.....</b>	<b>A1</b>
<b>Appendix B. Final Report for RR09 .....</b>	<b>B1</b>
<b>Appendix C. Updated “All-Lab Report” for RR09 .....</b>	<b>C1</b>
<b>Appendix D. Shipping Package Inserts for RR11.....</b>	<b>D1</b>
<b>Appendix E. Final Report for RR11 .....</b>	<b>E1</b>
<b>Appendix F. Updated “All-Lab Report” for RR11.....</b>	<b>F1</b>

## Introduction

From 1984 to 2017, the National Institute of Standards and Technology (NIST), formerly the National Bureau of Standards (NBS), coordinated the Micronutrients Measurement Quality Assurance Program (MMQAP) for laboratories that measure fat-soluble vitamins and carotenoids in human serum and plasma. The MMQAP provided participants with measurement comparability assessment through use of interlaboratory studies, Standard Reference Materials (SRMs) and control materials, and methods development and validation. Serum-based samples with assigned values for the target analytes (retinol, alpha-tocopherol, gamma/beta-tocopherol, *trans*- and total beta-carotene, and ascorbic acid) and performance-evaluation standards were distributed by NIST to laboratories for analysis.

Participants used the methodology of their choice to determine analyte content in the control and study materials. Participants provided their data to NIST, where it was compiled and evaluated for trueness relative to the NIST value, within-laboratory precision, and concordance within the participant community. NIST provided the participants with a technical summary report concerning their performance for each exercise and suggestions for methods development and refinement. Participants who had concerns regarding their laboratory's performance were encouraged to consult with the MMQAP coordinators.

All MMQAP interlaboratory studies consisted of individual units of batch-prepared samples that were distributed to each participant. For historical reasons these studies are referred to as "Round Robins". The MMQAP program and the nature of its studies are described elsewhere. [1]

### **Round Robin IX: Fat-Soluble Vitamins and Carotenoids in Human Serum**

Participants in the MMQAP Fat-Soluble Vitamins and Carotenoids in Human Serum Round Robin IX comparability study (hereafter referred to as RR09) received five lyophilized human serum test samples (sera 62 to 66) for analysis and three ethanolic calibration solutions. Unless multiple vials were previously requested, participants received one vial of each test sample. These sera were shipped on dry ice to participants in April 1987. The communication materials included in the sample shipment are described in Appendix A.

Participants were requested to report values for retinol,  $\alpha$ -tocopherol, and  $\beta$ -carotene. The isomeric form of the  $\beta$ -carotene, total or *trans*, was not specified. Not all participants reported values for all target analytes.

Our records for this study are incomplete. Appendix A presents 1) a fragmentary cover letter describing the test samples and calibration solutions and 2) an example of the data report form. Appendix B reproduces the data and its summary provided to all participants. Appendix C lists the measurement results reported for RR09 in a more accessible format.

### **Round Robin X: Selenium and Zinc in Human Serum**

The only analytes in MMQAP Round Robin X were selenium and zinc; the results from this study are not included in this report. The eleven participants each received five lyophilized human serum test samples (sera 67 to 71).

## **Round Robin XI: Fat-Soluble Vitamins and Carotenoids in Human Serum**

Five lyophilized human serum test samples (sera 72 to 76) were distributed in the MMQAP Fat-Soluble Vitamins and Carotenoids in Human Serum Round Robin XI comparability study (hereafter referred to as RR11). Twenty participants received all five sera; four received sera 73, 75, and 76; and five received sera 72 and 74. Most laboratories receiving partial sets were relatively new to the MMQAP. Unless multiple vials were previously requested, participants received one vial of each material. These sample materials were shipped on dry ice to participants in July 1987. The communication materials included in the sample shipment are described in Appendix D.

Participants were requested to report values for retinol,  $\alpha$ -tocopherol, and total and  $\beta$ -carotene. The isomeric form of the  $\beta$ -carotene, total or *trans*, was not specified. Not all participants reported values for all target analytes.

Our records for this study are incomplete. Appendix D presents an example of the data report form. Appendix E reproduces a letter sent to all participants that 1) describes revisions to the statistical summary data tables, 2) the design for Round Robin XII, and 3) a proposed way to clarify  $\beta$ -carotene reporting. Appendix E also includes the (revised) data and statistical summary tables. Appendix F lists the measurement results reported for RR11 in a more accessible format.

### **Reference**

- 1 Duewer DL, Brown Thomas J, Kline MC, MacCrehan WA, Schaffer R, Sharpless KE, May WE, Crowell JA. NIST/NCI Micronutrients Measurement Quality Assurance Program: Measurement Repeatabilities and Reproducibilities for Fat-Soluble Vitamin-Related Compounds in Human Sera. *Anal Chem* 1997;69(7):1406-1413.

## **Appendix A. Shipping Package Inserts for RR09**

Two items were attached to each package shipped to an RR09 participant:

- **Cover letter.** The original letter has been lost. It would have described the three ethanolic calibration solutions (solutions 7 to 9) and five lyophilized sample materials (sera 62 to 66) distributed for the study. Page A2 reproduces what appears to be a draft message about the use of the calibration solutions.
- **Datasheet.** Page A3 reproduces the report form.

Dear \_\_\_\_\_:

This shipment contains serum samples for RR IX and calibration solutions for beta-carotene, retinol and alpha-tocopherol measurements only. We are not concerning ourselves with the other analytes in this round robin. We want to learn whether interlaboratory measurement comparability would improve for these analytes if all the laboratories in the NBS/NCI program use identical calibration solutions in their analyses.

The calibration solutions were made from chromatographically purified (but not entirely purified) solutions of retinol and beta-carotene. The solutions were mixed, and then the alpha-tocopherol was added. The less concentrated solutions are dilutions of the most concentrated one. These solutions are to be employed specifically for determining the response factors you use for the RR IX-sample analyses. The concentrations assigned to these calibration solutions are correct enough for the purpose of this round robin. They could also serve as a rough check on the calibration solutions you prepare yourself. (However, we prefer that you not consider them to be reference calibration materials--)

The serum samples are similar to those you analyzed for RR VIII. They require the addition of 1.2 mL of water for reconstitution.

After the results from RR VIII were sent to everyone, some labs were sent additional RR VIII samples to measure, for them to use to find out why some of their results were 50% or more from the consensus. They were asked to telephone or write to let us know their new results. If yours is one of the few labs that has not yet responded, please do so before you run the samples for RR IX.

## Report on NBS/NCI Samples from Laboratory #\_\_\_\_\_

Results in mg/L

Based on NBS Calibration Solutions

Samples	Result 1	Result 2
Serum No: 62      Retinol β-Carotene Analysis Date      α-Tocopherol / /87		
Serum No: 63      Retinol β-Carotene Analysis Date      α-Tocopherol / /87		
Serum No: 64      Retinol β-Carotene Analysis Date      α-Tocopherol / /87		
Serum No: 65      Retinol β-Carotene Analysis Date      α-Tocopherol / /87		
Serum No: 66      Retinol β-Carotene Analysis Date      α-Tocopherol / /87		

Note: Add 1.20 mL of water to reconstitute  
freeze-dried samples.

## **Appendix B. Final Report for RR09**

No copy of the original cover letter and discussion of results is now available. The following ten pages list the following:

- The reported results for retinol,  $\alpha$ -tocopherol, and  $\beta$ -carotene. Due to the complex formatting used in the tables, the original laboratory codes have been deleted without replacement. Appendix C provides a complete listing of the RR09 results where the original codes have been altered to ensure confidentiality. Appendix C also provides more relevant summary statistics.
- The statistical summary table and its legend.

STANDARD NBS			ANALYTE=α-TOCOPHEROL SERUM #062 GRAND AVG = 10.830; NBS=11.10				
LAB	VIAL	USED	RESULT1	RESULT2	MEAN	SD MEAN	% BIAS
	205	YES	10.400	10.000	10.200	0.200	-5.82
	195	YES	12.800	12.500	12.650	0.150	16.80
	235	YES	10.870	11.030	10.950	0.080	1.11
	217	YES	9.040	8.590	8.815	0.225	-18.61
	218	YES	9.880	10.000	9.940	0.060	-8.22
	228	YES	11.900	12.300	12.100	0.200	11.73
	150	YES	10.300	10.800	10.550	0.250	-2.59
	258	YES	9.690	9.170	9.430	0.260	-12.93
	133	YES	9.780	10.250	10.015	0.235	-7.53
	222	YES	9.900	9.600	9.750	0.150	-9.97
	214	YES	10.160	10.590	10.375	0.215	-4.20
	206	YES	14.377	13.944	14.161	0.217	30.75
	219	YES	10.270	10.400	10.335	0.065	-4.57
	131	YES	9.380	9.560	9.470	0.090	-12.56
	148	YES	10.400	10.600	10.500	0.100	-3.05
	229	YES	9.990	10.390	10.190	0.200	-5.91
	140	YES	10.149	10.049	10.099	0.050	-6.75
		YES	12.046	13.609	12.828	0.781	18.44
	116	YES	11.200	11.200	11.200	0.000	3.42
	123	YES	14.000	13.200	13.600	0.400	25.58
	246	YES	10.100	10.450	10.275	0.175	-5.13
NBS	NBS		11.100		11.100	0.000	2.49
STANDARD NBS			ANALYTE=α-TOCOPHEROL SERUM #063 GRAND AVG = 7.107; NBS=6.67				
LAB	VIAL	USED	RESULT1	RESULT2	MEAN	SD MEAN	% BIAS
	215	YES	6.700	6.400	6.550	0.150	-7.83
	23	YES	8.300	8.500	8.400	0.100	18.20
	241	YES	6.870	6.950	6.910	0.040	-2.77
	235	YES	5.910	5.690	5.800	0.110	-18.39
	245	YES	6.670	6.930	6.800	0.130	-4.32
	242	YES	8.900	7.300	8.100	0.800	13.98
	138	YES	6.600	6.800	6.700	0.100	-5.72
	253	YES	6.320	5.730	6.025	0.295	-15.22
	134	YES	6.400	6.280	6.340	0.060	-10.79
	244	YES	6.200	6.200	6.200	0.000	-12.76
	199	YES	6.480	6.340	6.410	0.070	-9.80
	218	YES	9.433	9.805	9.619	0.186	35.35
	201	YES	6.790	6.960	6.875	0.085	-3.26
	109	YES	6.080	6.210	6.145	0.065	-13.53
	131	YES	7.100	6.800	6.950	0.150	-2.21
	210	YES	6.500	6.790	6.645	0.145	-6.50
	231	YES	6.594	6.686	6.640	0.046	-6.57
		YES	7.913	7.901	7.907	0.006	11.26
	118	YES	6.900	7.000	6.950	0.050	-2.21
	146	YES	9.800	10.800	10.300	0.500	44.93
	207	YES	7.200	6.750	6.975	0.225	-1.85
NBS	NBS		6.670		6.670	0.000	-6.15
STANDARD NBS			ANALYTE=α-TOCOPHEROL SERUM #064 GRAND AVG = 4.858; NBS=4.75				
LAB	VIAL	USED	RESULT1	RESULT2	MEAN	SD MEAN	% BIAS
	199	YES	4.900	4.600	4.750	0.150	-2.23

STANDARD NBS		ANALYTE=α-TOCOPHEROL SERUM #064 GRAND AVG = 4.858; NBS=4.75				
LAB	VIAL USED	RESULT1	RESULT2	MEAN	SD MEAN	% BIAS
215	YES	5.300	6.000	5.650	0.350	16.30
237	YES	5.000	4.900	4.950	0.050	1.89
218	YES	4.140	4.070	4.105	0.035	-15.51
220	YES	4.460	4.710	4.585	0.125	-5.63
238	YES	6.000	7.600	6.800	0.800	39.97
141	YES	4.400	4.700	4.550	0.150	-6.35
228	YES	4.820	3.970	4.395	0.425	-9.54
148	YES	4.520	4.270	4.395	0.125	-9.54
211	YES	4.400	4.300	4.350	0.050	-10.46
207	YES	4.440	4.540	4.490	0.050	-7.58
229	YES	7.069	7.305	7.187	0.118	47.93
204	YES	4.660	4.770	4.715	0.055	-2.95
219	YES	3.560	3.690	3.625	0.065	-25.39
124	YES	4.900	4.800	4.850	0.050	-0.17
194	YES	4.300	4.140	4.220	0.080	-13.14
139	YES	4.499	4.690	4.595	0.096	-5.43
	YES	5.579	5.581	5.580	0.001	14.85
119	YES	4.400	4.400	4.400	0.000	-9.43
158	OUT	9.000	9.500	9.250	0.250	90.39
235	YES	5.150	4.800	4.975	0.175	2.40
NBS	NBS	4.750		4.750	0.000	-2.23
STANDARD NBS		ANALYTE=α-TOCOPHEROL SERUM #065 GRAND AVG = 9.448; NBS=9.30				
LAB	VIAL USED	RESULT1	RESULT2	MEAN	SD MEAN	% BIAS
13	YES	9.000	8.900	8.950	0.050	-5.27
244	YES	9.900	9.800	9.850	0.050	4.25
04	YES	9.870	9.960	9.915	0.045	4.94
246	YES	7.970	7.900	7.935	0.035	-16.02
253	YES	9.150	9.190	9.170	0.020	-2.94
248	YES	9.600	10.200	9.900	0.300	4.78
188	YES	8.900	9.200	9.050	0.150	-4.21
257	YES	8.380	8.150	8.265	0.115	-12.52
178	YES	8.980	8.500	8.740	0.240	-7.50
012	YES	8.700	8.400	8.550	0.150	-9.51
217	YES	10.140	9.000	9.570	0.570	1.29
006	YES	12.556	12.660	12.608	0.052	33.44
249	YES	9.120	9.190	9.155	0.035	-3.10
172	YES	8.270	7.820	8.045	0.225	-14.85
065	YES	10.100	10.000	10.050	0.050	6.37
236	YES	8.410	8.260	8.335	0.075	-11.78
166	YES	8.993	9.734	9.363	0.371	-0.90
	YES	10.395	10.376	10.386	0.010	9.92
195	YES	9.000	8.800	8.900	0.100	-5.80
131	YES	11.500	12.200	11.850	0.350	25.42
232	YES	10.050	9.600	9.825	0.225	3.99
NBS	NBS	9.300		9.300	0.000	-1.57
STANDARD NBS		ANALYTE=α-TOCOPHEROL SERUM #066 GRAND AVG = 8.849; NBS=8.07				
LAB	VIAL USED	RESULT1	RESULT2	MEAN	SD MEAN	% BIAS
240	YES	8.400	8.300	8.350	0.050	-5.63

STANDARD NBS		ANALYTE= $\alpha$ -TOCOPHEROL SERUM #066 GRAND AVG = 8.849; NBS=8.07				
LAB	VIAL USED	RESULT1	RESULT2	MEAN	SD MEAN	% BIAS
245	YES	10.600	9.400	10.000	0.600	13.01
233	YES	8.970	9.090	9.030	0.060	2.05
242	YES	6.640	7.180	6.910	0.270	-21.91
239	YES	8.550	7.700	8.125	0.425	-8.18
249	YES	10.200	9.300	9.750	0.450	10.19
178	YES	8.000	8.400	8.200	0.200	-7.33
230	YES	8.480	7.340	7.910	0.570	-10.61
136	YES	8.090	8.270	8.180	0.090	-7.56
225	YES	7.900	8.100	8.000	0.100	-9.59
218	YES	8.980	8.600	8.790	0.190	-0.66
234	YES	12.251	12.166	12.209	0.043	37.97
215	YES	8.480	8.470	8.475	0.005	-4.22
142	YES	8.230	7.640	7.935	0.295	-10.32
150	YES	9.100	9.400	9.250	0.150	4.54
205	YES	8.410	7.510	7.960	0.450	-10.04
134	YES	7.977	8.169	8.073	0.096	-8.77
		9.732	10.116	9.924	0.192	12.15
138	YES	7.900	8.000	7.950	0.050	-10.16
171	YES	12.000	11.200	11.600	0.400	31.09
228	YES	9.150	9.250	9.200	0.050	3.97
NBS	NBS	8.070		8.070	0.000	-8.80
STANDARD NBS		ANALYTE= $\beta$ -CAROTENE SERUM #062 GRAND AVG = 0.155; NBS=0.137				
LAB	VIAL USED	RESULT1	RESULT2	MEAN	SD MEAN	% BIAS
201	YES	0.146	0.143	0.145	0.002	-6.92
205	YES	0.140	0.140	0.140	0.000	-9.81
195	YES	0.213	0.196	0.205	0.008	31.74
115	YES	0.142	0.148	0.145	0.003	-6.59
217	YES	0.149	0.145	0.147	0.002	-5.31
218	YES	0.230	0.250	0.240	0.010	54.60
150	YES	0.140	0.150	0.145	0.005	-6.59
133	YES	0.170	0.170	0.170	0.000	9.51
222	YES	0.155	0.159	0.157	0.002	1.14
131	YES	0.140	0.140	0.140	0.000	-9.81
148	YES	0.080	0.060	0.070	0.010	-54.91
229	YES	0.145	0.145	0.145	0.000	-6.59
245	YES	0.240	0.230	0.235	0.005	51.38
207	YES	0.146	0.149	0.147	0.002	-4.98
209	YES	0.173	0.177	0.175	0.002	12.73
		0.138	0.149	0.144	0.005	-7.56
116	YES	0.090	0.090	0.090	0.000	-42.02
NBS	NBS	0.137		0.137	0.000	-11.75
STANDARD NBS		ANALYTE= $\beta$ -CAROTENE SERUM #063 GRAND AVG = 0.960; NBS=0.877				
LAB	VIAL USED	RESULT1	RESULT2	MEAN	SD MEAN	% BIAS
222	YES	0.986	0.904	0.945	0.041	-1.56
215	YES	0.840	0.820	0.830	0.010	-13.54

STANDARD NBS		ANALYTE=β-CAROTENE SERUM #063 GRAND AVG = 0.960; NBS=0.877					
LAB	VIAL USED	RESULT1	RESULT2	MEAN	SD MEAN	% BIAS	
223	YES	1.286	1.241	1.264	0.023	31.62	
130	YES	1.292	1.131	1.212	0.081	26.20	
224	YES	0.883	0.874	0.879	0.005	-8.49	
235	YES	0.886	0.859	0.873	0.014	-9.11	
245	YES	0.830	0.780	0.805	0.025	-16.14	
138	YES	0.890	0.910	0.900	0.010	-6.25	
134	YES	1.040	1.070	1.055	0.015	9.90	
244	YES	1.093	1.061	1.077	0.016	12.19	
109	YES	0.970	0.980	0.975	0.005	1.57	
131	OUT	0.320	0.350	0.335	0.015	-65.10	
210	YES	0.832	0.870	0.851	0.019	-11.35	
229	YES	1.380	1.180	1.280	0.100	33.34	
257	YES	0.899	1.002	0.951	0.051	-0.99	
212	YES	0.969	0.978	0.974	0.004	1.41	
	YES	0.914	0.919	0.917	0.003	-4.53	
118	YES	0.540	0.530	0.535	0.005	-44.27	
NBS	NBS	0.877		0.877	0.000	-8.64	
STANDARD NBS		ANALYTE=β-CAROTENE SERUM #064 GRAND AVG = 0.067; NBS=0.070					
LAB	VIAL USED	RESULT1	RESULT2	MEAN	SD MEAN	% BIAS	
201	YES	0.066	0.062	0.064	0.002	-4.58	
199	YES	0.070	0.070	0.070	0.000	4.37	
215	YES	0.089	0.082	0.086	0.004	27.47	
122	YES	0.044	0.050	0.047	0.003	-29.93	
138	YES	0.064	0.058	0.061	0.003	-9.05	
218	YES	0.076	0.073	0.075	0.001	11.07	
220	YES	0.080	0.080	0.080	0.000	19.27	
141	YES	0.062	0.057	0.060	0.002	-11.29	
148	YES	0.060	0.070	0.065	0.005	-3.09	
211	YES	0.071	0.074	0.073	0.002	8.09	
219	YES	0.050	0.060	0.055	0.005	-18.00	
124	YES	0.060	0.050	0.055	0.005	-18.00	
194	YES	0.051	0.062	0.057	0.006	-15.76	
195	YES	0.100	0.090	0.095	0.005	41.64	
209	YES	0.063	0.060	0.061	0.001	-8.61	
197	YES	0.095	0.097	0.096	0.001	43.13	
	YES	0.070	0.069	0.070	0.001	3.62	
119	YES	0.040	0.040	0.040	0.000	-40.36	
NBS	NBS	0.070		0.070	0.000	4.37	
STANDARD NBS		ANALYTE=β-CAROTENE SERUM #065 GRAND AVG = 0.406; NBS=0.376					
LAB	VIAL USED	RESULT1	RESULT2	MEAN	SD MEAN	% BIAS	
254	YES	0.472	0.436	0.454	0.018	11.76	
13	YES	0.390	0.390	0.390	0.000	-3.99	
244	YES	0.496	0.523	0.510	0.013	25.42	
171	YES	0.452	0.472	0.462	0.010	13.73	
186	YES	0.411	0.416	0.414	0.002	1.79	

STANDARD NBS			ANALYTE=β-CAROTENE SERUM #065 GRAND AVG = 0.406; NBS=0.376					
LAB	VIAL	USED	RESULT1	RESULT2	MEAN	SD MEAN	% BIAS	
	246	YES	0.399	0.396	0.398	0.002	-2.15	
	253	YES	0.390	0.380	0.385	0.005	-5.22	
	188	YES	0.370	0.390	0.380	0.010	-6.46	
	178	YES	0.450	0.460	0.455	0.005	12.01	
	012	YES	0.458	0.444	0.451	0.007	11.02	
	172	YES	0.420	0.420	0.420	0.000	3.39	
	065	YES	0.180	0.190	0.185	0.005	-54.46	
	236	YES	0.371	0.328	0.350	0.022	-13.96	
	256	YES	0.520	0.590	0.555	0.035	36.62	
	009	YES	0.426	0.458	0.442	0.016	8.81	
	255	YES	0.437	0.447	0.442	0.005	8.81	
		YES	0.389	0.393	0.391	0.002	-3.75	
	195	YES	0.230	0.230	0.230	0.000	-43.38	
NBS		NBS	0.376		0.376	0.000	-7.44	
STANDARD NBS			ANALYTE=β-CAROTENE SERUM #066 GRAND AVG = 1.491; NBS=1.210					
LAB	VIAL	USED	RESULT1	RESULT2	MEAN	SD MEAN	% BIAS	
	206	YES	1.708	1.564	1.636	0.072	9.73	
	240	YES	1.250	1.240	1.245	0.005	-16.49	
	245	YES	1.897	1.915	1.906	0.009	27.84	
	162	YES	1.844	1.824	1.834	0.010	23.02	
	180	YES	1.372	1.347	1.360	0.012	-8.81	
	242	YES	1.204	1.288	1.246	0.042	-16.42	
	239	YES	1.160	1.160	1.160	0.000	-22.19	
	178	YES	1.330	1.400	1.365	0.035	-8.44	
	136	YES	1.520	1.500	1.510	0.010	1.28	
	225	YES	1.436	1.555	1.496	0.060	0.31	
	142	YES	1.630	1.480	1.555	0.075	4.30	
	150	OUT	0.470	0.510	0.490	0.020	-67.13	
	205	YES	1.260	1.210	1.235	0.025	-17.16	
	209	YES	1.820	2.040	1.930	0.110	29.45	
	254	YES	1.405	1.414	1.410	0.005	-5.46	
	221	YES	1.554	1.571	1.563	0.008	4.80	
		YES	1.408	1.402	1.405	0.003	-5.76	
	138	OUT	0.550	0.530	0.540	0.010	-63.78	
NBS		NBS	1.210		1.210	0.000	-18.84	
STANDARD NBS			ANALYTE=RETINOL SERUM #062 GRAND AVG = 0.299; NBS=0.327					
LAB	VIAL	USED	RESULT1	RESULT2	MEAN	SD MEAN	% BIAS	
	211	YES	0.261	0.240	0.251	0.011	-16.14	
	205	YES	0.330	0.330	0.330	0.000	10.47	
	195	YES	0.370	0.371	0.371	0.001	24.03	
		YES	0.279	0.264	0.272	0.008	-9.11	
	230	OUT	0.060	0.070	0.065	0.005	-78.24	
	235	YES	0.329	0.331	0.330	0.001	10.47	
	217	YES	0.280	0.269	0.275	0.005	-8.11	
	218	YES	0.375	0.372	0.374	0.002	25.03	

STANDARD NBS		ANALYTE=RETINOL		SERUM #062 GRAND AVG = 0.299; NBS=0.327		
LAB	VIAL USED	RESULT1	RESULT2	MEAN	SD MEAN	% BIAS
228	YES	0.275	0.250	0.263	0.013	-12.13
150	YES	0.250	0.280	0.265	0.015	-11.29
258	YES	0.260	0.260	0.260	0.000	-12.96
133	YES	0.250	0.250	0.250	0.000	-16.31
222	YES	0.260	0.260	0.260	0.000	-12.96
214	YES	0.225	0.227	0.226	0.001	-24.35
206	YES	0.374	0.376	0.375	0.001	25.53
131	YES	0.310	0.310	0.310	0.000	3.77
148	YES	0.340	0.360	0.350	0.010	17.16
229	YES	0.282	0.337	0.310	0.027	3.61
106	OUT	0.500	0.490	0.495	0.005	65.70
207	YES	0.321	0.330	0.326	0.005	8.96
140	YES	0.272	0.273	0.273	0.000	-8.78
	YES	0.376	0.375	0.376	0.001	25.70
116	YES	0.250	0.250	0.250	0.000	-16.31
123	YES	0.290	0.270	0.280	0.010	-6.27
NBS	NBS	0.327	0.327	0.000	9.46	
STANDARD NBS		ANALYTE=RETINOL		SERUM #063 GRAND AVG = 0.856; NBS=0.910		
LAB	VIAL USED	RESULT1	RESULT2	MEAN	SD MEAN	% BIAS
234	YES	0.667	0.680	0.674	0.007	-21.32
215	YES	0.940	0.950	0.945	0.005	10.39
223	YES	1.081	1.035	1.058	0.023	23.59
130	YES	0.842	0.841	0.842	0.001	-1.70
256	OUT	0.420	0.400	0.410	0.010	-52.10
241	YES	0.948	0.951	0.950	0.001	10.92
235	YES	0.795	0.770	0.783	0.012	-8.59
245	YES	0.863	0.899	0.881	0.018	2.92
242	YES	0.870	0.905	0.888	0.017	3.68
138	YES	0.790	0.830	0.810	0.020	-5.38
253	YES	0.800	0.800	0.800	0.000	-6.54
134	YES	0.700	0.700	0.700	0.000	-18.23
244	YES	0.760	0.780	0.770	0.010	-10.05
199	YES	0.861	0.888	0.875	0.014	2.16
218	YES	0.947	0.904	0.926	0.022	8.12
109	YES	0.960	0.970	0.965	0.005	12.73
131	YES	0.980	1.000	0.990	0.010	15.65
210	YES	0.865	0.908	0.887	0.022	3.56
150	OUT	1.400	1.410	1.405	0.005	64.13
257	YES	0.872	0.906	0.889	0.017	3.85
231	YES	0.763	0.763	0.763	0.000	-10.87
	YES	0.996	0.985	0.991	0.006	15.71
118	YES	0.720	0.730	0.725	0.005	-15.31
146	YES	0.730	0.720	0.725	0.005	-15.31
NBS	NBS	0.910	0.910	0.000	6.31	
STANDARD NBS		ANALYTE=RETINOL		SERUM #064 GRAND AVG = 0.540; NBS=0.597		
LAB	VIAL USED	RESULT1	RESULT2	MEAN	SD MEAN	% BIAS
212	YES	0.547	0.561	0.554	0.007	2.54

STANDARD NBS		ANALYTE=RETINOL		SERUM #064 GRAND AVG = 0.540; NBS=0.597		
LAB	VIAL USED	RESULT1	RESULT2	MEAN	SD MEAN	% BIAS
199	YES	0.620	0.620	0.620	0.000	14.75
215	YES	0.679	0.656	0.668	0.012	23.55
122	YES	0.544	0.559	0.552	0.007	2.08
205	YES	0.390	0.400	0.395	0.005	-26.89
237	YES	0.558	0.563	0.561	0.003	3.74
218	YES	0.510	0.489	0.500	0.010	-7.55
220	YES	0.609	0.642	0.626	0.016	15.77
238	YES	0.440	0.485	0.463	0.023	-14.40
141	YES	0.510	0.540	0.525	0.015	-2.83
228	YES	0.490	0.500	0.495	0.005	-8.38
148	YES	0.490	0.480	0.485	0.005	-10.23
211	YES	0.460	0.510	0.485	0.025	-10.23
207	YES	0.567	0.569	0.568	0.001	5.13
229	YES	0.597	0.593	0.595	0.002	10.13
219	YES	0.560	0.590	0.575	0.015	6.43
124	YES	0.690	0.670	0.680	0.010	25.86
194	YES	0.579	0.592	0.586	0.007	8.37
133	OUT	0.930	0.930	0.930	0.000	72.13
209	YES	0.590	0.550	0.570	0.020	5.50
139	YES	0.498	0.500	0.499	0.001	-7.64
	YES	0.639	0.687	0.663	0.024	22.71
119	YES	0.440	0.450	0.445	0.005	-17.64
158	YES	0.260	0.380	0.320	0.060	-40.77
NBS	NBS	0.597	0.597	0.597	0.000	10.50
STANDARD NBS		ANALYTE=RETINOL		SERUM #065 GRAND AVG = 0.440; NBS=0.479		
LAB	VIAL USED	RESULT1	RESULT2	MEAN	SD MEAN	% BIAS
251	YES	0.438	0.404	0.421	0.017	-4.38
13	YES	0.490	0.490	0.490	0.000	11.29
244	YES	0.499	0.529	0.514	0.015	16.74
171	YES	0.448	0.394	0.421	0.027	-4.38
226	YES	0.310	0.300	0.305	0.005	-30.73
04	YES	0.493	0.524	0.508	0.016	15.49
246	YES	0.416	0.407	0.412	0.005	-6.54
253	YES	0.490	0.510	0.500	0.010	13.56
248	YES	0.395	0.395	0.395	0.000	-10.29
188	YES	0.410	0.430	0.420	0.010	-4.61
257	YES	0.410	0.410	0.410	0.000	-6.88
178	YES	0.420	0.410	0.415	0.005	-5.75
012	YES	0.390	0.390	0.390	0.000	-11.42
217	YES	0.419	0.392	0.406	0.014	-7.90
006	YES	0.500	0.503	0.502	0.002	13.90
172	YES	0.480	0.460	0.470	0.010	6.74
065	YES	0.550	0.540	0.545	0.005	23.78
236	YES	0.473	0.474	0.474	0.000	7.54
132	OUT	0.750	0.750	0.750	0.000	70.34
009	YES	0.492	0.485	0.489	0.004	10.95

STANDARD NBS		ANALYTE=RETINOL		SERUM #065 GRAND AVG = 0.440; NBS=0.479		
LAB	VIAL USED	RESULT1	RESULT2	MEAN	SD MEAN	% BIAS
166	YES	0.397	0.383	0.390	0.007	-11.42
	YES	0.519	0.515	0.517	0.002	17.42
195	YES	0.360	0.360	0.360	0.000	-18.24
131	YES	0.400	0.350	0.375	0.025	-14.83
NBS	NBS	0.479		0.479	0.000	8.79
STANDARD NBS		ANALYTE=RETINOL		SERUM #066 GRAND AVG = 0.775; NBS=0.811		
LAB	VIAL USED	RESULT1	RESULT2	MEAN	SD MEAN	% BIAS
247	YES	0.661	0.698	0.680	0.019	-12.36
240	YES	0.850	0.850	0.850	0.000	9.63
245	YES	0.893	0.897	0.895	0.002	15.43
162	YES	0.770	0.752	0.761	0.009	-1.85
210	OUT	0.240	0.250	0.245	0.005	-68.40
233	YES	0.827	0.847	0.837	0.010	7.95
242	YES	0.639	0.696	0.668	0.028	-13.91
239	YES	0.770	0.776	0.773	0.003	-0.30
249	YES	0.645	0.640	0.643	0.003	-17.13
178	YES	0.700	0.740	0.720	0.020	-7.14
230	YES	0.800	0.710	0.755	0.045	-2.62
136	YES	0.640	0.660	0.650	0.010	-16.17
225	YES	0.710	0.670	0.690	0.020	-11.01
218	YES	0.854	0.876	0.865	0.011	11.56
234	YES	0.830	0.812	0.821	0.009	5.89
142	YES	0.910	0.880	0.895	0.015	15.43
150	YES	0.950	0.990	0.970	0.020	25.11
205	YES	0.805	0.750	0.778	0.028	0.28
143	OUT	1.260	1.260	1.260	0.000	62.51
254	YES	0.823	0.775	0.799	0.024	3.05
134	YES	0.695	0.673	0.684	0.011	-11.78
	YES	0.962	1.029	0.996	0.033	28.40
138	YES	0.650	0.650	0.650	0.000	-16.17
171	YES	0.660	0.700	0.680	0.020	-12.30
NBS	NBS	0.811		0.811	0.000	4.60

**Summary of RR IX Results  
"Fat Soluble Vitamins in Serum"**

**STANDARD NBS****ANALYTE= $\alpha$ -TOCOPHEROL**

SER	NBS	GRAND AVG	S.E. AVG	S WITHIN	S BETWEEN	% CV AVG	% CV X	FR REJECT
064	4.75	4.858	0.193	0.331	0.831	4.0	18.4	1/21
063	6.67	7.107	0.252	0.339	1.130	3.5	16.6	0/21
066	8.07	8.849	0.279	0.410	1.244	3.2	14.8	0/21
065	9.30	9.448	0.253	0.296	1.139	2.7	12.5	0/21
062	11.1	10.830	0.311	0.356	1.403	2.9	13.4	0/21

**STANDARD NBS****ANALYTE= $\beta$ -CAROTENE**

SER	NBS	GRAND AVG	S.E. AVG	S WITHIN	S BETWEEN	% CV AVG	% CV X	FR REJECT
064	0.070	0.067	0.004	0.004	0.015	5.3	23.1	0/18
062	0.137	0.155	0.010	0.007	0.043	6.7	27.7	0/17
065	0.376	0.406	0.021	0.018	0.087	5.1	21.8	0/18
063	0.877	0.960	0.044	0.052	0.179	4.6	19.4	1/18
066	1.210	1.491	0.060	0.062	0.235	4.0	16.3	2/18

**STANDARD NBS****ANALYTE=RETINOL**

SER	NBS	GRAND AVG	S.E. AVG	S WITHIN	S BETWEEN	% CV AVG	% CV X	FR REJECT
062	0.327	0.299	0.010	0.012	0.047	3.4	16.3	2/24
065	0.479	0.440	0.013	0.015	0.060	2.9	14.1	1/24
064	0.597	0.540	0.018	0.025	0.086	3.4	16.6	1/24
066	0.811	0.775	0.022	0.027	0.104	2.9	13.8	2/24
063	0.910	0.856	0.022	0.018	0.105	2.6	12.4	2/24

#### **Explanation of Column Heading on Statistical Summary**

GRAND AVG = grand average (consensus value from analysis of variance) =  $\bar{x}$  (mg/L)

S.E.AVG = one standard deviation (std. dev.) of  $\bar{x}$  =  $S_{\bar{x}}$

S WITHIN = within laboratory component of std. dev. =  $S_w$  (square root of within laboratory component of variance from the ANOVA; a pooled estimate of within laboratory imprecision)

S BETWEEN = between laboratory component of standard deviation =  $S_b$

% CV AVG = percent coefficient of variation of the grand average =  $100 S_{\bar{x}}/\bar{x}$

% CV X = percent coefficient of variation of a single measurement made by a single laboratory =  $100 S_x/\bar{x}$ , where  $S_x = \sqrt{S_w^2 + S_b^2}$  (NOTE:  $S_x$  is not equal to the std. dev. of all measurements)

FR REJECT = ratio of rejected to total laboratory results considered for statistical evaluation

## **Appendix C. Updated “All-Lab Report” for RR09**

The following three pages are modernized “All-Lab” report for RR09. This report has three parts:

- Page 1 lists results for all analytes reported.
- Page 2 provides the legend for page 1.
- Page 3 summarizes each participants’ performance for retinol,  $\alpha$ -tocopherol, and  $\beta$ -carotene, using the “Comparability Summary” calculations used from the 1999 to 2017 Round Robins.

To ensure confidentiality, the laboratory identifiers used in this “All-Lab Report” have been altered from those used in RR09. The only attributed results are those reported by NIST. The NIST results are not used in the assessment of the consensus summary results of the study.

## Round Robin IX Laboratory Results

Lab	Total Retinol, µg/mL					α-Tocopherol, µg/mL					Total β-Carotene, µg/mL				
	62	63	64	65	66	62	63	64	65	66	62	63	64	65	66
FSV-BA	0.273	0.763	0.499	0.390	0.684	10.10	6.64	4.59	9.36	8.07	0.175	0.974	0.096	0.442	1.563
FSV-BD	0.263	0.888	0.463	0.395	0.643	12.10	8.10	6.80	9.90	9.75	0.157	1.077	0.073	0.451	1.496
FSV-BF	0.260	0.770	0.485	0.390	0.690	9.75	6.20	4.35	8.55	8.00	0.205	1.264	0.086	0.510	1.906
FSV-BG	0.371	1.058	0.668	0.514	0.895	12.65	8.40	5.65	9.85	10.00	0.140	0.830	0.070	0.390	1.245
FSV-BI	0.330	0.945	0.620	0.490	0.850	10.20	6.55	4.75	8.95	8.35	0.090	0.535	0.040	0.230	0.540
FSV-BX	0.250	0.725	0.445	0.360	0.650	11.20	6.95	4.40	8.90	7.95	0.145	0.945	0.064	0.454	1.636
FSV-BY	0.251	0.674	0.554	0.421	0.680	13.60	10.30	9.25	11.85	11.60	0.145	0.879	0.061	0.414	1.360
FSV-BZ	0.280	0.725	0.320	0.375	0.680	10.95	6.91	4.95	9.92	9.03	0.144	0.917	0.070	0.391	1.405
FSV-CA	0.330	0.950	0.561	0.509	0.837	11.30	7.43	5.20	10.08	9.32	0.170	1.055	0.065	0.455	1.510
FSV-CJ	0.290	0.956	0.642	0.438	1.251	12.83	7.91	5.58	10.39	9.92	0.147	0.873	0.075	0.398	1.246
FSV-CK	0.376	0.991	0.663	0.517	0.996	10.02	6.34	4.40	8.74	8.18	0.235	1.280	0.095	0.555	1.930
FSV-CL	0.250	0.700	0.485	0.415	0.650	8.82	5.80	4.11	7.94	6.91	0.226	0.945	0.064	0.454	1.636
FSV-CO	0.275	0.783	0.500	0.412	0.668	10.34	6.88	4.72	9.16	8.48	0.140	0.975	0.055	0.420	1.555
FSV-DC	0.495	1.405	0.930	0.750	1.260	9.47	6.15	3.63	8.05	7.94	0.145	0.900	0.060	0.380	1.365
FSV-DE						9.43	6.03	4.40	8.27	7.91	0.133	0.721	0.059	0.358	1.222
FSV-DG	0.310	0.965	0.575	0.470	0.895	10.55	6.70	4.55	9.05	8.20	0.120	1.212	0.047	0.462	1.834
FSV-DH	0.260	0.800	0.495	0.410	0.755	10.38	6.41	4.49	9.57	8.79	0.148	0.951	0.061	0.442	1.410
FSV-DN	0.265	0.810	0.525	0.420	0.720	13.75	8.51	5.41	11.35	9.75	0.226	0.945	0.064	0.454	1.636
FSV-DO	0.065	0.410	0.395	0.305	0.245	9.80	5.63	4.07	8.47	7.73	0.226	0.945	0.064	0.454	1.636
FSV-DT	0.226	0.875	0.568	0.406	0.865	14.16	9.62	7.19	12.61	12.21	0.240	0.805	0.080	0.385	1.160
FSV-DZ	0.391	1.200	0.728	0.590	0.935	9.94	6.80	4.59	9.17	8.13	0.070	0.335	0.055	0.185	0.490
FSV-ED	0.319	0.872	0.590	0.458	0.812	10.50	6.95	4.85	10.05	9.25	0.145	0.851	0.057	0.350	1.235
FSV-EG	0.272	0.842	0.550	0.421	0.765	10.19	6.65	4.22	8.34	7.96	0.128	0.981	0.061	0.442	1.410
FSV-EO	0.326	0.889	0.570	0.489	0.799	10.28	6.98	4.98	9.83	9.20	0.145	0.851	0.057	0.350	1.235
n	27	27	27	27	27	24	24	24	24	24	19	19	19	19	19
Min	0.065	0.410	0.320	0.305	0.245	8.82	5.63	3.63	7.94	6.91	0.070	0.335	0.040	0.185	0.490
Median	0.290	0.881	0.568	0.438	0.778	10.36	6.84	4.65	9.27	8.41	0.145	0.917	0.064	0.414	1.405
Max	0.495	1.405	0.930	0.750	1.260	14.16	10.30	9.25	12.61	12.21	0.240	1.280	0.096	0.555	1.930
eSD	0.059	0.125	0.102	0.071	0.145	0.89	0.81	0.46	1.01	0.83	0.018	0.129	0.013	0.056	0.237
eCV	20	14	18	16	19	9	12	10	11	10	12	14	20	13	17
NISTa	0.332	0.900	0.558	0.501	0.781	10.01	6.37	4.29	8.46	7.53	0.129	0.793	0.070	0.389	1.104
NISTb	0.349	0.923	0.613	0.499	0.832	12.26	7.11	5.15	10.07	8.45	0.154	1.000	0.074	0.412	1.410
NNIST	4	4	4	4	4	4	4	4	4	4	4	4	4	4	4
Mean <sub>NIST</sub>	0.340	0.912	0.585	0.500	0.806	11.14	6.74	4.72	9.27	7.99	0.141	0.898	0.072	0.389	1.258
NAV	0.315	0.896	0.577	0.469	0.792	10.75	6.79	4.69	9.27	8.20	0.143	0.907	0.068	0.401	1.331
NAU	0.069	0.126	0.103	0.084	0.146	1.69	0.81	0.61	1.15	0.88	0.023	0.158	0.014	0.091	0.259

## Round Robin IX Laboratory Results

### Analytes Reported By One Laboratory Values in $\mu\text{g/mL}$

Analyte	Code	62	63	64	65	66
Total cis- $\beta$ -Carotene	NISTa	0.013	0.094	0.002	0.035	0.110
trans- $\beta$ -Carotene	NISTa	0.141	0.906	0.072	0.377	1.300

Table Legend <sup>a</sup>

Term	Definition
n	Number of (non-NIST) quantitative values reported for this analyte
Min	Minimum (non-NIST) quantitative value reported
Median	Median (non-NIST) quantitative value reported
Max	Maximum (non-NIST) quantitative value reported
eSD	Adjusted median absolute deviation from the median of the non-NIST results
eCV	Coefficient of Variation for (non-NIST) results: $100 \times \text{SD}/\text{Median}$
$N_{\text{NIST}}$	Number of results reported by NIST (NBS) analysts
$\text{Mean}_{\text{NIST}}$	Mean of NIST results
NAV	NIST Assigned Value: $(\text{Median} + N_{\text{NIST}})/2$
NAU	NIST Assigned Uncertainty: $\sqrt{(S^2 + S_{\text{btw}}^2)}$ S is the maximum of $(0.05 \times \text{NAV}, \text{eSD}, \text{SD}_{\text{LT}})$ and $S_{\text{btw}}$ is the standard deviation between Median and $\text{Mean}_{\text{NIST}}$ . The expected long-term SD, $\text{SD}_{\text{LT}}$ , is defined in: Duewer et al., Anal Chem 1997;69(7):1406-1413.

<sup>a</sup> What is now "NIST" was the National Bureau of Standards (NBS) when this study was conducted.

## Round Robin IX Laboratory Results

### Comparability Summary

Lab	TR	aT	bC
FSV-BA	1	1	2
FSV-BD	1	2	
FSV-BF	1	1	1
FSV-BG	2	2	2
FSV-BI	1	1	1
FSV-BX	1	1	3
FSV-BY	1		1
FSV-BZ	2	4	
FSV-CA	1	1	
FSV-CJ	2	1	1
FSV-CK	2	2	1
FSV-CL	1	1	1
FSV-CO	1	2	1
FSV-DC	4		3
FSV-DE		1	
FSV-DG	1	2	1
FSV-DH	1	1	
FSV-DN	1	1	1
FSV-DO	3		
FSV-DT	1	1	
FSV-DZ	2	2	
FSV-ED	1	1	1
FSV-EG	1		2
FSV-EO	1		1
FSV-EP	1	4	
FSV-ER	1	1	3
FSV-EU	2	1	3
FSV-EW	1	1	1
FSV-FI		1	
NISTa	1	1	1
NISTb	1	1	1
n	29	26	21

	TR	aT	bC
% 1	72	69	67
% 2	21	23	14
% 3	3	0	19
% 4	3	8	0

Label	Definition
Lab	Participant code
TR	Total Retinol
aT	$\alpha$ -Tocopherol
bC	$\beta$ -Carotene
n	Number of values
n	number of participants providing quantitative data
% 1	Percent of CS = 1 (within 1 SD of medians)
% 2	Percent of CS = 2 (within 2 SD of medians)
% 3	Percent of CS = 3 (within 3 SD of medians)
% 4	Percent of CS = 4 (3 or more SD from medians)

“Comparability Score”

The Comparability Score (CS) summarizes your measurement performance for a given analyte relative to the consensus medians in this study. CS is the average distance (in units of standard deviation) of your measurement performance characteristics from the consensus performance. CS is calculated when the number of quantitative values you reported,  $N_{you}$ , is at least two and at least six participants reported quantitative values for the analyte.

We define CS as follows:

$$CS = \text{MINIMUM}\left(4, \text{INTEGER}\left(1 + \sqrt{C^2 + AP^2}\right)\right)$$

$$C = \text{Concordance} = \frac{\sum_{i=1}^{N_{you}} \frac{You_i - \text{Median}_i}{NAU_i}}{N_{you}}$$

$$AP = \text{Apparent Precision} = \sqrt{\frac{\sum_{i=1}^{N_{you}} \left(\frac{You_i - \text{Median}_i}{NAU_i}\right)^2}{N_{you} - 1}}$$

NAU = NIST Assigned Uncertainty

For further details, please see

Duewer DL, Kline MC, Sharpless KE, Brown Thomas J, Gary KT. Micronutrients Measurement Quality Assurance Program: Helping participants use interlaboratory comparison exercise results to improve their long-term measurement performance. Anal Chem 1999;71(9):1870-8.

## **Appendix D. Shipping Package Inserts for RR11**

Two items were attached to each package shipped to an RR11 participant:

- **Cover letter.** The original letter as attached to the packages has been lost.
- **Datasheet.** Page D2 reproduces the form.

## REPORT ON NBS/NCI SAMPLES FROM LABORATORY #\_\_\_\_\_

DATE OF ANALYSIS \_\_\_\_\_

RESULTS IN mg/L

	SAMPLES	RESULT 1	RESULT 2
Serum 72	RETINOL		
VIAL #_____	B-CAROTENE		
	A-TOCOPHEROL		
Serum 73	RETINOL		
VIAL #_____	B-CAROTENE		
	A-TOCOPHEROL		
Serum 74	RETINOL		
VIAL #_____	B-CAROTENE		
	A-TOCOPHEROL		
Serum 75	RETINOL		
VIAL #_____	B-CAROTENE		
	A-TOCOPHEROL		
Serum 76	RETINOL		
VIAL #_____	B-CAROTENE		
	A-TOCOPHEROL		

## **Appendix E. Final Report for RR11**

The following 14 pages are the available fragments of the report for RR11 provided to participants. These fragments are:

- An individualized letter sent two months after the original letter that informed the participant of revisions to the summary table for  $\beta$ -carotene, identified a mechanism for chromatographically distinguishing total and *trans*- $\beta$ -carotene, and discussed plans for the next study, Round Robin XII.
- The reported results for retinol,  $\alpha$ -tocopherol, and  $\beta$ -carotene. Due to the complex formatting used in the tables, the original laboratory codes have been deleted without replacement. Appendix F provides a complete listing of the RR11 results where the original codes have been altered to ensure confidentiality. Appendix F also provides more relevant summary statistics.
- The statistical summary table and its legend.



UNITED STATES DEPARTMENT OF COMMERCE  
National Bureau of Standards  
Gaithersburg, Maryland 20899

December 29, 1987

Individualized  
name and address

Dear First Name:

Enclosed is a revision of the statistical summary for RR-XI mailed to you in November. The table for beta-carotene had errors that were introduced while rearranging the individual sample results in the order of increasing concentration. The tables for retinol and alpha-tocopherol were o.k.

Round-Robin XII will consist of proficiency testing samples for the fat-soluble vitamins, ascorbic acid, and the metals (zinc and selenium). The samples should be shipped on January 11, 1988.


The samples for fat-soluble vitamin analyses are freeze-dried plasma. Data entry forms are provided and please note that these samples are to be reconstituted with 1.00 mL of water and that you are to report only one value for each sample. Results must arrive at NBS by February 29, 1988 to be included in our statistical summaries.

In the past we have discussed the dilemma that we face in evaluating beta-carotene results. NBS and a few other labs feel that they have been reporting beta-carotene data based on the concentration of the all-trans isomer only. The majority of labs measure a single peak that contains both the all-trans and the cis-isomers of beta-carotene. In order to facilitate the interlaboratory comparison of data, we request that all laboratories report a value for total beta-carotene and that labs also report concentrations for all trans beta-carotene where possible. Laboratories that use "polymerically-bound" ODS columns should report data for both all-trans and total beta-carotene. Total beta-carotene should be determined by summing the all-trans and cis-isomers. A chromatogram representative of the analysis of an NBS sample on a "polymerically-bound" column is provided in figure 1. Laboratories that use "monomerically-bound" ODS columns (the vast majority) should, as they have in the past, report only total beta-carotene concentrations. Figure 2 is a chromatogram representative of the analysis of an NBS sample on a "monomerically-bound" column. Please feel free to contact me (301/975-3108) or Neal Craft (301/975-3111) if you have questions.

The samples for ascorbic acid analyses and those for analysis of zinc and selenium are similar to those sent previously. The metal samples will be shipped as frozen liquids. The ascorbic acid samples are freeze-dried and should be reconstituted with 1.00 mL of water. The reporting format and data forms are as before, i.e. results for each sample should be reported in duplicate. Results for the samples must arrive at NBS by March 15, 1988 to be included in our statistical summaries.

Dr. Robert Schaffer will retire within a few weeks. He will, however, continue to serve as Clinical Coordinator for this program. You should continue to communicate and interact with him, as in the past, through his NBS address and telephone number. However, official correspondence from NBS to you will come from me.

Sincerely,



Willie E. May, Ph.D.  
Chief  
Organic Analytical Research Division  
Center for Analytical Chemistry

Enclosures

cc: R. Schaffer  
M. Kline

## POLYMERICALLY-BOUND ODS COLUMN





Fig 2

# MONOMERICALLY-BOUND ODS COLUMN



STANDARD LAB		ANALYTE=α-TOCOPHEROL SERUM #72 GRAND AVG = 6.840; NBS = 6.76				
LAB	VIAL USED	RESULT1	RESULT2	MEAN	SD MEAN	% BIAS
48	YES	6.500	6.400	6.450	0.050	-5.70
16	YES	9.000	8.800	8.900	0.100	30.12
34	YES	6.920	6.780	6.850	0.070	0.15
46	YES	6.640	6.630	6.635	0.005	-3.00
02	YES	6.200	5.530	5.865	0.335	-14.25
26	YES	6.700	6.600	6.650	0.050	-2.78
14	OUT	11.400	11.800	11.600	0.200	69.59
28	YES	6.690	6.490	6.590	0.100	-3.65
06	YES	7.100	7.400	7.250	0.150	5.99
21	YES	6.860	6.500	6.680	0.180	-2.34
33	YES	6.250	6.360	6.305	0.055	-7.82
23	YES	6.290	6.240	6.265	0.025	-8.41
13	YES	6.050	6.570	6.310	0.260	-7.75
12	OUT	8.550	4.710	6.630	1.920	-3.07
11	YES	6.500	5.800	6.150	0.350	-10.09
04	NFI	5.476		5.476	0.000	-19.94
27	YES	7.400	7.300	7.350	0.050	7.46
15	YES	8.800	7.900	8.350	0.450	22.08

STANDARD LAB		ANALYTE=α-TOCOPHEROL SERUM #73 GRAND AVG = 11.241; NBS = 10.87				
LAB	VIAL USED	RESULT1	RESULT2	MEAN	SD MEAN	% BIAS
08	YES	10.200	10.200	10.200	0.000	-9.26
29	YES	14.300	15.900	15.100	0.800	34.33
26	YES	11.640	11.780	11.710	0.070	4.17
18	YES	10.460	10.790	10.625	0.165	-5.48
08	YES	9.230	9.610	9.420	0.190	-16.20
28	YES	11.700	11.100	11.400	0.300	1.41
35	OUT	18.100	19.200	18.650	0.550	65.91
20	YES	12.000	11.200	11.600	0.400	3.19
27	YES	12.990	11.870	12.430	0.560	10.57
24	YES	13.000	12.500	12.750	0.250	13.42
22	YES	12.500	13.400	12.950	0.450	15.20
04	YES	10.800	10.780	10.790	0.010	-4.01
32	YES	10.390	10.390	10.390	0.000	-7.57
06	YES	10.500	10.900	10.700	0.200	-4.82
41	YES	9.420	10.130	9.775	0.355	-13.04
49	OUT	6.000	6.330	6.165	0.165	-45.16
10	YES	11.380	10.240	10.810	0.570	-3.84
23	NFI	9.193		9.193	0.000	-18.22
14	YES	10.156	9.848	10.002	0.154	-11.02
27	YES	10.400	10.500	10.450	0.050	-7.04

STANDARD LAB		ANALYTE=α-TOCOPHEROL SERUM #74 GRAND AVG = 9.239; NBS = 9.36				
LAB	VIAL USED	RESULT1	RESULT2	MEAN	SD MEAN	% BIAS
46	YES	8.200	8.400	8.300	0.100	-10.16

STANDARD LAB ANALYTE= $\alpha$ -TOCOPHEROL SERUM #74 GRAND AVG = 9.239; NBS = 9.36  
 LAB VIAL USED RESULT1 RESULT2 MEAN SD MEAN % BIAS

21	YES	11.000	12.300	11.650	0.650	26.10
29	YES	9.330	9.290	9.310	0.020	0.77
35	YES	8.870	8.930	8.900	0.030	-3.67
42	YES	7.640	7.590	7.615	0.025	-17.57
31	YES	10.100	10.400	10.250	0.150	10.95
15	OUT	14.500	15.600	15.050	0.550	62.90
10	YES	9.430	9.880	9.655	0.225	4.51
43	YES	10.500	11.200	10.850	0.350	17.44
34	YES	6.900	8.300	7.600	0.700	-17.74
40	YES	8.710	9.120	8.915	0.205	-3.50
48	YES	8.430	8.360	8.395	0.035	-9.13
19	YES	7.450	8.050	7.750	0.300	-16.11
18	OUT	5.470	5.600	5.535	0.065	-40.09
32	YES	8.430	8.350	8.390	0.040	-9.19
45	NFI	7.483		7.483	0.000	-19.00
20	YES	10.100	10.000	10.050	0.050	8.78
26	YES	9.900	12.000	10.950	1.050	18.52

STANDARD LAB ANALYTE= $\alpha$ -TOCOPHEROL SERUM #75 GRAND AVG = 8.869; NBS = 8.98  
 LAB VIAL USED RESULT1 RESULT2 MEAN SD MEAN % BIAS

04	YES	8.600	8.500	8.550	0.050	-3.60
01	YES	11.900	12.900	12.400	0.500	39.81
10	YES	9.060	9.610	9.335	0.275	5.25
03	YES	8.810	8.490	8.650	0.160	-2.47
14	YES	7.700	7.510	7.605	0.095	-14.25
11	YES	8.600	9.400	9.000	0.400	1.47
18	OUT	14.800	14.700	14.750	0.050	66.30
23	YES	8.800	8.600	8.700	0.100	-1.91
22	YES	12.810	12.070	12.440	0.370	40.26
17	YES	10.200	10.500	10.350	0.150	16.70
09	YES	8.620	9.280	8.950	0.330	0.91
06	YES	8.550	8.320	8.435	0.115	-4.90
24	YES	8.340	8.480	8.410	0.070	-5.18
02	YES	8.000	7.900	7.950	0.050	-10.36
25	YES	7.590	8.190	7.890	0.300	-11.04
39	YES	6.060	5.600	5.830	0.230	-34.27
04	YES	9.390	8.670	9.030	0.360	1.81
07	NFI	7.651		7.651	0.000	-13.74
08	YES	7.618	7.625	7.622	0.003	-14.07
13	YES	8.500	8.500	8.500	0.000	-4.16

STANDARD LAB ANALYTE= $\alpha$ -TOCOPHEROL SERUM #76 GRAND AVG = 7.073; NBS = 7.25  
 LAB VIAL USED RESULT1 RESULT2 MEAN SD MEAN % BIAS

11	YES	7.000	6.800	6.900	0.100	-2.45
----	-----	-------	-------	-------	-------	-------

STANDARD LAB			ANALYTE= $\alpha$ -TOCOPHEROL SERUM #76 GRAND AVG = 7.073; NBS = 7.25					
LAB	VIAL	USED	RESULT1	RESULT2	MEAN	SD	MEAN	% BIAS
	42	YES	8.700	9.500	9.100	0.400		28.65
	30	YES	7.550	7.330	7.440	0.110		5.18
	35	YES	7.260	7.070	7.165	0.095		1.30
	32	YES	6.480	6.190	6.335	0.145		-10.44
	28	YES	7.200	5.900	6.550	0.650		-7.40
	41	OUT	11.800	12.500	12.150	0.350		71.77
	45	YES	7.600	7.100	7.350	0.250		3.91
	27	YES	9.030	9.480	9.255	0.225		30.84
	49	YES	8.200	8.700	8.450	0.250		19.46
	43	YES	6.200	6.630	6.415	0.215		-9.31
	36	YES	7.330	7.270	7.300	0.030		3.20
	23	YES	6.810	6.760	6.785	0.025		-4.08
	38	YES	6.800	6.900	6.850	0.050		-3.16
	25	YES	6.260	6.790	6.525	0.265		-7.75
	12	YES	4.400	4.910	4.655	0.255		-34.19
	44	YES	7.000	6.630	6.815	0.185		-3.65
	37	NFI	6.427		6.427	0.000		-9.14
	39	YES	6.641	6.518	6.580	0.061		-6.98
	34	YES	6.900	6.800	6.850	0.050		-3.16

STANDARD LAB			ANALYTE= $\beta$ -CAROTENE SERUM #72 GRAND AVG = 0.953; NBS = 0.864					
LAB	VIAL	USED	RESULT1	RESULT2	MEAN	SD	MEAN	% BIAS
015	10	NFI	0.897		0.897	0.000		-5.84
	48	YES	0.870	0.850	0.860	0.010		-9.72
	16	YES	1.078	1.004	1.041	0.037		9.28
	17	YES	0.756	0.752	0.754	0.002		-20.85
	38	YES	1.032	1.006	1.019	0.013		6.97
	46	YES	1.021	1.025	1.023	0.002		7.39
	02	YES	0.842	0.815	0.829	0.014		-13.03
	14	OUT	2.440	2.320	2.380	0.060		149.84
	06	YES	0.819	1.156	0.988	0.169		3.66
	13	YES	0.960	1.020	0.990	0.030		3.92
	11	YES	0.894	0.860	0.877	0.017		-7.94
	22	YES	0.977	0.977	0.977	0.000		2.56
	43	YES	1.104	1.196	1.150	0.046		20.72
	04	YES	0.917	0.932	0.925	0.008		-2.95
	27	OUT	1.600	1.600	1.600	0.000		67.96

STANDARD LAB			ANALYTE= $\beta$ -CAROTENE SERUM #73 GRAND AVG = 0.629; NBS = 0.554					
LAB	VIAL	USED	RESULT1	RESULT2	MEAN	SD	MEAN	% BIAS
	09	NFI	0.570		0.570	0.000		-9.35
	08	YES	0.560	0.540	0.550	0.010		-12.53

STANDARD LAB		ANALYTE=β-CAROTENE		SERUM #73	GRAND AVG = 0.629; NBS = 0.554
LAB	VIAL USED	RESULT1	RESULT2	MEAN	SD MEAN % BIAS
29	YES	0.678	0.675	0.677	0.001 7.59
11	YES	0.499	0.474	0.487	0.013 -22.63
19	YES	0.649	0.629	0.639	0.010 1.62
18	YES	0.693	0.695	0.694	0.001 10.37
08	YES	0.536	0.543	0.540	0.003 -14.20
35	OUT	1.660	1.560	1.610	0.050 156.05
20	YES	0.660	0.680	0.670	0.010 6.55
24	YES	0.580	0.709	0.645	0.064 2.50
41	YES	0.610	0.640	0.625	0.015 -0.60
10	YES	0.686	0.596	0.641	0.045 1.94
36	YES	0.723	0.737	0.730	0.007 16.10
23	YES	0.632	0.573	0.603	0.030 -4.18
14	YES	0.566	0.543	0.555	0.011 -11.81
27	YES	0.750	0.750	0.750	0.000 19.28

STANDARD LAB		ANALYTE=β-CAROTENE		SERUM #74	GRAND AVG = 0.418; NBS = 0.359
LAB	VIAL USED	RESULT1	RESULT2	MEAN	SD MEAN % BIAS
22	NFI	0.382		0.382	0.000 -8.69
46	YES	0.360	0.350	0.355	0.005 -15.15
21	YES	0.465	0.460	0.463	0.002 10.55
37	YES	0.318	0.312	0.315	0.003 -24.71
16	YES	0.444	0.439	0.442	0.002 5.53
35	YES	0.466	0.468	0.467	0.001 11.62
42	YES	0.359	0.351	0.355	0.004 -15.15
15	OUT	1.150	0.960	1.055	0.095 152.17
43	YES	0.434	0.518	0.476	0.042 13.77
19	YES	0.330	0.350	0.340	0.010 -18.73
32	YES	0.428	0.400	0.414	0.014 -1.05
25	YES	0.508	0.526	0.517	0.009 23.57
27	YES	0.423	0.470	0.447	0.024 6.72
45	YES	0.419	0.443	0.431	0.012 3.02
20	OUT	0.700	0.700	0.700	0.000 67.31

STANDARD LAB		ANALYTE=β-CAROTENE		SERUM #75	GRAND AVG = 0.083; NBS = 0.066
LAB	VIAL USED	RESULT1	RESULT2	MEAN	SD MEAN % BIAS
27	NFI	0.078		0.078	0.000 -5.54
04	YES	0.080	0.070	0.075	0.005 -9.17
01	YES	0.084	0.087	0.086	0.001 3.55
12	YES	0.042	0.038	0.040	0.002 -51.56
19	YES	0.090	0.087	0.089	0.002 7.18
03	YES	0.096	0.093	0.095	0.002 14.45
14	YES	0.049	0.059	0.054	0.005 -34.60

STANDARD LAB		ANALYTE=β-CAROTENE		SERUM #75 GRAND AVG = 0.083; NBS = 0.066		
LAB	VIAL USED	RESULT1	RESULT2	MEAN	SD MEAN	% BIAS
18	OUT	0.190	0.200	0.195	0.005	136.16
23	YES	0.100	0.090	0.095	0.005	15.05
17	YES	0.076	0.079	0.078	0.002	-6.14
25	YES	0.080	0.080	0.080	0.000	-3.11
04	YES	0.095	0.076	0.086	0.009	3.55
16	YES	0.114	0.112	0.113	0.001	36.85
07	YES	0.082	0.085	0.084	0.002	1.12
08	YES	0.060	0.068	0.064	0.004	-22.49
13	YES	0.120	0.120	0.120	0.000	45.33

STANDARD LAB		ANALYTE=β-CAROTENE		SERUM #76 GRAND AVG = 1.192; NBS = 1.01		
LAB	VIAL USED	RESULT1	RESULT2	MEAN	SD MEAN	% BIAS
33	NFI	1.084		1.084	0.000	-9.09
11	YES	1.020	1.020	1.020	0.000	-14.46
42	YES	1.265	1.353	1.309	0.044	9.78
31	YES	1.161	1.125	1.143	0.018	-4.14
47	YES	1.079	1.066	1.073	0.007	-10.05
35	YES	1.239	1.221	1.230	0.009	3.15
32	YES	1.010	1.010	1.010	0.000	-15.30
41	OUT	3.000	3.120	3.060	0.060	156.63
45	YES	1.420	1.400	1.410	0.010	18.25
49	YES	1.332	1.487	1.410	0.077	18.21
25	YES	1.110	1.180	1.145	0.035	-3.97
44	YES	1.160	1.040	1.100	0.060	-7.75
29	YES	1.330	1.332	1.331	0.001	11.62
37	YES	1.191	1.208	1.200	0.008	0.60
39	YES	1.035	1.013	1.024	0.011	-14.12
34	YES	1.310	1.270	1.290	0.020	8.19

STANDARD LAB		ANALYTE=RETINOL		SERUM #72 GRAND AVG = 0.874; NBS = 0.916		
LAB	VIAL USED	RESULT1	RESULT2	MEAN	SD MEAN	% BIAS
10	NFI	0.713		0.713	0.000	-18.39
48	YES	0.920	0.900	0.910	0.010	4.16
16	YES	0.912	0.882	0.897	0.015	2.67
17	YES	1.107	1.110	1.109	0.002	26.88
38	YES	1.035	1.042	1.038	0.004	18.86
46	YES	0.811	0.847	0.829	0.018	-5.11
02	YES	0.863	0.773	0.818	0.045	-6.37
26	YES	0.873	0.919	0.896	0.023	2.55
14	YES	0.600	0.620	0.610	0.010	-30.18
28	YES	0.920	0.850	0.885	0.035	1.30
06	YES	0.880	1.050	0.965	0.085	10.45

STANDARD LAB LAB	LAB VIAL USED	ANALYTE=RETINOL RESULT1	SERUM #72 RESULT2	MEAN	SD MEAN	GRAND AVG = 0.874; NBS = 0.916 % BIAS
21	YES	0.877	0.867	0.872	0.005	-0.19
33	YES	0.930	0.970	0.950	0.020	8.73
13	YES	0.840	0.800	0.820	0.020	-6.14
12	YES	0.930	0.980	0.955	0.025	9.31
11	YES	0.802	0.710	0.756	0.046	-13.47
22	YES	0.944	0.942	0.943	0.001	7.93
43	YES	0.998	1.086	1.042	0.044	19.27
04	NFI	0.785		0.785	0.000	-10.15
27	YES	0.630	0.650	0.640	0.010	-26.75
15	YES	0.670	0.660	0.665	0.005	-23.89

STANDARD LAB LAB	LAB VIAL USED	ANALYTE=RETINOL RESULT1	SERUM #73 RESULT2	MEAN	SD MEAN	GRAND AVG = 0.956; NBS = 0.967 % BIAS
09	NFI	0.840		0.840	0.000	-12.17
08	YES	1.010	1.020	1.015	0.005	6.12
29	YES	0.962	0.990	0.976	0.014	2.05
11	YES	1.260	1.198	1.229	0.031	28.50
19	YES	1.132	1.119	1.126	0.007	17.68
18	YES	0.904	0.909	0.907	0.003	-5.22
08	YES	0.824	0.831	0.828	0.003	-13.48
28	YES	0.972	0.963	0.968	0.005	1.16
35	YES	0.610	0.640	0.625	0.015	-34.65
20	YES	1.040	1.040	1.040	0.000	8.74
27	YES	0.990	0.940	0.965	0.025	0.90
24	YES	0.970	1.080	1.025	0.055	7.17
22	YES	0.897	0.841	0.869	0.028	-9.14
04	YES	0.960	0.940	0.950	0.010	-0.67
06	YES	0.934	0.922	0.928	0.006	-2.97
41	YES	0.920	0.870	0.895	0.025	-6.42
49	YES	0.990	0.980	0.985	0.005	2.99
10	YES	0.957	0.883	0.920	0.037	-3.81
36	YES	0.994	0.992	0.993	0.001	3.82
23	NFI	0.840		0.840	0.000	-12.17
14	YES	1.000	0.955	0.978	0.022	2.20
27	YES	0.906	0.912	0.909	0.003	-4.96

STANDARD LAB LAB	LAB VIAL USED	ANALYTE=RETINOL RESULT1	SERUM #74 RESULT2	MEAN	SD MEAN	GRAND AVG = 0.450; NBS = 0.469 % BIAS
22	NFI	0.410		0.410	0.000	-8.88
46	YES	0.450	0.440	0.445	0.005	-1.10
21	YES	0.494	0.486	0.490	0.004	8.90
37	YES	0.609	0.614	0.612	0.003	35.90

STANDARD LAB LAB	LAB VIAL USED	ANALYTE=RETINOL RESULT1	SERUM #74 RESULT2	MEAN	SD MEAN	GRAND AVG = 0.450; NBS = 0.469 % BIAS
16	YES	0.516	0.517	0.517	0.001	14.79
35	YES	0.442	0.466	0.454	0.012	0.90
42	YES	0.419	0.410	0.415	0.005	-7.88
31	YES	0.432	0.455	0.444	0.011	-1.43
15	YES	0.270	0.310	0.290	0.020	-35.55
10	YES	0.450	0.430	0.440	0.010	-2.21
43	YES	0.490	0.550	0.520	0.030	15.57
34	YES	0.500	0.444	0.472	0.028	4.90
40	YES	0.430	0.480	0.455	0.025	1.12
19	YES	0.380	0.360	0.370	0.010	-17.77
18	YES	0.560	0.480	0.520	0.040	15.57
32	YES	0.445	0.410	0.428	0.017	-4.99
25	YES	0.427	0.427	0.427	0.000	-5.10
27	YES	0.448	0.517	0.483	0.035	7.23
45	NFI	0.382		0.382	0.000	-15.10
20	YES	0.330	0.330	0.330	0.000	-26.66
26	YES	0.580	0.300	0.440	0.140	-2.21

STANDARD LAB LAB	LAB VIAL USED	ANALYTE=RETINOL RESULT1	SERUM #75 RESULT2	MEAN	SD MEAN	GRAND AVG = 0.395; NBS = 0.393 % BIAS
27	NFI	0.361		0.361	0.000	-8.55
04	YES	0.400	0.420	0.410	0.010	3.86
01	YES	0.406	0.398	0.402	0.004	1.84
12	YES	0.510	0.507	0.508	0.002	28.82
19	YES	0.444	0.453	0.449	0.005	13.62
03	YES	0.365	0.383	0.374	0.009	-5.26
14	YES	0.362	0.367	0.364	0.003	-7.66
11	YES	0.354	0.359	0.357	0.003	-9.69
18	YES	0.240	0.240	0.240	0.000	-39.20
23	YES	0.420	0.440	0.430	0.010	8.93
22	YES	0.390	0.380	0.385	0.005	-2.47
17	YES	0.430	0.450	0.440	0.010	11.46
09	YES	0.438	0.418	0.428	0.010	8.42
06	YES	0.400	0.420	0.410	0.010	3.86
02	YES	0.400	0.409	0.405	0.004	2.47
25	YES	0.360	0.350	0.355	0.005	-10.07
39	YES	0.420	0.390	0.405	0.015	2.60
04	YES	0.375	0.347	0.361	0.014	-8.55
16	YES	0.407	0.411	0.409	0.002	3.61
07	NFI	0.350		0.350	0.000	-11.34
08	YES	0.390	0.393	0.392	0.002	-0.82
13	YES	0.372	0.372	0.372	0.000	-5.76

STANDARD	LAB	ANALYTE=RETINOL	SERUM #76	GRAND AVG =	0.446; NBS = 0.437	
LAB	VIAL USED	RESULT1	RESULT2	MEAN	SD MEAN	% BIAS
33	NFI	0.405		0.405	0.000	-9.18
11	YES	0.480	0.470	0.475	0.005	6.52
42	YES	0.437	0.473	0.455	0.018	2.04
31	YES	0.648	0.656	0.652	0.004	46.21
47	YES	0.506	0.505	0.506	0.000	13.36
35	YES	0.424	0.433	0.429	0.005	-3.91
32	YES	0.422	0.403	0.413	0.009	-7.50
28	YES	0.410	0.356	0.383	0.027	-14.11
41	YES	0.260	0.270	0.265	0.005	-40.57
45	YES	0.480	0.470	0.475	0.005	6.52
27	YES	0.460	0.440	0.450	0.010	0.91
49	YES	0.500	0.520	0.510	0.010	14.37
43	YES	0.482	0.483	0.483	0.001	8.20
36	YES	0.410	0.420	0.415	0.005	-6.94
38	YES	0.414	0.411	0.413	0.002	-7.50
25	YES	0.420	0.400	0.410	0.010	-8.06
12	YES	0.470	0.470	0.470	0.000	5.40
44	YES	0.410	0.372	0.391	0.019	-12.32
29	YES	0.458	0.458	0.458	0.000	2.71
37	NFI	0.371		0.371	0.000	-16.80
39	YES	0.460	0.451	0.456	0.005	2.15
34	YES	0.414	0.411	0.413	0.002	-7.50

Round-Robin XI

STANDARD LAB		ANALYTE= $\alpha$ -TOCOPHEROL									
SER	NBS	GRAND AVG	S.E.	AVG	S WITHIN	S BETWEEN	% CV	AVG	% CV X	FR	REJECT
72	6.76	6.840	0.214	0.282	0.804	3.1	12.4	2/17			
76	7.25	7.073	0.249	0.339	1.028	3.5	15.3	1/19			
75	8.98	8.869	0.373	0.349	1.563	4.2	18.1	1/19			
74	9.36	9.239	0.333	0.561	1.227	3.6	14.6	2/17			
73	10.87	11.241	0.345	0.492	1.379	3.1	13.0	2/19			

STANDARD LAB		ANALYTE= $\beta$ -CAROTENE									
SER	NBS	GRAND AVG	S.E.	AVG	S WITHIN	S BETWEEN	% CV	AVG	% CV X	FR	REJECT
75	0.066	0.083	0.006	0.005	0.021	6.8	25.9	1/15			
76	0.359	1.192	0.038	0.045	0.137	3.2	12.1	1/15			
74	0.554	0.418	0.018	0.022	0.061	4.4	15.5	2/14			
73	0.864	0.629	0.020	0.034	0.072	3.2	12.6	1/15			
72	1.01	0.953	0.031	0.075	0.095	3.3	12.7	2/14			

STANDARD LAB		ANALYTE=RETINOL									
SER	NBS	GRAND AVG	S.E.	AVG	S WITHIN	S BETWEEN	% CV	AVG	% CV X	FR	REJECT
75	0.393	0.395	0.012	0.011	0.051	2.9	13.3	0/20			
76	0.437	0.446	0.016	0.014	0.072	3.6	16.4	0/20			
74	0.469	0.450	0.016	0.052	0.061	3.7	17.9	0/19			
72	0.916	0.874	0.031	0.043	0.132	3.5	15.8	0/19			
73	0.967	0.956	0.026	0.029	0.116	2.8	12.6	0/20			

#### **Explanation of Column Heading on Statistical Summary**

GRAND AVG = grand average (consensus value from analysis of variance) =  $\bar{x}$  (mg/L)

S.E.AVG = one standard deviation (std. dev.) of  $\bar{x}$  =  $S_{\bar{x}}$

S WITHIN = within laboratory component of std. dev. =  $S_w$  (square root of within laboratory component of variance from the ANOVA; a pooled estimate of within laboratory imprecision)

S BETWEEN = between laboratory component of standard deviation =  $S_b$

% CV AVG = percent coefficient of variation of the grand average =  $100 S_{\bar{x}} / \bar{x}$

% CV X = percent coefficient of variation of a single measurement made by a single laboratory =  $100 S_x / \bar{x}$ , where  $S_x = \sqrt{S_w^2 + S_b^2}$  (NOTE:  $S_x$  is not equal to the std. dev. of all measurements)

FR REJECT = ratio of rejected to total laboratory results considered for statistical evaluation

## **Appendix F. Updated “All-Lab Report” for RR11**

The following three pages are an updated version of an “All-Lab” report for RR11. This report has three parts:

- Page 1 lists results for all analytes.
- Page 2 provides a legend for page 1.
- Page 3 summarizes each participants’ performance for retinol,  $\alpha$ -tocopherol, and  $\beta$ -carotene, using the “Comparability Summary” calculations used from the 1999 to 2017 Round Robins.

To ensure confidentiality, the laboratory identifiers used in this “All-Lab Report” have been altered from those used in RR11. The only attributed results are those reported by NIST. The NIST results are not used in the assessment of the consensus summary results of the study.

## Round Robin XI Laboratory Results

Lab	Total Retinol, µg/mL					α-Tocopherol, µg/mL					β-Carotene, µg/mL				
	72	73	74	75	76	72	73	74	75	76	72	73	74	75	76
FSV-BA	0.785	0.840	0.382	0.350	0.371	5.48	9.19	7.48	7.65	6.43	0.925	0.603	0.431	0.084	1.200
FSV-BD	0.896	0.968	0.444	0.357	0.383	6.65	11.40	10.25	9.00	6.55	0.988	0.645	0.476	0.078	1.410
FSV-BE		0.928		0.405	0.413		10.70		7.95	6.85	1.041	0.677	0.463	0.086	1.309
FSV-BF	0.965	1.025	0.520	0.440	0.510	7.25	12.75	10.85	10.35	8.45	0.909	0.750	0.120	1.290	
FSV-BG	0.897	0.976	0.490	0.402	0.455	8.90	15.10	11.65	12.40	9.10	0.860	0.550	0.355	0.075	1.020
FSV-BH		0.909		0.372	0.413		10.45		8.50	6.85					
FSV-BI	0.910	1.015	0.445	0.410	0.475	6.45	10.20	8.30	8.55	6.90	0.897	0.570	0.382	0.078	1.084
FSV-BX	0.640		0.330			7.35		10.05			1.600		0.700		
FSV-BY	0.713	0.840	0.410	0.361	0.405						0.897				
FSV-BZ	0.665		0.440			8.35		10.95			1.019		0.639		
FSV-CJ	1.039	1.126	0.517	0.449	0.506	6.85	11.71	9.31	9.34	7.44	0.863		0.377		
FSV-CK	0.787		0.426			6.00		8.64			1.023		0.694		
FSV-CN		1.040		0.430	0.475		11.60		8.70	7.35	0.977		0.730		
FSV-CO	0.829	0.907	0.454	0.374	0.429	6.64	10.63	8.90	8.65	7.17	0.670		0.095		
FSV-DC	0.943	0.993	0.427	0.409	0.458						1.023		0.467		
FSV-DE						6.27	10.39	8.40	8.41	6.79	0.977		0.517		
FSV-DG	0.820	0.895	0.370	0.355	0.410	6.31	9.78	7.75	7.89	6.53	0.610		0.340		
FSV-DH	0.885	0.965	0.440	0.385	0.450	6.59	12.43	9.66	12.44	9.26	0.885		0.440		
FSV-DN	0.610	0.625	0.290	0.240	0.265	11.60	18.65	15.05	14.75	12.15	0.610		1.055		
FSV-DO	1.175	1.515	0.710	0.655	0.735						2.380		1.610		
FSV-DT	0.872	0.869	0.472	0.428	0.483	6.68	12.95	7.60	8.95	6.42	0.872		0.315		
FSV-ED		0.978		0.392	0.456		10.00		7.62	6.58	0.978		0.447		
FSV-EG	1.109	1.229	0.612	0.508	0.652						0.754		0.555		
FSV-EO	1.042		0.483			6.31	10.79	8.92	8.44	7.30	0.754		0.064		
FSV-EP	0.950	0.950	0.455	0.410	0.415	5.87	9.42	7.62	7.61	6.34	0.818		0.355		
FSV-ER						6.63	6.17	5.54	5.83	4.66	0.955		0.054		
FSV-EU	0.955	0.985	0.520	0.405	0.470	6.15	10.81	8.39	9.03	6.82	0.756		0.414		
FSV-EW	0.756	0.920	0.428	0.361	0.391	6.70		9.39			0.877		0.086		
FSV-FI											0.877		1.100		
n	23	22	23	22	22	20	20	20	20	20	16	16	16	16	16
Min	0.610	0.625	0.290	0.240	0.265	5.48	6.17	5.54	5.83	4.66	0.754	0.487	0.315	0.040	1.010
Median	0.885	0.958	0.444	0.397	0.439	6.63	10.75	8.91	8.60	6.85	0.982	0.640	0.436	0.085	1.171
Max	1.175	1.515	0.710	0.655	0.735	11.60	18.65	15.05	14.75	12.15	2.380	1.610	1.055	0.195	3.060
eSD	0.119	0.087	0.050	0.044	0.056	0.52	1.35	1.70	1.01	0.64	0.141	0.092	0.084	0.014	0.190
eCV	13	9	11	11	13	8	13	19	12	9	14	14	19	17	16
N <sub>RR</sub>	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1
N <sub>past</sub>	27	27	27	27	27	24	23	24	23	23	19	20	19	20	20
Median <sub>past</sub>	0.881	0.900	0.438	0.388	0.430	6.84	10.27	9.27	8.36	7.00	0.917	0.618	0.414	0.089	1.053
SD <sub>past</sub>	0.125	0.133	0.071	0.044	0.052	0.81	1.56	1.01	1.26	1.02	0.129	0.132	0.056	0.033	0.285
NIST	0.970	0.880	0.430	0.380	0.390	6.30	10.39	8.26	8.70	6.87	0.830	0.570	0.360	0.080	1.020
NAV	0.928	0.919	0.437	0.389	0.415	6.47	10.57	8.58	8.65	6.86	0.906	0.605	0.398	0.082	1.096
NAU	0.133	0.103	0.051	0.046	0.066	0.61	1.37	1.77	1.01	0.64	0.178	0.104	0.100	0.015	0.218

## Round Robin XI Laboratory Results

Table Legend <sup>a</sup>

Term	Definition
n	Number of (non-NIST) quantitative values reported for this analyte
Min	Minimum (non-NIST) quantitative value reported
Median	Median (non-NIST) quantitative value reported
Max	Maximum (non-NIST) quantitative value reported
eSD	Estimated standard deviation, calculated from the median absolute deviation from the median of the non-NIST results
eCV	Coefficient of Variation for (non-NIST) results: 100*eSD/Median
$N_{RR}$	Number of prior RR studies that distributed this serum
$N_{past}$	Maximum number of (non-NIST) quantitative values reported for this analyte in prior studies
$\text{Median}_{past}$	Mean of the median values from prior studies
$SD_{past}$	Pooled eSD of prior studies
NAV	NIST Assigned Value: (Median + NIST)/2
NAU	NIST Assigned Uncertainty: $\sqrt{(S^2 + S_{btw}^2)}$ S is the maximum of (0.05*NAV, eSD, $SD_{LT}$ ) and $S_{btw}$ is the standard deviation between Median and Mean <sub>NIST</sub> . The expected long-term SD, $SD_{LT}$ , is defined in: Duewer et al., Anal Chem 1997;69(7):1406-1413.

<sup>a</sup> What is now "NIST" was the National Bureau of Standards (NBS) when this study was conducted.

# Round Robin XI Laboratory Results

## Comparability Summary

Lab	TR	aT	bC	Label	Definition
FSV-BA	2	2	1	Lab	Participant code
FSV-BD	1	1		TR	Total Retinol
FSV-BE	1	1		aT	$\alpha$ -Tocopherol
FSV-BF	2	2	1	bC	$\beta$ -Carotene
FSV-BG	1	4	1	n	Number of values
FSV-BH	1	1	2		
FSV-BI	1	1	1	n	number of participants providing quantitative data
FSV-BX	3	1	4	% 1	Percent of CS = 1 (within 1 SD of medians)
FSV-BY	1		1	% 2	Percent of CS = 2 (within 2 SD of medians)
FSV-BZ	2	3		% 3	Percent of CS = 3 (within 3 SD of medians)
FSV-CJ	2	1	1	% 4	Percent of CS = 4 (3 or more SD from medians)
FSV-CK	1	1	1		
FSV-CN	1	1	1		
FSV-CO	1	1	1		
FSV-DC	1		2		
FSV-DE		1			The Comparability Score (CS) summarizes your measurement performance for a given analyte relative to the consensus medians in this study. CS is the average distance (in units of standard deviation) of your measurement performance characteristics from the consensus performance. CS is calculated when the number of quantitative values you reported, $N_{you}$ , is at least two and at least six participants reported quantitative values for the analyte.
FSV-DG	1	1	1		
FSV-DH	1	3			
FSV-DN	3	4	4		
FSV-DO	4				
FSV-DT	1	1			
FSV-ED	1	1	2		
FSV-EG	3		2		
FSV-EO	2		1		
FSV-EP	1	1			
FSV-ER	1	1	2		
FSV-EU	1	3			
FSV-EW	1	1	1		
FSV-FI		1			
NIST	1	1	1		
	n	28	25	20	

	TR	aT	bC
% 1	68	72	65
% 2	18	8	25
% 3	11	12	0
% 4	4	8	10

For further details, please see

Duewer DL, Kline MC, Sharpless KE, Brown Thomas J, Gary KT. Micronutrients Measurement Quality Assurance Program: Helping participants use interlaboratory comparison exercise results to improve their long-term measurement performance. Anal Chem 1999;71(9):1870-8.