



**NIST Interagency Report
NISTIR 8609-01**

NIST Time and Frequency Bulletin

Kelsey Rodriguez, Editor

This publication is available free of charge from:
<https://doi.org/10.6028/NIST.IR.8609-01>

**NIST Interagency Report
NISTIR 8609-01**

NIST Time and Frequency Bulletin

Kelsey Rodriguez, Editor
*Time and Frequency Division / HQ
Physical Measurement Laboratory*

This publication is available free of charge from:
<https://doi.org/10.6028/NIST.IR.8609-01>

January 2026
No. 817



U.S. Department of Commerce
Howard Lutnick, Secretary

National Institute of Standards and Technology
Craig Burkhardt, Acting NIST Director and Under Secretary of Commerce for Standards and Technology

NISTIR 8609-01
January 2026

Certain equipment, instruments, software, or materials, commercial or non-commercial, are identified in this paper in order to specify the experimental procedure adequately. Such identification does not imply recommendation or endorsement of any product or service by NIST, nor does it imply that the materials or equipment identified are necessarily the best available for the purpose.

NIST Technical Series Policies

[Copyright, Use, and Licensing Statements](#)

[NIST Technical Series Publication Identifier Syntax](#)

Publication History

Approved by the NIST Editorial Review Board on 2026-02-04

How to Cite this NIST Technical Series Publication

Rodriguez, K. (2026) NIST Time and Frequency Bulletin. (National Institute of Standards and Technology, Boulder, CO), NIST Interagency Report (NISTIR) 8609-01. <https://doi.org/10.6028/NIST.IR.8609-01>

Author ORCID iD

Kelsey Rodriguez: 0000-0001-8574-1155

Contact Information

This bulletin is published monthly. Address correspondence to:

Kelsey Rodriguez, Editor

Time and Frequency Division

National Institute of Standards and Technology

325 Broadway St

Boulder, CO 80305

(303) 497 5398

kelsey.rodriguez@nist.gov

Abstract

The Time and Frequency Bulletin provides information on performance of time scales and a variety of broadcasts (and related information) to users of the NIST services.

Keywords

Clocks; dissemination; frequency; GPS; oscillators; time.

Table of Contents

1. Time Scale Information.....	1
2. Broadcast Outages Over Five Minutes and WWVB Phase Perturbations.....	3
3. Notes on NIST Time Scales and Primary Standards	3
4. UTC NIST – AT1 Parameters	3
References.....	5
Appendix A. List of Symbols, Abbreviations, and Acronyms.....	6

List of Tables

Table 1. Variation in UT1 – UTC(NIST) and UTC(USNO, MC) – UTC(NIST) Time Scales.	1
Table 2. Corrections made to DUT1.....	1
Table 3. UTC – UTC(NIST).....	2
Table 4. Broadcast Outages and Phase Perturbations.	3
Table 5. UTC(NIST) - AT1.....	4

1. Time Scale Information

The values listed in the table below are based on data from the IERS, the USNO, and NIST. The UTC(USNO, MC) – UTC(NIST) values are obtained from the BIPM. UTC – UTC(NIST) data are on page 2.

Table 1. Variation in UT1 – UTC(NIST) and UTC(USNO, MC) – UTC(NIST) Time Scales.

0000 HOURS COORDINATED UNIVERSAL TIME			
December 2025	MJD	UT1 – UTC(NIST) (±1 ms)	UTC(USNO, MC) – UTC(NIST) (±5 ns)
5	61014	77.4 ms	0.1 ns
10	61019	77.0 ms	-0.3 ns
15	61024	74.9 ms	-0.2 ns
20	61029	76.2 ms	-0.1 ns
25	61034	77.4 ms	1.8 ns
30	61039	74.4 ms	3.2 ns

The clock pulses used by the WWV, WWVH, and WWVB time-code transmissions are referenced to the UTC (NIST) time scale. Occasionally, 1 s is added to the UTC time scale. This second is called a leap second. Its purpose is to keep the UTC time scale within ±0.9 s of the UT1 astronomical time scale, which changes slightly due to variations in Earth's period of rotation.

Positive leap seconds, beginning at 23 h 59 min 60 s UTC and ending at 0 h 0 min 0 s UTC, were inserted in the UTC time scale on 30 June 1972, 1981, 1983, 1985, 1992-1994, 1997, 2012, 2015 and on 31 December 1972-1979, 1987, 1989, 1990, 1995, 1998, 2005, 2008, 2016.

NO leap second will be introduced at the end of June 2026.

The insertion of leap seconds ensures that UT1 - UTC will always be held within ±0.9 s. The current value of UT1 – UTC is called the DUT1 correction. DUT1 corrections are broadcast by WWV, WWVH, WWVB, and ACTS and are printed below. These corrections may be added to the received UTC time signals in order to obtain UT1.

Table 2. Corrections made to DUT1.

DUT1 = UT1 – UTC =	+0.1 s beginning 0000 UTC 10 July 2025
	+0.0 s beginning 0000 UTC 26 December 2024
	+0.1 s beginning 0000 UTC 05 September 2024
	+0.0 s beginning 0000 UTC 28 July 2022
	-0.1 s beginning 0000 UTC 17 July 2021
	-0.2 s beginning 0000 UTC 02 May 2019
	-0.1 s beginning 0000 UTC 17 January 2019
	+0.0 s beginning 0000 UTC 21 September 2018
	+0.1 s beginning 0000 UTC 15 March 2018
	+0.2 s beginning 0000 UTC 30 November 2017
	+0.3 s beginning 0000 UTC 29 June 2017
	+0.4 s beginning 0000 UTC 30 March 2017
	+0.5 s beginning 0000 UTC 26 January 2017

The difference between UTC(NIST) and UTC has been within ± 100 ns since July 6, 1994. The table below shows values of UTC - UTC(NIST) as supplied by the BIPM in their *Circular T* publication for recent periods where data are available. Data are given at ten-day intervals. Five-day interval data are available in *Circular T*.

Table 3. UTC – UTC(NIST).

0000 Hours Coordinated Universal Time		
DATE	MJD	UTC-UTC(NIST), ns
Dec. 30, 2025	61039	2.3
Dec. 20, 2025	61029	-0.2
Dec. 10, 2025	61019	-0.2
Nov. 30, 2025	61009	-0.4
Nov. 20, 2025	60999	-0.0
Nov. 10, 2025	60989	0.6
Oct. 31, 2025	60979	0.6
Oct. 21, 2025	60969	1.8
Oct. 11, 2025	60959	1.9
Oct. 1, 2025	60949	1.0
Sep. 21, 2025	60939	0.6
Sep. 11, 2025	60929	0.2
Sep. 1, 2025	60919	-0.4
Aug. 22, 2025	60909	1.7
Aug. 12, 2025	60899	1.3
Aug. 2, 2025	60889	-0.2
Jul. 23, 2025	60879	-0.3
Jul. 13, 2025	60869	0.3
Jul. 3, 2025	60859	-0.2
Jun 23, 2025	60849	-1.1
Jun 13, 2025	60839	-1.5
Jun 3, 2025	60829	-0.9
May 24, 2025	60819	-0.0
May 14, 2025	60809	0.7
May 4, 2025	60799	-0.1
Apr. 24, 2025	60789	-0.6

2. Broadcast Outages Over Five Minutes and WWVB Phase Perturbations

Table 4. Broadcast Outages and Phase Perturbations.

OUTAGES OF 5 MINUTES OR MORE						PHASE PERTURBATIONS 2 ms			
Station	Dec 2025	MJD	Began UTC	Ended UTC	Freq.	Dec 2025	MJD	Began UTC	End UTC
WWVB	None					None			
WWV	None					None			
WWVH	None					None			

3. Notes on NIST Time Scales and Primary Standards

Primary frequency standards developed and operated by NIST are used to provide accuracy (rate) input to the BIPM and to provide the best possible realization of the SI second. Cold-atom cesium fountain frequency standards, currently NIST-F3 and NIST-F4, have served as the U.S. primary standards of time and frequency since 1999. The uncertainty of the primary standards is currently parts in 10^{16} .

The AT1 scale is run in real-time by use of data from an ensemble of cesium standards and hydrogen masers. It is a free-running scale whose frequency is maintained as nearly constant as possible by choosing the optimum weight for each clock that contributes to the computation.

UTC is generated at the BIPM by use of a post-processed time-scale algorithm and is not available in real-time. The parameters that we use to generate UTC(NIST) in real-time are therefore based on an extrapolation of UTC from the most recent available data.

UTC(NIST) is generated as an offset from our real-time scale AT1. Time steps are never used. Instead, the frequency is steered so that the time output remains close to UTC. This is accomplished by using data published by the BIPM in its *Circular T* and by weekly estimates of UTC, which are published by the BIPM as rapid UTC or UTCr. Changes in the frequency may be made as often as once per week and are limited to $\pm 2.3 \times 10^{-14}$. The frequency of UTC(NIST) is kept as stable as possible at other times.

4. UTC NIST – AT1 Parameters

Table 5 lists parameters that are used to define UTC(NIST) with respect to our real-time scale AT1. To find the value of UTC(NIST) – AT1 at any time T (expressed as a Modified Julian Date, including a fraction if needed), the appropriate equation to use is the one for which the desired T is greater than or equal to the entry in the T_0 column and less than the entry in the last column.

The values of x_{ls} , x , and y for that month are then used in the equation below to find the desired value. The parameters x and y represent the offsets in time and frequency, respectively,

between UTC(NIST) and AT1; the parameter x_{ls} is the number of leap seconds applied to both UTC(NIST) and UTC, as specified by the IERS. Leap seconds are not applied to AT1.

Table 5. UTC(NIST) - AT1.

UTC(NIST) - AT1 = $x_{ls} + x + y(T - T_0)$					
Month	x_{ls} (s)	x (ns)	y (ns/d)	T_0 (MJD)	Valid until 0000 on: (MJD)
Dec 25	-37	-562262.46	-38.63	61009	61040
Nov 25	-37	-562262.46	-38.63	61006	61009
Nov 25	-37	-561453.33	-38.53	60985	61006
Oct 25	-37	-561183.27	-38.58	60978	60985
Oct 25	-37	-560644.13	-38.51	60964	60978
Oct 25	-37	-560373.86	-38.61	60957	60964
Sep 25	-37	-559832.62	-38.66	60943	60957
Sep 25	-37	-559021.81	-38.61	60922	60943
Aug 25	-37	-558484.49	-38.38†	60908	60922
Aug 25	-37	-558214.78	-38.53	60901	60908
Jul 25	-37	-557672.56	-38.73†	60887	60901
Jul 25	-37	-556862.38	-38.58	60866	60887
Jun 25	-37	-555780.74	-38.63†	60838	60866
Jun 25	-37	-555356.36	-38.58	60827	60838
May 25	-37	-554469.02	-38.58†	60804	60827
May 25	-37	-554159.98	-38.63	60796	60804
Apr 25	-37	-553001.08	-38.63	60766	60796
Mar 25	-37	-551803.55	-38.63	60735	60766
Feb 25	-37	-550721.91	-38.63	60707	60735
Jan 25	-37	-549563.01	-38.63†	60677	60707
Jan 25	-37	-549524.23	-38.78	60676	60677
Dec 24	-37	-549020.09	-38.78†	60663	60676
Dec 24	-37	-548826.69	-38.68†	60658	60676
Dec 24	-37	-548480.82	-38.43†	60649	60658
Dec 24	-37	-548327.50	-38.33	60645	60649
Nov 24	-37	-548212.51	-38.33†	60642	60645
Nov 24	-37	-547942.80	-38.53†	60635	60642
Nov 24	-37	-547170.20	-38.63	60615	60635
Oct 24	-37	-547131.57	-38.63†	60614	60615
Oct 24	-37	-546587.25	-38.88†	60600	60614
Oct 24	-37	-546316.49	-38.68†	60592	60600
Oct 24	-37	-545969.72	-38.53	60584	60592

† Rate change in mid-month

* Provisional value

References

- [1] Levine, J (2012) The statistical modeling of atomic clocks and the design of time scales. *Review of Scientific Instruments*, 83: 0211101/1-28 (<https://dx.doi.org/10.1063/1.3681448>).
- [2] Parker, T; Jefferts, S; Heavner, T; and Donley, E (2005) Operation of the NIST-F1 caesium fountain primary frequency standard with a maser ensemble, including the impact of frequency transfer noise. *Metrologia* 42: 423-430 (<https://doi.org/10.1088/0026-1394/42/5/013>).
- [3] Sherman, J; Arissian, L; Brown, R; Deutch, M; Donley, E; Gerginov, V; Levine, J; Nelson, G; Novick, A; Patla, B; Parker, T; Stuhl, B; Sutton, D; Yao, J; Yates, W; Zhang, B; and Lombardi, M (2021) A Resilient Architecture for the Realization and Distribution of Coordinated Universal Time to Critical Infrastructure Systems in the United States: Methodologies and Recommendations from the National Institute of Standards and Technology (NIST). *NIST Technical Note 2187*, 189 p. (<https://doi.org/10.6028/NIST.TN.2187>).

Appendix A. List of Symbols, Abbreviations, and Acronyms

ACTS

Automated Computer Time Service

BIPM

Bureau International des Poids et Mesures

GPS

Global Positioning System

IERS

International Earth Rotation Service

MC

Master Clock

min

minute

MJD

Modified Julian Date

ms

Millisecond

NIST

National Institute of Standards and Technology

ns

Nanosecond

SI

International System of Units

TA

Atomic Time

TAI

International Atomic Time

s

Second

USNO

United States Naval Observatory

UT1

Universal Time (Astronomical)

UTC

Coordinated Universal Time

μs

Microsecond